

Continuing Calibration Report for SequenceID : FD050225AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049309.D

Aromatic C10-C12	5512093.000	40.000	4.431	6.176	137802.325	128531.556	-7.213
Aromatic C12-C16	8771780.000	60.000	6.177	8.805	146196.333	135727.245	-7.713
Aromatic C16-C21	11059605.000	80.000	8.806	13.095	138245.063	131120.372	-5.434
Aromatic C21-C36	21342567.000	180.000	13.096	18.516	118569.817	117353.852	-1.036
Aromatic EPH	46686045.000	360.000	4.431	18.516	129683.458	124717.278	-3.982