

Continuing Calibration Report for SequenceID : FD050225AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049315.D

Aromatic C10-C12	5180419.000	40.000	4.431	6.176	129510.475	128531.556	-0.762
Aromatic C12-C16	8441515.000	60.000	6.177	8.805	140691.917	135727.245	-3.658
Aromatic C16-C21	10922703.000	80.000	8.806	13.095	136533.788	131120.372	-4.129
Aromatic C21-C36	21099522.000	180.000	13.096	18.516	117219.567	117353.852	0.114
Aromatic EPH	45644159.000	360.000	4.431	18.516	126789.331	124717.278	-1.661