

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q1901 **SAS No :** Q1901 **SDG No:** Q1901
Sample No : PB167840BS **Datafile:** FC068744.D
Client ID : PB167840BS

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aliphatic C9-C12	300	185	62		(40-140)
Aliphatic C12-C16	200	178	89		(40-140)
Aliphatic C16-C21	300	318	106		(40-140)
Aliphatic C21-C28	400	426	106		(40-140)
Aliphatic C28-C40	600	637	106		(40-140)