

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q1901 **SAS No :** Q1901 **SDG No:** Q1901
Sample No : PB167840BSD **Datafile:** FD049312.D
Client ID : PB167840BSD

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aromatic C10-C12	200	163	82		3.8 (40-140)	50
Aromatic C12-C16	300	289	96		3.5 (40-140)	50
Aromatic C16-C21	500	624	124		4.8 (40-140)	50
Aromatic C21-C36	800	1040	129		5.1 (40-140)	50