

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q1901 **SAS No :** Q1901 **SDG No:** Q1901
Sample No : PB167903BS **Datafile:** FD049320.D
Client ID : PB167903BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aromatic C10-C12	6.7	5.23	78		(40-140)
Aromatic C12-C16	10	9.09	91		(40-140)
Aromatic C16-C21	16.7	18.9	113		(40-140)
Aromatic C21-C36	26.6	31.4	117		(40-140)