

RETENTION TIMES OF INITIAL CALIBRATION

Contract: PORT06

Lab Code: CHEM **Case No.:** Q1901 **SAS No.:** Q1901 **SDG NO.:** Q1901

Instrument ID: ECD_P **Calibration Date(s):** 04/22/2025 04/22/2025

Calibration Times: 10:29 17:49

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID: **RT 1000 =** PP071389.D **RT 750 =** PP071390.D

RT 500 = PP071391.D **RT 250 =** PP071392.D **RT 050 =** PP071393.D

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Aroclor-1016-1	(1)	5.66	5.67	5.67	5.67	5.67	5.67	5.57	5.77
Aroclor-1016-2	(2)	5.69	5.69	5.69	5.69	5.69	5.69	5.59	5.79
Aroclor-1016-3	(3)	5.75	5.75	5.75	5.75	5.75	5.75	5.65	5.85
Aroclor-1016-4	(4)	5.85	5.85	5.85	5.85	5.85	5.85	5.75	5.95
Aroclor-1016-5	(5)	6.14	6.14	6.15	6.14	6.14	6.14	6.04	6.24
Aroclor-1260-1	(1)	7.26	7.26	7.27	7.26	7.26	7.26	7.16	7.36
Aroclor-1260-2	(2)	7.51	7.51	7.52	7.52	7.52	7.52	7.42	7.62
Aroclor-1260-3	(3)	7.87	7.87	7.88	7.87	7.88	7.87	7.77	7.97
Aroclor-1260-4	(4)	8.10	8.10	8.10	8.10	8.10	8.10	8.00	8.20
Aroclor-1260-5	(5)	8.42	8.42	8.42	8.42	8.42	8.42	8.32	8.52
Decachlorobiphenyl		10.23	10.23	10.24	10.23	10.24	10.23	10.13	10.33
Tetrachloro-m-xylene		4.51	4.51	4.52	4.51	4.52	4.51	4.41	4.61
Aroclor-1242-1	(1)	5.67	5.67	5.67	5.67	5.67	5.67	5.57	5.77
Aroclor-1242-2	(2)	5.69	5.69	5.69	5.69	5.69	5.69	5.59	5.79
Aroclor-1242-3	(3)	5.75	5.75	5.75	5.75	5.75	5.75	5.65	5.85
Aroclor-1242-4	(4)	5.85	5.85	5.85	5.85	5.85	5.85	5.75	5.95
Aroclor-1242-5	(5)	6.58	6.58	6.58	6.58	6.58	6.58	6.48	6.68
Decachlorobiphenyl		10.23	10.24	10.24	10.23	10.24	10.24	10.14	10.34
Tetrachloro-m-xylene		4.51	4.52	4.51	4.51	4.52	4.51	4.41	4.61
Aroclor-1248-1	(1)	5.67	5.67	5.67	5.67	5.67	5.67	5.57	5.77
Aroclor-1248-2	(2)	5.94	5.94	5.94	5.94	5.94	5.94	5.84	6.04
Aroclor-1248-3	(3)	6.14	6.14	6.14	6.14	6.14	6.14	6.04	6.24
Aroclor-1248-4	(4)	6.54	6.54	6.54	6.54	6.54	6.54	6.44	6.64
Aroclor-1248-5	(5)	6.58	6.58	6.58	6.58	6.58	6.58	6.48	6.68
Decachlorobiphenyl		10.24	10.24	10.24	10.24	10.24	10.24	10.14	10.34
Tetrachloro-m-xylene		4.52	4.52	4.51	4.51	4.51	4.51	4.41	4.61
Aroclor-1254-1	(1)	6.52	6.52	6.52	6.52	6.52	6.52	6.42	6.62
Aroclor-1254-2	(2)	6.74	6.74	6.74	6.73	6.74	6.74	6.64	6.84
Aroclor-1254-3	(3)	7.10	7.10	7.10	7.10	7.10	7.10	7.00	7.20
Aroclor-1254-4	(4)	7.38	7.38	7.38	7.38	7.38	7.38	7.28	7.48
Aroclor-1254-5	(5)	7.80	7.80	7.80	7.80	7.80	7.80	7.70	7.90
Decachlorobiphenyl		10.24	10.24	10.24	10.23	10.24	10.24	10.14	10.34
Tetrachloro-m-xylene		4.52	4.52	4.52	4.51	4.51	4.51	4.41	4.61
Aroclor-1268-1	(1)	8.73	8.73	8.73	8.74	8.73	8.73	8.63	8.83
Aroclor-1268-2	(2)	8.83	8.83	8.83	8.83	8.83	8.83	8.73	8.93
Aroclor-1268-3	(3)	9.06	9.06	9.06	9.06	9.06	9.06	8.96	9.16
Aroclor-1268-4	(4)	9.48	9.48	9.48	9.48	9.48	9.48	9.38	9.58
Aroclor-1268-5	(5)	9.89	9.89	9.89	9.90	9.89	9.90	9.80	10.00

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	10.24	10.23	10.24	10.24	10.24	10.24	10.14	10.34
Tetrachloro-m-xylene	4.51	4.52	4.52	4.52	4.51	4.52	4.42	4.62

RETENTION TIMES OF INITIAL CALIBRATION

Contract:	<u>PORT06</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1901</u>	SAS No.:	<u>Q1901</u>	SDG NO.:	<u>Q1901</u>
Instrument ID:	<u>ECD_P</u>	Calibration Date(s):		<u>04/22/2025</u>	<u>04/22/2025</u>		
		Calibration Times:		<u>10:29</u>	<u>17:49</u>		

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PP071389.D RT 750 = PP071390.D
RT 500 = PP071391.D RT 250 = PP071392.D RT 050 = PP071393.D

[illegible]

[illegible]