

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL01 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO110846.D Time Analyzed: 09:24

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|----------------------|-------|--------------------|-------------------|------|
| Aroclor-1016-1       | 4.775 | 4.679                | 4.879 | 554.440            | 500.000           | 10.9 |
| Aroclor-1016-2       | 4.795 | 4.698                | 4.898 | 558.040            | 500.000           | 11.6 |
| Aroclor-1016-3       | 4.851 | 4.755                | 4.955 | 542.460            | 500.000           | 8.5  |
| Aroclor-1016-4       | 4.971 | 4.875                | 5.075 | 553.630            | 500.000           | 10.7 |
| Aroclor-1016-5       | 5.228 | 5.132                | 5.332 | 543.180            | 500.000           | 8.6  |
| Aroclor-1260-1       | 6.268 | 6.172                | 6.372 | 535.720            | 500.000           | 7.1  |
| Aroclor-1260-2       | 6.457 | 6.360                | 6.560 | 519.830            | 500.000           | 4.0  |
| Aroclor-1260-3       | 6.824 | 6.729                | 6.929 | 522.530            | 500.000           | 4.5  |
| Aroclor-1260-4       | 7.085 | 6.988                | 7.188 | 513.390            | 500.000           | 2.7  |
| Aroclor-1260-5       | 7.326 | 7.230                | 7.430 | 515.480            | 500.000           | 3.1  |
| Decachlorobiphenyl   | 8.727 | 8.632                | 8.832 | 47.400             | 50.000            | -5.2 |
| Tetrachloro-m-xylene | 3.686 | 3.588                | 3.788 | 56.210             | 50.000            | 12.4 |

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL01 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO110846.D Time Analyzed: 09:24

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.762 | 4.666     | 4.866 | 520.260            | 500.000           | 4.1  |
| Aroclor-1016-2       | 4.781 | 4.684     | 4.884 | 519.530            | 500.000           | 3.9  |
| Aroclor-1016-3       | 4.956 | 4.860     | 5.060 | 502.490            | 500.000           | 0.5  |
| Aroclor-1016-4       | 4.998 | 4.902     | 5.102 | 493.190            | 500.000           | -1.4 |
| Aroclor-1016-5       | 5.211 | 5.115     | 5.315 | 490.700            | 500.000           | -1.9 |
| Aroclor-1260-1       | 6.242 | 6.146     | 6.346 | 492.670            | 500.000           | -1.5 |
| Aroclor-1260-2       | 6.430 | 6.334     | 6.534 | 484.910            | 500.000           | -3.0 |
| Aroclor-1260-3       | 6.582 | 6.487     | 6.687 | 480.540            | 500.000           | -3.9 |
| Aroclor-1260-4       | 7.053 | 6.957     | 7.157 | 480.890            | 500.000           | -3.8 |
| Aroclor-1260-5       | 7.294 | 7.197     | 7.397 | 482.240            | 500.000           | -3.6 |
| Decachlorobiphenyl   | 8.679 | 8.584     | 8.784 | 46.500             | 50.000            | -7.0 |
| Tetrachloro-m-xylene | 3.683 | 3.586     | 3.786 | 53.130             | 50.000            | 6.3  |