

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL06 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PP071614.D Time Analyzed: 16:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.665	5.571	5.771	519.380	500.000	3.9
Aroclor-1016-2	5.688	5.593	5.793	509.990	500.000	2.0
Aroclor-1016-3	5.750	5.654	5.854	509.200	500.000	1.8
Aroclor-1016-4	5.847	5.752	5.952	513.170	500.000	2.6
Aroclor-1016-5	6.140	6.045	6.245	520.600	500.000	4.1
Aroclor-1260-1	7.259	7.165	7.365	536.190	500.000	7.2
Aroclor-1260-2	7.512	7.418	7.618	518.580	500.000	3.7
Aroclor-1260-3	7.870	7.777	7.977	532.700	500.000	6.5
Aroclor-1260-4	8.095	8.002	8.202	515.410	500.000	3.1
Aroclor-1260-5	8.415	8.322	8.522	534.510	500.000	6.9
Decachlorobiphenyl	10.230	10.137	10.337	54.370	50.000	8.7
Tetrachloro-m-xylene	4.512	4.417	4.617	51.110	50.000	2.2

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL06 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PP071614.D Time Analyzed: 16:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.895	4.798	4.998	517.640	500.000	3.5
Aroclor-1016-2	4.914	4.816	5.016	516.230	500.000	3.2
Aroclor-1016-3	5.090	4.993	5.193	534.980	500.000	7.0
Aroclor-1016-4	5.132	5.036	5.236	540.310	500.000	8.1
Aroclor-1016-5	5.347	5.250	5.450	579.870	500.000	16.0
Aroclor-1260-1	6.381	6.285	6.485	542.610	500.000	8.5
Aroclor-1260-2	6.569	6.473	6.673	540.120	500.000	8.0
Aroclor-1260-3	6.722	6.627	6.827	523.900	500.000	4.8
Aroclor-1260-4	7.193	7.098	7.298	538.880	500.000	7.8
Aroclor-1260-5	7.435	7.339	7.539	527.470	500.000	5.5
Decachlorobiphenyl	8.844	8.749	8.949	54.370	50.000	8.7
Tetrachloro-m-xylene	3.810	3.712	3.912	50.630	50.000	1.3