

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO041025

Review By	yogesh	Review On	4/10/2025 11:27:35 AM
Supervise By	mohammad	Supervise On	4/14/2025 12:40:29 AM
SubDirectory	PO041025	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO110347.D	10 Apr 2025 08:59	YP/AJ	Ok
2	I.BLK	PO110348.D	10 Apr 2025 09:17	YP/AJ	Ok
3	AR1660ICC1000	PO110349.D	10 Apr 2025 09:36	YP/AJ	Ok
4	AR1660ICC750	PO110350.D	10 Apr 2025 09:54	YP/AJ	Ok
5	AR1660ICC500	PO110351.D	10 Apr 2025 10:13	YP/AJ	Ok
6	AR1660ICC250	PO110352.D	10 Apr 2025 10:31	YP/AJ	Ok
7	AR1660ICC050	PO110353.D	10 Apr 2025 10:49	YP/AJ	Ok
8	AR1221ICC500	PO110354.D	10 Apr 2025 11:08	YP/AJ	Ok
9	AR1232ICC500	PO110355.D	10 Apr 2025 11:26	YP/AJ	Ok
10	AR1242ICC1000	PO110356.D	10 Apr 2025 11:44	YP/AJ	Ok
11	AR1242ICC750	PO110357.D	10 Apr 2025 12:03	YP/AJ	Ok
12	AR1242ICC500	PO110358.D	10 Apr 2025 12:21	YP/AJ	Ok
13	AR1242ICC250	PO110359.D	10 Apr 2025 12:39	YP/AJ	Ok
14	AR1242ICC050	PO110360.D	10 Apr 2025 12:58	YP/AJ	Ok
15	AR1248ICC1000	PO110361.D	10 Apr 2025 13:16	YP/AJ	Ok
16	AR1248ICC750	PO110362.D	10 Apr 2025 13:35	YP/AJ	Ok
17	AR1248ICC500	PO110363.D	10 Apr 2025 13:53	YP/AJ	Ok
18	AR1248ICC250	PO110364.D	10 Apr 2025 14:11	YP/AJ	Ok
19	AR1248ICC050	PO110365.D	10 Apr 2025 14:30	YP/AJ	Ok
20	AR1254ICC1000	PO110366.D	10 Apr 2025 14:48	YP/AJ	Ok
21	AR1254ICC750	PO110367.D	10 Apr 2025 15:06	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO041025

Review By	yogesh	Review On	4/10/2025 11:27:35 AM
Supervise By	mohammad	Supervise On	4/14/2025 12:40:29 AM
SubDirectory	PO041025	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PO110368.D	10 Apr 2025 15:25	YP/AJ	Ok
23	AR1254ICC250	PO110369.D	10 Apr 2025 15:43	YP/AJ	Ok
24	AR1254ICC050	PO110370.D	10 Apr 2025 16:02	YP/AJ	Ok
25	AR1262ICC500	PO110371.D	10 Apr 2025 16:20	YP/AJ	Ok
26	AR1268ICC1000	PO110372.D	10 Apr 2025 16:38	YP/AJ	Ok
27	AR1268ICC750	PO110373.D	10 Apr 2025 16:57	YP/AJ	Ok
28	AR1268ICC500	PO110374.D	10 Apr 2025 17:15	YP/AJ	Ok
29	AR1268ICC250	PO110375.D	10 Apr 2025 17:33	YP/AJ	Ok
30	AR1268ICC050	PO110376.D	10 Apr 2025 17:52	YP/AJ	Ok
31	PO041025ICV500	PO110377.D	10 Apr 2025 18:09	YP/AJ	Ok
32	AR1242ICV500	PO110378.D	10 Apr 2025 18:46	YP/AJ	Ok
33	AR1248ICV500	PO110379.D	10 Apr 2025 19:22	YP/AJ	Ok
34	AR1254ICV500	PO110380.D	10 Apr 2025 19:58	YP/AJ	Ok
35	AR1268ICV500	PO110381.D	10 Apr 2025 20:35	YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042925

Review By	yogesh	Review On	4/29/2025 11:07:42 AM
Supervise By	mohammad	Supervise On	5/2/2025 2:51:07 AM
SubDirectory	PO042925	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO110845.D	29 Apr 2025 09:06	YP/AJ	Ok
2	AR1660CCC500	PO110846.D	29 Apr 2025 09:24	YP/AJ	Ok
3	AR1242CCC500	PO110847.D	29 Apr 2025 09:45	YP/AJ	Ok
4	AR1248CCC500	PO110848.D	29 Apr 2025 10:03	YP/AJ	Ok
5	AR1254CCC500	PO110849.D	29 Apr 2025 10:21	YP/AJ	Ok
6	I.BLK	PO110850.D	29 Apr 2025 10:40	YP/AJ	Ok
7	PB167765BL	PO110851.D	29 Apr 2025 10:58	YP/AJ	Not Ok
8	Q1889-01	PO110852.D	29 Apr 2025 11:16	YP/AJ	Ok
9	Q1889-01MS	PO110853.D	29 Apr 2025 11:35	YP/AJ	Ok
10	Q1889-01MSD	PO110854.D	29 Apr 2025 11:53	YP/AJ	Ok
11	Q1860-34DL	PO110855.D	29 Apr 2025 12:11	YP/AJ	Ok,M
12	Q1901-05	PO110856.D	29 Apr 2025 13:36	YP/AJ	Ok
13	Q1902-02	PO110857.D	29 Apr 2025 13:54	YP/AJ	Ok,M
14	Q1903-01	PO110858.D	29 Apr 2025 14:12	YP/AJ	Ok
15	Q1903-02	PO110859.D	29 Apr 2025 14:31	YP/AJ	Ok
16	Q1903-03	PO110860.D	29 Apr 2025 14:49	YP/AJ	Ok
17	AR1660CCC500	PO110861.D	29 Apr 2025 15:58	YP/AJ	Ok
18	AR1242CCC500	PO110862.D	29 Apr 2025 16:35	YP/AJ	Ok
19	AR1248CCC500	PO110863.D	29 Apr 2025 16:53	YP/AJ	Ok
20	AR1254CCC500	PO110864.D	29 Apr 2025 17:11	YP/AJ	Ok
21	I.BLK	PO110865.D	29 Apr 2025 17:29	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042925

Review By	yogesh	Review On	4/29/2025 11:07:42 AM
Supervise By	mohammad	Supervise On	5/2/2025 2:51:07 AM
SubDirectory	PO042925	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q1904-01	PO110866.D	29 Apr 2025 17:48	YP/AJ	Ok
23	Q1904-01MS	PO110867.D	29 Apr 2025 18:05	YP/AJ	Ok
24	Q1904-01MSD	PO110868.D	29 Apr 2025 18:23	YP/AJ	Ok
25	Q1906-01	PO110869.D	29 Apr 2025 18:42	YP/AJ	Ok
26	Q1906-05	PO110870.D	29 Apr 2025 19:00	YP/AJ	Ok
27	Q1906-09	PO110871.D	29 Apr 2025 19:19	YP/AJ	Ok,M
28	Q1906-13	PO110872.D	29 Apr 2025 19:37	YP/AJ	Ok
29	Q1897-01	PO110873.D	29 Apr 2025 19:55	YP/AJ	Ok,M
30	Q1901-06	PO110874.D	29 Apr 2025 20:14	YP/AJ	Ok
31	Q1897-02	PO110875.D	29 Apr 2025 20:32	YP/AJ	Ok,M
32	AR1660CCC500	PO110876.D	29 Apr 2025 22:00	YP/AJ	Ok
33	AR1242CCC500	PO110877.D	29 Apr 2025 22:36	YP/AJ	Ok
34	AR1248CCC500	PO110878.D	29 Apr 2025 22:55	YP/AJ	Ok
35	AR1254CCC500	PO110879.D	29 Apr 2025 23:13	YP/AJ	Ok
36	I.BLK	PO110880.D	29 Apr 2025 23:31	YP/AJ	Ok
37	PB167790BL	PO110881.D	29 Apr 2025 23:50	YP/AJ	Ok
38	PB167790BS	PO110882.D	30 Apr 2025 00:08	YP/AJ	Ok
39	Q1909-01	PO110883.D	30 Apr 2025 00:26	YP/AJ	Ok,M
40	Q1909-02	PO110884.D	30 Apr 2025 00:45	YP/AJ	Ok,M
41	AR1660CCC500	PO110885.D	30 Apr 2025 02:11	YP/AJ	Ok
42	AR1242CCC500	PO110886.D	30 Apr 2025 02:48	YP/AJ	Ok
43	AR1248CCC500	PO110887.D	30 Apr 2025 03:06	YP/AJ	Ok
44	AR1254CCC500	PO110888.D	30 Apr 2025 03:25	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042925

Review By	yogesh	Review On	4/29/2025 11:07:42 AM
Supervise By	mohammad	Supervise On	5/2/2025 2:51:07 AM
SubDirectory	PO042925	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	I.BLK	PO110889.D	30 Apr 2025 03:43	YP/AJ	Ok
----	-------	------------	-------------------	-------	----

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP071387.D	22 Apr 2025 09:56	YP\AJ	Ok
2	I.BLK	PP071388.D	22 Apr 2025 10:13	YP\AJ	Ok
3	AR1660ICC1000	PP071389.D	22 Apr 2025 10:29	YP\AJ	Ok
4	AR1660ICC750	PP071390.D	22 Apr 2025 10:45	YP\AJ	Ok
5	AR1660ICC500	PP071391.D	22 Apr 2025 11:02	YP\AJ	Ok
6	AR1660ICC250	PP071392.D	22 Apr 2025 11:18	YP\AJ	Ok
7	AR1660ICC050	PP071393.D	22 Apr 2025 11:34	YP\AJ	Ok,M
8	AR1221ICC500	PP071394.D	22 Apr 2025 11:51	YP\AJ	Ok
9	AR1232ICC500	PP071395.D	22 Apr 2025 12:07	YP\AJ	Ok
10	AR1242ICC1000	PP071396.D	22 Apr 2025 12:23	YP\AJ	Ok
11	AR1242ICC750	PP071397.D	22 Apr 2025 12:39	YP\AJ	Ok
12	AR1242ICC500	PP071398.D	22 Apr 2025 12:56	YP\AJ	Ok
13	AR1242ICC250	PP071399.D	22 Apr 2025 13:12	YP\AJ	Ok
14	AR1242ICC050	PP071400.D	22 Apr 2025 13:28	YP\AJ	Ok
15	AR1248ICC1000	PP071401.D	22 Apr 2025 13:45	YP\AJ	Ok
16	AR1248ICC750	PP071402.D	22 Apr 2025 14:01	YP\AJ	Ok
17	AR1248ICC500	PP071403.D	22 Apr 2025 14:17	YP\AJ	Ok
18	AR1248ICC250	PP071404.D	22 Apr 2025 14:33	YP\AJ	Ok
19	AR1248ICC050	PP071405.D	22 Apr 2025 14:50	YP\AJ	Ok
20	AR1254ICC1000	PP071406.D	22 Apr 2025 15:06	YP\AJ	Ok
21	AR1254ICC750	PP071407.D	22 Apr 2025 15:22	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PP071408.D	22 Apr 2025 15:38	YP\AJ	Ok
23	AR1254ICC250	PP071409.D	22 Apr 2025 15:55	YP\AJ	Ok
24	AR1254ICC050	PP071410.D	22 Apr 2025 16:11	YP\AJ	Ok,M
25	AR1262ICC500	PP071411.D	22 Apr 2025 16:27	YP\AJ	Ok
26	AR1268ICC1000	PP071412.D	22 Apr 2025 16:44	YP\AJ	Ok,M
27	AR1268ICC750	PP071413.D	22 Apr 2025 17:00	YP\AJ	Ok,M
28	AR1268ICC500	PP071414.D	22 Apr 2025 17:16	YP\AJ	Ok,M
29	AR1268ICC250	PP071415.D	22 Apr 2025 17:33	YP\AJ	Ok,M
30	AR1268ICC050	PP071416.D	22 Apr 2025 17:49	YP\AJ	Ok,M
31	PP042225ICV500	PP071417.D	22 Apr 2025 18:05	YP\AJ	Ok
32	AR1242ICV500	PP071418.D	22 Apr 2025 18:38	YP\AJ	Ok
33	AR1248ICV500	PP071419.D	22 Apr 2025 19:27	YP\AJ	Ok
34	AR1254ICV500	PP071420.D	22 Apr 2025 19:43	YP\AJ	Ok
35	AR1268ICV500	PP071421.D	22 Apr 2025 20:16	YP\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042925

Review By	yogesh	Review On	4/29/2025 11:08:35 AM
Supervise By	mohammad	Supervise On	5/2/2025 2:54:00 AM
SubDirectory	PP042925	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP071598.D	29 Apr 2025 10:16	YP\AJ	Ok
2	AR1660CCC500	PP071599.D	29 Apr 2025 10:33	YP\AJ	Ok
3	AR1242CCC500	PP071600.D	29 Apr 2025 10:51	YP\AJ	Ok
4	AR1248CCC500	PP071601.D	29 Apr 2025 11:07	YP\AJ	Ok
5	AR1254CCC500	PP071602.D	29 Apr 2025 11:24	YP\AJ	Ok
6	I.BLK	PP071603.D	29 Apr 2025 11:40	YP\AJ	Ok
7	Q1860-35DL	PP071604.D	29 Apr 2025 11:56	YP\AJ	Ok,M
8	Q1884-01DL	PP071605.D	29 Apr 2025 12:12	YP\AJ	Ok,M
9	Q1884-02DL	PP071606.D	29 Apr 2025 12:29	YP\AJ	Ok,M
10	PB167776BL	PP071607.D	29 Apr 2025 13:30	YP\AJ	Ok
11	PB167776BS	PP071608.D	29 Apr 2025 13:46	YP\AJ	Ok
12	Q1883-01	PP071609.D	29 Apr 2025 14:02	YP\AJ	Ok,M
13	Q1883-03	PP071610.D	29 Apr 2025 14:19	YP\AJ	Ok
14	Q1883-05	PP071611.D	29 Apr 2025 14:35	YP\AJ	Ok
15	Q1883-07	PP071612.D	29 Apr 2025 14:51	YP\AJ	Ok
16	Q1883-09	PP071613.D	29 Apr 2025 15:08	YP\AJ	Ok
17	AR1660CCC500	PP071614.D	29 Apr 2025 16:29	YP\AJ	Ok
18	AR1242CCC500	PP071615.D	29 Apr 2025 17:02	YP\AJ	Ok
19	AR1248CCC500	PP071616.D	29 Apr 2025 17:18	YP\AJ	Ok
20	AR1254CCC500	PP071617.D	29 Apr 2025 17:34	YP\AJ	Ok
21	I.BLK	PP071618.D	29 Apr 2025 17:51	YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042925

Review By	yogesh	Review On	4/29/2025 11:08:35 AM
Supervise By	mohammad	Supervise On	5/2/2025 2:54:00 AM
SubDirectory	PP042925	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q1883-11	PP071619.D	29 Apr 2025 18:07	YP\AJ	Ok
23	Q1883-13	PP071620.D	29 Apr 2025 18:23	YP\AJ	Ok
24	Q1883-15	PP071621.D	29 Apr 2025 18:40	YP\AJ	Ok
25	Q1901-02	PP071622.D	29 Apr 2025 18:56	YP\AJ	Ok,M
26	Q1901-04	PP071623.D	29 Apr 2025 19:12	YP\AJ	Ok
27	PB167782BL	PP071624.D	29 Apr 2025 19:29	YP\AJ	Ok
28	PB167782BS	PP071625.D	29 Apr 2025 19:45	YP\AJ	Ok
29	PB167782BSD	PP071626.D	29 Apr 2025 20:01	YP\AJ	Ok
30	Q1890-01	PP071627.D	29 Apr 2025 20:18	YP\AJ	Ok
31	Q1896-03	PP071628.D	29 Apr 2025 20:34	YP\AJ	Ok,M
32	AR1660CCC500	PP071629.D	29 Apr 2025 21:55	YP\AJ	Ok
33	AR1242CCC500	PP071630.D	29 Apr 2025 22:28	YP\AJ	Ok
34	AR1248CCC500	PP071631.D	29 Apr 2025 22:44	YP\AJ	Ok
35	AR1254CCC500	PP071632.D	29 Apr 2025 23:01	YP\AJ	Ok,M
36	I.BLK	PP071633.D	29 Apr 2025 23:17	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP043025

Review By	yogesh	Review On	4/30/2025 11:33:57 AM
Supervise By	mohammad	Supervise On	5/2/2025 3:18:29 AM
SubDirectory	PP043025	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP071634.D	30 Apr 2025 08:51	YP\AJ	Ok
2	AR1660CCC500	PP071635.D	30 Apr 2025 09:07	YP\AJ	Ok
3	AR1242CCC500	PP071636.D	30 Apr 2025 09:28	YP\AJ	Ok
4	AR1248CCC500	PP071637.D	30 Apr 2025 09:45	YP\AJ	Ok
5	AR1254CCC500	PP071638.D	30 Apr 2025 10:01	YP\AJ	Ok,M
6	I.BLK	PP071639.D	30 Apr 2025 10:17	YP\AJ	Ok
7	Q1867-01DL	PP071640.D	30 Apr 2025 11:09	YP\AJ	Ok,M
8	PB167794BL	PP071641.D	30 Apr 2025 12:39	YP\AJ	Ok
9	PB167794BS	PP071642.D	30 Apr 2025 12:56	YP\AJ	Ok
10	Q1901-03	PP071643.D	30 Apr 2025 13:12	YP\AJ	Ok,M
11	Q1905-01	PP071644.D	30 Apr 2025 13:28	YP\AJ	Ok,M
12	Q1860-09	PP071645.D	30 Apr 2025 13:44	YP\AJ	Dilution
13	Q1860-09DL	PP071646.D	30 Apr 2025 14:01	YP\AJ	Ok,M
14	Q1910-01	PP071647.D	30 Apr 2025 14:33	YP\AJ	Ok,M
15	Q1910-01MS	PP071648.D	30 Apr 2025 14:49	YP\AJ	Ok,M
16	Q1910-01MSD	PP071649.D	30 Apr 2025 15:05	YP\AJ	Ok,M
17	AR1660CCC500	PP071650.D	30 Apr 2025 16:11	YP\AJ	Ok
18	AR1242CCC500	PP071651.D	30 Apr 2025 16:27	YP\AJ	Ok
19	AR1248CCC500	PP071652.D	30 Apr 2025 16:43	YP\AJ	Ok
20	AR1254CCC500	PP071653.D	30 Apr 2025 17:00	YP\AJ	Ok,M
21	I.BLK	PP071654.D	30 Apr 2025 17:16	YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP043025

Review By	yogesh	Review On	4/30/2025 11:33:57 AM
Supervise By	mohammad	Supervise On	5/2/2025 3:18:29 AM
SubDirectory	PP043025	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q1911-01	PP071655.D	30 Apr 2025 17:33	YPIAJ	Ok,M
23	Q1912-01	PP071656.D	30 Apr 2025 17:49	YPIAJ	Ok,M
24	Q1912-05	PP071657.D	30 Apr 2025 18:06	YPIAJ	Ok,M
25	Q1905-05	PP071658.D	30 Apr 2025 18:22	YPIAJ	Ok,M
26	Q1907-01	PP071659.D	30 Apr 2025 18:38	YPIAJ	Ok,M
27	Q1921-01	PP071660.D	30 Apr 2025 18:55	YPIAJ	Ok,M
28	Q1921-02	PP071661.D	30 Apr 2025 19:11	YPIAJ	Ok,M
29	Q1921-03	PP071662.D	30 Apr 2025 19:27	YPIAJ	Ok,M
30	Q1921-04	PP071663.D	30 Apr 2025 19:44	YPIAJ	Ok,M
31	Q1921-05	PP071664.D	30 Apr 2025 20:00	YPIAJ	Ok,M
32	AR1660CCC500	PP071665.D	30 Apr 2025 21:38	YPIAJ	Ok
33	AR1242CCC500	PP071666.D	30 Apr 2025 22:11	YPIAJ	Ok
34	AR1248CCC500	PP071667.D	30 Apr 2025 22:27	YPIAJ	Ok
35	AR1254CCC500	PP071668.D	30 Apr 2025 22:44	YPIAJ	Ok,M
36	I.BLK	PP071669.D	30 Apr 2025 23:00	YPIAJ	Ok,M
37	PB167809BL	PP071670.D	30 Apr 2025 23:16	YPIAJ	Ok
38	PB167809BS	PP071671.D	30 Apr 2025 23:33	YPIAJ	Ok
39	Q1921-06	PP071672.D	30 Apr 2025 23:49	YPIAJ	Ok
40	Q1921-07	PP071673.D	01 May 2025 00:05	YPIAJ	Ok,M
41	Q1921-08	PP071674.D	01 May 2025 00:22	YPIAJ	Ok,M
42	AR1660CCC500	PP071675.D	01 May 2025 02:00	YPIAJ	Ok
43	AR1242CCC500	PP071676.D	01 May 2025 02:49	YPIAJ	Ok,M
44	AR1248CCC500	PP071677.D	01 May 2025 03:05	YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP043025

Review By	yogesh	Review On	4/30/2025 11:33:57 AM
Supervise By	mohammad	Supervise On	5/2/2025 3:18:29 AM
SubDirectory	PP043025	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	AR1254CCC500	PP071678.D	01 May 2025 03:21	YP\AJ	Ok,M
46	I.BLK	PP071679.D	01 May 2025 03:38	YP\AJ	Ok,M

M : Manual Integration