



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: 8082A Datafile : PP071642.D

| Lab Sample ID | Parameter | Spike | Result | Units | Rec | RPD | Qual | RPD  |         | Limits    |     |
|---------------|-----------|-------|--------|-------|-----|-----|------|------|---------|-----------|-----|
|               |           |       |        |       |     |     |      | Qual | Low     | High      | RPD |
| PB167794BS    | AR1016    | 166.5 | 150    | ug/kg | 90  |     |      |      | 40 (71) | 140 (120) |     |
|               | AR1260    | 166.5 | 153    | ug/kg | 92  |     |      |      | 40 (65) | 140 (130) |     |