

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL02 Date Analyzed: 05/06/2025

Lab Sample No.: HSTDCCC750 Data File : PS030059.D Time Analyzed: 14:15

| COMPOUND         | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D  |
|------------------|-------|----------------------|-------|--------------------|-------------------|-----|
| 2,4,5-TP(Silvex) | 8.856 | 8.762                | 8.962 | 744.780            | 712.500           | 4.5 |
| 2,4-D            | 8.015 | 7.919                | 8.119 | 719.560            | 705.000           | 2.1 |
| 2,4-DCAA         | 6.938 | 6.842                | 7.042 | 763.780            | 750.000           | 1.8 |

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** PORT06

**Lab Code:** CHEM      **Case No.:** Q1901      **SAS No.:** Q1901      **SDG NO.:** Q1901

**GC Column:** RTX-CLP2      **ID:** 0.32 (mm)      **Initi. Calib. Date(s):** 04/23/2025      04/23/2025

**Client Sample No.:** CCAL02      **Date Analyzed:** 05/06/2025

**Lab Sample No.:** HSTDCCC750      **Data File :** PS030059.D      **Time Analyzed:** 14:15

| COMPOUND         | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D  |
|------------------|-------|-----------|-------|--------------------|-------------------|-----|
|                  |       | FROM      | TO    |                    |                   |     |
| 2,4,5-TP(Silvex) | 9.532 | 9.442     | 9.642 | 772.440            | 712.500           | 8.4 |
| 2,4-D            | 8.652 | 8.561     | 8.761 | 732.170            | 705.000           | 3.9 |
| 2,4-DCAA         | 7.461 | 7.367     | 7.567 | 793.080            | 750.000           | 5.7 |