

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL03 Date Analyzed: 05/06/2025

Lab Sample No.: HSTDCCC750 Data File : PS030070.D Time Analyzed: 18:40

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	8.855	8.762	8.962	756.820	712.500	6.2
2,4-D	8.014	7.919	8.119	729.360	705.000	3.5
2,4-DCAA	6.937	6.842	7.042	776.060	750.000	3.5

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL03 Date Analyzed: 05/06/2025

Lab Sample No.: HSTDCCC750 Data File : PS030070.D Time Analyzed: 18:40

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.531	9.442	9.642	791.940	712.500	11.1
2,4-D	8.653	8.561	8.761	755.570	705.000	7.2
2,4-DCAA	7.460	7.367	7.567	810.860	750.000	8.1