

Matrix:

TCLP



Lab Sample ID:

Report of Analysis

Client: Portal Partners Tri-Venture Date Collected: 04/23/25

Project: Amtrak Sawtooth Bridges 2025 Date Received: 04/23/25

Client Sample ID: PIBLK-PS029918.D SDG No.: Q1901

Analytical Method: SW8151A % Solid: 0 Decanted:

Sample Wt/Vol: 1000 Units: mL Final Vol: 10000 uL

Soil Aliquot Vol: uL Test: TCLP Herbicide

Extraction Type: Injection Volume:

GPC Factor: 1.0 PH:

I.BLK-PS029918.D

Prep Method: SW3510C

File ID/Qc Batch: Dilution: Prep Date Date Analyzed Prep Batch ID
PS029918.D 1 04/23/25 PS042325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						_
94-75-7	2,4-D	0.92	U	0.92	2.00	ug/L
93-72-1	2,4,5-TP (Silvex)	0.78	U	0.78	2.00	ug/L
SURROGATES	24 DG44	422		70 (20) 120 (175)	070/	CD1/ 500
	2,4,5-1P (Silvex) 2,4-DCAA	0.78	U	0.78 70 (39) - 130 (175)	2.00 87%	ug/L SPK:

Comments:

U = Not Detected

LOO = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit