

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: PS050625

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified			✓
2. PEM Meet Criteria.	✓		
3. Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis, 12 HOURS IF 8000 SERIES METHOD.			✓
3(a).Initial calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. COMMENTS:			✓
3(b).Continuous calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. COMMENTS:			✓
4. Blank Contamination - If yes, list compounds and concentrations in each blank: COMMENTS:		✓	
5. Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges. COMMENTS:			✓

6(a).Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

✓

If not met, list those compounds and their recoveries which fall outside the acceptable

COMMENTS:

Q1901-08MS -- Comp#5,8 recovery fail

Q1901-08MSD -- Comp#5,6,8 recovery fail , RPD Fail

6(b).Blank Spike/Blank Spike Duplicate Recoveries Meet Criteria

✓

If not met, list those compounds and their recoveries which fall outside the acceptable

COMMENTS:

7. Retention Time Shift Meet Criteria (if applicable)

✓

COMMENTS:

8. Extraction Holding Time Met

✓

If not met, list number of days exceeded for each sample:

COMMENTS:

9. Analysis Holding Time Met

✓

If not met, list those compounds and their recoveries which fall outside the acceptable range.

COMMENTS:

ADDITIONAL COMMENTS:

Helly

QA REVIEW

05/07/2025

Date