



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.:	Q1901				Order ID:	Q1901			
Client:	Portal Partners Tri-Venture				Project ID:	Amtrak Sawtooth Bridges 2025			
Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units	
Client ID :									
Total Concentration:				0.000					



SAMPLE DATA

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:		
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	05/01/25	
Client Sample ID:	PB167774TB		SDG No.:	Q1901	
Lab Sample ID:	PB167774TB		Matrix:	TCLP	
Analytical Method:	SW8081		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088372.D	1	05/01/25 08:56	05/01/25 14:06	PB167820

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.037	U	0.037	0.50	ug/L
76-44-8	Heptachlor	0.027	U	0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	0.096	U	0.096	0.50	ug/L
72-20-8	Endrin	0.032	U	0.032	0.50	ug/L
72-43-5	Methoxychlor	0.11	U	0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	19.1		30 (43) - 150 (140)	96%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.9		30 (77) - 150 (126)	94%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	04/26/25	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	04/28/25	
Client Sample ID:	B-167-SB01		SDG No.:	Q1901	
Lab Sample ID:	Q1901-08		Matrix:	TCLP	
Analytical Method:	SW8081		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088376.D	1	05/01/25 08:56	05/01/25 15:01	PB167820

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.037	U	0.037	0.50	ug/L
76-44-8	Heptachlor	0.027	U	0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	0.096	U	0.096	0.50	ug/L
72-20-8	Endrin	0.032	U	0.032	0.50	ug/L
72-43-5	Methoxychlor	0.11	U	0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	20.7		30 (43) - 150 (140)	104%	SPK: 20
877-09-8	Tetrachloro-m-xylene	19.6		30 (77) - 150 (126)	98%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	04/26/25	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	04/28/25	
Client Sample ID:	B-170-SB01		SDG No.:	Q1901	
Lab Sample ID:	Q1901-09		Matrix:	TCLP	
Analytical Method:	SW8081		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088379.D	1	05/01/25 08:56	05/01/25 15:42	PB167820

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.037	U	0.037	0.50	ug/L
76-44-8	Heptachlor	0.027	U	0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	0.096	U	0.096	0.50	ug/L
72-20-8	Endrin	0.032	U	0.032	0.50	ug/L
72-43-5	Methoxychlor	0.11	U	0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	19.9		30 (43) - 150 (140)	100%	SPK: 20
877-09-8	Tetrachloro-m-xylene	19.1		30 (77) - 150 (126)	95%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-167-SB02	SDG No.:	Q1901
Lab Sample ID:	Q1901-10	Matrix:	TCLP
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	TCLP Pesticide
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Decanted:	
		Final Vol:	10000
		Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088380.D	1	05/01/25 08:56	05/01/25 15:56	PB167820

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.037	U	0.037	0.50	ug/L
76-44-8	Heptachlor	0.027	U	0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	0.096	U	0.096	0.50	ug/L
72-20-8	Endrin	0.032	U	0.032	0.50	ug/L
72-43-5	Methoxychlor	0.11	U	0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	19.3		30 (43) - 150 (140)	96%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.5		30 (77) - 150 (126)	92%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-170-SB02	SDG No.:	Q1901
Lab Sample ID:	Q1901-11	Matrix:	TCLP
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Decanted:	
GPC Factor :	1.0	Final Vol:	10000
PH :		Test:	TCLP Pesticide
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088381.D	1	05/01/25 08:56	05/01/25 16:10	PB167820

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.037	U	0.037	0.50	ug/L
76-44-8	Heptachlor	0.027	U	0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	0.096	U	0.096	0.50	ug/L
72-20-8	Endrin	0.032	U	0.032	0.50	ug/L
72-43-5	Methoxychlor	0.11	U	0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	17.7		30 (43) - 150 (140)	89%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.1		30 (77) - 150 (126)	91%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



QC SUMMARY

Surrogate Summary

SDG No.: **Q1901**

Client: **Portal Partners Tri-Venture**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD088121.D	PIBLK-PD088121.D	Decachlorobiphenyl	1	20	22.9	115		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	20.1	101		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	22.7	113		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	20.9	104		30 (77)	150 (126)
I.BLK-PD088368.D	PIBLK-PD088368.D	Decachlorobiphenyl	1	20	20.7	103		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	20.4	102		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	18.9	94		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	19.8	99		30 (77)	150 (126)
PB167820BL	PB167820BL	Decachlorobiphenyl	1	20	18.6	93		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	18.0	90		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	16.8	84		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	18.5	92		30 (77)	150 (126)
PB167820BS	PB167820BS	Decachlorobiphenyl	1	20	20.5	102		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	20.0	100		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	18.7	93		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	20.5	103		30 (77)	150 (126)
PB167774TB	PB167774TB	Decachlorobiphenyl	1	20	19.1	96		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	18.4	92		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	17.2	86		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	18.9	94		30 (77)	150 (126)
Q1901-08	B-167-SB01	Decachlorobiphenyl	1	20	20.7	104		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	19.5	97		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	18.5	92		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	19.6	98		30 (77)	150 (126)
Q1901-08MS	B-167-SB01MS	Decachlorobiphenyl	1	20	20.8	104		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	18.4	92		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	19.0	95		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	18.8	94		30 (77)	150 (126)
Q1901-08MSD	B-167-SB01MSD	Decachlorobiphenyl	1	20	20.8	104		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	18.4	92		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	19.0	95		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	18.9	94		30 (77)	150 (126)
Q1901-09	B-170-SB01	Decachlorobiphenyl	1	20	19.9	100		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	18.9	95		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	18.1	91		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	19.1	95		30 (77)	150 (126)
Q1901-10	B-167-SB02	Decachlorobiphenyl	1	20	19.3	96		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	18.5	92		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	17.4	87		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	18.0	90		30 (77)	150 (126)
Q1901-11	B-170-SB02	Decachlorobiphenyl	1	20	17.7	89		30 (43)	150 (140)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1901-11	B-170-SB02	Tetrachloro-m-xylene	1	20	18.1	91		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	16.2	81		30 (43)	150 (140)
I.BLK-PD088383.D	PIBLK-PD088383.D	Tetrachloro-m-xylene	2	20	17.9	89		30 (77)	150 (126)
		Decachlorobiphenyl	1	20	21.0	105		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	20.4	102		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	19.9	99		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	19.9	100		30 (77)	150 (126)

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: 8081B

DataFile : PD088377.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Client Sample ID:	B-167-SB01MS											
Q1901-08MS	gamma-BHC (Lindane)	5	0	5.50	ug/L	110					30 (60)	150 (152)
	Heptachlor	5	0	5.40	ug/L	108					30 (56)	150 (147)
	Heptachlor epoxide	5	0	5.60	ug/L	112					30 (77)	150 (143)
	Endrin	5	0	5.50	ug/L	110					30 (76)	150 (144)
	Methoxychlor	5	0	5.10	ug/L	102					30 (70)	150 (142)

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: 8081B

DataFile : PD088378.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec Qual	RPD Qual	Low	Limits High	RPD
Client Sample ID: B-167-SB01MSD										
Q1901-08MSD	gamma-BHC (Lindane)	5	0	5.50	ug/L	110	0	30 (60)	150 (152)	20 (20)
	Heptachlor	5	0	5.40	ug/L	108	0	30 (56)	150 (147)	20 (20)
	Heptachlor epoxide	5	0	5.50	ug/L	110	2	30 (77)	150 (143)	20 (20)
	Endrin	5	0	5.50	ug/L	110	0	30 (76)	150 (144)	20 (20)
	Methoxychlor	5	0	5.00	ug/L	100	2	30 (70)	150 (142)	20 (20)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: **8081B** **Datafile :** PD088371.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167820BS	gamma-BHC (Lindane)	0.5	0.52	ug/L	104				40 (82)	140 (129)	
	Heptachlor	0.5	0.52	ug/L	104				40 (79)	140 (127)	
	Heptachlor epoxide	0.5	0.53	ug/L	106				40 (81)	140 (124)	
	Endrin	0.5	0.51	ug/L	102				40 (81)	140 (128)	
	Methoxychlor	0.5	0.46	ug/L	93				40 (78)	140 (108)	

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167820BL

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM Case No.: Q1901

SAS No.: Q1901 SDG NO.: Q1901

Lab Sample ID: PB167820BL

Lab File ID: PD088370.D

Matrix: (soil/water) water

Extraction: (Type) SEPF

Sulfur Cleanup: (Y/N) N

Date Extracted: 05/01/2025

Date Analyzed (1): 05/01/2025

Date Analyzed (2): 05/01/2025

Time Analyzed (1): 13:39

Time Analyzed (2): 13:39

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB167820BS	PB167820BS	PD088371.D	05/01/2025	05/01/2025
PB167774TB	PB167774TB	PD088372.D	05/01/2025	05/01/2025
B-167-SB01	Q1901-08	PD088376.D	05/01/2025	05/01/2025
B-167-SB01MS	Q1901-08MS	PD088377.D	05/01/2025	05/01/2025
B-167-SB01MSD	Q1901-08MSD	PD088378.D	05/01/2025	05/01/2025
B-170-SB01	Q1901-09	PD088379.D	05/01/2025	05/01/2025
B-167-SB02	Q1901-10	PD088380.D	05/01/2025	05/01/2025
B-170-SB02	Q1901-11	PD088381.D	05/01/2025	05/01/2025

COMMENTS: _____



QC SAMPLE DATA

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	
Client Sample ID:	PB167820BL	SDG No.:	Q1901
Lab Sample ID:	PB167820BL	Matrix:	TCLP
Analytical Method:	SW8081	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088370.D	1	05/01/25 08:56	05/01/25 13:39	PB167820

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	18.6		30 (43) - 150 (140)	93%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.5		30 (77) - 150 (126)	92%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/18/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/18/25
Client Sample ID:	PIBLK-PD088121.D	SDG No.:	Q1901
Lab Sample ID:	I.BLK-PD088121.D	Matrix:	TCLP
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	TCLP Pesticide
GPC Factor :	1.0	Injection Volume :	
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088121.D	1		04/18/25	PD041825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	22.9		30 (43) - 150 (140)	115%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.9		30 (77) - 150 (126)	104%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	05/01/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	05/01/25
Client Sample ID:	PIBLK-PD088368.D	SDG No.:	Q1901
Lab Sample ID:	I.BLK-PD088368.D	Matrix:	TCLP
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	TCLP Pesticide
GPC Factor :	1.0	Injection Volume :	
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088368.D	1		05/01/25	pd050125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	20.7		30 (43) - 150 (140)	103%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.4		30 (77) - 150 (126)	102%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	05/01/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	05/01/25
Client Sample ID:	PIBLK-PD088383.D	SDG No.:	Q1901
Lab Sample ID:	I.BLK-PD088383.D	Matrix:	TCLP
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	TCLP Pesticide
GPC Factor :	1.0	Injection Volume :	
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088383.D	1		05/01/25	pd050125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	21.0		30 (43) - 150 (140)	105%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.4		30 (77) - 150 (126)	102%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:		
Project:	Amtrak Sawtooth Bridges 2025		Date Received:		
Client Sample ID:	PB167820BS		SDG No.:	Q1901	
Lab Sample ID:	PB167820BS		Matrix:	TCLP	
Analytical Method:	SW8081		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088371.D	1	05/01/25 08:56	05/01/25 13:53	PB167820

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.52		0.0037	0.050	ug/L
76-44-8	Heptachlor	0.52		0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.53		0.0096	0.050	ug/L
72-20-8	Endrin	0.51		0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.46		0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	20.5		30 (43) - 150 (140)	102%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.5		30 (77) - 150 (126)	103%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MS	SDG No.:	Q1901
Lab Sample ID:	Q1901-08MS	Matrix:	TCLP
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Decanted:	
GPC Factor :	1.0	Final Vol:	10000
PH :		Test:	TCLP Pesticide
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088377.D	1	05/01/25 08:56	05/01/25 15:15	PB167820

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	5.50		0.037	0.50	ug/L
76-44-8	Heptachlor	5.40		0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	5.60		0.096	0.50	ug/L
72-20-8	Endrin	5.50		0.032	0.50	ug/L
72-43-5	Methoxychlor	5.10		0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	20.8		30 (43) - 150 (140)	104%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.8		30 (77) - 150 (126)	94%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MSD	SDG No.:	Q1901
Lab Sample ID:	Q1901-08MSD	Matrix:	TCLP
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Decanted:	
GPC Factor :	1.0	Final Vol:	10000
PH :		Test:	TCLP Pesticide
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088378.D	1	05/01/25 08:56	05/01/25 15:28	PB167820

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	5.50		0.037	0.50	ug/L
76-44-8	Heptachlor	5.40		0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	5.50		0.096	0.50	ug/L
72-20-8	Endrin	5.50		0.032	0.50	ug/L
72-43-5	Methoxychlor	5.00		0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	20.8		30 (43) - 150 (140)	104%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.9		30 (77) - 150 (126)	94%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



CALIBRATION SUMMARY

A
B
C
D
E
F
G
H
I
J
K
L

GC Column: ZB-MR1 ID: 0.32 (mm)

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

Instrument ID: ECD_D Calibration Date(s): 04/18/2025 04/18/2025

Calibration Times: 13:56 14:51

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID: RT 100 = PD088124.D RT 075 = PD088125.D

RT 050 = PD088126.D RT 025 = PD088127.D RT 005 = PD088128.D

COMPOUND	RT 100	RT 075	RT 050	RT 025	RT 005	MEAN RT	RT WINDOW	
							FROM	TO
Decachlorobiphenyl	8.09	8.08	8.08	8.08	8.08	8.08	7.98	8.18
Endrin	5.81	5.79	5.79	5.79	5.79	5.80	5.70	5.90
gamma-BHC (Lindane)	3.75	3.73	3.73	3.73	3.73	3.74	3.64	3.84
Heptachlor	4.10	4.09	4.09	4.09	4.09	4.09	3.99	4.19
Heptachlor epoxide	4.89	4.88	4.88	4.88	4.88	4.88	4.78	4.98
Methoxychlor	6.77	6.76	6.76	6.76	6.76	6.76	6.66	6.86
Tetrachloro-m-xylene	2.90	2.88	2.88	2.88	2.88	2.89	2.79	2.99

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: PORT06

Lab Code: CHEM **Case No.:** Q1901 **SAS No.:** Q1901 **SDG NO.:** Q1901

Instrument ID: ECD_D **Calibration Date(s):** 04/18/2025 04/18/2025
Calibration Times: 13:56 14:51

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID: CF 100 = PD088124.DCF 075 = PD088125.D CF 050 = PD088126.DCF 025 = PD088127.DCF 005 = PD088128.D							
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
Decachlorobiphenyl	3080820000	3141130000	3178140000	3290360000	3850090000	3308110000	9
Endrin	3128180000	3077300000	2935450000	2806460000	2965440000	2982570000	4
gamma-BHC (Lindane)	4507210000	4365130000	4155030000	3887470000	4022700000	4187510000	6
Heptachlor	4264490000	4159110000	3978190000	3781390000	4011590000	4038950000	5
Heptachlor epoxide	3670910000	3598060000	3486330000	3389810000	3724620000	3573950000	4
Methoxychlor	1461300000	1466360000	1467850000	1483290000	1618930000	1499540000	4
Tetrachloro-m-xylene	1982340000	2006790000	1938680000	1923660000	2135510000	1997400000	4

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: PORT06

Lab Code: CHEM **Case No.:** Q1901 **SAS No.:** Q1901 **SDG NO.:** Q1901

Instrument ID: ECD_D

Calibration Date(s): 04/18/2025 04/18/2025

Calibration Times: 13:56 14:51

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID: CF 100 = PD088124.DCF 075 = PD088125.DCF 050 = PD088126.DCF 025 = PD088127.DCF 005 = PD088128.D							
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
Decachlorobiphenyl	16767000000	17098200000	17470300000	18387600000	22674800000	18479600000	13
Endrin	16756200000	17001500000	17464500000	18215500000	21574100000	18202300000	11
gamma-BHC (Lindane)	20140800000	20226900000	20558100000	21137200000	25630400000	21538700000	11
Heptachlor	19826900000	20021500000	20522700000	21298800000	25320700000	21398100000	11
Heptachlor epoxide	17359900000	17576100000	18079700000	18879200000	22576700000	18894300000	11
Methoxychlor	8272330000	8467510000	8839240000	9290490000	10735700000	9121060000	11
Tetrachloro-m-xylene	13615300000	13685800000	14010800000	14551200000	17245000000	14621600000	10

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

Instrument ID: ECD_D Date(s) Analyzed: 04/18/2025 04/18/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	4.72	4.62	4.82	161995000
		2	5.24	5.14	5.34	160780000
		3	5.95	5.85	6.05	666185000
		4	6.03	5.93	6.13	797101000
		5	6.87	6.77	6.97	133051000
Toxaphene	500	1	6.24	6.14	6.34	23597900
		2	6.44	6.34	6.54	34064600
		3	7.15	7.05	7.25	65326300
		4	7.57	7.47	7.67	82045400
		5	7.93	7.83	8.03	47715200

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

Instrument ID: ECD_D Date(s) Analyzed: 04/18/2025 04/18/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	3.91	3.81	4.01	767527000
		2	4.49	4.39	4.59	783082000
		3	5.13	5.03	5.23	2396830000
		4	5.19	5.09	5.29	2031740000
		5	6.09	5.99	6.19	931551000
Toxaphene	500	1	5.48	5.38	5.58	130682000
		2	5.65	5.55	5.75	89513500
		3	6.76	6.66	6.86	417352000
		4	7.20	7.10	7.30	302911000
		5	7.33	7.23	7.43	207761000

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

Continuing Calib Date: 05/01/2025 Initial Calibration Date(s): 04/18/2025 04/18/2025

Continuing Calib Time: 12:09 Initial Calibration Time(s): 13:56 14:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.07	9.08	8.98	9.18	0.01
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.93	4.93	4.83	5.03	0.00
Heptachlor epoxide	5.69	5.69	5.59	5.79	0.00
Endrin	6.58	6.58	6.48	6.68	0.01
Methoxychlor	7.49	7.50	7.40	7.60	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

Continuing Calib Date: 05/01/2025 Initial Calibration Date(s): 04/18/2025 04/18/2025

Continuing Calib Time: 12:09 Initial Calibration Time(s): 13:56 14:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.08	8.08	7.98	8.18	0.01
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
gamma-BHC (Lindane)	3.73	3.73	3.63	3.83	0.00
Heptachlor	4.08	4.09	3.99	4.19	0.01
Heptachlor epoxide	4.88	4.88	4.78	4.98	0.00
Endrin	5.79	5.79	5.69	5.89	0.00
Methoxychlor	6.76	6.76	6.66	6.86	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL01 Date Analyzed: 05/01/2025

Lab Sample No.: PSTDCCC050 Data File : PD088369.D Time Analyzed: 12:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.073	8.975	9.175	46.430	50.000	-7.1
Endrin	6.575	6.476	6.676	52.080	50.000	4.2
gamma-BHC (Lindane)	4.331	4.232	4.432	54.230	50.000	8.5
Heptachlor	4.930	4.831	5.031	53.020	50.000	6.0
Heptachlor epoxide	5.691	5.592	5.792	53.870	50.000	7.7
Methoxychlor	7.494	7.395	7.595	45.160	50.000	-9.7
Tetrachloro-m-xylene	3.551	3.452	3.652	52.750	50.000	5.5

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL01 Date Analyzed: 05/01/2025

Lab Sample No.: PSTDCCC050 Data File : PD088369.D Time Analyzed: 12:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.075	7.977	8.177	41.500	50.000	-17.0
Endrin	5.791	5.693	5.893	46.350	50.000	-7.3
gamma-BHC (Lindane)	3.731	3.633	3.833	48.020	50.000	-4.0
Heptachlor	4.084	3.986	4.186	46.300	50.000	-7.4
Heptachlor epoxide	4.875	4.777	4.977	48.240	50.000	-3.5
Methoxychlor	6.756	6.659	6.859	41.540	50.000	-16.9
Tetrachloro-m-xylene	2.882	2.783	2.983	48.750	50.000	-2.5

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

Continuing Calib Date: 05/01/2025 Initial Calibration Date(s): 04/18/2025 04/18/2025

Continuing Calib Time: 17:19 Initial Calibration Time(s): 13:56 14:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.08	9.08	8.98	9.18	0.00
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.93	4.93	4.83	5.03	0.00
Heptachlor epoxide	5.69	5.69	5.59	5.79	0.00
Endrin	6.58	6.58	6.48	6.68	0.00
Methoxychlor	7.50	7.50	7.40	7.60	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

Continuing Calib Date: 05/01/2025 Initial Calibration Date(s): 04/18/2025 04/18/2025

Continuing Calib Time: 17:19 Initial Calibration Time(s): 13:56 14:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.08	8.08	7.98	8.18	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
gamma-BHC (Lindane)	3.73	3.73	3.63	3.83	0.00
Heptachlor	4.09	4.09	3.99	4.19	0.00
Heptachlor epoxide	4.88	4.88	4.78	4.98	0.00
Endrin	5.79	5.79	5.69	5.89	0.00
Methoxychlor	6.76	6.76	6.66	6.86	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL02 Date Analyzed: 05/01/2025

Lab Sample No.: PSTDCCC050 Data File : PD088384.D Time Analyzed: 17:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.077	8.975	9.175	47.420	50.000	-5.2
Endrin	6.577	6.476	6.676	52.460	50.000	4.9
gamma-BHC (Lindane)	4.333	4.232	4.432	53.820	50.000	7.6
Heptachlor	4.933	4.831	5.031	53.190	50.000	6.4
Heptachlor epoxide	5.694	5.592	5.792	53.690	50.000	7.4
Methoxychlor	7.496	7.395	7.595	47.380	50.000	-5.2
Tetrachloro-m-xylene	3.553	3.452	3.652	52.240	50.000	4.5

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL02 Date Analyzed: 05/01/2025

Lab Sample No.: PSTDCCC050 Data File : PD088384.D Time Analyzed: 17:19

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.076	7.977	8.177	43.900	50.000	-12.2
Endrin	5.791	5.693	5.893	47.310	50.000	-5.4
gamma-BHC (Lindane)	3.731	3.633	3.833	48.110	50.000	-3.8
Heptachlor	4.085	3.986	4.186	46.890	50.000	-6.2
Heptachlor epoxide	4.875	4.777	4.977	48.060	50.000	-3.9
Methoxychlor	6.757	6.659	6.859	44.580	50.000	-10.8
Tetrachloro-m-xylene	2.882	2.783	2.983	48.420	50.000	-3.2

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: **PORT06**

Lab Code: **CHEM** Case No.: **Q1901** SAS No.: **Q1901** SDG NO.: **Q1901**

GC Column: **ZB-MR1** ID: **0.32** (mm) Initi. Calib. Date(s): **04/18/2025** **04/18/2025**

Client Sample No. (PEM): **PEM - PD088122.D** Date Analyzed: **04/18/2025**

Lab Sample No.(PEM): **PEM** Time Analyzed: **13:29**

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	9.075	8.970	9.180	23.220	20.000	16.1
Tetrachloro-m-xylene	3.551	3.500	3.600	21.610	20.000	8.1
alpha-BHC	4.000	3.950	4.050	9.950	10.000	-0.5
beta-BHC	4.516	4.470	4.570	11.360	10.000	13.6
gamma-BHC (Lindane)	4.331	4.280	4.380	10.480	10.000	4.8
Endrin	6.576	6.510	6.650	51.530	50.000	3.1
4,4'-DDT	7.023	6.950	7.090	110.510	100.000	10.5
Methoxychlor	7.494	7.420	7.560	265.100	250.000	6.0

GC Column: **ZB-MR2** ID: **0.32** (mm) Initi. Calib. Date(s): **04/18/2025** **04/18/2025**

Client Sample No. (PEM): **PEM - PD088122.D** Date Analyzed: **04/18/2025**

Lab Sample No.(PEM): **PEM** Time Analyzed: **13:29**

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	8.076	7.980	8.180	22.950	20.000	14.8
Tetrachloro-m-xylene	2.883	2.830	2.930	22.200	20.000	11.0
alpha-BHC	3.396	3.350	3.450	11.720	10.000	17.2
beta-BHC	4.028	3.980	4.080	12.460	10.000	24.6
gamma-BHC (Lindane)	3.732	3.680	3.780	11.580	10.000	15.8
Endrin	5.793	5.720	5.860	49.780	50.000	-0.4
4,4'-DDT	6.187	6.120	6.260	102.610	100.000	2.6
Methoxychlor	6.758	6.690	6.830	215.580	250.000	-13.8

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: **PORT06**

Lab Code: **CHEM** Case No.: **Q1901** SAS No.: **Q1901** SDG NO.: **Q1901**

GC Column: **ZB-MR1** ID: **0.32** (mm) Initi. Calib. Date(s): **04/18/2025** **04/18/2025**

Client Sample No. (PEM): **PEM - PD088365.D** Date Analyzed: **05/01/2025**

Lab Sample No.(PEM): **PEM** Time Analyzed: **09:55**

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.073	8.970	9.170	20.730	20.000	3.7
Tetrachloro-m-xylene	3.550	3.500	3.600	21.750	20.000	8.8
alpha-BHC	3.999	3.950	4.050	9.880	10.000	-1.2
beta-BHC	4.515	4.460	4.570	11.390	10.000	13.9
gamma-BHC (Lindane)	4.330	4.280	4.380	10.360	10.000	3.6
Endrin	6.575	6.500	6.650	53.330	50.000	6.7
4,4'-DDT	7.021	6.950	7.090	95.920	100.000	-4.1
Methoxychlor	7.493	7.420	7.560	223.310	250.000	-10.7

GC Column: **ZB-MR2** ID: **0.32** (mm) Initi. Calib. Date(s): **04/18/2025** **04/18/2025**

Client Sample No. (PEM): **PEM - PD088365.D** Date Analyzed: **05/01/2025**

Lab Sample No.(PEM): **PEM** Time Analyzed: **09:55**

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.074	7.970	8.170	18.740	20.000	-6.3
Tetrachloro-m-xylene	2.881	2.830	2.930	21.090	20.000	5.5
alpha-BHC	3.394	3.340	3.440	10.870	10.000	8.7
beta-BHC	4.026	3.980	4.080	11.230	10.000	12.3
gamma-BHC (Lindane)	3.730	3.680	3.780	10.780	10.000	7.8
Endrin	5.790	5.720	5.860	46.960	50.000	-6.1
4,4'-DDT	6.185	6.110	6.260	83.920	100.000	-16.1
Methoxychlor	6.755	6.680	6.830	168.520	250.000	-32.6

Analytical Sequence

Client: Portal Partners Tri-Venture	SDG No.: Q1901
Project: Amtrak Sawtooth Bridges 2025	Instrument ID: ECD_D
GC Column: ZB-MR1	ID: 0.32 (mm) Inst. Calib. Date(s): 04/18/2025 04/18/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
LBLK	LBLK	04/18/2025	13:15	PD088121.D	9.07	3.55
PEM	PEM	04/18/2025	13:29	PD088122.D	9.08	3.55
RESCHK	RESCHK	04/18/2025	13:43	PD088123.D	9.08	3.55
PSTDICC100	PSTDICC100	04/18/2025	13:56	PD088124.D	9.08	3.55
PSTDICC075	PSTDICC075	04/18/2025	14:10	PD088125.D	9.07	3.55
PSTDICC050	PSTDICC050	04/18/2025	14:24	PD088126.D	9.08	3.55
PSTDICC025	PSTDICC025	04/18/2025	14:37	PD088127.D	9.08	3.55
PSTDICC005	PSTDICC005	04/18/2025	14:51	PD088128.D	9.08	3.55
PTOXICC500	PTOXICC500	04/18/2025	16:40	PD088136.D	9.07	3.55
PEM	PEM	05/01/2025	09:55	PD088365.D	9.07	3.55
LBLK	LBLK	05/01/2025	11:56	PD088368.D	9.07	3.55
PSTDCCC050	PSTDCCC050	05/01/2025	12:09	PD088369.D	9.07	3.55
PB167820BL	PB167820BL	05/01/2025	13:39	PD088370.D	9.09	3.56
PB167820BS	PB167820BS	05/01/2025	13:53	PD088371.D	9.08	3.55
PB167774TB	PB167774TB	05/01/2025	14:06	PD088372.D	9.07	3.55
B-167-SB01	Q1901-08	05/01/2025	15:01	PD088376.D	9.08	3.55
B-167-SB01MS	Q1901-08MS	05/01/2025	15:15	PD088377.D	9.07	3.55
B-167-SB01MSD	Q1901-08MSD	05/01/2025	15:28	PD088378.D	9.07	3.55
B-170-SB01	Q1901-09	05/01/2025	15:42	PD088379.D	9.07	3.55
B-167-SB02	Q1901-10	05/01/2025	15:56	PD088380.D	9.07	3.55
B-170-SB02	Q1901-11	05/01/2025	16:10	PD088381.D	9.07	3.55
LBLK	LBLK	05/01/2025	16:37	PD088383.D	9.07	3.55
PSTDCCC050	PSTDCCC050	05/01/2025	17:19	PD088384.D	9.08	3.55

Analytical Sequence

Client: Portal Partners Tri-Venture	SDG No.: Q1901	
Project: Amtrak Sawtooth Bridges 2025	Instrument ID: ECD_D	
GC Column: ZB-MR2	ID: 0.32 (mm)	Inst. Calib. Date(s): 04/18/2025 04/18/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	04/18/2025	13:15	PD088121.D	8.08	2.88
PEM	PEM	04/18/2025	13:29	PD088122.D	8.08	2.88
RESCHK	RESCHK	04/18/2025	13:43	PD088123.D	8.08	2.88
PSTDICC100	PSTDICC100	04/18/2025	13:56	PD088124.D	8.09	2.90
PSTDICC075	PSTDICC075	04/18/2025	14:10	PD088125.D	8.08	2.88
PSTDICC050	PSTDICC050	04/18/2025	14:24	PD088126.D	8.08	2.88
PSTDICC025	PSTDICC025	04/18/2025	14:37	PD088127.D	8.08	2.88
PSTDICC005	PSTDICC005	04/18/2025	14:51	PD088128.D	8.08	2.88
PTOXICC500	PTOXICC500	04/18/2025	16:40	PD088136.D	8.08	2.88
PEM	PEM	05/01/2025	09:55	PD088365.D	8.07	2.88
IBLK	IBLK	05/01/2025	11:56	PD088368.D	8.08	2.88
PSTDCCC050	PSTDCCC050	05/01/2025	12:09	PD088369.D	8.08	2.88
PB167820BL	PB167820BL	05/01/2025	13:39	PD088370.D	8.08	2.88
PB167820BS	PB167820BS	05/01/2025	13:53	PD088371.D	8.08	2.88
PB167774TB	PB167774TB	05/01/2025	14:06	PD088372.D	8.07	2.88
B-167-SB01	Q1901-08	05/01/2025	15:01	PD088376.D	8.08	2.88
B-167-SB01MS	Q1901-08MS	05/01/2025	15:15	PD088377.D	8.08	2.88
B-167-SB01MSD	Q1901-08MSD	05/01/2025	15:28	PD088378.D	8.08	2.88
B-170-SB01	Q1901-09	05/01/2025	15:42	PD088379.D	8.08	2.88
B-167-SB02	Q1901-10	05/01/2025	15:56	PD088380.D	8.08	2.88
B-170-SB02	Q1901-11	05/01/2025	16:10	PD088381.D	8.08	2.88
IBLK	IBLK	05/01/2025	16:37	PD088383.D	8.08	2.88
PSTDCCC050	PSTDCCC050	05/01/2025	17:19	PD088384.D	8.08	2.88

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

B-167-SB01MS

Contract: PORT06

Lab Code: CHEM Case No.: Q1901

SAS No.: Q1901 SDG NO.: Q1901

Lab Sample ID: Q1901-08MS

Date(s) Analyzed: 05/01/2025 05/01/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endrin	1	6.58	6.53	6.63	5.50	13.6
	2	5.79	5.74	5.84	4.80	
Methoxychlor	1	7.49	7.44	7.54	5.10	10.3
	2	6.76	6.71	6.81	4.60	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	5.50	11.5
	2	3.73	3.68	3.78	4.90	
Heptachlor	1	4.93	4.88	4.98	5.40	13.9
	2	4.09	4.04	4.14	4.70	
Heptachlor epoxide	1	5.69	5.64	5.74	5.60	13.3
	2	4.88	4.83	4.93	4.90	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

B-167-SB01MSD

Contract: PORT06

Lab Code: CHEM Case No.: Q1901

SAS No.: Q1901 SDG NO.: Q1901

Lab Sample ID: Q1901-08MSD

Date(s) Analyzed: 05/01/2025 05/01/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endrin	1	6.58	6.53	6.63	5.50	11.5
	2	5.79	5.74	5.84	4.90	
Methoxychlor	1	7.49	7.44	7.54	5.00	8.3
	2	6.76	6.71	6.81	4.60	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	5.50	11.5
	2	3.73	3.68	3.78	4.90	
Heptachlor	1	4.93	4.88	4.98	5.40	13.9
	2	4.09	4.04	4.14	4.70	
Heptachlor epoxide	1	5.69	5.64	5.74	5.50	11.5
	2	4.88	4.83	4.93	4.90	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB167820BS

Contract: PORT06

Lab Code: CHEM

Case No.: Q1901

SAS No.: Q1901

SDG NO.: Q1901

Lab Sample ID: PB167820BS

Date(s) Analyzed: 05/01/2025

05/01/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Methoxychlor	1	7.49	7.44	7.54	0.46	8.3
	2	6.76	6.71	6.81	0.43	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	0.52	8.6
	2	3.73	3.68	3.78	0.48	
Heptachlor	1	4.93	4.88	4.98	0.52	11.3
	2	4.08	4.03	4.13	0.46	
Heptachlor epoxide	1	5.69	5.64	5.74	0.53	9.9
	2	4.87	4.82	4.92	0.48	
Endrin	1	6.58	6.53	6.63	0.51	11.2
	2	5.79	5.74	5.84	0.46	



SAMPLE RAW DATA

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050125\
 Data File : PD088372.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2025 14:06
 Operator : AR\AJ
 Sample : PB167774TB
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB167774TB

A

B

C

D

E

F

G

H

I

J

K

L

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 01:31:36 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.550	2.881	36831889	275.8E6	18.440	18.865
28) SA Decachlor...	9.073	8.074	63249202	317.9E6	19.119	17.201

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050125\
Data File : PD088372.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2025 14:06
Operator : AR\AJ
Sample : PB167774TB
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB167774TB

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 01:31:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.550 min
Delta R.T.: -0.002 min
Response: 36831889
Conc: 18.44 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB167774TB




#1 Tetrachloro-m-xylene

R.T.: 2.881 min
Delta R.T.: -0.002 min
Response: 275830853
Conc: 18.86 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.073 min
Delta R.T.: -0.002 min
Response: 63249202
Conc: 19.12 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.074 min
Delta R.T.: -0.003 min
Response: 317876577
Conc: 17.20 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050125\
 Data File : PD088376.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2025 15:01
 Operator : AR\AJ
 Sample : Q1901-08
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 B-167-SB01

A
B
C
D
E
F
G
H
I
J
K
L

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 01:32:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.551	2.882	38920336	285.9E6	19.486	19.551
28) SA Decachlor...	9.075	8.076	68461731	341.7E6	20.695	18.492

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050125\
Data File : PD088376.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2025 15:01
Operator : AR\AJ
Sample : Q1901-08
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
B-167-SB01

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 01:32:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.551 min
Delta R.T.: -0.001 min
Response: 38920336
Conc: 19.49 ng/ml

Instrument :
ECD_D
ClientSampleId :
B-167-SB01




#1 Tetrachloro-m-xylene

R.T.: 2.882 min
Delta R.T.: 0.000 min
Response: 285868038
Conc: 19.55 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.075 min
Delta R.T.: 0.000 min
Response: 68461731
Conc: 20.70 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.076 min
Delta R.T.: -0.001 min
Response: 341722367
Conc: 18.49 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050125\
 Data File : PD088379.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2025 15:42
 Operator : AR\AJ
 Sample : Q1901-09
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 B-170-SB01

A

B

C

D

E

F

G

H

I

J

K

L

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 01:32:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.551	2.882	37826834	278.7E6	18.938	19.063
28) SA Decachlor...	9.074	8.075	65897934	335.1E6	19.920	18.131

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050125\
Data File : PD088379.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2025 15:42
Operator : AR\AJ
Sample : Q1901-09
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
B-170-SB01

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 01:32:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.551 min
Delta R.T.: -0.001 min
Response: 37826834
Conc: 18.94 ng/ml

Instrument :
ECD_D
ClientSampleId :
B-170-SB01




#1 Tetrachloro-m-xylene

R.T.: 2.882 min
Delta R.T.: 0.000 min
Response: 278728559
Conc: 19.06 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.074 min
Delta R.T.: -0.001 min
Response: 65897934
Conc: 19.92 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.075 min
Delta R.T.: -0.002 min
Response: 335062220
Conc: 18.13 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050125\
 Data File : PD088380.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2025 15:56
 Operator : AR\AJ
 Sample : Q1901-10
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 B-167-SB02

A
B
C
D
E
F
G
H
I
J
K
L

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 01:33:09 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.551	2.883	36913185	262.7E6	18.481	17.966
28) SA Decachlor...	9.073	8.075	63730197	321.6E6	19.265	17.403

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050125\
Data File : PD088380.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2025 15:56
Operator : AR\AJ
Sample : Q1901-10
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
B-167-SB02

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 01:33:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.551 min
Delta R.T.: 0.000 min
Response: 36913185
Conc: 18.48 ng/ml

Instrument :
ECD_D
ClientSampleId :
B-167-SB02




#1 Tetrachloro-m-xylene

R.T.: 2.883 min
Delta R.T.: 0.000 min
Response: 262693661
Conc: 17.97 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.073 min
Delta R.T.: -0.002 min
Response: 63730197
Conc: 19.26 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.075 min
Delta R.T.: -0.002 min
Response: 321607992
Conc: 17.40 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050125\
 Data File : PD088381.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2025 16:10
 Operator : AR\AJ
 Sample : Q1901-11
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 B-170-SB02

A
B
C
D
E
F
G
H
I
J
K
L

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 01:33:20 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.551	2.882	36221576	261.2E6	18.134	17.861
28) SA Decachlor...	9.074	8.075	58572360	299.6E6	17.706	16.214

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050125\
Data File : PD088381.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2025 16:10
Operator : AR\AJ
Sample : Q1901-11
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
B-170-SB02

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 01:33:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.551 min
Delta R.T.: -0.001 min
Response: 36221576
Conc: 18.13 ng/ml

Instrument :
ECD_D
ClientSampleId :
B-170-SB02




#1 Tetrachloro-m-xylene

R.T.: 2.882 min
Delta R.T.: 0.000 min
Response: 261153881
Conc: 17.86 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.074 min
Delta R.T.: -0.001 min
Response: 58572360
Conc: 17.71 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.075 min
Delta R.T.: -0.002 min
Response: 299625668
Conc: 16.21 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050125\
 Data File : PD088370.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2025 13:39
 Operator : AR\AJ
 Sample : PB167820BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB167820BL

A
B
C
D
E
F
G
H
I
J
K
L

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 01:31:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.558	2.881	35908305	270.0E6	17.978	18.464
28) SA Decachlor...	9.085	8.077	61588544	310.0E6	18.617	16.773

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050125\
Data File : PD088370.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2025 13:39
Operator : AR\AJ
Sample : PB167820BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB167820BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 01:31:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.558 min
Delta R.T.: 0.006 min
Response: 35908305
Conc: 17.98 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB167820BL




#1 Tetrachloro-m-xylene

R.T.: 2.881 min
Delta R.T.: -0.002 min
Response: 269968842
Conc: 18.46 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.085 min
Delta R.T.: 0.010 min
Response: 61588544
Conc: 18.62 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.077 min
Delta R.T.: 0.000 min
Response: 309963841
Conc: 16.77 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050125\
 Data File : PD088371.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2025 13:53
 Operator : AR\AJ
 Sample : PB167820BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB167820BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 01:31:21 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.550	2.881	39930635	300.2E6	19.991	20.532
28)	SA Decachlor...	9.075	8.075	67725971	345.4E6	20.473	18.691
Target Compounds							
2)	A alpha-BHC	4.000	3.394	225.0E6	1102.5E6	52.184	48.223
3)	MA gamma-BHC...	4.331	3.730	217.2E6	1024.5E6	51.873	47.567
4)	MA Heptachlor	4.931	4.084	210.0E6	993.9E6	51.998	46.449
5)	MB Aldrin	5.273	4.370	209.9E6	1002.9E6	53.154	48.208
6)	B beta-BHC	4.516	4.026	83660804	455.7E6	51.535	49.238
7)	B delta-BHC	4.765	4.263	225.9E6	1026.1E6	54.769	48.261
8)	B Heptachlo...	5.692	4.874	189.7E6	908.4E6	53.068	48.078
9)	A Endosulfan I	6.076	5.248	179.9E6	866.3E6	53.271	48.094
10)	B gamma-Chl...	5.948	5.127	191.7E6	978.5E6	52.935	48.211
11)	B alpha-Chl...	6.028	5.191	190.8E6	940.2E6	52.864	47.963
12)	B 4,4'-DDE	6.196	5.377	171.9E6	941.9E6	52.127	47.827
13)	MA Dieldrin	6.348	5.514	193.4E6	958.8E6	54.187	48.102
14)	MA Endrin	6.576	5.790	152.2E6	829.9E6	51.027	45.595
15)	B Endosulfa...	6.788	6.082	162.3E6	830.9E6	51.905	47.420
16)	A 4,4'-DDD	6.706	5.931	137.3E6	807.3E6	54.586	49.265
17)	MA 4,4'-DDT	7.022	6.185	137.1E6	755.9E6	49.352	44.424
18)	B Endrin al...	6.916	6.260	124.1E6	638.6E6	53.785	48.013
19)	B Endosulfa...	7.151	6.483	150.8E6	796.3E6	52.439	46.477
20)	A Methoxychlor	7.494	6.756	69710493	390.2E6	46.488	42.783
21)	B Endrin ke...	7.631	6.992	163.4E6	881.6E6	52.946	47.133
22)	Mirex	8.115	7.187	116.8E6	661.6E6	49.734	44.642



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050125\
Data File : PD088371.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2025 13:53
Operator : AR\AJ
Sample : PB167820BS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB167820BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 01:31:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050125\
 Data File : PD088377.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2025 15:15
 Operator : AR\AJ
 Sample : Q1901-08MS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

B-167-SB01MS

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 05/02/2025

Supervised By :mohammad ahmed 05/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 01:32:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.551	2.882	36861410	275.2E6	18.455	18.819
28)	SA Decachlor...	9.073	8.075	68809273	350.8E6	20.800	18.982
Target Compounds							
2)	A alpha-BHC	4.000	3.395	235.8E6	1126.2E6	54.674	49.263
3)	MA gamma-BHC...	4.331	3.731	232.4E6	1056.7E6	55.491	49.060
4)	MA Heptachlor	4.930	4.085	217.9E6	1014.4E6	53.951	47.407
5)	MB Aldrin	5.271	4.370	215.8E6	1006.6E6	54.637	48.387
6)	B beta-BHC	4.515	4.027	87031480	474.2E6	53.611	51.240
7)	B delta-BHC	4.763	4.264	254.2E6	1062.3E6	61.633m	49.963
8)	B Heptachlo...	5.691	4.875	198.5E6	930.3E6	55.538	49.239
9)	A Endosulfan I	6.075	5.249	189.2E6	886.2E6	56.020	49.201
10)	B gamma-Chl...	5.946	5.128	202.0E6	998.5E6	55.779	49.198
11)	B alpha-Chl...	6.027	5.192	200.5E6	961.6E6	55.553	49.054
12)	B 4,4'-DDE	6.196	5.377	180.1E6	963.7E6	54.597	48.935
13)	MA Dieldrin	6.347	5.515	202.3E6	981.3E6	56.695	49.229
14)	MA Endrin	6.575	5.791	165.3E6	881.8E6	55.424	48.443
15)	B Endosulfa...	6.786	6.082	170.8E6	862.0E6	54.646	49.197
16)	A 4,4'-DDD	6.705	5.931	144.9E6	830.1E6	57.621	50.658
17)	MA 4,4'-DDT	7.022	6.185	146.7E6	789.4E6	52.831	46.391
18)	B Endrin al...	6.915	6.260	128.7E6	653.1E6	55.767	49.098
19)	B Endosulfa...	7.149	6.484	160.3E6	829.6E6	55.736	48.419
20)	A Methoxychlor	7.493	6.756	75839303	419.5E6	50.575	45.993
21)	B Endrin ke...	7.630	6.993	171.9E6	903.8E6	55.689	48.320
22)	Mirex	8.114	7.187	123.2E6	682.4E6	52.456	46.045

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050125\
Data File : PD088377.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2025 15:15
Operator : AR\AJ
Sample : Q1901-08MS
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

B-167-SB01MS

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 05/02/2025

Supervised By :mohammad ahmed 05/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 01:32:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050125\
 Data File : PD088378.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2025 15:28
 Operator : AR\AJ
 Sample : Q1901-08MSD
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

B-167-SB01MSD

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 05/02/2025

Supervised By :mohammad ahmed 05/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 01:32:46 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.551	2.882	36653010	276.3E6	18.350	18.894
28)	SA Decachlor...	9.074	8.075	68849811	350.7E6	20.812	18.979
Target Compounds							
2)	A alpha-BHC	4.000	3.395	235.0E6	1131.9E6	54.489	49.509
3)	MA gamma-BHC...	4.331	3.731	231.7E6	1058.7E6	55.341	49.152
4)	MA Heptachlor	4.930	4.085	217.3E6	1013.1E6	53.810	47.344
5)	MB Aldrin	5.272	4.371	215.5E6	1009.1E6	54.569	48.507
6)	B beta-BHC	4.516	4.027	86867239	472.8E6	53.510	51.089
7)	B delta-BHC	4.763	4.263	252.8E6	1061.4E6	61.292m	49.922
8)	B Heptachlo...	5.692	4.875	197.9E6	929.5E6	55.376	49.196
9)	A Endosulfan I	6.075	5.249	188.6E6	887.4E6	55.851	49.264
10)	B gamma-Chl...	5.947	5.127	201.3E6	999.2E6	55.578	49.231
11)	B alpha-Chl...	6.028	5.192	199.7E6	962.1E6	55.328	49.079
12)	B 4,4'-DDE	6.196	5.378	179.8E6	966.8E6	54.527	49.092
13)	MA Dieldrin	6.347	5.515	202.1E6	983.2E6	56.622	49.325
14)	MA Endrin	6.575	5.791	164.9E6	885.9E6	55.274	48.670
15)	B Endosulfa...	6.787	6.082	170.3E6	867.2E6	54.489	49.494
16)	A 4,4'-DDD	6.705	5.932	144.6E6	832.4E6	57.492	50.799
17)	MA 4,4'-DDT	7.022	6.186	145.9E6	790.7E6	52.524	46.469
18)	B Endrin al...	6.915	6.260	129.0E6	656.5E6	55.889	49.360
19)	B Endosulfa...	7.150	6.484	160.3E6	833.1E6	55.756	48.625
20)	A Methoxychlor	7.494	6.756	75400976	420.1E6	50.283	46.062
21)	B Endrin ke...	7.631	6.993	171.7E6	911.1E6	55.616	48.714
22)	Mirex	8.114	7.188	122.8E6	686.0E6	52.306	46.286

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050125\
Data File : PD088378.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2025 15:28
Operator : AR\AJ
Sample : Q1901-08MSD
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

B-167-SB01MSD

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 05/02/2025

Supervised By :mohammad ahmed 05/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 01:32:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integration Report

Sequence:	PD041825	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PD088122.D	4,4"-DDD #2	Abdul	4/19/2025 2:06:43 PM	mohammad	4/22/2025 2:20:46	Peak Integrated by Software
PSTDICC100	PD088124.D	Heptachlor epoxide #2	Abdul	4/19/2025 2:06:47 PM	mohammad	4/22/2025 2:20:46	Peak Integrated by Software
PSTDICC005	PD088128.D	delta-BHC	Abdul	4/19/2025 2:20:23 PM	mohammad	4/22/2025 2:20:46	Peak Integrated by Software
PEM	PD088143.D	4,4"-DDD	Abdul	4/19/2025 2:07:12 PM	mohammad	4/22/2025 2:20:46	Peak Integrated by Software
PEM	PD088143.D	Endrin aldehyde	Abdul	4/19/2025 2:07:12 PM	mohammad	4/22/2025 2:20:46	Peak Integrated by Software

Manual Integration Report

Sequence:	pd050125	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PD088365.D	4,4"-DDE	Abdul	5/2/2025 8:29:00 AM	mohammad	5/5/2025 4:37:38	Peak Integrated by Software
PEM	PD088365.D	4,4"-DDE #2	Abdul	5/2/2025 8:29:00 AM	mohammad	5/5/2025 4:37:38	Peak Integrated by Software
PEM	PD088365.D	Endrin aldehyde #2	Abdul	5/2/2025 8:29:00 AM	mohammad	5/5/2025 4:37:38	Peak Integrated by Software
Q1901-08MS	PD088377.D	delta-BHC	Abdul	5/2/2025 8:29:14 AM	mohammad	5/5/2025 4:37:48	Peak Integrated by Software
Q1901-08MSD	PD088378.D	delta-BHC	Abdul	5/2/2025 8:29:17 AM	mohammad	5/5/2025 4:37:33	Peak Integrated by Software
PSTDCCC050	PD088384.D	4,4"-DDE #2	Abdul	5/2/2025 8:29:21 AM	mohammad	5/5/2025 4:37:35	Peak Integrated by Software

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD041825

Review By	Abdul	Review On	4/19/2025 2:07:56 PM
Supervise By	mohammad	Supervise On	4/22/2025 2:20:46 AM
SubDirectory	PD041825	HP Acquire Method	HP Processing Method pd041825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD088120.D	18 Apr 2025 13:02	AR\AJ	Ok
2	I.BLK	PD088121.D	18 Apr 2025 13:15	AR\AJ	Ok
3	PEM	PD088122.D	18 Apr 2025 13:29	AR\AJ	Ok,M
4	RESCHK	PD088123.D	18 Apr 2025 13:43	AR\AJ	Ok
5	PSTDICC100	PD088124.D	18 Apr 2025 13:56	AR\AJ	Ok,M
6	PSTDICC075	PD088125.D	18 Apr 2025 14:10	AR\AJ	Ok
7	PSTDICC050	PD088126.D	18 Apr 2025 14:24	AR\AJ	Ok
8	PSTDICC025	PD088127.D	18 Apr 2025 14:37	AR\AJ	Ok
9	PSTDICC005	PD088128.D	18 Apr 2025 14:51	AR\AJ	Ok,M
10	PCHLORICC1000	PD088129.D	18 Apr 2025 15:05	AR\AJ	Ok
11	PCHLORICC750	PD088130.D	18 Apr 2025 15:18	AR\AJ	Ok
12	PCHLORICC500	PD088131.D	18 Apr 2025 15:32	AR\AJ	Ok
13	PCHLORICC250	PD088132.D	18 Apr 2025 15:46	AR\AJ	Ok
14	PCHLORICC050	PD088133.D	18 Apr 2025 15:59	AR\AJ	Ok,M
15	PTOXICC1000	PD088134.D	18 Apr 2025 16:13	AR\AJ	Ok,M
16	PTOXICC750	PD088135.D	18 Apr 2025 16:27	AR\AJ	Ok
17	PTOXICC500	PD088136.D	18 Apr 2025 16:40	AR\AJ	Ok
18	PTOXICC250	PD088137.D	18 Apr 2025 16:54	AR\AJ	Ok,M
19	PTOXICC100	PD088138.D	18 Apr 2025 17:08	AR\AJ	Ok,M
20	PSTDICV050	PD088139.D	18 Apr 2025 17:21	AR\AJ	Ok
21	PCHLORICV500	PD088140.D	18 Apr 2025 17:35	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD041825

Review By	Abdul	Review On	4/19/2025 2:07:56 PM
Supervise By	mohammad	Supervise On	4/22/2025 2:20:46 AM
SubDirectory	PD041825	HP Acquire Method	HP Processing Method pd041825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PTOXICV500	PD088141.D	18 Apr 2025 17:49	AR\AJ	Ok
23	I.BLK	PD088142.D	18 Apr 2025 18:02	AR\AJ	Ok
24	PEM	PD088143.D	18 Apr 2025 18:16	AR\AJ	Ok,M
25	PSTDCCC050	PD088144.D	18 Apr 2025 18:30	AR\AJ	Not Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD050125

Review By	Abdul	Review On	5/2/2025 8:29:41 AM
Supervise By	mohammad	Supervise On	5/5/2025 4:38:37 AM
SubDirectory	PD050125	HP Acquire Method	HP Processing Method pd041825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD088363.D	01 May 2025 09:28	AR\AJ	Ok
2	I.BLK	PD088364.D	01 May 2025 09:42	AR\AJ	Ok
3	PEM	PD088365.D	01 May 2025 09:55	AR\AJ	Ok,M
4	PSTDCCC050	PD088366.D	01 May 2025 11:21	AR\AJ	Ok
5	Q1910-01DL	PD088367.D	01 May 2025 11:42	AR\AJ	Ok,M
6	I.BLK	PD088368.D	01 May 2025 11:56	AR\AJ	Ok
7	PSTDCCC050	PD088369.D	01 May 2025 12:09	AR\AJ	Ok
8	PB167820BL	PD088370.D	01 May 2025 13:39	AR\AJ	Ok
9	PB167820BS	PD088371.D	01 May 2025 13:53	AR\AJ	Ok
10	PB167774TB	PD088372.D	01 May 2025 14:06	AR\AJ	Ok
11	PB167805TB	PD088373.D	01 May 2025 14:20	AR\AJ	Ok
12	Q1913-02	PD088374.D	01 May 2025 14:34	AR\AJ	Ok,M
13	Q1913-04	PD088375.D	01 May 2025 14:47	AR\AJ	Ok,M
14	Q1901-08	PD088376.D	01 May 2025 15:01	AR\AJ	Ok
15	Q1901-08MS	PD088377.D	01 May 2025 15:15	AR\AJ	Ok,M
16	Q1901-08MSD	PD088378.D	01 May 2025 15:28	AR\AJ	Ok,M
17	Q1901-09	PD088379.D	01 May 2025 15:42	AR\AJ	Ok
18	Q1901-10	PD088380.D	01 May 2025 15:56	AR\AJ	Ok
19	Q1901-11	PD088381.D	01 May 2025 16:10	AR\AJ	Ok
20	Q1907-02	PD088382.D	01 May 2025 16:23	AR\AJ	Ok
21	I.BLK	PD088383.D	01 May 2025 16:37	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD050125

Review By	Abdul	Review On	5/2/2025 8:29:41 AM
Supervise By	mohammad	Supervise On	5/5/2025 4:38:37 AM
SubDirectory	PD050125	HP Acquire Method	HP Processing Method pd041825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PSTDCCC050	PD088384.D	01 May 2025 17:19	AR\AJ	Ok,M
----	------------	------------	-------------------	-------	------

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD041825

Review By	Abdul	Review On	4/19/2025 2:07:56 PM
Supervise By	mohammad	Supervise On	4/22/2025 2:20:46 AM
SubDirectory	PD041825	HP Acquire Method	HP Processing Method pd041825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24095
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD088120.D	18 Apr 2025 13:02		AR\AJ	Ok
2	I.BLK	I.BLK	PD088121.D	18 Apr 2025 13:15		AR\AJ	Ok
3	PEM	PEM	PD088122.D	18 Apr 2025 13:29		AR\AJ	Ok,M
4	RESCHK	RESCHK	PD088123.D	18 Apr 2025 13:43		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PD088124.D	18 Apr 2025 13:56		AR\AJ	Ok,M
6	PSTDICC075	PSTDICC075	PD088125.D	18 Apr 2025 14:10		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PD088126.D	18 Apr 2025 14:24		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PD088127.D	18 Apr 2025 14:37		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PD088128.D	18 Apr 2025 14:51		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PD088129.D	18 Apr 2025 15:05		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PD088130.D	18 Apr 2025 15:18		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PD088131.D	18 Apr 2025 15:32		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PD088132.D	18 Apr 2025 15:46		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PD088133.D	18 Apr 2025 15:59		AR\AJ	Ok,M
15	PTOXICC1000	PTOXICC1000	PD088134.D	18 Apr 2025 16:13		AR\AJ	Ok,M
16	PTOXICC750	PTOXICC750	PD088135.D	18 Apr 2025 16:27		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PD088136.D	18 Apr 2025 16:40		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PD088137.D	18 Apr 2025 16:54		AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD041825

Review By	Abdul	Review On	4/19/2025 2:07:56 PM
Supervise By	mohammad	Supervise On	4/22/2025 2:20:46 AM
SubDirectory	PD041825	HP Acquire Method	HP Processing Method
			pd041825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PD088138.D	18 Apr 2025 17:08		AR\AJ	Ok,M
20	PSTDICV050	ICVPD041825	PD088139.D	18 Apr 2025 17:21		AR\AJ	Ok
21	PCHLORICV500	ICVPD041825CHLOR	PD088140.D	18 Apr 2025 17:35		AR\AJ	Ok
22	PTOXICV500	ICVPD041825TOX	PD088141.D	18 Apr 2025 17:49		AR\AJ	Ok
23	I.BLK	I.BLK	PD088142.D	18 Apr 2025 18:02		AR\AJ	Ok
24	PEM	PEM	PD088143.D	18 Apr 2025 18:16		AR\AJ	Ok,M
25	PSTDCCC050	PSTDCCC050	PD088144.D	18 Apr 2025 18:30	ccc high	AR\AJ	Not Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD050125

Review By	Abdul	Review On	5/2/2025 8:29:41 AM
Supervise By	mohammad	Supervise On	5/5/2025 4:38:37 AM
SubDirectory	PD050125	HP Acquire Method	HP Processing Method pd041825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24095
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD088363.D	01 May 2025 09:28		AR\AJ	Ok
2	I.BLK	I.BLK	PD088364.D	01 May 2025 09:42		AR\AJ	Ok
3	PEM	PEM	PD088365.D	01 May 2025 09:55		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PD088366.D	01 May 2025 11:21		AR\AJ	Ok
5	Q1910-01DL	MOO-25-0123-27DL	PD088367.D	01 May 2025 11:42		AR\AJ	Ok,M
6	I.BLK	I.BLK	PD088368.D	01 May 2025 11:56		AR\AJ	Ok
7	PSTDCCC050	PSTDCCC050	PD088369.D	01 May 2025 12:09		AR\AJ	Ok
8	PB167820BL	PB167820BL	PD088370.D	01 May 2025 13:39		AR\AJ	Ok
9	PB167820BS	PB167820BS	PD088371.D	01 May 2025 13:53		AR\AJ	Ok
10	PB167774TB	PB167774TB	PD088372.D	01 May 2025 14:06		AR\AJ	Ok
11	PB167805TB	PB167805TB	PD088373.D	01 May 2025 14:20		AR\AJ	Ok
12	Q1913-02	WC-12-A-202504	PD088374.D	01 May 2025 14:34		AR\AJ	Ok,M
13	Q1913-04	WC-13-A-202504	PD088375.D	01 May 2025 14:47		AR\AJ	Ok,M
14	Q1901-08	B-167-SB01	PD088376.D	01 May 2025 15:01		AR\AJ	Ok
15	Q1901-08MS	B-167-SB01MS	PD088377.D	01 May 2025 15:15		AR\AJ	Ok,M
16	Q1901-08MSD	B-167-SB01MSD	PD088378.D	01 May 2025 15:28		AR\AJ	Ok,M
17	Q1901-09	B-170-SB01	PD088379.D	01 May 2025 15:42		AR\AJ	Ok
18	Q1901-10	B-167-SB02	PD088380.D	01 May 2025 15:56		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD050125

Review By	Abdul	Review On	5/2/2025 8:29:41 AM
Supervise By	mohammad	Supervise On	5/5/2025 4:38:37 AM
SubDirectory	PD050125	HP Acquire Method	HP Processing Method pd041825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q1901-11	B-170-SB02	PD088381.D	01 May 2025 16:10		AR\AJ	Ok
20	Q1907-02	CO-8R-WC	PD088382.D	01 May 2025 16:23		AR\AJ	Ok
21	I.BLK	I.BLK	PD088383.D	01 May 2025 16:37		AR\AJ	Ok
22	PSTDCCC050	PSTDCCC050	PD088384.D	01 May 2025 17:19		AR\AJ	Ok,M

M : Manual Integration

A
B
C
D
E
F
G
H
I
J
K
L

SOP ID :	M1311-TCLP-15	
SDG No :	N/A	Start Prep Date : 04/29/2025 Time : 17:20
Welgh By :	JP	End Prep Date : 04/30/2025 Time : 11:30
Balance ID :	WC SC-7	Combination Ratio : 20
pH Meter ID :	WC PH METER-1	ZHE Cleaning Batch : ¹⁰ N/A
Extraction By :	JP	Initial Room Temperature: 23 °C
Filter By :	JP	Final Room Temperature: 22 °C
Pippete ID :	WC	TCLP Technician Signature : <i>[Signature]</i>
Tumbler ID :	T-1 / T-2	Supervisor By : <i>12</i>
TCLP Filter ID :	115525	

Standarded Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
TCLP-FLUID-1	N/A	WP112795
HCL-TCLP,1N	N/A	WP112797
HNO3-TCLP,1N	N/A	WP112799
pH Strips	N/A	W1931,W1934,W3171,W3172
pH Strips	W1940,W1941,W1942	W3166,W1938,W1939,
1 Liter Amber	N/A	90924-08
120ml Plastic bottle	N/A	2738
1:1 HNO3	N/A	MP84041

Extraction Conformance/Non-Conformance Comments:

Matrix spikes are added after filtration and before preservation. TUMBLER T-1 /T-2 checked,30 rpm. Particle size reduction is not required. q1894-01,02 and 03 we receive limited volume so no fluid determentation. Q1901-08 IS USFD FOR MS-MSD.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
04/30/25 13:00	<i>JP</i> <i>Lab Room</i>	<i>[Signature]</i> <i>RJL/EHL</i>
	Preparation Group	Analysis Group <i>[Signature]</i>

TCLP EXTRACTION LOGPAGE

PB167774

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB167774TB	LEB774	21	N/A	2000	N/A	N/A	N/A	4.93	1.0	T-2
Q1894-01	PAINTED-GROUND-PIPE--1	01	25.00	500	N/A	N/A	N/A	5.5	1.0	T-1
Q1894-02	PAINTED-GROUND-PIPE-2	02	35.00	700	N/A	N/A	N/A	5.6	1.5	T-1
Q1894-03	PAINTED-GROUND-PIPE-3	03	34.00	680	N/A	N/A	N/A	5.6	1.5	T-1
Q1901-08	B-167-SB01	04	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-1
Q1901-09	B-170-SB01	05	100.03	2000	N/A	N/A	N/A	7.0	1.5	T-1
Q1901-10	B-167-SB02	06	100.04	2000	N/A	N/A	N/A	5.8	1.0	T-1
Q1901-11	B-170-SB02	07	100.02	2000	N/A	N/A	N/A	5.6	1.5	T-1
Q1904-02	VNJ-210	08	100.03	2000	N/A	N/A	N/A	6.0	1.0	T-1
Q1905-04	MH-G	09	100.02	2000	N/A	N/A	N/A	3.0	1.0	T-1
Q1905-08	MH-H	10	100.01	2000	N/A	N/A	N/A	3.0	1.5	T-1
Q1906-04	WC-4	11	100.02	2000	N/A	N/A	N/A	5.5	1.0	T-2
Q1906-08	WC-5	12	100.03	2000	N/A	N/A	N/A	5.6	1.5	T-2
Q1906-12	WC-6	13	100.02	2000	N/A	N/A	N/A	5.0	1.0	T-2
Q1906-16	WC-7	14	100.03	2000	N/A	N/A	N/A	5.0	1.5	T-2
Q1907-02	CO-8R-WC	15	100.02	2000	N/A	N/A	N/A	5.8	1.5	T-2
Q1912-04	MH-E	16	100.03	2000	N/A	N/A	N/A	4.5	1.0	T-2
Q1912-08	MH-F	17	100.02	2000	N/A	N/A	N/A	4.0	1.5	T-2
Q1913-01	WC-12-S-202504	18	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-2
Q1913-03	WC-13-S-202504	19	100.03	2000	N/A	N/A	N/A	6.2	1.5	T-2
Q1915-01	WC-04282025	20	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-2

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
PB167774TB	LEB774	N/A	N/A	N/A	N/A	N/A	N/A
Q1894-01	PAINTED-GROUND-PIPE--1	N/A	N/A	N/A	N/A	100	N/A
Q1894-02	PAINTED-GROUND-PIPE-2	N/A	N/A	N/A	N/A	100	N/A
Q1894-03	PAINTED-GROUND-PIPE-3	N/A	N/A	N/A	N/A	100	N/A
Q1901-08	B-167-SB01	N/A	N/A	N/A	N/A	100	N/A
Q1901-09	B-170-SB01	N/A	N/A	N/A	N/A	100	N/A
Q1901-10	B-167-SB02	N/A	N/A	N/A	N/A	100	N/A
Q1901-11	B-170-SB02	N/A	N/A	N/A	N/A	100	N/A
Q1904-02	VNJ-210	N/A	N/A	N/A	N/A	100	N/A
Q1905-04	MH-G	N/A	N/A	N/A	N/A	100	N/A
Q1905-08	MH-H	N/A	N/A	N/A	N/A	100	N/A
Q1906-04	WC-4	N/A	N/A	N/A	N/A	100	N/A
Q1906-08	WC-5	N/A	N/A	N/A	N/A	100	N/A
Q1906-12	WC-6	N/A	N/A	N/A	N/A	100	N/A
Q1906-16	WC-7	N/A	N/A	N/A	N/A	100	N/A
Q1907-02	CO-8R-WC	N/A	N/A	N/A	N/A	100	N/A
Q1912-04	MH-E	N/A	N/A	N/A	N/A	100	N/A
Q1912-08	MH-F	N/A	N/A	N/A	N/A	100	N/A
Q1913-01	WC-12-S-202504	N/A	N/A	N/A	N/A	100	N/A
Q1913-03	WC-13-S-202504	N/A	N/A	N/A	N/A	100	N/A
Q1915-01	WC-04282025	N/A	N/A	N/A	N/A	100	N/A

Hot Block ID : WC S-1 / WC S-2

Thermometer ID : FLASHPOINT

SampleID	ClientID	Sample Weight (g)	Volume DI Water (mL)	PH after 5 min stir	PH after 10 min stir	Extraction Fluid 1 or 2	pH Extraction Fluid
PB167774TB	LEB774	N/A	N/A	N/A	N/A	N/A	N/A
Q1894-01	PAINTED-GROUND-PIPE--1	N/A	N/A	N/A	N/A	N/A	N/A
Q1894-02	PAINTED-GROUND-PIPE-2	N/A	N/A	N/A	N/A	N/A	N/A
Q1894-03	PAINTED-GROUND-PIPE-3	N/A	N/A	N/A	N/A	N/A	N/A
Q1901-08	B-167-SB01	5.02	96.5	8.4	3.0	#1	4.93
Q1901-09	B-170-SB01	5.01	96.5	9.0	4.0	#1	4.93
Q1901-10	B-167-SB02	5.02	96.5	7.0	3.5	#1	4.93
Q1901-11	B-170-SB02	5.03	96.5	8.2	3.5	#1	4.93
Q1904-02	VNJ-210	5.04	96.5	8.4	3.0	#1	4.93
Q1905-04	MH-G	5.03	96.5	5.6	1.5	#1	4.93
Q1905-08	MH-H	5.02	96.5	6.0	2.0	#1	4.93
Q1906-04	WC-4	5.03	96.5	7.6	2.5	#1	4.93
Q1906-08	WC-5	5.02	96.5	7.0	2.5	#1	4.93
Q1906-12	WC-6	5.01	96.5	6.6	2.0	#1	4.93
Q1906-16	WC-7	5.02	96.5	6.6	2.0	#1	4.93
Q1907-02	CO-8R-WC	5.02	96.5	8.4	3.0	#1	4.93
Q1912-04	MH-E	5.01	96.5	6.6	2.0	#1	4.93
Q1912-08	MH-F	5.02	96.5	6.4	2.0	#1	4.93
Q1913-01	WC-12-S-202504	5.02	96.5	8.2	3.5	#1	4.93
Q1913-03	WC-13-S-202504	5.01	96.5	8.0	3.0	#1	4.93
Q1915-01	WC-04282025	5.02	96.5	8.4	3.5	#1	4.93

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TCLP Q1907 WorkList ID : 189188 Department : TCLP Extraction Date : 04-29-2025 07:58:37

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1913-01	WC-12-S-202504	Solid	TCLP Extraction	Cool 4 deg C	AECO02	L21	04/29/2025	1311
Q1913-03	WC-13-S-202504	Solid	TCLP Extraction	Cool 4 deg C	AECO02	L21	04/29/2025	1311
Q1915-01	WC-04282025	Solid	TCLP Extraction	Cool 4 deg C	CAMP02	L41	04/28/2025	1311
Q1901-08	B-167-SB01	Solid	TCLP Extraction	Cool 4 deg C	PORT06	L51	04/26/2025	1311
Q1901-09	B-170-SB01	Solid	TCLP Extraction	Cool 4 deg C	PORT06	L51	04/26/2025	1311
Q1901-10	B-167-SB02	Solid	TCLP Extraction	Cool 4 deg C	PORT06	L51	04/26/2025	1311
Q1901-11	B-170-SB02	Solid	TCLP Extraction	Cool 4 deg C	PORT06	L51	04/26/2025	1311
Q1894-01	PAINTED-GROUND-PIPE--1	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L31	04/28/2025	1311
Q1894-02	PAINTED-GROUND-PIPE-2	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L31	04/28/2025	1311
Q1894-03	PAINTED-GROUND-PIPE-3	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L31	04/28/2025	1311
Q1904-02	VNJ-210	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L41	04/28/2025	1311
Q1905-04	MH-G	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L51	04/28/2025	1311
Q1905-08	MH-H	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L51	04/28/2025	1311
Q1906-04	WC-4	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L41	04/28/2025	1311
Q1906-08	WC-5	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L41	04/28/2025	1311
Q1906-12	WC-6	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L41	04/28/2025	1311
Q1906-16	WC-7	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L41	04/28/2025	1311
Q1912-04	MH-E	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L51	04/29/2025	1311
Q1912-08	MH-F	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L51	04/29/2025	1311
Q1907-02	CO-8R-WC	Solid	TCLP Extraction	Cool 4 deg C	WALS01		04/28/2025	1311


Date/Time 04/29/25 17:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 04/29/25 19:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Analytical Method: M3510C,3580A-Extraction Pesticide-16

Concentration Date: 05/01/2025

Sample ID	Client Sample ID	Test	g / (mL)	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB167774TB	PB167774TB	TCLP Pesticide	100	6	RUPESH	Evelyn	10			SEP-1
PB167805TB	PB167805TB	TCLP Pesticide	100	6	RUPESH	Evelyn	10			2
PB167820BL	PBLK820	TCLP Pesticide	1000	6	RUPESH	Evelyn	10			3
PB167820BS	PLCS820	TCLP Pesticide	1000	6	RUPESH	Evelyn	10			4
Q1901-08	B-167-SB01	TCLP Pesticide	100	6	RUPESH	Evelyn	10	A		5
Q1901-08MS	B-167-SB01MS	TCLP Pesticide	100	6	RUPESH	Evelyn	10	A		6
Q1901-08MS D	B-167-SB01MSD	TCLP Pesticide	100	6	RUPESH	Evelyn	10	A		7
Q1901-09	B-170-SB01	TCLP Pesticide	100	6	RUPESH	Evelyn	10	A		8
Q1901-10	B-167-SB02	TCLP Pesticide	100	6	RUPESH	Evelyn	10	A		9
Q1901-11	B-170-SB02	TCLP Pesticide	100	6	RUPESH	Evelyn	10	A		10
Q1907-02	CO-8R-WC	TCLP Pesticide	100	6	RUPESH	Evelyn	10	A		11
Q1913-02	WC-12-A-202504	TCLP Pesticide	100	6	RUPESH	Evelyn	10	A		12
Q1913-04	WC-13-A-202504	TCLP Pesticide	100	6	RUPESH	Evelyn	10	A		13



TCLP EXTRACTION LOGPAGE

PB167774

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	PB Pos
PB167774TB	LEB774	21	N/A	2000	N/A	N/A	N/A	4.93	1.0	T-1
Q1894-01	PAINTED-GROUND-PIPE--1	01	25.00	500	N/A	N/A	N/A	5.5	1.0	T-1
Q1894-02	PAINTED-GROUND-PIPE-2	02	35.00	700	N/A	N/A	N/A	5.6	1.5	T-1
Q1894-03	PAINTED-GROUND-PIPE-3	03	34.00	680	N/A	N/A	N/A	5.6	1.5	T-1
Q1901-08	B-167-SB01	04	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-1
Q1901-09	B-170-SB01	05	100.03	2000	N/A	N/A	N/A	7.0	1.5	T-1
Q1901-10	B-167-SB02	06	100.04	2000	N/A	N/A	N/A	5.8	1.0	T-1
Q1901-11	B-170-SB02	07	100.02	2000	N/A	N/A	N/A	5.6	1.5	T-1
Q1904-02	VNJ-210	08	100.03	2000	N/A	N/A	N/A	6.0	1.0	T-1
Q1905-04	MH-G	09	100.02	2000	N/A	N/A	N/A	3.0	1.0	T-1
Q1905-08	MH-H	10	100.01	2000	N/A	N/A	N/A	3.0	1.5	T-1
Q1906-04	WC-4	11	100.02	2000	N/A	N/A	N/A	5.5	1.0	T-2
Q1906-08	WC-5	12	100.03	2000	N/A	N/A	N/A	5.6	1.5	T-2
Q1906-12	WC-6	13	100.02	2000	N/A	N/A	N/A	5.0	1.0	T-2
Q1906-16	WC-7	14	100.03	2000	N/A	N/A	N/A	5.0	1.5	T-2
Q1907-02	CO-8R-WC	15	100.02	2000	N/A	N/A	N/A	5.8	1.5	T-2
Q1912-04	MH-E	16	100.03	2000	N/A	N/A	N/A	4.5	1.0	T-2
Q1912-08	MH-F	17	100.02	2000	N/A	N/A	N/A	4.0	1.5	T-2
Q1913-01	WC-12-S-202504	18	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-2
Q1913-03	WC-13-S-202504	19	100.03	2000	N/A	N/A	N/A	6.2	1.5	T-2
Q1915-01	WC-04282025	20	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-2

04/30/25
13:00

TCLP EXTRACTION LOGPAGE

PB167805

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Pre Po.
PB167805TB	LEB805	N/A	N/A	N/A	N/A	N/A	N/A	4.93	1.0	N/A
Q1913-02	WC-12-A-202504	N/A	N/A	N/A	N/A	N/A	N/A	12.0	1.5	N/A
Q1913-04	WC-13-A-202504	N/A	N/A	N/A	N/A	N/A	N/A	12.0	1.5	N/A

04/30/25
131.00

LAB CHRONICLE

OrderID:	Q1901	OrderDate:	4/28/2025 10:24:00 AM
Client:	Portal Partners Tri-Venture	Project:	Amtrak Sawtooth Bridges 2025
Contact:	Joseph Krupansky	Location:	L51,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1901-02	B-167-SB01	SOIL	TPH GC	8015D	04/26/25	04/29/25	04/29/25	04/28/25
Q1901-03	B-170-SB01	SOIL	TPH GC	8015D	04/26/25	04/29/25	04/29/25	04/28/25
Q1901-04	B-167-SB02	SOIL	TPH GC	8015D	04/26/25	04/29/25	04/29/25	04/28/25
Q1901-05	B-170-SB02	SOIL	TPH GC	8015D	04/26/25	04/29/25	04/29/25	04/28/25
Q1901-08	B-167-SB01	TCLP	TCLP Pesticide	8081B	04/26/25	05/01/25	05/01/25	04/28/25
Q1901-09	B-170-SB01	TCLP	TCLP Pesticide	8081B	04/26/25	05/01/25	05/01/25	04/28/25
Q1901-10	B-167-SB02	TCLP	TCLP Pesticide	8081B	04/26/25	05/01/25	05/01/25	04/28/25
Q1901-11	B-170-SB02	TCLP	TCLP Pesticide	8081B	04/26/25	05/01/25	05/01/25	04/28/25