

Instrument ID: BNA\_F

**Daily Analysis Runlog For Sequence/QC Batch ID # BF043025**

|                          |          |                                                         |                     |                      |          |
|--------------------------|----------|---------------------------------------------------------|---------------------|----------------------|----------|
| Review By                | Rahul    | Review On                                               | 5/1/2025 1:20:20 PM |                      |          |
| Supervise By             | Jagrut   | Supervise On                                            | 5/1/2025 1:23:08 PM |                      |          |
| SubDirectory             | BF043025 | HP Acquire Method                                       | BNA_F               | HP Processing Method | BF043025 |
| STD. NAME                |          | STD REF.#                                               |                     |                      |          |
| Tune/Reschk              |          | SP6757                                                  |                     |                      |          |
| Initial Calibration Stds |          | SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729 |                     |                      |          |
| CCC                      |          | SP6725                                                  |                     |                      |          |
| Internal Standard/PEM    |          | S12661,10ul/1000ul sample                               |                     |                      |          |
| ICV/I.BLK                |          | SP6770                                                  |                     |                      |          |
| Surrogate Standard       |          |                                                         |                     |                      |          |
| MS/MSD Standard          |          |                                                         |                     |                      |          |
| LCS Standard             |          |                                                         |                     |                      |          |

| Sr# | SampleID    | ClientID    | Data File Name | Date-Time         | Comment                                           | Operator | Status |
|-----|-------------|-------------|----------------|-------------------|---------------------------------------------------|----------|--------|
| 1   | DFTPP       | DFTPP       | BF142238.D     | 30 Apr 2025 10:55 |                                                   | RC/JU    | Ok     |
| 2   | SSTDICC2.5  | SSTDICC2.5  | BF142239.D     | 30 Apr 2025 11:24 |                                                   | RC/JU    | Ok     |
| 3   | SSTDICC005  | SSTDICC005  | BF142240.D     | 30 Apr 2025 11:52 | Compound#32,54,56,65,70,85 removed from 5 ppm     | RC/JU    | Ok,M   |
| 4   | SSTDICC010  | SSTDICC010  | BF142241.D     | 30 Apr 2025 12:20 |                                                   | RC/JU    | Ok     |
| 5   | SSTDICC020  | SSTDICC020  | BF142242.D     | 30 Apr 2025 12:49 |                                                   | RC/JU    | Ok,M   |
| 6   | SSTDICCC040 | SSTDICCC040 | BF142243.D     | 30 Apr 2025 13:17 | Compound#32,48,51,53,57,62,65,80,84,85 Kept on LR | RC/JU    | Ok,M   |
| 7   | SSTDICC050  | SSTDICC050  | BF142244.D     | 30 Apr 2025 13:46 | Compound#26,54 Kept on QR                         | RC/JU    | Ok,M   |
| 8   | SSTDICC060  | SSTDICC060  | BF142245.D     | 30 Apr 2025 14:15 |                                                   | RC/JU    | Ok,M   |
| 9   | SSTDICC080  | SSTDICC080  | BF142246.D     | 30 Apr 2025 14:43 |                                                   | RC/JU    | Ok,M   |
| 10  | SSTDICV040  | ICVBF043025 | BF142247.D     | 30 Apr 2025 16:17 |                                                   | RC/JU    | Ok,M   |
| 11  | PB167726BL  | PB167726BL  | BF142248.D     | 30 Apr 2025 17:15 |                                                   | RC/JU    | Ok     |

M : Manual Integration

Instrument ID: BNA\_F

**Daily Analysis Runlog For Sequence/QC Batch ID # BF050125**

|              |          |                   |                      |                      |          |
|--------------|----------|-------------------|----------------------|----------------------|----------|
| Review By    | Rahul    | Review On         | 5/2/2025 10:46:38 AM |                      |          |
| Supervise By | Jagrut   | Supervise On      | 5/2/2025 4:49:12 PM  |                      |          |
| SubDirectory | BF050125 | HP Acquire Method | BNA_F                | HP Processing Method | BF043025 |

| STD. NAME                | STD REF.#                                               |
|--------------------------|---------------------------------------------------------|
| Tune/Reschk              | SP6757                                                  |
| Initial Calibration Stds | SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729 |
| CCC                      | SP6725                                                  |
| Internal Standard/PEM    | S12661,10ul/1000ul sample                               |
| ICV/I.BLK                | SP6770                                                  |
| Surrogate Standard       |                                                         |
| MS/MSD Standard          |                                                         |
| LCS Standard             |                                                         |

| Sr# | SampleID    | ClientID       | Data File Name | Date-Time         | Comment | Operator | Status |
|-----|-------------|----------------|----------------|-------------------|---------|----------|--------|
| 1   | DFTPP       | DFTPP          | BF142249.D     | 01 May 2025 09:48 |         | RC/JU    | Ok     |
| 2   | SSTDCCC040  | SSTDCCC040     | BF142250.D     | 01 May 2025 10:17 |         | RC/JU    | Ok,M   |
| 3   | PB167798BL  | PB167798BL     | BF142251.D     | 01 May 2025 10:45 |         | RC/JU    | Ok     |
| 4   | PB167798BS  | PB167798BS     | BF142252.D     | 01 May 2025 11:13 |         | RC/JU    | Ok,M   |
| 5   | PB167798BSD | PB167798BSD    | BF142253.D     | 01 May 2025 11:42 |         | RC/JU    | Ok,M   |
| 6   | PB167810BL  | PB167810BL     | BF142254.D     | 01 May 2025 12:10 |         | RC/JU    | Ok     |
| 7   | PB167810BS  | PB167810BS     | BF142255.D     | 01 May 2025 12:39 |         | RC/JU    | Ok,M   |
| 8   | PB167774TB  | PB167774TB     | BF142256.D     | 01 May 2025 13:07 |         | RC/JU    | Ok     |
| 9   | Q1890-01    | TW-WTS-07      | BF142257.D     | 01 May 2025 13:40 |         | RC/JU    | Ok     |
| 10  | Q1920-01    | 291            | BF142258.D     | 01 May 2025 14:09 |         | RC/JU    | Ok,M   |
| 11  | Q1914-14    | FB04282025     | BF142259.D     | 01 May 2025 14:37 |         | RC/JU    | Ok     |
| 12  | Q1913-04    | WC-13-A-202504 | BF142260.D     | 01 May 2025 15:06 |         | RC/JU    | Ok     |
| 13  | Q1915-01    | WC-04282025    | BF142261.D     | 01 May 2025 15:34 |         | RC/JU    | Ok     |
| 14  | Q1903-01    | COMP-4         | BF142262.D     | 01 May 2025 16:03 |         | RC/JU    | Ok     |
| 15  | Q1903-02    | COMP-5         | BF142263.D     | 01 May 2025 16:31 |         | RC/JU    | Ok     |
| 16  | Q1903-03    | COMP-6         | BF142264.D     | 01 May 2025 17:00 |         | RC/JU    | Ok     |
| 17  | Q1906-01    | WC-4           | BF142265.D     | 01 May 2025 17:29 |         | RC/JU    | Ok     |
| 18  | Q1906-13    | WC-7           | BF142266.D     | 01 May 2025 17:57 |         | RC/JU    | Ok     |

Instrument ID: BNA\_F

**Daily Analysis Runlog For Sequence/QC Batch ID # BF050125**

| Review By                | Rahul                                                   | Review On         | 5/2/2025 10:46:38 AM |                      |          |
|--------------------------|---------------------------------------------------------|-------------------|----------------------|----------------------|----------|
| Supervise By             | Jagrut                                                  | Supervise On      | 5/2/2025 4:49:12 PM  |                      |          |
| SubDirectory             | BF050125                                                | HP Acquire Method | BNA_F                | HP Processing Method | BF043025 |
| STD. NAME                | STD REF.#                                               |                   |                      |                      |          |
| Tune/Reschk              | SP6757                                                  |                   |                      |                      |          |
| Initial Calibration Stds | SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729 |                   |                      |                      |          |
| CCC                      | SP6725                                                  |                   |                      |                      |          |
| Internal Standard/PEM    | S12661,10ul/1000ul sample                               |                   |                      |                      |          |
| ICV/I.BLK                | SP6770                                                  |                   |                      |                      |          |
| Surrogate Standard       |                                                         |                   |                      |                      |          |
| MS/MSD Standard          |                                                         |                   |                      |                      |          |
| LCS Standard             |                                                         |                   |                      |                      |          |

|    |             |               |            |                   |  |       |      |
|----|-------------|---------------|------------|-------------------|--|-------|------|
| 19 | Q1901-06    | FB04262025    | BF142267.D | 01 May 2025 18:26 |  | RC/JU | Ok   |
| 20 | Q1901-02    | B-167-SB01    | BF142268.D | 01 May 2025 18:55 |  | RC/JU | Ok,M |
| 21 | Q1901-02MS  | B-167-SB01MS  | BF142269.D | 01 May 2025 19:24 |  | RC/JU | Ok,M |
| 22 | Q1901-02MSD | B-167-SB01MSD | BF142270.D | 01 May 2025 19:52 |  | RC/JU | Ok,M |
| 23 | Q1901-04    | B-167-SB02    | BF142271.D | 01 May 2025 20:20 |  | RC/JU | Ok   |
| 24 | Q1901-05    | B-170-SB02    | BF142272.D | 01 May 2025 20:49 |  | RC/JU | Ok   |

M : Manual Integration

Instrument ID: BNA\_M

**Daily Analysis Runlog For Sequence/QC Batch ID # BM042825**

|              |          |                      |                       |
|--------------|----------|----------------------|-----------------------|
| Review By    | Rahul    | Review On            | 4/29/2025 8:56:26 AM  |
| Supervise By | Jagrut   | Supervise On         | 4/29/2025 11:57:30 AM |
| SubDirectory | BM042825 | HP Acquire Method    | BNA_M                 |
|              |          | HP Processing Method | BM042825              |

| STD. NAME                | STD REF.#                                               |
|--------------------------|---------------------------------------------------------|
| Tune/Reschk              | SP6757                                                  |
| Initial Calibration Stds | SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729 |
| CCC                      | SP6725                                                  |
| Internal Standard/PEM    | S12661,10ul/1000ul sample                               |
| ICV/I.BLK                | SP6770                                                  |
| Surrogate Standard       |                                                         |
| MS/MSD Standard          |                                                         |
| LCS Standard             |                                                         |

| Sr# | SampleID    | ClientID    | Data File Name | Date-Time         | Comment                                              | Operator | Status |
|-----|-------------|-------------|----------------|-------------------|------------------------------------------------------|----------|--------|
| 1   | DFTPP       | DFTPP       | BM050023.D     | 28 Apr 2025 11:46 |                                                      | RC/JU    | Ok     |
| 2   | SSTDICC2.5  | SSTDICC2.5  | BM050024.D     | 28 Apr 2025 12:30 |                                                      | RC/JU    | Ok     |
| 3   | SSTDICC005  | SSTDICC005  | BM050025.D     | 28 Apr 2025 13:09 | Compound#32-41-54-56-65-70-7<br>7 removed from 5 ppm | RC/JU    | Ok,M   |
| 4   | SSTDICC010  | SSTDICC010  | BM050026.D     | 28 Apr 2025 13:48 |                                                      | RC/JU    | Ok,M   |
| 5   | SSTDICC020  | SSTDICC020  | BM050027.D     | 28 Apr 2025 14:27 |                                                      | RC/JU    | Ok,M   |
| 6   | SSTDICCC040 | SSTDICCC040 | BM050028.D     | 28 Apr 2025 15:06 | Compound#54 & 56 Kept on LR                          | RC/JU    | Ok     |
| 7   | SSTDICC050  | SSTDICC050  | BM050029.D     | 28 Apr 2025 15:45 |                                                      | RC/JU    | Ok     |
| 8   | SSTDICC060  | SSTDICC060  | BM050030.D     | 28 Apr 2025 16:24 |                                                      | RC/JU    | Ok     |
| 9   | SSTDICC080  | SSTDICC080  | BM050031.D     | 28 Apr 2025 17:04 |                                                      | RC/JU    | Ok     |
| 10  | SSTDICV040  | ICVBM042825 | BM050032.D     | 28 Apr 2025 17:43 |                                                      | RC/JU    | Ok,M   |
| 11  | PB167739BL  | PB167739BL  | BM050033.D     | 28 Apr 2025 19:40 | Analyzed for contamination<br>check                  | RC/JU    | Not Ok |

M : Manual Integration

Instrument ID: BNA\_M

**Daily Analysis Runlog For Sequence/QC Batch ID # BM050125**

|              |          |                      |                     |
|--------------|----------|----------------------|---------------------|
| Review By    | Rahul    | Review On            | 5/1/2025 3:03:06 PM |
| Supervise By | Jagrut   | Supervise On         | 5/2/2025 4:57:09 PM |
| SubDirectory | BM050125 | HP Acquire Method    | BNA_M               |
|              |          | HP Processing Method | BM042825            |

| STD. NAME                | STD REF.#                                               |
|--------------------------|---------------------------------------------------------|
| Tune/Reschk              | SP6757                                                  |
| Initial Calibration Stds | SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729 |
| CCC                      | SP6725                                                  |
| Internal Standard/PEM    | S12662,10ul/1000ul sample                               |
| ICV/I.BLK                | SP6770                                                  |
| Surrogate Standard       |                                                         |
| MS/MSD Standard          |                                                         |
| LCS Standard             |                                                         |

| Sr# | SampleID   | ClientID   | Data File Name | Date-Time         | Comment                  | Operator | Status   |
|-----|------------|------------|----------------|-------------------|--------------------------|----------|----------|
| 1   | DFTPP      | DFTPP      | BM050061.D     | 01 May 2025 12:24 |                          | RC/JU    | Ok       |
| 2   | SSTDCCC040 | SSTDCCC040 | BM050062.D     | 01 May 2025 13:25 |                          | RC/JU    | Ok       |
| 3   | PB167779BL | PB167779BL | BM050063.D     | 01 May 2025 14:04 |                          | RC/JU    | Ok       |
| 4   | PB167779BS | PB167779BS | BM050064.D     | 01 May 2025 14:43 |                          | RC/JU    | Ok       |
| 5   | Q1900-05   | WC-2       | BM050065.D     | 01 May 2025 15:26 |                          | RC/JU    | Ok       |
| 6   | Q1914-01   | SS-1       | BM050066.D     | 01 May 2025 16:05 |                          | RC/JU    | Ok       |
| 7   | Q1914-02   | SS-91      | BM050067.D     | 01 May 2025 16:44 |                          | RC/JU    | Ok       |
| 8   | Q1905-05   | MH-H       | BM050068.D     | 01 May 2025 17:23 |                          | RC/JU    | Ok,M     |
| 9   | Q1900-01   | WC-1       | BM050069.D     | 01 May 2025 18:02 |                          | RC/JU    | Ok,M     |
| 10  | Q1906-05   | WC-5       | BM050070.D     | 01 May 2025 18:41 |                          | RC/JU    | Ok,M     |
| 11  | Q1914-03   | SS-2       | BM050071.D     | 01 May 2025 19:20 |                          | RC/JU    | Ok       |
| 12  | Q1914-05   | SS-4       | BM050072.D     | 01 May 2025 20:00 |                          | RC/JU    | Ok,M     |
| 13  | Q1907-01   | CO-8R-WC   | BM050073.D     | 01 May 2025 20:39 | Need further 5X Dilution | RC/JU    | Dilution |
| 14  | Q1901-03   | B-170-SB01 | BM050074.D     | 01 May 2025 21:18 |                          | RC/JU    | Ok,M     |
| 15  | Q1900-09   | WC-3       | BM050075.D     | 01 May 2025 21:57 |                          | RC/JU    | Ok,M     |
| 16  | Q1905-01   | MH-G       | BM050076.D     | 01 May 2025 22:36 |                          | RC/JU    | Ok,M     |
| 17  | Q1906-09   | WC-6       | BM050077.D     | 01 May 2025 23:15 | Need Straight Run        | RC/JU    | Not Ok   |

M : Manual Integration