

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM

SAS No.: Q1901

SDG NO.: Q1901

Lab File ID: BF142249.D

DFTPP Injection Date: 05/01/2025

Instrument ID: BNA_F

DFTPP Injection Time: 09:48

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	39.4
68	Less than 2.0% of mass 69	0.5 (1.6) 1
69	Mass 69 relative abundance	31.6
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	42.4
197	Less than 2.0% of mass 198	0.6
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.9
275	10.0 - 60.0% of mass 198	23.8
365	Greater than 1% of mass 198	3.1
441	Present, but less than mass 443	15.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.1 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142250.D	05/01/2025	10:17
PB167798BL	PB167798BL	BF142251.D	05/01/2025	10:45
PB167798BS	PB167798BS	BF142252.D	05/01/2025	11:13
PB167798BSD	PB167798BSD	BF142253.D	05/01/2025	11:42
FB04262025	Q1901-06	BF142267.D	05/01/2025	18:26
B-167-SB01	Q1901-02	BF142268.D	05/01/2025	18:55
B-167-SB01MS	Q1901-02MS	BF142269.D	05/01/2025	19:24
B-167-SB01MSD	Q1901-02MSD	BF142270.D	05/01/2025	19:52
B-167-SB02	Q1901-04	BF142271.D	05/01/2025	20:20
B-170-SB02	Q1901-05	BF142272.D	05/01/2025	20:49