284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Cover Page

Order ID : Q1901

Project ID : Amtrak Sawtooth Bridges 2025

Client : Portal Partners Tri-Venture

Lab Sample Number	Client Sample Number
Q1901-01	B-170-SB00
Q1901-02	B-167-SB01
Q1901-03	B-170-SB01
Q1901-04	B-167-SB02
Q1901-05	B-170-SB02
Q1901-06	FB04262025
Q1901-07	TB04262025
Q1901-08	B-167-SB01
Q1901-09	B-170-SB01
Q1901-10	B-167-SB02
Q1901-11	B-170-SB02

I certify that the data package is in compliance with the terms and conditions of the contract,both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 5/6/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

284 Sheffield Street, Mountainside NJ 07092 Tel. 908-789-8900 Fax: 908-789-8922

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following " Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value**.** This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This is flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: **Q1901**

	Completed
For thorough review, the report must have the following:	
GENERAL:	
Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)	✓
Check chain-of-custody for proper relinquish/return of samples	✓
Is the chain of custody signed and complete	✓
Check internal chain-of-custody for proper relinquish/return of samples /sample extracts	✓
Collect information for each project id from server. Were all requirements followed	✓
COVER PAGE:	
Do numbers of samples correspond to the number of samples in the Chain of Custody on login page	✓
Do lab numbers and client Ids on cover page agree with the Chain of Custody	✓
CHAIN OF CUSTODY:	
Do requested analyses on Chain of Custody agree with form I results	✓
Do requested analyses on Chain of Custody agree with the log-in page	✓
Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody	✓
Were the samples received within hold time	✓
Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle	✓
ANALYTICAL:	
Was method requirement followed?	✓
Was client requirement followed?	✓
Does the case narrative summarize all QC failure?	✓
All runlogs and manual integration are reviewed for requirements	✓
All manual calculations and /or hand notations verified	✓

QA Review Signature: **MAHESH PATEL** **Date:** **05/06/2025**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LAB CHRONICLE

OrderID:	Q1901	**OrderDate:**	4/28/2025 10:24:00 AM
Client:	Portal Partners Tri-Venture	**Project:**	Amtrak Sawtooth Bridges 2025
Contact:	Joseph Krupansky	**Location:**	L51,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1901-02	**B-167-SB01**	**SOIL**			**04/26/25**			**04/28/25**
			SVOC-TCL BNA -20	8270E		04/29/25	05/01/25	
Q1901-03	**B-170-SB01**	**SOIL**			**04/26/25**			**04/28/25**
			SVOC-TCL BNA -20	8270E		04/29/25	05/01/25	
Q1901-04	**B-167-SB02**	**SOIL**			**04/26/25**			**04/28/25**
			SVOC-TCL BNA -20	8270E		04/29/25	05/01/25	
Q1901-05	**B-170-SB02**	**SOIL**			**04/26/25**			**04/28/25**
			SVOC-TCL BNA -20	8270E		04/29/25	05/01/25	
Q1901-06	**FB04262025**	**Water**			**04/26/25**			**04/28/25**
			SVOC-TCL BNA -20	8270E		04/30/25	05/01/25	
Q1901-08	**B-167-SB01**	**TCLP**			**04/26/25**			**04/28/25**
			TCLP BNA	8270E		04/30/25	05/02/25	
Q1901-09	**B-170-SB01**	**TCLP**			**04/26/25**			**04/28/25**
			TCLP BNA	8270E		04/30/25	05/02/25	
Q1901-10	**B-167-SB02**	**TCLP**			**04/26/25**			**04/28/25**
			TCLP BNA	8270E		04/30/25	05/02/25	
Q1901-10RE	**B-167-SB02RE**	**TCLP**			**04/26/25**			**04/28/25**
			TCLP BNA	8270E		04/30/25	05/05/25	
Q1901-11	**B-170-SB02**	**TCLP**			**04/26/25**			**04/28/25**
			TCLP BNA	8270E		04/30/25	05/02/25	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :	**B-167-SB01**							
Q1901-08	B-167-SB01	TCLP	Pyridine	24.100	J	12.8	50	ug/L
			Total Svoc :	**24.10**				
			Total Concentration:	**24.10**				



QC SUMMARY

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%) Low	Limits (%) High
PB167774TB	PB167774TB	2-Fluorophenol	150	116	78		15 (10)	110 (139)
		Phenol-d6	150	116	77		15 (10)	110 (134)
		Nitrobenzene-d5	100	87.3	87		30 (49)	130 (133)
		2-Fluorobiphenyl	100	72.1	72		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	131	87		15 (44)	110 (137)
		Terphenyl-d14	100	67.9	68		30 (48)	130 (125)
PB167810BL	PB167810BL	2-Fluorophenol	150	110	73		15 (10)	110 (139)
		Phenol-d6	150	109	73		15 (10)	110 (134)
		Nitrobenzene-d5	100	83.3	83		30 (49)	130 (133)
		2-Fluorobiphenyl	100	69.1	69		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	124	83		15 (44)	110 (137)
		Terphenyl-d14	100	63.3	63		30 (48)	130 (125)
PB167810BS	PB167810BS	2-Fluorophenol	150	111	74		15 (10)	110 (139)
		Phenol-d6	150	113	75		15 (10)	110 (134)
		Nitrobenzene-d5	100	82.4	82		30 (49)	130 (133)
		2-Fluorobiphenyl	100	68.5	69		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	132	88		15 (44)	110 (137)
		Terphenyl-d14	100	74.1	74		30 (48)	130 (125)
Q1901-08	B-167-SB01	2-Fluorophenol	150	103	69		15 (10)	110 (139)
		Phenol-d6	150	95.9	64		15 (10)	110 (134)
		Nitrobenzene-d5	100	88.3	88		30 (49)	130 (133)
		2-Fluorobiphenyl	100	69.9	70		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	139	92		15 (44)	110 (137)
		Terphenyl-d14	100	68.1	68		30 (48)	130 (125)
Q1901-08MS	B-167-SB01MS	2-Fluorophenol	150	96.6	64		15 (10)	110 (139)
		Phenol-d6	150	89.8	60		15 (10)	110 (134)
		Nitrobenzene-d5	100	84.0	84		30 (49)	130 (133)
		2-Fluorobiphenyl	100	70.7	71		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	126	84		15 (44)	110 (137)
		Terphenyl-d14	100	70.4	70		30 (48)	130 (125)
Q1901-08MSD	B-167-SB01MSD	2-Fluorophenol	150	99.4	66		15 (10)	110 (139)
		Phenol-d6	150	92.8	62		15 (10)	110 (134)
		Nitrobenzene-d5	100	87.8	88		30 (49)	130 (133)
		2-Fluorobiphenyl	100	72.3	72		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	133	88		15 (44)	110 (137)
		Terphenyl-d14	100	72.8	73		30 (48)	130 (125)
Q1901-09	B-170-SB01	2-Fluorophenol	150	101	67		15 (10)	110 (139)
		Phenol-d6	150	92.9	62		15 (10)	110 (134)
		Nitrobenzene-d5	100	87.7	88		30 (49)	130 (133)
		2-Fluorobiphenyl	100	70.7	71		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	130	87		15 (44)	110 (137)
		Terphenyl-d14	100	72.5	72		30 (48)	130 (125)
Q1901-10	B-167-SB02	2-Fluorophenol	150	101	67		15 (10)	110 (139)
		Phenol-d6	150	70.4	47		15 (10)	110 (134)
		Nitrobenzene-d5	100	90.2	90		30 (49)	130 (133)
		2-Fluorobiphenyl	100	114	114		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	212	141	*	15 (44)	110 (137)
		Terphenyl-d14	100	181	181	*	30 (48)	130 (125)
Q1901-10RE	B-167-SB02RE	2-Fluorophenol	150	115	77		15 (10)	110 (139)
		Phenol-d6	150	41.6	28		15 (10)	110 (134)
		Nitrobenzene-d5	100	93.0	93		30 (49)	130 (133)

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%) Low	Limits (%) High
Q1901-10RE	B-167-SB02RE	2-Fluorobiphenyl	100	314	313	*	30 (52)	130 (132)
		2,4,6-Tribromophenol	150	554	369	*	15 (44)	110 (137)
		Terphenyl-d14	100	902	902	*	30 (48)	130 (125)
Q1901-11	B-170-SB02	2-Fluorophenol	150	94.4	63		15 (10)	110 (139)
		Phenol-d6	150	88.5	59		15 (10)	110 (134)
		Nitrobenzene-d5	100	79.5	79		30 (49)	130 (133)
		2-Fluorobiphenyl	100	67.6	68		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	120	80		15 (44)	110 (137)
		Terphenyl-d14	100	63.8	64		30 (48)	130 (125)

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Limits Low	Limits High	RPD
Lab Sample ID: Q1901-08MS		**Client Sample ID: B-167-SB01MS**						**DataFile: BF142277.D**			
Pyridine	500	24.1	270	ug/L	49				20 (10)	160 (109)	
1,4-Dichlorobenzene	500	0	340	ug/L	68	*			70 (55)	130 (125)	
2-Methylphenol	500	0	360	ug/L	72				70 (60)	130 (131)	
3+4-Methylphenols	500	0	360	ug/L	72				20 (54)	160 (136)	
Hexachloroethane	500	0	350	ug/L	70				20 (19)	160 (146)	
Nitrobenzene	500	0	440	ug/L	88				70 (62)	130 (112)	
Hexachlorobutadiene	500	0	380	ug/L	76				70 (52)	130 (125)	
2,4,6-Trichlorophenol	500	0	470	ug/L	94				70 (78)	130 (112)	
2,4,5-Trichlorophenol	500	0	460	ug/L	92				70 (71)	130 (111)	
2,4-Dinitrotoluene	500	0	500	ug/L	100				70 (74)	130 (137)	
Hexachlorobenzene	500	0	440	ug/L	88				70 (72)	130 (115)	
Pentachlorophenol	1000	0	940	ug/L	94				20 (52)	160 (162)	

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Limits Low	Limits High	RPD
Lab Sample ID: Q1901-08MSD		**Client Sample ID: B-167-SB01MSD**						**DataFile: BF142278.D**			
Pyridine	500	24.1	280	ug/L	51		4		20 (10)	160 (109)	20 (20)
1,4-Dichlorobenzene	500	0	340	ug/L	68	*	0		70 (55)	130 (125)	20 (20)
2-Methylphenol	500	0	370	ug/L	74		3		70 (60)	130 (131)	20 (20)
3+4-Methylphenols	500	0	370	ug/L	74		3		20 (54)	160 (136)	20 (20)
Hexachloroethane	500	0	370	ug/L	74		6		20 (19)	160 (146)	20 (20)
Nitrobenzene	500	0	460	ug/L	92		4		70 (62)	130 (112)	20 (20)
Hexachlorobutadiene	500	0	390	ug/L	78		3		70 (52)	130 (125)	20 (20)
2,4,6-Trichlorophenol	500	0	480	ug/L	96		2		70 (78)	130 (112)	20 (20)
2,4,5-Trichlorophenol	500	0	470	ug/L	94		2		70 (71)	130 (111)	20 (20)
2,4-Dinitrotoluene	500	0	530	ug/L	106		6		70 (74)	130 (137)	20 (20)
Hexachlorobenzene	500	0	450	ug/L	90		2		70 (72)	130 (115)	20 (20)
Pentachlorophenol	1000	0	970	ug/L	97		3		20 (52)	160 (162)	20 (20)

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: 8270E **DataFile:** BF142255.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD Qual	Limits Low	Limits High	Limits RPD
PB167810BS	Pyridine	50	37.2	ug/L	74				20 (29)	160 (97)	
	1,4-Dichlorobenzene	50	39.6	ug/L	79				70 (76)	130 (103)	
	2-Methylphenol	50	41.0	ug/L	82				70 (69)	130 (109)	
	3+4-Methylphenols	50	40.8	ug/L	82				20 (67)	160 (106)	
	Hexachloroethane	50	41.8	ug/L	84				20 (76)	160 (118)	
	Nitrobenzene	50	45.8	ug/L	92				70 (58)	130 (106)	
	Hexachlorobutadiene	50	41.5	ug/L	83				70 (69)	130 (101)	
	2,4,6-Trichlorophenol	50	43.4	ug/L	87				70 (61)	130 (110)	
	2,4,5-Trichlorophenol	50	44.9	ug/L	90				70 (70)	130 (106)	
	2,4-Dinitrotoluene	50	49.8	ug/L	100				70 (60)	130 (115)	
	Hexachlorobenzene	50	41.6	ug/L	83				70 (73)	130 (106)	
	Pentachlorophenol	100	93.4	ug/L	93				20 (47)	160 (114)	

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167810BL

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

Lab File ID: BF142254.D Lab Sample ID: PB167810BL

Instrument ID: BNA_F Date Extracted: 04/30/2025

Matrix: (soil/water) water Date Analyzed: 05/01/2025

Level: (low/med) LOW Time Analyzed: 12:10

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB167810BS	PB167810BS	BF142255.D	05/01/2025
B-167-SB01	Q1901-08	BF142276.D	05/02/2025
B-167-SB01MS	Q1901-08MS	BF142277.D	05/02/2025
B-167-SB01MSD	Q1901-08MSD	BF142278.D	05/02/2025
B-170-SB01	Q1901-09	BF142279.D	05/02/2025
B-167-SB02	Q1901-10	BF142280.D	05/02/2025
B-170-SB02	Q1901-11	BF142281.D	05/02/2025
PB167774TB	PB167774TB	BF142256.D	05/01/2025

COMMENTS:

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM SAS No.: Q1901 SDG NO.: Q1901

Lab File ID: BF142238.D DFTPP Injection Date: 04/30/2025

Instrument ID: BNA_F DFTPP Injection Time: 10:55

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	32.6
68	Less than 2.0% of mass 69	0.5 (1.8) 1
69	Mass 69 relative abundance	27.3
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	37.1
197	Less than 2.0% of mass 198	0.7
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.3
275	10.0 - 60.0% of mass 198	22.8
365	Greater than 1% of mass 198	3
441	Present, but less than mass 443	15.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.9 (19.9) 2

1-Value is % mass 69 2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF142239.D	04/30/2025	11:24
SSTDICC005	SSTDICC005	BF142240.D	04/30/2025	11:52
SSTDICC010	SSTDICC010	BF142241.D	04/30/2025	12:20
SSTDICC020	SSTDICC020	BF142242.D	04/30/2025	12:49
SSTDICCC040	SSTDICCC040	BF142243.D	04/30/2025	13:17
SSTDICC050	SSTDICC050	BF142244.D	04/30/2025	13:46
SSTDICC060	SSTDICC060	BF142245.D	04/30/2025	14:15
SSTDICC080	SSTDICC080	BF142246.D	04/30/2025	14:43

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM SAS No.: Q1901 SDG NO.: Q1901

Lab File ID: BF142249.D DFTPP Injection Date: 05/01/2025

Instrument ID: BNA_F DFTPP Injection Time: 09:48

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	39.4
68	Less than 2.0% of mass 69	0.5 (1.6) 1
69	Mass 69 relative abundance	31.6
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	42.4
197	Less than 2.0% of mass 198	0.6
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.9
275	10.0 - 60.0% of mass 198	23.8
365	Greater than 1% of mass 198	3.1
441	Present, but less than mass 443	15.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.1 (19.1) 2

1-Value is % mass 69 2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142250.D	05/01/2025	10:17
PB167810BL	PB167810BL	BF142254.D	05/01/2025	12:10
PB167810BS	PB167810BS	BF142255.D	05/01/2025	12:39
PB167774TB	PB167774TB	BF142256.D	05/01/2025	13:07

Raw Data: **BF142273.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK

DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM SAS No.: Q1901 SDG NO.: Q1901

Lab File ID: BF142273.D DFTPP Injection Date: 05/02/2025

Instrument ID: BNA_F DFTPP Injection Time: 10:11

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	34.7
68	Less than 2.0% of mass 69	0.4 (1.6) 1
69	Mass 69 relative abundance	27.6
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	37.6
197	Less than 2.0% of mass 198	0.7
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.6
275	10.0 - 60.0% of mass 198	23.7
365	Greater than 1% of mass 198	3
441	Present, but less than mass 443	15.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.3 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142274.D	05/02/2025	10:39
B-167-SB01	Q1901-08	BF142276.D	05/02/2025	11:58
B-167-SB01MS	Q1901-08MS	BF142277.D	05/02/2025	12:27
B-167-SB01MSD	Q1901-08MSD	BF142278.D	05/02/2025	12:55
B-170-SB01	Q1901-09	BF142279.D	05/02/2025	13:24
B-167-SB02	Q1901-10	BF142280.D	05/02/2025	13:53
B-170-SB02	Q1901-11	BF142281.D	05/02/2025	14:22

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK

DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM SAS No.: Q1901 SDG NO.: Q1901

Lab File ID: BP024274.D DFTPP Injection Date: 04/14/2025

Instrument ID: BNA_P DFTPP Injection Time: 10:25

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	30.4
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	36
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	48.5
197	Less than 2.0% of mass 198	0.5
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	29.1
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	15.2
442	Greater than 50% of mass 198	99.1
443	15.0 - 24.0% of mass 442	19.2 (19.4) 2

1-Value is % mass 69 2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BP024275.D	04/14/2025	11:06
SSTDICC005	SSTDICC005	BP024276.D	04/14/2025	11:47
SSTDICC010	SSTDICC010	BP024277.D	04/14/2025	12:27
SSTDICC020	SSTDICC020	BP024278.D	04/14/2025	13:08
SSTDICCC040	SSTDICCC040	BP024279.D	04/14/2025	13:49
SSTDICC050	SSTDICC050	BP024280.D	04/14/2025	15:10
SSTDICC060	SSTDICC060	BP024281.D	04/14/2025	16:32
SSTDICC080	SSTDICC080	BP024282.D	04/14/2025	17:13

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM SAS No.: Q1901 SDG NO.: Q1901

Lab File ID: BP024521.D DFTPP Injection Date: 05/05/2025

Instrument ID: BNA_P DFTPP Injection Time: 11:57

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	20.8
68	Less than 2.0% of mass 69	0.2 (1.1) 1
69	Mass 69 relative abundance	23.1
70	Less than 2.0% of mass 69	0.1 (0.6) 1
127	10.0 - 80.0% of mass 198	33.6
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5
275	10.0 - 60.0% of mass 198	24.4
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	15.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.2 (19.2) 2

1-Value is % mass 69 2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP024522.D	05/05/2025	12:38
B-167-SB02RE	Q1901-10RE	BP024529.D	05/05/2025	17:26

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/01/2025

Lab File ID: BF142250.D Time Analyzed: 10:17

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

		IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
	12 HOUR STD	186578	6.91	719586	8.19	381198	9.95
	UPPER LIMIT	373156	7.41	1439170	8.692	762396	10.445
	LOWER LIMIT	93289	6.41	359793	7.692	190599	9.445
	EPA SAMPLE NO.						
01	PB167774TB	217516	6.90	839699	8.19	443070	9.95
02	PB167810BL	223997	6.90	857825	8.19	449926	9.95
03	PB167810BS	228833	6.91	892782	8.19	472284	9.95

IS1 (DCB) = 1,4-Dichlorobenzene-d4
IS2 (NPT) = Naphthalene-d8
IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/01/2025

Lab File ID: BF142250.D Time Analyzed: 10:17

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
	12 HOUR STD	647816	11.433	384442	14.074	349272	15.562
	UPPER LIMIT	1295630	11.933	768884	14.574	698544	16.062
	LOWER LIMIT	323908	10.933	192221	13.574	174636	15.062
	EPA SAMPLE NO.						
01	PB167774TB	773002	11.43	539235	14.07	394051	15.56
02	PB167810BL	794711	11.43	568207	14.07	404816	15.56
03	PB167810BS	804874	11.43	464382	14.07	446826	15.57

IS4 (PHN) = Phenanthrene-d10
IS5 (CRY) = Chrysene-d12
IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/02/2025

Lab File ID: BF142274.D Time Analyzed: 10:39

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

		IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
	12 HOUR STD	189457	6.91	747235	8.19	389025	9.95
	UPPER LIMIT	378914	7.41	1494470	8.692	778050	10.445
	LOWER LIMIT	94728.5	6.41	373618	7.692	194513	9.445
	EPA SAMPLE NO.						
01	B-167-SB01	222553	6.91	840174	8.19	447329	9.94
02	B-167-SB01MS	212656	6.91	808361	8.19	402059	9.95
03	B-167-SB01MSD	217028	6.91	832936	8.19	419407	9.95
04	B-170-SB01	207050	6.91	786560	8.19	402861	9.94
05	B-167-SB02	215163	6.90	812217	8.19	266125	9.94
06	B-170-SB02	213953	6.90	820195	8.19	417687	9.94

IS1 (DCB) = 1,4-Dichlorobenzene-d4
IS2 (NPT) = Naphthalene-d8
IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/02/2025

Lab File ID: BF142274.D Time Analyzed: 10:39

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #		RT #	IS6 (PRY) AREA #		RT #
	12 HOUR STD	652685	11.427	354752		14.068	327562		15.557
	UPPER LIMIT	1305370	11.927	709504		14.568	655124		16.057
	LOWER LIMIT	326343	10.927	177376		13.568	163781		15.057
	EPA SAMPLE NO.								
01	B-167-SB01	769739	11.43	529430		14.07	393063		15.56
02	B-167-SB01MS	629997	11.43	370460		14.07	379220		15.56
03	B-167-SB01MSD	684254	11.43	400079		14.07	401381		15.56
04	B-170-SB01	662438	11.43	411304		14.07	344110		15.56
05	B-167-SB02	611583	11.43	154727	*	14.06	3229	*	15.55
06	B-170-SB02	704817	11.43	439553		14.06	313049		15.55

IS4 (PHN) = Phenanthrene-d10
IS5 (CRY) = Chrysene-d12
IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/05/2025

Lab File ID: BP024522.D Time Analyzed: 12:38

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

		IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
	12 HOUR STD	132437	7.705	539349	10.48	358648	14.34
	UPPER LIMIT	264874	8.205	1078700	10.975	717296	14.84
	LOWER LIMIT	66218.5	7.205	269675	9.975	179324	13.84
	EPA SAMPLE NO.						
01	B-167-SB02RE	159015	7.71	625316	10.48	112223 *	14.34

IS1 (DCB) = 1,4-Dichlorobenzene-d4
IS2 (NPT) = Naphthalene-d8
IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/05/2025

Lab File ID: BP024522.D Time Analyzed: 12:38

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
	12 HOUR STD	715330	17.151	882732	21.604	1072390	24.927
	UPPER LIMIT	1430660	17.651	1765460	22.104	2144780	25.427
	LOWER LIMIT	357665	16.651	441366	21.104	536195	24.427
	EPA SAMPLE NO.						
01	B-167-SB02RE	431364	17.15	43410 *	21.59	0 *	0.00 *

IS4 (PHN) = Phenanthrene-d10
IS5 (CRY) = Chrysene-d12
IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.



SAMPLE DATA

Raw Data: **BF142276.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-167-SB01	SDG No.:	Q1901
Lab Sample ID:	Q1901-08	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142276.D	1	04/30/25 13:15	05/02/25 11:58	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	24.1	J	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/L
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	103		15 (10) - 110 (139)	69%	SPK: 150
13127-88-3	Phenol-d6	95.9		15 (10) - 110 (134)	64%	SPK: 150
4165-60-0	Nitrobenzene-d5	88.3		30 (49) - 130 (133)	88%	SPK: 100
321-60-8	2-Fluorobiphenyl	69.9		30 (52) - 130 (132)	70%	SPK: 100
118-79-6	2,4,6-Tribromophenol	139		15 (44) - 110 (137)	92%	SPK: 150
1718-51-0	Terphenyl-d14	68.1		30 (48) - 130 (125)	68%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	223000	6.91			
1146-65-2	Naphthalene-d8	840000	8.186			
15067-26-2	Acenaphthene-d10	447000	9.939			
1517-22-2	Phenanthrene-d10	770000	11.427			
1719-03-5	Chrysene-d12	529000	14.068			
1520-96-3	Perylene-d12	393000	15.557			

Raw Data: **BF142276.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-167-SB01	SDG No.:	Q1901
Lab Sample ID:	Q1901-08	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142276.D	1	04/30/25 13:15	05/02/25 11:58	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142276.D
Acq On : 02 May 2025 11:58
Operator : RC/JU
Sample : Q1901-08
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-167-SB01

Quant Time: May 02 12:28:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	222553	20.000	ng	0.00
21) Naphthalene-d8	8.186	136	840174	20.000	ng	0.00
39) Acenaphthene-d10	9.939	164	447329	20.000	ng	-0.01
64) Phenanthrene-d10	11.427	188	769739	20.000	ng	0.00
76) Chrysene-d12	14.068	240	529430	20.000	ng	0.00
86) Perylene-d12	15.557	264	393063	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.528	112	1438792	103.249	ng	0.01
7) Phenol-d6	6.522	99	1614400	95.937	ng	0.00
23) Nitrobenzene-d5	7.469	82	1129271	88.303	ng	0.00
42) 2,4,6-Tribromophenol	10.727	330	565380	138.641	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	2216915	69.876	ng	0.00
79) Terphenyl-d14	13.016	244	2479112	68.104	ng	0.00
Target Compounds						Qvalue
3) Pyridine	3.551	79	38955	2.408	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142276.D
Acq On : 02 May 2025 11:58
Operator : RC/JU
Sample : Q1901-08
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-167-SB01

Quant Time: May 02 12:28:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142276.D
Acq: 02 May 2025 11:58

Instrument :
BNA_F
ClientSampleId :
B-167-SB01

Tgt Ion:152 Resp: 222553

Ion	Ratio	Lower	Upper
152	100		
150	156.3	130.2	195.2
115	57.9	47.0	70.4







#3
Pyridine
Concen: 2.408 ng
RT: 3.551 min Scan# 231
Delta R.T. 0.088 min
Lab File: BF142276.D
Acq: 02 May 2025 11:58

Tgt Ion: 79 Resp: 38955

Ion	Ratio	Lower	Upper
79	100		
52	90.2	71.5	107.3
51	45.2	34.5	51.7







#5
2-Fluorophenol
Concen: 103.249 ng
RT: 5.528 min Scan# 5
Delta R.T. 0.012 min
Lab File: BF142276.D
Acq: 02 May 2025 11:58

Instrument :
BNA_F
ClientSampleId :
B-167-SB01

Tgt Ion:112 Resp: 1438792

Ion	Ratio	Lower	Upper
112	100		
64	60.5	48.4	72.6
63	31.3	24.8	37.2







#7
Phenol-d6
Concen: 95.937 ng
RT: 6.522 min Scan# 736
Delta R.T. -0.006 min
Lab File: BF142276.D
Acq: 02 May 2025 11:58

Tgt Ion: 99 Resp: 1614400

Ion	Ratio	Lower	Upper
99	100		
42	21.1	16.3	24.5
71	32.3	25.0	37.4







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.186 min Scan# 1
Delta R.T. -0.007 min
Lab File: BF142276.D
Acq: 02 May 2025 11:58

Instrument :
BNA_F
ClientSampleId :
B-167-SB01

Tgt Ion:136 Resp: 840174

Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.7	13.1
54	9.7	7.7	11.5
68	6.3	5.0	7.6







#23
Nitrobenzene-d5
Concen: 88.303 ng
RT: 7.469 min Scan# 897
Delta R.T. -0.006 min
Lab File: BF142276.D
Acq: 02 May 2025 11:58

Tgt Ion: 82 Resp: 1129271

Ion	Ratio	Lower	Upper
82	100		
128	47.9	37.0	55.6
54	60.7	48.0	72.0







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.939 min Scan# 1
Delta R.T. -0.012 min
Lab File: BF142276.D
Acq: 02 May 2025 11:58

Instrument :
BNA_F
ClientSampleId :
B-167-SB01

Tgt Ion:164 Resp: 447329

Ion	Ratio	Lower	Upper
164	100		
162	98.3	78.6	117.8
160	43.9	35.7	53.5







#42
2,4,6-Tribromophenol
Concen: 138.641 ng
RT: 10.727 min Scan# 1451
Delta R.T. -0.006 min
Lab File: BF142276.D
Acq: 02 May 2025 11:58

Tgt Ion:330 Resp: 565380

Ion	Ratio	Lower	Upper
330	100		
332	95.9	76.5	114.7
141	28.8	24.3	36.5







#45
2-Fluorobiphenyl
Concen: 69.876 ng
RT: 9.263 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142276.D
Acq: 02 May 2025 11:58

Instrument :
BNA_F
ClientSampleId :
B-167-SB01

Tgt Ion:172 Resp: 2216915

Ion	Ratio	Lower	Upper
172	100		
171	35.7	29.0	43.4
170	23.8	19.4	29.0







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.427 min Scan# 1570
Delta R.T. -0.006 min
Lab File: BF142276.D
Acq: 02 May 2025 11:58

Tgt Ion:188 Resp: 769739

Ion	Ratio	Lower	Upper
188	100		
94	9.9	8.2	12.2
80	10.6	8.6	12.8







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.068 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF142276.D
Acq: 02 May 2025 11:58

Instrument :
BNA_F
ClientSampleId :
B-167-SB01

Tgt Ion:240 Resp: 529430

Ion	Ratio	Lower	Upper
240	100		
120	11.7	10.1	15.1
236	25.5	20.5	30.7







#79
Terphenyl-d14
Concen: 68.104 ng
RT: 13.016 min Scan# 1840
Delta R.T. -0.006 min
Lab File: BF142276.D
Acq: 02 May 2025 11:58

Tgt Ion:244 Resp: 2479112

Ion	Ratio	Lower	Upper
244	100		
212	6.7	5.3	7.9
122	11.4	9.1	13.7






#86
Perylene-d12
Concen: 20.000 ng
RT: 15.557 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF142276.D
Acq: 02 May 2025 11:58

Instrument :
BNA_F
ClientSampleId :
B-167-SB01

Tgt Ion:264 Resp: 393063

Ion	Ratio	Lower	Upper
264	100		
260	23.3	18.4	27.6
265	21.5	17.3	25.9



Raw Data: **BF142279.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-170-SB01	SDG No.:	Q1901
Lab Sample ID:	Q1901-09	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142279.D	1	04/30/25 13:15	05/02/25 13:24	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	12.8	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/L
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	101		15 (10) - 110 (139)	67%	SPK: 150
13127-88-3	Phenol-d6	92.9		15 (10) - 110 (134)	62%	SPK: 150
4165-60-0	Nitrobenzene-d5	87.7		30 (49) - 130 (133)	88%	SPK: 100
321-60-8	2-Fluorobiphenyl	70.7		30 (52) - 130 (132)	71%	SPK: 100
118-79-6	2,4,6-Tribromophenol	130		15 (44) - 110 (137)	87%	SPK: 150
1718-51-0	Terphenyl-d14	72.5		30 (48) - 130 (125)	72%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	207000	6.91			
1146-65-2	Naphthalene-d8	787000	8.186			
15067-26-2	Acenaphthene-d10	403000	9.939			
1517-22-2	Phenanthrene-d10	662000	11.427			
1719-03-5	Chrysene-d12	411000	14.068			
1520-96-3	Perylene-d12	344000	15.557			

Raw Data: **BF142279.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-170-SB01	SDG No.:	Q1901
Lab Sample ID:	Q1901-09	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142279.D	1	04/30/25 13:15	05/02/25 13:24	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142279.D
Acq On : 02 May 2025 13:24
Operator : RC/JU
Sample : Q1901-09
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-170-SB01

Quant Time: May 02 14:19:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	207050	20.000	ng	0.00
21) Naphthalene-d8	8.186	136	786560	20.000	ng	0.00
39) Acenaphthene-d10	9.939	164	402861	20.000	ng	-0.01
64) Phenanthrene-d10	11.427	188	662438	20.000	ng	0.00
76) Chrysene-d12	14.068	240	411304	20.000	ng	0.00
86) Perylene-d12	15.557	264	344110	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.522	112	1304723	100.639	ng	0.00
7) Phenol-d6	6.522	99	1455008	92.939	ng	0.00
23) Nitrobenzene-d5	7.463	82	1050218	87.719	ng	-0.01
42) 2,4,6-Tribromophenol	10.727	330	478989	130.421	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	2020679	70.721	ng	0.00
79) Terphenyl-d14	13.016	244	2048893	72.451	ng	0.00
Target Compounds						Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142279.D
Acq On : 02 May 2025 13:24
Operator : RC/JU
Sample : Q1901-09
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-170-SB01

Quant Time: May 02 14:19:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142279.D
Acq: 02 May 2025 13:24

Instrument :
BNA_F
ClientSampleId :
B-170-SB01

Tgt Ion:152 Resp: 207050

Ion	Ratio	Lower	Upper
152	100		
150	155.6	130.2	195.2
115	56.4	47.0	70.4







#5
2-Fluorophenol
Concen: 100.639 ng
RT: 5.522 min Scan# 566
Delta R.T. 0.006 min
Lab File: BF142279.D
Acq: 02 May 2025 13:24

Tgt Ion:112 Resp: 1304723

Ion	Ratio	Lower	Upper
112	100		
64	61.5	48.4	72.6
63	31.8	24.8	37.2







#7
Phenol-d6
Concen: 92.939 ng
RT: 6.522 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF142279.D
Acq: 02 May 2025 13:24

Instrument :
BNA_F
ClientSampleId :
B-170-SB01

Tgt Ion: 99 Resp: 1455008

Ion	Ratio	Lower	Upper
99	100		
42	21.1	16.3	24.5
71	32.4	25.0	37.4







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.186 min Scan# 1019
Delta R.T. -0.007 min
Lab File: BF142279.D
Acq: 02 May 2025 13:24

Tgt Ion:136 Resp: 786560

Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.7	13.1
54	9.6	7.7	11.5
68	6.3	5.0	7.6







#23
Nitrobenzene-d5
Concen: 87.719 ng
RT: 7.463 min Scan# 8
Delta R.T. -0.012 min
Lab File: BF142279.D
Acq: 02 May 2025 13:24

Instrument :
BNA_F
ClientSampleId :
B-170-SB01

Tgt Ion: 82 Resp: 1050218

Ion	Ratio	Lower	Upper
82	100		
128	46.2	37.0	55.6
54	61.3	48.0	72.0







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.939 min Scan# 1317
Delta R.T. -0.012 min
Lab File: BF142279.D
Acq: 02 May 2025 13:24

Tgt Ion:164 Resp: 402861

Ion	Ratio	Lower	Upper
164	100		
162	100.7	78.6	117.8
160	44.5	35.7	53.5







#42
2,4,6-Tribromophenol
Concen: 130.421 ng
RT: 10.727 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF142279.D
Acq: 02 May 2025 13:24

Instrument :
BNA_F
ClientSampleId :
B-170-SB01

Tgt Ion:330 Resp: 478989

Ion	Ratio	Lower	Upper
330	100		
332	96.9	76.5	114.7
141	29.2	24.3	36.5







#45
2-Fluorobiphenyl
Concen: 70.721 ng
RT: 9.263 min Scan# 1202
Delta R.T. -0.006 min
Lab File: BF142279.D
Acq: 02 May 2025 13:24

Tgt Ion:172 Resp: 2020679

Ion	Ratio	Lower	Upper
172	100		
171	35.7	29.0	43.4
170	23.8	19.4	29.0







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.427 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142279.D
Acq: 02 May 2025 13:24

Instrument :
BNA_F
ClientSampleId :
B-170-SB01

Tgt Ion:188 Resp: 662438

Ion	Ratio	Lower	Upper
188	100		
94	10.0	8.2	12.2
80	10.3	8.6	12.8







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.068 min Scan# 2019
Delta R.T. -0.006 min
Lab File: BF142279.D
Acq: 02 May 2025 13:24

Tgt Ion:240 Resp: 411304

Ion	Ratio	Lower	Upper
240	100		
120	11.4	10.1	15.1
236	24.8	20.5	30.7







#79
Terphenyl-d14
Concen: 72.451 ng
RT: 13.016 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142279.D
Acq: 02 May 2025 13:24

Instrument :
BNA_F
ClientSampleId :
B-170-SB01

Tgt Ion:244 Resp: 2048893

Ion	Ratio	Lower	Upper
244	100		
212	6.7	5.3	7.9
122	11.3	9.1	13.7






#86
Perylene-d12
Concen: 20.000 ng
RT: 15.557 min Scan# 2272
Delta R.T. -0.012 min
Lab File: BF142279.D
Acq: 02 May 2025 13:24

Tgt Ion:264 Resp: 344110

Ion	Ratio	Lower	Upper
264	100		
260	23.1	18.4	27.6
265	21.1	17.3	25.9



Raw Data: **BF142280.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-167-SB02	SDG No.:	Q1901
Lab Sample ID:	Q1901-10	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142280.D	1	04/30/25 13:15	05/02/25 13:53	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	12.8	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/L
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	101		15 (10) - 110 (139)	67%	SPK: 150
13127-88-3	Phenol-d6	70.4		15 (10) - 110 (134)	47%	SPK: 150
4165-60-0	Nitrobenzene-d5	90.2		30 (49) - 130 (133)	90%	SPK: 100
321-60-8	2-Fluorobiphenyl	114		30 (52) - 130 (132)	114%	SPK: 100
118-79-6	2,4,6-Tribromophenol	212	*	15 (44) - 110 (137)	141%	SPK: 150
1718-51-0	Terphenyl-d14	181	*	30 (48) - 130 (125)	181%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	215000	6.904			
1146-65-2	Naphthalene-d8	812000	8.187			
15067-26-2	Acenaphthene-d10	266000	9.939			
1517-22-2	Phenanthrene-d10	612000	11.428			
1719-03-5	Chrysene-d12	155000	14.063			
1520-96-3	Perylene-d12	3230	15.551			

Raw Data: **BF142280.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-167-SB02	SDG No.:	Q1901
Lab Sample ID:	Q1901-10	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142280.D	1	04/30/25 13:15	05/02/25 13:53	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142280.D
Acq On : 02 May 2025 13:53
Operator : RC/JU
Sample : Q1901-10
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-167-SB02

Quant Time: May 02 14:22:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.904	152	215163	20.000	ng	0.00
21) Naphthalene-d8	8.187	136	812217	20.000	ng	0.00
39) Acenaphthene-d10	9.939	164	266125	20.000	ng	-0.01
64) Phenanthrene-d10	11.428	188	611583	20.000	ng	0.00
76) Chrysene-d12	14.063	240	154727	20.000	ng	-0.01
86) Perylene-d12	15.551	264	3229	20.000	ng	#-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.528	112	1362221	101.112	ng	0.01
7) Phenol-d6	6.522	99	1145189	70.391	ng	0.00
23) Nitrobenzene-d5	7.463	82	1114602	90.156	ng	-0.01
42) 2,4,6-Tribromophenol	10.728	330	514497	212.068	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	2159177	114.396	ng	0.00
79) Terphenyl-d14	13.016	244	1928369	181.263	ng	0.00
Target Compounds						Qvalue
32) Benzoic acid	7.869	122	28398	14.343	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed


Sample Results: **BF142280.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142280.D
Acq On : 02 May 2025 13:53
Operator : RC/JU
Sample : Q1901-10
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-167-SB02

Quant Time: May 02 14:22:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.904 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF142280.D
Acq: 02 May 2025 13:53

Instrument :
BNA_F
ClientSampleId :
B-167-SB02

Tgt Ion:152 Resp: 215163

Ion	Ratio	Lower	Upper
152	100		
150	156.8	130.2	195.2
115	59.4	47.0	70.4







#5
2-Fluorophenol
Concen: 101.112 ng
RT: 5.528 min Scan# 567
Delta R.T. 0.012 min
Lab File: BF142280.D
Acq: 02 May 2025 13:53

Tgt Ion:112 Resp: 1362221

Ion	Ratio	Lower	Upper
112	100		
64	59.8	48.4	72.6
63	30.7	24.8	37.2







#7
Phenol-d6
Concen: 70.391 ng
RT: 6.522 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF142280.D
Acq: 02 May 2025 13:53

Instrument :
BNA_F
ClientSampleId :
B-167-SB02

Tgt Ion: 99 Resp: 1145189

Ion	Ratio	Lower	Upper
99	100		
42	22.1	16.3	24.5
71	31.6	25.0	37.4







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.187 min Scan# 1019
Delta R.T. -0.006 min
Lab File: BF142280.D
Acq: 02 May 2025 13:53

Tgt Ion:136 Resp: 812217

Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.7	13.1
54	9.0	7.7	11.5
68	6.0	5.0	7.6







#23
Nitrobenzene-d5
Concen: 90.156 ng
RT: 7.463 min Scan# 8
Delta R.T. -0.012 min
Lab File: BF142280.D
Acq: 02 May 2025 13:53

Instrument :
BNA_F
ClientSampleId :
B-167-SB02

Tgt Ion: 82 Resp: 1114602

Ion	Ratio	Lower	Upper
82	100		
128	45.7	37.0	55.6
54	61.4	48.0	72.0







#32
Benzoic acid
Concen: 14.343 ng
RT: 7.869 min Scan# 965
Delta R.T. -0.065 min
Lab File: BF142280.D
Acq: 02 May 2025 13:53

Tgt Ion:122 Resp: 28398

Ion	Ratio	Lower	Upper
122	100		
105	118.5	98.2	138.2
77	86.1	66.8	106.8







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.939 min Scan# 1
Delta R.T. -0.012 min
Lab File: BF142280.D
Acq: 02 May 2025 13:53

Instrument :
BNA_F
ClientSampleId :
B-167-SB02

Tgt Ion:164 Resp: 266125

Ion	Ratio	Lower	Upper
164	100		
162	105.4	78.6	117.8
160	49.7	35.7	53.5







#42
2,4,6-Tribromophenol
Concen: 212.068 ng
RT: 10.728 min Scan# 1451
Delta R.T. -0.006 min
Lab File: BF142280.D
Acq: 02 May 2025 13:53

Tgt Ion:330 Resp: 514497

Ion	Ratio	Lower	Upper
330	100		
332	97.0	76.5	114.7
141	28.7	24.3	36.5







#45
2-Fluorobiphenyl
Concen: 114.396 ng
RT: 9.263 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142280.D
Acq: 02 May 2025 13:53

Instrument :
BNA_F
ClientSampleId :
B-167-SB02

Tgt Ion:172 Resp: 2159177

Ion	Ratio	Lower	Upper
172	100		
171	35.9	29.0	43.4
170	23.8	19.4	29.0







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.428 min Scan# 1570
Delta R.T. -0.006 min
Lab File: BF142280.D
Acq: 02 May 2025 13:53

Tgt Ion:188 Resp: 611583

Ion	Ratio	Lower	Upper
188	100		
94	9.6	8.2	12.2
80	10.0	8.6	12.8







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.063 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF142280.D
Acq: 02 May 2025 13:53

Instrument :
BNA_F
ClientSampleId :
B-167-SB02

Tgt Ion:240 Resp: 154727

Ion	Ratio	Lower	Upper
240	100		
120	12.4	10.1	15.1
236	28.4	20.5	30.7







#79
Terphenyl-d14
Concen: 181.263 ng
RT: 13.016 min Scan# 1840
Delta R.T. -0.006 min
Lab File: BF142280.D
Acq: 02 May 2025 13:53

Tgt Ion:244 Resp: 1928369

Ion	Ratio	Lower	Upper
244	100		
212	6.7	5.3	7.9
122	11.1	9.1	13.7






#86
Perylene-d12
Concen: 20.000 ng
RT: 15.551 min Scan# 2
Delta R.T. -0.018 min
Lab File: BF142280.D
Acq: 02 May 2025 13:53

Instrument :
BNA_F
ClientSampleId :
B-167-SB02

Tgt Ion:264 Resp: 3229

Ion	Ratio	Lower	Upper
264	100		
260	36.3	18.4	27.6#
265	32.4	17.3	25.9#



Raw Data: **BP024529.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-167-SB02RE	SDG No.:	Q1901
Lab Sample ID:	Q1901-10RE	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024529.D	1	04/30/25 13:15	05/05/25 17:26	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	12.8	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/L
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	115		15 (10) - 110 (139)	77%	SPK: 150
13127-88-3	Phenol-d6	41.6		15 (10) - 110 (134)	28%	SPK: 150
4165-60-0	Nitrobenzene-d5	93.0		30 (49) - 130 (133)	93%	SPK: 100
321-60-8	2-Fluorobiphenyl	314	*	30 (52) - 130 (132)	313%	SPK: 100
118-79-6	2,4,6-Tribromophenol	554	*	15 (44) - 110 (137)	369%	SPK: 150
1718-51-0	Terphenyl-d14	902	*	30 (48) - 130 (125)	902%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	159000	7.711			
1146-65-2	Naphthalene-d8	625000	10.481			
15067-26-2	Acenaphthene-d10	112000	14.34			
1517-22-2	Phenanthrene-d10	431000	17.145			
1719-03-5	Chrysene-d12	43400	21.586			
1520-96-3	Perylene-d12	0	0			

Raw Data: **BP024529.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025			Date Received:	04/28/25
Client Sample ID:	B-167-SB02RE			SDG No.:	Q1901
Lab Sample ID:	Q1901-10RE			Matrix:	TCLP
Analytical Method:	SW8270			% Solid:	0
Sample Wt/Vol:	100	Units:	mL	Final Vol:	1000 uL
Soil Aliquot Vol:			uL	Test:	TCLP BNA
Extraction Type :			Decanted : N	Level :	LOW
Injection Volume :			GPC Factor : 1.0	GPC Cleanup : N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024529.D	1	04/30/25 13:15	05/05/25 17:26	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Quantitation Report (QT Reviewed)

Sample Results: **BP024529.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050525\
Data File : BP024529.D
Acq On : 05 May 2025 17:26
Operator : RC/JU
Sample : Q1901-10RE
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
B-167-SB02RE

Quant Time: May 05 18:21:26 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.711	152	159015	20.000	ng	-0.01
21) Naphthalene-d8	10.481	136	625316	20.000	ng	#-0.02
39) Acenaphthene-d10	14.340	164	112223	20.000	ng	#-0.01
64) Phenanthrene-d10	17.145	188	431364	20.000	ng	0.00
76) Chrysene-d12	21.586	240	43410	20.000	ng	# 0.00
86) Perylene-d12	0.000	264	0m	20.000	ng	-24.93
System Monitoring Compounds						
5) 2-Fluorophenol	5.334	112	1105558	114.957	ng	-0.01
7) Phenol-d6	6.899	99	548086	41.621	ng	-0.01
23) Nitrobenzene-d5	8.858	82	1019694	92.984	ng	-0.01
42) 2,4,6-Tribromophenol	15.857	330	859661	553.668	ng	0.00
45) 2-Fluorobiphenyl	12.951	172	2297807	313.500	ng	-0.01
79) Terphenyl-d14	19.875	244	1941631	901.549	ng	0.00
Target Compounds						Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed


Sample Results: **BP024529.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050525\
Data File : BP024529.D
Acq On : 05 May 2025 17:26
Operator : RC/JU
Sample : Q1901-10RE
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
B-167-SB02RE

Quant Time: May 05 18:21:26 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.711 min Scan# 7
Delta R.T. -0.011 min
Lab File: BP024529.D
Acq: 05 May 2025 17:26

Instrument :
BNA_P
ClientSampleId :
B-167-SB02RE

Tgt Ion:152 Resp: 159015

Ion	Ratio	Lower	Upper
152	100		
150	153.8	124.9	187.3
115	62.1	47.7	71.5







#5
2-Fluorophenol
Concen: 114.957 ng
RT: 5.334 min Scan# 382
Delta R.T. -0.012 min
Lab File: BP024529.D
Acq: 05 May 2025 17:26

Tgt Ion:112 Resp: 1105558

Ion	Ratio	Lower	Upper
112	100		
64	54.2	41.1	61.7
63	28.7	20.6	30.8







#7
Phenol-d6
Concen: 41.621 ng
RT: 6.899 min Scan# 64
Delta R.T. -0.012 min
Lab File: BP024529.D
Acq: 05 May 2025 17:26

Instrument :
BNA_P
ClientSampleId :
B-167-SB02RE

Tgt Ion: 99 Resp: 548086

Ion	Ratio	Lower	Upper
99	100		
42	18.8	10.6	16.0#
71	35.2	23.7	35.5







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.481 min Scan# 1257
Delta R.T. -0.017 min
Lab File: BP024529.D
Acq: 05 May 2025 17:26

Tgt Ion:136 Resp: 625316

Ion	Ratio	Lower	Upper
136	100		
137	11.5	9.2	13.8
54	6.8	6.0	9.0
68	5.3	5.5	8.3#







#23
Nitrobenzene-d5
Concen: 92.984 ng
RT: 8.858 min Scan# 9
Delta R.T. -0.012 min
Lab File: BP024529.D
Acq: 05 May 2025 17:26

Instrument :
BNA_P
ClientSampleId :
B-167-SB02RE

Tgt Ion: 82 Resp: 1019694

Ion	Ratio	Lower	Upper
82	100		
128	46.6	36.5	54.7
54	46.9	33.4	50.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.340 min Scan# 1913
Delta R.T. -0.012 min
Lab File: BP024529.D
Acq: 05 May 2025 17:26

Tgt Ion:164 Resp: 112223

Ion	Ratio	Lower	Upper
164	100		
162	110.1	80.6	120.8
160	53.8	35.3	52.9#







#42
2,4,6-Tribromophenol
Concen: 553.668 ng
RT: 15.857 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP024529.D
Acq: 05 May 2025 17:26

Instrument :
BNA_P
ClientSampleId :
B-167-SB02RE

Tgt Ion:330 Resp: 859661

Ion	Ratio	Lower	Upper
330	100		
332	97.1	77.1	115.7
141	29.6	24.7	37.1







#45
2-Fluorobiphenyl
Concen: 313.500 ng
RT: 12.951 min Scan# 1677
Delta R.T. -0.012 min
Lab File: BP024529.D
Acq: 05 May 2025 17:26

Tgt Ion:172 Resp: 2297807

Ion	Ratio	Lower	Upper
172	100		
171	35.7	28.2	42.2
170	23.9	18.6	27.8







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.145 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024529.D
Acq: 05 May 2025 17:26

Instrument :
BNA_P
ClientSampleId :
B-167-SB02RE

Tgt Ion:188 Resp: 431364

Ion	Ratio	Lower	Upper
188	100		
94	9.6	8.6	13.0
80	10.6	9.8	14.6







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.586 min Scan# 3145
Delta R.T. -0.006 min
Lab File: BP024529.D
Acq: 05 May 2025 17:26

Tgt Ion:240 Resp: 43410

Ion	Ratio	Lower	Upper
240	100		
120	13.8	9.8	14.8
236	41.3	20.9	31.3#







#79
Terphenyl-d14
Concen: 901.549 ng
RT: 19.875 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024529.D
Acq: 05 May 2025 17:26

Instrument :
BNA_P
ClientSampleId :
B-167-SB02RE

Tgt Ion:244 Resp: 1941631

Ion	Ratio	Lower	Upper
244	100		
212	8.0	5.6	8.4
122	9.5	9.4	14.0





#86
Perylene-d12
Concen: 0.000 ng
Expected RT: 24.93 min

Lab File: BP024529.D
Acq: 05 May 2025 17:26

Tgt Ion: 264

Sig	Exp Ratio
264	100
260	23.6
265	22.2

Raw Data: **BF142281.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-170-SB02	SDG No.:	Q1901
Lab Sample ID:	Q1901-11	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142281.D	1	04/30/25 13:15	05/02/25 14:22	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	12.8	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/L
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	94.4		15 (10) - 110 (139)	63%	SPK: 150
13127-88-3	Phenol-d6	88.5		15 (10) - 110 (134)	59%	SPK: 150
4165-60-0	Nitrobenzene-d5	79.5		30 (49) - 130 (133)	79%	SPK: 100
321-60-8	2-Fluorobiphenyl	67.6		30 (52) - 130 (132)	68%	SPK: 100
118-79-6	2,4,6-Tribromophenol	120		15 (44) - 110 (137)	80%	SPK: 150
1718-51-0	Terphenyl-d14	63.8		30 (48) - 130 (125)	64%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	214000	6.904			
1146-65-2	Naphthalene-d8	820000	8.187			
15067-26-2	Acenaphthene-d10	418000	9.939			
1517-22-2	Phenanthrene-d10	705000	11.428			
1719-03-5	Chrysene-d12	440000	14.063			
1520-96-3	Perylene-d12	313000	15.551			

Raw Data: **BF142281.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-170-SB02	SDG No.:	Q1901
Lab Sample ID:	Q1901-11	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142281.D	1	04/30/25 13:15	05/02/25 14:22	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142281.D
Acq On : 02 May 2025 14:22
Operator : RC/JU
Sample : Q1901-11
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-170-SB02

Quant Time: May 02 15:21:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.904	152	213953	20.000	ng	0.00
21) Naphthalene-d8	8.187	136	820195	20.000	ng	0.00
39) Acenaphthene-d10	9.939	164	417687	20.000	ng	-0.01
64) Phenanthrene-d10	11.428	188	704817	20.000	ng	0.00
76) Chrysene-d12	14.063	240	439553	20.000	ng	-0.01
86) Perylene-d12	15.551	264	313049	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.522	112	1264763	94.409	ng	0.00
7) Phenol-d6	6.522	99	1431879	88.511	ng	0.00
23) Nitrobenzene-d5	7.463	82	992258	79.479	ng	-0.01
42) 2,4,6-Tribromophenol	10.728	330	458282	120.354	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	2003805	67.641	ng	0.00
79) Terphenyl-d14	13.016	244	1927099	63.764	ng	0.00
Target Compounds						Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142281.D
Acq On : 02 May 2025 14:22
Operator : RC/JU
Sample : Q1901-11
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-170-SB02

Quant Time: May 02 15:21:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.904 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF142281.D
Acq: 02 May 2025 14:22

Instrument :
BNA_F
ClientSampleId :
B-170-SB02

Tgt Ion:152 Resp: 213953

Ion	Ratio	Lower	Upper
152	100		
150	156.5	130.2	195.2
115	60.8	47.0	70.4







#5
2-Fluorophenol
Concen: 94.409 ng
RT: 5.522 min Scan# 566
Delta R.T. 0.006 min
Lab File: BF142281.D
Acq: 02 May 2025 14:22

Tgt Ion:112 Resp: 1264763

Ion	Ratio	Lower	Upper
112	100		
64	61.0	48.4	72.6
63	31.5	24.8	37.2







#7
Phenol-d6
Concen: 88.511 ng
RT: 6.522 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF142281.D
Acq: 02 May 2025 14:22

Instrument :
BNA_F
ClientSampleId :
B-170-SB02

Tgt Ion: 99 Resp: 1431879

Ion	Ratio	Lower	Upper
99	100		
42	20.7	16.3	24.5
71	32.0	25.0	37.4







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.187 min Scan# 1019
Delta R.T. -0.006 min
Lab File: BF142281.D
Acq: 02 May 2025 14:22

Tgt Ion:136 Resp: 820195

Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.7	13.1
54	9.3	7.7	11.5
68	6.1	5.0	7.6







#23
Nitrobenzene-d5
Concen: 79.479 ng
RT: 7.463 min Scan# 8
Delta R.T. -0.012 min
Lab File: BF142281.D
Acq: 02 May 2025 14:22

Instrument :
BNA_F
ClientSampleId :
B-170-SB02

Tgt Ion: 82 Resp: 992258

Ion	Ratio	Lower	Upper
82	100		
128	46.2	37.0	55.6
54	61.0	48.0	72.0







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.939 min Scan# 1317
Delta R.T. -0.012 min
Lab File: BF142281.D
Acq: 02 May 2025 14:22

Tgt Ion:164 Resp: 417687

Ion	Ratio	Lower	Upper
164	100		
162	100.3	78.6	117.8
160	45.0	35.7	53.5







#42
2,4,6-Tribromophenol
Concen: 120.354 ng
RT: 10.728 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF142281.D
Acq: 02 May 2025 14:22

Instrument :
BNA_F
ClientSampleId :
B-170-SB02

Tgt Ion:330 Resp: 458282

Ion	Ratio	Lower	Upper
330	100		
332	96.9	76.5	114.7
141	29.6	24.3	36.5







#45
2-Fluorobiphenyl
Concen: 67.641 ng
RT: 9.263 min Scan# 1202
Delta R.T. -0.006 min
Lab File: BF142281.D
Acq: 02 May 2025 14:22

Tgt Ion:172 Resp: 2003805

Ion	Ratio	Lower	Upper
172	100		
171	35.7	29.0	43.4
170	24.0	19.4	29.0







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.428 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142281.D
Acq: 02 May 2025 14:22

Instrument :
BNA_F
ClientSampleId :
B-170-SB02

Tgt Ion:188 Resp: 704817

Ion	Ratio	Lower	Upper
188	100		
94	9.8	8.2	12.2
80	10.1	8.6	12.8







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.063 min Scan# 2018
Delta R.T. -0.012 min
Lab File: BF142281.D
Acq: 02 May 2025 14:22

Tgt Ion:240 Resp: 439553

Ion	Ratio	Lower	Upper
240	100		
120	12.3	10.1	15.1
236	25.5	20.5	30.7







#79
Terphenyl-d14
Concen: 63.764 ng
RT: 13.016 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142281.D
Acq: 02 May 2025 14:22

Instrument :
BNA_F
ClientSampleId :
B-170-SB02

Tgt Ion:244 Resp: 1927099

Ion	Ratio	Lower	Upper
244	100		
212	6.5	5.3	7.9
122	10.9	9.1	13.7






#86
Perylene-d12
Concen: 20.000 ng
RT: 15.551 min Scan# 2271
Delta R.T. -0.018 min
Lab File: BF142281.D
Acq: 02 May 2025 14:22

Tgt Ion:264 Resp: 313049

Ion	Ratio	Lower	Upper
264	100		
260	23.2	18.4	27.6
265	21.4	17.3	25.9



CALIBRATION SUMMARY

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG No.: Q1901

Instrument ID: BNA_F Calibration Date(s): 04/30/2025 04/30/2025

Calibration Time(s): 11:24 14:43

LAB FILE ID: RRF2.5 = BF142239.D RRF005 = BF142240.D RRF010 = BF142241.D
RRF020 = BF142242.D RRF040 = BF142243.D RRF050 = BF142244.D

COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Pyridine		1.465	1.391	1.454	1.522	1.503	1.454	3.6
2-Fluorophenol		1.324	1.238	1.265	1.314	1.285	1.252	5.3
Phenol-d6		1.559	1.505	1.534	1.593	1.550	1.512	4.9
1,4-Dichlorobenzene		1.610	1.486	1.495	1.546	1.488	1.475	6.9
2-Methylphenol		1.085	1.029	1.068	1.121	1.084	1.060	4.2
3+4-Methylphenols		1.357	1.323	1.372	1.400	1.366	1.326	5.9
Nitrobenzene-d5		0.217	0.251	0.302	0.340	0.351	0.304	16.9
Hexachloroethane		0.468	0.466	0.481	0.511	0.514	0.483	4.8
Nitrobenzene		0.229	0.261	0.297	0.328	0.338	0.299	13.3
Hexachlorobutadiene		0.180	0.175	0.179	0.185	0.178	0.175	4.3
2,4,6-Trichlorophenol		0.275	0.329	0.353	0.370	0.375	0.349	10.5
2-Fluorobiphenyl		1.738	1.580	1.520	1.409	1.347	1.418	15.1
2,4,5-Trichlorophenol		0.322	0.330	0.362	0.404	0.411	0.367	9.2
2,4-Dinitrotoluene		0.104	0.145	0.213	0.285	0.311	0.239	36.0
2,4,6-Tribromophenol		0.139	0.158	0.184	0.204	0.205	0.182	13.7
Hexachlorobenzene		0.256	0.245	0.248	0.261	0.255	0.249	3.3
Pentachlorophenol			0.093	0.121	0.144	0.145	0.132	16.2
Terphenyl-d14		1.472	1.413	1.422	1.481	1.392	1.375	8.4

All other compounds must meet a minimum RRF of 0.010.

Response Factor Report BNA_F

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF043025.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Apr 30 16:00:01 2025
Response Via : Initial Calibration

Calibration Files
2.5 =BF142239.D 5 =BF142240.D 10 =BF142241.D 20 =BF142242.D 40 =BF142243.D 50 =BF142244.D 60 =BF142245.D 80 =BF1422
46.D

		Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1)	I	1,4-Dichlorobenzen...	----------ISTD----------									
2)		1,4-Dioxane		0.563	0.534	0.561	0.592	0.580	0.555	0.530	0.559	4.03
3)		Pyridine		1.465	1.391	1.454	1.522	1.503	1.460	1.380	1.454	3.62
4)		n-Nitrosodimet...		0.747	0.699	0.728	0.807	0.789	0.766	0.731	0.752	4.95
5)	S	2-Fluorophenol		1.324	1.238	1.265	1.314	1.285	1.210	1.131	1.252	5.35
6)		Aniline		2.212	2.085	2.157	2.245	2.203	2.101	1.981	2.140	4.29
7)	S	Phenol-d6		1.559	1.505	1.534	1.593	1.550	1.475	1.370	1.512	4.85
8)		2-Chlorophenol		1.252	1.232	1.304	1.379	1.354	1.313	1.225	1.294	4.64
9)		Benzaldehyde		1.154	1.088	1.095	1.117	1.078	0.993	0.875	1.057	8.91
10)	C	Phenol		1.679	1.588	1.657	1.715	1.687	1.603	1.468	1.628	5.15
11)		bis(2-Chloroet...		1.345	1.250	1.270	1.318	1.285	1.231	1.132	1.262	5.46
12)		1,3-Dichlorobe...		1.571	1.485	1.485	1.533	1.478	1.396	1.299	1.464	6.18
13)	C	1,4-Dichlorobe...		1.610	1.486	1.495	1.546	1.488	1.404	1.295	1.475	6.86
14)		1,2-Dichlorobe...		1.489	1.418	1.431	1.475	1.432	1.356	1.255	1.408	5.67
15)		Benzyl Alcohol		1.047	1.017	1.064	1.149	1.126	1.082	1.023	1.073	4.69
16)		2,2'-oxybis(1-...		2.368	2.287	2.303	2.386	2.344	2.216	2.043	2.278	5.19
17)		2-Methylphenol		1.085	1.029	1.068	1.121	1.084	1.050	0.983	1.060	4.19
18)		Hexachloroethane		0.468	0.466	0.481	0.511	0.514	0.489	0.451	0.483	4.84
19)	P	n-Nitroso-di-n...	0.953	0.987	0.937	0.949	0.967	0.939	0.898	0.844	0.934	4.77
20)		3+4-Methylphenols		1.357	1.323	1.372	1.400	1.366	1.296	1.166	1.326	5.90
21)	I	Naphthalene-d8	----------ISTD----------									
22)		Acetophenone		0.511	0.482	0.483	0.468	0.461	0.430	0.404	0.463	7.70
23)	S	Nitrobenzene-d5		0.217	0.251	0.302	0.340	0.351	0.339	0.330	0.304	16.88
24)		Nitrobenzene		0.229	0.261	0.297	0.328	0.338	0.323	0.314	0.299	13.31
25)		Isophorone		0.660	0.630	0.642	0.660	0.654	0.626	0.601	0.639	3.36
26)	C	2-Nitrophenol		0.048	0.057	0.083	0.115	0.132	0.136	0.142	0.102	38.38#
27)		2,4-Dimethylph...		0.315	0.319	0.333	0.337	0.333	0.315	0.302	0.322	3.96
28)		bis(2-Chloroet...		0.425	0.404	0.414	0.415	0.410	0.386	0.366	0.403	4.98
29)	C	2,4-Dichloroph...		0.246	0.253	0.274	0.291	0.286	0.272	0.263	0.269	6.12
30)		1,2,4-Trichlor...		0.320	0.303	0.305	0.312	0.304	0.287	0.273	0.301	5.28
31)		Naphthalene		1.106	1.040	1.035	1.020	0.999	0.924	0.858	0.997	8.21
32)		Benzoic acid			0.048	0.087	0.130	0.146	0.159	0.179	0.125	39.24
33)		4-Chloroaniline		0.446	0.414	0.429	0.429	0.423	0.396	0.371	0.415	5.98
34)	C	Hexachlorobuta...		0.180	0.175	0.179	0.185	0.178	0.170	0.161	0.175	4.35
35)		Caprolactam		0.068	0.073	0.081	0.087	0.089	0.085	0.083	0.081	9.45
36)	C	4-Chloro-3-met...		0.262	0.265	0.279	0.295	0.293	0.276	0.266	0.277	4.84
37)		2-Methylnaphth...		0.679	0.637	0.633	0.623	0.612	0.574	0.534	0.613	7.69
38)		1-Methylnaphth...		0.700	0.655	0.667	0.652	0.641	0.590	0.542	0.635	8.29


Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF043025.M

39)	I	Acenaphthene-d10	----------ISTD----------								
40)		1,2,4,5-Tetrac...	0.621	0.575	0.572	0.576	0.566	0.534	0.496	0.563	6.88
41)	P	Hexachlorocycl...	0.244	0.265	0.306	0.352	0.368	0.358	0.343	0.319	15.23
42)	S	2,4,6-Tribromo...	0.139	0.158	0.184	0.204	0.205	0.197	0.190	0.182	13.73
43)	C	2,4,6-Trichlor...	0.275	0.329	0.353	0.370	0.375	0.381	0.357	0.349	10.50
44)		2,4,5-Trichlor...	0.322	0.330	0.362	0.404	0.411	0.371	0.365	0.367	9.17
45)	S	2-Fluorobiphenyl	1.738	1.580	1.520	1.409	1.347	1.221	1.114	1.418	15.11
46)		1,1'-Biphenyl	1.773	1.621	1.612	1.581	1.547	1.435	1.316	1.555	9.38
47)		2-Chloronaphth...	1.289	1.176	1.191	1.185	1.157	1.085	1.014	1.157	7.51
48)		2-Nitroaniline	0.119	0.156	0.228	0.295	0.316	0.315	0.316	0.249	33.25
49)		Acenaphthylene	2.149	2.025	2.025	1.997	1.965	1.828	1.675	1.952	7.93
50)		Dimethylphthalate	1.362	1.284	1.301	1.321	1.309	1.227	1.152	1.280	5.42
51)		2,6-Dinitrotol...	0.098	0.137	0.193	0.241	0.259	0.254	0.249	0.204	31.45
52)	C	Acenaphthene	1.271	1.175	1.156	1.172	1.138	1.072	0.995	1.140	7.63
53)		3-Nitroaniline	0.143	0.187	0.252	0.298	0.311	0.310	0.299	0.257	26.09
54)	P	2,4-Dinitrophenol		0.025	0.036	0.057	0.069	0.076	0.085	0.058	40.36
55)		Dibenzofuran	1.930	1.775	1.751	1.743	1.713	1.581	1.466	1.709	8.67
56)	P	4-Nitrophenol		0.131	0.177	0.216	0.233	0.230	0.220	0.201	19.82
57)		2,4-Dinitrotol...	0.104	0.145	0.213	0.285	0.311	0.309	0.306	0.239	35.96
58)		Fluorene	1.493	1.403	1.351	1.317	1.273	1.186	1.071	1.299	10.74
59)		2,3,4,6-Tetrac...	0.252	0.267	0.312	0.334	0.334	0.323	0.305	0.304	10.65
60)		Diethylphthalate	1.324	1.264	1.302	1.313	1.300	1.215	1.126	1.263	5.61
61)		4-Chlorophenyl...	0.702	0.642	0.632	0.632	0.614	0.570	0.526	0.617	9.06
62)		4-Nitroaniline	0.144	0.177	0.233	0.272	0.287	0.281	0.269	0.238	23.57
63)		Azobenzene	1.346	1.260	1.275	1.279	1.278	1.198	1.098	1.248	6.33
64)	I	Phenanthrene-d10	----------ISTD----------								
65)		4,6-Dinitro-2-...		0.022	0.034	0.055	0.065	0.071	0.075	0.054	39.23
66)	c	n-Nitrosodiphe...	0.727	0.686	0.693	0.712	0.689	0.653	0.623	0.683	5.11
67)		4-Bromophenyl-...	0.224	0.219	0.223	0.236	0.225	0.220	0.211	0.223	3.39
68)		Hexachlorobenzene	0.256	0.245	0.248	0.261	0.255	0.244	0.237	0.249	3.28
69)		Atrazine	0.166	0.172	0.186	0.190	0.190	0.182	0.174	0.180	5.29
70)	C	Pentachlorophenol		0.093	0.121	0.144	0.145	0.144	0.144	0.132	16.25
71)		Phenanthrene	1.218	1.114	1.119	1.112	1.064	0.995	0.940	1.080	8.45
72)		Anthracene	1.209	1.161	1.151	1.128	1.099	1.025	0.959	1.105	7.77
73)		Carbazole	1.125	1.047	1.061	1.033	0.996	0.935	0.857	1.008	8.77
74)		Di-n-butylphth...	0.986	1.006	1.088	1.069	1.045	0.988	0.912	1.013	5.88
75)	C	Fluoranthene	1.221	1.144	1.155	1.085	1.035	0.957	0.878	1.068	11.24
76)	I	Chrysene-d12	----------ISTD----------								
77)		Benzidine	0.702	0.748	0.885	0.963	0.945	0.895	0.810	0.850	11.67
78)		Pyrene	1.833	1.801	1.849	2.026	1.937	1.833	1.635	1.845	6.54
79)	S	Terphenyl-d14	1.472	1.413	1.422	1.481	1.392	1.290	1.155	1.375	8.41
80)		Butylbenzylpht...	0.218	0.293	0.405	0.505	0.525	0.524	0.509	0.425	29.32
81)		Benzo(a)anthra...	1.379	1.320	1.332	1.438	1.388	1.316	1.225	1.343	5.06
82)		3,3'-Dichlorob...	0.286	0.317	0.380	0.442	0.450	0.451	0.435	0.394	17.37
83)		Chrysene	1.302	1.210	1.249	1.232	1.227	1.197	1.138	1.222	4.12
84)		Bis(2-ethylhex...	0.384	0.448	0.549	0.695	0.726	0.722	0.707	0.604	23.69
85)	c	Di-n-octyl pht...		0.668	0.875	1.262	1.320	1.347	1.329	1.133	25.55


Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF043025.M

86)	I	Perylene-d12	----------ISTD----------							
87)		Indeno(1,2,3-c...	1.323	1.384	1.466	1.613	1.556	1.490	1.422	1.465	6.78
88)		Benzo(b)fluora...	1.317	1.220	1.208	1.302	1.271	1.177	1.202	1.242	4.36
89)		Benzo(k)fluora...	1.211	1.159	1.222	1.145	1.146	1.110	0.962	1.136	7.59
90)	C	Benzo(a)pyrene	1.121	1.098	1.146	1.203	1.180	1.137	1.076	1.137	3.91
91)		Dibenzo(a,h)an...	1.075	1.131	1.197	1.312	1.271	1.192	1.144	1.189	6.90
92)		Benzo(g,h,i)pe...	1.076	1.138	1.204	1.308	1.267	1.213	1.162	1.195	6.55


(#) = Out of Range



Response = 1.466e+000 * Amt - 4.448e-001
Coef of Det (r^2) = 0.997236 Curve Fit: Linear
Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M
Calibration Table Last Updated: Wed Apr 30 16:00:01 2025




Response = 7.529e-001 * Amt - 1.300e-001
Coef of Det (r^2) = 0.997585 Curve Fit: Linear
Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M
Calibration Table Last Updated: Wed Apr 30 16:00:01 2025




Response = 5.468e-001 * Amt - 1.016e-001
Coef of Det (r^2) = 0.997533 Curve Fit: Linear
Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M
Calibration Table Last Updated: Wed Apr 30 16:00:01 2025


4,6-Dinitro-2-methylphenol




Response = 8.458e-002 * Amt - 4.487e-002
Coef of Det (r^2) = 0.991555 Curve Fit: Linear
Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M
Calibration Table Last Updated: Wed Apr 30 16:00:01 2025




Response = 2.879e-001 * Amt - 3.958e-002
Coef of Det (r^2) = 0.996450 Curve Fit: Linear
Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M
Calibration Table Last Updated: Wed Apr 30 16:00:01 2025




Response = 3.318e-001 * Amt - 8.340e-002
Coef of Det (r^2) = 0.997011 Curve Fit: Linear
Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M
Calibration Table Last Updated: Wed Apr 30 16:00:01 2025




R = 1.037e-002 A*A + 4.909e-002 A - 1.884e-002
Coef of Det (r^2) = 0.998097 Curve Fit: Quadratic
Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M
Calibration Table Last Updated: Wed Apr 30 16:00:01 2025




Response = 3.198e-001 * Amt - 5.099e-002
Coef of Det (r^2) = 0.997645 Curve Fit: Linear
Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M
Calibration Table Last Updated: Wed Apr 30 16:00:01 2025




Response = 2.679e-001 * Amt - 5.390e-002
Coef of Det (r^2) = 0.997116 Curve Fit: Linear
Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M
Calibration Table Last Updated: Wed Apr 30 16:00:01 2025



Response = 3.387e-001 * Amt - 8.076e-002
Coef of Det (r^2) = 0.998123 Curve Fit: Linear
Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M
Calibration Table Last Updated: Wed Apr 30 16:00:01 2025



Response = 1.978e-001 * Amt - 1.069e-001
Coef of Det (r^2) = 0.989640 Curve Fit: Linear
Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M
Calibration Table Last Updated: Wed Apr 30 16:00:01 2025



R = 8.463e-003 A*A + 1.177e-001 A - 2.955e-002
Coef of Det (r^2) = 0.997290 Curve Fit: Quadratic
Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M
Calibration Table Last Updated: Wed Apr 30 16:00:01 2025

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142239.D
Acq On : 30 Apr 2025 11:24
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Quant Time: Apr 30 15:56:10 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	204993	20.000	ng	0.00
21) Naphthalene-d8	8.192	136	780462	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	414360	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	737147	20.000	ng	0.00
76) Chrysene-d12	14.074	240	525717	20.000	ng	0.00
86) Perylene-d12	15.568	264	366865	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						Qvalue
19) n-Nitroso-di-n-propyla...	7.304	70	24421	2.550	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed


ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142239.D
Acq On : 30 Apr 2025 11:24
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Quant Time: Apr 30 15:56:10 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142239.D
Acq: 30 Apr 2025 11:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:152 Resp: 204993

Ion	Ratio	Lower	Upper
152	100		
150	154.4	130.2	195.2
115	57.4	47.0	70.4







#5
2-Fluorophenol
Concen: 0.000 ng
Expected RT: 5.52 min

Lab File: BF142239.D
Acq: 30 Apr 2025 11:24

Tgt Ion: 112

Sig	Exp Ratio
112	100
64	60.5
63	31.0






#7
Phenol-d6
Concen: 0.000 ng
Expected RT: 6.53 min

Lab File: BF142239.D
Acq: 30 Apr 2025 11:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion: 99


Sig	Exp Ratio
99	100
42	20.4
71	31.2




#19
n-Nitroso-di-n-propylamine
Concen: 2.550 ng
RT: 7.304 min Scan# 869
Delta R.T. -0.018 min
Lab File: BF142239.D
Acq: 30 Apr 2025 11:24

Tgt Ion: 70 Resp: 24421

Ion	Ratio	Lower	Upper
70	100		
42	67.3	53.4	80.2
101	9.1	7.7	11.5
130	20.5	17.2	25.8







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.192 min Scan# 1
Delta R.T. -0.001 min
Lab File: BF142239.D
Acq: 30 Apr 2025 11:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:136 Resp: 780462

Ion	Ratio	Lower	Upper
136	100		
137	10.7	8.7	13.1
54	9.2	7.7	11.5
68	6.2	5.0	7.6







#23
Nitrobenzene-d5
Concen: 0.000 ng
Expected RT: 7.47 min

Lab File: BF142239.D
Acq: 30 Apr 2025 11:24

Tgt Ion: 82

Sig	Exp Ratio
82	100
128	46.3
54	60.0







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.945 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142239.D
Acq: 30 Apr 2025 11:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:164 Resp: 414360

Ion	Ratio	Lower	Upper
164	100		
162	98.5	78.6	117.8
160	44.5	35.7	53.5







#42
2,4,6-Tribromophenol
Concen: 0.000 ng
Expected RT: 10.73 min

Lab File: BF142239.D
Acq: 30 Apr 2025 11:24

Tgt Ion: 330

Sig	Exp Ratio
330	100
332	95.6
141	30.4






#45
2-Fluorobiphenyl
Concen: 0.000 ng
Expected RT: 9.27 min

Lab File: BF142239.D
Acq: 30 Apr 2025 11:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion: 172


Sig	Exp Ratio
172	100
171	36.2
170	24.2




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BF142239.D
Acq: 30 Apr 2025 11:24

Tgt Ion:188 Resp: 737147

Ion	Ratio	Lower	Upper
188	100		
94	9.9	8.2	12.2
80	10.3	8.6	12.8







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.074 min Scan# 2(
Delta R.T. -0.000 min
Lab File: BF142239.D
Acq: 30 Apr 2025 11:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:240 Resp: 525717

Ion	Ratio	Lower	Upper
240	100		
120	11.8	10.1	15.1
236	24.9	20.5	30.7







#79
Terphenyl-d14
Concen: 0.000 ng
Expected RT: 13.02 min

Lab File: BF142239.D
Acq: 30 Apr 2025 11:24

Tgt Ion: 244

Sig	Exp Ratio
244	100
212	6.6
122	11.4






#86
Perylene-d12
Concen: 20.000 ng
RT: 15.568 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142239.D
Acq: 30 Apr 2025 11:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:264 Resp: 366865

Ion	Ratio	Lower	Upper
264	100		
260	23.7	18.4	27.6
265	21.3	17.3	25.9



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142240.D
Acq On : 30 Apr 2025 11:52
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Apr 30 15:40:24 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.910	152	187794	20.000	ng		0.00
21) Naphthalene-d8	8.192	136	726978	20.000	ng		0.00
39) Acenaphthene-d10	9.945	164	380022	20.000	ng		0.00
64) Phenanthrene-d10	11.433	188	664931	20.000	ng		0.00
76) Chrysene-d12	14.074	240	454911	20.000	ng		0.00
86) Perylene-d12	15.568	264	322126	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.510	112	124307	10.571	ng		0.00
7) Phenol-d6	6.510	99	146414	10.311	ng		-0.02
23) Nitrobenzene-d5	7.463	82	78765	7.118	ng		-0.01
42) 2,4,6-Tribromophenol	10.727	330	26375	7.613	ng		0.00
45) 2-Fluorobiphenyl	9.263	172	330323	12.256	ng		0.00
79) Terphenyl-d14	13.016	244	334852	10.706	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.699	88	26443	5.035	ng		97
3) Pyridine	3.469	79	68779	5.039	ng		99
4) n-Nitrosodimethylamine	3.393	42	35057	4.963	ng	#	97
6) Aniline	6.563	93	103866	5.168	ng		99
8) 2-Chlorophenol	6.687	128	58780	4.838	ng		99
9) Benzaldehyde	6.457	77	54160	5.457	ng		98
10) Phenol	6.528	94	78823	5.157	ng		96
11) bis(2-Chloroethyl)ether	6.634	93	63148	5.331	ng		99
12) 1,3-Dichlorobenzene	6.845	146	73777	5.367	ng		100
13) 1,4-Dichlorobenzene	6.928	146	75604	5.459	ng		97
14) 1,2-Dichlorobenzene	7.075	146	69920	5.289	ng		99
15) Benzyl Alcohol	7.039	79	49136	4.879	ng		97
16) 2,2'-oxybis(1-Chloropr...	7.181	45	111179	5.198	ng		100
17) 2-Methylphenol	7.145	107	50945	5.118	ng		98
18) Hexachloroethane	7.422	117	21958	4.844	ng		96
19) n-Nitroso-di-n-propyla...	7.310	70	46341	5.281	ng		99
20) 3+4-Methylphenols	7.292	107	63725	5.119	ng		93
22) Acetophenone	7.310	105	92806	5.518	ng		96
24) Nitrobenzene	7.481	77	41660	3.837	ng		94
25) Isophorone	7.722	82	119954	5.165	ng		100
26) 2-Nitrophenol	7.804	139	8775	6.901	ng		99
27) 2,4-Dimethylphenol	7.828	122	57217	4.887	ng		97
28) bis(2-Chloroethoxy)met...	7.934	93	77155	5.268	ng		100
29) 2,4-Dichlorophenol	8.034	162	44643	4.560	ng		98
30) 1,2,4-Trichlorobenzene	8.128	180	58177	5.326	ng		99
31) Naphthalene	8.210	128	200969	5.543	ng		99
33) 4-Chloroaniline	8.251	127	81024	5.367	ng		98
34) Hexachlorobutadiene	8.334	225	32676	5.127	ng		97
35) Caprolactam	8.575	113	12427	4.223	ng		100
36) 4-Chloro-3-methylphenol	8.722	107	47673	4.743	ng		99
37) 2-Methylnaphthalene	8.904	142	123458	5.539	ng		97
38) 1-Methylnaphthalene	8.998	142	127274	5.512	ng		99
40) 1,2,4,5-Tetrachloroben...	9.063	216	58967	5.513	ng		97
41) Hexachlorocyclopentadiene	9.057	237	23228	3.827	ng		98
43) 2,4,6-Trichlorophenol	9.175	196	26171	3.949	ng		98
44) 2,4,5-Trichlorophenol	9.204	196	30587	4.390	ng		95

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142240.D
Acq On : 30 Apr 2025 11:52
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025

Quant Time: Apr 30 15:40:24 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
46) 1,1'-Biphenyl	9.363	154	168456	5.702	ng		98
47) 2-Chloronaphthalene	9.386	162	122469	5.572	ng		98
48) 2-Nitroaniline	9.475	65	11346	6.532	ng		89
49) Acenaphthylene	9.804	152	204157	5.504	ng		99
50) Dimethylphthalate	9.657	163	129376	5.322	ng		100
51) 2,6-Dinitrotoluene	9.716	165	9271	5.846	ng	#	81
52) Acenaphthene	9.975	154	120775	5.577	ng		99
53) 3-Nitroaniline	9.886	138	13576	5.423	ng	#	88
55) Dibenzofuran	10.151	168	183405	5.649	ng		98
57) 2,4-Dinitrotoluene	10.116	165	9928	6.601	ng	#	85
58) Fluorene	10.492	166	141835	5.746	ng		98
59) 2,3,4,6-Tetrachlorophenol	10.263	232	23974m	4.154	ng		
60) Diethylphthalate	10.363	149	125748	5.238	ng		99
61) 4-Chlorophenyl-phenyle...	10.486	204	66676	5.688	ng		98
62) 4-Nitroaniline	10.492	138	13695	5.254	ng	#	76
63) Azobenzene	10.645	77	127873	5.394	ng		99
66) n-Nitrosodiphenylamine	10.598	169	120800	5.317	ng		100
67) 4-Bromophenyl-phenylether	10.975	248	37248	5.030	ng		97
68) Hexachlorobenzene	11.039	284	42558	5.131	ng		98
69) Atrazine	11.122	200	27629	4.614	ng		98
71) Phenanthrene	11.457	178	202460	5.637	ng		98
72) Anthracene	11.504	178	200946	5.472	ng		98
73) Carbazole	11.657	167	186963	5.581	ng		98
74) Di-n-butylphthalate	11.998	149	163860	4.864	ng		99
75) Fluoranthene	12.645	202	202911	5.715	ng		98
77) Benzidine	12.769	184	79877	4.132	ng		98
78) Pyrene	12.874	202	208462	4.968	ng		99
80) Butylbenzylphthalate	13.498	149	24840	5.714	ng		96
81) Benzo(a)anthracene	14.063	228	156840	5.136	ng		99
82) 3,3'-Dichlorobenzidine	14.027	252	32503	3.623	ng		94
83) Chrysene	14.098	228	148119	5.328	ng		98
84) Bis(2-ethylhexyl)phtha...	14.063	149	43692	6.006	ng		99
87) Indeno(1,2,3-cd)pyrene	17.068	276	106554	4.516	ng		98
88) Benzo(b)fluoranthene	15.121	252	106072	5.301	ng		100
89) Benzo(k)fluoranthene	15.151	252	97496	5.327	ng		99
90) Benzo(a)pyrene	15.498	252	90284	4.928	ng		99
91) Dibenzo(a,h)anthracene	17.086	278	86557	4.520	ng		100
92) Benzo(g,h,i)perylene	17.521	276	86631	4.500	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142240.D
Acq On : 30 Apr 2025 11:52
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1


Quant Time: Apr 30 15:40:24 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:152 Resp: 18779


Ion	Ratio	Lower	Upper
152	100		
150	157.6	130.2	195.2
115	58.9	47.0	70.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#2
1,4-Dioxane
Concen: 5.035 ng
RT: 2.699 min Scan# 86
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion: 88 Resp: 26443

Ion	Ratio	Lower	Upper
88	100		
58	91.8	72.2	108.2
43	37.4	27.0	40.6





#3
Pyridine
Concen: 5.039 ng
RT: 3.469 min Scan# 21
Delta R.T. 0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52


Tgt Ion: 79 Resp: 6877


Ion	Ratio	Lower	Upper
79	100		
52	88.7	71.5	107.3
51	42.1	34.5	51.7

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#4
n-Nitrosodimethylamine
Concen: 4.963 ng
RT: 3.393 min Scan# 204
Delta R.T. -0.018 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion: 42 Resp: 35057

Ion	Ratio	Lower	Upper
42	100		
74	110.4	89.5	134.3
44	18.5	6.6	9.8#






#5
2-Fluorophenol
Concen: 10.571 ng
RT: 5.510 min Scan# 5
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:112 Resp: 12430


Ion	Ratio	Lower	Upper
112	100		
64	58.2	48.4	72.6
63	30.8	24.8	37.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#6
Aniline
Concen: 5.168 ng
RT: 6.563 min Scan# 743
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion: 93 Resp: 103866

Ion	Ratio	Lower	Upper
93	100		
66	39.6	32.1	48.1
65	20.7	16.6	24.8






#7
Phenol-d6
Concen: 10.311 ng
RT: 6.510 min Scan# 7
Delta R.T. -0.018 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion: 99 Resp: 14641


Ion	Ratio	Lower	Upper
99	100		
42	21.1	16.3	24.5
71	30.6	25.0	37.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#8
2-Chlorophenol
Concen: 4.838 ng
RT: 6.687 min Scan# 764
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:128 Resp: 58780

Ion	Ratio	Lower	Upper
128	100		
130	33.7	13.1	53.1
64	49.0	28.6	68.6






#9
Benzaldehyde
Concen: 5.457 ng
RT: 6.457 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion: 77 Resp: 5416


Ion	Ratio	Lower	Upper
77	100		
105	99.9	79.1	119.1
106	98.3	76.0	116.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#10
Phenol
Concen: 5.157 ng
RT: 6.528 min Scan# 737
Delta R.T. -0.012 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion: 94 Resp: 78823

Ion	Ratio	Lower	Upper
94	100		
65	28.3	8.9	48.9
66	34.6	18.7	58.7






#11
bis(2-Chloroethyl)ether
Concen: 5.331 ng
RT: 6.634 min Scan# 7
Delta R.T. -0.012 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion: 93 Resp: 6314


Ion	Ratio	Lower	Upper
93	100		
63	81.0	62.2	102.2
95	32.0	12.7	52.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#12
1,3-Dichlorobenzene
Concen: 5.367 ng
RT: 6.845 min Scan# 791
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:146 Resp: 73777

Ion	Ratio	Lower	Upper
146	100		
148	62.7	50.5	75.7
75	29.2	23.2	34.8






#13
1,4-Dichlorobenzene
Concen: 5.459 ng
RT: 6.928 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:146 Resp: 7560


Ion	Ratio	Lower	Upper
146	100		
148	60.5	50.6	76.0
111	39.6	33.4	50.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#14
1,2-Dichlorobenzene
Concen: 5.289 ng
RT: 7.075 min Scan# 830
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:146 Resp: 69920

Ion	Ratio	Lower	Upper
146	100		
148	64.6	50.5	75.7
111	44.0	34.9	52.3






#15
Benzyl Alcohol
Concen: 4.879 ng
RT: 7.039 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion: 79 Resp: 4913


Ion	Ratio	Lower	Upper
79	100		
108	82.4	64.5	96.7
77	65.4	49.2	73.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 5.198 ng
RT: 7.181 min Scan# 848
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion: 45 Resp: 111179

Ion	Ratio	Lower	Upper
45	100		
77	11.5	0.0	31.3
79	8.5	0.0	28.5






#17
2-Methylphenol
Concen: 5.118 ng
RT: 7.145 min Scan# 84
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:107 Resp: 5094


Ion	Ratio	Lower	Upper
107	100		
108	111.4	91.0	136.4
77	41.1	34.4	51.6
79	42.1	34.4	51.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#18
Hexachloroethane
Concen: 4.844 ng
RT: 7.422 min Scan# 889
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:117 Resp: 21958

Ion	Ratio	Lower	Upper
117	100		
119	97.2	77.1	115.7
201	91.6	79.4	119.2






#19
n-Nitroso-di-n-propylamine
Concen: 5.281 ng
RT: 7.310 min Scan# 8
Delta R.T. -0.012 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion: 70 Resp: 4634


Ion	Ratio	Lower	Upper
70	100		
42	68.0	53.4	80.2
101	8.9	7.7	11.5
130	21.2	17.2	25.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#20
3+4-Methylphenols
Concen: 5.119 ng
RT: 7.292 min Scan# 867
Delta R.T. -0.012 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:107 Resp: 63725

Ion	Ratio	Lower	Upper
107	100		
108	91.8	71.2	111.2
77	48.0	41.1	81.1
79	26.0	8.1	48.1






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.192 min Scan# 1
Delta R.T. -0.001 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:136 Resp: 72697


Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.7	13.1
54	8.9	7.7	11.5
68	6.0	5.0	7.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#22
Acetophenone
Concen: 5.518 ng
RT: 7.310 min Scan# 870
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:105 Resp: 92806

Ion	Ratio	Lower	Upper
105	100		
71	7.3	5.7	8.5
51	33.0	30.1	45.1
120	23.8	19.0	28.4





#23
Nitrobenzene-d5
Concen: 7.118 ng
RT: 7.463 min Scan# 8
Delta R.T. -0.012 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52


Tgt Ion: 82 Resp: 7876


Ion	Ratio	Lower	Upper
82	100		
128	40.7	37.0	55.6
54	62.4	48.0	72.0

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#24
Nitrobenzene
Concen: 3.837 ng
RT: 7.481 min Scan# 899
Delta R.T. -0.012 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion: 77 Resp: 41660

Ion	Ratio	Lower	Upper
77	100		
123	42.7	38.6	58.0
65	14.1	11.5	17.3






#25
Isophorone
Concen: 5.165 ng
RT: 7.722 min Scan# 94
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion: 82 Resp: 11995


Ion	Ratio	Lower	Upper
82	100		
95	7.2	5.6	8.4
138	19.0	15.3	22.9



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#26
2-Nitrophenol
Concen: 6.901 ng
RT: 7.804 min Scan# 954
Delta R.T. 0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:139 Resp: 8775

Ion	Ratio	Lower	Upper
139	100		
109	27.5	21.8	32.8
65	52.6	41.1	61.7






#27
2,4-Dimethylphenol
Concen: 4.887 ng
RT: 7.828 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:122 Resp: 5721


Ion	Ratio	Lower	Upper
122	100		
107	111.3	86.0	129.0
121	59.1	46.3	69.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#28
bis(2-Chloroethoxy)methane
Concen: 5.268 ng
RT: 7.934 min Scan# 976
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion: 93 Resp: 77155

Ion	Ratio	Lower	Upper
93	100		
95	32.8	26.2	39.2
123	11.8	9.7	14.5





#29
2,4-Dichlorophenol
Concen: 4.560 ng
RT: 8.034 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52


Tgt Ion:162 Resp: 4464


Ion	Ratio	Lower	Upper
162	100		
164	60.8	43.0	83.0
98	36.0	16.7	56.7

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#30
1,2,4-Trichlorobenzene
Concen: 5.326 ng
RT: 8.128 min Scan# 1009
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:180 Resp: 58177

Ion	Ratio	Lower	Upper
180	100		
182	94.8	75.7	113.5
145	30.3	25.4	38.0






#31
Naphthalene
Concen: 5.543 ng
RT: 8.210 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:128 Resp: 20096


Ion	Ratio	Lower	Upper
128	100		
129	11.1	9.0	13.4
127	12.4	10.6	15.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#33
4-Chloroaniline
Concen: 5.367 ng
RT: 8.251 min Scan# 1030
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:127 Resp: 81024

Ion	Ratio	Lower	Upper
127	100		
129	31.2	26.1	39.1
65	32.3	25.5	38.3
92	18.7	15.4	23.2






#34
Hexachlorobutadiene
Concen: 5.127 ng
RT: 8.334 min Scan# 1(
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:225 Resp: 3267(


Ion	Ratio	Lower	Upper
225	100		
223	64.0	49.7	74.5
227	67.0	51.0	76.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#35
Caprolactam
Concen: 4.223 ng
RT: 8.575 min Scan# 1085
Delta R.T. -0.047 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:113 Resp: 12427

Ion	Ratio	Lower	Upper
113	100		
55	197.9	177.7	217.7
56	146.0	127.1	167.1






#36
4-Chloro-3-methylphenol
Concen: 4.743 ng
RT: 8.722 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:107 Resp: 4767


Ion	Ratio	Lower	Upper
107	100		
144	26.6	21.0	31.4
142	83.1	66.1	99.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#37
2-Methylnaphthalene
Concen: 5.539 ng
RT: 8.904 min Scan# 1141
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:142 Resp: 123458

Ion	Ratio	Lower	Upper
142	100		
141	85.8	70.5	105.7






#38
1-Methylnaphthalene
Concen: 5.512 ng
RT: 8.998 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:142 Resp: 12727


Ion	Ratio	Lower	Upper
142	100		
141	90.5	72.8	109.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.945 min Scan# 1318
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:164 Resp: 380022

Ion	Ratio	Lower	Upper
164	100		
162	100.5	78.6	117.8
160	44.8	35.7	53.5






#40
1,2,4,5-Tetrachlorobenzene
Concen: 5.513 ng
RT: 9.063 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:216 Resp: 5896


Ion	Ratio	Lower	Upper
216	100		
214	76.4	63.8	95.6
179	20.8	16.7	25.1
108	19.3	15.6	23.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#41
Hexachlorocyclopentadiene
Concen: 3.827 ng
RT: 9.057 min Scan# 1167
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:237 Resp: 23228

Ion	Ratio	Lower	Upper
237	100		
235	61.6	43.5	83.5
272	12.0	0.0	32.9






#42
2,4,6-Tribromophenol
Concen: 7.613 ng
RT: 10.727 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:330 Resp: 2637


Ion	Ratio	Lower	Upper
330	100		
332	94.2	76.5	114.7
141	34.3	24.3	36.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#43
2,4,6-Trichlorophenol
Concen: 3.949 ng
RT: 9.175 min Scan# 1187
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:196 Resp: 26171

Ion	Ratio	Lower	Upper
196	100		
198	98.3	77.4	116.2
200	33.2	24.6	37.0






#44
2,4,5-Trichlorophenol
Concen: 4.390 ng
RT: 9.204 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:196 Resp: 3058


Ion	Ratio	Lower	Upper
196	100		
198	99.9	75.2	112.8
97	52.4	39.8	59.8
132	32.5	23.8	35.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#45
2-Fluorobiphenyl
Concen: 12.256 ng
RT: 9.263 min Scan# 1202
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:172 Resp: 330323

Ion	Ratio	Lower	Upper
172	100		
171	35.0	29.0	43.4
170	22.9	19.4	29.0






#46
1,1'-Biphenyl
Concen: 5.702 ng
RT: 9.363 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:154 Resp: 16845


Ion	Ratio	Lower	Upper
154	100		
153	38.7	20.5	60.5
76	13.0	0.0	32.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#47
2-Chloronaphthalene
Concen: 5.572 ng
RT: 9.386 min Scan# 1223
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:162 Resp: 122469

Ion	Ratio	Lower	Upper
162	100		
127	39.2	32.6	49.0
164	32.9	26.1	39.1





#48
2-Nitroaniline
Concen: 6.532 ng
RT: 9.475 min Scan# 1
Delta R.T. -0.012 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52


Tgt Ion: 65 Resp: 1134


Ion	Ratio	Lower	Upper
65	100		
92	64.6	54.3	81.5
138	94.3	88.6	133.0

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#49
Acenaphthylene
Concen: 5.504 ng
RT: 9.804 min Scan# 1294
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:152 Resp: 204157

Ion	Ratio	Lower	Upper
152	100		
151	19.6	16.2	24.2
153	12.9	10.7	16.1





#50
Dimethylphthalate
Concen: 5.322 ng
RT: 9.657 min Scan# 1
Delta R.T. -0.012 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52


Tgt Ion:163 Resp: 12937


Ion	Ratio	Lower	Upper
163	100		
194	3.9	3.2	4.8
164	10.1	8.2	12.2

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#51
2,6-Dinitrotoluene
Concen: 5.846 ng
RT: 9.716 min Scan# 1279
Delta R.T. -0.012 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:165 Resp: 9271

Ion	Ratio	Lower	Upper
165	100		
63	77.9	48.5	72.7#
89	65.9	43.9	65.9#





#52
Acenaphthene
Concen: 5.577 ng
RT: 9.975 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52


Tgt Ion:154 Resp: 12077


Ion	Ratio	Lower	Upper
154	100		
153	110.1	88.6	133.0
152	52.8	42.1	63.1

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#53
3-Nitroaniline
Concen: 5.423 ng
RT: 9.886 min Scan# 1308
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:138 Resp: 13576

Ion	Ratio	Lower	Upper
138	100		
108	37.8	11.7	17.5#
92	120.9	91.2	136.8






#55
Dibenzofuran
Concen: 5.649 ng
RT: 10.151 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:168 Resp: 18340


Ion	Ratio	Lower	Upper
168	100		
139	35.1	29.4	44.0
169	13.2	11.0	16.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#57
2,4-Dinitrotoluene
Concen: 6.601 ng
RT: 10.116 min Scan# 1347
Delta R.T. -0.012 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:165 Resp: 9928

Ion	Ratio	Lower	Upper
165	100		
63	55.7	38.8	58.2
89	86.5	57.1	85.7#
182	2.6	2.2	3.2






#58
Fluorene
Concen: 5.746 ng
RT: 10.492 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:166 Resp: 14183


Ion	Ratio	Lower	Upper
166	100		
165	94.0	73.9	110.9
167	13.3	10.8	16.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#59
2,3,4,6-Tetrachlorophenol
Concen: 4.154 ng m
RT: 10.263 min Scan# 1372
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:232 Resp: 23974

Ion	Ratio	Lower	Upper
232	100		
131	33.9	35.9	53.9#
130	2.0	1.8	2.6
166	20.6	24.0	36.0#






#60
Diethylphthalate
Concen: 5.238 ng
RT: 10.363 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:149 Resp: 12574


Ion	Ratio	Lower	Upper
149	100		
177	21.0	17.6	26.4
150	12.8	10.2	15.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#61
4-Chlorophenyl-phenylether
Concen: 5.688 ng
RT: 10.486 min Scan# 1410
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:204 Resp: 66676

Ion	Ratio	Lower	Upper
204	100		
206	31.7	25.6	38.4
141	58.7	48.9	73.3






#62
4-Nitroaniline
Concen: 5.254 ng
RT: 10.492 min Scan# 14
Delta R.T. -0.012 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:138 Resp: 1369


Ion	Ratio	Lower	Upper
138	100		
92	54.7	29.2	69.2
108	97.8	49.1	89.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#63
Azobenzene
Concen: 5.394 ng
RT: 10.645 min Scan# 1437
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion: 77 Resp: 127873

Ion	Ratio	Lower	Upper
77	100		
182	26.1	7.3	47.3
105	22.7	2.5	42.5
51	34.7	14.4	54.4






#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:188 Resp: 66493


Ion	Ratio	Lower	Upper
188	100		
94	10.0	8.2	12.2
80	10.5	8.6	12.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#66
n-Nitrosodiphenylamine
Concen: 5.317 ng
RT: 10.598 min Scan# 1429
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:169 Resp: 120800

Ion	Ratio	Lower	Upper
169	100		
168	66.4	53.0	79.6
167	36.3	28.8	43.2





#67
4-Bromophenyl-phenylether
Concen: 5.030 ng
RT: 10.975 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52


Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:248 Resp: 3724

Ion	Ratio	Lower	Upper
248	100		
250	97.7	77.2	115.8
141	77.2	58.4	87.6

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#68
Hexachlorobenzene
Concen: 5.131 ng
RT: 11.039 min Scan# 1504
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:284 Resp: 42558

Ion	Ratio	Lower	Upper
284	100		
142	32.4	24.6	37.0
249	28.6	23.2	34.8






#69
Atrazine
Concen: 4.614 ng
RT: 11.122 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:200 Resp: 2762


Ion	Ratio	Lower	Upper
200	100		
173	28.4	8.6	48.6
215	46.6	28.5	68.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#71
Phenanthrene
Concen: 5.637 ng
RT: 11.457 min Scan# 1575
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:178 Resp: 202460

Ion	Ratio	Lower	Upper
178	100		
176	18.6	15.5	23.3
179	14.7	12.6	19.0






#72
Anthracene
Concen: 5.472 ng
RT: 11.504 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:178 Resp: 20094


Ion	Ratio	Lower	Upper
178	100		
176	18.2	15.3	22.9
179	15.3	12.7	19.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#73
Carbazole
Concen: 5.581 ng
RT: 11.657 min Scan# 1609
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:167 Resp: 186963

Ion	Ratio	Lower	Upper
167	100		
166	20.5	17.3	25.9
139	12.3	10.3	15.5






#74
Di-n-butylphthalate
Concen: 4.864 ng
RT: 11.998 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:149 Resp: 16386


Ion	Ratio	Lower	Upper
149	100		
150	8.9	7.5	11.3
104	4.6	3.6	5.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#75
Fluoranthene
Concen: 5.715 ng
RT: 12.645 min Scan# 1777
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:202 Resp: 202911

Ion	Ratio	Lower	Upper
202	100		
101	11.3	0.0	32.4
203	17.2	0.0	37.5






#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.074 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:240 Resp: 45491


Ion	Ratio	Lower	Upper
240	100		
120	12.1	10.1	15.1
236	25.2	20.5	30.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#77
Benzidine
Concen: 4.132 ng
RT: 12.769 min Scan# 1798
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:184 Resp: 79877

Ion	Ratio	Lower	Upper
184	100		
185	13.8	11.8	17.6
183	10.7	8.9	13.3






#78
Pyrene
Concen: 4.968 ng
RT: 12.874 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:202 Resp: 20846


Ion	Ratio	Lower	Upper
202	100		
200	20.0	16.1	24.1
203	17.2	14.1	21.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#79
Terphenyl-d14
Concen: 10.706 ng
RT: 13.016 min Scan# 1840
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:244 Resp: 334852

Ion	Ratio	Lower	Upper
244	100		
212	6.3	5.3	7.9
122	12.4	9.1	13.7






#80
Butylbenzylphthalate
Concen: 5.714 ng
RT: 13.498 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:149 Resp: 2484


Ion	Ratio	Lower	Upper
149	100		
91	71.9	54.4	81.6
206	20.4	15.9	23.9



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#81
Benzo(a)anthracene
Concen: 5.136 ng
RT: 14.063 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:228 Resp: 156840

Ion	Ratio	Lower	Upper
228	100		
226	27.1	21.9	32.9
229	18.8	16.0	24.0





#82
3,3'-Dichlorobenzidine
Concen: 3.623 ng
RT: 14.027 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52


Tgt Ion:252 Resp: 3250


Ion	Ratio	Lower	Upper
252	100		
254	68.6	51.0	76.6
126	14.1	9.4	14.2

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#83
Chrysene
Concen: 5.328 ng
RT: 14.098 min Scan# 2024
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:228 Resp: 148119

Ion	Ratio	Lower	Upper
228	100		
226	29.2	24.0	36.0
229	18.8	15.7	23.5





#84
Bis(2-ethylhexyl)phthalate
Concen: 6.006 ng
RT: 14.063 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52


Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:149 Resp: 4369

Ion	Ratio	Lower	Upper
149	100		
167	26.5	21.7	32.5
279	3.3	3.0	4.4

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.568 min Scan# 2274
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:264 Resp: 322126

Ion	Ratio	Lower	Upper
264	100		
260	22.9	18.4	27.6
265	22.0	17.3	25.9






#87
Indeno(1,2,3-cd)pyrene
Concen: 4.516 ng
RT: 17.068 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:276 Resp: 10655


Ion	Ratio	Lower	Upper
276	100		
138	29.4	22.6	33.8
277	24.7	20.1	30.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#88
Benzo(b)fluoranthene
Concen: 5.301 ng
RT: 15.121 min Scan# 2198
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:252 Resp: 106072

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.7	26.5
125	11.6	9.0	13.4






#89
Benzo(k)fluoranthene
Concen: 5.327 ng
RT: 15.151 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:252 Resp: 9749


Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.8	26.8
125	10.8	9.0	13.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#90
Benzo(a)pyrene
Concen: 4.928 ng
RT: 15.498 min Scan# 2262
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:252 Resp: 90284

Ion	Ratio	Lower	Upper
252	100		
253	21.1	17.6	26.4
125	11.9	9.5	14.3





#91
Dibenzo(a,h)anthracene
Concen: 4.520 ng
RT: 17.086 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52


Tgt Ion:278 Resp: 8655


Ion	Ratio	Lower	Upper
278	100		
139	19.0	15.2	22.8
279	23.7	19.0	28.6

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025






#92
Benzo(g,h,i)perylene
Concen: 4.500 ng
RT: 17.521 min Scan# 2606
Delta R.T. -0.018 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:276 Resp: 86631

Ion	Ratio	Lower	Upper
276	100		
277	23.3	18.7	28.1
138	25.0	19.4	29.0



ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142241.D
Acq On : 30 Apr 2025 12:20
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Quant Time: Apr 30 15:41:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.910	152	198828	20.000	ng		0.00
21) Naphthalene-d8	8.192	136	764114	20.000	ng		0.00
39) Acenaphthene-d10	9.945	164	402628	20.000	ng		0.00
64) Phenanthrene-d10	11.433	188	692116	20.000	ng		0.00
76) Chrysene-d12	14.074	240	449534	20.000	ng		0.00
86) Perylene-d12	15.568	264	332067	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.510	112	246136	19.771	ng		0.00
7) Phenol-d6	6.516	99	299244	19.905	ng		-0.01
23) Nitrobenzene-d5	7.463	82	192172	16.523	ng		-0.01
42) 2,4,6-Tribromophenol	10.727	330	63470	17.292	ng		0.00
45) 2-Fluorobiphenyl	9.263	172	636078	22.275	ng		0.00
79) Terphenyl-d14	13.016	244	635233	20.552	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.699	88	53052	9.541	ng		97
3) Pyridine	3.469	79	138279	9.569	ng		99
4) n-Nitrosodimethylamine	3.399	42	69527	9.297	ng	#	98
6) Aniline	6.563	93	207232	9.739	ng		100
8) 2-Chlorophenol	6.687	128	122451	9.519	ng		98
9) Benzaldehyde	6.457	77	108198	10.297	ng		99
10) Phenol	6.528	94	157877	9.755	ng		99
11) bis(2-Chloroethyl)ether	6.634	93	124281	9.909	ng		99
12) 1,3-Dichlorobenzene	6.851	146	147587	10.141	ng		99
13) 1,4-Dichlorobenzene	6.928	146	147728	10.075	ng		99
14) 1,2-Dichlorobenzene	7.081	146	140999	10.073	ng		99
15) Benzyl Alcohol	7.039	79	101094	9.481	ng		98
16) 2,2'-oxybis(1-Chloropr...	7.181	45	227340	10.038	ng		99
17) 2-Methylphenol	7.145	107	102299	9.707	ng		99
18) Hexachloroethane	7.422	117	46369	9.662	ng		94
19) n-Nitroso-di-n-propyla...	7.310	70	93168	10.029	ng		99
20) 3+4-Methylphenols	7.298	107	131546	9.981	ng		97
22) Acetophenone	7.310	105	184253	10.423	ng		96
24) Nitrobenzene	7.481	77	99900	8.754	ng		95
25) Isophorone	7.722	82	240659	9.859	ng		100
26) 2-Nitrophenol	7.804	139	21744	9.530	ng		98
27) 2,4-Dimethylphenol	7.834	122	121707	9.889	ng		97
28) bis(2-Chloroethoxy)met...	7.934	93	154349	10.026	ng		99
29) 2,4-Dichlorophenol	8.034	162	96761	9.404	ng		98
30) 1,2,4-Trichlorobenzene	8.128	180	115815	10.087	ng		99
31) Naphthalene	8.210	128	397159	10.422	ng		99
32) Benzoic acid	7.881	122	18272	13.045	ng		96
33) 4-Chloroaniline	8.257	127	157999	9.957	ng		99
34) Hexachlorobutadiene	8.334	225	66829	9.976	ng		98
35) Caprolactam	8.586	113	27800	8.989	ng		98
36) 4-Chloro-3-methylphenol	8.722	107	101179	9.576	ng		99
37) 2-Methylnaphthalene	8.904	142	243429	10.390	ng		99
38) 1-Methylnaphthalene	9.004	142	250299	10.312	ng		99
40) 1,2,4,5-Tetrachloroben...	9.069	216	115685	10.209	ng		98
41) Hexachlorocyclopentadiene	9.057	237	53365	8.298	ng		98
43) 2,4,6-Trichlorophenol	9.175	196	66271	9.438	ng		97

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142241.D
Acq On : 30 Apr 2025 12:20
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Quant Time: Apr 30 15:41:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.210	196	66509	9.009	ng		98
46) 1,1'-Biphenyl	9.363	154	326241	10.423	ng		99
47) 2-Chloronaphthalene	9.392	162	236725	10.165	ng		99
48) 2-Nitroaniline	9.480	65	31455	9.382	ng		98
49) Acenaphthylene	9.804	152	407630	10.373	ng		98
50) Dimethylphthalate	9.657	163	258512	10.036	ng		99
51) 2,6-Dinitrotoluene	9.722	165	27624	9.147	ng		91
52) Acenaphthene	9.980	154	236444	10.305	ng		99
53) 3-Nitroaniline	9.886	138	37727	9.049	ng	#	97
54) 2,4-Dinitrophenol	9.986	184	4993	11.362	ng	#	1
55) Dibenzofuran	10.151	168	357431	10.392	ng		99
56) 4-Nitrophenol	10.028	139	26399	6.520	ng		91
57) 2,4-Dinitrotoluene	10.122	165	29245	9.405	ng		94
58) Fluorene	10.492	166	282449	10.800	ng		99
59) 2,3,4,6-Tetrachlorophenol	10.263	232	53660	8.775	ng		99
60) Diethylphthalate	10.363	149	254431	10.003	ng		98
61) 4-Chlorophenyl-phenyle...	10.486	204	129149	10.398	ng		99
62) 4-Nitroaniline	10.492	138	35703	8.911	ng		88
63) Azobenzene	10.645	77	253731	10.102	ng		99
65) 4,6-Dinitro-2-methylph...	10.527	198	7762	13.263	ng		95
66) n-Nitrosodiphenylamine	10.598	169	237471	10.042	ng		100
67) 4-Bromophenyl-phenylether	10.975	248	75692	9.820	ng		97
68) Hexachlorobenzene	11.039	284	84881	9.832	ng		96
69) Atrazine	11.122	200	59553	9.554	ng		98
70) Pentachlorophenol	11.227	266	32058	7.029	ng		99
71) Phenanthrene	11.457	178	385677	10.316	ng		99
72) Anthracene	11.510	178	401700	10.509	ng		98
73) Carbazole	11.657	167	362249	10.389	ng		98
74) Di-n-butylphthalate	11.998	149	348111	9.927	ng		99
75) Fluoranthene	12.645	202	395757	10.709	ng		99
77) Benzidine	12.769	184	168038	8.797	ng		99
78) Pyrene	12.874	202	404860	9.763	ng		100
80) Butylbenzylphthalate	13.498	149	65964	9.084	ng		98
81) Benzo(a)anthracene	14.063	228	296710	9.832	ng		99
82) 3,3'-Dichlorobenzidine	14.027	252	71321	8.046	ng		98
83) Chrysene	14.098	228	271875	9.897	ng		100
84) Bis(2-ethylhexyl)phtha...	14.063	149	100777	9.409	ng		99
85) Di-n-octyl phthalate	14.680	149	150072	10.622	ng		99
87) Indeno(1,2,3-cd)pyrene	17.068	276	229818	9.449	ng		100
88) Benzo(b)fluoranthene	15.121	252	202615	9.823	ng		99
89) Benzo(k)fluoranthene	15.157	252	192500	10.203	ng		98
90) Benzo(a)pyrene	15.498	252	182258	9.651	ng		99
91) Dibenzo(a,h)anthracene	17.092	278	187788	9.513	ng		99
92) Benzo(g,h,i)perylene	17.521	276	188891	9.518	ng		99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142241.D
Acq On : 30 Apr 2025 12:20
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Quant Time: Apr 30 15:41:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:152 Resp: 198828

Ion	Ratio	Lower	Upper
152	100		
150	155.6	130.2	195.2
115	58.2	47.0	70.4







#2
1,4-Dioxane
Concen: 9.541 ng
RT: 2.699 min Scan# 86
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion: 88 Resp: 53052

Ion	Ratio	Lower	Upper
88	100		
58	92.8	72.2	108.2
43	35.3	27.0	40.6







#3
Pyridine
Concen: 9.569 ng
RT: 3.469 min Scan# 21
Delta R.T. 0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 79 Resp: 138279

Ion	Ratio	Lower	Upper
79	100		
52	89.5	71.5	107.3
51	41.3	34.5	51.7







#4
n-Nitrosodimethylamine
Concen: 9.297 ng
RT: 3.399 min Scan# 205
Delta R.T. -0.012 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion: 42 Resp: 69527

Ion	Ratio	Lower	Upper
42	100		
74	113.7	89.5	134.3
44	12.7	6.6	9.8#







#5
2-Fluorophenol
Concen: 19.771 ng
RT: 5.510 min Scan# 5
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:112 Resp: 246136

Ion	Ratio	Lower	Upper
112	100		
64	60.2	48.4	72.6
63	31.3	24.8	37.2







#6
Aniline
Concen: 9.739 ng
RT: 6.563 min Scan# 743
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion: 93 Resp: 207232

Ion	Ratio	Lower	Upper
93	100		
66	40.2	32.1	48.1
65	20.5	16.6	24.8







#7
Phenol-d6
Concen: 19.905 ng
RT: 6.516 min Scan# 7
Delta R.T. -0.012 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 99 Resp: 299244

Ion	Ratio	Lower	Upper
99	100		
42	20.6	16.3	24.5
71	30.9	25.0	37.4







#8
2-Chlorophenol
Concen: 9.519 ng
RT: 6.687 min Scan# 764
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:128 Resp: 122451

Ion	Ratio	Lower	Upper
128	100		
130	31.1	13.1	53.1
64	49.2	28.6	68.6







#9
Benzaldehyde
Concen: 10.297 ng
RT: 6.457 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 77 Resp: 108198

Ion	Ratio	Lower	Upper
77	100		
105	100.7	79.1	119.1
106	95.9	76.0	116.0







#10
Phenol
Concen: 9.755 ng
RT: 6.528 min Scan# 737
Delta R.T. -0.012 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion: 94 Resp: 157877

Ion	Ratio	Lower	Upper
94	100		
65	28.9	8.9	48.9
66	37.6	18.7	58.7







#11
bis(2-Chloroethyl)ether
Concen: 9.909 ng
RT: 6.634 min Scan# 7
Delta R.T. -0.012 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 93 Resp: 124281

Ion	Ratio	Lower	Upper
93	100		
63	83.6	62.2	102.2
95	32.8	12.7	52.7







#12
1,3-Dichlorobenzene
Concen: 10.141 ng
RT: 6.851 min Scan# 792
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:146 Resp: 147587

Ion	Ratio	Lower	Upper
146	100		
148	63.5	50.5	75.7
75	28.0	23.2	34.8







#13
1,4-Dichlorobenzene
Concen: 10.075 ng
RT: 6.928 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:146 Resp: 147728

Ion	Ratio	Lower	Upper
146	100		
148	62.2	50.6	76.0
111	41.4	33.4	50.0







#14
1,2-Dichlorobenzene
Concen: 10.073 ng
RT: 7.081 min Scan# 831
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:146 Resp: 140999

Ion	Ratio	Lower	Upper
146	100		
148	63.7	50.5	75.7
111	42.8	34.9	52.3







#15
Benzyl Alcohol
Concen: 9.481 ng
RT: 7.039 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 79 Resp: 101094

Ion	Ratio	Lower	Upper
79	100		
108	80.3	64.5	96.7
77	63.9	49.2	73.8







#16
2,2'-oxybis(1-Chloropropane)
Concen: 10.038 ng
RT: 7.181 min Scan# 848
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion: 45 Resp: 227340

Ion	Ratio	Lower	Upper
45	100		
77	10.8	0.0	31.3
79	8.2	0.0	28.5







#17
2-Methylphenol
Concen: 9.707 ng
RT: 7.145 min Scan# 84
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:107 Resp: 102299

Ion	Ratio	Lower	Upper
107	100		
108	113.6	91.0	136.4
77	42.5	34.4	51.6
79	42.0	34.4	51.6







#18
Hexachloroethane
Concen: 9.662 ng
RT: 7.422 min Scan# 889
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:117 Resp: 46369

Ion	Ratio	Lower	Upper
117	100		
119	96.1	77.1	115.7
201	88.1	79.4	119.2







#19
n-Nitroso-di-n-propylamine
Concen: 10.029 ng
RT: 7.310 min Scan# 8
Delta R.T. -0.012 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 70 Resp: 93168

Ion	Ratio	Lower	Upper
70	100		
42	68.0	53.4	80.2
101	9.5	7.7	11.5
130	21.4	17.2	25.8







#20
3+4-Methylphenols
Concen: 9.981 ng
RT: 7.298 min Scan# 868
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:107 Resp: 131546

Ion	Ratio	Lower	Upper
107	100		
108	91.7	71.2	111.2
77	68.1	41.1	81.1
79	27.8	8.1	48.1







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.192 min Scan# 1
Delta R.T. -0.001 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:136 Resp: 764114

Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.7	13.1
54	9.1	7.7	11.5
68	6.1	5.0	7.6







#22
Acetophenone
Concen: 10.423 ng
RT: 7.310 min Scan# 870
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:105 Resp: 184253

Ion	Ratio	Lower	Upper
105	100		
71	8.1	5.7	8.5
51	34.2	30.1	45.1
120	22.7	19.0	28.4







#23
Nitrobenzene-d5
Concen: 16.523 ng
RT: 7.463 min Scan# 8
Delta R.T. -0.012 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 82 Resp: 192172

Ion	Ratio	Lower	Upper
82	100		
128	41.8	37.0	55.6
54	60.8	48.0	72.0







#24
Nitrobenzene
Concen: 8.754 ng
RT: 7.481 min Scan# 899
Delta R.T. -0.012 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion: 77 Resp: 99900

Ion	Ratio	Lower	Upper
77	100		
123	43.7	38.6	58.0
65	14.3	11.5	17.3







#25
Isophorone
Concen: 9.859 ng
RT: 7.722 min Scan# 94
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 82 Resp: 240659

Ion	Ratio	Lower	Upper
82	100		
95	6.9	5.6	8.4
138	18.8	15.3	22.9







#26
2-Nitrophenol
Concen: 9.530 ng
RT: 7.804 min Scan# 954
Delta R.T. 0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:139 Resp: 21744

Ion	Ratio	Lower	Upper
139	100		
109	26.6	21.8	32.8
65	53.0	41.1	61.7







#27
2,4-Dimethylphenol
Concen: 9.889 ng
RT: 7.834 min Scan# 9
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:122 Resp: 121707

Ion	Ratio	Lower	Upper
122	100		
107	102.8	86.0	129.0
121	57.1	46.3	69.5







#28
bis(2-Chloroethoxy)methane
Concen: 10.026 ng
RT: 7.934 min Scan# 976
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion: 93 Resp: 154349

Ion	Ratio	Lower	Upper
93	100		
95	32.9	26.2	39.2
123	11.2	9.7	14.5







#29
2,4-Dichlorophenol
Concen: 9.404 ng
RT: 8.034 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:162 Resp: 96761

Ion	Ratio	Lower	Upper
162	100		
164	63.7	43.0	83.0
98	38.4	16.7	56.7







#30
1,2,4-Trichlorobenzene
Concen: 10.087 ng
RT: 8.128 min Scan# 1009
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:180 Resp: 115815

Ion	Ratio	Lower	Upper
180	100		
182	93.3	75.7	113.5
145	31.6	25.4	38.0







#31
Naphthalene
Concen: 10.422 ng
RT: 8.210 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:128 Resp: 397159

Ion	Ratio	Lower	Upper
128	100		
129	11.0	9.0	13.4
127	12.8	10.6	15.8







#32
Benzoic acid
Concen: 13.045 ng
RT: 7.881 min Scan# 967
Delta R.T. -0.053 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:122 Resp: 18272

Ion	Ratio	Lower	Upper
122	100		
105	120.0	98.2	138.2
77	92.6	66.8	106.8







#33
4-Chloroaniline
Concen: 9.957 ng
RT: 8.257 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:127 Resp: 157999

Ion	Ratio	Lower	Upper
127	100		
129	31.9	26.1	39.1
65	30.9	25.5	38.3
92	18.9	15.4	23.2







#34
Hexachlorobutadiene
Concen: 9.976 ng
RT: 8.334 min Scan# 1044
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:225 Resp: 66829

Ion	Ratio	Lower	Upper
225	100		
223	63.9	49.7	74.5
227	62.6	51.0	76.4







#35
Caprolactam
Concen: 8.989 ng
RT: 8.586 min Scan# 1
Delta R.T. -0.035 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:113 Resp: 27800

Ion	Ratio	Lower	Upper
113	100		
55	194.0	177.7	217.7
56	144.7	127.1	167.1







#36
4-Chloro-3-methylphenol
Concen: 9.576 ng
RT: 8.722 min Scan# 1110
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:107 Resp: 101179

Ion	Ratio	Lower	Upper
107	100		
144	27.3	21.0	31.4
142	82.8	66.1	99.1







#37
2-Methylnaphthalene
Concen: 10.390 ng
RT: 8.904 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:142 Resp: 243429

Ion	Ratio	Lower	Upper
142	100		
141	86.8	70.5	105.7







#38
1-Methylnaphthalene
Concen: 10.312 ng
RT: 9.004 min Scan# 1158
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:142 Resp: 250299

Ion	Ratio	Lower	Upper
142	100		
141	92.1	72.8	109.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.945 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:164 Resp: 402628

Ion	Ratio	Lower	Upper
164	100		
162	99.1	78.6	117.8
160	43.8	35.7	53.5







#40
1,2,4,5-Tetrachlorobenzene
Concen: 10.209 ng
RT: 9.069 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:216 Resp: 115685

Ion	Ratio	Lower	Upper
216	100		
214	77.5	63.8	95.6
179	21.0	16.7	25.1
108	18.9	15.6	23.4







#41
Hexachlorocyclopentadiene
Concen: 8.298 ng
RT: 9.057 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:237 Resp: 53365

Ion	Ratio	Lower	Upper
237	100		
235	62.1	43.5	83.5
272	12.8	0.0	32.9







#42
2,4,6-Tribromophenol
Concen: 17.292 ng
RT: 10.727 min Scan# 1451
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:330 Resp: 63470

Ion	Ratio	Lower	Upper
330	100		
332	95.0	76.5	114.7
141	32.9	24.3	36.5







#43
2,4,6-Trichlorophenol
Concen: 9.438 ng
RT: 9.175 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:196 Resp: 66271

Ion	Ratio	Lower	Upper
196	100		
198	92.9	77.4	116.2
200	30.2	24.6	37.0







#44
2,4,5-Trichlorophenol
Concen: 9.009 ng
RT: 9.210 min Scan# 1193
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:196 Resp: 66509

Ion	Ratio	Lower	Upper
196	100		
198	95.6	75.2	112.8
97	47.5	39.8	59.8
132	29.2	23.8	35.8







#45
2-Fluorobiphenyl
Concen: 22.275 ng
RT: 9.263 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:172 Resp: 636078

Ion	Ratio	Lower	Upper
172	100		
171	35.4	29.0	43.4
170	22.6	19.4	29.0







#46
1,1'-Biphenyl
Concen: 10.423 ng
RT: 9.363 min Scan# 1219
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:154 Resp: 326241

Ion	Ratio	Lower	Upper
154	100		
153	39.5	20.5	60.5
76	12.7	0.0	32.8







#47
2-Chloronaphthalene
Concen: 10.165 ng
RT: 9.392 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:162 Resp: 236725

Ion	Ratio	Lower	Upper
162	100		
127	40.0	32.6	49.0
164	33.3	26.1	39.1







#48
2-Nitroaniline
Concen: 9.382 ng
RT: 9.480 min Scan# 1239
Delta R.T. -0.007 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion: 65 Resp: 31455

Ion	Ratio	Lower	Upper
65	100		
92	68.4	54.3	81.5
138	108.0	88.6	133.0







#49
Acenaphthylene
Concen: 10.373 ng
RT: 9.804 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:152 Resp: 407630

Ion	Ratio	Lower	Upper
152	100		
151	19.3	16.2	24.2
153	13.1	10.7	16.1







#50
Dimethylphthalate
Concen: 10.036 ng
RT: 9.657 min Scan# 1269
Delta R.T. -0.012 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:163 Resp: 258512

Ion	Ratio	Lower	Upper
163	100		
194	3.8	3.2	4.8
164	10.0	8.2	12.2







#51
2,6-Dinitrotoluene
Concen: 9.147 ng
RT: 9.722 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:165 Resp: 27624

Ion	Ratio	Lower	Upper
165	100		
63	66.0	48.5	72.7
89	63.2	43.9	65.9







#52
Acenaphthene
Concen: 10.305 ng
RT: 9.980 min Scan# 1324
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:154 Resp: 236444

Ion	Ratio	Lower	Upper
154	100		
153	110.2	88.6	133.0
152	52.1	42.1	63.1







#53
3-Nitroaniline
Concen: 9.049 ng
RT: 9.886 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:138 Resp: 37727

Ion	Ratio	Lower	Upper
138	100		
108	25.3	11.7	17.5#
92	113.9	91.2	136.8







#54
2,4-Dinitrophenol
Concen: 11.362 ng
RT: 9.986 min Scan# 1325
Delta R.T. -0.012 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:184 Resp: 4993

Ion	Ratio	Lower	Upper
184	100		
63	302.8	69.2	103.8#
154	3030.9	271.4	407.2#







#55
Dibenzofuran
Concen: 10.392 ng
RT: 10.151 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:168 Resp: 357431

Ion	Ratio	Lower	Upper
168	100		
139	35.7	29.4	44.0
169	13.3	11.0	16.4







#56
4-Nitrophenol
Concen: 6.520 ng
RT: 10.028 min Scan# 1332
Delta R.T. -0.011 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:139 Resp: 26399

Ion	Ratio	Lower	Upper
139	100		
109	77.2	44.9	84.9
65	99.0	74.7	114.7







#57
2,4-Dinitrotoluene
Concen: 9.405 ng
RT: 10.122 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:165 Resp: 29245

Ion	Ratio	Lower	Upper
165	100		
63	51.5	38.8	58.2
89	76.8	57.1	85.7
182	2.7	2.2	3.2







#58
Fluorene
Concen: 10.800 ng
RT: 10.492 min Scan# 1411
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:166 Resp: 282449

Ion	Ratio	Lower	Upper
166	100		
165	93.3	73.9	110.9
167	13.4	10.8	16.2







#59
2,3,4,6-Tetrachlorophenol
Concen: 8.775 ng
RT: 10.263 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:232 Resp: 53660

Ion	Ratio	Lower	Upper
232	100		
131	44.3	35.9	53.9
130	2.4	1.8	2.6
166	30.8	24.0	36.0







#60
Diethylphthalate
Concen: 10.003 ng
RT: 10.363 min Scan# 1389
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:149 Resp: 254431

Ion	Ratio	Lower	Upper
149	100		
177	21.2	17.6	26.4
150	12.3	10.2	15.2







#61
4-Chlorophenyl-phenylether
Concen: 10.398 ng
RT: 10.486 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:204 Resp: 129149

Ion	Ratio	Lower	Upper
204	100		
206	31.9	25.6	38.4
141	60.4	48.9	73.3







#62
4-Nitroaniline
Concen: 8.911 ng
RT: 10.492 min Scan# 1411
Delta R.T. -0.012 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:138 Resp: 35703

Ion	Ratio	Lower	Upper
138	100		
92	52.3	29.2	69.2
108	82.7	49.1	89.1







#63
Azobenzene
Concen: 10.102 ng
RT: 10.645 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 77 Resp: 253731

Ion	Ratio	Lower	Upper
77	100		
182	25.8	7.3	47.3
105	22.6	2.5	42.5
51	34.7	14.4	54.4







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:188 Resp: 692116

Ion	Ratio	Lower	Upper
188	100		
94	10.1	8.2	12.2
80	10.4	8.6	12.8







#65
4,6-Dinitro-2-methylphenol
Concen: 13.263 ng
RT: 10.527 min Scan# 14
Delta R.T. -0.007 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:198 Resp: 7762

Ion	Ratio	Lower	Upper
198	100		
51	65.4	40.6	80.6
105	52.1	29.2	69.2







#66
n-Nitrosodiphenylamine
Concen: 10.042 ng
RT: 10.598 min Scan# 1429
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:169 Resp: 237471

Ion	Ratio	Lower	Upper
169	100		
168	66.5	53.0	79.6
167	35.9	28.8	43.2







#67
4-Bromophenyl-phenylether
Concen: 9.820 ng
RT: 10.975 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:248 Resp: 75692

Ion	Ratio	Lower	Upper
248	100		
250	96.8	77.2	115.8
141	79.1	58.4	87.6







#68
Hexachlorobenzene
Concen: 9.832 ng
RT: 11.039 min Scan# 1504
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:284 Resp: 84881

Ion	Ratio	Lower	Upper
284	100		
142	33.6	24.6	37.0
249	30.2	23.2	34.8







#69
Atrazine
Concen: 9.554 ng
RT: 11.122 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:200 Resp: 59553

Ion	Ratio	Lower	Upper
200	100		
173	26.8	8.6	48.6
215	47.6	28.5	68.5







#70
Pentachlorophenol
Concen: 7.029 ng
RT: 11.227 min Scan# 1536
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:266 Resp: 32058

Ion	Ratio	Lower	Upper
266	100		
268	63.5	50.0	75.0
264	61.4	49.3	73.9







#71
Phenanthrene
Concen: 10.316 ng
RT: 11.457 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:178 Resp: 385677

Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.5	23.3
179	15.2	12.6	19.0







#72
Anthracene
Concen: 10.509 ng
RT: 11.510 min Scan# 1584
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:178 Resp: 401700

Ion	Ratio	Lower	Upper
178	100		
176	18.2	15.3	22.9
179	15.1	12.7	19.1







#73
Carbazole
Concen: 10.389 ng
RT: 11.657 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:167 Resp: 362249

Ion	Ratio	Lower	Upper
167	100		
166	20.7	17.3	25.9
139	12.5	10.3	15.5







#74
Di-n-butylphthalate
Concen: 9.927 ng
RT: 11.998 min Scan# 1667
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:149 Resp: 348111

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.5	11.3
104	4.4	3.6	5.4







#75
Fluoranthene
Concen: 10.709 ng
RT: 12.645 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:202 Resp: 395757

Ion	Ratio	Lower	Upper
202	100		
101	11.6	0.0	32.4
203	17.3	0.0	37.5







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.074 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:240 Resp: 449534

Ion	Ratio	Lower	Upper
240	100		
120	12.0	10.1	15.1
236	25.6	20.5	30.7







#77
Benzidine
Concen: 8.797 ng
RT: 12.769 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:184 Resp: 168038

Ion	Ratio	Lower	Upper
184	100		
185	14.3	11.8	17.6
183	10.4	8.9	13.3







#78
Pyrene
Concen: 9.763 ng
RT: 12.874 min Scan# 1816
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:202 Resp: 404860

Ion	Ratio	Lower	Upper
202	100		
200	19.8	16.1	24.1
203	17.6	14.1	21.1







#79
Terphenyl-d14
Concen: 20.552 ng
RT: 13.016 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:244 Resp: 635233

Ion	Ratio	Lower	Upper
244	100		
212	6.5	5.3	7.9
122	12.3	9.1	13.7







#80
Butylbenzylphthalate
Concen: 9.084 ng
RT: 13.498 min Scan# 1922
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:149 Resp: 65964

Ion	Ratio	Lower	Upper
149	100		
91	66.6	54.4	81.6
206	20.5	15.9	23.9







#81
Benzo(a)anthracene
Concen: 9.832 ng
RT: 14.063 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:228 Resp: 296710

Ion	Ratio	Lower	Upper
228	100		
226	27.0	21.9	32.9
229	19.7	16.0	24.0







#82
3,3'-Dichlorobenzidine
Concen: 8.046 ng
RT: 14.027 min Scan# 2012
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:252 Resp: 71321

Ion	Ratio	Lower	Upper
252	100		
254	65.6	51.0	76.6
126	12.5	9.4	14.2







#83
Chrysene
Concen: 9.897 ng
RT: 14.098 min Scan# 2(
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:228 Resp: 271875

Ion	Ratio	Lower	Upper
228	100		
226	30.3	24.0	36.0
229	19.8	15.7	23.5







#84
Bis(2-ethylhexyl)phthalate
Concen: 9.409 ng
RT: 14.063 min Scan# 2018
Delta R.T. 0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:149 Resp: 100777

Ion	Ratio	Lower	Upper
149	100		
167	26.3	21.7	32.5
279	4.0	3.0	4.4







#85
Di-n-octyl phthalate
Concen: 10.622 ng
RT: 14.680 min Scan# 21
Delta R.T. 0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:149 Resp: 150072

Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.1	1.7
43	12.1	9.3	13.9







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.568 min Scan# 2274
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:264 Resp: 332067

Ion	Ratio	Lower	Upper
264	100		
260	23.4	18.4	27.6
265	21.0	17.3	25.9







#87
Indeno(1,2,3-cd)pyrene
Concen: 9.449 ng
RT: 17.068 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:276 Resp: 229818

Ion	Ratio	Lower	Upper
276	100		
138	28.5	22.6	33.8
277	25.1	20.1	30.1







#88
Benzo(b)fluoranthene
Concen: 9.823 ng
RT: 15.121 min Scan# 2198
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:252 Resp: 202615

Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.7	26.5
125	11.1	9.0	13.4







#89
Benzo(k)fluoranthene
Concen: 10.203 ng
RT: 15.157 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:252 Resp: 192500

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.8	26.8
125	10.4	9.0	13.6







#90
Benzo(a)pyrene
Concen: 9.651 ng
RT: 15.498 min Scan# 2262
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:252 Resp: 182258

Ion	Ratio	Lower	Upper
252	100		
253	21.3	17.6	26.4
125	12.5	9.5	14.3







#91
Dibenzo(a,h)anthracene
Concen: 9.513 ng
RT: 17.092 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:278 Resp: 187788

Ion	Ratio	Lower	Upper
278	100		
139	19.3	15.2	22.8
279	24.0	19.0	28.6






#92
Benzo(g,h,i)perylene
Concen: 9.518 ng
RT: 17.521 min Scan# 2606
Delta R.T. -0.018 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:276 Resp: 188891

Ion	Ratio	Lower	Upper
276	100		
277	23.9	18.7	28.1
138	24.4	19.4	29.0



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142242.D
Acq On : 30 Apr 2025 12:49
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025

Quant Time: Apr 30 15:42:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	194641	20.000	ng	0.00
21) Naphthalene-d8	8.193	136	746567	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	398805	20.000	ng	0.00
64) Phenanthrene-d10	11.434	188	679379	20.000	ng	0.00
76) Chrysene-d12	14.075	240	430351	20.000	ng	0.00
86) Perylene-d12	15.569	264	350840	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.510	112	492356	40.399	ng	0.00
7) Phenol-d6	6.516	99	597221	40.580	ng	-0.01
23) Nitrobenzene-d5	7.469	82	450837	39.673	ng	0.00
42) 2,4,6-Tribromophenol	10.734	330	146779	40.372	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	1212743	42.876	ng	0.00
79) Terphenyl-d14	13.022	244	1223735	41.357	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.693	88	109201	20.062	ng	100
3) Pyridine	3.458	79	282987	20.004	ng	100
4) n-Nitrosodimethylamine	3.399	42	141737	19.361	ng	97
6) Aniline	6.569	93	419797	20.152	ng	99
8) 2-Chlorophenol	6.687	128	253840	20.156	ng	99
9) Benzaldehyde	6.457	77	213052	20.712	ng	98
10) Phenol	6.534	94	322435	20.352	ng	98
11) bis(2-Chloroethyl)ether	6.640	93	247179	20.132	ng	99
12) 1,3-Dichlorobenzene	6.851	146	289129	20.294	ng	100
13) 1,4-Dichlorobenzene	6.928	146	290927	20.268	ng	99
14) 1,2-Dichlorobenzene	7.081	146	278592	20.331	ng	99
15) Benzyl Alcohol	7.040	79	207042	19.835	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.181	45	448250	20.218	ng	99
17) 2-Methylphenol	7.146	107	207957	20.157	ng	99
18) Hexachloroethane	7.422	117	93572	19.917	ng	96
19) n-Nitroso-di-n-propyla...	7.316	70	184776	20.318	ng	99
20) 3+4-Methylphenols	7.298	107	267057	20.700	ng	96
22) Acetophenone	7.310	105	360484	20.872	ng	98
24) Nitrobenzene	7.487	77	221499	19.865	ng	98
25) Isophorone	7.722	82	479096	20.088	ng	99
26) 2-Nitrophenol	7.804	139	61615	17.893	ng	98
27) 2,4-Dimethylphenol	7.834	122	248864	20.696	ng	97
28) bis(2-Chloroethoxy)met...	7.934	93	309073	20.549	ng	99
29) 2,4-Dichlorophenol	8.040	162	204821	20.374	ng	99
30) 1,2,4-Trichlorobenzene	8.128	180	227898	20.315	ng	100
31) Naphthalene	8.210	128	772839	20.758	ng	100
32) Benzoic acid	7.904	122	64656m	19.412	ng	
33) 4-Chloroaniline	8.257	127	320218	20.654	ng	99
34) Hexachlorobutadiene	8.334	225	133474	20.393	ng	99
35) Caprolactam	8.604	113	60791	20.118	ng	99
36) 4-Chloro-3-methylphenol	8.722	107	208419	20.190	ng	99
37) 2-Methylnaphthalene	8.904	142	472896	20.658	ng	100
38) 1-Methylnaphthalene	9.004	142	497635	20.985	ng	99
40) 1,2,4,5-Tetrachloroben...	9.069	216	228311	20.342	ng	99
41) Hexachlorocyclopentadiene	9.057	237	121953	19.145	ng	99
43) 2,4,6-Trichlorophenol	9.175	196	140807	20.246	ng	100

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142242.D
Acq On : 30 Apr 2025 12:49
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025

Quant Time: Apr 30 15:42:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.210	196	144462	19.756	ng		99
46) 1,1'-Biphenyl	9.369	154	643017	20.740	ng		99
47) 2-Chloronaphthalene	9.392	162	474899	20.588	ng		98
48) 2-Nitroaniline	9.481	65	90798	18.213	ng		97
49) Acenaphthylene	9.810	152	807654	20.749	ng		99
50) Dimethylphthalate	9.663	163	518802	20.334	ng		100
51) 2,6-Dinitrotoluene	9.722	165	76955	18.432	ng		93
52) Acenaphthene	9.981	154	460927	20.281	ng		100
53) 3-Nitroaniline	9.892	138	100560	18.959	ng	#	95
54) 2,4-Dinitrophenol	9.992	184	14553	18.804	ng	#	1
55) Dibenzofuran	10.151	168	698155	20.493	ng		100
56) 4-Nitrophenol	10.034	139	70423	17.561	ng		97
57) 2,4-Dinitrotoluene	10.122	165	84812	17.844	ng		95
58) Fluorene	10.492	166	538647	20.793	ng		99
59) 2,3,4,6-Tetrachlorophenol	10.263	232	124353	20.530	ng		99
60) Diethylphthalate	10.363	149	519294	20.611	ng		98
61) 4-Chlorophenyl-phenyle...	10.486	204	252213	20.502	ng		99
62) 4-Nitroaniline	10.498	138	92884	18.933	ng		97
63) Azobenzene	10.645	77	508482	20.438	ng		98
65) 4,6-Dinitro-2-methylph...	10.528	198	23400	18.756	ng		98
66) n-Nitrosodiphenylamine	10.604	169	470767	20.280	ng		100
67) 4-Bromophenyl-phenylether	10.981	248	151720	20.053	ng		99
68) Hexachlorobenzene	11.039	284	168767	19.916	ng		97
69) Atrazine	11.128	200	126444	20.665	ng		99
70) Pentachlorophenol	11.228	266	81929	18.301	ng		99
71) Phenanthrene	11.457	178	760476	20.723	ng		99
72) Anthracene	11.510	178	782215	20.848	ng		99
73) Carbazole	11.657	167	720775	21.059	ng		99
74) Di-n-butylphthalate	11.998	149	739011	21.469	ng		99
75) Fluoranthene	12.645	202	784880	21.636	ng		100
77) Benzidine	12.769	184	381034	20.837	ng		99
78) Pyrene	12.875	202	795806	20.046	ng		100
80) Butylbenzylphthalate	13.498	149	174268	18.529	ng		99
81) Benzo(a)anthracene	14.063	228	573438	19.849	ng		99
82) 3,3'-Dichlorobenzidine	14.027	252	163359	19.251	ng		100
83) Chrysene	14.098	228	537388	20.435	ng		100
84) Bis(2-ethylhexyl)phtha...	14.057	149	236325	18.042	ng		100
85) Di-n-octyl phthalate	14.674	149	376362	17.998	ng		98
87) Indeno(1,2,3-cd)pyrene	17.074	276	514496	20.021	ng		99
88) Benzo(b)fluoranthene	15.127	252	423694	19.442	ng		98
89) Benzo(k)fluoranthene	15.157	252	428646	21.504	ng		98
90) Benzo(a)pyrene	15.504	252	402101	20.154	ng		99
91) Dibenzo(a,h)anthracene	17.092	278	420029	20.139	ng		98
92) Benzo(g,h,i)perylene	17.527	276	422497	20.150	ng		100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142242.D
Acq On : 30 Apr 2025 12:49
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1


Quant Time: Apr 30 15:42:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:152 Resp: 19464


Ion	Ratio	Lower	Upper
152	100		
150	157.2	130.2	195.2
115	57.3	47.0	70.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#2
1,4-Dioxane
Concen: 20.062 ng
RT: 2.693 min Scan# 85
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion: 88 Resp: 109201

Ion	Ratio	Lower	Upper
88	100		
58	90.0	72.2	108.2
43	34.2	27.0	40.6






#3
Pyridine
Concen: 20.004 ng
RT: 3.458 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion: 79 Resp: 28298


Ion	Ratio	Lower	Upper
79	100		
52	89.0	71.5	107.3
51	43.1	34.5	51.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#4
n-Nitrosodimethylamine
Concen: 19.361 ng
RT: 3.399 min Scan# 205
Delta R.T. -0.012 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion: 42 Resp: 141737

Ion	Ratio	Lower	Upper
42	100		
74	114.9	89.5	134.3
44	9.6	6.6	9.8





#5
2-Fluorophenol
Concen: 40.399 ng
RT: 5.510 min Scan# 5
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49


Tgt Ion:112 Resp: 49235


Ion	Ratio	Lower	Upper
112	100		
64	60.7	48.4	72.6
63	31.7	24.8	37.2

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#6
Aniline
Concen: 20.152 ng
RT: 6.569 min Scan# 744
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion: 93 Resp: 419797

Ion	Ratio	Lower	Upper
93	100		
66	39.3	32.1	48.1
65	20.1	16.6	24.8






#7
Phenol-d6
Concen: 40.580 ng
RT: 6.516 min Scan# 7
Delta R.T. -0.012 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion: 99 Resp: 59722


Ion	Ratio	Lower	Upper
99	100		
42	21.3	16.3	24.5
71	31.6	25.0	37.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#8
2-Chlorophenol
Concen: 20.156 ng
RT: 6.687 min Scan# 764
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:128 Resp: 253840

Ion	Ratio	Lower	Upper
128	100		
130	32.8	13.1	53.1
64	50.0	28.6	68.6






#9
Benzaldehyde
Concen: 20.712 ng
RT: 6.457 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion: 77 Resp: 21305


Ion	Ratio	Lower	Upper
77	100		
105	101.1	79.1	119.1
106	97.4	76.0	116.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#10
Phenol
Concen: 20.352 ng
RT: 6.534 min Scan# 738
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion: 94 Resp: 322435

Ion	Ratio	Lower	Upper
94	100		
65	28.4	8.9	48.9
66	37.1	18.7	58.7






#11
bis(2-Chloroethyl)ether
Concen: 20.132 ng
RT: 6.640 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion: 93 Resp: 24717


Ion	Ratio	Lower	Upper
93	100		
63	83.3	62.2	102.2
95	32.3	12.7	52.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#12
1,3-Dichlorobenzene
Concen: 20.294 ng
RT: 6.851 min Scan# 792
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:146 Resp: 289129

Ion	Ratio	Lower	Upper
146	100		
148	63.2	50.5	75.7
75	28.6	23.2	34.8






#13
1,4-Dichlorobenzene
Concen: 20.268 ng
RT: 6.928 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:146 Resp: 29092


Ion	Ratio	Lower	Upper
146	100		
148	64.1	50.6	76.0
111	41.6	33.4	50.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#14
1,2-Dichlorobenzene
Concen: 20.331 ng
RT: 7.081 min Scan# 831
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:146 Resp: 278592

Ion	Ratio	Lower	Upper
146	100		
148	63.6	50.5	75.7
111	42.7	34.9	52.3





#15
Benzyl Alcohol
Concen: 19.835 ng
RT: 7.040 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49


Tgt Ion: 79 Resp: 20704


Ion	Ratio	Lower	Upper
79	100		
108	81.7	64.5	96.7
77	63.5	49.2	73.8

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#16
2,2'-oxybis(1-Chloropropane)
Concen: 20.218 ng
RT: 7.181 min Scan# 848
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion: 45 Resp: 448250

Ion	Ratio	Lower	Upper
45	100		
77	10.9	0.0	31.3
79	8.4	0.0	28.5






#17
2-Methylphenol
Concen: 20.157 ng
RT: 7.146 min Scan# 84
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:107 Resp: 20795


Ion	Ratio	Lower	Upper
107	100		
108	114.6	91.0	136.4
77	44.2	34.4	51.6
79	42.9	34.4	51.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#18
Hexachloroethane
Concen: 19.917 ng
RT: 7.422 min Scan# 889
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:117 Resp: 93572

Ion	Ratio	Lower	Upper
117	100		
119	97.2	77.1	115.7
201	92.2	79.4	119.2






#19
n-Nitroso-di-n-propylamine
Concen: 20.318 ng
RT: 7.316 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion: 70 Resp: 18477


Ion	Ratio	Lower	Upper
70	100		
42	67.2	53.4	80.2
101	9.2	7.7	11.5
130	21.3	17.2	25.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#20
3+4-Methylphenols
Concen: 20.700 ng
RT: 7.298 min Scan# 868
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:107 Resp: 267057

Ion	Ratio	Lower	Upper
107	100		
108	89.7	71.2	111.2
77	54.6	41.1	81.1
79	26.8	8.1	48.1






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.193 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:136 Resp: 74656


Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.7	13.1
54	9.1	7.7	11.5
68	6.1	5.0	7.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#22
Acetophenone
Concen: 20.872 ng
RT: 7.310 min Scan# 870
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:105 Resp: 360484

Ion	Ratio	Lower	Upper
105	100		
71	8.3	5.7	8.5
51	36.4	30.1	45.1
120	23.4	19.0	28.4






#23
Nitrobenzene-d5
Concen: 39.673 ng
RT: 7.469 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion: 82 Resp: 45083


Ion	Ratio	Lower	Upper
82	100		
128	44.3	37.0	55.6
54	61.1	48.0	72.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#24
Nitrobenzene
Concen: 19.865 ng
RT: 7.487 min Scan# 900
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion: 77 Resp: 221499

Ion	Ratio	Lower	Upper
77	100		
123	46.9	38.6	58.0
65	14.3	11.5	17.3





#25
Isophorone
Concen: 20.088 ng
RT: 7.722 min Scan# 94
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49


Tgt Ion: 82 Resp: 47909


Ion	Ratio	Lower	Upper
82	100		
95	6.9	5.6	8.4
138	19.4	15.3	22.9

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#26
2-Nitrophenol
Concen: 17.893 ng
RT: 7.804 min Scan# 954
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:139 Resp: 61615

Ion	Ratio	Lower	Upper
139	100		
109	26.4	21.8	32.8
65	49.6	41.1	61.7






#27
2,4-Dimethylphenol
Concen: 20.696 ng
RT: 7.834 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:122 Resp: 24886


Ion	Ratio	Lower	Upper
122	100		
107	104.1	86.0	129.0
121	56.5	46.3	69.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#28
bis(2-Chloroethoxy)methane
Concen: 20.549 ng
RT: 7.934 min Scan# 976
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion: 93 Resp: 309073

Ion	Ratio	Lower	Upper
93	100		
95	32.6	26.2	39.2
123	11.5	9.7	14.5






#29
2,4-Dichlorophenol
Concen: 20.374 ng
RT: 8.040 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:162 Resp: 20482


Ion	Ratio	Lower	Upper
162	100		
164	64.3	43.0	83.0
98	36.1	16.7	56.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#30
1,2,4-Trichlorobenzene
Concen: 20.315 ng
RT: 8.128 min Scan# 1009
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:180 Resp: 227898

Ion	Ratio	Lower	Upper
180	100		
182	94.7	75.7	113.5
145	31.7	25.4	38.0






#31
Naphthalene
Concen: 20.758 ng
RT: 8.210 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:128 Resp: 77283


Ion	Ratio	Lower	Upper
128	100		
129	11.2	9.0	13.4
127	13.1	10.6	15.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#32
Benzoic acid
Concen: 19.412 ng m
RT: 7.904 min Scan# 971
Delta R.T. -0.030 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:122 Resp: 64656

Ion	Ratio	Lower	Upper
122	100		
105	117.1	98.2	138.2
77	84.0	66.8	106.8






#33
4-Chloroaniline
Concen: 20.654 ng
RT: 8.257 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:127 Resp: 32021


Ion	Ratio	Lower	Upper
127	100		
129	32.2	26.1	39.1
65	30.7	25.5	38.3
92	18.9	15.4	23.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#34
Hexachlorobutadiene
Concen: 20.393 ng
RT: 8.334 min Scan# 1044
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:225 Resp: 133474

Ion	Ratio	Lower	Upper
225	100		
223	62.8	49.7	74.5
227	63.4	51.0	76.4






#35
Caprolactam
Concen: 20.118 ng
RT: 8.604 min Scan# 1
Delta R.T. -0.018 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:113 Resp: 6079


Ion	Ratio	Lower	Upper
113	100		
55	198.2	177.7	217.7
56	144.8	127.1	167.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#36
4-Chloro-3-methylphenol
Concen: 20.190 ng
RT: 8.722 min Scan# 1110
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:107 Resp: 208419

Ion	Ratio	Lower	Upper
107	100		
144	26.4	21.0	31.4
142	83.4	66.1	99.1






#37
2-Methylnaphthalene
Concen: 20.658 ng
RT: 8.904 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:142 Resp: 47289


Ion	Ratio	Lower	Upper
142	100		
141	87.8	70.5	105.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#38
1-Methylnaphthalene
Concen: 20.985 ng
RT: 9.004 min Scan# 1158
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:142 Resp: 497635

Ion	Ratio	Lower	Upper
142	100		
141	90.1	72.8	109.2






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.945 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:164 Resp: 39880


Ion	Ratio	Lower	Upper
164	100		
162	98.3	78.6	117.8
160	44.1	35.7	53.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 20.342 ng
RT: 9.069 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:216 Resp: 228311

Ion	Ratio	Lower	Upper
216	100		
214	78.9	63.8	95.6
179	20.5	16.7	25.1
108	19.4	15.6	23.4






#41
Hexachlorocyclopentadiene
Concen: 19.145 ng
RT: 9.057 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:237 Resp: 12195


Ion	Ratio	Lower	Upper
237	100		
235	62.4	43.5	83.5
272	12.8	0.0	32.9



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#42
2,4,6-Tribromophenol
Concen: 40.372 ng
RT: 10.734 min Scan# 1452
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:330 Resp: 146779

Ion	Ratio	Lower	Upper
330	100		
332	96.7	76.5	114.7
141	32.8	24.3	36.5






#43
2,4,6-Trichlorophenol
Concen: 20.246 ng
RT: 9.175 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:196 Resp: 14080


Ion	Ratio	Lower	Upper
196	100		
198	96.4	77.4	116.2
200	30.7	24.6	37.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#44
2,4,5-Trichlorophenol
Concen: 19.756 ng
RT: 9.210 min Scan# 1193
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:196 Resp: 144462

Ion	Ratio	Lower	Upper
196	100		
198	96.1	75.2	112.8
97	49.7	39.8	59.8
132	29.1	23.8	35.8






#45
2-Fluorobiphenyl
Concen: 42.876 ng
RT: 9.263 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:172 Resp: 121274


Ion	Ratio	Lower	Upper
172	100		
171	34.7	29.0	43.4
170	23.3	19.4	29.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#46
1,1'-Biphenyl
Concen: 20.740 ng
RT: 9.369 min Scan# 1220
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:154 Resp: 643017

Ion	Ratio	Lower	Upper
154	100		
153	39.9	20.5	60.5
76	12.0	0.0	32.8






#47
2-Chloronaphthalene
Concen: 20.588 ng
RT: 9.392 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:162 Resp: 474899


Ion	Ratio	Lower	Upper
162	100		
127	39.1	32.6	49.0
164	32.2	26.1	39.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#48
2-Nitroaniline
Concen: 18.213 ng
RT: 9.481 min Scan# 1239
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion: 65 Resp: 90798

Ion	Ratio	Lower	Upper
65	100		
92	66.4	54.3	81.5
138	107.2	88.6	133.0






#49
Acenaphthylene
Concen: 20.749 ng
RT: 9.810 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:152 Resp: 80765


Ion	Ratio	Lower	Upper
152	100		
151	19.6	16.2	24.2
153	13.0	10.7	16.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#50
Dimethylphthalate
Concen: 20.334 ng
RT: 9.663 min Scan# 1270
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:163 Resp: 518802

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.2	4.8
164	10.0	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 18.432 ng
RT: 9.722 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49


Tgt Ion:165 Resp: 7695


Ion	Ratio	Lower	Upper
165	100		
63	66.2	48.5	72.7
89	59.6	43.9	65.9

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#52
Acenaphthene
Concen: 20.281 ng
RT: 9.981 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:154 Resp: 460927

Ion	Ratio	Lower	Upper
154	100		
153	110.6	88.6	133.0
152	52.4	42.1	63.1





#53
3-Nitroaniline
Concen: 18.959 ng
RT: 9.892 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49


Tgt Ion:138 Resp: 10056


Ion	Ratio	Lower	Upper
138	100		
108	19.4	11.7	17.5#
92	110.1	91.2	136.8

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#54
2,4-Dinitrophenol
Concen: 18.804 ng
RT: 9.992 min Scan# 1326
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:184 Resp: 14553

Ion	Ratio	Lower	Upper
184	100		
63	131.7	69.2	103.8#
154	869.6	271.4	407.2#






#55
Dibenzofuran
Concen: 20.493 ng
RT: 10.151 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:168 Resp: 69815


Ion	Ratio	Lower	Upper
168	100		
139	36.8	29.4	44.0
169	13.3	11.0	16.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#56
4-Nitrophenol
Concen: 17.561 ng
RT: 10.034 min Scan# 1333
Delta R.T. -0.005 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:139 Resp: 70423

Ion	Ratio	Lower	Upper
139	100		
109	71.2	44.9	84.9
65	94.8	74.7	114.7






#57
2,4-Dinitrotoluene
Concen: 17.844 ng
RT: 10.122 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:165 Resp: 8481


Ion	Ratio	Lower	Upper
165	100		
63	50.7	38.8	58.2
89	76.2	57.1	85.7
182	2.8	2.2	3.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#58
Fluorene
Concen: 20.793 ng
RT: 10.492 min Scan# 1411
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:166 Resp: 538647

Ion	Ratio	Lower	Upper
166	100		
165	93.4	73.9	110.9
167	13.6	10.8	16.2






#59
2,3,4,6-Tetrachlorophenol
Concen: 20.530 ng
RT: 10.263 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:232 Resp: 12435


Ion	Ratio	Lower	Upper
232	100		
131	45.0	35.9	53.9
130	2.1	1.8	2.6
166	28.8	24.0	36.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#60
Diethylphthalate
Concen: 20.611 ng
RT: 10.363 min Scan# 1389
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:149 Resp: 519294

Ion	Ratio	Lower	Upper
149	100		
177	21.3	17.6	26.4
150	12.0	10.2	15.2





#61
4-Chlorophenyl-phenylether
Concen: 20.502 ng
RT: 10.486 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49


Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion:204 Resp: 25221

Ion	Ratio	Lower	Upper
204	100		
206	33.0	25.6	38.4
141	60.1	48.9	73.3

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#62
4-Nitroaniline
Concen: 18.933 ng
RT: 10.498 min Scan# 1412
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:138 Resp: 92884

Ion	Ratio	Lower	Upper
138	100		
92	49.5	29.2	69.2
108	73.2	49.1	89.1






#63
Azobenzene
Concen: 20.438 ng
RT: 10.645 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion: 77 Resp: 50848


Ion	Ratio	Lower	Upper
77	100		
182	25.1	7.3	47.3
105	22.2	2.5	42.5
51	34.2	14.4	54.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.434 min Scan# 1571
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:188 Resp: 679379

Ion	Ratio	Lower	Upper
188	100		
94	9.8	8.2	12.2
80	10.5	8.6	12.8






#65
4,6-Dinitro-2-methylphenol
Concen: 18.756 ng
RT: 10.528 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:198 Resp: 2340


Ion	Ratio	Lower	Upper
198	100		
51	62.5	40.6	80.6
105	50.4	29.2	69.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#66
n-Nitrosodiphenylamine
Concen: 20.280 ng
RT: 10.604 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:169 Resp: 470767

Ion	Ratio	Lower	Upper
169	100		
168	66.2	53.0	79.6
167	36.4	28.8	43.2






#67
4-Bromophenyl-phenylether
Concen: 20.053 ng
RT: 10.981 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:248 Resp: 15172


Ion	Ratio	Lower	Upper
248	100		
250	98.1	77.2	115.8
141	72.5	58.4	87.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#68
Hexachlorobenzene
Concen: 19.916 ng
RT: 11.039 min Scan# 1504
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:284 Resp: 168767

Ion	Ratio	Lower	Upper
284	100		
142	33.8	24.6	37.0
249	29.3	23.2	34.8





#69
Atrazine
Concen: 20.665 ng
RT: 11.128 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49


Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion:200 Resp: 12644

Ion	Ratio	Lower	Upper
200	100		
173	27.9	8.6	48.6
215	49.1	28.5	68.5

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#70
Pentachlorophenol
Concen: 18.301 ng
RT: 11.228 min Scan# 1536
Delta R.T. -0.005 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:266 Resp: 81929

Ion	Ratio	Lower	Upper
266	100		
268	63.6	50.0	75.0
264	62.8	49.3	73.9






#71
Phenanthrene
Concen: 20.723 ng
RT: 11.457 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:178 Resp: 76047


Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.5	23.3
179	15.3	12.6	19.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#72
Anthracene
Concen: 20.848 ng
RT: 11.510 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:178 Resp: 782215

Ion	Ratio	Lower	Upper
178	100		
176	18.4	15.3	22.9
179	15.5	12.7	19.1






#73
Carbazole
Concen: 21.059 ng
RT: 11.657 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:167 Resp: 72077


Ion	Ratio	Lower	Upper
167	100		
166	21.2	17.3	25.9
139	12.7	10.3	15.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#74
Di-n-butylphthalate
Concen: 21.469 ng
RT: 11.998 min Scan# 1667
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:149 Resp: 739011

Ion	Ratio	Lower	Upper
149	100		
150	8.9	7.5	11.3
104	4.3	3.6	5.4






#75
Fluoranthene
Concen: 21.636 ng
RT: 12.645 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:202 Resp: 78488


Ion	Ratio	Lower	Upper
202	100		
101	12.1	0.0	32.4
203	17.6	0.0	37.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.075 min Scan# 2020
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:240 Resp: 430351

Ion	Ratio	Lower	Upper
240	100		
120	12.4	10.1	15.1
236	25.6	20.5	30.7






#77
Benzidine
Concen: 20.837 ng
RT: 12.769 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:184 Resp: 38103


Ion	Ratio	Lower	Upper
184	100		
185	14.2	11.8	17.6
183	11.1	8.9	13.3



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#78
Pyrene
Concen: 20.046 ng
RT: 12.875 min Scan# 1816
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:202 Resp: 795806

Ion	Ratio	Lower	Upper
202	100		
200	20.1	16.1	24.1
203	17.9	14.1	21.1






#79
Terphenyl-d14
Concen: 41.357 ng
RT: 13.022 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:244 Resp: 122373


Ion	Ratio	Lower	Upper
244	100		
212	6.3	5.3	7.9
122	11.1	9.1	13.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#80
Butylbenzylphthalate
Concen: 18.529 ng
RT: 13.498 min Scan# 1922
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:149 Resp: 174268

Ion	Ratio	Lower	Upper
149	100		
91	67.9	54.4	81.6
206	20.8	15.9	23.9






#81
Benzo(a)anthracene
Concen: 19.849 ng
RT: 14.063 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:228 Resp: 57343


Ion	Ratio	Lower	Upper
228	100		
226	26.7	21.9	32.9
229	19.7	16.0	24.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#82
3,3'-Dichlorobenzidine
Concen: 19.251 ng
RT: 14.027 min Scan# 2012
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:252 Resp: 163359

Ion	Ratio	Lower	Upper
252	100		
254	63.8	51.0	76.6
126	11.8	9.4	14.2






#83
Chrysene
Concen: 20.435 ng
RT: 14.098 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:228 Resp: 53738


Ion	Ratio	Lower	Upper
228	100		
226	30.1	24.0	36.0
229	19.6	15.7	23.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 18.042 ng
RT: 14.057 min Scan# 2017
Delta R.T. -0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:149 Resp: 236325

Ion	Ratio	Lower	Upper
149	100		
167	27.1	21.7	32.5
279	3.8	3.0	4.4





#85
Di-n-octyl phthalate
Concen: 17.998 ng
RT: 14.674 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49


Tgt Ion:149 Resp: 37636


Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.1	1.7
43	12.4	9.3	13.9

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.569 min Scan# 2274
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:264 Resp: 350840

Ion	Ratio	Lower	Upper
264	100		
260	23.9	18.4	27.6
265	21.8	17.3	25.9






#87
Indeno(1,2,3-cd)pyrene
Concen: 20.021 ng
RT: 17.074 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:276 Resp: 51449


Ion	Ratio	Lower	Upper
276	100		
138	28.9	22.6	33.8
277	25.1	20.1	30.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#88
Benzo(b)fluoranthene
Concen: 19.442 ng
RT: 15.127 min Scan# 2199
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:252 Resp: 423694

Ion	Ratio	Lower	Upper
252	100		
253	21.3	17.7	26.5
125	10.8	9.0	13.4





#89
Benzo(k)fluoranthene
Concen: 21.504 ng
RT: 15.157 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49


Tgt Ion:252 Resp: 42864


Ion	Ratio	Lower	Upper
252	100		
253	21.5	17.8	26.8
125	10.5	9.0	13.6

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#90
Benzo(a)pyrene
Concen: 20.154 ng
RT: 15.504 min Scan# 2263
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:252 Resp: 402101

Ion	Ratio	Lower	Upper
252	100		
253	21.5	17.6	26.4
125	11.7	9.5	14.3






#91
Dibenzo(a,h)anthracene
Concen: 20.139 ng
RT: 17.092 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:278 Resp: 42002


Ion	Ratio	Lower	Upper
278	100		
139	18.2	15.2	22.8
279	23.0	19.0	28.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#92
Benzo(g,h,i)perylene
Concen: 20.150 ng
RT: 17.527 min Scan# 2607
Delta R.T. -0.012 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:276 Resp: 422497

Ion	Ratio	Lower	Upper
276	100		
277	23.3	18.7	28.1
138	24.4	19.4	29.0



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142243.D
Acq On : 30 Apr 2025 13:17
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Apr 30 15:43:02 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	180007	20.000	ng	0.00
21) Naphthalene-d8	8.193	136	706244	20.000	ng	0.00
39) Acenaphthene-d10	9.951	164	370869	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	616408	20.000	ng	0.00
76) Chrysene-d12	14.074	240	331192	20.000	ng	0.00
86) Perylene-d12	15.568	264	328405	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	946150	83.944	ng	0.00
7) Phenol-d6	6.528	99	1146926	84.267	ng	0.00
23) Nitrobenzene-d5	7.475	82	961009	89.396	ng	0.00
42) 2,4,6-Tribromophenol	10.734	330	302223	89.389	ng	0.00
45) 2-Fluorobiphenyl	9.269	172	2089707	79.446	ng	0.00
79) Terphenyl-d14	13.022	244	1962470	86.181	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.699	88	213029	42.319	ng	100
3) Pyridine	3.463	79	547989	41.886	ng	100
4) n-Nitrosodimethylamine	3.410	42	290357	42.887	ng	100
6) Aniline	6.569	93	808338	41.959	ng	100
8) 2-Chlorophenol	6.693	128	496290	42.612	ng	100
9) Benzaldehyde	6.457	77	402047	42.263	ng	100
10) Phenol	6.540	94	617257	42.128	ng	100
11) bis(2-Chloroethyl)ether	6.646	93	474404	41.780	ng	100
12) 1,3-Dichlorobenzene	6.851	146	551813	41.881	ng	100
13) 1,4-Dichlorobenzene	6.928	146	556691	41.937	ng	100
14) 1,2-Dichlorobenzene	7.081	146	530895	41.892	ng	100
15) Benzyl Alcohol	7.046	79	413830	42.870	ng	100
16) 2,2'-oxybis(1-Chloropr...	7.187	45	858875	41.888	ng	100
17) 2-Methylphenol	7.151	107	403467	42.287	ng	100
18) Hexachloroethane	7.428	117	183881	42.322	ng	100
19) n-Nitroso-di-n-propyla...	7.322	70	348288	41.410	ng	100
20) 3+4-Methylphenols	7.304	107	503856	42.229	ng	100
22) Acetophenone	7.316	105	660959	40.454	ng	100
24) Nitrobenzene	7.493	77	463357	43.929	ng	100
25) Isophorone	7.728	82	931763	41.299	ng	100
26) 2-Nitrophenol	7.804	139	162563	38.737	ng	100
27) 2,4-Dimethylphenol	7.834	122	476186	41.862	ng	100
28) bis(2-Chloroethoxy)met...	7.940	93	586887	41.247	ng	100
29) 2,4-Dichlorophenol	8.040	162	410487	43.163	ng	100
30) 1,2,4-Trichlorobenzene	8.134	180	440825	41.539	ng	100
31) Naphthalene	8.216	128	1441240	40.920	ng	100
32) Benzoic acid	7.934	122	184052m	37.083	ng	
33) 4-Chloroaniline	8.257	127	606136	41.328	ng	100
34) Hexachlorobutadiene	8.334	225	260821	42.125	ng	100
35) Caprolactam	8.622	113	122598	42.889	ng	100
36) 4-Chloro-3-methylphenol	8.728	107	417126	42.715	ng	100
37) 2-Methylnaphthalene	8.904	142	880315	40.652	ng	100
38) 1-Methylnaphthalene	9.004	142	920983	41.054	ng	100
40) 1,2,4,5-Tetrachloroben...	9.069	216	427286	40.938	ng	100
41) Hexachlorocyclopentadiene	9.057	237	260979	44.056	ng	100
43) 2,4,6-Trichlorophenol	9.175	196	274555	42.450	ng	100

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142243.D
Acq On : 30 Apr 2025 13:17
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Apr 30 15:43:02 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.210	196	299967	44.113	ng	100
46) 1,1'-Biphenyl	9.369	154	1172400	40.664	ng	100
47) 2-Chloronaphthalene	9.392	162	879152	40.984	ng	100
48) 2-Nitroaniline	9.487	65	218766	39.600	ng	100
49) Acenaphthylene	9.810	152	1481543	40.929	ng	100
50) Dimethylphthalate	9.669	163	980172	41.312	ng	100
51) 2,6-Dinitrotoluene	9.728	165	178867	40.034	ng	100
52) Acenaphthene	9.981	154	869604	41.145	ng	100
53) 3-Nitroaniline	9.892	138	221168	40.485	ng	100
54) 2,4-Dinitrophenol	9.998	184	42260	38.466	ng	100
55) Dibenzofuran	10.151	168	1293035	40.813	ng	100
56) 4-Nitrophenol	10.039	139	159989	42.900	ng	100
57) 2,4-Dinitrotoluene	10.128	165	211218	39.352	ng	100
58) Fluorene	10.498	166	976948	40.554	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.269	232	247655	43.967	ng	100
60) Diethylphthalate	10.369	149	974036	41.573	ng	100
61) 4-Chlorophenyl-phenyle...	10.486	204	468898	40.986	ng	100
62) 4-Nitroaniline	10.504	138	201491	40.498	ng	100
63) Azobenzene	10.651	77	948440	40.994	ng	100
65) 4,6-Dinitro-2-methylph...	10.534	198	67678	36.574	ng	100
66) n-Nitrosodiphenylamine	10.604	169	877338	41.656	ng	100
67) 4-Bromophenyl-phenylether	10.981	248	291063	42.400	ng	100
68) Hexachlorobenzene	11.045	284	321448	41.809	ng	100
69) Atrazine	11.128	200	234825	42.299	ng	100
70) Pentachlorophenol	11.233	266	177755	43.764	ng	100
71) Phenanthrene	11.457	178	1371009	41.176	ng	100
72) Anthracene	11.510	178	1390308	40.840	ng	100
73) Carbazole	11.663	167	1273184	40.999	ng	100
74) Di-n-butylphthalate	11.998	149	1317363	42.180	ng	100
75) Fluoranthene	12.645	202	1338006	40.652	ng	100
77) Benzidine	12.769	184	638142	45.346	ng	100
78) Pyrene	12.880	202	1341800	43.920	ng	100
80) Butylbenzylphthalate	13.498	149	334175	40.624	ng	100
81) Benzo(a)anthracene	14.063	228	952290	42.831	ng	100
82) 3,3'-Dichlorobenzidine	14.027	252	292958	44.859	ng	100
83) Chrysene	14.104	228	816041	40.321	ng	100
84) Bis(2-ethylhexyl)phtha...	14.057	149	460074	40.356	ng	100
85) Di-n-octyl phthalate	14.680	149	836078	40.504	ng	100
87) Indeno(1,2,3-cd)pyrene	17.080	276	1059591	44.049	ng	100
88) Benzo(b)fluoranthene	15.127	252	854971	41.912	ng	100
89) Benzo(k)fluoranthene	15.157	252	752135	40.310	ng	100
90) Benzo(a)pyrene	15.504	252	790330	42.318	ng	100
91) Dibenzo(a,h)anthracene	17.098	278	861783	44.142	ng	100
92) Benzo(g,h,i)perylene	17.539	276	858804	43.758	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142243.D
Acq On : 30 Apr 2025 13:17
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1


Quant Time: Apr 30 15:43:02 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:152 Resp: 18000


Ion	Ratio	Lower	Upper
152	100		
150	162.7	130.2	195.2
115	58.7	47.0	70.4



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#2
1,4-Dioxane
Concen: 42.319 ng
RT: 2.699 min Scan# 86
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion: 88 Resp: 213029

Ion	Ratio	Lower	Upper
88	100		
58	90.2	72.2	108.2
43	33.8	27.0	40.6






#3
Pyridine
Concen: 41.886 ng
RT: 3.463 min Scan# 21
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion: 79 Resp: 54798


Ion	Ratio	Lower	Upper
79	100		
52	89.4	71.5	107.3
51	43.1	34.5	51.7



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#4
n-Nitrosodimethylamine
Concen: 42.887 ng
RT: 3.410 min Scan# 207
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion: 42 Resp: 290357

Ion	Ratio	Lower	Upper
42	100		
74	111.9	89.5	134.3
44	8.2	6.6	9.8






#5
2-Fluorophenol
Concen: 83.944 ng
RT: 5.516 min Scan# 5
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:112 Resp: 94615


Ion	Ratio	Lower	Upper
112	100		
64	60.5	48.4	72.6
63	31.0	24.8	37.2



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#6
Aniline
Concen: 41.959 ng
RT: 6.569 min Scan# 744
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion: 93 Resp: 808338

Ion	Ratio	Lower	Upper
93	100		
66	40.1	32.1	48.1
65	20.7	16.6	24.8






#7
Phenol-d6
Concen: 84.267 ng
RT: 6.528 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion: 99 Resp: 114692


Ion	Ratio	Lower	Upper
99	100		
42	20.4	16.3	24.5
71	31.2	25.0	37.4



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#8
2-Chlorophenol
Concen: 42.612 ng
RT: 6.693 min Scan# 765
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:128 Resp: 496290

Ion	Ratio	Lower	Upper
128	100		
130	33.1	13.1	53.1
64	48.6	28.6	68.6






#9
Benzaldehyde
Concen: 42.263 ng
RT: 6.457 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion: 77 Resp: 40204


Ion	Ratio	Lower	Upper
77	100		
105	99.1	79.1	119.1
106	96.0	76.0	116.0



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#10
Phenol
Concen: 42.128 ng
RT: 6.540 min Scan# 739
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion: 94 Resp: 617257

Ion	Ratio	Lower	Upper
94	100		
65	28.9	8.9	48.9
66	38.7	18.7	58.7






#11
bis(2-Chloroethyl)ether
Concen: 41.780 ng
RT: 6.646 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion: 93 Resp: 47440


Ion	Ratio	Lower	Upper
93	100		
63	82.2	62.2	102.2
95	32.7	12.7	52.7



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#12
1,3-Dichlorobenzene
Concen: 41.881 ng
RT: 6.851 min Scan# 792
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:146 Resp: 551813

Ion	Ratio	Lower	Upper
146	100		
148	63.1	50.5	75.7
75	29.0	23.2	34.8






#13
1,4-Dichlorobenzene
Concen: 41.937 ng
RT: 6.928 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:146 Resp: 55669


Ion	Ratio	Lower	Upper
146	100		
148	63.3	50.6	76.0
111	41.7	33.4	50.0



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#14
1,2-Dichlorobenzene
Concen: 41.892 ng
RT: 7.081 min Scan# 831
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:146 Resp: 530895

Ion	Ratio	Lower	Upper
146	100		
148	63.1	50.5	75.7
111	43.6	34.9	52.3





#15
Benzyl Alcohol
Concen: 42.870 ng
RT: 7.046 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Tgt Ion: 79 Resp: 41383


Ion	Ratio	Lower	Upper
79	100		
108	80.6	64.5	96.7
77	61.5	49.2	73.8

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#16
2,2'-oxybis(1-Chloropropane)
Concen: 41.888 ng
RT: 7.187 min Scan# 849
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion: 45 Resp: 858875

Ion	Ratio	Lower	Upper
45	100		
77	11.3	0.0	31.3
79	8.5	0.0	28.5






#17
2-Methylphenol
Concen: 42.287 ng
RT: 7.151 min Scan# 84
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:107 Resp: 40346


Ion	Ratio	Lower	Upper
107	100		
108	113.7	91.0	136.4
77	43.0	34.4	51.6
79	43.0	34.4	51.6



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#18
Hexachloroethane
Concen: 42.322 ng
RT: 7.428 min Scan# 890
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:117 Resp: 183881

Ion	Ratio	Lower	Upper
117	100		
119	96.4	77.1	115.7
201	99.3	79.4	119.2






#19
n-Nitroso-di-n-propylamine
Concen: 41.410 ng
RT: 7.322 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion: 70 Resp: 34828


Ion	Ratio	Lower	Upper
70	100		
42	66.8	53.4	80.2
101	9.6	7.7	11.5
130	21.5	17.2	25.8



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#20
3+4-Methylphenols
Concen: 42.229 ng
RT: 7.304 min Scan# 869
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:107 Resp: 503856

Ion	Ratio	Lower	Upper
107	100		
108	91.2	71.2	111.2
77	61.1	41.1	81.1
79	28.1	8.1	48.1






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.193 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:136 Resp: 70624


Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.7	13.1
54	9.6	7.7	11.5
68	6.3	5.0	7.6



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#22
Acetophenone
Concen: 40.454 ng
RT: 7.316 min Scan# 871
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:105 Resp: 660959

Ion	Ratio	Lower	Upper
105	100		
71	7.1	5.7	8.5
51	37.6	30.1	45.1
120	23.7	19.0	28.4





#23
Nitrobenzene-d5
Concen: 89.396 ng
RT: 7.475 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Tgt Ion: 82 Resp: 96100


Ion	Ratio	Lower	Upper
82	100		
128	46.3	37.0	55.6
54	60.0	48.0	72.0

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#24
Nitrobenzene
Concen: 43.929 ng
RT: 7.493 min Scan# 901
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion: 77 Resp: 463357

Ion	Ratio	Lower	Upper
77	100		
123	48.3	38.6	58.0
65	14.4	11.5	17.3






#25
Isophorone
Concen: 41.299 ng
RT: 7.728 min Scan# 94
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion: 82 Resp: 93176


Ion	Ratio	Lower	Upper
82	100		
95	7.0	5.6	8.4
138	19.1	15.3	22.9



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#26
2-Nitrophenol
Concen: 38.737 ng
RT: 7.804 min Scan# 954
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:139 Resp: 162563

Ion	Ratio	Lower	Upper
139	100		
109	27.3	21.8	32.8
65	51.4	41.1	61.7






#27
2,4-Dimethylphenol
Concen: 41.862 ng
RT: 7.834 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:122 Resp: 47618


Ion	Ratio	Lower	Upper
122	100		
107	107.5	86.0	129.0
121	57.9	46.3	69.5



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#28
bis(2-Chloroethoxy)methane
Concen: 41.247 ng
RT: 7.940 min Scan# 977
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion: 93 Resp: 586887

Ion	Ratio	Lower	Upper
93	100		
95	32.7	26.2	39.2
123	12.1	9.7	14.5






#29
2,4-Dichlorophenol
Concen: 43.163 ng
RT: 8.040 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:162 Resp: 41048


Ion	Ratio	Lower	Upper
162	100		
164	63.0	43.0	83.0
98	36.7	16.7	56.7



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#30
1,2,4-Trichlorobenzene
Concen: 41.539 ng
RT: 8.134 min Scan# 1010
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:180 Resp: 440825

Ion	Ratio	Lower	Upper
180	100		
182	94.6	75.7	113.5
145	31.7	25.4	38.0






#31
Naphthalene
Concen: 40.920 ng
RT: 8.216 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:128 Resp: 144124


Ion	Ratio	Lower	Upper
128	100		
129	11.2	9.0	13.4
127	13.2	10.6	15.8



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#32
Benzoic acid
Concen: 37.083 ng m
RT: 7.934 min Scan# 976
Delta R.T. -0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:122 Resp: 184052

Ion	Ratio	Lower	Upper
122	100		
105	118.2	98.2	138.2
77	86.8	66.8	106.8






#33
4-Chloroaniline
Concen: 41.328 ng
RT: 8.257 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:127 Resp: 60613


Ion	Ratio	Lower	Upper
127	100		
129	32.6	26.1	39.1
65	31.9	25.5	38.3
92	19.3	15.4	23.2



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#34
Hexachlorobutadiene
Concen: 42.125 ng
RT: 8.334 min Scan# 1044
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:225 Resp: 260821

Ion	Ratio	Lower	Upper
225	100		
223	62.1	49.7	74.5
227	63.7	51.0	76.4





#35
Caprolactam
Concen: 42.889 ng
RT: 8.622 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Tgt Ion:113 Resp: 12259


Ion	Ratio	Lower	Upper
113	100		
55	197.7	177.7	217.7
56	147.1	127.1	167.1

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#36
4-Chloro-3-methylphenol
Concen: 42.715 ng
RT: 8.728 min Scan# 1111
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:107 Resp: 417126

Ion	Ratio	Lower	Upper
107	100		
144	26.2	21.0	31.4
142	82.6	66.1	99.1






#37
2-Methylnaphthalene
Concen: 40.652 ng
RT: 8.904 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:142 Resp: 88031


Ion	Ratio	Lower	Upper
142	100		
141	88.1	70.5	105.7



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#38
1-Methylnaphthalene
Concen: 41.054 ng
RT: 9.004 min Scan# 1158
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:142 Resp: 920983

Ion	Ratio	Lower	Upper
142	100		
141	91.0	72.8	109.2






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.951 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:164 Resp: 370869


Ion	Ratio	Lower	Upper
164	100		
162	98.2	78.6	117.8
160	44.6	35.7	53.5



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 40.938 ng
RT: 9.069 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:216 Resp: 427286

Ion	Ratio	Lower	Upper
216	100		
214	79.7	63.8	95.6
179	20.9	16.7	25.1
108	19.5	15.6	23.4





#41
Hexachlorocyclopentadiene
Concen: 44.056 ng
RT: 9.057 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Instrument :
BNA_F
ClientSampleId :
SSTDICCC040


Tgt Ion:237 Resp: 26097

Ion	Ratio	Lower	Upper
237	100		
235	63.5	43.5	83.5
272	12.9	0.0	32.9

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#42
2,4,6-Tribromophenol
Concen: 89.389 ng
RT: 10.734 min Scan# 1452
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:330 Resp: 302223

Ion	Ratio	Lower	Upper
330	100		
332	95.6	76.5	114.7
141	30.4	24.3	36.5






#43
2,4,6-Trichlorophenol
Concen: 42.450 ng
RT: 9.175 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:196 Resp: 27455


Ion	Ratio	Lower	Upper
196	100		
198	96.8	77.4	116.2
200	30.8	24.6	37.0



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#44
2,4,5-Trichlorophenol
Concen: 44.113 ng
RT: 9.210 min Scan# 1193
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:196 Resp: 299967

Ion	Ratio	Lower	Upper
196	100		
198	94.0	75.2	112.8
97	49.8	39.8	59.8
132	29.8	23.8	35.8






#45
2-Fluorobiphenyl
Concen: 79.446 ng
RT: 9.269 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:172 Resp: 208970


Ion	Ratio	Lower	Upper
172	100		
171	36.2	29.0	43.4
170	24.2	19.4	29.0



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#46
1,1'-Biphenyl
Concen: 40.664 ng
RT: 9.369 min Scan# 1220
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:154 Resp: 1172400

Ion	Ratio	Lower	Upper
154	100		
153	40.5	20.5	60.5
76	12.8	0.0	32.8






#47
2-Chloronaphthalene
Concen: 40.984 ng
RT: 9.392 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:162 Resp: 87915


Ion	Ratio	Lower	Upper
162	100		
127	40.8	32.6	49.0
164	32.6	26.1	39.1



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#48
2-Nitroaniline
Concen: 39.600 ng
RT: 9.487 min Scan# 1240
Delta R.T. -0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion: 65 Resp: 218766

Ion	Ratio	Lower	Upper
65	100		
92	67.9	54.3	81.5
138	110.8	88.6	133.0





#49
Acenaphthylene
Concen: 40.929 ng
RT: 9.810 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Instrument :
BNA_F
ClientSampleId :
SSTDICCC040


Tgt Ion:152 Resp: 148154

Ion	Ratio	Lower	Upper
152	100		
151	20.2	16.2	24.2
153	13.4	10.7	16.1

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#50
Dimethylphthalate
Concen: 41.312 ng
RT: 9.669 min Scan# 1271
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:163 Resp: 980172

Ion	Ratio	Lower	Upper
163	100		
194	4.0	3.2	4.8
164	10.2	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 40.034 ng
RT: 9.728 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Tgt Ion:165 Resp: 17886


Ion	Ratio	Lower	Upper
165	100		
63	60.6	48.5	72.7
89	54.9	43.9	65.9

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#52
Acenaphthene
Concen: 41.145 ng
RT: 9.981 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:154 Resp: 869604

Ion	Ratio	Lower	Upper
154	100		
153	110.8	88.6	133.0
152	52.6	42.1	63.1






#53
3-Nitroaniline
Concen: 40.485 ng
RT: 9.892 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:138 Resp: 22116


Ion	Ratio	Lower	Upper
138	100		
108	14.6	11.7	17.5
92	114.0	91.2	136.8



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#54
2,4-Dinitrophenol
Concen: 38.466 ng
RT: 9.998 min Scan# 1327
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:184 Resp: 42260

Ion	Ratio	Lower	Upper
184	100		
63	86.5	69.2	103.8
154	339.3	271.4	407.2






#55
Dibenzofuran
Concen: 40.813 ng
RT: 10.151 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:168 Resp: 129303


Ion	Ratio	Lower	Upper
168	100		
139	36.7	29.4	44.0
169	13.7	11.0	16.4



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#56
4-Nitrophenol
Concen: 42.900 ng
RT: 10.039 min Scan# 1334
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:139 Resp: 159989

Ion	Ratio	Lower	Upper
139	100		
109	64.9	44.9	84.9
65	94.7	74.7	114.7






#57
2,4-Dinitrotoluene
Concen: 39.352 ng
RT: 10.128 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:165 Resp: 21121


Ion	Ratio	Lower	Upper
165	100		
63	48.5	38.8	58.2
89	71.4	57.1	85.7
182	2.7	2.2	3.2



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#58
Fluorene
Concen: 40.554 ng
RT: 10.498 min Scan# 1412
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:166 Resp: 976948

Ion	Ratio	Lower	Upper
166	100		
165	92.4	73.9	110.9
167	13.5	10.8	16.2





#59
2,3,4,6-Tetrachlorophenol
Concen: 43.967 ng
RT: 10.269 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Instrument :
BNA_F
ClientSampleId :
SSTDICCC040


Tgt Ion:232 Resp: 24765

Ion	Ratio	Lower	Upper
232	100		
131	44.9	35.9	53.9
130	2.2	1.8	2.6
166	30.0	24.0	36.0

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#60
Diethylphthalate
Concen: 41.573 ng
RT: 10.369 min Scan# 1390
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:149 Resp: 974036

Ion	Ratio	Lower	Upper
149	100		
177	22.0	17.6	26.4
150	12.7	10.2	15.2





#61
4-Chlorophenyl-phenylether
Concen: 40.986 ng
RT: 10.486 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Tgt Ion:204 Resp: 46889


Ion	Ratio	Lower	Upper
204	100		
206	32.0	25.6	38.4
141	61.1	48.9	73.3

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#62
4-Nitroaniline
Concen: 40.498 ng
RT: 10.504 min Scan# 1413
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:138 Resp: 201491

Ion	Ratio	Lower	Upper
138	100		
92	49.2	29.2	69.2
108	69.1	49.1	89.1





#63
Azobenzene
Concen: 40.994 ng
RT: 10.651 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Instrument :
BNA_F
ClientSampleId :
SSTDICCC040


Tgt Ion: 77 Resp: 94844

Ion	Ratio	Lower	Upper
77	100		
182	27.3	7.3	47.3
105	22.5	2.5	42.5
51	34.4	14.4	54.4

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:188 Resp: 616408

Ion	Ratio	Lower	Upper
188	100		
94	10.2	8.2	12.2
80	10.7	8.6	12.8





#65
4,6-Dinitro-2-methylphenol
Concen: 36.574 ng
RT: 10.534 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Tgt Ion:198 Resp: 6767


Ion	Ratio	Lower	Upper
198	100		
51	60.6	40.6	80.6
105	49.2	29.2	69.2

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#66
n-Nitrosodiphenylamine
Concen: 41.656 ng
RT: 10.604 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:169 Resp: 877338

Ion	Ratio	Lower	Upper
169	100		
168	66.3	53.0	79.6
167	36.0	28.8	43.2






#67
4-Bromophenyl-phenylether
Concen: 42.400 ng
RT: 10.981 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:248 Resp: 29106


Ion	Ratio	Lower	Upper
248	100		
250	96.5	77.2	115.8
141	73.0	58.4	87.6



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#68
Hexachlorobenzene
Concen: 41.809 ng
RT: 11.045 min Scan# 1505
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:284 Resp: 321448

Ion	Ratio	Lower	Upper
284	100		
142	30.8	24.6	37.0
249	29.0	23.2	34.8





#69
Atrazine
Concen: 42.299 ng
RT: 11.128 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Instrument :
BNA_F
ClientSampleId :
SSTDICCC040


Tgt Ion:200 Resp: 23482

Ion	Ratio	Lower	Upper
200	100		
173	28.6	8.6	48.6
215	48.5	28.5	68.5

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#70
Pentachlorophenol
Concen: 43.764 ng
RT: 11.233 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:266 Resp: 177755

Ion	Ratio	Lower	Upper
266	100		
268	62.5	50.0	75.0
264	61.6	49.3	73.9






#71
Phenanthrene
Concen: 41.176 ng
RT: 11.457 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:178 Resp: 137100


Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.5	23.3
179	15.8	12.6	19.0



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#72
Anthracene
Concen: 40.840 ng
RT: 11.510 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:178 Resp: 1390308

Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.3	22.9
179	15.9	12.7	19.1






#73
Carbazole
Concen: 40.999 ng
RT: 11.663 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:167 Resp: 127318


Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.3	25.9
139	12.9	10.3	15.5



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#74
Di-n-butylphthalate
Concen: 42.180 ng
RT: 11.998 min Scan# 1667
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:149 Resp: 1317363

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.5	11.3
104	4.5	3.6	5.4






#75
Fluoranthene
Concen: 40.652 ng
RT: 12.645 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:202 Resp: 133800


Ion	Ratio	Lower	Upper
202	100		
101	12.4	0.0	32.4
203	17.5	0.0	37.5



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.074 min Scan# 2020
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:240 Resp: 331192

Ion	Ratio	Lower	Upper
240	100		
120	12.6	10.1	15.1
236	25.6	20.5	30.7






#77
Benzidine
Concen: 45.346 ng
RT: 12.769 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:184 Resp: 63814


Ion	Ratio	Lower	Upper
184	100		
185	14.7	11.8	17.6
183	11.1	8.9	13.3



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#78
Pyrene
Concen: 43.920 ng
RT: 12.880 min Scan# 1817
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:202 Resp: 1341800

Ion	Ratio	Lower	Upper
202	100		
200	20.1	16.1	24.1
203	17.6	14.1	21.1






#79
Terphenyl-d14
Concen: 86.181 ng
RT: 13.022 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:244 Resp: 196247


Ion	Ratio	Lower	Upper
244	100		
212	6.6	5.3	7.9
122	11.4	9.1	13.7



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#80
Butylbenzylphthalate
Concen: 40.624 ng
RT: 13.498 min Scan# 1922
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:149 Resp: 334175

Ion	Ratio	Lower	Upper
149	100		
91	68.0	54.4	81.6
206	19.9	15.9	23.9






#81
Benzo(a)anthracene
Concen: 42.831 ng
RT: 14.063 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:228 Resp: 95229


Ion	Ratio	Lower	Upper
228	100		
226	27.4	21.9	32.9
229	20.0	16.0	24.0



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#82
3,3'-Dichlorobenzidine
Concen: 44.859 ng
RT: 14.027 min Scan# 2012
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:252 Resp: 292958

Ion	Ratio	Lower	Upper
252	100		
254	63.8	51.0	76.6
126	11.8	9.4	14.2






#83
Chrysene
Concen: 40.321 ng
RT: 14.104 min Scan# 2(
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:228 Resp: 81604


Ion	Ratio	Lower	Upper
228	100		
226	30.0	24.0	36.0
229	19.6	15.7	23.5



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 40.356 ng
RT: 14.057 min Scan# 2017
Delta R.T. -0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:149 Resp: 460074

Ion	Ratio	Lower	Upper
149	100		
167	27.1	21.7	32.5
279	3.7	3.0	4.4






#85
Di-n-octyl phthalate
Concen: 40.504 ng
RT: 14.680 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:149 Resp: 83607


Ion	Ratio	Lower	Upper
149	100		
167	1.4	1.1	1.7
43	11.6	9.3	13.9



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 15.568 min Scan# 2274
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:264 Resp: 328405

Ion	Ratio	Lower	Upper
264	100		
260	23.0	18.4	27.6
265	21.6	17.3	25.9






#87
Indeno(1,2,3-cd)pyrene
Concen: 44.049 ng
RT: 17.080 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:276 Resp: 105959


Ion	Ratio	Lower	Upper
276	100		
138	28.2	22.6	33.8
277	25.1	20.1	30.1



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#88
Benzo(b)fluoranthene
Concen: 41.912 ng
RT: 15.127 min Scan# 2199
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:252 Resp: 854971

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	11.2	9.0	13.4






#89
Benzo(k)fluoranthene
Concen: 40.310 ng
RT: 15.157 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:252 Resp: 75213


Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.8	26.8
125	11.3	9.0	13.6



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#90
Benzo(a)pyrene
Concen: 42.318 ng
RT: 15.504 min Scan# 2263
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:252 Resp: 790330

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.6	26.4
125	11.9	9.5	14.3





#91
Dibenzo(a,h)anthracene
Concen: 44.142 ng
RT: 17.098 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Instrument :
BNA_F
ClientSampleId :
SSTDICCC040


Tgt Ion:278 Resp: 86178

Ion	Ratio	Lower	Upper
278	100		
139	19.0	15.2	22.8
279	23.8	19.0	28.6

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025






#92
Benzo(g,h,i)perylene
Concen: 43.758 ng
RT: 17.539 min Scan# 2609
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:276 Resp: 858804

Ion	Ratio	Lower	Upper
276	100		
277	23.4	18.7	28.1
138	24.2	19.4	29.0



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142244.D
Acq On : 30 Apr 2025 13:46
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Apr 30 15:43:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.910	152	179073	20.000	ng		0.00
21) Naphthalene-d8	8.192	136	691236	20.000	ng		0.00
39) Acenaphthene-d10	9.951	164	360144	20.000	ng		0.00
64) Phenanthrene-d10	11.433	188	604171	20.000	ng		0.00
76) Chrysene-d12	14.074	240	322957	20.000	ng		0.00
86) Perylene-d12	15.568	264	337726	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.516	112	1150166	102.577	ng		0.00
7) Phenol-d6	6.528	99	1387418	102.468	ng		0.00
23) Nitrobenzene-d5	7.475	82	1213856	115.369	ng		0.00
42) 2,4,6-Tribromophenol	10.739	330	369823	112.641	ng		0.00
45) 2-Fluorobiphenyl	9.269	172	2425708	94.966	ng		0.00
79) Terphenyl-d14	13.021	244	2247982	101.236	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.693	88	259869	51.894	ng		99
3) Pyridine	3.463	79	672711	51.688	ng		100
4) n-Nitrosodimethylamine	3.416	42	353028	52.416	ng		96
6) Aniline	6.575	93	986118	51.454	ng		99
8) 2-Chlorophenol	6.692	128	606252	52.325	ng		99
9) Benzaldehyde	6.463	77	482619	50.998	ng		98
10) Phenol	6.539	94	755326m	51.820	ng		
11) bis(2-Chloroethyl)ether	6.645	93	575347	50.934	ng		100
12) 1,3-Dichlorobenzene	6.851	146	661612	50.476	ng		99
13) 1,4-Dichlorobenzene	6.928	146	666083	50.439	ng		99
14) 1,2-Dichlorobenzene	7.081	146	640983	50.843	ng		100
15) Benzyl Alcohol	7.045	79	504297	52.514	ng		99
16) 2,2'-oxybis(1-Chloropr...	7.186	45	1049319	51.444	ng		100
17) 2-Methylphenol	7.151	107	485145	51.113	ng		100
18) Hexachloroethane	7.428	117	229911	53.193	ng		97
19) n-Nitroso-di-n-propyla...	7.328	70	420535	50.261	ng		99
20) 3+4-Methylphenols	7.310	107	611404	51.510	ng	#	89
22) Acetophenone	7.322	105	796460	49.805	ng		96
24) Nitrobenzene	7.492	77	584670	56.634	ng		99
25) Isophorone	7.734	82	1129933	51.170	ng		99
26) 2-Nitrophenol	7.810	139	228021	51.526	ng		97
27) 2,4-Dimethylphenol	7.839	122	576054	51.741	ng		100
28) bis(2-Chloroethoxy)met...	7.939	93	708439	50.870	ng		99
29) 2,4-Dichlorophenol	8.045	162	494488	53.125	ng		99
30) 1,2,4-Trichlorobenzene	8.133	180	524508	50.497	ng		100
31) Naphthalene	8.216	128	1725818	50.064	ng		100
32) Benzoic acid	7.939	122	252184	47.667	ng		98
33) 4-Chloroaniline	8.257	127	730860	50.914	ng		99
34) Hexachlorobutadiene	8.333	225	307118	50.679	ng		99
35) Caprolactam	8.628	113	154579	55.251	ng		97
36) 4-Chloro-3-methylphenol	8.733	107	505527	52.891	ng		100
37) 2-Methylnaphthalene	8.904	142	1057924	49.915	ng		100
38) 1-Methylnaphthalene	9.004	142	1107217	50.427	ng		98
40) 1,2,4,5-Tetrachloroben...	9.069	216	509426	50.261	ng		99
41) Hexachlorocyclopentadiene	9.063	237	331287	57.591	ng		99
43) 2,4,6-Trichlorophenol	9.180	196	337910	53.801	ng		100

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142244.D
Acq On : 30 Apr 2025 13:46
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Apr 30 15:43:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.216	196	370489	56.106	ng	98
46) 1,1'-Biphenyl	9.369	154	1392513	49.736	ng	100
47) 2-Chloronaphthalene	9.398	162	1041842	50.015	ng	100
48) 2-Nitroaniline	9.486	65	284370	51.394	ng	100
49) Acenaphthylene	9.810	152	1769441	50.339	ng	100
50) Dimethylphthalate	9.669	163	1178638	51.156	ng	99
51) 2,6-Dinitrotoluene	9.727	165	232906	52.309	ng	99
52) Acenaphthene	9.986	154	1024409	49.912	ng	100
53) 3-Nitroaniline	9.898	138	279608	51.745	ng	97
54) 2,4-Dinitrophenol	9.998	184	62086	50.725	ng	98
55) Dibenzofuran	10.157	168	1542097	50.123	ng	100
56) 4-Nitrophenol	10.039	139	210153	58.029	ng	99
57) 2,4-Dinitrotoluene	10.127	165	280193	51.918	ng	98
58) Fluorene	10.498	166	1146269	49.000	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.269	232	300782	54.989	ng	99
60) Diethylphthalate	10.369	149	1170793	51.459	ng	99
61) 4-Chlorophenyl-phenyle...	10.492	204	553271	49.802	ng	98
62) 4-Nitroaniline	10.510	138	258117	52.547	ng	100
63) Azobenzene	10.651	77	1150899	51.226	ng	99
65) 4,6-Dinitro-2-methylph...	10.539	198	98114	49.012	ng	95
66) n-Nitrosodiphenylamine	10.610	169	1041277	50.442	ng	100
67) 4-Bromophenyl-phenylether	10.980	248	340298	50.576	ng	97
68) Hexachlorobenzene	11.045	284	384807	51.064	ng	99
69) Atrazine	11.133	200	287405	52.819	ng	100
70) Pentachlorophenol	11.233	266	219414	55.114	ng	98
71) Phenanthrene	11.463	178	1607091	49.244	ng	99
72) Anthracene	11.510	178	1659788	49.744	ng	100
73) Carbazole	11.663	167	1504715	49.436	ng	100
74) Di-n-butylphthalate	11.998	149	1578769	51.573	ng	100
75) Fluoranthene	12.651	202	1563998	48.481	ng	99
77) Benzidine	12.768	184	762858	55.590	ng	99
78) Pyrene	12.880	202	1563568	52.484	ng	100
80) Butylbenzylphthalate	13.498	149	423529	51.685	ng	100
81) Benzo(a)anthracene	14.068	228	1120699	51.691	ng	100
82) 3,3'-Dichlorobenzidine	14.027	252	363124	57.021	ng	98
83) Chrysene	14.104	228	990991	50.214	ng	100
84) Bis(2-ethylhexyl)phtha...	14.057	149	585800	51.638	ng	100
85) Di-n-octyl phthalate	14.680	149	1065626	51.078	ng	99
87) Indeno(1,2,3-cd)pyrene	17.080	276	1313336	53.091	ng	99
88) Benzo(b)fluoranthene	15.127	252	1073096	51.153	ng	100
89) Benzo(k)fluoranthene	15.162	252	967248	50.408	ng	98
90) Benzo(a)pyrene	15.504	252	996599	51.890	ng	100
91) Dibenzo(a,h)anthracene	17.103	278	1073246	53.457	ng	99
92) Benzo(g,h,i)perylene	17.539	276	1069402	52.984	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142244.D
Acq On : 30 Apr 2025 13:46
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1


Quant Time: Apr 30 15:43:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:152 Resp: 17907


Ion	Ratio	Lower	Upper
152	100		
150	157.6	130.2	195.2
115	59.6	47.0	70.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#2
1,4-Dioxane
Concen: 51.894 ng
RT: 2.693 min Scan# 85
Delta R.T. -0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion: 88 Resp: 259869

Ion	Ratio	Lower	Upper
88	100		
58	90.5	72.2	108.2
43	35.0	27.0	40.6






#3
Pyridine
Concen: 51.688 ng
RT: 3.463 min Scan# 21
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion: 79 Resp: 67271


Ion	Ratio	Lower	Upper
79	100		
52	89.2	71.5	107.3
51	42.6	34.5	51.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#4
n-Nitrosodimethylamine
Concen: 52.416 ng
RT: 3.416 min Scan# 208
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion: 42 Resp: 353028

Ion	Ratio	Lower	Upper
42	100		
74	116.0	89.5	134.3
44	7.9	6.6	9.8





#5
2-Fluorophenol
Concen: 102.577 ng
RT: 5.516 min Scan# 5
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46


Tgt Ion:112 Resp: 115016


Ion	Ratio	Lower	Upper
112	100		
64	60.8	48.4	72.6
63	31.3	24.8	37.2

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#6
Aniline
Concen: 51.454 ng
RT: 6.575 min Scan# 745
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion: 93 Resp: 986118

Ion	Ratio	Lower	Upper
93	100		
66	39.7	32.1	48.1
65	20.2	16.6	24.8






#7
Phenol-d6
Concen: 102.468 ng
RT: 6.528 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion: 99 Resp: 138741


Ion	Ratio	Lower	Upper
99	100		
42	20.5	16.3	24.5
71	31.7	25.0	37.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#8
2-Chlorophenol
Concen: 52.325 ng
RT: 6.692 min Scan# 765
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:128 Resp: 606252

Ion	Ratio	Lower	Upper
128	100		
130	32.3	13.1	53.1
64	48.9	28.6	68.6






#9
Benzaldehyde
Concen: 50.998 ng
RT: 6.463 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion: 77 Resp: 48261


Ion	Ratio	Lower	Upper
77	100		
105	100.3	79.1	119.1
106	97.8	76.0	116.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#10
Phenol
Concen: 51.820 ng m
RT: 6.539 min Scan# 739
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion: 94 Resp: 755326

Ion	Ratio	Lower	Upper
94	100		
65	28.8	8.9	48.9
66	40.2	18.7	58.7






#11
bis(2-Chloroethyl)ether
Concen: 50.934 ng
RT: 6.645 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion: 93 Resp: 57534


Ion	Ratio	Lower	Upper
93	100		
63	82.3	62.2	102.2
95	32.5	12.7	52.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#12
1,3-Dichlorobenzene
Concen: 50.476 ng
RT: 6.851 min Scan# 792
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:146 Resp: 661612

Ion	Ratio	Lower	Upper
146	100		
148	63.6	50.5	75.7
75	30.6	23.2	34.8






#13
1,4-Dichlorobenzene
Concen: 50.439 ng
RT: 6.928 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:146 Resp: 66608


Ion	Ratio	Lower	Upper
146	100		
148	63.4	50.6	76.0
111	42.4	33.4	50.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#14
1,2-Dichlorobenzene
Concen: 50.843 ng
RT: 7.081 min Scan# 831
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:146 Resp: 640983

Ion	Ratio	Lower	Upper
146	100		
148	63.4	50.5	75.7
111	43.7	34.9	52.3





#15
Benzyl Alcohol
Concen: 52.514 ng
RT: 7.045 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46


Tgt Ion: 79 Resp: 50429


Ion	Ratio	Lower	Upper
79	100		
108	80.2	64.5	96.7
77	62.5	49.2	73.8

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#16
2,2'-oxybis(1-Chloropropane)
Concen: 51.444 ng
RT: 7.186 min Scan# 849
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion: 45 Resp: 1049319

Ion	Ratio	Lower	Upper
45	100		
77	11.3	0.0	31.3
79	8.6	0.0	28.5






#17
2-Methylphenol
Concen: 51.113 ng
RT: 7.151 min Scan# 84
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:107 Resp: 48514


Ion	Ratio	Lower	Upper
107	100		
108	113.7	91.0	136.4
77	43.6	34.4	51.6
79	43.1	34.4	51.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#18
Hexachloroethane
Concen: 53.193 ng
RT: 7.428 min Scan# 890
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:117 Resp: 229911

Ion	Ratio	Lower	Upper
117	100		
119	95.9	77.1	115.7
201	94.8	79.4	119.2






#19
n-Nitroso-di-n-propylamine
Concen: 50.261 ng
RT: 7.328 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion: 70 Resp: 42053


Ion	Ratio	Lower	Upper
70	100		
42	66.4	53.4	80.2
101	9.9	7.7	11.5
130	22.3	17.2	25.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#20
3+4-Methylphenols
Concen: 51.510 ng
RT: 7.310 min Scan# 870
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:107 Resp: 611404

Ion	Ratio	Lower	Upper
107	100		
108	91.7	71.2	111.2
77	84.0	41.1	81.1#
79	28.3	8.1	48.1






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.192 min Scan# 1(
Delta R.T. -0.001 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:136 Resp: 69123(


Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.7	13.1
54	9.7	7.7	11.5
68	6.3	5.0	7.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#22
Acetophenone
Concen: 49.805 ng
RT: 7.322 min Scan# 872
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:105 Resp: 796460

Ion	Ratio	Lower	Upper
105	100		
71	7.9	5.7	8.5
51	33.2	30.1	45.1
120	23.5	19.0	28.4





#23
Nitrobenzene-d5
Concen: 115.369 ng
RT: 7.475 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46


Tgt Ion: 82 Resp: 121385


Ion	Ratio	Lower	Upper
82	100		
128	46.3	37.0	55.6
54	59.8	48.0	72.0

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#24
Nitrobenzene
Concen: 56.634 ng
RT: 7.492 min Scan# 901
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion: 77 Resp: 584670

Ion	Ratio	Lower	Upper
77	100		
123	47.9	38.6	58.0
65	14.1	11.5	17.3






#25
Isophorone
Concen: 51.170 ng
RT: 7.734 min Scan# 94
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion: 82 Resp: 112993


Ion	Ratio	Lower	Upper
82	100		
95	7.1	5.6	8.4
138	19.9	15.3	22.9



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#26
2-Nitrophenol
Concen: 51.526 ng
RT: 7.810 min Scan# 955
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:139 Resp: 228021

Ion	Ratio	Lower	Upper
139	100		
109	26.7	21.8	32.8
65	48.1	41.1	61.7






#27
2,4-Dimethylphenol
Concen: 51.741 ng
RT: 7.839 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:122 Resp: 57605


Ion	Ratio	Lower	Upper
122	100		
107	107.3	86.0	129.0
121	57.9	46.3	69.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#28
bis(2-Chloroethoxy)methane
Concen: 50.870 ng
RT: 7.939 min Scan# 977
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion: 93 Resp: 708439

Ion	Ratio	Lower	Upper
93	100		
95	32.9	26.2	39.2
123	12.4	9.7	14.5






#29
2,4-Dichlorophenol
Concen: 53.125 ng
RT: 8.045 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:162 Resp: 49448


Ion	Ratio	Lower	Upper
162	100		
164	64.7	43.0	83.0
98	36.5	16.7	56.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#30
1,2,4-Trichlorobenzene
Concen: 50.497 ng
RT: 8.133 min Scan# 1010
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:180 Resp: 524508

Ion	Ratio	Lower	Upper
180	100		
182	94.9	75.7	113.5
145	32.0	25.4	38.0






#31
Naphthalene
Concen: 50.064 ng
RT: 8.216 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:128 Resp: 172581


Ion	Ratio	Lower	Upper
128	100		
129	11.4	9.0	13.4
127	13.3	10.6	15.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#32
Benzoic acid
Concen: 47.667 ng
RT: 7.939 min Scan# 977
Delta R.T. 0.005 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:122 Resp: 252184

Ion	Ratio	Lower	Upper
122	100		
105	120.1	98.2	138.2
77	89.3	66.8	106.8






#33
4-Chloroaniline
Concen: 50.914 ng
RT: 8.257 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:127 Resp: 73086


Ion	Ratio	Lower	Upper
127	100		
129	32.2	26.1	39.1
65	32.0	25.5	38.3
92	18.9	15.4	23.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#34
Hexachlorobutadiene
Concen: 50.679 ng
RT: 8.333 min Scan# 1044
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:225 Resp: 307118

Ion	Ratio	Lower	Upper
225	100		
223	63.2	49.7	74.5
227	63.8	51.0	76.4






#35
Caprolactam
Concen: 55.251 ng
RT: 8.628 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:113 Resp: 15457


Ion	Ratio	Lower	Upper
113	100		
55	194.4	177.7	217.7
56	142.7	127.1	167.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#36
4-Chloro-3-methylphenol
Concen: 52.891 ng
RT: 8.733 min Scan# 1112
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:107 Resp: 505527

Ion	Ratio	Lower	Upper
107	100		
144	26.9	21.0	31.4
142	82.6	66.1	99.1






#37
2-Methylnaphthalene
Concen: 49.915 ng
RT: 8.904 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:142 Resp: 105792


Ion	Ratio	Lower	Upper
142	100		
141	87.8	70.5	105.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#38
1-Methylnaphthalene
Concen: 50.427 ng
RT: 9.004 min Scan# 1158
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:142 Resp: 1107217

Ion	Ratio	Lower	Upper
142	100		
141	89.6	72.8	109.2





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.951 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46


Tgt Ion:164 Resp: 36014


Ion	Ratio	Lower	Upper
164	100		
162	99.8	78.6	117.8
160	44.5	35.7	53.5

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#40
1,2,4,5-Tetrachlorobenzene
Concen: 50.261 ng
RT: 9.069 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:216 Resp: 509426

Ion	Ratio	Lower	Upper
216	100		
214	80.5	63.8	95.6
179	21.3	16.7	25.1
108	19.3	15.6	23.4






#41
Hexachlorocyclopentadiene
Concen: 57.591 ng
RT: 9.063 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:237 Resp: 33128


Ion	Ratio	Lower	Upper
237	100		
235	62.8	43.5	83.5
272	12.9	0.0	32.9



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#42
2,4,6-Tribromophenol
Concen: 112.641 ng
RT: 10.739 min Scan# 1453
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:330 Resp: 369823

Ion	Ratio	Lower	Upper
330	100		
332	97.4	76.5	114.7
141	30.5	24.3	36.5






#43
2,4,6-Trichlorophenol
Concen: 53.801 ng
RT: 9.180 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:196 Resp: 33791


Ion	Ratio	Lower	Upper
196	100		
198	96.7	77.4	116.2
200	30.9	24.6	37.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#44
2,4,5-Trichlorophenol
Concen: 56.106 ng
RT: 9.216 min Scan# 1194
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:196 Resp: 370489

Ion	Ratio	Lower	Upper
196	100		
198	94.7	75.2	112.8
97	47.7	39.8	59.8
132	28.5	23.8	35.8






#45
2-Fluorobiphenyl
Concen: 94.966 ng
RT: 9.269 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:172 Resp: 242570


Ion	Ratio	Lower	Upper
172	100		
171	36.1	29.0	43.4
170	24.1	19.4	29.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#46
1,1'-Biphenyl
Concen: 49.736 ng
RT: 9.369 min Scan# 1220
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:154 Resp: 1392513

Ion	Ratio	Lower	Upper
154	100		
153	40.6	20.5	60.5
76	13.0	0.0	32.8






#47
2-Chloronaphthalene
Concen: 50.015 ng
RT: 9.398 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:162 Resp: 104184


Ion	Ratio	Lower	Upper
162	100		
127	40.6	32.6	49.0
164	32.9	26.1	39.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#48
2-Nitroaniline
Concen: 51.394 ng
RT: 9.486 min Scan# 1240
Delta R.T. -0.001 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion: 65 Resp: 284370

Ion	Ratio	Lower	Upper
65	100		
92	67.9	54.3	81.5
138	110.9	88.6	133.0






#49
Acenaphthylene
Concen: 50.339 ng
RT: 9.810 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:152 Resp: 176944


Ion	Ratio	Lower	Upper
152	100		
151	20.3	16.2	24.2
153	13.7	10.7	16.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#50
Dimethylphthalate
Concen: 51.156 ng
RT: 9.669 min Scan# 1271
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:163 Resp: 1178638

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.2	4.8
164	10.6	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 52.309 ng
RT: 9.727 min Scan# 1
Delta R.T. -0.001 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46


Tgt Ion:165 Resp: 23290


Ion	Ratio	Lower	Upper
165	100		
63	61.1	48.5	72.7
89	55.8	43.9	65.9

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#52
Acenaphthene
Concen: 49.912 ng
RT: 9.986 min Scan# 1325
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:154 Resp: 1024409

Ion	Ratio	Lower	Upper
154	100		
153	110.9	88.6	133.0
152	52.7	42.1	63.1






#53
3-Nitroaniline
Concen: 51.745 ng
RT: 9.898 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:138 Resp: 27960


Ion	Ratio	Lower	Upper
138	100		
108	13.4	11.7	17.5
92	111.0	91.2	136.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#54
2,4-Dinitrophenol
Concen: 50.725 ng
RT: 9.998 min Scan# 1327
Delta R.T. 0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:184 Resp: 62086

Ion	Ratio	Lower	Upper
184	100		
63	92.9	69.2	103.8
154	342.0	271.4	407.2






#55
Dibenzofuran
Concen: 50.123 ng
RT: 10.157 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:168 Resp: 154209


Ion	Ratio	Lower	Upper
168	100		
139	36.7	29.4	44.0
169	13.6	11.0	16.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#56
4-Nitrophenol
Concen: 58.029 ng
RT: 10.039 min Scan# 1334
Delta R.T. 0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:139 Resp: 210153

Ion	Ratio	Lower	Upper
139	100		
109	65.7	44.9	84.9
65	94.5	74.7	114.7






#57
2,4-Dinitrotoluene
Concen: 51.918 ng
RT: 10.127 min Scan# 1
Delta R.T. -0.001 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:165 Resp: 28019


Ion	Ratio	Lower	Upper
165	100		
63	50.2	38.8	58.2
89	72.8	57.1	85.7
182	2.8	2.2	3.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#58
Fluorene
Concen: 49.000 ng
RT: 10.498 min Scan# 1412
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:166 Resp: 1146269

Ion	Ratio	Lower	Upper
166	100		
165	93.2	73.9	110.9
167	13.6	10.8	16.2






#59
2,3,4,6-Tetrachlorophenol
Concen: 54.989 ng
RT: 10.269 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:232 Resp: 30078


Ion	Ratio	Lower	Upper
232	100		
131	44.4	35.9	53.9
130	2.3	1.8	2.6
166	29.3	24.0	36.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#60
Diethylphthalate
Concen: 51.459 ng
RT: 10.369 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:149 Resp: 1170793

Ion	Ratio	Lower	Upper
149	100		
177	21.7	17.6	26.4
150	12.8	10.2	15.2





#61
4-Chlorophenyl-phenylether
Concen: 49.802 ng
RT: 10.492 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46


Instrument :
BNA_F
ClientSampleId :
SSTDICC050


Tgt Ion:204 Resp: 55327

Ion	Ratio	Lower	Upper
204	100		
206	32.6	25.6	38.4
141	58.9	48.9	73.3

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#62
4-Nitroaniline
Concen: 52.547 ng
RT: 10.510 min Scan# 1414
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:138 Resp: 258117

Ion	Ratio	Lower	Upper
138	100		
92	49.0	29.2	69.2
108	69.4	49.1	89.1






#63
Azobenzene
Concen: 51.226 ng
RT: 10.651 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion: 77 Resp: 1150899


Ion	Ratio	Lower	Upper
77	100		
182	26.8	7.3	47.3
105	22.6	2.5	42.5
51	33.9	14.4	54.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:188 Resp: 604171

Ion	Ratio	Lower	Upper
188	100		
94	10.2	8.2	12.2
80	10.8	8.6	12.8





#65
4,6-Dinitro-2-methylphenol
Concen: 49.012 ng
RT: 10.539 min Scan# 14
Delta R.T. 0.005 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46


Tgt Ion:198 Resp: 9811


Ion	Ratio	Lower	Upper
198	100		
51	55.5	40.6	80.6
105	47.3	29.2	69.2

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#66
n-Nitrosodiphenylamine
Concen: 50.442 ng
RT: 10.610 min Scan# 1431
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:169 Resp: 1041277

Ion	Ratio	Lower	Upper
169	100		
168	65.9	53.0	79.6
167	36.1	28.8	43.2






#67
4-Bromophenyl-phenylether
Concen: 50.576 ng
RT: 10.980 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:248 Resp: 34029


Ion	Ratio	Lower	Upper
248	100		
250	100.0	77.2	115.8
141	75.5	58.4	87.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#68
Hexachlorobenzene
Concen: 51.064 ng
RT: 11.045 min Scan# 1505
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:284 Resp: 384807

Ion	Ratio	Lower	Upper
284	100		
142	31.4	24.6	37.0
249	30.0	23.2	34.8






#69
Atrazine
Concen: 52.819 ng
RT: 11.133 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:200 Resp: 28740


Ion	Ratio	Lower	Upper
200	100		
173	28.4	8.6	48.6
215	48.4	28.5	68.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#70
Pentachlorophenol
Concen: 55.114 ng
RT: 11.233 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:266 Resp: 219414

Ion	Ratio	Lower	Upper
266	100		
268	63.5	50.0	75.0
264	63.2	49.3	73.9





#71
Phenanthrene
Concen: 49.244 ng
RT: 11.463 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46


Tgt Ion:178 Resp: 160709


Ion	Ratio	Lower	Upper
178	100		
176	19.9	15.5	23.3
179	15.7	12.6	19.0

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#72
Anthracene
Concen: 49.744 ng
RT: 11.510 min Scan# 1584
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:178 Resp: 1659788

Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.3	22.9
179	15.9	12.7	19.1






#73
Carbazole
Concen: 49.436 ng
RT: 11.663 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:167 Resp: 150471


Ion	Ratio	Lower	Upper
167	100		
166	21.4	17.3	25.9
139	13.0	10.3	15.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#74
Di-n-butylphthalate
Concen: 51.573 ng
RT: 11.998 min Scan# 1667
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:149 Resp: 1578769

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.5	11.3
104	4.6	3.6	5.4






#75
Fluoranthene
Concen: 48.481 ng
RT: 12.651 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:202 Resp: 156399


Ion	Ratio	Lower	Upper
202	100		
101	11.7	0.0	32.4
203	17.9	0.0	37.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.074 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:240 Resp: 322957

Ion	Ratio	Lower	Upper
240	100		
120	12.6	10.1	15.1
236	25.3	20.5	30.7






#77
Benzidine
Concen: 55.590 ng
RT: 12.768 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:184 Resp: 76285


Ion	Ratio	Lower	Upper
184	100		
185	15.0	11.8	17.6
183	10.8	8.9	13.3



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#78
Pyrene
Concen: 52.484 ng
RT: 12.880 min Scan# 1817
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:202 Resp: 1563568

Ion	Ratio	Lower	Upper
202	100		
200	20.3	16.1	24.1
203	17.7	14.1	21.1






#79
Terphenyl-d14
Concen: 101.236 ng
RT: 13.021 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:244 Resp: 224798


Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.3	7.9
122	11.6	9.1	13.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#80
Butylbenzylphthalate
Concen: 51.685 ng
RT: 13.498 min Scan# 1922
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:149 Resp: 423529

Ion	Ratio	Lower	Upper
149	100		
91	67.9	54.4	81.6
206	20.3	15.9	23.9






#81
Benzo(a)anthracene
Concen: 51.691 ng
RT: 14.068 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:228 Resp: 1120699


Ion	Ratio	Lower	Upper
228	100		
226	27.3	21.9	32.9
229	19.8	16.0	24.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#82
3,3'-Dichlorobenzidine
Concen: 57.021 ng
RT: 14.027 min Scan# 2012
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:252 Resp: 363124

Ion	Ratio	Lower	Upper
252	100		
254	65.7	51.0	76.6
126	12.3	9.4	14.2






#83
Chrysene
Concen: 50.214 ng
RT: 14.104 min Scan# 2(
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:228 Resp: 99099


Ion	Ratio	Lower	Upper
228	100		
226	29.9	24.0	36.0
229	19.7	15.7	23.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 51.638 ng
RT: 14.057 min Scan# 2017
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:149 Resp: 585800

Ion	Ratio	Lower	Upper
149	100		
167	27.4	21.7	32.5
279	3.7	3.0	4.4





#85
Di-n-octyl phthalate
Concen: 51.078 ng
RT: 14.680 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46


Tgt Ion:149 Resp: 106562


Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.1	1.7
43	11.3	9.3	13.9

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.568 min Scan# 2274
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:264 Resp: 337726

Ion	Ratio	Lower	Upper
264	100		
260	23.0	18.4	27.6
265	21.4	17.3	25.9






#87
Indeno(1,2,3-cd)pyrene
Concen: 53.091 ng
RT: 17.080 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:276 Resp: 131333


Ion	Ratio	Lower	Upper
276	100		
138	29.0	22.6	33.8
277	25.1	20.1	30.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#88
Benzo(b)fluoranthene
Concen: 51.153 ng
RT: 15.127 min Scan# 2199
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:252 Resp: 1073096

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	11.2	9.0	13.4






#89
Benzo(k)fluoranthene
Concen: 50.408 ng
RT: 15.162 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:252 Resp: 96724


Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.8	26.8
125	10.0	9.0	13.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#90
Benzo(a)pyrene
Concen: 51.890 ng
RT: 15.504 min Scan# 2263
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:252 Resp: 996599

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.6	26.4
125	12.2	9.5	14.3






#91
Dibenzo(a,h)anthracene
Concen: 53.457 ng
RT: 17.103 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:278 Resp: 107324


Ion	Ratio	Lower	Upper
278	100		
139	18.2	15.2	22.8
279	23.7	19.0	28.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#92
Benzo(g,h,i)perylene
Concen: 52.984 ng
RT: 17.539 min Scan# 2609
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:276 Resp: 1069402

Ion	Ratio	Lower	Upper
276	100		
277	23.8	18.7	28.1
138	25.3	19.4	29.0



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142245.D
Acq On : 30 Apr 2025 14:15
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Apr 30 15:46:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	187963	20.000	ng	0.00
21) Naphthalene-d8	8.192	136	730767	20.000	ng	0.00
39) Acenaphthene-d10	9.951	164	379270	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	624729	20.000	ng	0.00
76) Chrysene-d12	14.080	240	328990	20.000	ng	0.00
86) Perylene-d12	15.568	264	361305	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	1364962	115.976	ng	0.00
7) Phenol-d6	6.528	99	1663210	117.026	ng	0.00
23) Nitrobenzene-d5	7.475	82	1488103	133.783	ng	0.00
42) 2,4,6-Tribromophenol	10.739	330	447862	129.531	ng	0.00
45) 2-Fluorobiphenyl	9.269	172	2777485	103.255	ng	0.00
79) Terphenyl-d14	13.022	244	2546694	112.585	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.699	88	312897	59.528	ng	99
3) Pyridine	3.469	79	823442	60.277	ng	98
4) n-Nitrosodimethylamine	3.422	42	431685	61.063	ng	98
6) Aniline	6.575	93	1184649	58.890	ng	100
8) 2-Chlorophenol	6.693	128	740327	60.874	ng	99
9) Benzaldehyde	6.463	77	559766	56.352	ng	98
10) Phenol	6.545	94	903976m	59.085	ng	
11) bis(2-Chloroethyl)ether	6.646	93	693997	58.532	ng	99
12) 1,3-Dichlorobenzene	6.851	146	787463	57.236	ng	99
13) 1,4-Dichlorobenzene	6.928	146	791509	57.102	ng	99
14) 1,2-Dichlorobenzene	7.081	146	764397	57.765	ng	99
15) Benzyl Alcohol	7.045	79	609871	60.504	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.187	45	1249716	58.370	ng	100
17) 2-Methylphenol	7.151	107	592171	59.438	ng	99
18) Hexachloroethane	7.428	117	275661	60.761	ng	99
19) n-Nitroso-di-n-propyla...	7.328	70	506389	57.660	ng	100
20) 3+4-Methylphenols	7.310	107	730966	58.670	ng	94
22) Acetophenone	7.322	105	943225	55.792	ng	97
24) Nitrobenzene	7.498	77	708847	64.948	ng	99
25) Isophorone	7.734	82	1371824	58.764	ng	100
26) 2-Nitrophenol	7.810	139	299107	61.132	ng	97
27) 2,4-Dimethylphenol	7.840	122	691268	58.731	ng	100
28) bis(2-Chloroethoxy)met...	7.940	93	847188	57.543	ng	100
29) 2,4-Dichlorophenol	8.045	162	596890	60.658	ng	99
30) 1,2,4-Trichlorobenzene	8.134	180	628720	57.256	ng	100
31) Naphthalene	8.216	128	2026390	55.603	ng	99
32) Benzoic acid	7.951	122	349577	59.198	ng	99
33) 4-Chloroaniline	8.263	127	868158	57.207	ng	99
34) Hexachlorobutadiene	8.334	225	373284	58.266	ng	100
35) Caprolactam	8.639	113	186124	62.928	ng	94
36) 4-Chloro-3-methylphenol	8.734	107	605133	59.887	ng	99
37) 2-Methylnaphthalene	8.904	142	1257822	56.136	ng	99
38) 1-Methylnaphthalene	9.004	142	1294318	55.760	ng	100
40) 1,2,4,5-Tetrachloroben...	9.069	216	607614	56.925	ng	100
41) Hexachlorocyclopentadiene	9.063	237	407387	67.248	ng	100
43) 2,4,6-Trichlorophenol	9.181	196	433323	65.514	ng	99

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142245.D
Acq On : 30 Apr 2025 14:15
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025

Quant Time: Apr 30 15:46:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.216	196	422145	60.705	ng	98
46) 1,1'-Biphenyl	9.369	154	1632628	55.372	ng	100
47) 2-Chloronaphthalene	9.398	162	1234670	56.283	ng	99
48) 2-Nitroaniline	9.486	65	357897	60.490	ng	99
49) Acenaphthylene	9.810	152	2079452	56.175	ng	99
50) Dimethylphthalate	9.675	163	1395869	57.529	ng	100
51) 2,6-Dinitrotoluene	9.728	165	289353	60.986	ng	98
52) Acenaphthene	9.986	154	1219341	56.414	ng	99
53) 3-Nitroaniline	9.898	138	352271	61.278	ng	97
54) 2,4-Dinitrophenol	9.998	184	86520	61.133	ng	83
55) Dibenzofuran	10.157	168	1798627	55.513	ng	99
56) 4-Nitrophenol	10.045	139	261987	68.694	ng	98
57) 2,4-Dinitrotoluene	10.134	165	352001	60.964	ng	98
58) Fluorene	10.498	166	1348902	54.754	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.269	232	367314	63.765	ng	99
60) Diethylphthalate	10.369	149	1382905	57.717	ng	100
61) 4-Chlorophenyl-phenyle...	10.492	204	648920	55.466	ng	97
62) 4-Nitroaniline	10.516	138	319292	61.243	ng	99
63) Azobenzene	10.651	77	1362906	57.603	ng	99
65) 4,6-Dinitro-2-methylph...	10.539	198	132189	60.646	ng	96
66) n-Nitrosodiphenylamine	10.610	169	1224189	57.351	ng	99
67) 4-Bromophenyl-phenylether	10.980	248	413069	59.371	ng	100
68) Hexachlorobenzene	11.045	284	457317	58.689	ng	99
69) Atrazine	11.133	200	340471	60.513	ng	99
70) Pentachlorophenol	11.233	266	270497	65.710	ng	98
71) Phenanthrene	11.463	178	1864413	55.249	ng	100
72) Anthracene	11.516	178	1920817	55.673	ng	100
73) Carbazole	11.663	167	1751538	55.651	ng	100
74) Di-n-butylphthalate	11.998	149	1852105	58.511	ng	100
75) Fluoranthene	12.651	202	1794425	53.793	ng	98
77) Benzidine	12.769	184	883568	63.206	ng	99
78) Pyrene	12.880	202	1809576	59.627	ng	99
80) Butylbenzylphthalate	13.498	149	516723	61.167	ng	99
81) Benzo(a)anthracene	14.069	228	1299270	58.829	ng	100
82) 3,3'-Dichlorobenzidine	14.027	252	444901	68.582	ng	99
83) Chrysene	14.104	228	1181729	58.781	ng	99
84) Bis(2-ethylhexyl)phtha...	14.057	149	712716	61.002	ng	99
85) Di-n-octyl phthalate	14.680	149	1329733	61.204	ng	99
87) Indeno(1,2,3-cd)pyrene	17.086	276	1614650	61.012	ng	100
88) Benzo(b)fluoranthene	15.127	252	1275552	56.836	ng	99
89) Benzo(k)fluoranthene	15.163	252	1202952	58.601	ng	99
90) Benzo(a)pyrene	15.510	252	1232454	59.982	ng	100
91) Dibenzo(a,h)anthracene	17.110	278	1292219	60.163	ng	99
92) Benzo(g,h,i)perylene	17.545	276	1315066	60.904	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142245.D
Acq On : 30 Apr 2025 14:15
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1


Quant Time: Apr 30 15:46:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:152 Resp: 18796


Ion	Ratio	Lower	Upper
152	100		
150	160.9	130.2	195.2
115	59.8	47.0	70.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#2
1,4-Dioxane
Concen: 59.528 ng
RT: 2.699 min Scan# 86
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion: 88 Resp: 312897

Ion	Ratio	Lower	Upper
88	100		
58	90.8	72.2	108.2
43	35.0	27.0	40.6






#3
Pyridine
Concen: 60.277 ng
RT: 3.469 min Scan# 21
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion: 79 Resp: 82344


Ion	Ratio	Lower	Upper
79	100		
52	87.7	71.5	107.3
51	41.9	34.5	51.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#4
n-Nitrosodimethylamine
Concen: 61.063 ng
RT: 3.422 min Scan# 209
Delta R.T. 0.012 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion: 42 Resp: 431685

Ion	Ratio	Lower	Upper
42	100		
74	113.5	89.5	134.3
44	7.5	6.6	9.8





#5
2-Fluorophenol
Concen: 115.976 ng
RT: 5.516 min Scan# 5
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15


Tgt Ion:112 Resp: 136496


Ion	Ratio	Lower	Upper
112	100		
64	61.6	48.4	72.6
63	31.5	24.8	37.2

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#6
Aniline
Concen: 58.890 ng
RT: 6.575 min Scan# 745
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion: 93 Resp: 1184649

Ion	Ratio	Lower	Upper
93	100		
66	40.1	32.1	48.1
65	20.2	16.6	24.8






#7
Phenol-d6
Concen: 117.026 ng
RT: 6.528 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion: 99 Resp: 166321


Ion	Ratio	Lower	Upper
99	100		
42	20.7	16.3	24.5
71	32.3	25.0	37.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#8
2-Chlorophenol
Concen: 60.874 ng
RT: 6.693 min Scan# 765
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:128 Resp: 740327

Ion	Ratio	Lower	Upper
128	100		
130	32.8	13.1	53.1
64	49.4	28.6	68.6






#9
Benzaldehyde
Concen: 56.352 ng
RT: 6.463 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion: 77 Resp: 55976


Ion	Ratio	Lower	Upper
77	100		
105	100.6	79.1	119.1
106	98.1	76.0	116.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#10
Phenol
Concen: 59.085 ng m
RT: 6.545 min Scan# 740
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion: 94 Resp: 903976

Ion	Ratio	Lower	Upper
94	100		
65	29.0	8.9	48.9
66	39.2	18.7	58.7






#11
bis(2-Chloroethyl)ether
Concen: 58.532 ng
RT: 6.646 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion: 93 Resp: 69399


Ion	Ratio	Lower	Upper
93	100		
63	83.0	62.2	102.2
95	32.3	12.7	52.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#12
1,3-Dichlorobenzene
Concen: 57.236 ng
RT: 6.851 min Scan# 792
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:146 Resp: 787463

Ion	Ratio	Lower	Upper
146	100		
148	63.6	50.5	75.7
75	30.1	23.2	34.8






#13
1,4-Dichlorobenzene
Concen: 57.102 ng
RT: 6.928 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:146 Resp: 79150


Ion	Ratio	Lower	Upper
146	100		
148	64.1	50.6	76.0
111	42.7	33.4	50.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#14
1,2-Dichlorobenzene
Concen: 57.765 ng
RT: 7.081 min Scan# 831
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:146 Resp: 764397

Ion	Ratio	Lower	Upper
146	100		
148	63.9	50.5	75.7
111	44.2	34.9	52.3






#15
Benzyl Alcohol
Concen: 60.504 ng
RT: 7.045 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion: 79 Resp: 60987


Ion	Ratio	Lower	Upper
79	100		
108	79.3	64.5	96.7
77	61.4	49.2	73.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 58.370 ng
RT: 7.187 min Scan# 849
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion: 45 Resp: 1249716

Ion	Ratio	Lower	Upper
45	100		
77	11.2	0.0	31.3
79	8.5	0.0	28.5






#17
2-Methylphenol
Concen: 59.438 ng
RT: 7.151 min Scan# 84
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:107 Resp: 59217


Ion	Ratio	Lower	Upper
107	100		
108	113.3	91.0	136.4
77	43.7	34.4	51.6
79	43.5	34.4	51.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#18
Hexachloroethane
Concen: 60.761 ng
RT: 7.428 min Scan# 890
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:117 Resp: 275661

Ion	Ratio	Lower	Upper
117	100		
119	96.1	77.1	115.7
201	97.0	79.4	119.2






#19
n-Nitroso-di-n-propylamine
Concen: 57.660 ng
RT: 7.328 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion: 70 Resp: 50638


Ion	Ratio	Lower	Upper
70	100		
42	66.9	53.4	80.2
101	10.0	7.7	11.5
130	21.9	17.2	25.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#20
3+4-Methylphenols
Concen: 58.670 ng
RT: 7.310 min Scan# 870
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:107 Resp: 730966

Ion	Ratio	Lower	Upper
107	100		
108	90.6	71.2	111.2
77	73.0	41.1	81.1
79	28.2	8.1	48.1






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.192 min Scan# 1
Delta R.T. -0.001 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:136 Resp: 73076


Ion	Ratio	Lower	Upper
136	100		
137	11.2	8.7	13.1
54	9.7	7.7	11.5
68	6.3	5.0	7.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#22
Acetophenone
Concen: 55.792 ng
RT: 7.322 min Scan# 872
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:105 Resp: 943225

Ion	Ratio	Lower	Upper
105	100		
71	6.5	5.7	8.5
51	34.7	30.1	45.1
120	23.8	19.0	28.4





#23
Nitrobenzene-d5
Concen: 133.783 ng
RT: 7.475 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15


Tgt Ion: 82 Resp: 148810


Ion	Ratio	Lower	Upper
82	100		
128	46.0	37.0	55.6
54	60.6	48.0	72.0

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#24
Nitrobenzene
Concen: 64.948 ng
RT: 7.498 min Scan# 902
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion: 77 Resp: 708847

Ion	Ratio	Lower	Upper
77	100		
123	48.8	38.6	58.0
65	14.4	11.5	17.3






#25
Isophorone
Concen: 58.764 ng
RT: 7.734 min Scan# 94
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion: 82 Resp: 137182


Ion	Ratio	Lower	Upper
82	100		
95	7.2	5.6	8.4
138	19.2	15.3	22.9



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#26
2-Nitrophenol
Concen: 61.132 ng
RT: 7.810 min Scan# 955
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:139 Resp: 299107

Ion	Ratio	Lower	Upper
139	100		
109	26.5	21.8	32.8
65	49.1	41.1	61.7






#27
2,4-Dimethylphenol
Concen: 58.731 ng
RT: 7.840 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:122 Resp: 69126


Ion	Ratio	Lower	Upper
122	100		
107	107.8	86.0	129.0
121	57.5	46.3	69.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#28
bis(2-Chloroethoxy)methane
Concen: 57.543 ng
RT: 7.940 min Scan# 977
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion: 93 Resp: 847188

Ion	Ratio	Lower	Upper
93	100		
95	32.9	26.2	39.2
123	12.3	9.7	14.5





#29
2,4-Dichlorophenol
Concen: 60.658 ng
RT: 8.045 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15


Tgt Ion:162 Resp: 59689


Ion	Ratio	Lower	Upper
162	100		
164	64.4	43.0	83.0
98	36.7	16.7	56.7

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#30
1,2,4-Trichlorobenzene
Concen: 57.256 ng
RT: 8.134 min Scan# 1010
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:180 Resp: 628720

Ion	Ratio	Lower	Upper
180	100		
182	94.6	75.7	113.5
145	32.1	25.4	38.0






#31
Naphthalene
Concen: 55.603 ng
RT: 8.216 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:128 Resp: 202639


Ion	Ratio	Lower	Upper
128	100		
129	11.6	9.0	13.4
127	13.2	10.6	15.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#32
Benzoic acid
Concen: 59.198 ng
RT: 7.951 min Scan# 979
Delta R.T. 0.017 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:122 Resp: 349577

Ion	Ratio	Lower	Upper
122	100		
105	118.0	98.2	138.2
77	88.8	66.8	106.8






#33
4-Chloroaniline
Concen: 57.207 ng
RT: 8.263 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:127 Resp: 86815


Ion	Ratio	Lower	Upper
127	100		
129	32.8	26.1	39.1
65	30.5	25.5	38.3
92	18.8	15.4	23.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#34
Hexachlorobutadiene
Concen: 58.266 ng
RT: 8.334 min Scan# 1044
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:225 Resp: 373284

Ion	Ratio	Lower	Upper
225	100		
223	62.5	49.7	74.5
227	63.6	51.0	76.4






#35
Caprolactam
Concen: 62.928 ng
RT: 8.639 min Scan# 1
Delta R.T. 0.018 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:113 Resp: 18612


Ion	Ratio	Lower	Upper
113	100		
55	189.5	177.7	217.7
56	139.8	127.1	167.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#36
4-Chloro-3-methylphenol
Concen: 59.887 ng
RT: 8.734 min Scan# 1112
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:107 Resp: 605133

Ion	Ratio	Lower	Upper
107	100		
144	26.6	21.0	31.4
142	83.2	66.1	99.1






#37
2-Methylnaphthalene
Concen: 56.136 ng
RT: 8.904 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:142 Resp: 125782


Ion	Ratio	Lower	Upper
142	100		
141	87.6	70.5	105.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#38
1-Methylnaphthalene
Concen: 55.760 ng
RT: 9.004 min Scan# 1158
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:142 Resp: 1294318

Ion	Ratio	Lower	Upper
142	100		
141	90.6	72.8	109.2






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.951 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:164 Resp: 37927


Ion	Ratio	Lower	Upper
164	100		
162	99.2	78.6	117.8
160	44.3	35.7	53.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 56.925 ng
RT: 9.069 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:216 Resp: 607614

Ion	Ratio	Lower	Upper
216	100		
214	79.4	63.8	95.6
179	20.9	16.7	25.1
108	19.1	15.6	23.4





#41
Hexachlorocyclopentadiene
Concen: 67.248 ng
RT: 9.063 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15


Tgt Ion:237 Resp: 40738


Ion	Ratio	Lower	Upper
237	100		
235	63.6	43.5	83.5
272	13.1	0.0	32.9

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#42
2,4,6-Tribromophenol
Concen: 129.531 ng
RT: 10.739 min Scan# 1453
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:330 Resp: 447862

Ion	Ratio	Lower	Upper
330	100		
332	96.2	76.5	114.7
141	29.9	24.3	36.5






#43
2,4,6-Trichlorophenol
Concen: 65.514 ng
RT: 9.181 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:196 Resp: 43332


Ion	Ratio	Lower	Upper
196	100		
198	95.9	77.4	116.2
200	30.8	24.6	37.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#44
2,4,5-Trichlorophenol
Concen: 60.705 ng
RT: 9.216 min Scan# 1194
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:196 Resp: 422145

Ion	Ratio	Lower	Upper
196	100		
198	95.8	75.2	112.8
97	48.1	39.8	59.8
132	29.4	23.8	35.8






#45
2-Fluorobiphenyl
Concen: 103.255 ng
RT: 9.269 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:172 Resp: 277748


Ion	Ratio	Lower	Upper
172	100		
171	36.5	29.0	43.4
170	24.4	19.4	29.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#46
1,1'-Biphenyl
Concen: 55.372 ng
RT: 9.369 min Scan# 1220
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:154 Resp: 1632628

Ion	Ratio	Lower	Upper
154	100		
153	40.7	20.5	60.5
76	13.3	0.0	32.8






#47
2-Chloronaphthalene
Concen: 56.283 ng
RT: 9.398 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:162 Resp: 123467


Ion	Ratio	Lower	Upper
162	100		
127	41.6	32.6	49.0
164	33.2	26.1	39.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#48
2-Nitroaniline
Concen: 60.490 ng
RT: 9.486 min Scan# 1240
Delta R.T. -0.001 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion: 65 Resp: 357897

Ion	Ratio	Lower	Upper
65	100		
92	68.1	54.3	81.5
138	111.9	88.6	133.0






#49
Acenaphthylene
Concen: 56.175 ng
RT: 9.810 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:152 Resp: 207945


Ion	Ratio	Lower	Upper
152	100		
151	20.5	16.2	24.2
153	13.6	10.7	16.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#50
Dimethylphthalate
Concen: 57.529 ng
RT: 9.675 min Scan# 1272
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:163 Resp: 1395869

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.2	4.8
164	10.2	8.2	12.2






#51
2,6-Dinitrotoluene
Concen: 60.986 ng
RT: 9.728 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:165 Resp: 28935


Ion	Ratio	Lower	Upper
165	100		
63	62.4	48.5	72.7
89	55.4	43.9	65.9



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#52
Acenaphthene
Concen: 56.414 ng
RT: 9.986 min Scan# 1325
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:154 Resp: 1219341

Ion	Ratio	Lower	Upper
154	100		
153	109.8	88.6	133.0
152	52.4	42.1	63.1





#53
3-Nitroaniline
Concen: 61.278 ng
RT: 9.898 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15


Tgt Ion:138 Resp: 35227


Ion	Ratio	Lower	Upper
138	100		
108	12.4	11.7	17.5
92	110.7	91.2	136.8

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#54
2,4-Dinitrophenol
Concen: 61.133 ng
RT: 9.998 min Scan# 1327
Delta R.T. 0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:184 Resp: 86520

Ion	Ratio	Lower	Upper
184	100		
63	93.7	69.2	103.8
154	379.9	271.4	407.2






#55
Dibenzofuran
Concen: 55.513 ng
RT: 10.157 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:168 Resp: 179862


Ion	Ratio	Lower	Upper
168	100		
139	37.1	29.4	44.0
169	13.9	11.0	16.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#56
4-Nitrophenol
Concen: 68.694 ng
RT: 10.045 min Scan# 1335
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:139 Resp: 261987

Ion	Ratio	Lower	Upper
139	100		
109	64.4	44.9	84.9
65	91.8	74.7	114.7






#57
2,4-Dinitrotoluene
Concen: 60.964 ng
RT: 10.134 min Scan# 1
Delta R.T. 0.005 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:165 Resp: 35200


Ion	Ratio	Lower	Upper
165	100		
63	47.5	38.8	58.2
89	70.1	57.1	85.7
182	2.6	2.2	3.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#58
Fluorene
Concen: 54.754 ng
RT: 10.498 min Scan# 1412
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:166 Resp: 1348902

Ion	Ratio	Lower	Upper
166	100		
165	92.8	73.9	110.9
167	13.8	10.8	16.2






#59
2,3,4,6-Tetrachlorophenol
Concen: 63.765 ng
RT: 10.269 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:232 Resp: 36731


Ion	Ratio	Lower	Upper
232	100		
131	44.2	35.9	53.9
130	2.2	1.8	2.6
166	29.4	24.0	36.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#60
Diethylphthalate
Concen: 57.717 ng
RT: 10.369 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:149 Resp: 1382905

Ion	Ratio	Lower	Upper
149	100		
177	21.8	17.6	26.4
150	12.8	10.2	15.2





#61
4-Chlorophenyl-phenylether
Concen: 55.466 ng
RT: 10.492 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15


Instrument :
BNA_F
ClientSampleId :
SSTDICC060


Tgt Ion:204 Resp: 64892

Ion	Ratio	Lower	Upper
204	100		
206	33.2	25.6	38.4
141	58.4	48.9	73.3

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#62
4-Nitroaniline
Concen: 61.243 ng
RT: 10.516 min Scan# 1415
Delta R.T. 0.012 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:138 Resp: 319292

Ion	Ratio	Lower	Upper
138	100		
92	48.4	29.2	69.2
108	68.2	49.1	89.1






#63
Azobenzene
Concen: 57.603 ng
RT: 10.651 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion: 77 Resp: 136290


Ion	Ratio	Lower	Upper
77	100		
182	26.7	7.3	47.3
105	22.5	2.5	42.5
51	33.9	14.4	54.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:188 Resp: 624729

Ion	Ratio	Lower	Upper
188	100		
94	10.4	8.2	12.2
80	11.0	8.6	12.8





#65
4,6-Dinitro-2-methylphenol
Concen: 60.646 ng
RT: 10.539 min Scan# 14
Delta R.T. 0.005 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15


Tgt Ion:198 Resp: 13218


Ion	Ratio	Lower	Upper
198	100		
51	57.5	40.6	80.6
105	45.8	29.2	69.2

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#66
n-Nitrosodiphenylamine
Concen: 57.351 ng
RT: 10.610 min Scan# 1431
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:169 Resp: 1224189

Ion	Ratio	Lower	Upper
169	100		
168	66.7	53.0	79.6
167	36.3	28.8	43.2





#67
4-Bromophenyl-phenylether
Concen: 59.371 ng
RT: 10.980 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15


Tgt Ion:248 Resp: 41306


Ion	Ratio	Lower	Upper
248	100		
250	97.0	77.2	115.8
141	73.0	58.4	87.6

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#68
Hexachlorobenzene
Concen: 58.689 ng
RT: 11.045 min Scan# 1505
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:284 Resp: 457317

Ion	Ratio	Lower	Upper
284	100		
142	31.6	24.6	37.0
249	29.6	23.2	34.8






#69
Atrazine
Concen: 60.513 ng
RT: 11.133 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:200 Resp: 34047


Ion	Ratio	Lower	Upper
200	100		
173	28.7	8.6	48.6
215	49.3	28.5	68.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#70
Pentachlorophenol
Concen: 65.710 ng
RT: 11.233 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:266 Resp: 270497

Ion	Ratio	Lower	Upper
266	100		
268	64.1	50.0	75.0
264	63.1	49.3	73.9





#71
Phenanthrene
Concen: 55.249 ng
RT: 11.463 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15


Tgt Ion:178 Resp: 186441


Ion	Ratio	Lower	Upper
178	100		
176	19.7	15.5	23.3
179	15.7	12.6	19.0

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#72
Anthracene
Concen: 55.673 ng
RT: 11.516 min Scan# 1585
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:178 Resp: 1920817

Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.3	22.9
179	16.0	12.7	19.1






#73
Carbazole
Concen: 55.651 ng
RT: 11.663 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:167 Resp: 175153


Ion	Ratio	Lower	Upper
167	100		
166	21.5	17.3	25.9
139	12.9	10.3	15.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#74
Di-n-butylphthalate
Concen: 58.511 ng
RT: 11.998 min Scan# 1667
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:149 Resp: 1852105

Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.5	11.3
104	4.6	3.6	5.4






#75
Fluoranthene
Concen: 53.793 ng
RT: 12.651 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:202 Resp: 179442


Ion	Ratio	Lower	Upper
202	100		
101	11.7	0.0	32.4
203	18.0	0.0	37.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.080 min Scan# 2021
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:240 Resp: 328990

Ion	Ratio	Lower	Upper
240	100		
120	11.4	10.1	15.1
236	25.2	20.5	30.7






#77
Benzidine
Concen: 63.206 ng
RT: 12.769 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:184 Resp: 88356


Ion	Ratio	Lower	Upper
184	100		
185	15.2	11.8	17.6
183	10.8	8.9	13.3



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#78
Pyrene
Concen: 59.627 ng
RT: 12.880 min Scan# 1817
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:202 Resp: 1809576

Ion	Ratio	Lower	Upper
202	100		
200	20.5	16.1	24.1
203	17.8	14.1	21.1






#79
Terphenyl-d14
Concen: 112.585 ng
RT: 13.022 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:244 Resp: 254669


Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.3	7.9
122	11.5	9.1	13.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#80
Butylbenzylphthalate
Concen: 61.167 ng
RT: 13.498 min Scan# 1922
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:149 Resp: 516723

Ion	Ratio	Lower	Upper
149	100		
91	68.3	54.4	81.6
206	20.6	15.9	23.9






#81
Benzo(a)anthracene
Concen: 58.829 ng
RT: 14.069 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:228 Resp: 129927


Ion	Ratio	Lower	Upper
228	100		
226	27.5	21.9	32.9
229	20.0	16.0	24.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#82
3,3'-Dichlorobenzidine
Concen: 68.582 ng
RT: 14.027 min Scan# 2012
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:252 Resp: 444901

Ion	Ratio	Lower	Upper
252	100		
254	64.5	51.0	76.6
126	12.4	9.4	14.2






#83
Chrysene
Concen: 58.781 ng
RT: 14.104 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:228 Resp: 118172


Ion	Ratio	Lower	Upper
228	100		
226	30.6	24.0	36.0
229	19.7	15.7	23.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 61.002 ng
RT: 14.057 min Scan# 2017
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:149 Resp: 712716

Ion	Ratio	Lower	Upper
149	100		
167	26.8	21.7	32.5
279	3.8	3.0	4.4






#85
Di-n-octyl phthalate
Concen: 61.204 ng
RT: 14.680 min Scan# 21
Delta R.T. 0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:149 Resp: 132973


Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.1	1.7
43	11.3	9.3	13.9



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 15.568 min Scan# 2274
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:264 Resp: 361305

Ion	Ratio	Lower	Upper
264	100		
260	23.6	18.4	27.6
265	21.9	17.3	25.9






#87
Indeno(1,2,3-cd)pyrene
Concen: 61.012 ng
RT: 17.086 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:276 Resp: 161465


Ion	Ratio	Lower	Upper
276	100		
138	28.5	22.6	33.8
277	25.1	20.1	30.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#88
Benzo(b)fluoranthene
Concen: 56.836 ng
RT: 15.127 min Scan# 2199
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:252 Resp: 1275552

Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.7	26.5
125	11.5	9.0	13.4






#89
Benzo(k)fluoranthene
Concen: 58.601 ng
RT: 15.163 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:252 Resp: 120295


Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.8	26.8
125	10.8	9.0	13.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#90
Benzo(a)pyrene
Concen: 59.982 ng
RT: 15.510 min Scan# 2264
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:252 Resp: 1232454

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.6	26.4
125	11.5	9.5	14.3






#91
Dibenzo(a,h)anthracene
Concen: 60.163 ng
RT: 17.110 min Scan# 2
Delta R.T. 0.012 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:278 Resp: 129221


Ion	Ratio	Lower	Upper
278	100		
139	18.4	15.2	22.8
279	23.9	19.0	28.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#92
Benzo(g,h,i)perylene
Concen: 60.904 ng
RT: 17.545 min Scan# 2610
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:276 Resp: 1315066

Ion	Ratio	Lower	Upper
276	100		
277	23.2	18.7	28.1
138	24.7	19.4	29.0



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142246.D
Acq On : 30 Apr 2025 14:43
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Quant Time: Apr 30 15:46:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.910	152	192490	20.000	ng		0.00
21) Naphthalene-d8	8.192	136	730774	20.000	ng		0.00
39) Acenaphthene-d10	9.951	164	381378	20.000	ng		0.00
64) Phenanthrene-d10	11.433	188	608907	20.000	ng		0.00
76) Chrysene-d12	14.080	240	323925	20.000	ng		0.00
86) Perylene-d12	15.568	264	374102	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.522	112	1741015	144.449	ng		0.00
7) Phenol-d6	6.534	99	2109624	144.946	ng		0.00
23) Nitrobenzene-d5	7.481	82	1929643	173.477	ng		0.00
42) 2,4,6-Tribromophenol	10.739	330	579431	166.657	ng		0.00
45) 2-Fluorobiphenyl	9.275	172	3399868	125.694	ng		0.00
79) Terphenyl-d14	13.027	244	2993632	134.413	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.699	88	408111	75.816	ng		99
3) Pyridine	3.469	79	1062667	75.959	ng		99
4) n-Nitrosodimethylamine	3.434	42	562539	77.702	ng		98
6) Aniline	6.575	93	1525082	74.030	ng		99
8) 2-Chlorophenol	6.692	128	942996	75.716	ng		99
9) Benzaldehyde	6.463	77	673391	66.197	ng		99
10) Phenol	6.551	94	1130209m	72.134	ng		
11) bis(2-Chloroethyl)ether	6.651	93	871905	71.808	ng		99
12) 1,3-Dichlorobenzene	6.851	146	1000058	70.979	ng		99
13) 1,4-Dichlorobenzene	6.928	146	997448	70.267	ng		99
14) 1,2-Dichlorobenzene	7.081	146	966587	71.326	ng		99
15) Benzyl Alcohol	7.051	79	787715	76.310	ng		99
16) 2,2'-oxybis(1-Chloropr...	7.186	45	1573151	71.749	ng		99
17) 2-Methylphenol	7.157	107	757239	74.219	ng		100
18) Hexachloroethane	7.428	117	347325	74.757	ng		99
19) n-Nitroso-di-n-propyla...	7.334	70	650174	72.291	ng		99
20) 3+4-Methylphenols	7.316	107	897475	70.340	ng	#	87
22) Acetophenone	7.328	105	1180939	69.853	ng		97
24) Nitrobenzene	7.498	77	917286	84.046	ng		100
25) Isophorone	7.739	82	1758014	75.306	ng		99
26) 2-Nitrophenol	7.810	139	416387	79.254	ng		98
27) 2,4-Dimethylphenol	7.839	122	883922	75.098	ng		100
28) bis(2-Chloroethoxy)met...	7.945	93	1070383	72.702	ng		99
29) 2,4-Dichlorophenol	8.045	162	768570	78.103	ng		99
30) 1,2,4-Trichlorobenzene	8.133	180	797476	72.623	ng		100
31) Naphthalene	8.216	128	2508082	68.820	ng		99
32) Benzoic acid	7.969	122	523215m	83.328	ng		
33) 4-Chloroaniline	8.263	127	1084543	71.465	ng		99
34) Hexachlorobutadiene	8.333	225	471301	73.564	ng		100
35) Caprolactam	8.645	113	242414	81.958	ng		97
36) 4-Chloro-3-methylphenol	8.739	107	776483	76.844	ng		99
37) 2-Methylnaphthalene	8.910	142	1559858	69.615	ng		99
38) 1-Methylnaphthalene	9.010	142	1583955	68.237	ng		99
40) 1,2,4,5-Tetrachloroben...	9.075	216	757210	70.548	ng		100
41) Hexachlorocyclopentadiene	9.063	237	523186	85.886	ng		99
43) 2,4,6-Trichlorophenol	9.180	196	545358	81.996	ng		98

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142246.D
Acq On : 30 Apr 2025 14:43
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Quant Time: Apr 30 15:46:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.216	196	557542	79.732	ng		99
46) 1,1'-Biphenyl	9.375	154	2006955	67.691	ng		99
47) 2-Chloronaphthalene	9.398	162	1547207	70.140	ng		99
48) 2-Nitroaniline	9.486	65	481954	79.390	ng		97
49) Acenaphthylene	9.816	152	2555385	68.650	ng		99
50) Dimethylphthalate	9.675	163	1757977	72.052	ng		99
51) 2,6-Dinitrotoluene	9.733	165	379127	78.246	ng		97
52) Acenaphthene	9.986	154	1517991	69.843	ng		99
53) 3-Nitroaniline	9.904	138	456571	78.061	ng		96
54) 2,4-Dinitrophenol	10.004	184	129817	79.532	ng	#	28
55) Dibenzofuran	10.157	168	2237085	68.664	ng		98
56) 4-Nitrophenol	10.045	139	335015	87.357	ng		97
57) 2,4-Dinitrotoluene	10.133	165	467545	78.915	ng		99
58) Fluorene	10.498	166	1634486	65.980	ng		99
59) 2,3,4,6-Tetrachlorophenol	10.269	232	465050	80.286	ng		99
60) Diethylphthalate	10.375	149	1717661	71.292	ng		99
61) 4-Chlorophenyl-phenyle...	10.492	204	802224	68.190	ng		98
62) 4-Nitroaniline	10.522	138	410928	77.614	ng		98
63) Azobenzene	10.651	77	1674434	70.379	ng		99
65) 4,6-Dinitro-2-methylph...	10.545	198	183184	81.749	ng		96
66) n-Nitrosodiphenylamine	10.610	169	1518420	72.983	ng		99
67) 4-Bromophenyl-phenylether	10.980	248	514462	75.866	ng		97
68) Hexachlorobenzene	11.045	284	577203	75.999	ng		98
69) Atrazine	11.133	200	423886	77.296	ng		99
70) Pentachlorophenol	11.233	266	350009	87.235	ng		99
71) Phenanthrene	11.463	178	2288481	69.578	ng		99
72) Anthracene	11.516	178	2336551	69.482	ng		100
73) Carbazole	11.663	167	2087769	68.058	ng		99
74) Di-n-butylphthalate	11.998	149	2221417	72.002	ng		100
75) Fluoranthene	12.651	202	2137636	65.747	ng		98
77) Benzidine	12.768	184	1049467	76.247	ng		99
78) Pyrene	12.880	202	2118847	70.910	ng		99
80) Butylbenzylphthalate	13.498	149	659574	78.197	ng		98
81) Benzo(a)anthracene	14.068	228	1586778	72.970	ng		100
82) 3,3'-Dichlorobenzidine	14.033	252	563738	88.259	ng		99
83) Chrysene	14.104	228	1474094	74.470	ng		99
84) Bis(2-ethylhexyl)phtha...	14.057	149	915684	78.547	ng		100
85) Di-n-octyl phthalate	14.680	149	1722215	78.594	ng		100
87) Indeno(1,2,3-cd)pyrene	17.092	276	2128475	77.677	ng		99
88) Benzo(b)fluoranthene	15.133	252	1798083	77.378	ng		99
89) Benzo(k)fluoranthene	15.162	252	1439304	67.716	ng		100
90) Benzo(a)pyrene	15.509	252	1610193	75.685	ng		99
91) Dibenzo(a,h)anthracene	17.115	278	1712182	76.989	ng		99
92) Benzo(g,h,i)perylene	17.551	276	1738450	77.757	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142246.D
Acq On : 30 Apr 2025 14:43
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1


Quant Time: Apr 30 15:46:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:152 Resp: 19249


Ion	Ratio	Lower	Upper
152	100		
150	159.4	130.2	195.2
115	58.9	47.0	70.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#2
1,4-Dioxane
Concen: 75.816 ng
RT: 2.699 min Scan# 86
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion: 88 Resp: 408111

Ion	Ratio	Lower	Upper
88	100		
58	90.8	72.2	108.2
43	34.2	27.0	40.6






#3
Pyridine
Concen: 75.959 ng
RT: 3.469 min Scan# 21
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion: 79 Resp: 106266


Ion	Ratio	Lower	Upper
79	100		
52	88.8	71.5	107.3
51	42.4	34.5	51.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#4
n-Nitrosodimethylamine
Concen: 77.702 ng
RT: 3.434 min Scan# 211
Delta R.T. 0.023 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion: 42 Resp: 562539

Ion	Ratio	Lower	Upper
42	100		
74	113.7	89.5	134.3
44	7.5	6.6	9.8






#5
2-Fluorophenol
Concen: 144.449 ng
RT: 5.522 min Scan# 5
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:112 Resp: 174101


Ion	Ratio	Lower	Upper
112	100		
64	60.3	48.4	72.6
63	31.1	24.8	37.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#6
Aniline
Concen: 74.030 ng
RT: 6.575 min Scan# 745
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion: 93 Resp: 1525082

Ion	Ratio	Lower	Upper
93	100		
66	40.8	32.1	48.1
65	20.5	16.6	24.8






#7
Phenol-d6
Concen: 144.946 ng
RT: 6.534 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion: 99 Resp: 210962


Ion	Ratio	Lower	Upper
99	100		
42	20.7	16.3	24.5
71	32.3	25.0	37.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#8
2-Chlorophenol
Concen: 75.716 ng
RT: 6.692 min Scan# 765
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:128 Resp: 942996

Ion	Ratio	Lower	Upper
128	100		
130	32.6	13.1	53.1
64	49.8	28.6	68.6






#9
Benzaldehyde
Concen: 66.197 ng
RT: 6.463 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion: 77 Resp: 67339


Ion	Ratio	Lower	Upper
77	100		
105	100.2	79.1	119.1
106	97.0	76.0	116.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#10
Phenol
Concen: 72.134 ng m
RT: 6.551 min Scan# 741
Delta R.T. 0.012 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion: 94 Resp: 1130209

Ion	Ratio	Lower	Upper
94	100		
65	30.1	8.9	48.9
66	41.8	18.7	58.7






#11
bis(2-Chloroethyl)ether
Concen: 71.808 ng
RT: 6.651 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion: 93 Resp: 87190


Ion	Ratio	Lower	Upper
93	100		
63	82.9	62.2	102.2
95	32.7	12.7	52.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#12
1,3-Dichlorobenzene
Concen: 70.979 ng
RT: 6.851 min Scan# 792
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:146 Resp: 1000058

Ion	Ratio	Lower	Upper
146	100		
148	63.6	50.5	75.7
75	30.3	23.2	34.8






#13
1,4-Dichlorobenzene
Concen: 70.267 ng
RT: 6.928 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:146 Resp: 99744


Ion	Ratio	Lower	Upper
146	100		
148	63.7	50.6	76.0
111	42.7	33.4	50.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#14
1,2-Dichlorobenzene
Concen: 71.326 ng
RT: 7.081 min Scan# 831
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:146 Resp: 966587

Ion	Ratio	Lower	Upper
146	100		
148	63.8	50.5	75.7
111	44.6	34.9	52.3






#15
Benzyl Alcohol
Concen: 76.310 ng
RT: 7.051 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion: 79 Resp: 78771


Ion	Ratio	Lower	Upper
79	100		
108	81.5	64.5	96.7
77	62.3	49.2	73.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 71.749 ng
RT: 7.186 min Scan# 849
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion: 45 Resp: 1573151

Ion	Ratio	Lower	Upper
45	100		
77	11.5	0.0	31.3
79	8.8	0.0	28.5






#17
2-Methylphenol
Concen: 74.219 ng
RT: 7.157 min Scan# 84
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:107 Resp: 75723


Ion	Ratio	Lower	Upper
107	100		
108	113.7	91.0	136.4
77	43.0	34.4	51.6
79	43.6	34.4	51.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#18
Hexachloroethane
Concen: 74.757 ng
RT: 7.428 min Scan# 890
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:117 Resp: 347325

Ion	Ratio	Lower	Upper
117	100		
119	96.3	77.1	115.7
201	97.4	79.4	119.2






#19
n-Nitroso-di-n-propylamine
Concen: 72.291 ng
RT: 7.334 min Scan# 8
Delta R.T. 0.012 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion: 70 Resp: 65017


Ion	Ratio	Lower	Upper
70	100		
42	66.3	53.4	80.2
101	10.0	7.7	11.5
130	22.1	17.2	25.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#20
3+4-Methylphenols
Concen: 70.340 ng
RT: 7.316 min Scan# 871
Delta R.T. 0.012 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:107 Resp: 897475

Ion	Ratio	Lower	Upper
107	100		
108	91.7	71.2	111.2
77	89.8	41.1	81.1#
79	28.5	8.1	48.1






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.192 min Scan# 1
Delta R.T. -0.001 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:136 Resp: 73077


Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.7	13.1
54	10.0	7.7	11.5
68	6.4	5.0	7.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#22
Acetophenone
Concen: 69.853 ng
RT: 7.328 min Scan# 873
Delta R.T. 0.012 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:105 Resp: 1180939

Ion	Ratio	Lower	Upper
105	100		
71	7.0	5.7	8.5
51	33.9	30.1	45.1
120	23.8	19.0	28.4





#23
Nitrobenzene-d5
Concen: 173.477 ng
RT: 7.481 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43


Tgt Ion: 82 Resp: 192964


Ion	Ratio	Lower	Upper
82	100		
128	47.0	37.0	55.6
54	59.7	48.0	72.0

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#24
Nitrobenzene
Concen: 84.046 ng
RT: 7.498 min Scan# 902
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion: 77 Resp: 917286

Ion	Ratio	Lower	Upper
77	100		
123	48.4	38.6	58.0
65	14.4	11.5	17.3






#25
Isophorone
Concen: 75.306 ng
RT: 7.739 min Scan# 94
Delta R.T. 0.012 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion: 82 Resp: 175801


Ion	Ratio	Lower	Upper
82	100		
95	7.1	5.6	8.4
138	19.4	15.3	22.9



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#26
2-Nitrophenol
Concen: 79.254 ng
RT: 7.810 min Scan# 955
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:139 Resp: 416387

Ion	Ratio	Lower	Upper
139	100		
109	26.8	21.8	32.8
65	49.4	41.1	61.7






#27
2,4-Dimethylphenol
Concen: 75.098 ng
RT: 7.839 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:122 Resp: 88392


Ion	Ratio	Lower	Upper
122	100		
107	108.1	86.0	129.0
121	57.8	46.3	69.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#28
bis(2-Chloroethoxy)methane
Concen: 72.702 ng
RT: 7.945 min Scan# 978
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion: 93 Resp: 1070383

Ion	Ratio	Lower	Upper
93	100		
95	33.4	26.2	39.2
123	12.5	9.7	14.5






#29
2,4-Dichlorophenol
Concen: 78.103 ng
RT: 8.045 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:162 Resp: 76857


Ion	Ratio	Lower	Upper
162	100		
164	63.7	43.0	83.0
98	37.3	16.7	56.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#30
1,2,4-Trichlorobenzene
Concen: 72.623 ng
RT: 8.133 min Scan# 1010
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:180 Resp: 797476

Ion	Ratio	Lower	Upper
180	100		
182	95.0	75.7	113.5
145	31.7	25.4	38.0






#31
Naphthalene
Concen: 68.820 ng
RT: 8.216 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:128 Resp: 250808


Ion	Ratio	Lower	Upper
128	100		
129	11.6	9.0	13.4
127	13.7	10.6	15.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#32
Benzoic acid
Concen: 83.328 ng m
RT: 7.969 min Scan# 982
Delta R.T. 0.035 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:122 Resp: 523215

Ion	Ratio	Lower	Upper
122	100		
105	118.4	98.2	138.2
77	89.2	66.8	106.8






#33
4-Chloroaniline
Concen: 71.465 ng
RT: 8.263 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:127 Resp: 108454


Ion	Ratio	Lower	Upper
127	100		
129	32.7	26.1	39.1
65	31.4	25.5	38.3
92	19.6	15.4	23.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#34
Hexachlorobutadiene
Concen: 73.564 ng
RT: 8.333 min Scan# 1044
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:225 Resp: 471301

Ion	Ratio	Lower	Upper
225	100		
223	62.5	49.7	74.5
227	63.6	51.0	76.4






#35
Caprolactam
Concen: 81.958 ng
RT: 8.645 min Scan# 1
Delta R.T. 0.023 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:113 Resp: 24241


Ion	Ratio	Lower	Upper
113	100		
55	194.1	177.7	217.7
56	142.5	127.1	167.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#36
4-Chloro-3-methylphenol
Concen: 76.844 ng
RT: 8.739 min Scan# 1113
Delta R.T. 0.012 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:107 Resp: 776483

Ion	Ratio	Lower	Upper
107	100		
144	26.9	21.0	31.4
142	82.9	66.1	99.1






#37
2-Methylnaphthalene
Concen: 69.615 ng
RT: 8.910 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:142 Resp: 155985


Ion	Ratio	Lower	Upper
142	100		
141	87.0	70.5	105.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#38
1-Methylnaphthalene
Concen: 68.237 ng
RT: 9.010 min Scan# 1159
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:142 Resp: 1583955

Ion	Ratio	Lower	Upper
142	100		
141	91.8	72.8	109.2






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.951 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:164 Resp: 38137


Ion	Ratio	Lower	Upper
164	100		
162	98.2	78.6	117.8
160	44.2	35.7	53.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 70.548 ng
RT: 9.075 min Scan# 1170
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:216 Resp: 757210

Ion	Ratio	Lower	Upper
216	100		
214	79.7	63.8	95.6
179	21.0	16.7	25.1
108	18.9	15.6	23.4






#41
Hexachlorocyclopentadiene
Concen: 85.886 ng
RT: 9.063 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:237 Resp: 52318


Ion	Ratio	Lower	Upper
237	100		
235	64.0	43.5	83.5
272	13.5	0.0	32.9



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#42
2,4,6-Tribromophenol
Concen: 166.657 ng
RT: 10.739 min Scan# 1453
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:330 Resp: 579431

Ion	Ratio	Lower	Upper
330	100		
332	97.1	76.5	114.7
141	28.6	24.3	36.5






#43
2,4,6-Trichlorophenol
Concen: 81.996 ng
RT: 9.180 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:196 Resp: 54535


Ion	Ratio	Lower	Upper
196	100		
198	93.9	77.4	116.2
200	30.6	24.6	37.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#44
2,4,5-Trichlorophenol
Concen: 79.732 ng
RT: 9.216 min Scan# 1194
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:196 Resp: 557542

Ion	Ratio	Lower	Upper
196	100		
198	95.4	75.2	112.8
97	50.7	39.8	59.8
132	29.8	23.8	35.8






#45
2-Fluorobiphenyl
Concen: 125.694 ng
RT: 9.275 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:172 Resp: 339986


Ion	Ratio	Lower	Upper
172	100		
171	36.9	29.0	43.4
170	24.6	19.4	29.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#46
1,1'-Biphenyl
Concen: 67.691 ng
RT: 9.375 min Scan# 1221
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:154 Resp: 2006955

Ion	Ratio	Lower	Upper
154	100		
153	40.9	20.5	60.5
76	12.9	0.0	32.8






#47
2-Chloronaphthalene
Concen: 70.140 ng
RT: 9.398 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:162 Resp: 154720


Ion	Ratio	Lower	Upper
162	100		
127	41.9	32.6	49.0
164	33.0	26.1	39.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#48
2-Nitroaniline
Concen: 79.390 ng
RT: 9.486 min Scan# 1240
Delta R.T. -0.001 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion: 65 Resp: 481954

Ion	Ratio	Lower	Upper
65	100		
92	67.1	54.3	81.5
138	106.4	88.6	133.0






#49
Acenaphthylene
Concen: 68.650 ng
RT: 9.816 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:152 Resp: 255538


Ion	Ratio	Lower	Upper
152	100		
151	20.9	16.2	24.2
153	13.7	10.7	16.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#50
Dimethylphthalate
Concen: 72.052 ng
RT: 9.675 min Scan# 1272
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:163 Resp: 1757977

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.2	4.8
164	10.5	8.2	12.2






#51
2,6-Dinitrotoluene
Concen: 78.246 ng
RT: 9.733 min Scan# 1
Delta R.T. 0.005 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:165 Resp: 37912


Ion	Ratio	Lower	Upper
165	100		
63	58.4	48.5	72.7
89	52.7	43.9	65.9



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#52
Acenaphthene
Concen: 69.843 ng
RT: 9.986 min Scan# 1325
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:154 Resp: 1517991

Ion	Ratio	Lower	Upper
154	100		
153	110.2	88.6	133.0
152	53.1	42.1	63.1





#53
3-Nitroaniline
Concen: 78.061 ng
RT: 9.904 min Scan# 1
Delta R.T. 0.012 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43


Tgt Ion:138 Resp: 45657


Ion	Ratio	Lower	Upper
138	100		
108	11.8	11.7	17.5
92	109.7	91.2	136.8

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#54
2,4-Dinitrophenol
Concen: 79.532 ng
RT: 10.004 min Scan# 1328
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:184 Resp: 129817

Ion	Ratio	Lower	Upper
184	100		
63	73.9	69.2	103.8
154	153.5	271.4	407.2#






#55
Dibenzofuran
Concen: 68.664 ng
RT: 10.157 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:168 Resp: 223708


Ion	Ratio	Lower	Upper
168	100		
139	38.0	29.4	44.0
169	14.1	11.0	16.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#56
4-Nitrophenol
Concen: 87.357 ng
RT: 10.045 min Scan# 1335
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:139 Resp: 335015

Ion	Ratio	Lower	Upper
139	100		
109	65.2	44.9	84.9
65	98.8	74.7	114.7






#57
2,4-Dinitrotoluene
Concen: 78.915 ng
RT: 10.133 min Scan# 1
Delta R.T. 0.005 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:165 Resp: 46754


Ion	Ratio	Lower	Upper
165	100		
63	49.4	38.8	58.2
89	72.5	57.1	85.7
182	2.6	2.2	3.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#58
Fluorene
Concen: 65.980 ng
RT: 10.498 min Scan# 1412
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:166 Resp: 1634486

Ion	Ratio	Lower	Upper
166	100		
165	93.0	73.9	110.9
167	13.9	10.8	16.2






#59
2,3,4,6-Tetrachlorophenol
Concen: 80.286 ng
RT: 10.269 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:232 Resp: 46505


Ion	Ratio	Lower	Upper
232	100		
131	44.0	35.9	53.9
130	2.2	1.8	2.6
166	29.4	24.0	36.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#60
Diethylphthalate
Concen: 71.292 ng
RT: 10.375 min Scan# 1391
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:149 Resp: 1717661

Ion	Ratio	Lower	Upper
149	100		
177	22.5	17.6	26.4
150	12.8	10.2	15.2





#61
4-Chlorophenyl-phenylether
Concen: 68.190 ng
RT: 10.492 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43


Instrument :
BNA_F
ClientSampleId :
SSTDICC080


Tgt Ion:204 Resp: 80222

Ion	Ratio	Lower	Upper
204	100		
206	32.4	25.6	38.4
141	58.7	48.9	73.3

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#62
4-Nitroaniline
Concen: 77.614 ng
RT: 10.522 min Scan# 1416
Delta R.T. 0.018 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:138 Resp: 410928

Ion	Ratio	Lower	Upper
138	100		
92	48.3	29.2	69.2
108	67.6	49.1	89.1






#63
Azobenzene
Concen: 70.379 ng
RT: 10.651 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion: 77 Resp: 167443


Ion	Ratio	Lower	Upper
77	100		
182	26.0	7.3	47.3
105	22.6	2.5	42.5
51	34.6	14.4	54.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:188 Resp: 608907

Ion	Ratio	Lower	Upper
188	100		
94	10.7	8.2	12.2
80	11.0	8.6	12.8





#65
4,6-Dinitro-2-methylphenol
Concen: 81.749 ng
RT: 10.545 min Scan# 14
Delta R.T. 0.011 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43


Tgt Ion:198 Resp: 18318


Ion	Ratio	Lower	Upper
198	100		
51	56.4	40.6	80.6
105	47.1	29.2	69.2

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#66
n-Nitrosodiphenylamine
Concen: 72.983 ng
RT: 10.610 min Scan# 1431
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:169 Resp: 1518420

Ion	Ratio	Lower	Upper
169	100		
168	66.6	53.0	79.6
167	36.8	28.8	43.2






#67
4-Bromophenyl-phenylether
Concen: 75.866 ng
RT: 10.980 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:248 Resp: 51446


Ion	Ratio	Lower	Upper
248	100		
250	99.1	77.2	115.8
141	76.7	58.4	87.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#68
Hexachlorobenzene
Concen: 75.999 ng
RT: 11.045 min Scan# 1505
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:284 Resp: 577203

Ion	Ratio	Lower	Upper
284	100		
142	31.8	24.6	37.0
249	30.0	23.2	34.8






#69
Atrazine
Concen: 77.296 ng
RT: 11.133 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:200 Resp: 42388


Ion	Ratio	Lower	Upper
200	100		
173	28.7	8.6	48.6
215	49.0	28.5	68.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#70
Pentachlorophenol
Concen: 87.235 ng
RT: 11.233 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:266 Resp: 350009

Ion	Ratio	Lower	Upper
266	100		
268	62.7	50.0	75.0
264	63.2	49.3	73.9






#71
Phenanthrene
Concen: 69.578 ng
RT: 11.463 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:178 Resp: 228848


Ion	Ratio	Lower	Upper
178	100		
176	20.1	15.5	23.3
179	16.1	12.6	19.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#72
Anthracene
Concen: 69.482 ng
RT: 11.516 min Scan# 1585
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:178 Resp: 2336551

Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.3	22.9
179	15.9	12.7	19.1






#73
Carbazole
Concen: 68.058 ng
RT: 11.663 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:167 Resp: 208776


Ion	Ratio	Lower	Upper
167	100		
166	21.8	17.3	25.9
139	13.4	10.3	15.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#74
Di-n-butylphthalate
Concen: 72.002 ng
RT: 11.998 min Scan# 1667
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:149 Resp: 2221417

Ion	Ratio	Lower	Upper
149	100		
150	9.6	7.5	11.3
104	4.6	3.6	5.4






#75
Fluoranthene
Concen: 65.747 ng
RT: 12.651 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:202 Resp: 213763


Ion	Ratio	Lower	Upper
202	100		
101	11.8	0.0	32.4
203	18.2	0.0	37.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.080 min Scan# 2021
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:240 Resp: 323925

Ion	Ratio	Lower	Upper
240	100		
120	12.2	10.1	15.1
236	25.4	20.5	30.7






#77
Benzidine
Concen: 76.247 ng
RT: 12.768 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:184 Resp: 104946


Ion	Ratio	Lower	Upper
184	100		
185	15.3	11.8	17.6
183	11.0	8.9	13.3



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#78
Pyrene
Concen: 70.910 ng
RT: 12.880 min Scan# 1817
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:202 Resp: 2118847

Ion	Ratio	Lower	Upper
202	100		
200	20.7	16.1	24.1
203	18.3	14.1	21.1






#79
Terphenyl-d14
Concen: 134.413 ng
RT: 13.027 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:244 Resp: 299363


Ion	Ratio	Lower	Upper
244	100		
212	6.7	5.3	7.9
122	10.4	9.1	13.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#80
Butylbenzylphthalate
Concen: 78.197 ng
RT: 13.498 min Scan# 1922
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:149 Resp: 659574

Ion	Ratio	Lower	Upper
149	100		
91	70.0	54.4	81.6
206	20.7	15.9	23.9





#81
Benzo(a)anthracene
Concen: 72.970 ng
RT: 14.068 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43


Tgt Ion:228 Resp: 158677


Ion	Ratio	Lower	Upper
228	100		
226	27.5	21.9	32.9
229	20.2	16.0	24.0

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#82
3,3'-Dichlorobenzidine
Concen: 88.259 ng
RT: 14.033 min Scan# 2013
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:252 Resp: 563738

Ion	Ratio	Lower	Upper
252	100		
254	64.9	51.0	76.6
126	11.5	9.4	14.2






#83
Chrysene
Concen: 74.470 ng
RT: 14.104 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:228 Resp: 147409


Ion	Ratio	Lower	Upper
228	100		
226	30.5	24.0	36.0
229	20.2	15.7	23.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 78.547 ng
RT: 14.057 min Scan# 2017
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:149 Resp: 915684

Ion	Ratio	Lower	Upper
149	100		
167	27.3	21.7	32.5
279	3.9	3.0	4.4





#85
Di-n-octyl phthalate
Concen: 78.594 ng
RT: 14.680 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43


Tgt Ion:149 Resp: 172221


Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.1	1.7
43	11.4	9.3	13.9

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.568 min Scan# 2274
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:264 Resp: 374102

Ion	Ratio	Lower	Upper
264	100		
260	23.6	18.4	27.6
265	21.8	17.3	25.9






#87
Indeno(1,2,3-cd)pyrene
Concen: 77.677 ng
RT: 17.092 min Scan# 2
Delta R.T. 0.012 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:276 Resp: 212847


Ion	Ratio	Lower	Upper
276	100		
138	28.4	22.6	33.8
277	25.4	20.1	30.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#88
Benzo(b)fluoranthene
Concen: 77.378 ng
RT: 15.133 min Scan# 2200
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:252 Resp: 1798083

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.7	26.5
125	10.6	9.0	13.4






#89
Benzo(k)fluoranthene
Concen: 67.716 ng
RT: 15.162 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:252 Resp: 143930


Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.8	26.8
125	11.2	9.0	13.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#90
Benzo(a)pyrene
Concen: 75.685 ng
RT: 15.509 min Scan# 2264
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:252 Resp: 1610193

Ion	Ratio	Lower	Upper
252	100		
253	22.4	17.6	26.4
125	12.2	9.5	14.3






#91
Dibenzo(a,h)anthracene
Concen: 76.989 ng
RT: 17.115 min Scan# 2
Delta R.T. 0.017 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:278 Resp: 171218


Ion	Ratio	Lower	Upper
278	100		
139	17.9	15.2	22.8
279	24.1	19.0	28.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#92
Benzo(g,h,i)perylene
Concen: 77.757 ng
RT: 17.551 min Scan# 2611
Delta R.T. 0.012 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:276 Resp: 1738450

Ion	Ratio	Lower	Upper
276	100		
277	23.9	18.7	28.1
138	25.1	19.4	29.0



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142247.D
Acq On : 30 Apr 2025 16:17
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025

Quant Time: Apr 30 16:36:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	185327	20.000	ng	0.00
21) Naphthalene-d8	8.192	136	725482	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	381170	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	633040	20.000	ng	0.00
76) Chrysene-d12	14.074	240	345729	20.000	ng	0.00
86) Perylene-d12	15.562	264	342826	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	987704	85.116	ng	0.00
7) Phenol-d6	6.522	99	1201328	85.730	ng	0.00
23) Nitrobenzene-d5	7.469	82	1091522	98.845	ng	0.00
42) 2,4,6-Tribromophenol	10.733	330	323319	93.044	ng	0.00
45) 2-Fluorobiphenyl	9.269	172	2157499	79.807	ng	0.00
79) Terphenyl-d14	13.021	244	2037561	85.716	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.699	88	219939	42.438	ng	98
3) Pyridine	3.469	79	572034	42.469	ng	99
4) n-Nitrosodimethylamine	3.416	42	299316	42.942	ng	96
6) Aniline	6.569	93	844910	42.598	ng	99
8) 2-Chlorophenol	6.692	128	530021	44.202	ng	98
9) Benzaldehyde	6.457	77	420408	42.925	ng	99
10) Phenol	6.539	94	645155m	42.765	ng	
11) bis(2-Chloroethyl)ether	6.639	93	494843	42.329	ng	98
12) 1,3-Dichlorobenzene	6.851	146	567280	41.819	ng	99
13) 1,4-Dichlorobenzene	6.928	146	572974	41.924	ng	100
14) 1,2-Dichlorobenzene	7.081	146	551216	42.247	ng	99
15) Benzyl Alcohol	7.045	79	431108	43.378	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.181	45	901053	42.684	ng	99
17) 2-Methylphenol	7.151	107	415403	42.288	ng	98
18) Hexachloroethane	7.422	117	197959	44.255	ng	95
19) n-Nitroso-di-n-propyla...	7.322	70	363621	41.992	ng	99
20) 3+4-Methylphenols	7.304	107	539514	43.919	ng	98
22) Acetophenone	7.316	105	695539	41.441	ng	98
24) Nitrobenzene	7.492	77	513604	47.402	ng	99
25) Isophorone	7.728	82	968502	41.789	ng	100
26) 2-Nitrophenol	7.804	139	219881	48.175	ng	100
27) 2,4-Dimethylphenol	7.833	122	502502	43.004	ng	100
28) bis(2-Chloroethoxy)met...	7.939	93	610669	41.780	ng	99
29) 2,4-Dichlorophenol	8.039	162	429967	44.013	ng	99
30) 1,2,4-Trichlorobenzene	8.133	180	454848	41.724	ng	100
31) Naphthalene	8.216	128	1483568	41.005	ng	100
32) Benzoic acid	7.933	122	246172m	45.113	ng	
33) 4-Chloroaniline	8.257	127	622894	41.344	ng	100
34) Hexachlorobutadiene	8.333	225	267940	42.127	ng	99
35) Caprolactam	8.622	113	130473	44.434	ng	98
36) 4-Chloro-3-methylphenol	8.728	107	436215	43.485	ng	100
37) 2-Methylnaphthalene	8.904	142	915474	41.155	ng	100
38) 1-Methylnaphthalene	9.004	142	950163	41.232	ng	100
40) 1,2,4,5-Tetrachloroben...	9.069	216	443561	41.349	ng	100
41) Hexachlorocyclopentadiene	9.057	237	288102	47.321	ng	100
43) 2,4,6-Trichlorophenol	9.175	196	303872	45.713	ng	99

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142247.D
Acq On : 30 Apr 2025 16:17
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025

Quant Time: Apr 30 16:36:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.210	196	312237	44.676	ng		99
46) 1,1'-Biphenyl	9.369	154	1213047	40.936	ng		100
47) 2-Chloronaphthalene	9.392	162	903974	41.003	ng		99
48) 2-Nitroaniline	9.480	65	269528	46.523	ng		95
49) Acenaphthylene	9.810	152	1533164	41.211	ng		100
50) Dimethylphthalate	9.669	163	1025498	42.054	ng		100
51) 2,6-Dinitrotoluene	9.727	165	211107	45.376	ng		96
52) Acenaphthene	9.980	154	904325	41.631	ng		99
53) 3-Nitroaniline	9.892	138	257926	45.509	ng		96
54) 2,4-Dinitrophenol	9.992	184	62568	49.090	ng	#	6
55) Dibenzofuran	10.151	168	1340934	41.181	ng		100
56) 4-Nitrophenol	10.033	139	191463	49.952	ng		99
57) 2,4-Dinitrotoluene	10.127	165	265412	46.994	ng		96
58) Fluorene	10.492	166	1008032	40.714	ng		99
59) 2,3,4,6-Tetrachlorophenol	10.263	232	264709	45.724	ng		100
60) Diethylphthalate	10.369	149	1016722	42.222	ng		100
61) 4-Chlorophenyl-phenyle...	10.486	204	477688	40.626	ng		99
62) 4-Nitroaniline	10.504	138	232165	45.070	ng		98
63) Azobenzene	10.645	77	984443	41.400	ng		99
65) 4,6-Dinitro-2-methylph...	10.533	198	96341	46.598	ng		95
66) n-Nitrosodiphenylamine	10.604	169	904466	41.816	ng		100
67) 4-Bromophenyl-phenylether	10.980	248	299605	42.497	ng		98
68) Hexachlorobenzene	11.039	284	330048	41.800	ng		97
69) Atrazine	11.127	200	255030	44.732	ng		99
70) Pentachlorophenol	11.233	266	197476	47.342	ng		99
71) Phenanthrene	11.457	178	1400439	40.955	ng		100
72) Anthracene	11.510	178	1463247	41.854	ng		100
73) Carbazole	11.663	167	1332856	41.793	ng		99
74) Di-n-butylphthalate	11.992	149	1412919	44.051	ng		100
75) Fluoranthene	12.645	202	1409189	41.690	ng		99
77) Benzidine	12.768	184	682155	46.435	ng		99
78) Pyrene	12.874	202	1405875	44.082	ng		100
80) Butylbenzylphthalate	13.492	149	371146	42.984	ng		98
81) Benzo(a)anthracene	14.062	228	960484	41.383	ng		99
82) 3,3'-Dichlorobenzidine	14.027	252	310153	45.495	ng		98
83) Chrysene	14.098	228	890671	42.158	ng		100
84) Bis(2-ethylhexyl)phtha...	14.057	149	487122	40.882	ng		99
85) Di-n-octyl phthalate	14.674	149	889844	41.178	ng		100
87) Indeno(1,2,3-cd)pyrene	17.074	276	1098774	43.757	ng		99
88) Benzo(b)fluoranthene	15.121	252	860370	40.403	ng		100
89) Benzo(k)fluoranthene	15.157	252	811540	41.665	ng		99
90) Benzo(a)pyrene	15.498	252	826250	42.380	ng		99
91) Dibenzo(a,h)anthracene	17.098	278	887855	43.565	ng		99
92) Benzo(g,h,i)perylene	17.527	276	899474	43.902	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142247.D
Acq On : 30 Apr 2025 16:17
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1


Quant Time: Apr 30 16:36:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:152 Resp: 18532


Ion	Ratio	Lower	Upper
152	100		
150	162.7	130.2	195.2
115	59.5	47.0	70.4



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#2
1,4-Dioxane
Concen: 42.438 ng
RT: 2.699 min Scan# 86
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion: 88 Resp: 219939

Ion	Ratio	Lower	Upper
88	100		
58	91.6	72.2	108.2
43	34.8	27.0	40.6






#3
Pyridine
Concen: 42.469 ng
RT: 3.469 min Scan# 21
Delta R.T. 0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion: 79 Resp: 57203


Ion	Ratio	Lower	Upper
79	100		
52	88.3	71.5	107.3
51	41.9	34.5	51.7



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#4
n-Nitrosodimethylamine
Concen: 42.942 ng
RT: 3.416 min Scan# 208
Delta R.T. 0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion: 42 Resp: 299316

Ion	Ratio	Lower	Upper
42	100		
74	116.5	89.5	134.3
44	7.9	6.6	9.8





#5
2-Fluorophenol
Concen: 85.116 ng
RT: 5.516 min Scan# 5
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17


Tgt Ion:112 Resp: 98770


Ion	Ratio	Lower	Upper
112	100		
64	60.5	48.4	72.6
63	31.2	24.8	37.2

Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#6
Aniline
Concen: 42.598 ng
RT: 6.569 min Scan# 744
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion: 93 Resp: 844910

Ion	Ratio	Lower	Upper
93	100		
66	40.4	32.1	48.1
65	20.3	16.6	24.8






#7
Phenol-d6
Concen: 85.730 ng
RT: 6.522 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion: 99 Resp: 120132


Ion	Ratio	Lower	Upper
99	100		
42	21.4	16.3	24.5
71	32.2	25.0	37.4



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#8
2-Chlorophenol
Concen: 44.202 ng
RT: 6.692 min Scan# 765
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:128 Resp: 530021

Ion	Ratio	Lower	Upper
128	100		
130	32.7	13.1	53.1
64	47.1	28.6	68.6






#9
Benzaldehyde
Concen: 42.925 ng
RT: 6.457 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion: 77 Resp: 42040


Ion	Ratio	Lower	Upper
77	100		
105	98.5	79.1	119.1
106	95.0	76.0	116.0



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#10
Phenol
Concen: 42.765 ng m
RT: 6.539 min Scan# 739
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion: 94 Resp: 645155

Ion	Ratio	Lower	Upper
94	100		
65	29.1	8.9	48.9
66	38.7	18.7	58.7






#11
bis(2-Chloroethyl)ether
Concen: 42.329 ng
RT: 6.639 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion: 93 Resp: 49484


Ion	Ratio	Lower	Upper
93	100		
63	84.5	62.2	102.2
95	32.4	12.7	52.7



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#12
1,3-Dichlorobenzene
Concen: 41.819 ng
RT: 6.851 min Scan# 792
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:146 Resp: 567280

Ion	Ratio	Lower	Upper
146	100		
148	63.8	50.5	75.7
75	29.2	23.2	34.8






#13
1,4-Dichlorobenzene
Concen: 41.924 ng
RT: 6.928 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:146 Resp: 57297


Ion	Ratio	Lower	Upper
146	100		
148	63.3	50.6	76.0
111	41.8	33.4	50.0



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#14
1,2-Dichlorobenzene
Concen: 42.247 ng
RT: 7.081 min Scan# 831
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:146 Resp: 551216

Ion	Ratio	Lower	Upper
146	100		
148	64.4	50.5	75.7
111	43.8	34.9	52.3






#15
Benzyl Alcohol
Concen: 43.378 ng
RT: 7.045 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion: 79 Resp: 43110


Ion	Ratio	Lower	Upper
79	100		
108	81.5	64.5	96.7
77	62.1	49.2	73.8



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 42.684 ng
RT: 7.181 min Scan# 848
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion: 45 Resp: 901053

Ion	Ratio	Lower	Upper
45	100		
77	10.9	0.0	31.3
79	8.2	0.0	28.5






#17
2-Methylphenol
Concen: 42.288 ng
RT: 7.151 min Scan# 84
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:107 Resp: 41540


Ion	Ratio	Lower	Upper
107	100		
108	116.0	91.0	136.4
77	43.7	34.4	51.6
79	42.7	34.4	51.6



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#18
Hexachloroethane
Concen: 44.255 ng
RT: 7.422 min Scan# 889
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:117 Resp: 197959

Ion	Ratio	Lower	Upper
117	100		
119	96.3	77.1	115.7
201	89.8	79.4	119.2






#19
n-Nitroso-di-n-propylamine
Concen: 41.992 ng
RT: 7.322 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion: 70 Resp: 36362


Ion	Ratio	Lower	Upper
70	100		
42	66.9	53.4	80.2
101	9.7	7.7	11.5
130	22.4	17.2	25.8



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#20
3+4-Methylphenols
Concen: 43.919 ng
RT: 7.304 min Scan# 869
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:107 Resp: 539514

Ion	Ratio	Lower	Upper
107	100		
108	90.2	71.2	111.2
77	63.5	41.1	81.1
79	27.5	8.1	48.1






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.192 min Scan# 1(
Delta R.T. -0.001 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:136 Resp: 72548


Ion	Ratio	Lower	Upper
136	100		
137	11.2	8.7	13.1
54	9.7	7.7	11.5
68	6.1	5.0	7.6



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#22
Acetophenone
Concen: 41.441 ng
RT: 7.316 min Scan# 871
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:105 Resp: 695539

Ion	Ratio	Lower	Upper
105	100		
71	7.3	5.7	8.5
51	35.9	30.1	45.1
120	23.5	19.0	28.4





#23
Nitrobenzene-d5
Concen: 98.845 ng
RT: 7.469 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17


Tgt Ion: 82 Resp: 109152


Ion	Ratio	Lower	Upper
82	100		
128	45.7	37.0	55.6
54	61.2	48.0	72.0

Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#24
Nitrobenzene
Concen: 47.402 ng
RT: 7.492 min Scan# 901
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion: 77 Resp: 513604

Ion	Ratio	Lower	Upper
77	100		
123	48.8	38.6	58.0
65	14.5	11.5	17.3






#25
Isophorone
Concen: 41.789 ng
RT: 7.728 min Scan# 9
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion: 82 Resp: 96850


Ion	Ratio	Lower	Upper
82	100		
95	7.0	5.6	8.4
138	19.2	15.3	22.9



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#26
2-Nitrophenol
Concen: 48.175 ng
RT: 7.804 min Scan# 954
Delta R.T. 0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:139 Resp: 219881

Ion	Ratio	Lower	Upper
139	100		
109	27.6	21.8	32.8
65	51.2	41.1	61.7






#27
2,4-Dimethylphenol
Concen: 43.004 ng
RT: 7.833 min Scan# 9
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:122 Resp: 50250


Ion	Ratio	Lower	Upper
122	100		
107	108.0	86.0	129.0
121	58.1	46.3	69.5



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#28
bis(2-Chloroethoxy)methane
Concen: 41.780 ng
RT: 7.939 min Scan# 977
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion: 93 Resp: 610669

Ion	Ratio	Lower	Upper
93	100		
95	33.1	26.2	39.2
123	12.6	9.7	14.5






#29
2,4-Dichlorophenol
Concen: 44.013 ng
RT: 8.039 min Scan# 9
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:162 Resp: 42996


Ion	Ratio	Lower	Upper
162	100		
164	63.9	43.0	83.0
98	37.3	16.7	56.7



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#30
1,2,4-Trichlorobenzene
Concen: 41.724 ng
RT: 8.133 min Scan# 1010
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:180 Resp: 454848

Ion	Ratio	Lower	Upper
180	100		
182	94.8	75.7	113.5
145	31.1	25.4	38.0






#31
Naphthalene
Concen: 41.005 ng
RT: 8.216 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:128 Resp: 148356


Ion	Ratio	Lower	Upper
128	100		
129	11.3	9.0	13.4
127	13.3	10.6	15.8



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#32
Benzoic acid
Concen: 45.113 ng m
RT: 7.933 min Scan# 976
Delta R.T. -0.001 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:122 Resp: 246172

Ion	Ratio	Lower	Upper
122	100		
105	117.2	98.2	138.2
77	86.9	66.8	106.8






#33
4-Chloroaniline
Concen: 41.344 ng
RT: 8.257 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:127 Resp: 62289


Ion	Ratio	Lower	Upper
127	100		
129	32.7	26.1	39.1
65	32.0	25.5	38.3
92	19.0	15.4	23.2



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#34
Hexachlorobutadiene
Concen: 42.127 ng
RT: 8.333 min Scan# 1044
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:225 Resp: 267940

Ion	Ratio	Lower	Upper
225	100		
223	63.0	49.7	74.5
227	63.9	51.0	76.4





#35
Caprolactam
Concen: 44.434 ng
RT: 8.622 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17


Tgt Ion:113 Resp: 13047


Ion	Ratio	Lower	Upper
113	100		
55	199.6	177.7	217.7
56	144.7	127.1	167.1

Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#36
4-Chloro-3-methylphenol
Concen: 43.485 ng
RT: 8.728 min Scan# 1111
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:107 Resp: 436215

Ion	Ratio	Lower	Upper
107	100		
144	26.7	21.0	31.4
142	82.5	66.1	99.1






#37
2-Methylnaphthalene
Concen: 41.155 ng
RT: 8.904 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:142 Resp: 91547


Ion	Ratio	Lower	Upper
142	100		
141	87.6	70.5	105.7



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#38
1-Methylnaphthalene
Concen: 41.232 ng
RT: 9.004 min Scan# 1158
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:142 Resp: 950163

Ion	Ratio	Lower	Upper
142	100		
141	90.9	72.8	109.2






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.945 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:164 Resp: 38117


Ion	Ratio	Lower	Upper
164	100		
162	98.0	78.6	117.8
160	44.5	35.7	53.5



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 41.349 ng
RT: 9.069 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:216 Resp: 443561

Ion	Ratio	Lower	Upper
216	100		
214	79.6	63.8	95.6
179	21.0	16.7	25.1
108	19.2	15.6	23.4





#41
Hexachlorocyclopentadiene
Concen: 47.321 ng
RT: 9.057 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17


Tgt Ion:237 Resp: 28810


Ion	Ratio	Lower	Upper
237	100		
235	63.3	43.5	83.5
272	13.2	0.0	32.9

Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#42
2,4,6-Tribromophenol
Concen: 93.044 ng
RT: 10.733 min Scan# 1452
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:330 Resp: 323319

Ion	Ratio	Lower	Upper
330	100		
332	96.9	76.5	114.7
141	31.4	24.3	36.5






#43
2,4,6-Trichlorophenol
Concen: 45.713 ng
RT: 9.175 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:196 Resp: 30387


Ion	Ratio	Lower	Upper
196	100		
198	95.5	77.4	116.2
200	30.1	24.6	37.0



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#44
2,4,5-Trichlorophenol
Concen: 44.676 ng
RT: 9.210 min Scan# 1193
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:196 Resp: 312237

Ion	Ratio	Lower	Upper
196	100		
198	96.3	75.2	112.8
97	49.7	39.8	59.8
132	29.6	23.8	35.8






#45
2-Fluorobiphenyl
Concen: 79.807 ng
RT: 9.269 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:172 Resp: 215749


Ion	Ratio	Lower	Upper
172	100		
171	36.1	29.0	43.4
170	24.0	19.4	29.0



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#46
1,1'-Biphenyl
Concen: 40.936 ng
RT: 9.369 min Scan# 1220
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:154 Resp: 1213047

Ion	Ratio	Lower	Upper
154	100		
153	40.7	20.5	60.5
76	12.6	0.0	32.8






#47
2-Chloronaphthalene
Concen: 41.003 ng
RT: 9.392 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:162 Resp: 90397


Ion	Ratio	Lower	Upper
162	100		
127	41.2	32.6	49.0
164	32.8	26.1	39.1



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#48
2-Nitroaniline
Concen: 46.523 ng
RT: 9.480 min Scan# 1239
Delta R.T. -0.007 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion: 65 Resp: 269528

Ion	Ratio	Lower	Upper
65	100		
92	65.5	54.3	81.5
138	104.3	88.6	133.0






#49
Acenaphthylene
Concen: 41.211 ng
RT: 9.810 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:152 Resp: 153316


Ion	Ratio	Lower	Upper
152	100		
151	20.2	16.2	24.2
153	13.4	10.7	16.1



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#50
Dimethylphthalate
Concen: 42.054 ng
RT: 9.669 min Scan# 1271
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:163 Resp: 1025498

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.2	4.8
164	10.1	8.2	12.2






#51
2,6-Dinitrotoluene
Concen: 45.376 ng
RT: 9.727 min Scan# 1
Delta R.T. -0.001 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:165 Resp: 21110


Ion	Ratio	Lower	Upper
165	100		
63	58.0	48.5	72.7
89	52.2	43.9	65.9



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#52
Acenaphthene
Concen: 41.631 ng
RT: 9.980 min Scan# 1324
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:154 Resp: 904325

Ion	Ratio	Lower	Upper
154	100		
153	110.3	88.6	133.0
152	53.0	42.1	63.1





#53
3-Nitroaniline
Concen: 45.509 ng
RT: 9.892 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17


Tgt Ion:138 Resp: 25792


Ion	Ratio	Lower	Upper
138	100		
108	11.7	11.7	17.5
92	109.7	91.2	136.8

Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#54
2,4-Dinitrophenol
Concen: 49.090 ng
RT: 9.992 min Scan# 1326
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:184 Resp: 62568

Ion	Ratio	Lower	Upper
184	100		
63	109.3	69.2	103.8#
154	577.6	271.4	407.2#






#55
Dibenzofuran
Concen: 41.181 ng
RT: 10.151 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:168 Resp: 134093


Ion	Ratio	Lower	Upper
168	100		
139	36.8	29.4	44.0
169	13.7	11.0	16.4



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#56
4-Nitrophenol
Concen: 49.952 ng
RT: 10.033 min Scan# 1333
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:139 Resp: 191463

Ion	Ratio	Lower	Upper
139	100		
109	65.2	44.9	84.9
65	96.4	74.7	114.7






#57
2,4-Dinitrotoluene
Concen: 46.994 ng
RT: 10.127 min Scan# 1
Delta R.T. -0.001 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:165 Resp: 26541


Ion	Ratio	Lower	Upper
165	100		
63	45.6	38.8	58.2
89	68.8	57.1	85.7
182	2.8	2.2	3.2



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#58
Fluorene
Concen: 40.714 ng
RT: 10.492 min Scan# 1411
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:166 Resp: 1008032

Ion	Ratio	Lower	Upper
166	100		
165	93.0	73.9	110.9
167	13.7	10.8	16.2






#59
2,3,4,6-Tetrachlorophenol
Concen: 45.724 ng
RT: 10.263 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:232 Resp: 26470


Ion	Ratio	Lower	Upper
232	100		
131	44.8	35.9	53.9
130	2.3	1.8	2.6
166	29.7	24.0	36.0



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#60
Diethylphthalate
Concen: 42.222 ng
RT: 10.369 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:149 Resp: 1016722

Ion	Ratio	Lower	Upper
149	100		
177	22.1	17.6	26.4
150	12.7	10.2	15.2





#61
4-Chlorophenyl-phenylether
Concen: 40.626 ng
RT: 10.486 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17


Instrument :
BNA_F
ClientSampleId :
ICVBF043025


Tgt Ion:204 Resp: 47768

Ion	Ratio	Lower	Upper
204	100		
206	32.7	25.6	38.4
141	61.2	48.9	73.3

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#62
4-Nitroaniline
Concen: 45.070 ng
RT: 10.504 min Scan# 1413
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:138 Resp: 232165

Ion	Ratio	Lower	Upper
138	100		
92	48.8	29.2	69.2
108	67.1	49.1	89.1






#63
Azobenzene
Concen: 41.400 ng
RT: 10.645 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion: 77 Resp: 98444


Ion	Ratio	Lower	Upper
77	100		
182	25.8	7.3	47.3
105	22.4	2.5	42.5
51	34.9	14.4	54.4



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:188 Resp: 633040

Ion	Ratio	Lower	Upper
188	100		
94	10.0	8.2	12.2
80	10.8	8.6	12.8





#65
4,6-Dinitro-2-methylphenol
Concen: 46.598 ng
RT: 10.533 min Scan# 14
Delta R.T. -0.001 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17


Tgt Ion:198 Resp: 9634


Ion	Ratio	Lower	Upper
198	100		
51	56.1	40.6	80.6
105	46.7	29.2	69.2

Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025







#66
n-Nitrosodiphenylamine
Concen: 41.816 ng
RT: 10.604 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:169 Resp: 904466

Ion	Ratio	Lower	Upper
169	100		
168	66.4	53.0	79.6
167	36.3	28.8	43.2






#67
4-Bromophenyl-phenylether
Concen: 42.497 ng
RT: 10.980 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:248 Resp: 29960


Ion	Ratio	Lower	Upper
248	100		
250	96.6	77.2	115.8
141	70.0	58.4	87.6



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#68
Hexachlorobenzene
Concen: 41.800 ng
RT: 11.039 min Scan# 1504
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:284 Resp: 330048

Ion	Ratio	Lower	Upper
284	100		
142	33.5	24.6	37.0
249	30.1	23.2	34.8






#69
Atrazine
Concen: 44.732 ng
RT: 11.127 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:200 Resp: 25503


Ion	Ratio	Lower	Upper
200	100		
173	27.9	8.6	48.6
215	48.4	28.5	68.5



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#70
Pentachlorophenol
Concen: 47.342 ng
RT: 11.233 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:266 Resp: 197476

Ion	Ratio	Lower	Upper
266	100		
268	62.3	50.0	75.0
264	62.9	49.3	73.9






#71
Phenanthrene
Concen: 40.955 ng
RT: 11.457 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:178 Resp: 140043


Ion	Ratio	Lower	Upper
178	100		
176	19.2	15.5	23.3
179	15.6	12.6	19.0



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#72
Anthracene
Concen: 41.854 ng
RT: 11.510 min Scan# 1584
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:178 Resp: 1463247

Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.3	22.9
179	15.8	12.7	19.1






#73
Carbazole
Concen: 41.793 ng
RT: 11.663 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:167 Resp: 133285


Ion	Ratio	Lower	Upper
167	100		
166	21.4	17.3	25.9
139	12.5	10.3	15.5



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#74
Di-n-butylphthalate
Concen: 44.051 ng
RT: 11.992 min Scan# 1666
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:149 Resp: 1412919

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.5	11.3
104	4.6	3.6	5.4






#75
Fluoranthene
Concen: 41.690 ng
RT: 12.645 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:202 Resp: 140918


Ion	Ratio	Lower	Upper
202	100		
101	12.0	0.0	32.4
203	17.8	0.0	37.5



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.074 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:240 Resp: 345729

Ion	Ratio	Lower	Upper
240	100		
120	12.1	10.1	15.1
236	25.9	20.5	30.7






#77
Benzidine
Concen: 46.435 ng
RT: 12.768 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:184 Resp: 68215


Ion	Ratio	Lower	Upper
184	100		
185	15.3	11.8	17.6
183	10.9	8.9	13.3



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#78
Pyrene
Concen: 44.082 ng
RT: 12.874 min Scan# 1816
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:202 Resp: 1405875

Ion	Ratio	Lower	Upper
202	100		
200	20.2	16.1	24.1
203	17.8	14.1	21.1






#79
Terphenyl-d14
Concen: 85.716 ng
RT: 13.021 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:244 Resp: 203756


Ion	Ratio	Lower	Upper
244	100		
212	6.6	5.3	7.9
122	10.9	9.1	13.7



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#80
Butylbenzylphthalate
Concen: 42.984 ng
RT: 13.492 min Scan# 1921
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:149 Resp: 371146

Ion	Ratio	Lower	Upper
149	100		
91	69.8	54.4	81.6
206	19.9	15.9	23.9






#81
Benzo(a)anthracene
Concen: 41.383 ng
RT: 14.062 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:228 Resp: 96048


Ion	Ratio	Lower	Upper
228	100		
226	27.0	21.9	32.9
229	19.6	16.0	24.0



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#82
3,3'-Dichlorobenzidine
Concen: 45.495 ng
RT: 14.027 min Scan# 2012
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:252 Resp: 310153

Ion	Ratio	Lower	Upper
252	100		
254	65.0	51.0	76.6
126	11.3	9.4	14.2






#83
Chrysene
Concen: 42.158 ng
RT: 14.098 min Scan# 2(
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:228 Resp: 89067


Ion	Ratio	Lower	Upper
228	100		
226	30.0	24.0	36.0
229	19.8	15.7	23.5



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 40.882 ng
RT: 14.057 min Scan# 2017
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:149 Resp: 487122

Ion	Ratio	Lower	Upper
149	100		
167	26.8	21.7	32.5
279	3.9	3.0	4.4






#85
Di-n-octyl phthalate
Concen: 41.178 ng
RT: 14.674 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:149 Resp: 88984


Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.1	1.7
43	11.4	9.3	13.9



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 15.562 min Scan# 2273
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:264 Resp: 342826

Ion	Ratio	Lower	Upper
264	100		
260	23.6	18.4	27.6
265	22.0	17.3	25.9






#87
Indeno(1,2,3-cd)pyrene
Concen: 43.757 ng
RT: 17.074 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:276 Resp: 109877


Ion	Ratio	Lower	Upper
276	100		
138	28.7	22.6	33.8
277	25.1	20.1	30.1



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#88
Benzo(b)fluoranthene
Concen: 40.403 ng
RT: 15.121 min Scan# 2198
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:252 Resp: 860370

Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.7	26.5
125	11.4	9.0	13.4






#89
Benzo(k)fluoranthene
Concen: 41.665 ng
RT: 15.157 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:252 Resp: 81154


Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.8	26.8
125	10.9	9.0	13.6



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#90
Benzo(a)pyrene
Concen: 42.380 ng
RT: 15.498 min Scan# 2262
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:252 Resp: 826250

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.6	26.4
125	12.3	9.5	14.3






#91
Dibenzo(a,h)anthracene
Concen: 43.565 ng
RT: 17.098 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:278 Resp: 88785


Ion	Ratio	Lower	Upper
278	100		
139	18.2	15.2	22.8
279	23.7	19.0	28.6



Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#92
Benzo(g,h,i)perylene
Concen: 43.902 ng
RT: 17.527 min Scan# 2607
Delta R.T. -0.012 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:276 Resp: 899474

Ion	Ratio	Lower	Upper
276	100		
277	23.4	18.7	28.1
138	24.8	19.4	29.0



Evaluate Continuing Calibration Report

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142247.D
Acq On : 30 Apr 2025 16:17
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Quant Time: Apr 30 16:36:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2		1,4-Dioxane	0.559	0.593	-6.1	103	0.00
3		Pyridine	1.454	1.543	-6.1	104	0.00
4		n-Nitrosodimethylamine	0.752	0.808	-7.4	103	0.00
5	S	2-Fluorophenol	1.252	1.332	-6.4	104	0.00
6		Aniline	2.140	2.280	-6.5	105	0.00
7	S	Phenol-d6	1.512	1.621	-7.2	105	0.00
8		2-Chlorophenol	1.294	1.430	-10.5	107	0.00
9		Benzaldehyde	1.057	1.134	-7.3	105	0.00
10	C	Phenol	1.628	1.741	-6.9	105	0.00
11		bis(2-Chloroethyl)ether	1.262	1.335	-5.8	104	0.00
12		1,3-Dichlorobenzene	1.464	1.530	-4.5	103	0.00
13	C	1,4-Dichlorobenzene	1.475	1.546	-4.8	103	0.00
14		1,2-Dichlorobenzene	1.408	1.487	-5.6	104	0.00
15		Benzyl Alcohol	1.073	1.163	-8.4	104	0.00
16		2,2'-oxybis(1-Chloropropane	2.278	2.431	-6.7	105	0.00
17		2-Methylphenol	1.060	1.121	-5.8	103	0.00
18		Hexachloroethane	0.483	0.534	-10.6	108	0.00
19	P	n-Nitroso-di-n-propylamine	0.934	0.981	-5.0	104	0.00
20		3+4-Methylphenols	1.326	1.456	-9.8	107	0.00
21	I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
22		Acetophenone	0.463	0.479	-3.5	105	0.00
23	S	Nitrobenzene-d5	0.304	0.376	-23.7	114	0.00
24		Nitrobenzene	0.299	0.354	-18.4	111	0.00
25		Isophorone	0.639	0.667	-4.4	104	0.00
26	C	2-Nitrophenol	0.102	0.152	-49.0#	135	0.00
27		2,4-Dimethylphenol	0.322	0.346	-7.5	106	0.00
28		bis(2-Chloroethoxy)methane	0.403	0.421	-4.5	104	0.00
29	C	2,4-Dichlorophenol	0.269	0.296	-10.0	105	0.00
30		1,2,4-Trichlorobenzene	0.301	0.313	-4.0	103	0.00
31		Naphthalene	0.997	1.022	-2.5	103	0.00
32		Benzoic acid	0.125	0.170	-36.0#	134	0.00
33		4-Chloroaniline	0.415	0.429	-3.4	103	0.00
34	C	Hexachlorobutadiene	0.175	0.185	-5.7	103	0.00
35		Caprolactam	0.081	0.090	-11.1	106	0.00
36	C	4-Chloro-3-methylphenol	0.277	0.301	-8.7	105	0.00
37		2-Methylnaphthalene	0.613	0.631	-2.9	104	0.00
38		1-Methylnaphthalene	0.635	0.655	-3.1	103	0.00
39	I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
40		1,2,4,5-Tetrachlorobenzene	0.563	0.582	-3.4	104	0.00
41	P	Hexachlorocyclopentadiene	0.319	0.378	-18.5	110	0.00
42	S	2,4,6-Tribromophenol	0.182	0.212	-16.5	107	0.00
43	C	2,4,6-Trichlorophenol	0.349	0.399	-14.3	111	0.00
44		2,4,5-Trichlorophenol	0.367	0.410	-11.7	104	0.00
45	S	2-Fluorobiphenyl	1.418	1.415	0.2	103	0.00
46		1,1'-Biphenyl	1.555	1.591	-2.3	103	0.00
47		2-Chloronaphthalene	1.157	1.186	-2.5	103	0.00

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142247.D
Acq On : 30 Apr 2025 16:17
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Quant Time: Apr 30 16:36:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.249	0.354	-42.2#	123	0.00
49		Acenaphthylene	1.952	2.011	-3.0	103	0.00
50		Dimethylphthalate	1.280	1.345	-5.1	105	0.00
51		2,6-Dinitrotoluene	0.204	0.277	-35.8#	118	0.00
52	C	Acenaphthene	1.140	1.186	-4.0	104	0.00
53		3-Nitroaniline	0.257	0.338	-31.5#	117	0.00
54	P	2,4-Dinitrophenol	0.058	0.082	-41.4#	148	0.00
55		Dibenzofuran	1.709	1.759	-2.9	104	0.00
56	P	4-Nitrophenol	0.201	0.251	-24.9	120	0.00
57		2,4-Dinitrotoluene	0.239	0.348	-45.6#	126	0.00
58		Fluorene	1.299	1.322	-1.8	103	0.00
59		2,3,4,6-Tetrachlorophenol	0.304	0.347	-14.1	107	0.00
60		Diethylphthalate	1.263	1.334	-5.6	104	0.00
61		4-Chlorophenyl-phenylether	0.617	0.627	-1.6	102	0.00
62		4-Nitroaniline	0.238	0.305	-28.2#	115	0.00
63		Azobenzene	1.248	1.291	-3.4	104	0.00
64	I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
65		4,6-Dinitro-2-methylphenol	0.054	0.076	-40.7#	142	0.00
66	c	n-Nitrosodiphenylamine	0.683	0.714	-4.5	103	0.00
67		4-Bromophenyl-phenylether	0.223	0.237	-6.3	103	0.00
68		Hexachlorobenzene	0.249	0.261	-4.8	103	0.00
69		Atrazine	0.180	0.201	-11.7	109	0.00
70	C	Pentachlorophenol	0.132	0.156	-18.2	111	0.00
71		Phenanthrene	1.080	1.106	-2.4	102	0.00
72		Anthracene	1.105	1.156	-4.6	105	0.00
73		Carbazole	1.008	1.053	-4.5	105	0.00
74		Di-n-butylphthalate	1.013	1.116	-10.2	107	0.00
75	C	Fluoranthene	1.068	1.113	-4.2	105	0.00
76	I	Chrysene-d12	1.000	1.000	0.0	104	0.00
77		Benzidine	0.850	0.987	-16.1	107	0.00
78		Pyrene	1.845	2.033	-10.2	105	0.00
79	S	Terphenyl-d14	1.375	1.473	-7.1	104	0.00
80		Butylbenzylphthalate	0.425	0.537	-26.4#	111	0.00
81		Benzo(a)anthracene	1.343	1.389	-3.4	101	0.00
82		3,3'-Dichlorobenzidine	0.394	0.449	-14.0	106	0.00
83		Chrysene	1.222	1.288	-5.4	109	0.00
84		Bis(2-ethylhexyl)phthalate	0.604	0.704	-16.6	106	0.00
85	c	Di-n-octyl phthalate	1.133	1.287	-13.6	106	0.00
86	I	Perylene-d12	1.000	1.000	0.0	104	0.00
87		Indeno(1,2,3-cd)pyrene	1.465	1.603	-9.4	104	0.00
88		Benzo(b)fluoranthene	1.242	1.255	-1.0	101	0.00
89		Benzo(k)fluoranthene	1.136	1.184	-4.2	108	0.00
90	C	Benzo(a)pyrene	1.137	1.205	-6.0	105	0.00
91		Dibenzo(a,h)anthracene	1.189	1.295	-8.9	103	0.00
92		Benzo(g,h,i)perylene	1.195	1.312	-9.8	105	-0.01

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142247.D
Acq On : 30 Apr 2025 16:17
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Quant Time: Apr 30 16:36:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 1

Evaluate Continuing Calibration Report

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142247.D
Acq On : 30 Apr 2025 16:17
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Quant Time: Apr 30 16:36:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	103	0.00
2		1,4-Dioxane	40.000	42.438	-6.1	103	0.00
3		Pyridine	40.000	42.469	-6.2	104	0.00
4		n-Nitrosodimethylamine	40.000	42.942	-7.4	103	0.00
5	S	2-Fluorophenol	80.000	85.116	-6.4	104	0.00
6		Aniline	40.000	42.598	-6.5	105	0.00
7	S	Phenol-d6	80.000	85.730	-7.2	105	0.00
8		2-Chlorophenol	40.000	44.202	-10.5	107	0.00
9		Benzaldehyde	40.000	42.925	-7.3	105	0.00
10	C	Phenol	40.000	42.765	-6.9	105	0.00
11		bis(2-Chloroethyl)ether	40.000	42.329	-5.8	104	0.00
12		1,3-Dichlorobenzene	40.000	41.819	-4.5	103	0.00
13	C	1,4-Dichlorobenzene	40.000	41.924	-4.8	103	0.00
14		1,2-Dichlorobenzene	40.000	42.247	-5.6	104	0.00
15		Benzyl Alcohol	40.000	43.378	-8.4	104	0.00
16		2,2'-oxybis(1-Chloropropane	40.000	42.684	-6.7	105	0.00
17		2-Methylphenol	40.000	42.288	-5.7	103	0.00
18		Hexachloroethane	40.000	44.255	-10.6	108	0.00
19	P	n-Nitroso-di-n-propylamine	40.000	41.992	-5.0	104	0.00
20		3+4-Methylphenols	40.000	43.919	-9.8	107	0.00
21	I	Naphthalene-d8	20.000	20.000	0.0	103	0.00
22		Acetophenone	40.000	41.441	-3.6	105	0.00
23	S	Nitrobenzene-d5	80.000	98.845	-23.6	114	0.00
24		Nitrobenzene	40.000	47.402	-18.5	111	0.00
25		Isophorone	40.000	41.789	-4.5	104	0.00
26	C	2-Nitrophenol	40.000	48.175	-20.4#	135	0.00
27		2,4-Dimethylphenol	40.000	43.004	-7.5	106	0.00
28		bis(2-Chloroethoxy)methane	40.000	41.780	-4.5	104	0.00
29	C	2,4-Dichlorophenol	40.000	44.013	-10.0	105	0.00
30		1,2,4-Trichlorobenzene	40.000	41.724	-4.3	103	0.00
31		Naphthalene	40.000	41.005	-2.5	103	0.00
32		Benzoic acid	40.000	45.113	-12.8	134	0.00
33		4-Chloroaniline	40.000	41.344	-3.4	103	0.00
34	C	Hexachlorobutadiene	40.000	42.127	-5.3	103	0.00
35		Caprolactam	40.000	44.434	-11.1	106	0.00
36	C	4-Chloro-3-methylphenol	40.000	43.485	-8.7	105	0.00
37		2-Methylnaphthalene	40.000	41.155	-2.9	104	0.00
38		1-Methylnaphthalene	40.000	41.232	-3.1	103	0.00
39	I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40		1,2,4,5-Tetrachlorobenzene	40.000	41.349	-3.4	104	0.00
41	P	Hexachlorocyclopentadiene	40.000	47.321	-18.3	110	0.00
42	S	2,4,6-Tribromophenol	80.000	93.044	-16.3	107	0.00
43	C	2,4,6-Trichlorophenol	40.000	45.713	-14.3	111	0.00
44		2,4,5-Trichlorophenol	40.000	44.676	-11.7	104	0.00
45	S	2-Fluorobiphenyl	80.000	79.807	0.2	103	0.00
46		1,1'-Biphenyl	40.000	40.936	-2.3	103	0.00
47		2-Chloronaphthalene	40.000	41.003	-2.5	103	0.00

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142247.D
Acq On : 30 Apr 2025 16:17
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Quant Time: Apr 30 16:36:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	46.523	-16.3	123	0.00
49		Acenaphthylene	40.000	41.211	-3.0	103	0.00
50		Dimethylphthalate	40.000	42.054	-5.1	105	0.00
51		2,6-Dinitrotoluene	40.000	45.376	-13.4	118	0.00
52	C	Acenaphthene	40.000	41.631	-4.1	104	0.00
53		3-Nitroaniline	40.000	45.509	-13.8	117	0.00
54	P	2,4-Dinitrophenol	40.000	49.090	-22.7	148	0.00
55		Dibenzofuran	40.000	41.181	-3.0	104	0.00
56	P	4-Nitrophenol	40.000	49.952	-24.9	120	0.00
57		2,4-Dinitrotoluene	40.000	46.994	-17.5	126	0.00
58		Fluorene	40.000	40.714	-1.8	103	0.00
59		2,3,4,6-Tetrachlorophenol	40.000	45.724	-14.3	107	0.00
60		Diethylphthalate	40.000	42.222	-5.6	104	0.00
61		4-Chlorophenyl-phenylether	40.000	40.626	-1.6	102	0.00
62		4-Nitroaniline	40.000	45.070	-12.7	115	0.00
63		Azobenzene	40.000	41.400	-3.5	104	0.00
64	I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65		4,6-Dinitro-2-methylphenol	40.000	46.598	-16.5	142	0.00
66	c	n-Nitrosodiphenylamine	40.000	41.816	-4.5	103	0.00
67		4-Bromophenyl-phenylether	40.000	42.497	-6.2	103	0.00
68		Hexachlorobenzene	40.000	41.800	-4.5	103	0.00
69		Atrazine	40.000	44.732	-11.8	109	0.00
70	C	Pentachlorophenol	40.000	47.342	-18.4	111	0.00
71		Phenanthrene	40.000	40.955	-2.4	102	0.00
72		Anthracene	40.000	41.854	-4.6	105	0.00
73		Carbazole	40.000	41.793	-4.5	105	0.00
74		Di-n-butylphthalate	40.000	44.051	-10.1	107	0.00
75	C	Fluoranthene	40.000	41.690	-4.2	105	0.00
76	I	Chrysene-d12	20.000	20.000	0.0	104	0.00
77		Benzidine	40.000	46.435	-16.1	107	0.00
78		Pyrene	40.000	44.082	-10.2	105	0.00
79	S	Terphenyl-d14	80.000	85.716	-7.1	104	0.00
80		Butylbenzylphthalate	40.000	42.984	-7.5	111	0.00
81		Benzo(a)anthracene	40.000	41.383	-3.5	101	0.00
82		3,3'-Dichlorobenzidine	40.000	45.495	-13.7	106	0.00
83		Chrysene	40.000	42.158	-5.4	109	0.00
84		Bis(2-ethylhexyl)phthalate	40.000	40.882	-2.2	106	0.00
85	c	Di-n-octyl phthalate	40.000	41.178	-2.9	106	0.00
86	I	Perylene-d12	20.000	20.000	0.0	104	0.00
87		Indeno(1,2,3-cd)pyrene	40.000	43.757	-9.4	104	0.00
88		Benzo(b)fluoranthene	40.000	40.403	-1.0	101	0.00
89		Benzo(k)fluoranthene	40.000	41.665	-4.2	108	0.00
90	C	Benzo(a)pyrene	40.000	42.380	-6.0	105	0.00
91		Dibenzo(a,h)anthracene	40.000	43.565	-8.9	103	0.00
92		Benzo(g,h,i)perylene	40.000	43.902	-9.8	105	-0.01

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142247.D
Acq On : 30 Apr 2025 16:17
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Quant Time: Apr 30 16:36:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 1

Raw Data

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG No.: Q1901

Instrument ID: BNA_P Calibration Date(s): 04/14/2025 04/14/2025

Calibration Time(s): 11:06 17:13

LAB FILE ID: RRF2.5 = BP024275.D RRF005 = BP024276.D RRF010 = BP024277.D
RRF020 = BP024278.D RRF040 = BP024279.D RRF050 = BP024280.D

COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Pyridine		1.265	1.280	1.422	1.396	1.333	1.351	4.4
2-Fluorophenol		1.123	1.145	1.262	1.254	1.210	1.210	4.9
Phenol-d6		1.463	1.552	1.718	1.748	1.697	1.656	6.7
1,4-Dichlorobenzene		1.496	1.489	1.550	1.494	1.436	1.475	3.3
2-Methylphenol		0.915	1.002	1.123	1.140	1.105	1.075	8.0
3+4-Methylphenols		1.224	1.365	1.560	1.592	1.567	1.491	9.6
Nitrobenzene-d5		0.322	0.337	0.367	0.368	0.357	0.351	4.8
Hexachloroethane		0.532	0.537	0.558	0.542	0.521	0.533	2.9
Nitrobenzene		0.328	0.334	0.362	0.364	0.349	0.347	4.0
Hexachlorobutadiene		0.181	0.181	0.186	0.189	0.181	0.183	1.6
2,4,6-Trichlorophenol		0.301	0.335	0.369	0.389	0.386	0.366	9.5
2-Fluorobiphenyl		1.332	1.334	1.359	1.366	1.263	1.306	4.1
2,4,5-Trichlorophenol		0.332	0.371	0.409	0.427	0.429	0.405	9.6
2,4-Dinitrotoluene		0.334	0.368	0.439	0.448	0.442	0.421	11.8
2,4,6-Tribromophenol		0.235	0.256	0.278	0.289	0.289	0.277	8.3
Hexachlorobenzene		0.247	0.247	0.260	0.271	0.261	0.260	3.7
Pentachlorophenol		0.143	0.155	0.175	0.196	0.197	0.181	13.3
Terphenyl-d14		0.967	0.998	1.055	1.032	1.005	0.992	4.4

All other compounds must meet a minimum RRF of 0.010.

Response Factor Report MSBNA_P

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP041425.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Apr 15 04:48:42 2025
Response Via : Initial Calibration

Calibration Files
2.5 =BP024275.D 5 =BP024276.D 10 =BP024277.D 20 =BP024278.D 40 =BP024279.D 50 =BP024280.D 60 =BP024281.D 80 =BP0242
82.D

		Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1)	I	1,4-Dichlorobenzen...	----------ISTD----------									
2)		1,4-Dioxane		0.551	0.517	0.552	0.512	0.481	0.490	0.477	0.512	6.10
3)		Pyridine		1.265	1.280	1.422	1.396	1.333	1.383	1.378	1.351	4.43
4)		n-Nitrosodimet...		0.424	0.443	0.469	0.457	0.442	0.460	0.444	0.448	3.29
5)	S	2-Fluorophenol		1.123	1.145	1.262	1.254	1.210	1.277	1.196	1.210	4.93
6)		Aniline		1.439	1.487	1.636	1.587	1.403	1.428	1.380	1.480	6.52
7)	S	Phenol-d6		1.463	1.552	1.718	1.748	1.697	1.769	1.647	1.656	6.75
8)		2-Chlorophenol		1.260	1.297	1.412	1.399	1.359	1.400	1.328	1.350	4.30
9)		Benzaldehyde		0.943	0.934	0.942	0.836	0.732	0.726	0.621	0.819	15.70
10)	C	Phenol		1.475	1.565	1.720	1.750	1.698	1.779	1.643	1.661	6.54
11)		bis(2-Chloroet...		1.278	1.317	1.399	1.398	1.321	1.358	1.300	1.339	3.56
12)		1,3-Dichlorobe...		1.464	1.486	1.534	1.486	1.402	1.433	1.373	1.454	3.79
13)	C	1,4-Dichlorobe...		1.496	1.489	1.550	1.494	1.436	1.463	1.399	1.475	3.28
14)		1,2-Dichlorobe...		1.472	1.441	1.502	1.445	1.362	1.411	1.339	1.424	4.10
15)		Benzyl Alcohol		0.857	0.935	1.106	1.158	1.142	1.178	1.122	1.071	11.58
16)		2,2'-oxybis(1-...		1.452	1.496	1.517	1.514	1.405	1.397	1.348	1.447	4.54
17)		2-Methylphenol		0.915	1.002	1.123	1.140	1.105	1.150	1.087	1.075	7.97
18)		Hexachloroethane		0.532	0.537	0.558	0.542	0.521	0.533	0.510	0.533	2.88
19)	P	n-Nitroso-di-n...	0.941	1.008	1.008	1.088	1.084	1.041	1.038	0.989	1.025	4.76
20)		3+4-Methylphenols		1.224	1.365	1.560	1.592	1.567	1.607	1.525	1.491	9.57
21)	I	Naphthalene-d8	----------ISTD----------									
22)		Acetophenone		0.470	0.487	0.522	0.519	0.490	0.495	0.483	0.495	3.85
23)	S	Nitrobenzene-d5		0.322	0.337	0.367	0.368	0.357	0.356	0.347	0.351	4.76
24)		Nitrobenzene		0.328	0.334	0.362	0.364	0.349	0.353	0.340	0.347	3.96
25)		Isophorone		0.575	0.601	0.653	0.673	0.650	0.654	0.639	0.635	5.41
26)	C	2-Nitrophenol		0.130	0.143	0.172	0.183	0.184	0.191	0.186	0.170	13.97
27)		2,4-Dimethylph...		0.180	0.194	0.219	0.226	0.225	0.228	0.221	0.213	8.81
28)		bis(2-Chloroet...		0.412	0.426	0.444	0.452	0.430	0.425	0.412	0.429	3.51
29)	C	2,4-Dichloroph...		0.242	0.268	0.297	0.308	0.307	0.311	0.302	0.291	8.94
30)		1,2,4-Trichlor...		0.308	0.305	0.318	0.321	0.309	0.315	0.306	0.311	2.00
31)		Naphthalene		1.044	1.054	1.091	1.079	1.037	1.048	1.023	1.054	2.27
32)		Benzoic acid			0.185	0.226	0.242	0.270	0.285	0.282	0.248	15.56
33)		4-Chloroaniline		0.328	0.349	0.389	0.397	0.378	0.379	0.378	0.371	6.46
34)	C	Hexachlorobuta...		0.181	0.181	0.186	0.189	0.181	0.182	0.182	0.183	1.62
35)		Caprolactam		0.085	0.092	0.111	0.112	0.115	0.121	0.115	0.107	12.53
36)	C	4-Chloro-3-met...		0.284	0.310	0.349	0.357	0.356	0.363	0.351	0.339	8.82
37)		2-Methylnaphth...		0.700	0.720	0.755	0.755	0.733	0.740	0.712	0.731	2.89
38)		1-Methylnaphth...		0.700	0.714	0.737	0.731	0.712	0.718	0.691	0.715	2.25

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP041425.M

39)	I	Acenaphthene-d10	----------ISTD----------								
40)		1,2,4,5-Tetrac...	0.522	0.538	0.568	0.576	0.544	0.549	0.552	0.550	3.33
41)	P	Hexachlorocycl...	0.162	0.185	0.197	0.218	0.199	0.197	0.204	0.195	8.92
42)	S	2,4,6-Tribromo...	0.235	0.256	0.278	0.289	0.289	0.296	0.294	0.277	8.29
43)	C	2,4,6-Trichlor...	0.301	0.335	0.369	0.389	0.386	0.390	0.390	0.366	9.51
44)		2,4,5-Trichlor...	0.332	0.371	0.409	0.427	0.429	0.436	0.430	0.405	9.62
45)	S	2-Fluorobiphenyl	1.332	1.334	1.359	1.366	1.263	1.251	1.240	1.306	4.08
46)		1,1'-Biphenyl	1.454	1.489	1.528	1.529	1.434	1.428	1.429	1.470	3.07
47)		2-Chloronaphth...	1.063	1.100	1.135	1.139	1.085	1.090	1.078	1.099	2.60
48)		2-Nitroaniline	0.250	0.270	0.317	0.330	0.324	0.335	0.331	0.308	10.98
49)		Acenaphthylene	1.604	1.680	1.793	1.814	1.726	1.763	1.730	1.730	4.11
50)		Dimethylphthalate	1.429	1.416	1.495	1.495	1.460	1.460	1.426	1.454	2.22
51)		2,6-Dinitrotol...	0.264	0.290	0.319	0.325	0.319	0.328	0.317	0.309	7.58
52)	C	Acenaphthene	1.087	1.087	1.141	1.126	1.081	1.079	1.076	1.097	2.35
53)		3-Nitroaniline	0.250	0.289	0.338	0.354	0.342	0.340	0.358	0.325	12.25
54)	P	2,4-Dinitrophenol		0.125	0.163	0.186	0.196	0.210	0.211	0.182	18.22
55)		Dibenzofuran	1.801	1.807	1.871	1.833	1.746	1.768	1.723	1.793	2.84
56)	P	4-Nitrophenol	0.190	0.231	0.289	0.305	0.306	0.323	0.317	0.280	17.90
57)		2,4-Dinitrotol...	0.334	0.368	0.439	0.448	0.442	0.463	0.456	0.421	11.82
58)		Fluorene	1.376	1.397	1.454	1.398	1.368	1.366	1.356	1.388	2.40
59)		2,3,4,6-Tetrac...	0.327	0.351	0.380	0.389	0.387	0.390	0.389	0.373	6.57
60)		Diethylphthalate	1.451	1.484	1.528	1.550	1.499	1.499	1.481	1.499	2.16
61)		4-Chlorophenyl...	0.667	0.675	0.699	0.680	0.667	0.670	0.659	0.674	1.95
62)		4-Nitroaniline	0.263	0.300	0.353	0.369	0.364	0.380	0.376	0.344	12.99
63)		Azobenzene	1.269	1.383	1.446	1.431	1.385	1.385	1.344	1.378	4.26
64)	I	Phenanthrene-d10	----------ISTD----------								
65)		4,6-Dinitro-2-...		0.095	0.118	0.131	0.134	0.140	0.138	0.126	13.50
66)	c	n-Nitrosodiphe...	0.567	0.585	0.616	0.630	0.591	0.608	0.580	0.597	3.71
67)		4-Bromophenyl-...	0.203	0.212	0.220	0.230	0.220	0.225	0.222	0.219	4.11
68)		Hexachlorobenzene	0.247	0.247	0.260	0.271	0.261	0.269	0.263	0.260	3.67
69)		Atrazine	0.176	0.167	0.134	0.121	0.162			0.152	15.57
70)	C	Pentachlorophenol	0.143	0.155	0.175	0.196	0.197	0.203	0.201	0.181	13.35
71)		Phenanthrene	1.098	1.086	1.117	1.127	1.070	1.082	1.058	1.091	2.27
72)		Anthracene	0.995	1.017	1.084	1.111	1.053	1.071	1.030	1.052	3.85
73)		Carbazole	0.968	0.998	1.060	1.084	1.014	1.037	1.031	1.027	3.78
74)		Di-n-butylphth...	1.147	1.230	1.253	1.398	1.310	1.338	1.284	1.280	6.30
75)	C	Fluoranthene	1.289	1.274	1.316	1.347	1.280	1.304	1.274	1.298	2.07
76)	I	Chrysene-d12	----------ISTD----------								
77)		Benzidine	0.144	0.141	0.109	0.436	0.333	0.302	0.307	0.254	48.39
78)		Pyrene	1.192	1.236	1.345	1.305	1.294	1.305	1.219	1.271	4.38
79)	S	Terphenyl-d14	0.967	0.998	1.055	1.032	1.005	0.964	0.926	0.992	4.42
80)		Butylbenzylpht...	0.434	0.488	0.546	0.596	0.580	0.593	0.574	0.544	11.25
81)		Benzo(a)anthra...	1.195	1.214	1.292	1.293	1.234	1.253	1.221	1.243	3.08
82)		3,3'-Dichlorob...	0.345	0.383	0.431	0.486	0.463	0.485	0.460	0.436	12.30
83)		Chrysene	1.180	1.181	1.229	1.213	1.173	1.193	1.167	1.191	1.88
84)		Bis(2-ethylhex...	0.639	0.735	0.779	0.893	0.853	0.867	0.819	0.798	11.09
85)	c	Di-n-octyl pht...	0.945	1.093	1.222	1.461	1.430	1.486	1.446	1.297	16.46

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP041425.M

86)	I	Perylene-d12	----------ISTD----------							
87)		Indeno(1,2,3-c...	1.252	1.304	1.407	1.444	1.401	1.445	1.466	1.388	5.77
88)		Benzo(b)fluora...	1.130	1.174	1.217	1.258	1.196	1.205	1.212	1.199	3.30
89)		Benzo(k)fluora...	1.088	1.131	1.217	1.212	1.154	1.166	1.138	1.158	3.94
90)	C	Benzo(a)pyrene	0.939	0.971	1.057	1.091	1.053	1.070	1.058	1.034	5.44
91)		Dibenzo(a,h)an...	1.046	1.089	1.172	1.203	1.169	1.186	1.202	1.152	5.28
92)		Benzo(g,h,i)pe...	1.073	1.110	1.198	1.218	1.180	1.209	1.224	1.173	4.99

(#) = Out of Range

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024275.D
Acq On : 14 Apr 2025 11:06
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Quant Time: Apr 14 17:26:12 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.734	152	229109	20.000	ng	0.00
21) Naphthalene-d8	10.505	136	962244	20.000	ng	0.00
39) Acenaphthene-d10	14.363	164	614109	20.000	ng	0.00
64) Phenanthrene-d10	17.181	188	1238206	20.000	ng	0.02
76) Chrysene-d12	21.616	240	1367665	20.000	ng	0.00
86) Perylene-d12	24.980	264	1473609	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						Qvalue
19) n-Nitroso-di-n-propyla...	8.528	70	26959	2.285	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed


ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024275.D
Acq On : 14 Apr 2025 11:06
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Quant Time: Apr 14 17:26:12 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.734 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion:152 Resp: 229109

Ion	Ratio	Lower	Upper
152	100		
150	155.3	124.9	187.3
115	60.4	47.7	71.5







#5
2-Fluorophenol
Concen: 0.000 ng
Expected RT: 5.35 min

Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Tgt Ion: 112

Sig	Exp Ratio
112	100
64	51.4
63	25.7






#7
Phenol-d6
Concen: 0.000 ng
Expected RT: 6.92 min

Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion: 99


Sig	Exp Ratio
99	100
42	13.3
71	29.6




#19
n-Nitroso-di-n-propylamine
Concen: 2.285 ng
RT: 8.528 min Scan# 925
Delta R.T. 0.000 min
Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Tgt Ion: 70 Resp: 26959

Ion	Ratio	Lower	Upper
70	100		
42	50.7	37.3	55.9
101	11.7	8.2	12.4
130	23.9	18.6	28.0







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.505 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion:136 Resp: 962244

Ion	Ratio	Lower	Upper
136	100		
137	10.9	9.2	13.8
54	7.2	6.0	9.0
68	6.7	5.5	8.3







#23
Nitrobenzene-d5
Concen: 0.000 ng
Expected RT: 8.88 min

Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Tgt Ion: 82

Sig	Exp Ratio
82	100
128	45.6
54	41.8







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.363 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion:164 Resp: 614109

Ion	Ratio	Lower	Upper
164	100		
162	100.1	80.6	120.8
160	44.5	35.3	52.9







#42
2,4,6-Tribromophenol
Concen: 0.000 ng
Expected RT: 15.88 min

Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Tgt Ion: 330

Sig	Exp Ratio
330	100
332	96.4
141	30.9






#45
2-Fluorobiphenyl
Concen: 0.000 ng
Expected RT: 12.97 min

Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion: 172


Sig	Exp Ratio
172	100
171	35.2
170	23.2




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.181 min Scan# 2396
Delta R.T. 0.018 min
Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Tgt Ion:188 Resp: 1238206

Ion	Ratio	Lower	Upper
188	100		
94	10.5	8.6	13.0
80	11.3	9.8	14.6







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.616 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion:240 Resp: 1367665

Ion	Ratio	Lower	Upper
240	100		
120	12.7	9.8	14.8
236	25.9	20.9	31.3







#79
Terphenyl-d14
Concen: 0.000 ng
Expected RT: 19.90 min

Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Tgt Ion: 244

Sig	Exp Ratio
244	100
212	7.0
122	11.7






#86
Perylene-d12
Concen: 20.000 ng
RT: 24.980 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion:264 Resp: 1473609

Ion	Ratio	Lower	Upper
264	100		
260	23.6	18.9	28.3
265	22.3	17.8	26.6



ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024276.D
Acq On : 14 Apr 2025 11:47
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Quant Time: Apr 14 17:26:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.734	152	223010	20.000	ng		0.00
21) Naphthalene-d8	10.510	136	946483	20.000	ng		0.00
39) Acenaphthene-d10	14.369	164	609743	20.000	ng		0.00
64) Phenanthrene-d10	17.163	188	1234074	20.000	ng		0.00
76) Chrysene-d12	21.610	240	1401979	20.000	ng		-0.01
86) Perylene-d12	24.968	264	1519057	20.000	ng		-0.02
System Monitoring Compounds							
5) 2-Fluorophenol	5.352	112	125173	9.263	ng		0.00
7) Phenol-d6	6.916	99	163122	8.824	ng		0.00
23) Nitrobenzene-d5	8.875	82	152286	9.160	ng		0.00
42) 2,4,6-Tribromophenol	15.875	330	71560	8.571	ng		0.00
45) 2-Fluorobiphenyl	12.975	172	406060	10.111	ng		0.00
79) Terphenyl-d14	19.886	244	677812	9.638	ng		-0.01
Target Compounds							Qvalue
2) 1,4-Dioxane	3.293	88	30738	5.328	ng		100
3) Pyridine	3.699	79	70539	4.697	ng		97
4) n-Nitrosodimethylamine	3.599	42	23634	4.720	ng	#	90
6) Aniline	7.069	93	80239	4.808	ng		99
8) 2-Chlorophenol	7.305	128	70239	4.651	ng		98
9) Benzaldehyde	6.887	77	52595m	5.493	ng		
10) Phenol	6.940	94	82245	4.431	ng		97
11) bis(2-Chloroethyl)ether	7.163	93	71228	4.749	ng		97
12) 1,3-Dichlorobenzene	7.622	146	81600	4.987	ng		98
13) 1,4-Dichlorobenzene	7.769	146	83408	5.027	ng		99
14) 1,2-Dichlorobenzene	8.081	146	82066	5.115	ng		98
15) Benzyl Alcohol	7.975	79	47777	4.032	ng		98
16) 2,2'-oxybis(1-Chloropr...	8.258	45	80975	4.962	ng		97
17) 2-Methylphenol	8.175	107	51039	4.267	ng		99
18) Hexachloroethane	8.805	117	29643	4.950	ng		99
19) n-Nitroso-di-n-propyla...	8.528	70	56212	4.896	ng		96
20) 3+4-Methylphenols	8.505	107	68264	4.120	ng		99
22) Acetophenone	8.546	105	111159	4.725	ng		99
24) Nitrobenzene	8.922	77	77593	4.709	ng		96
25) Isophorone	9.446	82	136090	4.534	ng		99
26) 2-Nitrophenol	9.628	139	30712	3.882	ng		99
27) 2,4-Dimethylphenol	9.693	122	42510	4.239	ng		94
28) bis(2-Chloroethoxy)met...	9.922	93	97512	4.774	ng		98
29) 2,4-Dichlorophenol	10.169	162	57301	4.193	ng		97
30) 1,2,4-Trichlorobenzene	10.369	180	72789	4.923	ng		98
31) Naphthalene	10.557	128	246999	4.930	ng		99
33) 4-Chloroaniline	10.669	127	77667	4.437	ng		98
34) Hexachlorobutadiene	10.846	225	42723	4.926	ng		99
35) Caprolactam	11.446	113	20082	4.002	ng		94
36) 4-Chloro-3-methylphenol	11.804	107	67116	4.214	ng		100
37) 2-Methylnaphthalene	12.169	142	165617	4.770	ng		99
38) 1-Methylnaphthalene	12.399	142	165656	4.870	ng		99
40) 1,2,4,5-Tetrachloroben...	12.540	216	79501	4.744	ng		100
41) Hexachlorocyclopentadiene	12.528	237	24696	4.198	ng		96
43) 2,4,6-Trichlorophenol	12.787	196	45925	4.165	ng		95
44) 2,4,5-Trichlorophenol	12.863	196	50675	4.147	ng		99

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024276.D
Acq On : 14 Apr 2025 11:47
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Quant Time: Apr 14 17:26:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration


	Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
46)	1,1'-Biphenyl	13.187	154	221653	4.922	ng		99
47)	2-Chloronaphthalene	13.234	162	162099	4.824	ng		99
48)	2-Nitroaniline	13.440	65	38152	4.111	ng		90
49)	Acenaphthylene	14.087	152	244566	4.637	ng		99
50)	Dimethylphthalate	13.822	163	217848	4.897	ng		100
51)	2,6-Dinitrotoluene	13.934	165	40182	4.287	ng		94
52)	Acenaphthene	14.428	154	165694	4.940	ng		97
53)	3-Nitroaniline	14.275	138	38178	3.926	ng	#	94
55)	Dibenzofuran	14.769	168	274506	4.991	ng		99
56)	4-Nitrophenol	14.610	139	29033	3.473	ng		98
57)	2,4-Dinitrotoluene	14.745	165	50843	4.013	ng		97
58)	Fluorene	15.428	166	209763	4.939	ng		97
59)	2,3,4,6-Tetrachlorophenol	15.004	232	49911	4.415	ng		98
60)	Diethylphthalate	15.204	149	221119	4.830	ng		99
61)	4-Chlorophenyl-phenyle...	15.422	204	101636	4.928	ng		99
62)	4-Nitroaniline	15.445	138	40139	3.893	ng		96
63)	Azobenzene	15.722	77	193402	4.586	ng		99
66)	n-Nitrosodiphenylamine	15.645	169	174867	4.726	ng		99
67)	4-Bromophenyl-phenylether	16.334	248	62558	4.647	ng		97
68)	Hexachlorobenzene	16.451	284	76267	4.766	ng		96
69)	Atrazine	16.616	200	54439	5.803	ng		98
70)	Pentachlorophenol	16.810	266	44119	4.015	ng		98
71)	Phenanthrene	17.204	178	338611	5.005	ng		100
72)	Anthracene	17.292	178	307065	4.716	ng		99
73)	Carbazole	17.586	167	298495	4.711	ng		100
74)	Di-n-butylphthalate	18.175	149	354008	4.484	ng		100
75)	Fluoranthene	19.292	202	397601	4.951	ng		99
77)	Benzidine	19.498	184	50615m	2.925	ng		
78)	Pyrene	19.663	202	417746	4.657	ng		99
80)	Butylbenzylphthalate	20.622	149	152196	4.024	ng		99
81)	Benzo(a)anthracene	21.592	228	418745	4.791	ng		100
82)	3,3'-Dichlorobenzidine	21.516	252	121047	3.994	ng		99
83)	Chrysene	21.657	228	413570	4.937	ng		99
84)	Bis(2-ethylhexyl)phtha...	21.533	149	223960	4.021	ng		98
85)	Di-n-octyl phthalate	22.816	149	331185	3.712	ng		99
87)	Indeno(1,2,3-cd)pyrene	28.762	276	475578	4.552	ng	#	92
88)	Benzo(b)fluoranthene	23.904	252	429214	4.723	ng		99
89)	Benzo(k)fluoranthene	23.968	252	413302	4.686	ng		99
90)	Benzo(a)pyrene	24.810	252	356658	4.559	ng		98
91)	Dibenzo(a,h)anthracene	28.845	278	397247	4.571	ng		100
92)	Benzo(g,h,i)perylene	29.951	276	407522	4.607	ng		99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024276.D
Acq On : 14 Apr 2025 11:47
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Quant Time: Apr 14 17:26:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.734 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:152 Resp: 223010

Ion	Ratio	Lower	Upper
152	100		
150	153.5	124.9	187.3
115	59.8	47.7	71.5







#2
1,4-Dioxane
Concen: 5.328 ng
RT: 3.293 min Scan# 35
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion: 88 Resp: 30738

Ion	Ratio	Lower	Upper
88	100		
58	64.3	51.4	77.2
43	22.3	17.1	25.7







#3
Pyridine
Concen: 4.697 ng
RT: 3.699 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 79 Resp: 70539

Ion	Ratio	Lower	Upper
79	100		
52	56.9	45.4	68.2
51	32.8	22.6	33.8







#4
n-Nitrosodimethylamine
Concen: 4.720 ng
RT: 3.599 min Scan# 87
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion: 42 Resp: 23634

Ion	Ratio	Lower	Upper
42	100		
74	161.0	139.5	209.3
44	17.2	5.8	8.8#







#5
2-Fluorophenol
Concen: 9.263 ng
RT: 5.352 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:112 Resp: 125173

Ion	Ratio	Lower	Upper
112	100		
64	51.4	41.1	61.7
63	27.1	20.6	30.8







#6
Aniline
Concen: 4.808 ng
RT: 7.069 min Scan# 677
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion: 93 Resp: 80239

Ion	Ratio	Lower	Upper
93	100		
66	36.8	29.1	43.7
65	18.0	14.7	22.1







#7
Phenol-d6
Concen: 8.824 ng
RT: 6.916 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 99 Resp: 163122

Ion	Ratio	Lower	Upper
99	100		
42	14.7	10.6	16.0
71	30.7	23.7	35.5







#8
2-Chlorophenol
Concen: 4.651 ng
RT: 7.305 min Scan# 717
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:128 Resp: 70239

Ion	Ratio	Lower	Upper
128	100		
130	33.5	12.9	52.9
64	44.2	22.7	62.7







#9
Benzaldehyde
Concen: 5.493 ng m
RT: 6.887 min Scan# 64
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 77 Resp: 52595

Ion	Ratio	Lower	Upper
77	100		
105	100.6	79.2	119.2
106	97.4	76.2	116.2







#10
Phenol
Concen: 4.431 ng
RT: 6.940 min Scan# 655
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion: 94 Resp: 82245

Ion	Ratio	Lower	Upper
94	100		
65	26.3	4.5	44.5
66	33.7	12.4	52.4







#11
bis(2-Chloroethyl)ether
Concen: 4.749 ng
RT: 7.163 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 93 Resp: 71228

Ion	Ratio	Lower	Upper
93	100		
63	71.6	48.6	88.6
95	32.9	12.3	52.3







#12
1,3-Dichlorobenzene
Concen: 4.987 ng
RT: 7.622 min Scan# 771
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:146 Resp: 81600

Ion	Ratio	Lower	Upper
146	100		
148	62.9	51.4	77.0
75	29.6	22.9	34.3







#13
1,4-Dichlorobenzene
Concen: 5.027 ng
RT: 7.769 min Scan# 7
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:146 Resp: 83408

Ion	Ratio	Lower	Upper
146	100		
148	64.2	51.1	76.7
111	42.2	32.6	48.8







#14
1,2-Dichlorobenzene
Concen: 5.115 ng
RT: 8.081 min Scan# 849
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:146 Resp: 82066

Ion	Ratio	Lower	Upper
146	100		
148	62.6	51.0	76.4
111	41.2	33.8	50.8







#15
Benzyl Alcohol
Concen: 4.032 ng
RT: 7.975 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 79 Resp: 47777

Ion	Ratio	Lower	Upper
79	100		
108	79.4	65.1	97.7
77	65.4	50.7	76.1







#16
2,2'-oxybis(1-Chloropropane)
Concen: 4.962 ng
RT: 8.258 min Scan# 879
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion: 45 Resp: 80975

Ion	Ratio	Lower	Upper
45	100		
77	19.4	0.0	37.7
79	14.4	0.0	33.5







#17
2-Methylphenol
Concen: 4.267 ng
RT: 8.175 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:107 Resp: 51039

Ion	Ratio	Lower	Upper
107	100		
108	112.3	89.8	134.6
77	44.7	34.2	51.4
79	43.3	33.8	50.8







#18
Hexachloroethane
Concen: 4.950 ng
RT: 8.805 min Scan# 972
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:117 Resp: 29643

Ion	Ratio	Lower	Upper
117	100		
119	96.4	77.8	116.6
201	96.8	75.9	113.9







#19
n-Nitroso-di-n-propylamine
Concen: 4.896 ng
RT: 8.528 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 70 Resp: 56212

Ion	Ratio	Lower	Upper
70	100		
42	43.5	37.3	55.9
101	9.4	8.2	12.4
130	22.0	18.6	28.0







#20
3+4-Methylphenols
Concen: 4.120 ng
RT: 8.505 min Scan# 921
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:107 Resp: 68264

Ion	Ratio	Lower	Upper
107	100		
108	89.1	69.8	109.8
77	30.9	11.5	51.5
79	28.1	7.8	47.8







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.510 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:136 Resp: 946483

Ion	Ratio	Lower	Upper
136	100		
137	10.8	9.2	13.8
54	7.1	6.0	9.0
68	6.5	5.5	8.3







#22
Acetophenone
Concen: 4.725 ng
RT: 8.546 min Scan# 928
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:105 Resp: 111159

Ion	Ratio	Lower	Upper
105	100		
71	1.9	1.7	2.5
51	22.3	16.9	25.3
120	23.5	18.6	28.0







#23
Nitrobenzene-d5
Concen: 9.160 ng
RT: 8.875 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 82 Resp: 152286

Ion	Ratio	Lower	Upper
82	100		
128	43.4	36.5	54.7
54	41.8	33.4	50.2







#24
Nitrobenzene
Concen: 4.709 ng
RT: 8.922 min Scan# 992
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion: 77 Resp: 77593

Ion	Ratio	Lower	Upper
77	100		
123	45.4	39.2	58.8
65	12.3	9.8	14.6







#25
Isophorone
Concen: 4.534 ng
RT: 9.446 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 82 Resp: 136090

Ion	Ratio	Lower	Upper
82	100		
95	6.8	5.4	8.2
138	17.6	14.4	21.6







#26
2-Nitrophenol
Concen: 3.882 ng
RT: 9.628 min Scan# 1112
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:139 Resp: 30712

Ion	Ratio	Lower	Upper
139	100		
109	28.1	21.8	32.6
65	45.7	36.2	54.2







#27
2,4-Dimethylphenol
Concen: 4.239 ng
RT: 9.693 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:122 Resp: 42510

Ion	Ratio	Lower	Upper
122	100		
107	109.1	93.7	140.5
121	56.5	48.2	72.2







#28
bis(2-Chloroethoxy)methane
Concen: 4.774 ng
RT: 9.922 min Scan# 1162
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion: 93 Resp: 97512

Ion	Ratio	Lower	Upper
93	100		
95	34.0	26.0	39.0
123	11.2	9.1	13.7







#29
2,4-Dichlorophenol
Concen: 4.193 ng
RT: 10.169 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:162 Resp: 57301

Ion	Ratio	Lower	Upper
162	100		
164	64.2	43.3	83.3
98	41.0	16.6	56.6







#30
1,2,4-Trichlorobenzene
Concen: 4.923 ng
RT: 10.369 min Scan# 1238
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:180 Resp: 72789

Ion	Ratio	Lower	Upper
180	100		
182	95.1	77.3	115.9
145	30.3	25.1	37.7







#31
Naphthalene
Concen: 4.930 ng
RT: 10.557 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:128 Resp: 246999

Ion	Ratio	Lower	Upper
128	100		
129	11.5	8.6	13.0
127	13.1	10.3	15.5







#33
4-Chloroaniline
Concen: 4.437 ng
RT: 10.669 min Scan# 1289
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:127 Resp: 77667

Ion	Ratio	Lower	Upper
127	100		
129	31.4	26.1	39.1
65	28.1	22.1	33.1
92	20.0	15.0	22.4







#34
Hexachlorobutadiene
Concen: 4.926 ng
RT: 10.846 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:225 Resp: 42723

Ion	Ratio	Lower	Upper
225	100		
223	62.1	49.5	74.3
227	61.5	50.9	76.3







#35
Caprolactam
Concen: 4.002 ng
RT: 11.446 min Scan# 1421
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:113 Resp: 20082

Ion	Ratio	Lower	Upper
113	100		
55	146.2	115.9	155.9
56	104.7	81.1	121.1







#36
4-Chloro-3-methylphenol
Concen: 4.214 ng
RT: 11.804 min Scan# 14
Delta R.T. 0.012 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:107 Resp: 67116

Ion	Ratio	Lower	Upper
107	100		
144	25.9	20.5	30.7
142	79.3	63.4	95.2







#37
2-Methylnaphthalene
Concen: 4.770 ng
RT: 12.169 min Scan# 1544
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:142 Resp: 165617

Ion	Ratio	Lower	Upper
142	100		
141	89.1	70.3	105.5
115	33.8	26.9	40.3







#38
1-Methylnaphthalene
Concen: 4.870 ng
RT: 12.399 min Scan# 1
Delta R.T. 0.018 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:142 Resp: 165656

Ion	Ratio	Lower	Upper
142	100		
141	91.3	73.6	110.4
116	4.0	3.0	4.6







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.369 min Scan# 1918
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:164 Resp: 609743

Ion	Ratio	Lower	Upper
164	100		
162	100.9	80.6	120.8
160	44.4	35.3	52.9







#40
1,2,4,5-Tetrachlorobenzene
Concen: 4.744 ng
RT: 12.540 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:216 Resp: 79501

Ion	Ratio	Lower	Upper
216	100		
214	78.0	62.6	93.8
179	21.7	17.2	25.8
108	19.0	15.0	22.6







#41
Hexachlorocyclopentadiene
Concen: 4.198 ng
RT: 12.528 min Scan# 1605
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:237 Resp: 24696

Ion	Ratio	Lower	Upper
237	100		
235	60.2	43.9	83.9
272	14.0	0.0	33.6







#42
2,4,6-Tribromophenol
Concen: 8.571 ng
RT: 15.875 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:330 Resp: 71560

Ion	Ratio	Lower	Upper
330	100		
332	99.6	77.1	115.7
141	33.2	24.7	37.1







#43
2,4,6-Trichlorophenol
Concen: 4.165 ng
RT: 12.787 min Scan# 1649
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:196 Resp: 45925

Ion	Ratio	Lower	Upper
196	100		
198	97.9	74.3	111.5
200	32.2	24.0	36.0







#44
2,4,5-Trichlorophenol
Concen: 4.147 ng
RT: 12.863 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:196 Resp: 50675

Ion	Ratio	Lower	Upper
196	100		
198	97.0	77.3	115.9
97	48.1	40.4	60.6
132	28.5	23.4	35.0







#45
2-Fluorobiphenyl
Concen: 10.111 ng
RT: 12.975 min Scan# 1681
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:172 Resp: 406060

Ion	Ratio	Lower	Upper
172	100		
171	35.9	28.2	42.2
170	23.1	18.6	27.8







#46
1,1'-Biphenyl
Concen: 4.922 ng
RT: 13.187 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:154 Resp: 221653

Ion	Ratio	Lower	Upper
154	100		
153	39.5	20.3	60.3
76	13.7	0.0	33.5







#47
2-Chloronaphthalene
Concen: 4.824 ng
RT: 13.234 min Scan# 1725
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:162 Resp: 162099

Ion	Ratio	Lower	Upper
162	100		
127	39.0	31.9	47.9
164	32.5	26.2	39.4







#48
2-Nitroaniline
Concen: 4.111 ng
RT: 13.440 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 65 Resp: 38152

Ion	Ratio	Lower	Upper
65	100		
92	71.5	62.3	93.5
138	107.8	97.2	145.8







#49
Acenaphthylene
Concen: 4.637 ng
RT: 14.087 min Scan# 1870
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:152 Resp: 244566

Ion	Ratio	Lower	Upper
152	100		
151	20.4	15.9	23.9
153	13.2	10.6	15.8







#50
Dimethylphthalate
Concen: 4.897 ng
RT: 13.822 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:163 Resp: 217848

Ion	Ratio	Lower	Upper
163	100		
194	4.3	3.5	5.3
164	9.8	8.0	12.0







#51
2,6-Dinitrotoluene
Concen: 4.287 ng
RT: 13.934 min Scan# 1844
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:165 Resp: 40182

Ion	Ratio	Lower	Upper
165	100		
63	50.3	38.0	57.0
89	57.0	41.7	62.5







#52
Acenaphthene
Concen: 4.940 ng
RT: 14.428 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:154 Resp: 165694

Ion	Ratio	Lower	Upper
154	100		
153	115.2	89.8	134.6
152	54.6	42.6	63.8







#53
3-Nitroaniline
Concen: 3.926 ng
RT: 14.275 min Scan# 1902
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:138 Resp: 38178

Ion	Ratio	Lower	Upper
138	100		
108	12.8	8.3	12.5#
92	118.4	89.8	134.8







#55
Dibenzofuran
Concen: 4.991 ng
RT: 14.769 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:168 Resp: 274506

Ion	Ratio	Lower	Upper
168	100		
139	37.3	30.3	45.5
169	12.9	10.6	16.0







#56
4-Nitrophenol
Concen: 3.473 ng
RT: 14.610 min Scan# 1959
Delta R.T. 0.018 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:139 Resp: 29033

Ion	Ratio	Lower	Upper
139	100		
109	67.8	46.1	86.1
65	83.9	65.0	105.0







#57
2,4-Dinitrotoluene
Concen: 4.013 ng
RT: 14.745 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:165 Resp: 50843

Ion	Ratio	Lower	Upper
165	100		
63	40.8	30.4	45.6
89	69.9	54.6	82.0
182	2.8	2.2	3.2







#58
Fluorene
Concen: 4.939 ng
RT: 15.428 min Scan# 2098
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:166 Resp: 209763

Ion	Ratio	Lower	Upper
166	100		
165	101.0	78.1	117.1
167	13.9	10.9	16.3







#59
2,3,4,6-Tetrachlorophenol
Concen: 4.415 ng
RT: 15.004 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:232 Resp: 49911

Ion	Ratio	Lower	Upper
232	100		
131	48.4	36.8	55.2
130	2.5	1.9	2.9
166	31.3	25.0	37.6







#60
Diethylphthalate
Concen: 4.830 ng
RT: 15.204 min Scan# 2060
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:149 Resp: 221119

Ion	Ratio	Lower	Upper
149	100		
177	21.3	17.4	26.2
150	12.7	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 4.928 ng
RT: 15.422 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:204 Resp: 101636

Ion	Ratio	Lower	Upper
204	100		
206	32.7	26.5	39.7
141	62.0	49.4	74.0







#62
4-Nitroaniline
Concen: 3.893 ng
RT: 15.445 min Scan# 2101
Delta R.T. -0.012 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:138 Resp: 40139

Ion	Ratio	Lower	Upper
138	100		
92	51.4	27.8	67.8
108	71.8	48.9	88.9







#63
Azobenzene
Concen: 4.586 ng
RT: 15.722 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 77 Resp: 193402

Ion	Ratio	Lower	Upper
77	100		
182	26.4	7.3	47.3
105	22.3	2.3	42.3
51	23.2	3.4	43.4







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.163 min Scan# 2393
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:188 Resp: 1234074

Ion	Ratio	Lower	Upper
188	100		
94	11.1	8.6	13.0
80	12.1	9.8	14.6







#66
n-Nitrosodiphenylamine
Concen: 4.726 ng
RT: 15.645 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:169 Resp: 174867

Ion	Ratio	Lower	Upper
169	100		
168	66.8	53.0	79.4
167	35.6	28.9	43.3







#67
4-Bromophenyl-phenylether
Concen: 4.647 ng
RT: 16.334 min Scan# 2252
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:248 Resp: 62558

Ion	Ratio	Lower	Upper
248	100		
250	101.6	78.2	117.4
141	74.2	57.7	86.5







#68
Hexachlorobenzene
Concen: 4.766 ng
RT: 16.451 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:284 Resp: 76267

Ion	Ratio	Lower	Upper
284	100		
142	35.1	24.6	37.0
249	30.1	23.8	35.6







#69
Atrazine
Concen: 5.803 ng
RT: 16.616 min Scan# 2300
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:200 Resp: 54439

Ion	Ratio	Lower	Upper
200	100		
173	25.6	7.5	47.5
215	49.4	28.8	68.8







#70
Pentachlorophenol
Concen: 4.015 ng
RT: 16.810 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:266 Resp: 44119

Ion	Ratio	Lower	Upper
266	100		
268	67.1	51.0	76.6
264	61.9	49.8	74.6







#71
Phenanthrene
Concen: 5.005 ng
RT: 17.204 min Scan# 2400
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:178 Resp: 338611

Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.4	23.0
179	15.5	12.5	18.7







#72
Anthracene
Concen: 4.716 ng
RT: 17.292 min Scan# 24
Delta R.T. -0.012 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:178 Resp: 307065

Ion	Ratio	Lower	Upper
178	100		
176	18.2	14.7	22.1
179	14.8	12.4	18.6







#73
Carbazole
Concen: 4.711 ng
RT: 17.586 min Scan# 2465
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:167 Resp: 298495

Ion	Ratio	Lower	Upper
167	100		
166	21.1	17.1	25.7
139	13.3	10.6	16.0







#74
Di-n-butylphthalate
Concen: 4.484 ng
RT: 18.175 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:149 Resp: 354008

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.4	11.2
104	5.1	3.9	5.9







#75
Fluoranthene
Concen: 4.951 ng
RT: 19.292 min Scan# 2755
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:202 Resp: 397601

Ion	Ratio	Lower	Upper
202	100		
101	12.5	0.0	33.2
203	17.3	0.0	37.7







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.610 min Scan# 3
Delta R.T. -0.012 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:240 Resp: 1401979

Ion	Ratio	Lower	Upper
240	100		
120	12.2	9.8	14.8
236	25.8	20.9	31.3







#77
Benzidine
Concen: 2.925 ng m
RT: 19.498 min Scan# 2790
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:184 Resp: 50615

Ion	Ratio	Lower	Upper
184	100		
185	15.8	11.6	17.4
183	12.4	9.2	13.8







#78
Pyrene
Concen: 4.657 ng
RT: 19.663 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:202 Resp: 417746

Ion	Ratio	Lower	Upper
202	100		
200	21.2	17.2	25.8
203	17.5	14.4	21.6







#79
Terphenyl-d14
Concen: 9.638 ng
RT: 19.886 min Scan# 2856
Delta R.T. -0.012 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:244 Resp: 677812

Ion	Ratio	Lower	Upper
244	100		
212	6.4	5.6	8.4
122	11.7	9.4	14.0







#80
Butylbenzylphthalate
Concen: 4.024 ng
RT: 20.622 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:149 Resp: 152196

Ion	Ratio	Lower	Upper
149	100		
91	69.9	56.4	84.6
206	21.6	16.3	24.5







#81
Benzo(a)anthracene
Concen: 4.791 ng
RT: 21.592 min Scan# 3146
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:228 Resp: 418745

Ion	Ratio	Lower	Upper
228	100		
226	27.3	21.8	32.8
229	20.0	15.9	23.9







#82
3,3'-Dichlorobenzidine
Concen: 3.994 ng
RT: 21.516 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:252 Resp: 121047

Ion	Ratio	Lower	Upper
252	100		
254	64.7	51.7	77.5
126	12.8	8.8	13.2







#83
Chrysene
Concen: 4.937 ng
RT: 21.657 min Scan# 3157
Delta R.T. -0.018 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:228 Resp: 413570

Ion	Ratio	Lower	Upper
228	100		
226	28.9	24.2	36.2
229	19.4	15.6	23.4







#84
Bis(2-ethylhexyl)phthalate
Concen: 4.021 ng
RT: 21.533 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:149 Resp: 223960

Ion	Ratio	Lower	Upper
149	100		
167	27.0	22.4	33.6
279	5.0	4.0	6.0







#85
Di-n-octyl phthalate
Concen: 3.712 ng
RT: 22.816 min Scan# 3354
Delta R.T. -0.012 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:149 Resp: 331185

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	7.9	6.0	9.0







#86
Perylene-d12
Concen: 20.000 ng
RT: 24.968 min Scan# 3
Delta R.T. -0.018 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:264 Resp: 1519057

Ion	Ratio	Lower	Upper
264	100		
260	23.4	18.9	28.3
265	21.6	17.8	26.6







#87
Indeno(1,2,3-cd)pyrene
Concen: 4.552 ng
RT: 28.762 min Scan# 4365
Delta R.T. -0.047 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:276 Resp: 475578

Ion	Ratio	Lower	Upper
276	100		
138	16.7	14.5	21.7
277	19.3	20.2	30.2#







#88
Benzo(b)fluoranthene
Concen: 4.723 ng
RT: 23.904 min Scan# 3
Delta R.T. -0.024 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:252 Resp: 429214

Ion	Ratio	Lower	Upper
252	100		
253	21.3	17.7	26.5
125	11.9	9.4	14.0







#89
Benzo(k)fluoranthene
Concen: 4.686 ng
RT: 23.968 min Scan# 3550
Delta R.T. -0.024 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:252 Resp: 413302

Ion	Ratio	Lower	Upper
252	100		
253	21.4	17.7	26.5
125	11.4	9.1	13.7







#90
Benzo(a)pyrene
Concen: 4.559 ng
RT: 24.810 min Scan# 3
Delta R.T. -0.024 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:252 Resp: 356658

Ion	Ratio	Lower	Upper
252	100		
253	22.7	17.6	26.4
125	13.6	10.1	15.1







#91
Dibenzo(a,h)anthracene
Concen: 4.571 ng
RT: 28.845 min Scan# 4379
Delta R.T. -0.059 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:278 Resp: 397247

Ion	Ratio	Lower	Upper
278	100		
139	16.2	12.6	19.0
279	23.1	18.5	27.7






#92
Benzo(g,h,i)perylene
Concen: 4.607 ng
RT: 29.951 min Scan# 4
Delta R.T. -0.035 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:276 Resp: 407522

Ion	Ratio	Lower	Upper
276	100		
277	23.9	18.9	28.3
138	23.8	18.7	28.1



ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024277.D
Acq On : 14 Apr 2025 12:27
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Quant Time: Apr 14 17:26:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.734	152	237860	20.000	ng		0.00
21) Naphthalene-d8	10.510	136	1022895	20.000	ng		0.00
39) Acenaphthene-d10	14.363	164	658652	20.000	ng		0.00
64) Phenanthrene-d10	17.163	188	1310701	20.000	ng		0.00
76) Chrysene-d12	21.616	240	1408913	20.000	ng		0.00
86) Perylene-d12	24.980	264	1508235	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.352	112	272237	18.889	ng		0.00
7) Phenol-d6	6.910	99	369199	18.725	ng		0.00
23) Nitrobenzene-d5	8.881	82	345094	19.207	ng		0.00
42) 2,4,6-Tribromophenol	15.886	330	168790	18.715	ng		0.00
45) 2-Fluorobiphenyl	12.975	172	878366	20.247	ng		0.00
79) Terphenyl-d14	19.892	244	1405643	19.888	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.293	88	61433	9.984	ng		98
3) Pyridine	3.693	79	152243	9.506	ng		99
4) n-Nitrosodimethylamine	3.599	42	52702	9.868	ng	#	98
6) Aniline	7.069	93	176840	9.935	ng		99
8) 2-Chlorophenol	7.305	128	154194	9.573	ng		100
9) Benzaldehyde	6.887	77	111093m	10.879	ng		
10) Phenol	6.940	94	186113	9.401	ng		97
11) bis(2-Chloroethyl)ether	7.157	93	156684	9.794	ng		98
12) 1,3-Dichlorobenzene	7.622	146	176712	10.126	ng		97
13) 1,4-Dichlorobenzene	7.769	146	177067	10.006	ng		99
14) 1,2-Dichlorobenzene	8.081	146	171376	10.016	ng		99
15) Benzyl Alcohol	7.969	79	111176	8.797	ng		99
16) 2,2'-oxybis(1-Chloropr...	8.252	45	177896	10.221	ng		98
17) 2-Methylphenol	8.175	107	119161	9.340	ng		98
18) Hexachloroethane	8.804	117	63912	10.007	ng		98
19) n-Nitroso-di-n-propyla...	8.528	70	119861	9.787	ng		100
20) 3+4-Methylphenols	8.499	107	162332	9.187	ng		99
22) Acetophenone	8.546	105	248823	9.787	ng		99
24) Nitrobenzene	8.922	77	170770	9.590	ng		98
25) Isophorone	9.446	82	307272	9.473	ng		99
26) 2-Nitrophenol	9.628	139	73320	8.575	ng		96
27) 2,4-Dimethylphenol	9.693	122	99186	9.152	ng		97
28) bis(2-Chloroethoxy)met...	9.922	93	217866	9.869	ng		99
29) 2,4-Dichlorophenol	10.163	162	136816	9.263	ng		99
30) 1,2,4-Trichlorobenzene	10.369	180	155965	9.760	ng		100
31) Naphthalene	10.557	128	538892	9.953	ng		99
32) Benzoic acid	9.781	122	94803m	8.095	ng		
33) 4-Chloroaniline	10.669	127	178451	9.433	ng		99
34) Hexachlorobutadiene	10.845	225	92662	9.885	ng		99
35) Caprolactam	11.440	113	47014	8.670	ng		98
36) 4-Chloro-3-methylphenol	11.798	107	158563	9.213	ng		99
37) 2-Methylnaphthalene	12.169	142	368122	9.810	ng		98
38) 1-Methylnaphthalene	12.387	142	365139	9.933	ng		100
40) 1,2,4,5-Tetrachloroben...	12.540	216	177188	9.789	ng		100
41) Hexachlorocyclopentadiene	12.516	237	60914	9.586	ng		96
43) 2,4,6-Trichlorophenol	12.781	196	110203	9.253	ng		99

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024277.D
Acq On : 14 Apr 2025 12:27
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Quant Time: Apr 14 17:26:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	12.857	196	122104	9.251	ng		98
46) 1,1'-Biphenyl	13.187	154	490325	10.080	ng		100
47) 2-Chloronaphthalene	13.228	162	362419	9.985	ng		98
48) 2-Nitroaniline	13.434	65	88889	8.866	ng		95
49) Acenaphthylene	14.081	152	553215	9.711	ng		100
50) Dimethylphthalate	13.816	163	466314	9.704	ng		99
51) 2,6-Dinitrotoluene	13.934	165	95596	9.442	ng		96
52) Acenaphthene	14.428	154	357828	9.877	ng		99
53) 3-Nitroaniline	14.269	138	95237	9.067	ng		97
54) 2,4-Dinitrophenol	14.481	184	41132	7.588	ng		97
55) Dibenzofuran	14.769	168	594953	10.013	ng		99
56) 4-Nitrophenol	14.598	139	75995	8.416	ng		95
57) 2,4-Dinitrotoluene	14.734	165	121276	8.862	ng	#	95
58) Fluorene	15.428	166	459952	10.026	ng		100
59) 2,3,4,6-Tetrachlorophenol	15.004	232	115645	9.470	ng		100
60) Diethylphthalate	15.198	149	488847	9.884	ng		100
61) 4-Chlorophenyl-phenyle...	15.428	204	222437	9.985	ng		97
62) 4-Nitroaniline	15.451	138	98638	8.857	ng		93
63) Azobenzene	15.722	77	455459	9.999	ng		96
65) 4,6-Dinitro-2-methylph...	15.516	198	62421	8.151	ng		96
66) n-Nitrosodiphenylamine	15.639	169	383558	9.759	ng		99
67) 4-Bromophenyl-phenylether	16.339	248	138724	9.703	ng		97
68) Hexachlorobenzene	16.457	284	161944	9.528	ng		98
69) Atrazine	16.628	200	109503	10.990	ng		98
70) Pentachlorophenol	16.810	266	101318	8.682	ng		96
71) Phenanthrene	17.210	178	711407	9.900	ng		99
72) Anthracene	17.304	178	666269	9.635	ng		100
73) Carbazole	17.586	167	654205	9.721	ng		99
74) Di-n-butylphthalate	18.180	149	806280	9.616	ng		99
75) Fluoranthene	19.292	202	835153	9.791	ng		98
77) Benzidine	19.498	184	99659	5.730	ng		99
78) Pyrene	19.669	202	870767	9.659	ng		99
80) Butylbenzylphthalate	20.627	149	343670	9.043	ng		99
81) Benzo(a)anthracene	21.598	228	854983	9.734	ng		98
82) 3,3'-Dichlorobenzidine	21.521	252	269772	8.857	ng		98
83) Chrysene	21.668	228	832263	9.887	ng		100
84) Bis(2-ethylhexyl)phtha...	21.539	149	517894	9.253	ng		99
85) Di-n-octyl phthalate	22.827	149	769654	8.584	ng		99
87) Indeno(1,2,3-cd)pyrene	28.792	276	983323	9.480	ng		99
88) Benzo(b)fluoranthene	23.915	252	885212	9.810	ng		99
89) Benzo(k)fluoranthene	23.986	252	852606	9.735	ng		99
90) Benzo(a)pyrene	24.815	252	732209	9.426	ng		99
91) Dibenzo(a,h)anthracene	28.862	278	821160	9.517	ng		99
92) Benzo(g,h,i)perylene	29.962	276	836993	9.530	ng		99


(#) = qualifier out of range (m) = manual integration (+) = signals summed


ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024277.D
Acq On : 14 Apr 2025 12:27
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Quant Time: Apr 14 17:26:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.734 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:152 Resp: 237860

Ion	Ratio	Lower	Upper
152	100		
150	153.1	124.9	187.3
115	59.6	47.7	71.5







#2
1,4-Dioxane
Concen: 9.984 ng
RT: 3.293 min Scan# 35
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion: 88 Resp: 61433

Ion	Ratio	Lower	Upper
88	100		
58	65.6	51.4	77.2
43	22.9	17.1	25.7







#3
Pyridine
Concen: 9.506 ng
RT: 3.693 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 79 Resp: 152243

Ion	Ratio	Lower	Upper
79	100		
52	57.2	45.4	68.2
51	29.7	22.6	33.8







#4
n-Nitrosodimethylamine
Concen: 9.868 ng
RT: 3.599 min Scan# 87
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion: 42 Resp: 52702

Ion	Ratio	Lower	Upper
42	100		
74	172.0	139.5	209.3
44	10.8	5.8	8.8#







#5
2-Fluorophenol
Concen: 18.889 ng
RT: 5.352 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:112 Resp: 272237

Ion	Ratio	Lower	Upper
112	100		
64	51.6	41.1	61.7
63	26.0	20.6	30.8







#6
Aniline
Concen: 9.935 ng
RT: 7.069 min Scan# 677
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion: 93 Resp: 176840

Ion	Ratio	Lower	Upper
93	100		
66	36.8	29.1	43.7
65	18.6	14.7	22.1







#7
Phenol-d6
Concen: 18.725 ng
RT: 6.910 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 99 Resp: 369199

Ion	Ratio	Lower	Upper
99	100		
42	13.5	10.6	16.0
71	31.0	23.7	35.5







#8
2-Chlorophenol
Concen: 9.573 ng
RT: 7.305 min Scan# 717
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:128 Resp: 154194

Ion	Ratio	Lower	Upper
128	100		
130	32.7	12.9	52.9
64	42.8	22.7	62.7







#9
Benzaldehyde
Concen: 10.879 ng m
RT: 6.887 min Scan# 64
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 77 Resp: 111093

Ion	Ratio	Lower	Upper
77	100		
105	98.4	79.2	119.2
106	93.5	76.2	116.2







#10
Phenol
Concen: 9.401 ng
RT: 6.940 min Scan# 655
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion: 94 Resp: 186113

Ion	Ratio	Lower	Upper
94	100		
65	25.7	4.5	44.5
66	34.4	12.4	52.4







#11
bis(2-Chloroethyl)ether
Concen: 9.794 ng
RT: 7.157 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 93 Resp: 156684

Ion	Ratio	Lower	Upper
93	100		
63	70.5	48.6	88.6
95	31.6	12.3	52.3







#12
1,3-Dichlorobenzene
Concen: 10.126 ng
RT: 7.622 min Scan# 771
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:146 Resp: 176712

Ion	Ratio	Lower	Upper
146	100		
148	60.8	51.4	77.0
75	28.4	22.9	34.3







#13
1,4-Dichlorobenzene
Concen: 10.006 ng
RT: 7.769 min Scan# 7
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:146 Resp: 177067

Ion	Ratio	Lower	Upper
146	100		
148	64.9	51.1	76.7
111	41.1	32.6	48.8







#14
1,2-Dichlorobenzene
Concen: 10.016 ng
RT: 8.081 min Scan# 849
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:146 Resp: 171376

Ion	Ratio	Lower	Upper
146	100		
148	64.0	51.0	76.4
111	44.3	33.8	50.8







#15
Benzyl Alcohol
Concen: 8.797 ng
RT: 7.969 min Scan# 8
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 79 Resp: 111176

Ion	Ratio	Lower	Upper
79	100		
108	81.3	65.1	97.7
77	64.5	50.7	76.1







#16
2,2'-oxybis(1-Chloropropane)
Concen: 10.221 ng
RT: 8.252 min Scan# 878
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion: 45 Resp: 177896

Ion	Ratio	Lower	Upper
45	100		
77	18.4	0.0	37.7
79	14.1	0.0	33.5







#17
2-Methylphenol
Concen: 9.340 ng
RT: 8.175 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:107 Resp: 119161

Ion	Ratio	Lower	Upper
107	100		
108	109.3	89.8	134.6
77	43.9	34.2	51.4
79	43.8	33.8	50.8







#18
Hexachloroethane
Concen: 10.007 ng
RT: 8.804 min Scan# 972
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:117 Resp: 63912

Ion	Ratio	Lower	Upper
117	100		
119	95.8	77.8	116.6
201	92.4	75.9	113.9







#19
n-Nitroso-di-n-propylamine
Concen: 9.787 ng
RT: 8.528 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 70 Resp: 119861

Ion	Ratio	Lower	Upper
70	100		
42	46.2	37.3	55.9
101	10.3	8.2	12.4
130	23.3	18.6	28.0







#20
3+4-Methylphenols
Concen: 9.187 ng
RT: 8.499 min Scan# 920
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:107 Resp: 162332

Ion	Ratio	Lower	Upper
107	100		
108	90.5	69.8	109.8
77	32.3	11.5	51.5
79	27.4	7.8	47.8







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.510 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:136 Resp: 1022895

Ion	Ratio	Lower	Upper
136	100		
137	10.7	9.2	13.8
54	6.6	6.0	9.0
68	6.5	5.5	8.3







#22
Acetophenone
Concen: 9.787 ng
RT: 8.546 min Scan# 928
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:105 Resp: 248823

Ion	Ratio	Lower	Upper
105	100		
71	2.1	1.7	2.5
51	21.7	16.9	25.3
120	23.2	18.6	28.0







#23
Nitrobenzene-d5
Concen: 19.207 ng
RT: 8.881 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 82 Resp: 345094

Ion	Ratio	Lower	Upper
82	100		
128	43.8	36.5	54.7
54	40.9	33.4	50.2







#24
Nitrobenzene
Concen: 9.590 ng
RT: 8.922 min Scan# 992
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion: 77 Resp: 170770

Ion	Ratio	Lower	Upper
77	100		
123	47.0	39.2	58.8
65	12.1	9.8	14.6







#25
Isophorone
Concen: 9.473 ng
RT: 9.446 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 82 Resp: 307272

Ion	Ratio	Lower	Upper
82	100		
95	6.6	5.4	8.2
138	18.3	14.4	21.6







#26
2-Nitrophenol
Concen: 8.575 ng
RT: 9.628 min Scan# 1112
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:139 Resp: 73320

Ion	Ratio	Lower	Upper
139	100		
109	29.2	21.8	32.6
65	48.2	36.2	54.2







#27
2,4-Dimethylphenol
Concen: 9.152 ng
RT: 9.693 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:122 Resp: 99186

Ion	Ratio	Lower	Upper
122	100		
107	112.4	93.7	140.5
121	59.0	48.2	72.2







#28
bis(2-Chloroethoxy)methane
Concen: 9.869 ng
RT: 9.922 min Scan# 1162
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion: 93 Resp: 217866

Ion	Ratio	Lower	Upper
93	100		
95	31.8	26.0	39.0
123	11.7	9.1	13.7







#29
2,4-Dichlorophenol
Concen: 9.263 ng
RT: 10.163 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:162 Resp: 136816

Ion	Ratio	Lower	Upper
162	100		
164	62.8	43.3	83.3
98	37.9	16.6	56.6







#30
1,2,4-Trichlorobenzene
Concen: 9.760 ng
RT: 10.369 min Scan# 1238
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:180 Resp: 155965

Ion	Ratio	Lower	Upper
180	100		
182	96.5	77.3	115.9
145	31.1	25.1	37.7







#31
Naphthalene
Concen: 9.953 ng
RT: 10.557 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:128 Resp: 538892

Ion	Ratio	Lower	Upper
128	100		
129	10.5	8.6	13.0
127	13.1	10.3	15.5







#32
Benzoic acid
Concen: 8.095 ng m
RT: 9.781 min Scan# 1138
Delta R.T. -0.047 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:122 Resp: 94803

Ion	Ratio	Lower	Upper
122	100		
105	123.3	104.8	144.8
77	92.9	69.2	109.2







#33
4-Chloroaniline
Concen: 9.433 ng
RT: 10.669 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:127 Resp: 178451

Ion	Ratio	Lower	Upper
127	100		
129	32.9	26.1	39.1
65	28.1	22.1	33.1
92	19.2	15.0	22.4







#34
Hexachlorobutadiene
Concen: 9.885 ng
RT: 10.845 min Scan# 1319
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:225 Resp: 92662

Ion	Ratio	Lower	Upper
225	100		
223	61.4	49.5	74.3
227	65.0	50.9	76.3







#35
Caprolactam
Concen: 8.670 ng
RT: 11.440 min Scan# 14
Delta R.T. -0.012 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:113 Resp: 47014

Ion	Ratio	Lower	Upper
113	100		
55	134.3	115.9	155.9
56	97.4	81.1	121.1







#36
4-Chloro-3-methylphenol
Concen: 9.213 ng
RT: 11.798 min Scan# 1481
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:107 Resp: 158563

Ion	Ratio	Lower	Upper
107	100		
144	25.8	20.5	30.7
142	80.6	63.4	95.2







#37
2-Methylnaphthalene
Concen: 9.810 ng
RT: 12.169 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:142 Resp: 368122

Ion	Ratio	Lower	Upper
142	100		
141	89.7	70.3	105.5
115	32.5	26.9	40.3







#38
1-Methylnaphthalene
Concen: 9.933 ng
RT: 12.387 min Scan# 1581
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:142 Resp: 365139

Ion	Ratio	Lower	Upper
142	100		
141	92.1	73.6	110.4
116	3.5	3.0	4.6







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.363 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:164 Resp: 658652

Ion	Ratio	Lower	Upper
164	100		
162	100.6	80.6	120.8
160	45.0	35.3	52.9







#40
1,2,4,5-Tetrachlorobenzene
Concen: 9.789 ng
RT: 12.540 min Scan# 1607
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:216 Resp: 177188

Ion	Ratio	Lower	Upper
216	100		
214	78.3	62.6	93.8
179	21.5	17.2	25.8
108	19.8	15.0	22.6







#41
Hexachlorocyclopentadiene
Concen: 9.586 ng
RT: 12.516 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:237 Resp: 60914

Ion	Ratio	Lower	Upper
237	100		
235	60.0	43.9	83.9
272	13.3	0.0	33.6







#42
2,4,6-Tribromophenol
Concen: 18.715 ng
RT: 15.886 min Scan# 2176
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:330 Resp: 168790

Ion	Ratio	Lower	Upper
330	100		
332	95.5	77.1	115.7
141	31.2	24.7	37.1







#43
2,4,6-Trichlorophenol
Concen: 9.253 ng
RT: 12.781 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:196 Resp: 110203

Ion	Ratio	Lower	Upper
196	100		
198	93.9	74.3	111.5
200	29.3	24.0	36.0







#44
2,4,5-Trichlorophenol
Concen: 9.251 ng
RT: 12.857 min Scan# 1661
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:196 Resp: 122104

Ion	Ratio	Lower	Upper
196	100		
198	94.4	77.3	115.9
97	50.8	40.4	60.6
132	29.9	23.4	35.0







#45
2-Fluorobiphenyl
Concen: 20.247 ng
RT: 12.975 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:172 Resp: 878366

Ion	Ratio	Lower	Upper
172	100		
171	34.7	28.2	42.2
170	23.0	18.6	27.8







#46
1,1'-Biphenyl
Concen: 10.080 ng
RT: 13.187 min Scan# 1717
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:154 Resp: 490325

Ion	Ratio	Lower	Upper
154	100		
153	40.0	20.3	60.3
76	13.6	0.0	33.5







#47
2-Chloronaphthalene
Concen: 9.985 ng
RT: 13.228 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:162 Resp: 362419

Ion	Ratio	Lower	Upper
162	100		
127	38.3	31.9	47.9
164	32.5	26.2	39.4







#48
2-Nitroaniline
Concen: 8.866 ng
RT: 13.434 min Scan# 1759
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion: 65 Resp: 88889

Ion	Ratio	Lower	Upper
65	100		
92	77.0	62.3	93.5
138	113.7	97.2	145.8







#49
Acenaphthylene
Concen: 9.711 ng
RT: 14.081 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:152 Resp: 553215

Ion	Ratio	Lower	Upper
152	100		
151	19.8	15.9	23.9
153	13.1	10.6	15.8







#50
Dimethylphthalate
Concen: 9.704 ng
RT: 13.816 min Scan# 1824
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:163 Resp: 466314

Ion	Ratio	Lower	Upper
163	100		
194	4.2	3.5	5.3
164	9.8	8.0	12.0







#51
2,6-Dinitrotoluene
Concen: 9.442 ng
RT: 13.934 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:165 Resp: 95596

Ion	Ratio	Lower	Upper
165	100		
63	50.6	38.0	57.0
89	54.0	41.7	62.5







#52
Acenaphthene
Concen: 9.877 ng
RT: 14.428 min Scan# 1928
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:154 Resp: 357828

Ion	Ratio	Lower	Upper
154	100		
153	111.6	89.8	134.6
152	53.6	42.6	63.8







#53
3-Nitroaniline
Concen: 9.067 ng
RT: 14.269 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:138 Resp: 95237

Ion	Ratio	Lower	Upper
138	100		
108	11.7	8.3	12.5
92	116.0	89.8	134.8







#54
2,4-Dinitrophenol
Concen: 7.588 ng
RT: 14.481 min Scan# 1937
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:184 Resp: 41132

Ion	Ratio	Lower	Upper
184	100		
63	57.2	42.6	64.0
154	58.9	47.2	70.8







#55
Dibenzofuran
Concen: 10.013 ng
RT: 14.769 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:168 Resp: 594953

Ion	Ratio	Lower	Upper
168	100		
139	37.5	30.3	45.5
169	13.1	10.6	16.0







#56
4-Nitrophenol
Concen: 8.416 ng
RT: 14.598 min Scan# 1957
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:139 Resp: 75995

Ion	Ratio	Lower	Upper
139	100		
109	67.3	46.1	86.1
65	92.4	65.0	105.0







#57
2,4-Dinitrotoluene
Concen: 8.862 ng
RT: 14.734 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:165 Resp: 121276

Ion	Ratio	Lower	Upper
165	100		
63	39.5	30.4	45.6
89	73.4	54.6	82.0
182	3.4	2.2	3.2#







#58
Fluorene
Concen: 10.026 ng
RT: 15.428 min Scan# 2098
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:166 Resp: 459952

Ion	Ratio	Lower	Upper
166	100		
165	97.2	78.1	117.1
167	13.4	10.9	16.3







#59
2,3,4,6-Tetrachlorophenol
Concen: 9.470 ng
RT: 15.004 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:232 Resp: 115645

Ion	Ratio	Lower	Upper
232	100		
131	45.6	36.8	55.2
130	2.5	1.9	2.9
166	31.3	25.0	37.6







#60
Diethylphthalate
Concen: 9.884 ng
RT: 15.198 min Scan# 2059
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:149 Resp: 488847

Ion	Ratio	Lower	Upper
149	100		
177	21.9	17.4	26.2
150	12.6	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 9.985 ng
RT: 15.428 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:204 Resp: 222437

Ion	Ratio	Lower	Upper
204	100		
206	32.9	26.5	39.7
141	58.8	49.4	74.0







#62
4-Nitroaniline
Concen: 8.857 ng
RT: 15.451 min Scan# 2102
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:138 Resp: 98638

Ion	Ratio	Lower	Upper
138	100		
92	53.3	27.8	67.8
108	74.5	48.9	88.9







#63
Azobenzene
Concen: 9.999 ng
RT: 15.722 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 77 Resp: 455459

Ion	Ratio	Lower	Upper
77	100		
182	22.7	7.3	47.3
105	20.9	2.3	42.3
51	23.4	3.4	43.4







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.163 min Scan# 2393
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:188 Resp: 1310701

Ion	Ratio	Lower	Upper
188	100		
94	10.9	8.6	13.0
80	12.1	9.8	14.6







#65
4,6-Dinitro-2-methylphenol
Concen: 8.151 ng
RT: 15.516 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:198 Resp: 62421

Ion	Ratio	Lower	Upper
198	100		
51	41.0	18.4	58.4
105	49.0	26.5	66.5







#66
n-Nitrosodiphenylamine
Concen: 9.759 ng
RT: 15.639 min Scan# 2134
Delta R.T. -0.012 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:169 Resp: 383558

Ion	Ratio	Lower	Upper
169	100		
168	67.2	53.0	79.4
167	37.1	28.9	43.3







#67
4-Bromophenyl-phenylether
Concen: 9.703 ng
RT: 16.339 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:248 Resp: 138724

Ion	Ratio	Lower	Upper
248	100		
250	97.2	78.2	117.4
141	76.6	57.7	86.5







#68
Hexachlorobenzene
Concen: 9.528 ng
RT: 16.457 min Scan# 2273
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:284 Resp: 161944

Ion	Ratio	Lower	Upper
284	100		
142	32.4	24.6	37.0
249	29.1	23.8	35.6







#69
Atrazine
Concen: 10.990 ng
RT: 16.628 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:200 Resp: 109503

Ion	Ratio	Lower	Upper
200	100		
173	28.8	7.5	47.5
215	49.5	28.8	68.8







#70
Pentachlorophenol
Concen: 8.682 ng
RT: 16.810 min Scan# 2333
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:266 Resp: 101318

Ion	Ratio	Lower	Upper
266	100		
268	67.1	51.0	76.6
264	65.3	49.8	74.6







#71
Phenanthrene
Concen: 9.900 ng
RT: 17.210 min Scan# 24
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:178 Resp: 711407

Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.4	23.0
179	15.5	12.5	18.7







#72
Anthracene
Concen: 9.635 ng
RT: 17.304 min Scan# 2417
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:178 Resp: 666269

Ion	Ratio	Lower	Upper
178	100		
176	18.5	14.7	22.1
179	15.4	12.4	18.6







#73
Carbazole
Concen: 9.721 ng
RT: 17.586 min Scan# 24
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:167 Resp: 654205

Ion	Ratio	Lower	Upper
167	100		
166	20.8	17.1	25.7
139	12.5	10.6	16.0







#74
Di-n-butylphthalate
Concen: 9.616 ng
RT: 18.180 min Scan# 2566
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:149 Resp: 806280

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.4	11.2
104	4.9	3.9	5.9







#75
Fluoranthene
Concen: 9.791 ng
RT: 19.292 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:202 Resp: 835153

Ion	Ratio	Lower	Upper
202	100		
101	12.3	0.0	33.2
203	17.1	0.0	37.7







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.616 min Scan# 3150
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:240 Resp: 1408913

Ion	Ratio	Lower	Upper
240	100		
120	12.7	9.8	14.8
236	25.0	20.9	31.3







#77
Benzidine
Concen: 5.730 ng
RT: 19.498 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:184 Resp: 99659

Ion	Ratio	Lower	Upper
184	100		
185	14.7	11.6	17.4
183	10.6	9.2	13.8







#78
Pyrene
Concen: 9.659 ng
RT: 19.669 min Scan# 2819
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:202 Resp: 870767

Ion	Ratio	Lower	Upper
202	100		
200	21.1	17.2	25.8
203	17.4	14.4	21.6







#79
Terphenyl-d14
Concen: 19.888 ng
RT: 19.892 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:244 Resp: 1405643

Ion	Ratio	Lower	Upper
244	100		
212	6.4	5.6	8.4
122	11.8	9.4	14.0







#80
Butylbenzylphthalate
Concen: 9.043 ng
RT: 20.627 min Scan# 2982
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:149 Resp: 343670

Ion	Ratio	Lower	Upper
149	100		
91	69.9	56.4	84.6
206	19.7	16.3	24.5







#81
Benzo(a)anthracene
Concen: 9.734 ng
RT: 21.598 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:228 Resp: 854983

Ion	Ratio	Lower	Upper
228	100		
226	26.1	21.8	32.8
229	20.2	15.9	23.9







#82
3,3'-Dichlorobenzidine
Concen: 8.857 ng
RT: 21.521 min Scan# 3134
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:252 Resp: 269772

Ion	Ratio	Lower	Upper
252	100		
254	63.4	51.7	77.5
126	12.2	8.8	13.2







#83
Chrysene
Concen: 9.887 ng
RT: 21.668 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:228 Resp: 832263

Ion	Ratio	Lower	Upper
228	100		
226	30.1	24.2	36.2
229	19.7	15.6	23.4







#84
Bis(2-ethylhexyl)phthalate
Concen: 9.253 ng
RT: 21.539 min Scan# 3137
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:149 Resp: 517894

Ion	Ratio	Lower	Upper
149	100		
167	27.4	22.4	33.6
279	4.4	4.0	6.0







#85
Di-n-octyl phthalate
Concen: 8.584 ng
RT: 22.827 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:149 Resp: 769654

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	7.9	6.0	9.0







#86
Perylene-d12
Concen: 20.000 ng
RT: 24.980 min Scan# 3722
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:264 Resp: 1508235

Ion	Ratio	Lower	Upper
264	100		
260	23.7	18.9	28.3
265	21.8	17.8	26.6







#87
Indeno(1,2,3-cd)pyrene
Concen: 9.480 ng
RT: 28.792 min Scan# 4
Delta R.T. -0.018 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:276 Resp: 983323

Ion	Ratio	Lower	Upper
276	100		
138	19.2	14.5	21.7
277	25.4	20.2	30.2







#88
Benzo(b)fluoranthene
Concen: 9.810 ng
RT: 23.915 min Scan# 3541
Delta R.T. -0.012 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:252 Resp: 885212

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.7	26.5
125	12.2	9.4	14.0







#89
Benzo(k)fluoranthene
Concen: 9.735 ng
RT: 23.986 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:252 Resp: 852606

Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.7	26.5
125	11.5	9.1	13.7







#90
Benzo(a)pyrene
Concen: 9.426 ng
RT: 24.815 min Scan# 3694
Delta R.T. -0.018 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:252 Resp: 732209

Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.6	26.4
125	13.0	10.1	15.1







#91
Dibenzo(a,h)anthracene
Concen: 9.517 ng
RT: 28.862 min Scan# 4
Delta R.T. -0.041 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:278 Resp: 821160

Ion	Ratio	Lower	Upper
278	100		
139	16.8	12.6	19.0
279	23.4	18.5	27.7






#92
Benzo(g,h,i)perylene
Concen: 9.530 ng
RT: 29.962 min Scan# 4569
Delta R.T. -0.024 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:276 Resp: 836993

Ion	Ratio	Lower	Upper
276	100		
277	24.1	18.9	28.3
138	22.5	18.7	28.1



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024278.D
Acq On : 14 Apr 2025 13:08
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Quant Time: Apr 14 17:26:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.734	152	251998	20.000	ng	0.00
21) Naphthalene-d8	10.505	136	1081087	20.000	ng	0.00
39) Acenaphthene-d10	14.369	164	688775	20.000	ng	0.00
64) Phenanthrene-d10	17.169	188	1378480	20.000	ng	0.00
76) Chrysene-d12	21.616	240	1415765	20.000	ng	0.00
86) Perylene-d12	24.986	264	1549781	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.352	112	636245	41.668	ng	0.00
7) Phenol-d6	6.911	99	866004	41.458	ng	0.00
23) Nitrobenzene-d5	8.875	82	794485	41.839	ng	0.00
42) 2,4,6-Tribromophenol	15.875	330	383088	40.617	ng	0.00
45) 2-Fluorobiphenyl	12.975	172	1871534	41.254	ng	0.00
79) Terphenyl-d14	19.898	244	2987245	42.062	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.293	88	139222	21.358	ng	99
3) Pyridine	3.693	79	358370	21.120	ng	99
4) n-Nitrosodimethylamine	3.599	42	118077	20.869	ng	98
6) Aniline	7.069	93	412168	21.857	ng	99
8) 2-Chlorophenol	7.305	128	355815	20.852	ng	99
9) Benzaldehyde	6.881	77	237371	21.940	ng	100
10) Phenol	6.940	94	433531	20.670	ng	98
11) bis(2-Chloroethyl)ether	7.164	93	352530	20.800	ng	100
12) 1,3-Dichlorobenzene	7.622	146	386470	20.903	ng	100
13) 1,4-Dichlorobenzene	7.763	146	390483	20.828	ng	100
14) 1,2-Dichlorobenzene	8.081	146	378518	20.880	ng	99
15) Benzyl Alcohol	7.969	79	278608	20.809	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.258	45	382308	20.734	ng	99
17) 2-Methylphenol	8.175	107	283057	20.942	ng	99
18) Hexachloroethane	8.805	117	140558	20.773	ng	99
19) n-Nitroso-di-n-propyla...	8.528	70	274118	21.127	ng	100
20) 3+4-Methylphenols	8.499	107	393131	20.999	ng	99
22) Acetophenone	8.546	105	563994	20.991	ng	99
24) Nitrobenzene	8.916	77	390911	20.770	ng	98
25) Isophorone	9.446	82	705771	20.588	ng	98
26) 2-Nitrophenol	9.628	139	185740	20.553	ng	99
27) 2,4-Dimethylphenol	9.687	122	236436	20.643	ng	98
28) bis(2-Chloroethoxy)met...	9.916	93	480036	20.575	ng	100
29) 2,4-Dichlorophenol	10.163	162	320579	20.536	ng	99
30) 1,2,4-Trichlorobenzene	10.369	180	343641	20.348	ng	99
31) Naphthalene	10.557	128	1179848	20.618	ng	100
32) Benzoic acid	9.805	122	244213m	19.731	ng	
33) 4-Chloroaniline	10.669	127	420086	21.010	ng	100
34) Hexachlorobutadiene	10.846	225	200785	20.267	ng	99
35) Caprolactam	11.446	113	119912	20.922	ng	98
36) 4-Chloro-3-methylphenol	11.799	107	377768	20.767	ng	99
37) 2-Methylnaphthalene	12.169	142	816314	20.583	ng	100
38) 1-Methylnaphthalene	12.387	142	796550	20.503	ng	100
40) 1,2,4,5-Tetrachloroben...	12.540	216	391241	20.669	ng	100
41) Hexachlorocyclopentadiene	12.522	237	135654	20.414	ng	97
43) 2,4,6-Trichlorophenol	12.787	196	254189	20.410	ng	97

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024278.D
Acq On : 14 Apr 2025 13:08
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Quant Time: Apr 14 17:26:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.863	196	281887	20.424	ng	99
46) 1,1'-Biphenyl	13.187	154	1052618	20.694	ng	100
47) 2-Chloronaphthalene	13.228	162	782040	20.603	ng	99
48) 2-Nitroaniline	13.440	65	218194	20.811	ng	99
49) Acenaphthylene	14.087	152	1234728	20.726	ng	99
50) Dimethylphthalate	13.822	163	1029599	20.489	ng	100
51) 2,6-Dinitrotoluene	13.940	165	219492	20.731	ng	98
52) Acenaphthene	14.434	154	785789	20.741	ng	99
53) 3-Nitroaniline	14.269	138	233006	21.212	ng	99
54) 2,4-Dinitrophenol	14.481	184	112157	19.785	ng	94
55) Dibenzofuran	14.769	168	1288363	20.735	ng	99
56) 4-Nitrophenol	14.593	139	199315	21.109	ng	99
57) 2,4-Dinitrotoluene	14.734	165	302332	21.126	ng	98
58) Fluorene	15.428	166	1001515	20.876	ng	100
59) 2,3,4,6-Tetrachlorophenol	14.998	232	262077	20.523	ng	99
60) Diethylphthalate	15.198	149	1052183	20.344	ng	100
61) 4-Chlorophenyl-phenyle...	15.422	204	481631	20.675	ng	98
62) 4-Nitroaniline	15.445	138	243257	20.888	ng	96
63) Azobenzene	15.722	77	995912	20.907	ng	100
65) 4,6-Dinitro-2-methylph...	15.510	198	162501	20.176	ng	98
66) n-Nitrosodiphenylamine	15.640	169	849476	20.552	ng	100
67) 4-Bromophenyl-phenylether	16.334	248	302601	20.125	ng	99
68) Hexachlorobenzene	16.451	284	358889	20.077	ng	100
69) Atrazine	16.616	200	184151	17.573	ng	98
70) Pentachlorophenol	16.810	266	241815	19.702	ng	98
71) Phenanthrene	17.210	178	1540112	20.378	ng	100
72) Anthracene	17.298	178	1494006	20.543	ng	99
73) Carbazole	17.581	167	1461729	20.653	ng	100
74) Di-n-butylphthalate	18.175	149	1727074	19.585	ng	100
75) Fluoranthene	19.292	202	1813662	20.218	ng	99
77) Benzidine	19.498	184	154964	8.867	ng	99
78) Pyrene	19.669	202	1904884	21.028	ng	99
80) Butylbenzylphthalate	20.633	149	772776	20.235	ng	96
81) Benzo(a)anthracene	21.598	228	1829798	20.732	ng	99
82) 3,3'-Dichlorobenzidine	21.527	252	610749	19.956	ng	99
83) Chrysene	21.669	228	1739548	20.565	ng	99
84) Bis(2-ethylhexyl)phtha...	21.545	149	1103263	19.616	ng	99
85) Di-n-octyl phthalate	22.827	149	1730436	19.207	ng	100
87) Indeno(1,2,3-cd)pyrene	28.792	276	2180177	20.454	ng	99
88) Benzo(b)fluoranthene	23.916	252	1885654	20.337	ng	99
89) Benzo(k)fluoranthene	23.980	252	1885747	20.955	ng	99
90) Benzo(a)pyrene	24.816	252	1637648	20.518	ng	98
91) Dibenzo(a,h)anthracene	28.868	278	1815743	20.480	ng	98
92) Benzo(g,h,i)perylene	29.951	276	1856380	20.571	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024278.D
Acq On : 14 Apr 2025 13:08
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Quant Time: Apr 14 17:26:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.734 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:152 Resp: 251998

Ion	Ratio	Lower	Upper
152	100		
150	155.4	124.9	187.3
115	59.4	47.7	71.5







#2
1,4-Dioxane
Concen: 21.358 ng
RT: 3.293 min Scan# 35
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion: 88 Resp: 139222

Ion	Ratio	Lower	Upper
88	100		
58	64.8	51.4	77.2
43	22.2	17.1	25.7







#3
Pyridine
Concen: 21.120 ng
RT: 3.693 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 79 Resp: 358370

Ion	Ratio	Lower	Upper
79	100		
52	55.9	45.4	68.2
51	28.3	22.6	33.8







#4
n-Nitrosodimethylamine
Concen: 20.869 ng
RT: 3.599 min Scan# 87
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion: 42 Resp: 118077

Ion	Ratio	Lower	Upper
42	100		
74	171.1	139.5	209.3
44	8.3	5.8	8.8







#5
2-Fluorophenol
Concen: 41.668 ng
RT: 5.352 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:112 Resp: 636245

Ion	Ratio	Lower	Upper
112	100		
64	52.4	41.1	61.7
63	25.4	20.6	30.8







#6
Aniline
Concen: 21.857 ng
RT: 7.069 min Scan# 677
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion: 93 Resp: 412168

Ion	Ratio	Lower	Upper
93	100		
66	35.2	29.1	43.7
65	18.5	14.7	22.1







#7
Phenol-d6
Concen: 41.458 ng
RT: 6.911 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 99 Resp: 866004

Ion	Ratio	Lower	Upper
99	100		
42	13.7	10.6	16.0
71	30.2	23.7	35.5







#8
2-Chlorophenol
Concen: 20.852 ng
RT: 7.305 min Scan# 717
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:128 Resp: 355815

Ion	Ratio	Lower	Upper
128	100		
130	32.2	12.9	52.9
64	42.7	22.7	62.7







#9
Benzaldehyde
Concen: 21.940 ng
RT: 6.881 min Scan# 64
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 77 Resp: 237371

Ion	Ratio	Lower	Upper
77	100		
105	99.1	79.2	119.2
106	95.6	76.2	116.2







#10
Phenol
Concen: 20.670 ng
RT: 6.940 min Scan# 655
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion: 94 Resp: 433531

Ion	Ratio	Lower	Upper
94	100		
65	25.8	4.5	44.5
66	32.8	12.4	52.4







#11
bis(2-Chloroethyl)ether
Concen: 20.800 ng
RT: 7.164 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 93 Resp: 352530

Ion	Ratio	Lower	Upper
93	100		
63	68.7	48.6	88.6
95	32.7	12.3	52.3







#12
1,3-Dichlorobenzene
Concen: 20.903 ng
RT: 7.622 min Scan# 771
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:146 Resp: 386470

Ion	Ratio	Lower	Upper
146	100		
148	64.1	51.4	77.0
75	28.4	22.9	34.3







#13
1,4-Dichlorobenzene
Concen: 20.828 ng
RT: 7.763 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:146 Resp: 390483

Ion	Ratio	Lower	Upper
146	100		
148	63.6	51.1	76.7
111	41.1	32.6	48.8







#14
1,2-Dichlorobenzene
Concen: 20.880 ng
RT: 8.081 min Scan# 849
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:146 Resp: 378518

Ion	Ratio	Lower	Upper
146	100		
148	64.6	51.0	76.4
111	42.3	33.8	50.8







#15
Benzyl Alcohol
Concen: 20.809 ng
RT: 7.969 min Scan# 8
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 79 Resp: 278608

Ion	Ratio	Lower	Upper
79	100		
108	79.3	65.1	97.7
77	63.4	50.7	76.1







#16
2,2'-oxybis(1-Chloropropane)
Concen: 20.734 ng
RT: 8.258 min Scan# 879
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion: 45 Resp: 382308

Ion	Ratio	Lower	Upper
45	100		
77	18.0	0.0	37.7
79	14.4	0.0	33.5







#17
2-Methylphenol
Concen: 20.942 ng
RT: 8.175 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:107 Resp: 283057

Ion	Ratio	Lower	Upper
107	100		
108	112.9	89.8	134.6
77	43.0	34.2	51.4
79	41.4	33.8	50.8







#18
Hexachloroethane
Concen: 20.773 ng
RT: 8.805 min Scan# 972
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:117 Resp: 140558

Ion	Ratio	Lower	Upper
117	100		
119	97.5	77.8	116.6
201	96.0	75.9	113.9







#19
n-Nitroso-di-n-propylamine
Concen: 21.127 ng
RT: 8.528 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 70 Resp: 274118

Ion	Ratio	Lower	Upper
70	100		
42	46.5	37.3	55.9
101	10.3	8.2	12.4
130	22.8	18.6	28.0







#20
3+4-Methylphenols
Concen: 20.999 ng
RT: 8.499 min Scan# 920
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:107 Resp: 393131

Ion	Ratio	Lower	Upper
107	100		
108	90.3	69.8	109.8
77	32.3	11.5	51.5
79	26.9	7.8	47.8







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.505 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:136 Resp: 1081087

Ion	Ratio	Lower	Upper
136	100		
137	10.9	9.2	13.8
54	7.3	6.0	9.0
68	6.8	5.5	8.3







#22
Acetophenone
Concen: 20.991 ng
RT: 8.546 min Scan# 928
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:105 Resp: 563994

Ion	Ratio	Lower	Upper
105	100		
71	1.7	1.7	2.5
51	21.2	16.9	25.3
120	23.9	18.6	28.0







#23
Nitrobenzene-d5
Concen: 41.839 ng
RT: 8.875 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 82 Resp: 794485

Ion	Ratio	Lower	Upper
82	100		
128	45.1	36.5	54.7
54	41.7	33.4	50.2







#24
Nitrobenzene
Concen: 20.770 ng
RT: 8.916 min Scan# 991
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion: 77 Resp: 390911

Ion	Ratio	Lower	Upper
77	100		
123	46.9	39.2	58.8
65	12.3	9.8	14.6







#25
Isophorone
Concen: 20.588 ng
RT: 9.446 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 82 Resp: 705771

Ion	Ratio	Lower	Upper
82	100		
95	6.7	5.4	8.2
138	19.2	14.4	21.6







#26
2-Nitrophenol
Concen: 20.553 ng
RT: 9.628 min Scan# 1112
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:139 Resp: 185740

Ion	Ratio	Lower	Upper
139	100		
109	28.3	21.8	32.6
65	45.4	36.2	54.2







#27
2,4-Dimethylphenol
Concen: 20.643 ng
RT: 9.687 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:122 Resp: 236436

Ion	Ratio	Lower	Upper
122	100		
107	114.9	93.7	140.5
121	59.2	48.2	72.2







#28
bis(2-Chloroethoxy)methane
Concen: 20.575 ng
RT: 9.916 min Scan# 1161
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion: 93 Resp: 480036

Ion	Ratio	Lower	Upper
93	100		
95	32.3	26.0	39.0
123	11.8	9.1	13.7







#29
2,4-Dichlorophenol
Concen: 20.536 ng
RT: 10.163 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:162 Resp: 320579

Ion	Ratio	Lower	Upper
162	100		
164	63.7	43.3	83.3
98	35.6	16.6	56.6







#30
1,2,4-Trichlorobenzene
Concen: 20.348 ng
RT: 10.369 min Scan# 1238
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:180 Resp: 343641

Ion	Ratio	Lower	Upper
180	100		
182	95.5	77.3	115.9
145	32.3	25.1	37.7







#31
Naphthalene
Concen: 20.618 ng
RT: 10.557 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:128 Resp: 1179848

Ion	Ratio	Lower	Upper
128	100		
129	10.8	8.6	13.0
127	13.2	10.3	15.5







#32
Benzoic acid
Concen: 19.731 ng m
RT: 9.805 min Scan# 1142
Delta R.T. -0.024 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:122 Resp: 244213

Ion	Ratio	Lower	Upper
122	100		
105	124.1	104.8	144.8
77	89.8	69.2	109.2







#33
4-Chloroaniline
Concen: 21.010 ng
RT: 10.669 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:127 Resp: 420086

Ion	Ratio	Lower	Upper
127	100		
129	32.4	26.1	39.1
65	27.2	22.1	33.1
92	18.7	15.0	22.4







#34
Hexachlorobutadiene
Concen: 20.267 ng
RT: 10.846 min Scan# 1319
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:225 Resp: 200785

Ion	Ratio	Lower	Upper
225	100		
223	63.4	49.5	74.3
227	63.3	50.9	76.3







#35
Caprolactam
Concen: 20.922 ng
RT: 11.446 min Scan# 14
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:113 Resp: 119912

Ion	Ratio	Lower	Upper
113	100		
55	134.1	115.9	155.9
56	103.0	81.1	121.1







#36
4-Chloro-3-methylphenol
Concen: 20.767 ng
RT: 11.799 min Scan# 1481
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:107 Resp: 377768

Ion	Ratio	Lower	Upper
107	100		
144	26.5	20.5	30.7
142	80.4	63.4	95.2







#37
2-Methylnaphthalene
Concen: 20.583 ng
RT: 12.169 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:142 Resp: 816314

Ion	Ratio	Lower	Upper
142	100		
141	87.5	70.3	105.5
115	33.7	26.9	40.3







#38
1-Methylnaphthalene
Concen: 20.503 ng
RT: 12.387 min Scan# 1581
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:142 Resp: 796550

Ion	Ratio	Lower	Upper
142	100		
141	92.2	73.6	110.4
116	3.8	3.0	4.6







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.369 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:164 Resp: 688775

Ion	Ratio	Lower	Upper
164	100		
162	101.2	80.6	120.8
160	45.0	35.3	52.9







#40
1,2,4,5-Tetrachlorobenzene
Concen: 20.669 ng
RT: 12.540 min Scan# 1607
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:216 Resp: 391241

Ion	Ratio	Lower	Upper
216	100		
214	77.8	62.6	93.8
179	21.2	17.2	25.8
108	18.9	15.0	22.6







#41
Hexachlorocyclopentadiene
Concen: 20.414 ng
RT: 12.522 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:237 Resp: 135654

Ion	Ratio	Lower	Upper
237	100		
235	61.3	43.9	83.9
272	13.5	0.0	33.6







#42
2,4,6-Tribromophenol
Concen: 40.617 ng
RT: 15.875 min Scan# 2174
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:330 Resp: 383088

Ion	Ratio	Lower	Upper
330	100		
332	97.5	77.1	115.7
141	30.9	24.7	37.1







#43
2,4,6-Trichlorophenol
Concen: 20.410 ng
RT: 12.787 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:196 Resp: 254189

Ion	Ratio	Lower	Upper
196	100		
198	96.6	74.3	111.5
200	31.2	24.0	36.0







#44
2,4,5-Trichlorophenol
Concen: 20.424 ng
RT: 12.863 min Scan# 1662
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:196 Resp: 281887

Ion	Ratio	Lower	Upper
196	100		
198	96.9	77.3	115.9
97	48.9	40.4	60.6
132	29.0	23.4	35.0







#45
2-Fluorobiphenyl
Concen: 41.254 ng
RT: 12.975 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:172 Resp: 1871534

Ion	Ratio	Lower	Upper
172	100		
171	35.4	28.2	42.2
170	23.4	18.6	27.8







#46
1,1'-Biphenyl
Concen: 20.694 ng
RT: 13.187 min Scan# 1717
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:154 Resp: 1052618

Ion	Ratio	Lower	Upper
154	100		
153	40.1	20.3	60.3
76	13.7	0.0	33.5







#47
2-Chloronaphthalene
Concen: 20.603 ng
RT: 13.228 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:162 Resp: 782040

Ion	Ratio	Lower	Upper
162	100		
127	40.4	31.9	47.9
164	32.6	26.2	39.4







#48
2-Nitroaniline
Concen: 20.811 ng
RT: 13.440 min Scan# 1760
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion: 65 Resp: 218194

Ion	Ratio	Lower	Upper
65	100		
92	74.9	62.3	93.5
138	121.6	97.2	145.8







#49
Acenaphthylene
Concen: 20.726 ng
RT: 14.087 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:152 Resp: 1234728

Ion	Ratio	Lower	Upper
152	100		
151	19.9	15.9	23.9
153	12.7	10.6	15.8







#50
Dimethylphthalate
Concen: 20.489 ng
RT: 13.822 min Scan# 1825
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:163 Resp: 1029599

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.5	5.3
164	10.1	8.0	12.0







#51
2,6-Dinitrotoluene
Concen: 20.731 ng
RT: 13.940 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:165 Resp: 219492

Ion	Ratio	Lower	Upper
165	100		
63	48.3	38.0	57.0
89	53.4	41.7	62.5







#52
Acenaphthene
Concen: 20.741 ng
RT: 14.434 min Scan# 1929
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:154 Resp: 785789

Ion	Ratio	Lower	Upper
154	100		
153	113.9	89.8	134.6
152	53.8	42.6	63.8







#53
3-Nitroaniline
Concen: 21.212 ng
RT: 14.269 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:138 Resp: 233006

Ion	Ratio	Lower	Upper
138	100		
108	10.5	8.3	12.5
92	111.7	89.8	134.8







#54
2,4-Dinitrophenol
Concen: 19.785 ng
RT: 14.481 min Scan# 1937
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:184 Resp: 112157

Ion	Ratio	Lower	Upper
184	100		
63	61.0	42.6	64.0
154	60.0	47.2	70.8







#55
Dibenzofuran
Concen: 20.735 ng
RT: 14.769 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:168 Resp: 1288363

Ion	Ratio	Lower	Upper
168	100		
139	37.3	30.3	45.5
169	13.2	10.6	16.0







#56
4-Nitrophenol
Concen: 21.109 ng
RT: 14.593 min Scan# 1956
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:139 Resp: 199315

Ion	Ratio	Lower	Upper
139	100		
109	67.0	46.1	86.1
65	86.5	65.0	105.0







#57
2,4-Dinitrotoluene
Concen: 21.126 ng
RT: 14.734 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:165 Resp: 302332

Ion	Ratio	Lower	Upper
165	100		
63	37.0	30.4	45.6
89	66.2	54.6	82.0
182	3.0	2.2	3.2







#58
Fluorene
Concen: 20.876 ng
RT: 15.428 min Scan# 2098
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:166 Resp: 1001515

Ion	Ratio	Lower	Upper
166	100		
165	97.3	78.1	117.1
167	13.2	10.9	16.3







#59
2,3,4,6-Tetrachlorophenol
Concen: 20.523 ng
RT: 14.998 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:232 Resp: 262077

Ion	Ratio	Lower	Upper
232	100		
131	45.5	36.8	55.2
130	2.4	1.9	2.9
166	31.1	25.0	37.6







#60
Diethylphthalate
Concen: 20.344 ng
RT: 15.198 min Scan# 2059
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:149 Resp: 1052183

Ion	Ratio	Lower	Upper
149	100		
177	22.0	17.4	26.2
150	12.8	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 20.675 ng
RT: 15.422 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:204 Resp: 481631

Ion	Ratio	Lower	Upper
204	100		
206	32.6	26.5	39.7
141	59.2	49.4	74.0







#62
4-Nitroaniline
Concen: 20.888 ng
RT: 15.445 min Scan# 2101
Delta R.T. -0.012 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:138 Resp: 243257

Ion	Ratio	Lower	Upper
138	100		
92	51.2	27.8	67.8
108	72.0	48.9	88.9







#63
Azobenzene
Concen: 20.907 ng
RT: 15.722 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 77 Resp: 995912

Ion	Ratio	Lower	Upper
77	100		
182	27.3	7.3	47.3
105	22.0	2.3	42.3
51	23.1	3.4	43.4







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.169 min Scan# 2394
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:188 Resp: 1378480

Ion	Ratio	Lower	Upper
188	100		
94	10.2	8.6	13.0
80	11.1	9.8	14.6







#65
4,6-Dinitro-2-methylphenol
Concen: 20.176 ng
RT: 15.510 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:198 Resp: 162501

Ion	Ratio	Lower	Upper
198	100		
51	38.2	18.4	58.4
105	43.8	26.5	66.5







#66
n-Nitrosodiphenylamine
Concen: 20.552 ng
RT: 15.640 min Scan# 2134
Delta R.T. -0.012 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:169 Resp: 849476

Ion	Ratio	Lower	Upper
169	100		
168	66.6	53.0	79.4
167	36.1	28.9	43.3







#67
4-Bromophenyl-phenylether
Concen: 20.125 ng
RT: 16.334 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:248 Resp: 302601

Ion	Ratio	Lower	Upper
248	100		
250	96.8	78.2	117.4
141	71.8	57.7	86.5







#68
Hexachlorobenzene
Concen: 20.077 ng
RT: 16.451 min Scan# 2272
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:284 Resp: 358889

Ion	Ratio	Lower	Upper
284	100		
142	31.0	24.6	37.0
249	29.5	23.8	35.6







#69
Atrazine
Concen: 17.573 ng
RT: 16.616 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:200 Resp: 184151

Ion	Ratio	Lower	Upper
200	100		
173	26.9	7.5	47.5
215	47.4	28.8	68.8







#70
Pentachlorophenol
Concen: 19.702 ng
RT: 16.810 min Scan# 2333
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:266 Resp: 241815

Ion	Ratio	Lower	Upper
266	100		
268	65.1	51.0	76.6
264	63.6	49.8	74.6







#71
Phenanthrene
Concen: 20.378 ng
RT: 17.210 min Scan# 24
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:178 Resp: 1540112

Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.4	23.0
179	15.6	12.5	18.7







#72
Anthracene
Concen: 20.543 ng
RT: 17.298 min Scan# 2416
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:178 Resp: 1494006

Ion	Ratio	Lower	Upper
178	100		
176	18.9	14.7	22.1
179	15.1	12.4	18.6







#73
Carbazole
Concen: 20.653 ng
RT: 17.581 min Scan# 24
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:167 Resp: 1461729

Ion	Ratio	Lower	Upper
167	100		
166	21.4	17.1	25.7
139	12.8	10.6	16.0







#74
Di-n-butylphthalate
Concen: 19.585 ng
RT: 18.175 min Scan# 2565
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:149 Resp: 1727074

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.4	11.2
104	4.8	3.9	5.9







#75
Fluoranthene
Concen: 20.218 ng
RT: 19.292 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:202 Resp: 1813662

Ion	Ratio	Lower	Upper
202	100		
101	12.3	0.0	33.2
203	17.6	0.0	37.7







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.616 min Scan# 3150
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:240 Resp: 1415765

Ion	Ratio	Lower	Upper
240	100		
120	13.0	9.8	14.8
236	25.2	20.9	31.3







#77
Benzidine
Concen: 8.867 ng
RT: 19.498 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:184 Resp: 154964

Ion	Ratio	Lower	Upper
184	100		
185	14.9	11.6	17.4
183	11.1	9.2	13.8







#78
Pyrene
Concen: 21.028 ng
RT: 19.669 min Scan# 2819
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:202 Resp: 1904884

Ion	Ratio	Lower	Upper
202	100		
200	21.2	17.2	25.8
203	17.6	14.4	21.6







#79
Terphenyl-d14
Concen: 42.062 ng
RT: 19.898 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:244 Resp: 2987245

Ion	Ratio	Lower	Upper
244	100		
212	6.6	5.6	8.4
122	11.7	9.4	14.0







#80
Butylbenzylphthalate
Concen: 20.235 ng
RT: 20.633 min Scan# 2983
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:149 Resp: 772776

Ion	Ratio	Lower	Upper
149	100		
91	67.1	56.4	84.6
206	21.2	16.3	24.5







#81
Benzo(a)anthracene
Concen: 20.732 ng
RT: 21.598 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:228 Resp: 1829798

Ion	Ratio	Lower	Upper
228	100		
226	26.7	21.8	32.8
229	19.6	15.9	23.9







#82
3,3'-Dichlorobenzidine
Concen: 19.956 ng
RT: 21.527 min Scan# 3135
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:252 Resp: 610749

Ion	Ratio	Lower	Upper
252	100		
254	64.5	51.7	77.5
126	12.4	8.8	13.2







#83
Chrysene
Concen: 20.565 ng
RT: 21.669 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:228 Resp: 1739548

Ion	Ratio	Lower	Upper
228	100		
226	29.8	24.2	36.2
229	19.8	15.6	23.4







#84
Bis(2-ethylhexyl)phthalate
Concen: 19.616 ng
RT: 21.545 min Scan# 3138
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:149 Resp: 1103263

Ion	Ratio	Lower	Upper
149	100		
167	27.5	22.4	33.6
279	4.9	4.0	6.0







#85
Di-n-octyl phthalate
Concen: 19.207 ng
RT: 22.827 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:149 Resp: 1730436

Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	7.7	6.0	9.0







#86
Perylene-d12
Concen: 20.000 ng
RT: 24.986 min Scan# 3723
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:264 Resp: 1549781

Ion	Ratio	Lower	Upper
264	100		
260	23.4	18.9	28.3
265	21.7	17.8	26.6







#87
Indeno(1,2,3-cd)pyrene
Concen: 20.454 ng
RT: 28.792 min Scan# 4
Delta R.T. -0.018 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:276 Resp: 2180177

Ion	Ratio	Lower	Upper
276	100		
138	16.7	14.5	21.7
277	25.2	20.2	30.2







#88
Benzo(b)fluoranthene
Concen: 20.337 ng
RT: 23.916 min Scan# 3541
Delta R.T. -0.012 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:252 Resp: 1885654

Ion	Ratio	Lower	Upper
252	100		
253	21.4	17.7	26.5
125	11.7	9.4	14.0







#89
Benzo(k)fluoranthene
Concen: 20.955 ng
RT: 23.980 min Scan# 3
Delta R.T. -0.012 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:252 Resp: 1885747

Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.7	26.5
125	11.4	9.1	13.7







#90
Benzo(a)pyrene
Concen: 20.518 ng
RT: 24.816 min Scan# 3694
Delta R.T. -0.018 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:252 Resp: 1637648

Ion	Ratio	Lower	Upper
252	100		
253	21.3	17.6	26.4
125	13.2	10.1	15.1







#91
Dibenzo(a,h)anthracene
Concen: 20.480 ng
RT: 28.868 min Scan# 4
Delta R.T. -0.035 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:278 Resp: 1815743

Ion	Ratio	Lower	Upper
278	100		
139	16.4	12.6	19.0
279	24.2	18.5	27.7






#92
Benzo(g,h,i)perylene
Concen: 20.571 ng
RT: 29.951 min Scan# 4567
Delta R.T. -0.035 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:276 Resp: 1856380

Ion	Ratio	Lower	Upper
276	100		
277	24.1	18.9	28.3
138	22.6	18.7	28.1



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024279.D
Acq On : 14 Apr 2025 13:49
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Quant Time: Apr 14 17:26:48 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.734	152	246363	20.000	ng	0.00
21) Naphthalene-d8	10.505	136	1046876	20.000	ng	0.00
39) Acenaphthene-d10	14.363	164	657357	20.000	ng	0.00
64) Phenanthrene-d10	17.163	188	1277198	20.000	ng	0.00
76) Chrysene-d12	21.622	240	1349860	20.000	ng	0.00
86) Perylene-d12	24.986	264	1485135	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.352	112	1236220	82.813	ng	0.00
7) Phenol-d6	6.916	99	1722202	84.332	ng	0.00
23) Nitrobenzene-d5	8.875	82	1540790	83.791	ng	0.00
42) 2,4,6-Tribromophenol	15.881	330	759528	84.379	ng	0.00
45) 2-Fluorobiphenyl	12.969	172	3591010	82.939	ng	0.00
79) Terphenyl-d14	19.898	244	5570563	82.266	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.293	88	252519	39.624	ng	100
3) Pyridine	3.693	79	687730	41.457	ng	100
4) n-Nitrosodimethylamine	3.599	42	225273	40.725	ng	100
6) Aniline	7.069	93	781907	42.413	ng	100
8) 2-Chlorophenol	7.305	128	689257	41.317	ng	100
9) Benzaldehyde	6.881	77	411846	38.937	ng	100
10) Phenol	6.940	94	862244	42.050	ng	100
11) bis(2-Chloroethyl)ether	7.164	93	688989	41.581	ng	100
12) 1,3-Dichlorobenzene	7.622	146	732329	40.516	ng	100
13) 1,4-Dichlorobenzene	7.763	146	736270	40.171	ng	100
14) 1,2-Dichlorobenzene	8.081	146	711951	40.172	ng	100
15) Benzyl Alcohol	7.969	79	570437	43.581	ng	100
16) 2,2'-oxybis(1-Chloropr...	8.252	45	746041	41.386	ng	100
17) 2-Methylphenol	8.169	107	561621	42.503	ng	100
18) Hexachloroethane	8.799	117	266915	40.349	ng	100
19) n-Nitroso-di-n-propyla...	8.528	70	533912	42.092	ng	100
20) 3+4-Methylphenols	8.499	107	784336	42.854	ng	100
22) Acetophenone	8.546	105	1086788	41.770	ng	100
24) Nitrobenzene	8.922	77	762756	41.852	ng	100
25) Isophorone	9.440	82	1408146	42.419	ng	100
26) 2-Nitrophenol	9.628	139	383751	43.852	ng	100
27) 2,4-Dimethylphenol	9.687	122	472488	42.600	ng	100
28) bis(2-Chloroethoxy)met...	9.922	93	947250	41.926	ng	100
29) 2,4-Dichlorophenol	10.157	162	645849	42.725	ng	100
30) 1,2,4-Trichlorobenzene	10.369	180	671264	41.046	ng	100
31) Naphthalene	10.557	128	2258924	40.765	ng	100
32) Benzoic acid	9.828	122	506813	42.286	ng	100
33) 4-Chloroaniline	10.669	127	830870	42.913	ng	100
34) Hexachlorobutadiene	10.846	225	394737	41.146	ng	100
35) Caprolactam	11.452	113	234519	42.256	ng	100
36) 4-Chloro-3-methylphenol	11.793	107	747541	42.438	ng	100
37) 2-Methylnaphthalene	12.163	142	1580497	41.154	ng	100
38) 1-Methylnaphthalene	12.381	142	1530982	40.695	ng	100
40) 1,2,4,5-Tetrachloroben...	12.540	216	757748	41.945	ng	100
41) Hexachlorocyclopentadiene	12.522	237	286135	45.116	ng	100
43) 2,4,6-Trichlorophenol	12.781	196	511566	43.038	ng	100

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024279.D
Acq On : 14 Apr 2025 13:49
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Quant Time: Apr 14 17:26:48 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.857	196	561631	42.637	ng	100
46) 1,1'-Biphenyl	13.187	154	2010029	41.404	ng	100
47) 2-Chloronaphthalene	13.228	162	1497152	41.328	ng	100
48) 2-Nitroaniline	13.434	65	434050	43.377	ng	100
49) Acenaphthylene	14.081	152	2384785	41.944	ng	100
50) Dimethylphthalate	13.822	163	1965142	40.976	ng	100
51) 2,6-Dinitrotoluene	13.934	165	427248	42.283	ng	100
52) Acenaphthene	14.434	154	1480526	40.947	ng	100
53) 3-Nitroaniline	14.269	138	465347	44.389	ng	100
54) 2,4-Dinitrophenol	14.481	184	244714	45.231	ng	100
55) Dibenzofuran	14.775	168	2409526	40.633	ng	100
56) 4-Nitrophenol	14.593	139	401591	44.564	ng	100
57) 2,4-Dinitrotoluene	14.734	165	588798	43.110	ng	100
58) Fluorene	15.434	166	1838264	40.150	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.004	232	511714	41.987	ng	100
60) Diethylphthalate	15.204	149	2037863	41.286	ng	100
61) 4-Chlorophenyl-phenyle...	15.428	204	894350	40.226	ng	100
62) 4-Nitroaniline	15.457	138	485622	43.693	ng	100
63) Azobenzene	15.728	77	1881773	41.392	ng	100
65) 4,6-Dinitro-2-methylph...	15.510	198	335347	44.938	ng	100
66) n-Nitrosodiphenylamine	15.651	169	1609251	42.021	ng	100
67) 4-Bromophenyl-phenylether	16.340	248	587416	42.165	ng	100
68) Hexachlorobenzene	16.457	284	692546	41.815	ng	100
69) Atrazine	16.622	200	308618	31.786	ng	100
70) Pentachlorophenol	16.810	266	499858	43.955	ng	100
71) Phenanthrene	17.210	178	2879498	41.121	ng	100
72) Anthracene	17.304	178	2838724	42.130	ng	100
73) Carbazole	17.587	167	2770072	42.242	ng	100
74) Di-n-butylphthalate	18.175	149	3571220	43.708	ng	100
75) Fluoranthene	19.298	202	3441282	41.404	ng	100
77) Benzidine	19.498	184	1178258	70.714	ng	100
78) Pyrene	19.669	202	3523104	40.790	ng	100
80) Butylbenzylphthalate	20.628	149	1607881	44.158	ng	100
81) Benzo(a)anthracene	21.598	228	3489938	41.472	ng	100
82) 3,3'-Dichlorobenzidine	21.522	252	1311770	44.953	ng	100
83) Chrysene	21.675	228	3274153	40.596	ng	100
84) Bis(2-ethylhexyl)phtha...	21.539	149	2411998	44.979	ng	100
85) Di-n-octyl phthalate	22.827	149	3943939	45.913	ng	100
87) Indeno(1,2,3-cd)pyrene	28.809	276	4288630	41.987	ng	100
88) Benzo(b)fluoranthene	23.927	252	3737035	42.058	ng	100
89) Benzo(k)fluoranthene	23.992	252	3600626	41.753	ng	100
90) Benzo(a)pyrene	24.833	252	3240684	42.369	ng	100
91) Dibenzo(a,h)anthracene	28.904	278	3573562	42.061	ng	100
92) Benzo(g,h,i)perylene	29.986	276	3617069	41.826	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed


ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024279.D
Acq On : 14 Apr 2025 13:49
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Quant Time: Apr 14 17:26:48 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.734 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:152 Resp: 246363

Ion	Ratio	Lower	Upper
152	100		
150	156.1	124.9	187.3
115	59.6	47.7	71.5







#2
1,4-Dioxane
Concen: 39.624 ng
RT: 3.293 min Scan# 35
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion: 88 Resp: 252519

Ion	Ratio	Lower	Upper
88	100		
58	64.3	51.4	77.2
43	21.4	17.1	25.7







#3
Pyridine
Concen: 41.457 ng
RT: 3.693 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 79 Resp: 687730

Ion	Ratio	Lower	Upper
79	100		
52	56.8	45.4	68.2
51	28.2	22.6	33.8







#4
n-Nitrosodimethylamine
Concen: 40.725 ng
RT: 3.599 min Scan# 87
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion: 42 Resp: 225273

Ion	Ratio	Lower	Upper
42	100		
74	174.4	139.5	209.3
44	7.3	5.8	8.8







#5
2-Fluorophenol
Concen: 82.813 ng
RT: 5.352 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:112 Resp: 1236220

Ion	Ratio	Lower	Upper
112	100		
64	51.4	41.1	61.7
63	25.7	20.6	30.8







#6
Aniline
Concen: 42.413 ng
RT: 7.069 min Scan# 677
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion: 93 Resp: 781907

Ion	Ratio	Lower	Upper
93	100		
66	36.4	29.1	43.7
65	18.4	14.7	22.1







#7
Phenol-d6
Concen: 84.332 ng
RT: 6.916 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 99 Resp: 1722202

Ion	Ratio	Lower	Upper
99	100		
42	13.3	10.6	16.0
71	29.6	23.7	35.5







#8
2-Chlorophenol
Concen: 41.317 ng
RT: 7.305 min Scan# 717
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:128 Resp: 689257

Ion	Ratio	Lower	Upper
128	100		
130	32.9	12.9	52.9
64	42.7	22.7	62.7







#9
Benzaldehyde
Concen: 38.937 ng
RT: 6.881 min Scan# 64
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 77 Resp: 411846

Ion	Ratio	Lower	Upper
77	100		
105	99.2	79.2	119.2
106	96.2	76.2	116.2







#10
Phenol
Concen: 42.050 ng
RT: 6.940 min Scan# 655
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion: 94 Resp: 862244

Ion	Ratio	Lower	Upper
94	100		
65	24.5	4.5	44.5
66	32.4	12.4	52.4







#11
bis(2-Chloroethyl)ether
Concen: 41.581 ng
RT: 7.164 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 93 Resp: 688989

Ion	Ratio	Lower	Upper
93	100		
63	68.6	48.6	88.6
95	32.3	12.3	52.3







#12
1,3-Dichlorobenzene
Concen: 40.516 ng
RT: 7.622 min Scan# 771
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:146 Resp: 732329

Ion	Ratio	Lower	Upper
146	100		
148	64.2	51.4	77.0
75	28.6	22.9	34.3







#13
1,4-Dichlorobenzene
Concen: 40.171 ng
RT: 7.763 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:146 Resp: 736270

Ion	Ratio	Lower	Upper
146	100		
148	63.9	51.1	76.7
111	40.7	32.6	48.8







#14
1,2-Dichlorobenzene
Concen: 40.172 ng
RT: 8.081 min Scan# 849
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:146 Resp: 711951

Ion	Ratio	Lower	Upper
146	100		
148	63.7	51.0	76.4
111	42.3	33.8	50.8







#15
Benzyl Alcohol
Concen: 43.581 ng
RT: 7.969 min Scan# 8
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 79 Resp: 570437

Ion	Ratio	Lower	Upper
79	100		
108	81.4	65.1	97.7
77	63.4	50.7	76.1







#16
2,2'-oxybis(1-Chloropropane)
Concen: 41.386 ng
RT: 8.252 min Scan# 878
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion: 45 Resp: 746041

Ion	Ratio	Lower	Upper
45	100		
77	17.7	0.0	37.7
79	13.5	0.0	33.5







#17
2-Methylphenol
Concen: 42.503 ng
RT: 8.169 min Scan# 8
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:107 Resp: 561621

Ion	Ratio	Lower	Upper
107	100		
108	112.2	89.8	134.6
77	42.8	34.2	51.4
79	42.3	33.8	50.8







#18
Hexachloroethane
Concen: 40.349 ng
RT: 8.799 min Scan# 971
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:117 Resp: 266915

Ion	Ratio	Lower	Upper
117	100		
119	97.2	77.8	116.6
201	94.9	75.9	113.9







#19
n-Nitroso-di-n-propylamine
Concen: 42.092 ng
RT: 8.528 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 70 Resp: 533912

Ion	Ratio	Lower	Upper
70	100		
42	46.6	37.3	55.9
101	10.3	8.2	12.4
130	23.3	18.6	28.0







#20
3+4-Methylphenols
Concen: 42.854 ng
RT: 8.499 min Scan# 920
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:107 Resp: 784336

Ion	Ratio	Lower	Upper
107	100		
108	89.8	69.8	109.8
77	31.5	11.5	51.5
79	27.8	7.8	47.8







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.505 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:136 Resp: 1046876

Ion	Ratio	Lower	Upper
136	100		
137	11.5	9.2	13.8
54	7.5	6.0	9.0
68	6.9	5.5	8.3







#22
Acetophenone
Concen: 41.770 ng
RT: 8.546 min Scan# 928
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:105 Resp: 1086788

Ion	Ratio	Lower	Upper
105	100		
71	2.1	1.7	2.5
51	21.1	16.9	25.3
120	23.3	18.6	28.0







#23
Nitrobenzene-d5
Concen: 83.791 ng
RT: 8.875 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 82 Resp: 1540790

Ion	Ratio	Lower	Upper
82	100		
128	45.6	36.5	54.7
54	41.8	33.4	50.2







#24
Nitrobenzene
Concen: 41.852 ng
RT: 8.922 min Scan# 992
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion: 77 Resp: 762756

Ion	Ratio	Lower	Upper
77	100		
123	49.0	39.2	58.8
65	12.2	9.8	14.6







#25
Isophorone
Concen: 42.419 ng
RT: 9.440 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 82 Resp: 1408146

Ion	Ratio	Lower	Upper
82	100		
95	6.8	5.4	8.2
138	18.0	14.4	21.6







#26
2-Nitrophenol
Concen: 43.852 ng
RT: 9.628 min Scan# 1112
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:139 Resp: 383751

Ion	Ratio	Lower	Upper
139	100		
109	27.2	21.8	32.6
65	45.2	36.2	54.2







#27
2,4-Dimethylphenol
Concen: 42.600 ng
RT: 9.687 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:122 Resp: 472488

Ion	Ratio	Lower	Upper
122	100		
107	117.1	93.7	140.5
121	60.2	48.2	72.2







#28
bis(2-Chloroethoxy)methane
Concen: 41.926 ng
RT: 9.922 min Scan# 1162
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion: 93 Resp: 947250

Ion	Ratio	Lower	Upper
93	100		
95	32.5	26.0	39.0
123	11.4	9.1	13.7







#29
2,4-Dichlorophenol
Concen: 42.725 ng
RT: 10.157 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:162 Resp: 645849

Ion	Ratio	Lower	Upper
162	100		
164	63.3	43.3	83.3
98	36.6	16.6	56.6







#30
1,2,4-Trichlorobenzene
Concen: 41.046 ng
RT: 10.369 min Scan# 1238
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:180 Resp: 671264

Ion	Ratio	Lower	Upper
180	100		
182	96.6	77.3	115.9
145	31.4	25.1	37.7







#31
Naphthalene
Concen: 40.765 ng
RT: 10.557 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:128 Resp: 2258924

Ion	Ratio	Lower	Upper
128	100		
129	10.8	8.6	13.0
127	12.9	10.3	15.5







#32
Benzoic acid
Concen: 42.286 ng
RT: 9.828 min Scan# 1146
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:122 Resp: 506813

Ion	Ratio	Lower	Upper
122	100		
105	124.8	104.8	144.8
77	89.2	69.2	109.2







#33
4-Chloroaniline
Concen: 42.913 ng
RT: 10.669 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:127 Resp: 830870

Ion	Ratio	Lower	Upper
127	100		
129	32.6	26.1	39.1
65	27.6	22.1	33.1
92	18.7	15.0	22.4







#34
Hexachlorobutadiene
Concen: 41.146 ng
RT: 10.846 min Scan# 1319
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:225 Resp: 394737

Ion	Ratio	Lower	Upper
225	100		
223	61.9	49.5	74.3
227	63.6	50.9	76.3







#35
Caprolactam
Concen: 42.256 ng
RT: 11.452 min Scan# 14
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:113 Resp: 234519

Ion	Ratio	Lower	Upper
113	100		
55	135.9	115.9	155.9
56	101.1	81.1	121.1







#36
4-Chloro-3-methylphenol
Concen: 42.438 ng
RT: 11.793 min Scan# 1480
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:107 Resp: 747541

Ion	Ratio	Lower	Upper
107	100		
144	25.6	20.5	30.7
142	79.3	63.4	95.2







#37
2-Methylnaphthalene
Concen: 41.154 ng
RT: 12.163 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:142 Resp: 1580497

Ion	Ratio	Lower	Upper
142	100		
141	87.9	70.3	105.5
115	33.6	26.9	40.3







#38
1-Methylnaphthalene
Concen: 40.695 ng
RT: 12.381 min Scan# 1580
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:142 Resp: 1530982

Ion	Ratio	Lower	Upper
142	100		
141	92.0	73.6	110.4
116	3.8	3.0	4.6







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.363 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:164 Resp: 657357

Ion	Ratio	Lower	Upper
164	100		
162	100.7	80.6	120.8
160	44.1	35.3	52.9







#40
1,2,4,5-Tetrachlorobenzene
Concen: 41.945 ng
RT: 12.540 min Scan# 1607
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:216 Resp: 757748

Ion	Ratio	Lower	Upper
216	100		
214	78.2	62.6	93.8
179	21.5	17.2	25.8
108	18.8	15.0	22.6







#41
Hexachlorocyclopentadiene
Concen: 45.116 ng
RT: 12.522 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:237 Resp: 286135

Ion	Ratio	Lower	Upper
237	100		
235	63.9	43.9	83.9
272	13.6	0.0	33.6







#42
2,4,6-Tribromophenol
Concen: 84.379 ng
RT: 15.881 min Scan# 2175
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:330 Resp: 759528

Ion	Ratio	Lower	Upper
330	100		
332	96.4	77.1	115.7
141	30.9	24.7	37.1







#43
2,4,6-Trichlorophenol
Concen: 43.038 ng
RT: 12.781 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:196 Resp: 511566

Ion	Ratio	Lower	Upper
196	100		
198	92.9	74.3	111.5
200	30.0	24.0	36.0







#44
2,4,5-Trichlorophenol
Concen: 42.637 ng
RT: 12.857 min Scan# 1661
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:196 Resp: 561631

Ion	Ratio	Lower	Upper
196	100		
198	96.6	77.3	115.9
97	50.5	40.4	60.6
132	29.2	23.4	35.0







#45
2-Fluorobiphenyl
Concen: 82.939 ng
RT: 12.969 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:172 Resp: 3591010

Ion	Ratio	Lower	Upper
172	100		
171	35.2	28.2	42.2
170	23.2	18.6	27.8







#46
1,1'-Biphenyl
Concen: 41.404 ng
RT: 13.187 min Scan# 1717
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:154 Resp: 2010029

Ion	Ratio	Lower	Upper
154	100		
153	40.3	20.3	60.3
76	13.5	0.0	33.5







#47
2-Chloronaphthalene
Concen: 41.328 ng
RT: 13.228 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:162 Resp: 1497152

Ion	Ratio	Lower	Upper
162	100		
127	39.9	31.9	47.9
164	32.8	26.2	39.4







#48
2-Nitroaniline
Concen: 43.377 ng
RT: 13.434 min Scan# 1759
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion: 65 Resp: 434050

Ion	Ratio	Lower	Upper
65	100		
92	77.9	62.3	93.5
138	121.5	97.2	145.8







#49
Acenaphthylene
Concen: 41.944 ng
RT: 14.081 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:152 Resp: 2384785

Ion	Ratio	Lower	Upper
152	100		
151	19.9	15.9	23.9
153	13.2	10.6	15.8







#50
Dimethylphthalate
Concen: 40.976 ng
RT: 13.822 min Scan# 1825
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:163 Resp: 1965142

Ion	Ratio	Lower	Upper
163	100		
194	4.4	3.5	5.3
164	10.0	8.0	12.0







#51
2,6-Dinitrotoluene
Concen: 42.283 ng
RT: 13.934 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:165 Resp: 427248

Ion	Ratio	Lower	Upper
165	100		
63	47.5	38.0	57.0
89	52.1	41.7	62.5







#52
Acenaphthene
Concen: 40.947 ng
RT: 14.434 min Scan# 1929
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:154 Resp: 1480526

Ion	Ratio	Lower	Upper
154	100		
153	112.2	89.8	134.6
152	53.2	42.6	63.8







#53
3-Nitroaniline
Concen: 44.389 ng
RT: 14.269 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:138 Resp: 465347

Ion	Ratio	Lower	Upper
138	100		
108	10.4	8.3	12.5
92	112.3	89.8	134.8







#54
2,4-Dinitrophenol
Concen: 45.231 ng
RT: 14.481 min Scan# 1937
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:184 Resp: 244714

Ion	Ratio	Lower	Upper
184	100		
63	53.3	42.6	64.0
154	59.0	47.2	70.8







#55
Dibenzofuran
Concen: 40.633 ng
RT: 14.775 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:168 Resp: 2409526

Ion	Ratio	Lower	Upper
168	100		
139	37.9	30.3	45.5
169	13.3	10.6	16.0







#56
4-Nitrophenol
Concen: 44.564 ng
RT: 14.593 min Scan# 1956
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:139 Resp: 401591

Ion	Ratio	Lower	Upper
139	100		
109	66.1	46.1	86.1
65	85.0	65.0	105.0







#57
2,4-Dinitrotoluene
Concen: 43.110 ng
RT: 14.734 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:165 Resp: 588798

Ion	Ratio	Lower	Upper
165	100		
63	38.0	30.4	45.6
89	68.3	54.6	82.0
182	2.7	2.2	3.2







#58
Fluorene
Concen: 40.150 ng
RT: 15.434 min Scan# 2099
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:166 Resp: 1838264

Ion	Ratio	Lower	Upper
166	100		
165	97.6	78.1	117.1
167	13.6	10.9	16.3







#59
2,3,4,6-Tetrachlorophenol
Concen: 41.987 ng
RT: 15.004 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:232 Resp: 511714

Ion	Ratio	Lower	Upper
232	100		
131	46.0	36.8	55.2
130	2.4	1.9	2.9
166	31.3	25.0	37.6







#60
Diethylphthalate
Concen: 41.286 ng
RT: 15.204 min Scan# 2060
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:149 Resp: 2037863

Ion	Ratio	Lower	Upper
149	100		
177	21.8	17.4	26.2
150	12.6	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 40.226 ng
RT: 15.428 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:204 Resp: 894350

Ion	Ratio	Lower	Upper
204	100		
206	33.1	26.5	39.7
141	61.7	49.4	74.0







#62
4-Nitroaniline
Concen: 43.693 ng
RT: 15.457 min Scan# 2103
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:138 Resp: 485622

Ion	Ratio	Lower	Upper
138	100		
92	47.8	27.8	67.8
108	68.9	48.9	88.9







#63
Azobenzene
Concen: 41.392 ng
RT: 15.728 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 77 Resp: 1881773

Ion	Ratio	Lower	Upper
77	100		
182	27.3	7.3	47.3
105	22.3	2.3	42.3
51	23.4	3.4	43.4







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.163 min Scan# 2393
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:188 Resp: 1277198

Ion	Ratio	Lower	Upper
188	100		
94	10.8	8.6	13.0
80	12.2	9.8	14.6







#65
4,6-Dinitro-2-methylphenol
Concen: 44.938 ng
RT: 15.510 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:198 Resp: 335347

Ion	Ratio	Lower	Upper
198	100		
51	38.4	18.4	58.4
105	46.5	26.5	66.5







#66
n-Nitrosodiphenylamine
Concen: 42.021 ng
RT: 15.651 min Scan# 2136
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:169 Resp: 1609251

Ion	Ratio	Lower	Upper
169	100		
168	66.2	53.0	79.4
167	36.1	28.9	43.3







#67
4-Bromophenyl-phenylether
Concen: 42.165 ng
RT: 16.340 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:248 Resp: 587416

Ion	Ratio	Lower	Upper
248	100		
250	97.8	78.2	117.4
141	72.1	57.7	86.5







#68
Hexachlorobenzene
Concen: 41.815 ng
RT: 16.457 min Scan# 2273
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:284 Resp: 692546

Ion	Ratio	Lower	Upper
284	100		
142	30.8	24.6	37.0
249	29.7	23.8	35.6







#69
Atrazine
Concen: 31.786 ng
RT: 16.622 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:200 Resp: 308618

Ion	Ratio	Lower	Upper
200	100		
173	27.5	7.5	47.5
215	48.8	28.8	68.8







#70
Pentachlorophenol
Concen: 43.955 ng
RT: 16.810 min Scan# 2333
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:266 Resp: 499858

Ion	Ratio	Lower	Upper
266	100		
268	63.8	51.0	76.6
264	62.2	49.8	74.6







#71
Phenanthrene
Concen: 41.121 ng
RT: 17.210 min Scan# 24
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:178 Resp: 2879498

Ion	Ratio	Lower	Upper
178	100		
176	19.2	15.4	23.0
179	15.6	12.5	18.7







#72
Anthracene
Concen: 42.130 ng
RT: 17.304 min Scan# 2417
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:178 Resp: 2838724

Ion	Ratio	Lower	Upper
178	100		
176	18.4	14.7	22.1
179	15.5	12.4	18.6







#73
Carbazole
Concen: 42.242 ng
RT: 17.587 min Scan# 24
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:167 Resp: 2770072

Ion	Ratio	Lower	Upper
167	100		
166	21.4	17.1	25.7
139	13.3	10.6	16.0







#74
Di-n-butylphthalate
Concen: 43.708 ng
RT: 18.175 min Scan# 2565
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:149 Resp: 3571220

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.4	11.2
104	4.9	3.9	5.9







#75
Fluoranthene
Concen: 41.404 ng
RT: 19.298 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:202 Resp: 3441282

Ion	Ratio	Lower	Upper
202	100		
101	13.2	0.0	33.2
203	17.7	0.0	37.7







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.622 min Scan# 3151
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:240 Resp: 1349860

Ion	Ratio	Lower	Upper
240	100		
120	12.3	9.8	14.8
236	26.1	20.9	31.3







#77
Benzidine
Concen: 70.714 ng
RT: 19.498 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:184 Resp: 1178258

Ion	Ratio	Lower	Upper
184	100		
185	14.5	11.6	17.4
183	11.5	9.2	13.8







#78
Pyrene
Concen: 40.790 ng
RT: 19.669 min Scan# 2819
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:202 Resp: 3523104

Ion	Ratio	Lower	Upper
202	100		
200	21.5	17.2	25.8
203	18.0	14.4	21.6







#79
Terphenyl-d14
Concen: 82.266 ng
RT: 19.898 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:244 Resp: 5570563

Ion	Ratio	Lower	Upper
244	100		
212	7.0	5.6	8.4
122	11.7	9.4	14.0







#80
Butylbenzylphthalate
Concen: 44.158 ng
RT: 20.628 min Scan# 2982
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:149 Resp: 1607881

Ion	Ratio	Lower	Upper
149	100		
91	70.5	56.4	84.6
206	20.4	16.3	24.5







#81
Benzo(a)anthracene
Concen: 41.472 ng
RT: 21.598 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:228 Resp: 3489938

Ion	Ratio	Lower	Upper
228	100		
226	27.3	21.8	32.8
229	19.9	15.9	23.9







#82
3,3'-Dichlorobenzidine
Concen: 44.953 ng
RT: 21.522 min Scan# 3134
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:252 Resp: 1311770

Ion	Ratio	Lower	Upper
252	100		
254	64.6	51.7	77.5
126	11.0	8.8	13.2







#83
Chrysene
Concen: 40.596 ng
RT: 21.675 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:228 Resp: 3274153

Ion	Ratio	Lower	Upper
228	100		
226	30.2	24.2	36.2
229	19.5	15.6	23.4







#84
Bis(2-ethylhexyl)phthalate
Concen: 44.979 ng
RT: 21.539 min Scan# 3137
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:149 Resp: 2411998

Ion	Ratio	Lower	Upper
149	100		
167	28.0	22.4	33.6
279	5.0	4.0	6.0







#85
Di-n-octyl phthalate
Concen: 45.913 ng
RT: 22.827 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:149 Resp: 3943939

Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	7.5	6.0	9.0







#86
Perylene-d12
Concen: 20.000 ng
RT: 24.986 min Scan# 3723
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:264 Resp: 1485135

Ion	Ratio	Lower	Upper
264	100		
260	23.6	18.9	28.3
265	22.2	17.8	26.6







#87
Indeno(1,2,3-cd)pyrene
Concen: 41.987 ng
RT: 28.809 min Scan# 4
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:276 Resp: 4288630

Ion	Ratio	Lower	Upper
276	100		
138	18.1	14.5	21.7
277	25.2	20.2	30.2







#88
Benzo(b)fluoranthene
Concen: 42.058 ng
RT: 23.927 min Scan# 3543
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:252 Resp: 3737035

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	11.7	9.4	14.0







#89
Benzo(k)fluoranthene
Concen: 41.753 ng
RT: 23.992 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:252 Resp: 3600626

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	11.4	9.1	13.7







#90
Benzo(a)pyrene
Concen: 42.369 ng
RT: 24.833 min Scan# 3697
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:252 Resp: 3240684

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.6	26.4
125	12.6	10.1	15.1







#91
Dibenzo(a,h)anthracene
Concen: 42.061 ng
RT: 28.904 min Scan# 4
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:278 Resp: 3573562

Ion	Ratio	Lower	Upper
278	100		
139	15.8	12.6	19.0
279	23.1	18.5	27.7






#92
Benzo(g,h,i)perylene
Concen: 41.826 ng
RT: 29.986 min Scan# 4573
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:276 Resp: 3617069

Ion	Ratio	Lower	Upper
276	100		
277	23.6	18.9	28.3
138	23.4	18.7	28.1



ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024280.D
Acq On : 14 Apr 2025 15:10
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Quant Time: Apr 14 17:26:57 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.734	152	277364	20.000	ng		0.00
21) Naphthalene-d8	10.504	136	1176925	20.000	ng		0.00
39) Acenaphthene-d10	14.363	164	760235	20.000	ng		0.00
64) Phenanthrene-d10	17.163	188	1505097	20.000	ng		0.00
76) Chrysene-d12	21.616	240	1527193	20.000	ng		0.00
86) Perylene-d12	24.986	264	1706545	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.352	112	1677842	99.834	ng		0.00
7) Phenol-d6	6.916	99	2353543	102.366	ng		0.00
23) Nitrobenzene-d5	8.881	82	2100100	101.588	ng		0.00
42) 2,4,6-Tribromophenol	15.881	330	1098486	105.521	ng		0.00
45) 2-Fluorobiphenyl	12.975	172	4800845	95.877	ng		0.00
79) Terphenyl-d14	19.898	244	7671902	100.142	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.293	88	333474	46.479	ng		100
3) Pyridine	3.687	79	924382m	49.495	ng		
4) n-Nitrosodimethylamine	3.593	42	306279	49.180	ng		99
6) Aniline	7.075	93	972988	46.879	ng		100
8) 2-Chlorophenol	7.310	128	942078	50.160	ng		99
9) Benzaldehyde	6.887	77	507606	42.627	ng		97
10) Phenol	6.946	94	1177623	51.012	ng		99
11) bis(2-Chloroethyl)ether	7.163	93	915679	49.085	ng		100
12) 1,3-Dichlorobenzene	7.622	146	972089	47.770	ng		100
13) 1,4-Dichlorobenzene	7.769	146	995872	48.262	ng		99
14) 1,2-Dichlorobenzene	8.087	146	944314	47.327	ng		99
15) Benzyl Alcohol	7.975	79	791998	53.745	ng		99
16) 2,2'-oxybis(1-Chloropr...	8.257	45	973910	47.988	ng		99
17) 2-Methylphenol	8.175	107	766536	51.526	ng		100
18) Hexachloroethane	8.805	117	360999	48.472	ng		98
19) n-Nitroso-di-n-propyla...	8.534	70	722064	50.562	ng		98
20) 3+4-Methylphenols	8.505	107	1086273	52.718	ng		99
22) Acetophenone	8.546	105	1441728	49.289	ng	#	99
24) Nitrobenzene	8.922	77	1025983	50.074	ng		99
25) Isophorone	9.446	82	1913172	51.264	ng		99
26) 2-Nitrophenol	9.622	139	541332	55.024	ng		98
27) 2,4-Dimethylphenol	9.687	122	662183	53.106	ng		99
28) bis(2-Chloroethoxy)met...	9.922	93	1265001	49.803	ng		100
29) 2,4-Dichlorophenol	10.157	162	904519	53.225	ng		100
30) 1,2,4-Trichlorobenzene	10.369	180	908732	49.426	ng		98
31) Naphthalene	10.557	128	3049887	48.957	ng		100
32) Benzoic acid	9.851	122	794832	58.988	ng		99
33) 4-Chloroaniline	10.669	127	1110772	51.030	ng		99
34) Hexachlorobutadiene	10.840	225	533336	49.450	ng		99
35) Caprolactam	11.463	113	339834	54.466	ng		100
36) 4-Chloro-3-methylphenol	11.798	107	1046795	52.860	ng		99
37) 2-Methylnaphthalene	12.169	142	2156247	49.942	ng		99
38) 1-Methylnaphthalene	12.393	142	2095184	49.538	ng		99
40) 1,2,4,5-Tetrachloroben...	12.540	216	1034638	49.522	ng		99
41) Hexachlorocyclopentadiene	12.522	237	378578	51.614	ng		99
43) 2,4,6-Trichlorophenol	12.787	196	732902	53.315	ng		96

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024280.D
Acq On : 14 Apr 2025 15:10
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Quant Time: Apr 14 17:26:57 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.863	196	814885	53.491	ng	98
46) 1,1'-Biphenyl	13.187	154	2725840	48.551	ng	99
47) 2-Chloronaphthalene	13.234	162	2062098	49.220	ng	100
48) 2-Nitroaniline	13.434	65	616599	53.282	ng	98
49) Acenaphthylene	14.081	152	3279666	49.877	ng	100
50) Dimethylphthalate	13.822	163	2775759	50.046	ng	100
51) 2,6-Dinitrotoluene	13.934	165	606525	51.903	ng	98
52) Acenaphthene	14.428	154	2054333	49.128	ng	100
53) 3-Nitroaniline	14.275	138	649071	53.536	ng	98
54) 2,4-Dinitrophenol	14.481	184	373165	59.640	ng	99
55) Dibenzofuran	14.769	168	3318798	48.392	ng	100
56) 4-Nitrophenol	14.592	139	581023	55.750	ng	97
57) 2,4-Dinitrotoluene	14.734	165	839701	53.160	ng	99
58) Fluorene	15.428	166	2599105	49.085	ng	100
59) 2,3,4,6-Tetrachlorophenol	14.998	232	734741	52.128	ng	100
60) Diethylphthalate	15.204	149	2849294	49.914	ng	100
61) 4-Chlorophenyl-phenyle...	15.422	204	1267291	49.287	ng	99
62) 4-Nitroaniline	15.451	138	691081	53.764	ng	96
63) Azobenzene	15.722	77	2632843	50.076	ng	99
65) 4,6-Dinitro-2-methylph...	15.510	198	503765	57.285	ng	97
66) n-Nitrosodiphenylamine	15.645	169	2225475	49.312	ng	100
67) 4-Bromophenyl-phenylether	16.334	248	827245	50.389	ng	99
68) Hexachlorobenzene	16.457	284	983768	50.405	ng	98
69) Atrazine	16.622	200	610479	53.356	ng	100
70) Pentachlorophenol	16.810	266	739779	55.203	ng	99
71) Phenanthrene	17.210	178	4024282	48.768	ng	100
72) Anthracene	17.298	178	3960913	49.883	ng	100
73) Carbazole	17.581	167	3814743	49.364	ng	100
74) Di-n-butylphthalate	18.180	149	4930617	51.208	ng	100
75) Fluoranthene	19.298	202	4815021	49.160	ng	99
77) Benzidine	19.498	184	1272924	67.525	ng	99
78) Pyrene	19.675	202	4942328	50.577	ng	100
80) Butylbenzylphthalate	20.627	149	2215843	53.788	ng	97
81) Benzo(a)anthracene	21.598	228	4711301	49.485	ng	99
82) 3,3'-Dichlorobenzidine	21.527	252	1768950	53.582	ng	99
83) Chrysene	21.669	228	4480150	49.099	ng	100
84) Bis(2-ethylhexyl)phtha...	21.539	149	3256722	53.679	ng	99
85) Di-n-octyl phthalate	22.821	149	5458484	56.166	ng	99
87) Indeno(1,2,3-cd)pyrene	28.792	276	5977144m	50.926	ng	
88) Benzo(b)fluoranthene	23.915	252	5102465	49.975	ng	99
89) Benzo(k)fluoranthene	23.986	252	4923203	49.683	ng	99
90) Benzo(a)pyrene	24.821	252	4490517	51.093	ng	100
91) Dibenzo(a,h)anthracene	28.880	278	4987219	51.084	ng	98
92) Benzo(g,h,i)perylene	29.980	276	5033807	50.656	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024280.D
Acq On : 14 Apr 2025 15:10
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Quant Time: Apr 14 17:26:57 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.734 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:152 Resp: 277364

Ion	Ratio	Lower	Upper
152	100		
150	156.9	124.9	187.3
115	58.9	47.7	71.5







#2
1,4-Dioxane
Concen: 46.479 ng
RT: 3.293 min Scan# 35
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion: 88 Resp: 333474

Ion	Ratio	Lower	Upper
88	100		
58	64.0	51.4	77.2
43	21.2	17.1	25.7







#3
Pyridine
Concen: 49.495 ng m
RT: 3.687 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 79 Resp: 924382

Ion	Ratio	Lower	Upper
79	100		
52	58.2	45.4	68.2
51	27.7	22.6	33.8







#4
n-Nitrosodimethylamine
Concen: 49.180 ng
RT: 3.593 min Scan# 86
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion: 42 Resp: 306279

Ion	Ratio	Lower	Upper
42	100		
74	172.6	139.5	209.3
44	7.3	5.8	8.8







#5
2-Fluorophenol
Concen: 99.834 ng
RT: 5.352 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:112 Resp: 1677842

Ion	Ratio	Lower	Upper
112	100		
64	52.2	41.1	61.7
63	25.9	20.6	30.8







#6
Aniline
Concen: 46.879 ng
RT: 7.075 min Scan# 678
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion: 93 Resp: 972988

Ion	Ratio	Lower	Upper
93	100		
66	36.4	29.1	43.7
65	18.2	14.7	22.1







#7
Phenol-d6
Concen: 102.366 ng
RT: 6.916 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 99 Resp: 2353543

Ion	Ratio	Lower	Upper
99	100		
42	13.7	10.6	16.0
71	30.4	23.7	35.5







#8
2-Chlorophenol
Concen: 50.160 ng
RT: 7.310 min Scan# 718
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:128 Resp: 942078

Ion	Ratio	Lower	Upper
128	100		
130	32.7	12.9	52.9
64	41.2	22.7	62.7







#9
Benzaldehyde
Concen: 42.627 ng
RT: 6.887 min Scan# 64
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 77 Resp: 507606

Ion	Ratio	Lower	Upper
77	100		
105	101.5	79.2	119.2
106	99.2	76.2	116.2







#10
Phenol
Concen: 51.012 ng
RT: 6.946 min Scan# 656
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion: 94 Resp: 1177623

Ion	Ratio	Lower	Upper
94	100		
65	24.8	4.5	44.5
66	32.8	12.4	52.4







#11
bis(2-Chloroethyl)ether
Concen: 49.085 ng
RT: 7.163 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 93 Resp: 915679

Ion	Ratio	Lower	Upper
93	100		
63	68.6	48.6	88.6
95	31.9	12.3	52.3







#12
1,3-Dichlorobenzene
Concen: 47.770 ng
RT: 7.622 min Scan# 771
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:146 Resp: 972089

Ion	Ratio	Lower	Upper
146	100		
148	64.2	51.4	77.0
75	28.7	22.9	34.3







#13
1,4-Dichlorobenzene
Concen: 48.262 ng
RT: 7.769 min Scan# 7
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:146 Resp: 995872

Ion	Ratio	Lower	Upper
146	100		
148	64.4	51.1	76.7
111	40.3	32.6	48.8







#14
1,2-Dichlorobenzene
Concen: 47.327 ng
RT: 8.087 min Scan# 850
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:146 Resp: 944314

Ion	Ratio	Lower	Upper
146	100		
148	64.1	51.0	76.4
111	41.8	33.8	50.8







#15
Benzyl Alcohol
Concen: 53.745 ng
RT: 7.975 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 79 Resp: 791998

Ion	Ratio	Lower	Upper
79	100		
108	81.9	65.1	97.7
77	63.9	50.7	76.1







#16
2,2'-oxybis(1-Chloropropane)
Concen: 47.988 ng
RT: 8.257 min Scan# 879
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion: 45 Resp: 973910

Ion	Ratio	Lower	Upper
45	100		
77	17.6	0.0	37.7
79	14.2	0.0	33.5







#17
2-Methylphenol
Concen: 51.526 ng
RT: 8.175 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:107 Resp: 766536

Ion	Ratio	Lower	Upper
107	100		
108	111.5	89.8	134.6
77	42.7	34.2	51.4
79	42.2	33.8	50.8







#18
Hexachloroethane
Concen: 48.472 ng
RT: 8.805 min Scan# 972
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:117 Resp: 360999

Ion	Ratio	Lower	Upper
117	100		
119	96.9	77.8	116.6
201	98.2	75.9	113.9







#19
n-Nitroso-di-n-propylamine
Concen: 50.562 ng
RT: 8.534 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 70 Resp: 722064

Ion	Ratio	Lower	Upper
70	100		
42	45.2	37.3	55.9
101	10.6	8.2	12.4
130	23.7	18.6	28.0







#20
3+4-Methylphenols
Concen: 52.718 ng
RT: 8.505 min Scan# 921
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:107 Resp: 1086273

Ion	Ratio	Lower	Upper
107	100		
108	91.1	69.8	109.8
77	31.0	11.5	51.5
79	28.1	7.8	47.8







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.504 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:136 Resp: 1176925

Ion	Ratio	Lower	Upper
136	100		
137	10.9	9.2	13.8
54	6.9	6.0	9.0
68	6.4	5.5	8.3







#22
Acetophenone
Concen: 49.289 ng
RT: 8.546 min Scan# 928
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:105 Resp: 1441728

Ion	Ratio	Lower	Upper
105	100		
71	3.6	1.7	2.5#
51	21.1	16.9	25.3
120	23.9	18.6	28.0







#23
Nitrobenzene-d5
Concen: 101.588 ng
RT: 8.881 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 82 Resp: 2100100

Ion	Ratio	Lower	Upper
82	100		
128	46.9	36.5	54.7
54	41.8	33.4	50.2







#24
Nitrobenzene
Concen: 50.074 ng
RT: 8.922 min Scan# 992
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion: 77 Resp: 1025983

Ion	Ratio	Lower	Upper
77	100		
123	48.7	39.2	58.8
65	12.7	9.8	14.6







#25
Isophorone
Concen: 51.264 ng
RT: 9.446 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 82 Resp: 1913172

Ion	Ratio	Lower	Upper
82	100		
95	7.0	5.4	8.2
138	18.7	14.4	21.6







#26
2-Nitrophenol
Concen: 55.024 ng
RT: 9.622 min Scan# 1111
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:139 Resp: 541332

Ion	Ratio	Lower	Upper
139	100		
109	27.5	21.8	32.6
65	46.7	36.2	54.2







#27
2,4-Dimethylphenol
Concen: 53.106 ng
RT: 9.687 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:122 Resp: 662183

Ion	Ratio	Lower	Upper
122	100		
107	116.6	93.7	140.5
121	58.7	48.2	72.2







#28
bis(2-Chloroethoxy)methane
Concen: 49.803 ng
RT: 9.922 min Scan# 1162
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion: 93 Resp: 1265001

Ion	Ratio	Lower	Upper
93	100		
95	32.4	26.0	39.0
123	11.8	9.1	13.7







#29
2,4-Dichlorophenol
Concen: 53.225 ng
RT: 10.157 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:162 Resp: 904519

Ion	Ratio	Lower	Upper
162	100		
164	63.5	43.3	83.3
98	36.6	16.6	56.6







#30
1,2,4-Trichlorobenzene
Concen: 49.426 ng
RT: 10.369 min Scan# 1238
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:180 Resp: 908732

Ion	Ratio	Lower	Upper
180	100		
182	94.5	77.3	115.9
145	31.4	25.1	37.7







#31
Naphthalene
Concen: 48.957 ng
RT: 10.557 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:128 Resp: 3049887

Ion	Ratio	Lower	Upper
128	100		
129	10.8	8.6	13.0
127	13.1	10.3	15.5







#32
Benzoic acid
Concen: 58.988 ng
RT: 9.851 min Scan# 1150
Delta R.T. 0.023 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:122 Resp: 794832

Ion	Ratio	Lower	Upper
122	100		
105	124.7	104.8	144.8
77	90.5	69.2	109.2







#33
4-Chloroaniline
Concen: 51.030 ng
RT: 10.669 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:127 Resp: 1110772

Ion	Ratio	Lower	Upper
127	100		
129	31.7	26.1	39.1
65	26.8	22.1	33.1
92	18.4	15.0	22.4







#34
Hexachlorobutadiene
Concen: 49.450 ng
RT: 10.840 min Scan# 1318
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:225 Resp: 533336

Ion	Ratio	Lower	Upper
225	100		
223	63.1	49.5	74.3
227	64.1	50.9	76.3







#35
Caprolactam
Concen: 54.466 ng
RT: 11.463 min Scan# 14
Delta R.T. 0.012 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:113 Resp: 339834

Ion	Ratio	Lower	Upper
113	100		
55	136.1	115.9	155.9
56	100.2	81.1	121.1







#36
4-Chloro-3-methylphenol
Concen: 52.860 ng
RT: 11.798 min Scan# 1481
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:107 Resp: 1046795

Ion	Ratio	Lower	Upper
107	100		
144	25.4	20.5	30.7
142	80.1	63.4	95.2







#37
2-Methylnaphthalene
Concen: 49.942 ng
RT: 12.169 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:142 Resp: 2156247

Ion	Ratio	Lower	Upper
142	100		
141	88.8	70.3	105.5
115	34.5	26.9	40.3







#38
1-Methylnaphthalene
Concen: 49.538 ng
RT: 12.393 min Scan# 1582
Delta R.T. 0.012 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:142 Resp: 2095184

Ion	Ratio	Lower	Upper
142	100		
141	91.2	73.6	110.4
116	3.7	3.0	4.6







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.363 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:164 Resp: 760235

Ion	Ratio	Lower	Upper
164	100		
162	101.6	80.6	120.8
160	44.8	35.3	52.9







#40
1,2,4,5-Tetrachlorobenzene
Concen: 49.522 ng
RT: 12.540 min Scan# 1607
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:216 Resp: 1034638

Ion	Ratio	Lower	Upper
216	100		
214	78.9	62.6	93.8
179	21.4	17.2	25.8
108	19.0	15.0	22.6







#41
Hexachlorocyclopentadiene
Concen: 51.614 ng
RT: 12.522 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:237 Resp: 378578

Ion	Ratio	Lower	Upper
237	100		
235	63.4	43.9	83.9
272	13.7	0.0	33.6







#42
2,4,6-Tribromophenol
Concen: 105.521 ng
RT: 15.881 min Scan# 2175
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:330 Resp: 1098486

Ion	Ratio	Lower	Upper
330	100		
332	96.8	77.1	115.7
141	29.5	24.7	37.1







#43
2,4,6-Trichlorophenol
Concen: 53.315 ng
RT: 12.787 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:196 Resp: 732902

Ion	Ratio	Lower	Upper
196	100		
198	96.7	74.3	111.5
200	31.2	24.0	36.0







#44
2,4,5-Trichlorophenol
Concen: 53.491 ng
RT: 12.863 min Scan# 1662
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:196 Resp: 814885

Ion	Ratio	Lower	Upper
196	100		
198	94.8	77.3	115.9
97	48.1	40.4	60.6
132	28.3	23.4	35.0







#45
2-Fluorobiphenyl
Concen: 95.877 ng
RT: 12.975 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:172 Resp: 4800845

Ion	Ratio	Lower	Upper
172	100		
171	35.8	28.2	42.2
170	23.9	18.6	27.8







#46
1,1'-Biphenyl
Concen: 48.551 ng
RT: 13.187 min Scan# 1717
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:154 Resp: 2725840

Ion	Ratio	Lower	Upper
154	100		
153	40.6	20.3	60.3
76	13.9	0.0	33.5







#47
2-Chloronaphthalene
Concen: 49.220 ng
RT: 13.234 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:162 Resp: 2062098

Ion	Ratio	Lower	Upper
162	100		
127	39.5	31.9	47.9
164	32.7	26.2	39.4







#48
2-Nitroaniline
Concen: 53.282 ng
RT: 13.434 min Scan# 1759
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion: 65 Resp: 616599

Ion	Ratio	Lower	Upper
65	100		
92	76.6	62.3	93.5
138	123.3	97.2	145.8







#49
Acenaphthylene
Concen: 49.877 ng
RT: 14.081 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:152 Resp: 3279666

Ion	Ratio	Lower	Upper
152	100		
151	19.9	15.9	23.9
153	13.0	10.6	15.8







#50
Dimethylphthalate
Concen: 50.046 ng
RT: 13.822 min Scan# 1825
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:163 Resp: 2775759

Ion	Ratio	Lower	Upper
163	100		
194	4.3	3.5	5.3
164	10.0	8.0	12.0







#51
2,6-Dinitrotoluene
Concen: 51.903 ng
RT: 13.934 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:165 Resp: 606525

Ion	Ratio	Lower	Upper
165	100		
63	49.3	38.0	57.0
89	53.4	41.7	62.5







#52
Acenaphthene
Concen: 49.128 ng
RT: 14.428 min Scan# 1928
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:154 Resp: 2054333

Ion	Ratio	Lower	Upper
154	100		
153	112.5	89.8	134.6
152	53.4	42.6	63.8







#53
3-Nitroaniline
Concen: 53.536 ng
RT: 14.275 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:138 Resp: 649071

Ion	Ratio	Lower	Upper
138	100		
108	10.2	8.3	12.5
92	110.0	89.8	134.8







#54
2,4-Dinitrophenol
Concen: 59.640 ng
RT: 14.481 min Scan# 1937
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:184 Resp: 373165

Ion	Ratio	Lower	Upper
184	100		
63	53.7	42.6	64.0
154	57.8	47.2	70.8







#55
Dibenzofuran
Concen: 48.392 ng
RT: 14.769 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:168 Resp: 3318798

Ion	Ratio	Lower	Upper
168	100		
139	38.1	30.3	45.5
169	13.4	10.6	16.0







#56
4-Nitrophenol
Concen: 55.750 ng
RT: 14.592 min Scan# 1956
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:139 Resp: 581023

Ion	Ratio	Lower	Upper
139	100		
109	68.1	46.1	86.1
65	87.3	65.0	105.0







#57
2,4-Dinitrotoluene
Concen: 53.160 ng
RT: 14.734 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:165 Resp: 839701

Ion	Ratio	Lower	Upper
165	100		
63	37.5	30.4	45.6
89	69.1	54.6	82.0
182	2.9	2.2	3.2







#58
Fluorene
Concen: 49.085 ng
RT: 15.428 min Scan# 2098
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:166 Resp: 2599105

Ion	Ratio	Lower	Upper
166	100		
165	97.2	78.1	117.1
167	13.5	10.9	16.3







#59
2,3,4,6-Tetrachlorophenol
Concen: 52.128 ng
RT: 14.998 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:232 Resp: 734741

Ion	Ratio	Lower	Upper
232	100		
131	46.0	36.8	55.2
130	2.3	1.9	2.9
166	31.2	25.0	37.6







#60
Diethylphthalate
Concen: 49.914 ng
RT: 15.204 min Scan# 2060
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:149 Resp: 2849294

Ion	Ratio	Lower	Upper
149	100		
177	22.1	17.4	26.2
150	12.6	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 49.287 ng
RT: 15.422 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:204 Resp: 1267291

Ion	Ratio	Lower	Upper
204	100		
206	32.9	26.5	39.7
141	62.8	49.4	74.0







#62
4-Nitroaniline
Concen: 53.764 ng
RT: 15.451 min Scan# 2102
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:138 Resp: 691081

Ion	Ratio	Lower	Upper
138	100		
92	50.2	27.8	67.8
108	72.0	48.9	88.9







#63
Azobenzene
Concen: 50.076 ng
RT: 15.722 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 77 Resp: 2632843

Ion	Ratio	Lower	Upper
77	100		
182	26.3	7.3	47.3
105	22.1	2.3	42.3
51	23.9	3.4	43.4







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.163 min Scan# 2393
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:188 Resp: 1505097

Ion	Ratio	Lower	Upper
188	100		
94	10.8	8.6	13.0
80	12.1	9.8	14.6







#65
4,6-Dinitro-2-methylphenol
Concen: 57.285 ng
RT: 15.510 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:198 Resp: 503765

Ion	Ratio	Lower	Upper
198	100		
51	37.1	18.4	58.4
105	44.5	26.5	66.5







#66
n-Nitrosodiphenylamine
Concen: 49.312 ng
RT: 15.645 min Scan# 2135
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:169 Resp: 2225475

Ion	Ratio	Lower	Upper
169	100		
168	66.2	53.0	79.4
167	35.9	28.9	43.3







#67
4-Bromophenyl-phenylether
Concen: 50.389 ng
RT: 16.334 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:248 Resp: 827245

Ion	Ratio	Lower	Upper
248	100		
250	97.5	78.2	117.4
141	73.8	57.7	86.5







#68
Hexachlorobenzene
Concen: 50.405 ng
RT: 16.457 min Scan# 2273
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:284 Resp: 983768

Ion	Ratio	Lower	Upper
284	100		
142	32.4	24.6	37.0
249	29.7	23.8	35.6







#69
Atrazine
Concen: 53.356 ng
RT: 16.622 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:200 Resp: 610479

Ion	Ratio	Lower	Upper
200	100		
173	27.4	7.5	47.5
215	49.1	28.8	68.8







#70
Pentachlorophenol
Concen: 55.203 ng
RT: 16.810 min Scan# 2333
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:266 Resp: 739779

Ion	Ratio	Lower	Upper
266	100		
268	64.9	51.0	76.6
264	62.9	49.8	74.6







#71
Phenanthrene
Concen: 48.768 ng
RT: 17.210 min Scan# 24
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:178 Resp: 4024282

Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.4	23.0
179	15.5	12.5	18.7







#72
Anthracene
Concen: 49.883 ng
RT: 17.298 min Scan# 2416
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:178 Resp: 3960913

Ion	Ratio	Lower	Upper
178	100		
176	18.7	14.7	22.1
179	15.5	12.4	18.6







#73
Carbazole
Concen: 49.364 ng
RT: 17.581 min Scan# 24
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:167 Resp: 3814743

Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.1	25.7
139	13.3	10.6	16.0







#74
Di-n-butylphthalate
Concen: 51.208 ng
RT: 18.180 min Scan# 2566
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:149 Resp: 4930617

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.4	11.2
104	5.0	3.9	5.9







#75
Fluoranthene
Concen: 49.160 ng
RT: 19.298 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:202 Resp: 4815021

Ion	Ratio	Lower	Upper
202	100		
101	13.0	0.0	33.2
203	17.4	0.0	37.7







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.616 min Scan# 3150
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:240 Resp: 1527193

Ion	Ratio	Lower	Upper
240	100		
120	11.7	9.8	14.8
236	25.6	20.9	31.3







#77
Benzidine
Concen: 67.525 ng
RT: 19.498 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:184 Resp: 1272924

Ion	Ratio	Lower	Upper
184	100		
185	14.2	11.6	17.4
183	11.7	9.2	13.8







#78
Pyrene
Concen: 50.577 ng
RT: 19.675 min Scan# 2820
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:202 Resp: 4942328

Ion	Ratio	Lower	Upper
202	100		
200	21.4	17.2	25.8
203	17.7	14.4	21.6







#79
Terphenyl-d14
Concen: 100.142 ng
RT: 19.898 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:244 Resp: 7671902

Ion	Ratio	Lower	Upper
244	100		
212	7.1	5.6	8.4
122	11.2	9.4	14.0







#80
Butylbenzylphthalate
Concen: 53.788 ng
RT: 20.627 min Scan# 2982
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:149 Resp: 2215843

Ion	Ratio	Lower	Upper
149	100		
91	67.6	56.4	84.6
206	21.3	16.3	24.5







#81
Benzo(a)anthracene
Concen: 49.485 ng
RT: 21.598 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:228 Resp: 4711301

Ion	Ratio	Lower	Upper
228	100		
226	27.7	21.8	32.8
229	20.0	15.9	23.9







#82
3,3'-Dichlorobenzidine
Concen: 53.582 ng
RT: 21.527 min Scan# 3135
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:252 Resp: 1768950

Ion	Ratio	Lower	Upper
252	100		
254	64.2	51.7	77.5
126	11.3	8.8	13.2







#83
Chrysene
Concen: 49.099 ng
RT: 21.669 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:228 Resp: 4480150

Ion	Ratio	Lower	Upper
228	100		
226	30.2	24.2	36.2
229	19.7	15.6	23.4







#84
Bis(2-ethylhexyl)phthalate
Concen: 53.679 ng
RT: 21.539 min Scan# 3137
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:149 Resp: 3256722

Ion	Ratio	Lower	Upper
149	100		
167	28.3	22.4	33.6
279	5.1	4.0	6.0







#85
Di-n-octyl phthalate
Concen: 56.166 ng
RT: 22.821 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:149 Resp: 5458484

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	7.2	6.0	9.0







#86
Perylene-d12
Concen: 20.000 ng
RT: 24.986 min Scan# 3723
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:264 Resp: 1706545

Ion	Ratio	Lower	Upper
264	100		
260	23.4	18.9	28.3
265	21.5	17.8	26.6







#87
Indeno(1,2,3-cd)pyrene
Concen: 50.926 ng m
RT: 28.792 min Scan# 4
Delta R.T. -0.018 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:276 Resp: 5977144

Ion	Ratio	Lower	Upper
276	100		
138	17.7	14.5	21.7
277	20.8	20.2	30.2







#88
Benzo(b)fluoranthene
Concen: 49.975 ng
RT: 23.915 min Scan# 3541
Delta R.T. -0.012 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:252 Resp: 5102465

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.7	26.5
125	10.7	9.4	14.0







#89
Benzo(k)fluoranthene
Concen: 49.683 ng
RT: 23.986 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:252 Resp: 4923203

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.7	26.5
125	11.0	9.1	13.7







#90
Benzo(a)pyrene
Concen: 51.093 ng
RT: 24.821 min Scan# 3695
Delta R.T. -0.012 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:252 Resp: 4490517

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.6	26.4
125	12.5	10.1	15.1







#91
Dibenzo(a,h)anthracene
Concen: 51.084 ng
RT: 28.880 min Scan# 4
Delta R.T. -0.024 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:278 Resp: 4987219

Ion	Ratio	Lower	Upper
278	100		
139	17.1	12.6	19.0
279	24.0	18.5	27.7






#92
Benzo(g,h,i)perylene
Concen: 50.656 ng
RT: 29.980 min Scan# 4572
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:276 Resp: 5033807

Ion	Ratio	Lower	Upper
276	100		
277	23.7	18.9	28.3
138	22.4	18.7	28.1



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024281.D
Acq On : 14 Apr 2025 16:32
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Quant Time: Apr 14 17:27:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.728	152	283549	20.000	ng		0.00
21) Naphthalene-d8	10.510	136	1206231	20.000	ng		0.00
39) Acenaphthene-d10	14.363	164	777459	20.000	ng		0.00
64) Phenanthrene-d10	17.169	188	1536805	20.000	ng		0.00
76) Chrysene-d12	21.621	240	1596148	20.000	ng		0.00
86) Perylene-d12	24.986	264	1858365	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.352	112	2173071	126.481	ng		0.00
7) Phenol-d6	6.916	99	3009690	128.049	ng		0.00
23) Nitrobenzene-d5	8.881	82	2579126	121.729	ng		0.00
42) 2,4,6-Tribromophenol	15.881	330	1382037	129.818	ng		0.00
45) 2-Fluorobiphenyl	12.975	172	5835791	113.964	ng		0.00
79) Terphenyl-d14	19.898	244	9228969	115.262	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.293	88	417124	56.869	ng		100
3) Pyridine	3.687	79	1176097m	61.599	ng		
4) n-Nitrosodimethylamine	3.599	42	391176	61.443	ng		99
6) Aniline	7.069	93	1214636	57.245	ng		100
8) 2-Chlorophenol	7.305	128	1190812	62.021	ng		99
9) Benzaldehyde	6.881	77	617878	50.755	ng		99
10) Phenol	6.946	94	1513303	64.123	ng		99
11) bis(2-Chloroethyl)ether	7.163	93	1155297	60.579	ng		100
12) 1,3-Dichlorobenzene	7.622	146	1218832	58.589	ng		100
13) 1,4-Dichlorobenzene	7.763	146	1244205	58.981	ng		100
14) 1,2-Dichlorobenzene	8.081	146	1199988	58.830	ng		100
15) Benzyl Alcohol	7.975	79	1002308	66.533	ng		100
16) 2,2'-oxybis(1-Chloropr...	8.257	45	1187992	57.259	ng		99
17) 2-Methylphenol	8.175	107	978531	64.342	ng		99
18) Hexachloroethane	8.804	117	453349	59.545	ng		98
19) n-Nitroso-di-n-propyla...	8.534	70	883042	60.486	ng		99
20) 3+4-Methylphenols	8.499	107	1367005	64.895	ng		99
22) Acetophenone	8.546	105	1792666	59.797	ng	#	99
24) Nitrobenzene	8.922	77	1276229	60.775	ng		100
25) Isophorone	9.446	82	2365423	61.842	ng		99
26) 2-Nitrophenol	9.628	139	690735	68.504	ng		99
27) 2,4-Dimethylphenol	9.693	122	826335	64.660	ng		98
28) bis(2-Chloroethoxy)met...	9.922	93	1539129	59.124	ng		99
29) 2,4-Dichlorophenol	10.157	162	1124170	64.543	ng		99
30) 1,2,4-Trichlorobenzene	10.369	180	1138896	60.440	ng		99
31) Naphthalene	10.557	128	3790924	59.373	ng		99
32) Benzoic acid	9.869	122	1032910m	74.795	ng		
33) 4-Chloroaniline	10.675	127	1372601	61.527	ng		100
34) Hexachlorobutadiene	10.845	225	660129	59.719	ng		99
35) Caprolactam	11.475	113	437691	68.445	ng		99
36) 4-Chloro-3-methylphenol	11.804	107	1314421	64.762	ng		98
37) 2-Methylnaphthalene	12.169	142	2676555	60.487	ng		99
38) 1-Methylnaphthalene	12.387	142	2599119	59.960	ng		100
40) 1,2,4,5-Tetrachloroben...	12.540	216	1281552	59.981	ng		100
41) Hexachlorocyclopentadiene	12.522	237	459486	61.257	ng		99
43) 2,4,6-Trichlorophenol	12.787	196	909999	64.732	ng		98

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024281.D
Acq On : 14 Apr 2025 16:32
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Quant Time: Apr 14 17:27:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.863	196	1017414	65.306	ng	99
46) 1,1'-Biphenyl	13.187	154	3330176	58.001	ng	99
47) 2-Chloronaphthalene	13.228	162	2542212	59.336	ng	99
48) 2-Nitroaniline	13.439	65	781584	66.042	ng	98
49) Acenaphthylene	14.087	152	4111553	61.143	ng	99
50) Dimethylphthalate	13.828	163	3404463	60.022	ng	100
51) 2,6-Dinitrotoluene	13.939	165	764810	63.998	ng	98
52) Acenaphthene	14.434	154	2516981	58.858	ng	99
53) 3-Nitroaniline	14.275	138	793796	64.022	ng	99
54) 2,4-Dinitrophenol	14.486	184	490497	76.655	ng	97
55) Dibenzofuran	14.775	168	4124640	58.810	ng	100
56) 4-Nitrophenol	14.598	139	754038	70.748	ng	99
57) 2,4-Dinitrotoluene	14.739	165	1079633	66.836	ng	98
58) Fluorene	15.434	166	3184960	58.817	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.004	232	909497	63.097	ng	100
60) Diethylphthalate	15.204	149	3495577	59.879	ng	100
61) 4-Chlorophenyl-phenyle...	15.428	204	1563042	59.442	ng	99
62) 4-Nitroaniline	15.457	138	886163	67.414	ng	98
63) Azobenzene	15.728	77	3229873	60.070	ng	98
65) 4,6-Dinitro-2-methylph...	15.516	198	647253	72.083	ng	94
66) n-Nitrosodiphenylamine	15.651	169	2805137	60.874	ng	100
67) 4-Bromophenyl-phenylether	16.339	248	1038071	61.926	ng	98
68) Hexachlorobenzene	16.457	284	1239550	62.200	ng	97
70) Pentachlorophenol	16.816	266	936562	68.445	ng	100
71) Phenanthrene	17.222	178	4988840	59.209	ng	99
72) Anthracene	17.310	178	4938117	60.907	ng	99
73) Carbazole	17.592	167	4780408	60.584	ng	99
74) Di-n-butylphthalate	18.186	149	6167206	62.730	ng	100
75) Fluoranthene	19.304	202	6009748	60.092	ng	98
77) Benzidine	19.504	184	1446166	73.401	ng	100
78) Pyrene	19.680	202	6250781	61.204	ng	100
80) Butylbenzylphthalate	20.621	149	2840001	65.961	ng	98
81) Benzo(a)anthracene	21.604	228	6002059	60.318	ng	99
82) 3,3'-Dichlorobenzidine	21.521	252	2323294	67.333	ng	100
83) Chrysene	21.674	228	5714775	59.924	ng	99
84) Bis(2-ethylhexyl)phtha...	21.545	149	4153128	65.497	ng	99
85) Di-n-octyl phthalate	22.839	149	7116217	70.060	ng	99
87) Indeno(1,2,3-cd)pyrene	28.821	276	8057576m	63.043	ng	
88) Benzo(b)fluoranthene	23.927	252	6715444	60.400	ng	99
89) Benzo(k)fluoranthene	24.004	252	6500782	60.244	ng	99
90) Benzo(a)pyrene	24.827	252	5964163	62.316	ng	99
91) Dibenzo(a,h)anthracene	28.892	278	6614173	62.214	ng	98
92) Benzo(g,h,i)perylene	30.003	276	6740929	62.294	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed


ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024281.D
Acq On : 14 Apr 2025 16:32
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Quant Time: Apr 14 17:27:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.728 min Scan# 7
Delta R.T. -0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:152 Resp: 283549

Ion	Ratio	Lower	Upper
152	100		
150	156.8	124.9	187.3
115	61.7	47.7	71.5







#2
1,4-Dioxane
Concen: 56.869 ng
RT: 3.293 min Scan# 35
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion: 88 Resp: 417124

Ion	Ratio	Lower	Upper
88	100		
58	64.4	51.4	77.2
43	21.4	17.1	25.7







#3
Pyridine
Concen: 61.599 ng m
RT: 3.687 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 79 Resp: 1176097

Ion	Ratio	Lower	Upper
79	100		
52	57.7	45.4	68.2
51	28.4	22.6	33.8







#4
n-Nitrosodimethylamine
Concen: 61.443 ng
RT: 3.599 min Scan# 87
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion: 42 Resp: 391176

Ion	Ratio	Lower	Upper
42	100		
74	176.0	139.5	209.3
44	6.7	5.8	8.8







#5
2-Fluorophenol
Concen: 126.481 ng
RT: 5.352 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:112 Resp: 2173071

Ion	Ratio	Lower	Upper
112	100		
64	51.6	41.1	61.7
63	25.5	20.6	30.8







#6
Aniline
Concen: 57.245 ng
RT: 7.069 min Scan# 677
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion: 93 Resp: 1214636

Ion	Ratio	Lower	Upper
93	100		
66	36.6	29.1	43.7
65	18.1	14.7	22.1







#7
Phenol-d6
Concen: 128.049 ng
RT: 6.916 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 99 Resp: 3009690

Ion	Ratio	Lower	Upper
99	100		
42	13.8	10.6	16.0
71	30.8	23.7	35.5







#8
2-Chlorophenol
Concen: 62.021 ng
RT: 7.305 min Scan# 717
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:128 Resp: 1190812

Ion	Ratio	Lower	Upper
128	100		
130	32.3	12.9	52.9
64	42.3	22.7	62.7







#9
Benzaldehyde
Concen: 50.755 ng
RT: 6.881 min Scan# 64
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 77 Resp: 617878

Ion	Ratio	Lower	Upper
77	100		
105	98.5	79.2	119.2
106	94.7	76.2	116.2







#10
Phenol
Concen: 64.123 ng
RT: 6.946 min Scan# 656
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion: 94 Resp: 1513303

Ion	Ratio	Lower	Upper
94	100		
65	24.8	4.5	44.5
66	32.8	12.4	52.4







#11
bis(2-Chloroethyl)ether
Concen: 60.579 ng
RT: 7.163 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 93 Resp: 1155297

Ion	Ratio	Lower	Upper
93	100		
63	68.8	48.6	88.6
95	32.2	12.3	52.3







#12
1,3-Dichlorobenzene
Concen: 58.589 ng
RT: 7.622 min Scan# 771
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:146 Resp: 1218832

Ion	Ratio	Lower	Upper
146	100		
148	64.1	51.4	77.0
75	28.7	22.9	34.3







#13
1,4-Dichlorobenzene
Concen: 58.981 ng
RT: 7.763 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:146 Resp: 1244205

Ion	Ratio	Lower	Upper
146	100		
148	63.9	51.1	76.7
111	41.2	32.6	48.8







#14
1,2-Dichlorobenzene
Concen: 58.830 ng
RT: 8.081 min Scan# 849
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:146 Resp: 1199988

Ion	Ratio	Lower	Upper
146	100		
148	63.5	51.0	76.4
111	42.0	33.8	50.8







#15
Benzyl Alcohol
Concen: 66.533 ng
RT: 7.975 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 79 Resp: 1002308

Ion	Ratio	Lower	Upper
79	100		
108	81.4	65.1	97.7
77	62.9	50.7	76.1







#16
2,2'-oxybis(1-Chloropropane)
Concen: 57.259 ng
RT: 8.257 min Scan# 879
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion: 45 Resp: 1187992

Ion	Ratio	Lower	Upper
45	100		
77	17.9	0.0	37.7
79	14.3	0.0	33.5







#17
2-Methylphenol
Concen: 64.342 ng
RT: 8.175 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:107 Resp: 978531

Ion	Ratio	Lower	Upper
107	100		
108	111.3	89.8	134.6
77	43.4	34.2	51.4
79	42.4	33.8	50.8







#18
Hexachloroethane
Concen: 59.545 ng
RT: 8.804 min Scan# 972
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:117 Resp: 453349

Ion	Ratio	Lower	Upper
117	100		
119	96.7	77.8	116.6
201	99.1	75.9	113.9







#19
n-Nitroso-di-n-propylamine
Concen: 60.486 ng
RT: 8.534 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 70 Resp: 883042

Ion	Ratio	Lower	Upper
70	100		
42	45.8	37.3	55.9
101	10.7	8.2	12.4
130	23.9	18.6	28.0







#20
3+4-Methylphenols
Concen: 64.895 ng
RT: 8.499 min Scan# 920
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:107 Resp: 1367005

Ion	Ratio	Lower	Upper
107	100		
108	89.0	69.8	109.8
77	31.1	11.5	51.5
79	26.7	7.8	47.8







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.510 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:136 Resp: 1206231

Ion	Ratio	Lower	Upper
136	100		
137	10.6	9.2	13.8
54	6.7	6.0	9.0
68	6.0	5.5	8.3







#22
Acetophenone
Concen: 59.797 ng
RT: 8.546 min Scan# 928
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:105 Resp: 1792666

Ion	Ratio	Lower	Upper
105	100		
71	3.4	1.7	2.5#
51	21.4	16.9	25.3
120	23.4	18.6	28.0







#23
Nitrobenzene-d5
Concen: 121.729 ng
RT: 8.881 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 82 Resp: 2579126

Ion	Ratio	Lower	Upper
82	100		
128	47.8	36.5	54.7
54	41.9	33.4	50.2







#24
Nitrobenzene
Concen: 60.775 ng
RT: 8.922 min Scan# 992
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion: 77 Resp: 1276229

Ion	Ratio	Lower	Upper
77	100		
123	49.1	39.2	58.8
65	12.4	9.8	14.6







#25
Isophorone
Concen: 61.842 ng
RT: 9.446 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 82 Resp: 2365423

Ion	Ratio	Lower	Upper
82	100		
95	7.1	5.4	8.2
138	18.7	14.4	21.6







#26
2-Nitrophenol
Concen: 68.504 ng
RT: 9.628 min Scan# 1112
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:139 Resp: 690735

Ion	Ratio	Lower	Upper
139	100		
109	27.1	21.8	32.6
65	44.6	36.2	54.2







#27
2,4-Dimethylphenol
Concen: 64.660 ng
RT: 9.693 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:122 Resp: 826335

Ion	Ratio	Lower	Upper
122	100		
107	115.3	93.7	140.5
121	59.3	48.2	72.2







#28
bis(2-Chloroethoxy)methane
Concen: 59.124 ng
RT: 9.922 min Scan# 1162
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion: 93 Resp: 1539129

Ion	Ratio	Lower	Upper
93	100		
95	32.8	26.0	39.0
123	11.9	9.1	13.7







#29
2,4-Dichlorophenol
Concen: 64.543 ng
RT: 10.157 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:162 Resp: 1124170

Ion	Ratio	Lower	Upper
162	100		
164	64.5	43.3	83.3
98	37.0	16.6	56.6







#30
1,2,4-Trichlorobenzene
Concen: 60.440 ng
RT: 10.369 min Scan# 1238
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:180 Resp: 1138896

Ion	Ratio	Lower	Upper
180	100		
182	95.5	77.3	115.9
145	30.8	25.1	37.7







#31
Naphthalene
Concen: 59.373 ng
RT: 10.557 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:128 Resp: 3790924

Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.6	13.0
127	13.2	10.3	15.5







#32
Benzoic acid
Concen: 74.795 ng m
RT: 9.869 min Scan# 1153
Delta R.T. 0.041 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:122 Resp: 1032910

Ion	Ratio	Lower	Upper
122	100		
105	123.4	104.8	144.8
77	89.7	69.2	109.2







#33
4-Chloroaniline
Concen: 61.527 ng
RT: 10.675 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:127 Resp: 1372601

Ion	Ratio	Lower	Upper
127	100		
129	32.4	26.1	39.1
65	27.2	22.1	33.1
92	18.9	15.0	22.4







#34
Hexachlorobutadiene
Concen: 59.719 ng
RT: 10.845 min Scan# 1319
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:225 Resp: 660129

Ion	Ratio	Lower	Upper
225	100		
223	62.2	49.5	74.3
227	64.1	50.9	76.3







#35
Caprolactam
Concen: 68.445 ng
RT: 11.475 min Scan# 14
Delta R.T. 0.023 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:113 Resp: 437691

Ion	Ratio	Lower	Upper
113	100		
55	134.0	115.9	155.9
56	101.6	81.1	121.1







#36
4-Chloro-3-methylphenol
Concen: 64.762 ng
RT: 11.804 min Scan# 1482
Delta R.T. 0.012 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:107 Resp: 1314421

Ion	Ratio	Lower	Upper
107	100		
144	26.3	20.5	30.7
142	81.7	63.4	95.2







#37
2-Methylnaphthalene
Concen: 60.487 ng
RT: 12.169 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:142 Resp: 2676555

Ion	Ratio	Lower	Upper
142	100		
141	89.0	70.3	105.5
115	34.4	26.9	40.3







#38
1-Methylnaphthalene
Concen: 59.960 ng
RT: 12.387 min Scan# 1581
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:142 Resp: 2599119

Ion	Ratio	Lower	Upper
142	100		
141	92.3	73.6	110.4
116	3.7	3.0	4.6







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.363 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:164 Resp: 777459

Ion	Ratio	Lower	Upper
164	100		
162	100.8	80.6	120.8
160	44.6	35.3	52.9







#40
1,2,4,5-Tetrachlorobenzene
Concen: 59.981 ng
RT: 12.540 min Scan# 1607
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:216 Resp: 1281552

Ion	Ratio	Lower	Upper
216	100		
214	78.7	62.6	93.8
179	21.4	17.2	25.8
108	18.7	15.0	22.6







#41
Hexachlorocyclopentadiene
Concen: 61.257 ng
RT: 12.522 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:237 Resp: 459486

Ion	Ratio	Lower	Upper
237	100		
235	62.7	43.9	83.9
272	13.7	0.0	33.6







#42
2,4,6-Tribromophenol
Concen: 129.818 ng
RT: 15.881 min Scan# 2175
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:330 Resp: 1382037

Ion	Ratio	Lower	Upper
330	100		
332	96.7	77.1	115.7
141	30.0	24.7	37.1







#43
2,4,6-Trichlorophenol
Concen: 64.732 ng
RT: 12.787 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:196 Resp: 909999

Ion	Ratio	Lower	Upper
196	100		
198	95.1	74.3	111.5
200	31.2	24.0	36.0







#44
2,4,5-Trichlorophenol
Concen: 65.306 ng
RT: 12.863 min Scan# 1662
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:196 Resp: 1017414

Ion	Ratio	Lower	Upper
196	100		
198	96.2	77.3	115.9
97	48.5	40.4	60.6
132	28.9	23.4	35.0







#45
2-Fluorobiphenyl
Concen: 113.964 ng
RT: 12.975 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:172 Resp: 5835791

Ion	Ratio	Lower	Upper
172	100		
171	35.7	28.2	42.2
170	23.8	18.6	27.8







#46
1,1'-Biphenyl
Concen: 58.001 ng
RT: 13.187 min Scan# 1717
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:154 Resp: 3330176

Ion	Ratio	Lower	Upper
154	100		
153	40.6	20.3	60.3
76	13.8	0.0	33.5







#47
2-Chloronaphthalene
Concen: 59.336 ng
RT: 13.228 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:162 Resp: 2542212

Ion	Ratio	Lower	Upper
162	100		
127	41.3	31.9	47.9
164	32.8	26.2	39.4







#48
2-Nitroaniline
Concen: 66.042 ng
RT: 13.439 min Scan# 1760
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion: 65 Resp: 781584

Ion	Ratio	Lower	Upper
65	100		
92	75.9	62.3	93.5
138	123.3	97.2	145.8







#49
Acenaphthylene
Concen: 61.143 ng
RT: 14.087 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:152 Resp: 4111553

Ion	Ratio	Lower	Upper
152	100		
151	20.2	15.9	23.9
153	13.1	10.6	15.8







#50
Dimethylphthalate
Concen: 60.022 ng
RT: 13.828 min Scan# 1826
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:163 Resp: 3404463

Ion	Ratio	Lower	Upper
163	100		
194	4.3	3.5	5.3
164	10.1	8.0	12.0







#51
2,6-Dinitrotoluene
Concen: 63.998 ng
RT: 13.939 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:165 Resp: 764810

Ion	Ratio	Lower	Upper
165	100		
63	45.9	38.0	57.0
89	51.1	41.7	62.5







#52
Acenaphthene
Concen: 58.858 ng
RT: 14.434 min Scan# 1929
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:154 Resp: 2516981

Ion	Ratio	Lower	Upper
154	100		
153	113.6	89.8	134.6
152	53.7	42.6	63.8







#53
3-Nitroaniline
Concen: 64.022 ng
RT: 14.275 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:138 Resp: 793796

Ion	Ratio	Lower	Upper
138	100		
108	10.7	8.3	12.5
92	111.3	89.8	134.8







#54
2,4-Dinitrophenol
Concen: 76.655 ng
RT: 14.486 min Scan# 1938
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:184 Resp: 490497

Ion	Ratio	Lower	Upper
184	100		
63	50.8	42.6	64.0
154	57.7	47.2	70.8







#55
Dibenzofuran
Concen: 58.810 ng
RT: 14.775 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:168 Resp: 4124640

Ion	Ratio	Lower	Upper
168	100		
139	37.9	30.3	45.5
169	13.3	10.6	16.0







#56
4-Nitrophenol
Concen: 70.748 ng
RT: 14.598 min Scan# 1957
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:139 Resp: 754038

Ion	Ratio	Lower	Upper
139	100		
109	66.1	46.1	86.1
65	82.9	65.0	105.0







#57
2,4-Dinitrotoluene
Concen: 66.836 ng
RT: 14.739 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:165 Resp: 1079633

Ion	Ratio	Lower	Upper
165	100		
63	36.6	30.4	45.6
89	66.4	54.6	82.0
182	2.9	2.2	3.2







#58
Fluorene
Concen: 58.817 ng
RT: 15.434 min Scan# 2099
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:166 Resp: 3184960

Ion	Ratio	Lower	Upper
166	100		
165	97.2	78.1	117.1
167	13.4	10.9	16.3







#59
2,3,4,6-Tetrachlorophenol
Concen: 63.097 ng
RT: 15.004 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:232 Resp: 909497

Ion	Ratio	Lower	Upper
232	100		
131	46.4	36.8	55.2
130	2.3	1.9	2.9
166	31.4	25.0	37.6







#60
Diethylphthalate
Concen: 59.879 ng
RT: 15.204 min Scan# 2060
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:149 Resp: 3495577

Ion	Ratio	Lower	Upper
149	100		
177	21.9	17.4	26.2
150	12.9	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 59.442 ng
RT: 15.428 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:204 Resp: 1563042

Ion	Ratio	Lower	Upper
204	100		
206	33.3	26.5	39.7
141	61.1	49.4	74.0







#62
4-Nitroaniline
Concen: 67.414 ng
RT: 15.457 min Scan# 2103
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:138 Resp: 886163

Ion	Ratio	Lower	Upper
138	100		
92	48.6	27.8	67.8
108	70.7	48.9	88.9







#63
Azobenzene
Concen: 60.070 ng
RT: 15.728 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 77 Resp: 3229873

Ion	Ratio	Lower	Upper
77	100		
182	25.5	7.3	47.3
105	22.1	2.3	42.3
51	24.3	3.4	43.4







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.169 min Scan# 2394
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:188 Resp: 1536805

Ion	Ratio	Lower	Upper
188	100		
94	10.7	8.6	13.0
80	12.0	9.8	14.6







#65
4,6-Dinitro-2-methylphenol
Concen: 72.083 ng
RT: 15.516 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:198 Resp: 647253

Ion	Ratio	Lower	Upper
198	100		
51	34.1	18.4	58.4
105	42.6	26.5	66.5







#66
n-Nitrosodiphenylamine
Concen: 60.874 ng
RT: 15.651 min Scan# 2136
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:169 Resp: 2805137

Ion	Ratio	Lower	Upper
169	100		
168	66.6	53.0	79.4
167	36.3	28.9	43.3







#67
4-Bromophenyl-phenylether
Concen: 61.926 ng
RT: 16.339 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:248 Resp: 1038071

Ion	Ratio	Lower	Upper
248	100		
250	96.7	78.2	117.4
141	74.2	57.7	86.5







#68
Hexachlorobenzene
Concen: 62.200 ng
RT: 16.457 min Scan# 2273
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:284 Resp: 1239550

Ion	Ratio	Lower	Upper
284	100		
142	32.9	24.6	37.0
249	30.7	23.8	35.6







#70
Pentachlorophenol
Concen: 68.445 ng
RT: 16.816 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:266 Resp: 936562

Ion	Ratio	Lower	Upper
266	100		
268	64.2	51.0	76.6
264	62.3	49.8	74.6







#71
Phenanthrene
Concen: 59.209 ng
RT: 17.222 min Scan# 2403
Delta R.T. 0.012 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:178 Resp: 4988840

Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.4	23.0
179	15.3	12.5	18.7







#72
Anthracene
Concen: 60.907 ng
RT: 17.310 min Scan# 24
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:178 Resp: 4938117

Ion	Ratio	Lower	Upper
178	100		
176	18.9	14.7	22.1
179	15.7	12.4	18.6







#73
Carbazole
Concen: 60.584 ng
RT: 17.592 min Scan# 2466
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:167 Resp: 4780408

Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.1	25.7
139	13.7	10.6	16.0







#74
Di-n-butylphthalate
Concen: 62.730 ng
RT: 18.186 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:149 Resp: 6167206

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.4	11.2
104	5.0	3.9	5.9







#75
Fluoranthene
Concen: 60.092 ng
RT: 19.304 min Scan# 2757
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:202 Resp: 6009748

Ion	Ratio	Lower	Upper
202	100		
101	11.9	0.0	33.2
203	17.6	0.0	37.7







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.621 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:240 Resp: 1596148

Ion	Ratio	Lower	Upper
240	100		
120	12.5	9.8	14.8
236	25.8	20.9	31.3







#77
Benzidine
Concen: 73.401 ng
RT: 19.504 min Scan# 2791
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:184 Resp: 1446166

Ion	Ratio	Lower	Upper
184	100		
185	14.4	11.6	17.4
183	11.5	9.2	13.8







#78
Pyrene
Concen: 61.204 ng
RT: 19.680 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:202 Resp: 6250781

Ion	Ratio	Lower	Upper
202	100		
200	21.8	17.2	25.8
203	17.9	14.4	21.6







#79
Terphenyl-d14
Concen: 115.262 ng
RT: 19.898 min Scan# 2858
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:244 Resp: 9228969

Ion	Ratio	Lower	Upper
244	100		
212	7.6	5.6	8.4
122	10.9	9.4	14.0







#80
Butylbenzylphthalate
Concen: 65.961 ng
RT: 20.621 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:149 Resp: 2840001

Ion	Ratio	Lower	Upper
149	100		
91	72.0	56.4	84.6
206	20.7	16.3	24.5







#81
Benzo(a)anthracene
Concen: 60.318 ng
RT: 21.604 min Scan# 3148
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:228 Resp: 6002059

Ion	Ratio	Lower	Upper
228	100		
226	27.9	21.8	32.8
229	20.1	15.9	23.9







#82
3,3'-Dichlorobenzidine
Concen: 67.333 ng
RT: 21.521 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:252 Resp: 2323294

Ion	Ratio	Lower	Upper
252	100		
254	64.8	51.7	77.5
126	11.5	8.8	13.2







#83
Chrysene
Concen: 59.924 ng
RT: 21.674 min Scan# 3160
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:228 Resp: 5714775

Ion	Ratio	Lower	Upper
228	100		
226	30.8	24.2	36.2
229	20.2	15.6	23.4







#84
Bis(2-ethylhexyl)phthalate
Concen: 65.497 ng
RT: 21.545 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:149 Resp: 4153128

Ion	Ratio	Lower	Upper
149	100		
167	28.6	22.4	33.6
279	5.3	4.0	6.0







#85
Di-n-octyl phthalate
Concen: 70.060 ng
RT: 22.839 min Scan# 3358
Delta R.T. 0.012 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:149 Resp: 7116217

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	7.2	6.0	9.0







#86
Perylene-d12
Concen: 20.000 ng
RT: 24.986 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:264 Resp: 1858365

Ion	Ratio	Lower	Upper
264	100		
260	23.3	18.9	28.3
265	21.8	17.8	26.6







#87
Indeno(1,2,3-cd)pyrene
Concen: 63.043 ng m
RT: 28.821 min Scan# 4375
Delta R.T. 0.012 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:276 Resp: 8057576

Ion	Ratio	Lower	Upper
276	100		
138	18.2	14.5	21.7
277	20.7	20.2	30.2







#88
Benzo(b)fluoranthene
Concen: 60.400 ng
RT: 23.927 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:252 Resp: 6715444

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	11.0	9.4	14.0







#89
Benzo(k)fluoranthene
Concen: 60.244 ng
RT: 24.004 min Scan# 3556
Delta R.T. 0.012 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:252 Resp: 6500782

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.7	26.5
125	10.4	9.1	13.7







#90
Benzo(a)pyrene
Concen: 62.316 ng
RT: 24.827 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:252 Resp: 5964163

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.6	26.4
125	11.8	10.1	15.1







#91
Dibenzo(a,h)anthracene
Concen: 62.214 ng
RT: 28.892 min Scan# 4387
Delta R.T. -0.012 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:278 Resp: 6614173

Ion	Ratio	Lower	Upper
278	100		
139	16.6	12.6	19.0
279	24.1	18.5	27.7






#92
Benzo(g,h,i)perylene
Concen: 62.294 ng
RT: 30.003 min Scan# 4
Delta R.T. 0.017 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:276 Resp: 6740929

Ion	Ratio	Lower	Upper
276	100		
277	24.0	18.9	28.3
138	21.5	18.7	28.1



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024282.D
Acq On : 14 Apr 2025 17:13
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Quant Time: Apr 14 18:14:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.728	152	244732	20.000	ng		0.00
21) Naphthalene-d8	10.504	136	1024316	20.000	ng		0.00
39) Acenaphthene-d10	14.369	164	641442	20.000	ng		0.00
64) Phenanthrene-d10	17.180	188	1279313	20.000	ng		0.02
76) Chrysene-d12	21.621	240	1378105	20.000	ng		0.00
86) Perylene-d12	24.986	264	1583281	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.352	112	2341577	157.905	ng		0.00
7) Phenol-d6	6.916	99	3224074	158.927	ng		0.00
23) Nitrobenzene-d5	8.881	82	2847003	158.236	ng		0.00
42) 2,4,6-Tribromophenol	15.892	330	1507534	171.633	ng		0.01
45) 2-Fluorobiphenyl	12.981	172	6362479	150.596	ng		0.01
79) Terphenyl-d14	19.922	244	10209149	147.678	ng		0.02
Target Compounds							Qvalue
2) 1,4-Dioxane	3.293	88	467234	73.805	ng		99
3) Pyridine	3.687	79	1349044m	81.864	ng		
4) n-Nitrosodimethylamine	3.593	42	434861	79.138	ng		97
6) Aniline	7.069	93	1351337	73.789	ng		100
8) 2-Chlorophenol	7.305	128	1299769	78.433	ng		99
9) Benzaldehyde	6.881	77	608272	57.891	ng		99
10) Phenol	6.946	94	1608027	78.944	ng		98
11) bis(2-Chloroethyl)ether	7.163	93	1272327	77.297	ng		99
12) 1,3-Dichlorobenzene	7.622	146	1344292	74.869	ng		100
13) 1,4-Dichlorobenzene	7.763	146	1369087	75.195	ng		99
14) 1,2-Dichlorobenzene	8.075	146	1310407	74.432	ng		100
15) Benzyl Alcohol	7.975	79	1098602	84.492	ng		99
16) 2,2'-oxybis(1-Chloropr...	8.257	45	1319830	73.704	ng		98
17) 2-Methylphenol	8.175	107	1064555	81.101	ng		100
18) Hexachloroethane	8.799	117	498897	75.920	ng		99
19) n-Nitroso-di-n-propyla...	8.534	70	968278	76.844	ng		99
20) 3+4-Methylphenols	8.499	107	1492527	82.092	ng		99
22) Acetophenone	8.546	105	1977578	77.680	ng	#	100
24) Nitrobenzene	8.922	77	1392093	78.066	ng		98
25) Isophorone	9.446	82	2617853	80.596	ng		99
26) 2-Nitrophenol	9.628	139	760077	88.768	ng		98
27) 2,4-Dimethylphenol	9.693	122	905891	83.475	ng		99
28) bis(2-Chloroethoxy)met...	9.922	93	1689686	76.434	ng		99
29) 2,4-Dichlorophenol	10.157	162	1238253	83.719	ng		99
30) 1,2,4-Trichlorobenzene	10.369	180	1252476	78.272	ng		99
31) Naphthalene	10.557	128	4190486	77.287	ng		100
32) Benzoic acid	9.875	122	1154243m	98.425	ng		
33) 4-Chloroaniline	10.669	127	1547427	81.682	ng		99
34) Hexachlorobutadiene	10.840	225	743811	79.240	ng		99
35) Caprolactam	11.481	113	469910	86.534	ng		98
36) 4-Chloro-3-methylphenol	11.804	107	1438601	83.468	ng		98
37) 2-Methylnaphthalene	12.169	142	2918685	77.673	ng		99
38) 1-Methylnaphthalene	12.393	142	2831173	76.912	ng		100
40) 1,2,4,5-Tetrachloroben...	12.545	216	1416894	80.378	ng		100
41) Hexachlorocyclopentadiene	12.528	237	523136	84.532	ng		99
43) 2,4,6-Trichlorophenol	12.787	196	1000309	86.244	ng		97

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024282.D
Acq On : 14 Apr 2025 17:13
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Quant Time: Apr 14 18:14:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.869	196	1103073	85.819	ng	98
46) 1,1'-Biphenyl	13.198	154	3665607	77.381	ng	99
47) 2-Chloronaphthalene	13.234	162	2766302	78.257	ng	99
48) 2-Nitroaniline	13.445	65	848619	86.912	ng	98
49) Acenaphthylene	14.087	152	4438648	80.004	ng	100
50) Dimethylphthalate	13.834	163	3659579	78.201	ng	99
51) 2,6-Dinitrotoluene	13.945	165	814314	82.589	ng	98
52) Acenaphthene	14.439	154	2761536	78.270	ng	100
53) 3-Nitroaniline	14.281	138	919775	89.913	ng	99
54) 2,4-Dinitrophenol	14.492	184	541483	102.567	ng	99
55) Dibenzofuran	14.781	168	4420987	76.402	ng	99
56) 4-Nitrophenol	14.604	139	814261	92.599	ng	98
57) 2,4-Dinitrotoluene	14.745	165	1169554	87.755	ng	98
58) Fluorene	15.439	166	3479094	77.873	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.004	232	999335	84.031	ng	100
60) Diethylphthalate	15.222	149	3800651	78.910	ng	99
61) 4-Chlorophenyl-phenyle...	15.439	204	1689620	77.881	ng	100
62) 4-Nitroaniline	15.469	138	965926	89.063	ng	98
63) Azobenzene	15.739	77	3448680	77.741	ng	98
65) 4,6-Dinitro-2-methylph...	15.528	198	705506	94.385	ng	92
66) n-Nitrosodiphenylamine	15.657	169	2970516	77.438	ng	99
67) 4-Bromophenyl-phenylether	16.345	248	1137128	81.488	ng	99
68) Hexachlorobenzene	16.469	284	1347567	81.231	ng	97
70) Pentachlorophenol	16.828	266	1027691	90.221	ng	99
71) Phenanthrene	17.228	178	5415046	77.203	ng	100
72) Anthracene	17.316	178	5270987	78.097	ng	99
73) Carbazole	17.604	167	5275449	80.314	ng	100
74) Di-n-butylphthalate	18.192	149	6569078	80.266	ng	100
75) Fluoranthene	19.316	202	6520527	78.322	ng	99
77) Benzidine	19.510	184	1694694m	99.624	ng	
78) Pyrene	19.686	202	6718789	76.195	ng	100
80) Butylbenzylphthalate	20.645	149	3163633	85.104	ng	96
81) Benzo(a)anthracene	21.604	228	6729971	78.335	ng	99
82) 3,3'-Dichlorobenzidine	21.539	252	2536509	85.143	ng	99
83) Chrysene	21.680	228	6433173	78.130	ng	99
84) Bis(2-ethylhexyl)phtha...	21.545	149	4516949	82.506	ng	98
85) Di-n-octyl phthalate	22.833	149	7969779	90.878	ng	99
87) Indeno(1,2,3-cd)pyrene	28.815	276	9286186	85.280	ng	99
88) Benzo(b)fluoranthene	23.933	252	7677613	81.051	ng	98
89) Benzo(k)fluoranthene	23.998	252	7207214	78.395	ng	100
90) Benzo(a)pyrene	24.833	252	6703562	82.211	ng	99
91) Dibenzo(a,h)anthracene	28.886	278	7614901	84.071	ng	98
92) Benzo(g,h,i)perylene	29.997	276	7750941	84.072	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024282.D
Acq On : 14 Apr 2025 17:13
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Quant Time: Apr 14 18:14:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.728 min Scan# 7
Delta R.T. -0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:152 Resp: 244732

Ion	Ratio	Lower	Upper
152	100		
150	157.4	124.9	187.3
115	61.0	47.7	71.5







#2
1,4-Dioxane
Concen: 73.805 ng
RT: 3.293 min Scan# 35
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion: 88 Resp: 467234

Ion	Ratio	Lower	Upper
88	100		
58	64.8	51.4	77.2
43	21.7	17.1	25.7







#3
Pyridine
Concen: 81.864 ng m
RT: 3.687 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 79 Resp: 1349044

Ion	Ratio	Lower	Upper
79	100		
52	57.7	45.4	68.2
51	28.3	22.6	33.8







#4
n-Nitrosodimethylamine
Concen: 79.138 ng
RT: 3.593 min Scan# 86
Delta R.T. -0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion: 42 Resp: 434861

Ion	Ratio	Lower	Upper
42	100		
74	170.0	139.5	209.3
44	6.8	5.8	8.8







#5
2-Fluorophenol
Concen: 157.905 ng
RT: 5.352 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:112 Resp: 2341577

Ion	Ratio	Lower	Upper
112	100		
64	51.6	41.1	61.7
63	26.0	20.6	30.8







#6
Aniline
Concen: 73.789 ng
RT: 7.069 min Scan# 677
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion: 93 Resp: 1351337

Ion	Ratio	Lower	Upper
93	100		
66	36.1	29.1	43.7
65	18.2	14.7	22.1







#7
Phenol-d6
Concen: 158.927 ng
RT: 6.916 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 99 Resp: 3224074

Ion	Ratio	Lower	Upper
99	100		
42	13.9	10.6	16.0
71	31.1	23.7	35.5







#8
2-Chlorophenol
Concen: 78.433 ng
RT: 7.305 min Scan# 717
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:128 Resp: 1299769

Ion	Ratio	Lower	Upper
128	100		
130	32.6	12.9	52.9
64	42.1	22.7	62.7







#9
Benzaldehyde
Concen: 57.891 ng
RT: 6.881 min Scan# 64
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 77 Resp: 608272

Ion	Ratio	Lower	Upper
77	100		
105	101.2	79.2	119.2
106	96.2	76.2	116.2







#10
Phenol
Concen: 78.944 ng
RT: 6.946 min Scan# 656
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion: 94 Resp: 1608027

Ion	Ratio	Lower	Upper
94	100		
65	25.1	4.5	44.5
66	33.5	12.4	52.4







#11
bis(2-Chloroethyl)ether
Concen: 77.297 ng
RT: 7.163 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 93 Resp: 1272327

Ion	Ratio	Lower	Upper
93	100		
63	68.2	48.6	88.6
95	31.9	12.3	52.3







#12
1,3-Dichlorobenzene
Concen: 74.869 ng
RT: 7.622 min Scan# 771
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:146 Resp: 1344292

Ion	Ratio	Lower	Upper
146	100		
148	64.3	51.4	77.0
75	28.4	22.9	34.3







#13
1,4-Dichlorobenzene
Concen: 75.195 ng
RT: 7.763 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:146 Resp: 1369087

Ion	Ratio	Lower	Upper
146	100		
148	64.5	51.1	76.7
111	40.7	32.6	48.8







#14
1,2-Dichlorobenzene
Concen: 74.432 ng
RT: 8.075 min Scan# 848
Delta R.T. -0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:146 Resp: 1310407

Ion	Ratio	Lower	Upper
146	100		
148	63.7	51.0	76.4
111	42.7	33.8	50.8







#15
Benzyl Alcohol
Concen: 84.492 ng
RT: 7.975 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 79 Resp: 1098602

Ion	Ratio	Lower	Upper
79	100		
108	82.4	65.1	97.7
77	62.4	50.7	76.1







#16
2,2'-oxybis(1-Chloropropane)
Concen: 73.704 ng
RT: 8.257 min Scan# 879
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion: 45 Resp: 1319830

Ion	Ratio	Lower	Upper
45	100		
77	18.1	0.0	37.7
79	14.7	0.0	33.5







#17
2-Methylphenol
Concen: 81.101 ng
RT: 8.175 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:107 Resp: 1064555

Ion	Ratio	Lower	Upper
107	100		
108	111.5	89.8	134.6
77	43.0	34.2	51.4
79	42.3	33.8	50.8







#18
Hexachloroethane
Concen: 75.920 ng
RT: 8.799 min Scan# 971
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:117 Resp: 498897

Ion	Ratio	Lower	Upper
117	100		
119	98.3	77.8	116.6
201	96.5	75.9	113.9







#19
n-Nitroso-di-n-propylamine
Concen: 76.844 ng
RT: 8.534 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 70 Resp: 968278

Ion	Ratio	Lower	Upper
70	100		
42	46.3	37.3	55.9
101	11.0	8.2	12.4
130	24.9	18.6	28.0







#20
3+4-Methylphenols
Concen: 82.092 ng
RT: 8.499 min Scan# 920
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:107 Resp: 1492527

Ion	Ratio	Lower	Upper
107	100		
108	89.0	69.8	109.8
77	30.4	11.5	51.5
79	26.5	7.8	47.8







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.504 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:136 Resp: 1024316

Ion	Ratio	Lower	Upper
136	100		
137	10.7	9.2	13.8
54	6.9	6.0	9.0
68	6.2	5.5	8.3







#22
Acetophenone
Concen: 77.680 ng
RT: 8.546 min Scan# 928
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:105 Resp: 1977578

Ion	Ratio	Lower	Upper
105	100		
71	4.0	1.7	2.5#
51	21.2	16.9	25.3
120	23.3	18.6	28.0







#23
Nitrobenzene-d5
Concen: 158.236 ng
RT: 8.881 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 82 Resp: 2847003

Ion	Ratio	Lower	Upper
82	100		
128	47.0	36.5	54.7
54	42.3	33.4	50.2







#24
Nitrobenzene
Concen: 78.066 ng
RT: 8.922 min Scan# 992
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion: 77 Resp: 1392093

Ion	Ratio	Lower	Upper
77	100		
123	50.5	39.2	58.8
65	12.9	9.8	14.6







#25
Isophorone
Concen: 80.596 ng
RT: 9.446 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 82 Resp: 2617853

Ion	Ratio	Lower	Upper
82	100		
95	7.0	5.4	8.2
138	18.4	14.4	21.6







#26
2-Nitrophenol
Concen: 88.768 ng
RT: 9.628 min Scan# 1112
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:139 Resp: 760077

Ion	Ratio	Lower	Upper
139	100		
109	27.5	21.8	32.6
65	43.6	36.2	54.2







#27
2,4-Dimethylphenol
Concen: 83.475 ng
RT: 9.693 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:122 Resp: 905891

Ion	Ratio	Lower	Upper
122	100		
107	116.1	93.7	140.5
121	59.3	48.2	72.2







#28
bis(2-Chloroethoxy)methane
Concen: 76.434 ng
RT: 9.922 min Scan# 1162
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion: 93 Resp: 1689686

Ion	Ratio	Lower	Upper
93	100		
95	32.4	26.0	39.0
123	12.2	9.1	13.7







#29
2,4-Dichlorophenol
Concen: 83.719 ng
RT: 10.157 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:162 Resp: 1238253

Ion	Ratio	Lower	Upper
162	100		
164	64.3	43.3	83.3
98	37.2	16.6	56.6







#30
1,2,4-Trichlorobenzene
Concen: 78.272 ng
RT: 10.369 min Scan# 1238
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:180 Resp: 1252476

Ion	Ratio	Lower	Upper
180	100		
182	95.5	77.3	115.9
145	30.9	25.1	37.7







#31
Naphthalene
Concen: 77.287 ng
RT: 10.557 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:128 Resp: 4190486

Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.6	13.0
127	13.0	10.3	15.5







#32
Benzoic acid
Concen: 98.425 ng m
RT: 9.875 min Scan# 1154
Delta R.T. 0.047 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:122 Resp: 1154243

Ion	Ratio	Lower	Upper
122	100		
105	122.9	104.8	144.8
77	89.8	69.2	109.2







#33
4-Chloroaniline
Concen: 81.682 ng
RT: 10.669 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:127 Resp: 1547427

Ion	Ratio	Lower	Upper
127	100		
129	32.0	26.1	39.1
65	26.8	22.1	33.1
92	18.5	15.0	22.4







#34
Hexachlorobutadiene
Concen: 79.240 ng
RT: 10.840 min Scan# 1318
Delta R.T. -0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:225 Resp: 743811

Ion	Ratio	Lower	Upper
225	100		
223	63.0	49.5	74.3
227	64.3	50.9	76.3







#35
Caprolactam
Concen: 86.534 ng
RT: 11.481 min Scan# 14
Delta R.T. 0.029 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:113 Resp: 469910

Ion	Ratio	Lower	Upper
113	100		
55	131.9	115.9	155.9
56	99.8	81.1	121.1







#36
4-Chloro-3-methylphenol
Concen: 83.468 ng
RT: 11.804 min Scan# 1482
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:107 Resp: 1438601

Ion	Ratio	Lower	Upper
107	100		
144	26.4	20.5	30.7
142	81.0	63.4	95.2







#37
2-Methylnaphthalene
Concen: 77.673 ng
RT: 12.169 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:142 Resp: 2918685

Ion	Ratio	Lower	Upper
142	100		
141	88.4	70.3	105.5
115	34.9	26.9	40.3







#38
1-Methylnaphthalene
Concen: 76.912 ng
RT: 12.393 min Scan# 1582
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:142 Resp: 2831173

Ion	Ratio	Lower	Upper
142	100		
141	91.8	73.6	110.4
116	3.8	3.0	4.6







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.369 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:164 Resp: 641442

Ion	Ratio	Lower	Upper
164	100		
162	99.6	80.6	120.8
160	43.9	35.3	52.9







#40
1,2,4,5-Tetrachlorobenzene
Concen: 80.378 ng
RT: 12.545 min Scan# 1608
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:216 Resp: 1416894

Ion	Ratio	Lower	Upper
216	100		
214	78.4	62.6	93.8
179	21.5	17.2	25.8
108	18.6	15.0	22.6







#41
Hexachlorocyclopentadiene
Concen: 84.532 ng
RT: 12.528 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:237 Resp: 523136

Ion	Ratio	Lower	Upper
237	100		
235	62.9	43.9	83.9
272	13.7	0.0	33.6







#42
2,4,6-Tribromophenol
Concen: 171.633 ng
RT: 15.892 min Scan# 2177
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:330 Resp: 1507534

Ion	Ratio	Lower	Upper
330	100		
332	96.5	77.1	115.7
141	29.7	24.7	37.1







#43
2,4,6-Trichlorophenol
Concen: 86.244 ng
RT: 12.787 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:196 Resp: 1000309

Ion	Ratio	Lower	Upper
196	100		
198	96.5	74.3	111.5
200	31.2	24.0	36.0







#44
2,4,5-Trichlorophenol
Concen: 85.819 ng
RT: 12.869 min Scan# 1663
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:196 Resp: 1103073

Ion	Ratio	Lower	Upper
196	100		
198	95.0	77.3	115.9
97	47.7	40.4	60.6
132	28.6	23.4	35.0







#45
2-Fluorobiphenyl
Concen: 150.596 ng
RT: 12.981 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:172 Resp: 6362479

Ion	Ratio	Lower	Upper
172	100		
171	35.8	28.2	42.2
170	23.8	18.6	27.8







#46
1,1'-Biphenyl
Concen: 77.381 ng
RT: 13.198 min Scan# 1719
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:154 Resp: 3665607

Ion	Ratio	Lower	Upper
154	100		
153	40.7	20.3	60.3
76	13.2	0.0	33.5







#47
2-Chloronaphthalene
Concen: 78.257 ng
RT: 13.234 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:162 Resp: 2766302

Ion	Ratio	Lower	Upper
162	100		
127	40.9	31.9	47.9
164	32.8	26.2	39.4







#48
2-Nitroaniline
Concen: 86.912 ng
RT: 13.445 min Scan# 1761
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion: 65 Resp: 848619

Ion	Ratio	Lower	Upper
65	100		
92	76.5	62.3	93.5
138	123.5	97.2	145.8







#49
Acenaphthylene
Concen: 80.004 ng
RT: 14.087 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:152 Resp: 4438648

Ion	Ratio	Lower	Upper
152	100		
151	20.2	15.9	23.9
153	13.2	10.6	15.8







#50
Dimethylphthalate
Concen: 78.201 ng
RT: 13.834 min Scan# 1827
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:163 Resp: 3659579

Ion	Ratio	Lower	Upper
163	100		
194	4.4	3.5	5.3
164	10.3	8.0	12.0







#51
2,6-Dinitrotoluene
Concen: 82.589 ng
RT: 13.945 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:165 Resp: 814314

Ion	Ratio	Lower	Upper
165	100		
63	46.4	38.0	57.0
89	51.0	41.7	62.5







#52
Acenaphthene
Concen: 78.270 ng
RT: 14.439 min Scan# 1930
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:154 Resp: 2761536

Ion	Ratio	Lower	Upper
154	100		
153	112.7	89.8	134.6
152	53.5	42.6	63.8







#53
3-Nitroaniline
Concen: 89.913 ng
RT: 14.281 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:138 Resp: 919775

Ion	Ratio	Lower	Upper
138	100		
108	10.5	8.3	12.5
92	111.0	89.8	134.8







#54
2,4-Dinitrophenol
Concen: 102.567 ng
RT: 14.492 min Scan# 1939
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:184 Resp: 541483

Ion	Ratio	Lower	Upper
184	100		
63	53.6	42.6	64.0
154	58.2	47.2	70.8







#55
Dibenzofuran
Concen: 76.402 ng
RT: 14.781 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:168 Resp: 4420987

Ion	Ratio	Lower	Upper
168	100		
139	38.7	30.3	45.5
169	13.3	10.6	16.0







#56
4-Nitrophenol
Concen: 92.599 ng
RT: 14.604 min Scan# 1958
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:139 Resp: 814261

Ion	Ratio	Lower	Upper
139	100		
109	68.4	46.1	86.1
65	86.6	65.0	105.0







#57
2,4-Dinitrotoluene
Concen: 87.755 ng
RT: 14.745 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:165 Resp: 1169554

Ion	Ratio	Lower	Upper
165	100		
63	36.2	30.4	45.6
89	66.4	54.6	82.0
182	2.9	2.2	3.2







#58
Fluorene
Concen: 77.873 ng
RT: 15.439 min Scan# 2100
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:166 Resp: 3479094

Ion	Ratio	Lower	Upper
166	100		
165	96.6	78.1	117.1
167	13.7	10.9	16.3







#59
2,3,4,6-Tetrachlorophenol
Concen: 84.031 ng
RT: 15.004 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:232 Resp: 999335

Ion	Ratio	Lower	Upper
232	100		
131	46.1	36.8	55.2
130	2.4	1.9	2.9
166	31.4	25.0	37.6







#60
Diethylphthalate
Concen: 78.910 ng
RT: 15.222 min Scan# 2063
Delta R.T. 0.018 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:149 Resp: 3800651

Ion	Ratio	Lower	Upper
149	100		
177	22.6	17.4	26.2
150	12.6	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 77.881 ng
RT: 15.439 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:204 Resp: 1689620

Ion	Ratio	Lower	Upper
204	100		
206	33.3	26.5	39.7
141	61.4	49.4	74.0







#62
4-Nitroaniline
Concen: 89.063 ng
RT: 15.469 min Scan# 2105
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:138 Resp: 965926

Ion	Ratio	Lower	Upper
138	100		
92	48.5	27.8	67.8
108	71.7	48.9	88.9







#63
Azobenzene
Concen: 77.741 ng
RT: 15.739 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 77 Resp: 3448680

Ion	Ratio	Lower	Upper
77	100		
182	29.0	7.3	47.3
105	22.7	2.3	42.3
51	24.5	3.4	43.4







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.180 min Scan# 2396
Delta R.T. 0.018 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:188 Resp: 1279313

Ion	Ratio	Lower	Upper
188	100		
94	10.4	8.6	13.0
80	11.2	9.8	14.6







#65
4,6-Dinitro-2-methylphenol
Concen: 94.385 ng
RT: 15.528 min Scan# 21
Delta R.T. 0.018 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:198 Resp: 705506

Ion	Ratio	Lower	Upper
198	100		
51	34.3	18.4	58.4
105	41.0	26.5	66.5







#66
n-Nitrosodiphenylamine
Concen: 77.438 ng
RT: 15.657 min Scan# 2137
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:169 Resp: 2970516

Ion	Ratio	Lower	Upper
169	100		
168	66.7	53.0	79.4
167	36.3	28.9	43.3







#67
4-Bromophenyl-phenylether
Concen: 81.488 ng
RT: 16.345 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:248 Resp: 1137128

Ion	Ratio	Lower	Upper
248	100		
250	98.5	78.2	117.4
141	73.2	57.7	86.5







#68
Hexachlorobenzene
Concen: 81.231 ng
RT: 16.469 min Scan# 2275
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:284 Resp: 1347567

Ion	Ratio	Lower	Upper
284	100		
142	33.3	24.6	37.0
249	30.5	23.8	35.6







#70
Pentachlorophenol
Concen: 90.221 ng
RT: 16.828 min Scan# 2
Delta R.T. 0.017 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:266 Resp: 1027691

Ion	Ratio	Lower	Upper
266	100		
268	64.0	51.0	76.6
264	63.1	49.8	74.6







#71
Phenanthrene
Concen: 77.203 ng
RT: 17.228 min Scan# 2404
Delta R.T. 0.018 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:178 Resp: 5415046

Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.4	23.0
179	15.6	12.5	18.7







#72
Anthracene
Concen: 78.097 ng
RT: 17.316 min Scan# 24
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:178 Resp: 5270987

Ion	Ratio	Lower	Upper
178	100		
176	19.0	14.7	22.1
179	15.6	12.4	18.6







#73
Carbazole
Concen: 80.314 ng
RT: 17.604 min Scan# 2468
Delta R.T. 0.017 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:167 Resp: 5275449

Ion	Ratio	Lower	Upper
167	100		
166	21.7	17.1	25.7
139	13.2	10.6	16.0







#74
Di-n-butylphthalate
Concen: 80.266 ng
RT: 18.192 min Scan# 2
Delta R.T. 0.017 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:149 Resp: 6569078

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.4	11.2
104	4.9	3.9	5.9







#75
Fluoranthene
Concen: 78.322 ng
RT: 19.316 min Scan# 2759
Delta R.T. 0.017 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:202 Resp: 6520527

Ion	Ratio	Lower	Upper
202	100		
101	12.6	0.0	33.2
203	17.9	0.0	37.7







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.621 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:240 Resp: 1378105

Ion	Ratio	Lower	Upper
240	100		
120	11.8	9.8	14.8
236	25.5	20.9	31.3







#77
Benzidine
Concen: 99.624 ng m
RT: 19.510 min Scan# 2792
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:184 Resp: 1694694

Ion	Ratio	Lower	Upper
184	100		
185	14.3	11.6	17.4
183	11.8	9.2	13.8







#78
Pyrene
Concen: 76.195 ng
RT: 19.686 min Scan# 2
Delta R.T. 0.018 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:202 Resp: 6718789

Ion	Ratio	Lower	Upper
202	100		
200	21.8	17.2	25.8
203	18.1	14.4	21.6







#79
Terphenyl-d14
Concen: 147.678 ng
RT: 19.922 min Scan# 2862
Delta R.T. 0.023 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:244 Resp:10209149

Ion	Ratio	Lower	Upper
244	100		
212	7.5	5.6	8.4
122	10.4	9.4	14.0







#80
Butylbenzylphthalate
Concen: 85.104 ng
RT: 20.645 min Scan# 2
Delta R.T. 0.018 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:149 Resp: 3163633

Ion	Ratio	Lower	Upper
149	100		
91	67.1	56.4	84.6
206	22.0	16.3	24.5







#81
Benzo(a)anthracene
Concen: 78.335 ng
RT: 21.604 min Scan# 3148
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:228 Resp: 6729971

Ion	Ratio	Lower	Upper
228	100		
226	27.8	21.8	32.8
229	20.1	15.9	23.9







#82
3,3'-Dichlorobenzidine
Concen: 85.143 ng
RT: 21.539 min Scan# 3
Delta R.T. 0.018 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:252 Resp: 2536509

Ion	Ratio	Lower	Upper
252	100		
254	65.1	51.7	77.5
126	10.0	8.8	13.2







#83
Chrysene
Concen: 78.130 ng
RT: 21.680 min Scan# 3161
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:228 Resp: 6433173

Ion	Ratio	Lower	Upper
228	100		
226	30.5	24.2	36.2
229	20.0	15.6	23.4







#84
Bis(2-ethylhexyl)phthalate
Concen: 82.506 ng
RT: 21.545 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:149 Resp: 4516949

Ion	Ratio	Lower	Upper
149	100		
167	28.8	22.4	33.6
279	5.4	4.0	6.0







#85
Di-n-octyl phthalate
Concen: 90.878 ng
RT: 22.833 min Scan# 3357
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:149 Resp: 7969779

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	7.3	6.0	9.0







#86
Perylene-d12
Concen: 20.000 ng
RT: 24.986 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:264 Resp: 1583281

Ion	Ratio	Lower	Upper
264	100		
260	23.8	18.9	28.3
265	22.3	17.8	26.6







#87
Indeno(1,2,3-cd)pyrene
Concen: 85.280 ng
RT: 28.815 min Scan# 4374
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:276 Resp: 9286186

Ion	Ratio	Lower	Upper
276	100		
138	17.4	14.5	21.7
277	25.4	20.2	30.2







#88
Benzo(b)fluoranthene
Concen: 81.051 ng
RT: 23.933 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:252 Resp: 7677613

Ion	Ratio	Lower	Upper
252	100		
253	22.5	17.7	26.5
125	10.4	9.4	14.0







#89
Benzo(k)fluoranthene
Concen: 78.395 ng
RT: 23.998 min Scan# 3555
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:252 Resp: 7207214

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	11.1	9.1	13.7







#90
Benzo(a)pyrene
Concen: 82.211 ng
RT: 24.833 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:252 Resp: 6703562

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.6	26.4
125	11.8	10.1	15.1







#91
Dibenzo(a,h)anthracene
Concen: 84.071 ng
RT: 28.886 min Scan# 4386
Delta R.T. -0.018 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:278 Resp: 7614901

Ion	Ratio	Lower	Upper
278	100		
139	16.2	12.6	19.0
279	24.2	18.5	27.7






#92
Benzo(g,h,i)perylene
Concen: 84.072 ng
RT: 29.997 min Scan# 4
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:276 Resp: 7750941

Ion	Ratio	Lower	Upper
276	100		
277	23.8	18.9	28.3
138	22.4	18.7	28.1



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024283.D
Acq On : 14 Apr 2025 18:35
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Quant Time: Apr 15 04:54:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 15 04:48:42 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.734	152	232241	20.000	ng	0.00
21) Naphthalene-d8	10.504	136	956448	20.000	ng	0.00
39) Acenaphthene-d10	14.363	164	600031	20.000	ng	0.00
64) Phenanthrene-d10	17.163	188	1178916	20.000	ng	-0.02
76) Chrysene-d12	21.616	240	1288125	20.000	ng	0.00
86) Perylene-d12	24.986	264	1476872	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.346	112	1141956	81.302	ng	0.00
7) Phenol-d6	6.911	99	1559050	81.063	ng	0.00
23) Nitrobenzene-d5	8.881	82	1401115	83.531	ng	0.00
42) 2,4,6-Tribromophenol	15.875	330	709703	85.488	ng	-0.02
45) 2-Fluorobiphenyl	12.975	172	3319933	84.715	ng	0.00
79) Terphenyl-d14	19.892	244	5217888	81.649	ng	-0.03
Target Compounds						Qvalue
2) 1,4-Dioxane	3.293	88	240091	40.412	ng	100
3) Pyridine	3.687	79	626813	39.955	ng	99
4) n-Nitrosodimethylamine	3.593	42	209326	40.205	ng	98
6) Aniline	7.069	93	710981	41.370	ng	100
8) 2-Chlorophenol	7.305	128	632515	40.334	ng	99
9) Benzaldehyde	6.881	77	399803	42.024	ng	98
10) Phenol	6.940	94	777631	40.306	ng	100
11) bis(2-Chloroethyl)ether	7.163	93	620261	39.902	ng	99
12) 1,3-Dichlorobenzene	7.622	146	680343	40.298	ng	100
13) 1,4-Dichlorobenzene	7.769	146	694020	40.516	ng	99
14) 1,2-Dichlorobenzene	8.081	146	659947	39.898	ng	99
15) Benzyl Alcohol	7.969	79	517678	41.621	ng	100
16) 2,2'-oxybis(1-Chloropr...	8.258	45	674999	40.173	ng	99
17) 2-Methylphenol	8.175	107	512292	41.046	ng	100
18) Hexachloroethane	8.799	117	250852	40.522	ng	99
19) n-Nitroso-di-n-propyla...	8.528	70	481540	40.471	ng	99
20) 3+4-Methylphenols	8.499	107	706733	40.810	ng	99
22) Acetophenone	8.546	105	986019	41.652	ng	100
24) Nitrobenzene	8.922	77	689231	41.537	ng	100
25) Isophorone	9.446	82	1264480	41.648	ng	99
26) 2-Nitrophenol	9.628	139	352229	43.376	ng	99
27) 2,4-Dimethylphenol	9.687	122	425062	41.689	ng	99
28) bis(2-Chloroethoxy)met...	9.922	93	849161	41.402	ng	99
29) 2,4-Dichlorophenol	10.163	162	590306	42.461	ng	98
30) 1,2,4-Trichlorobenzene	10.369	180	619818	41.612	ng	99
31) Naphthalene	10.557	128	2085090	41.386	ng	99
32) Benzoic acid	9.834	122	486627	40.959	ng	98
33) 4-Chloroaniline	10.669	127	761817	42.937	ng	99
34) Hexachlorobutadiene	10.846	225	367684	42.007	ng	100
35) Caprolactam	11.457	113	216229	42.152	ng	99
36) 4-Chloro-3-methylphenol	11.799	107	675857	41.737	ng	98
37) 2-Methylnaphthalene	12.169	142	1446647	41.403	ng	100
38) 1-Methylnaphthalene	12.387	142	1408633	41.210	ng	99
40) 1,2,4,5-Tetrachloroben...	12.540	216	697580	42.275	ng	100
41) Hexachlorocyclopentadiene	12.516	237	254676	43.639	ng	99
43) 2,4,6-Trichlorophenol	12.781	196	475349	43.329	ng	97

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024283.D
Acq On : 14 Apr 2025 18:35
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Quant Time: Apr 15 04:54:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 15 04:48:42 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.863	196	520298	42.828	ng	97
46) 1,1'-Biphenyl	13.187	154	1834570	41.595	ng	99
47) 2-Chloronaphthalene	13.228	162	1377424	41.786	ng	100
48) 2-Nitroaniline	13.434	65	392742	42.475	ng	98
49) Acenaphthylene	14.081	152	2168474	41.783	ng	99
50) Dimethylphthalate	13.816	163	1809923	41.478	ng	99
51) 2,6-Dinitrotoluene	13.940	165	398345	42.990	ng	97
52) Acenaphthene	14.428	154	1344151	40.853	ng	99
53) 3-Nitroaniline	14.275	138	427828	43.931	ng	98
54) 2,4-Dinitrophenol	14.481	184	231882	42.485	ng	98
55) Dibenzofuran	14.763	168	2187786	40.679	ng	99
56) 4-Nitrophenol	14.592	139	378214	44.968	ng	99
57) 2,4-Dinitrotoluene	14.734	165	550175	43.527	ng	97
58) Fluorene	15.428	166	1738215	41.750	ng	99
59) 2,3,4,6-Tetrachlorophenol	14.998	232	470547	41.995	ng	99
60) Diethylphthalate	15.198	149	1869274	41.570	ng	99
61) 4-Chlorophenyl-phenyle...	15.422	204	844629	41.777	ng	99
62) 4-Nitroaniline	15.451	138	444482	43.114	ng	96
63) Azobenzene	15.722	77	1722900	41.686	ng	98
65) 4,6-Dinitro-2-methylph...	15.510	198	317548	42.724	ng	95
66) n-Nitrosodiphenylamine	15.645	169	1460282	41.500	ng	100
67) 4-Bromophenyl-phenylether	16.334	248	545579	42.314	ng	97
68) Hexachlorobenzene	16.451	284	642095	41.909	ng	96
69) Atrazine	16.622	200	282908	31.567	ng	98
70) Pentachlorophenol	16.810	266	458529	42.900	ng	99
71) Phenanthrene	17.204	178	2612052	40.615	ng	100
72) Anthracene	17.298	178	2606508	42.051	ng	100
73) Carbazole	17.581	167	2564955	42.351	ng	99
74) Di-n-butylphthalate	18.175	149	3360635	44.539	ng	100
75) Fluoranthene	19.298	202	3195045	41.771	ng	99
77) Benzidine	19.492	184	588487	36.041	ng	100
78) Pyrene	19.675	202	3276290	40.022	ng	99
80) Butylbenzylphthalate	20.627	149	1511473	43.107	ng	97
81) Benzo(a)anthracene	21.598	228	3299250	41.207	ng	99
82) 3,3'-Dichlorobenzidine	21.516	252	1212290	43.139	ng	99
83) Chrysene	21.669	228	3136332	40.888	ng	100
84) Bis(2-ethylhexyl)phtha...	21.539	149	2282222	44.400	ng	100
85) Di-n-octyl phthalate	22.827	149	3803266	45.513	ng	100
87) Indeno(1,2,3-cd)pyrene	28.798	276	4313548m	42.071	ng	
88) Benzo(b)fluoranthene	23.933	252	3654984	41.288	ng	99
89) Benzo(k)fluoranthene	23.998	252	3512834	41.081	ng	100
90) Benzo(a)pyrene	24.833	252	3196130	41.855	ng	99
91) Dibenzo(a,h)anthracene	28.880	278	3566000	41.902	ng	98
92) Benzo(g,h,i)perylene	29.992	276	3590938	41.455	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed


ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024283.D
Acq On : 14 Apr 2025 18:35
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Quant Time: Apr 15 04:54:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 15 04:48:42 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.734 min Scan# 7
Delta R.T. 0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:152 Resp: 232241

Ion	Ratio	Lower	Upper
152	100		
150	156.8	124.9	187.3
115	58.6	47.7	71.5







#2
1,4-Dioxane
Concen: 40.412 ng
RT: 3.293 min Scan# 35
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion: 88 Resp: 240091

Ion	Ratio	Lower	Upper
88	100		
58	64.4	51.4	77.2
43	21.7	17.1	25.7







#3
Pyridine
Concen: 39.955 ng
RT: 3.687 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion: 79 Resp: 626813

Ion	Ratio	Lower	Upper
79	100		
52	58.1	45.4	68.2
51	28.0	22.6	33.8







#4
n-Nitrosodimethylamine
Concen: 40.205 ng
RT: 3.593 min Scan# 86
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion: 42 Resp: 209326

Ion	Ratio	Lower	Upper
42	100		
74	171.4	139.5	209.3
44	7.3	5.8	8.8







#5
2-Fluorophenol
Concen: 81.302 ng
RT: 5.346 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:112 Resp: 1141956

Ion	Ratio	Lower	Upper
112	100		
64	53.2	41.1	61.7
63	26.8	20.6	30.8







#6
Aniline
Concen: 41.370 ng
RT: 7.069 min Scan# 677
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion: 93 Resp: 710981

Ion	Ratio	Lower	Upper
93	100		
66	36.2	29.1	43.7
65	18.2	14.7	22.1







#7
Phenol-d6
Concen: 81.063 ng
RT: 6.911 min Scan# 6
Delta R.T. -0.005 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion: 99 Resp: 1559050

Ion	Ratio	Lower	Upper
99	100		
42	13.6	10.6	16.0
71	30.9	23.7	35.5







#8
2-Chlorophenol
Concen: 40.334 ng
RT: 7.305 min Scan# 717
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:128 Resp: 632515

Ion	Ratio	Lower	Upper
128	100		
130	32.3	12.9	52.9
64	41.9	22.7	62.7







#9
Benzaldehyde
Concen: 42.024 ng
RT: 6.881 min Scan# 64
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion: 77 Resp: 399803

Ion	Ratio	Lower	Upper
77	100		
105	101.7	79.2	119.2
106	98.0	76.2	116.2







#10
Phenol
Concen: 40.306 ng
RT: 6.940 min Scan# 655
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion: 94 Resp: 777631

Ion	Ratio	Lower	Upper
94	100		
65	24.7	4.5	44.5
66	32.5	12.4	52.4







#11
bis(2-Chloroethyl)ether
Concen: 39.902 ng
RT: 7.163 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion: 93 Resp: 620261

Ion	Ratio	Lower	Upper
93	100		
63	69.2	48.6	88.6
95	32.0	12.3	52.3







#12
1,3-Dichlorobenzene
Concen: 40.298 ng
RT: 7.622 min Scan# 771
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:146 Resp: 680343

Ion	Ratio	Lower	Upper
146	100		
148	64.5	51.4	77.0
75	28.2	22.9	34.3







#13
1,4-Dichlorobenzene
Concen: 40.516 ng
RT: 7.769 min Scan# 7
Delta R.T. 0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:146 Resp: 694020

Ion	Ratio	Lower	Upper
146	100		
148	63.3	51.1	76.7
111	39.9	32.6	48.8







#14
1,2-Dichlorobenzene
Concen: 39.898 ng
RT: 8.081 min Scan# 849
Delta R.T. 0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:146 Resp: 659947

Ion	Ratio	Lower	Upper
146	100		
148	64.4	51.0	76.4
111	43.3	33.8	50.8







#15
Benzyl Alcohol
Concen: 41.621 ng
RT: 7.969 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion: 79 Resp: 517678

Ion	Ratio	Lower	Upper
79	100		
108	81.4	65.1	97.7
77	63.1	50.7	76.1







#16
2,2'-oxybis(1-Chloropropane)
Concen: 40.173 ng
RT: 8.258 min Scan# 879
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion: 45 Resp: 674999

Ion	Ratio	Lower	Upper
45	100		
77	17.8	0.0	37.7
79	13.9	0.0	33.5







#17
2-Methylphenol
Concen: 41.046 ng
RT: 8.175 min Scan# 8
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:107 Resp: 512292

Ion	Ratio	Lower	Upper
107	100		
108	111.8	89.8	134.6
77	43.2	34.2	51.4
79	42.5	33.8	50.8







#18
Hexachloroethane
Concen: 40.522 ng
RT: 8.799 min Scan# 971
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:117 Resp: 250852

Ion	Ratio	Lower	Upper
117	100		
119	95.2	77.8	116.6
201	95.1	75.9	113.9







#19
n-Nitroso-di-n-propylamine
Concen: 40.471 ng
RT: 8.528 min Scan# 9
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion: 70 Resp: 481540

Ion	Ratio	Lower	Upper
70	100		
42	47.1	37.3	55.9
101	10.4	8.2	12.4
130	24.3	18.6	28.0







#20
3+4-Methylphenols
Concen: 40.810 ng
RT: 8.499 min Scan# 920
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:107 Resp: 706733

Ion	Ratio	Lower	Upper
107	100		
108	90.2	69.8	109.8
77	30.7	11.5	51.5
79	27.6	7.8	47.8







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.504 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:136 Resp: 956448

Ion	Ratio	Lower	Upper
136	100		
137	10.9	9.2	13.8
54	6.8	6.0	9.0
68	6.6	5.5	8.3







#22
Acetophenone
Concen: 41.652 ng
RT: 8.546 min Scan# 928
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:105 Resp: 986019

Ion	Ratio	Lower	Upper
105	100		
71	1.8	1.7	2.5
51	20.8	16.9	25.3
120	23.2	18.6	28.0







#23
Nitrobenzene-d5
Concen: 83.531 ng
RT: 8.881 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion: 82 Resp: 1401115

Ion	Ratio	Lower	Upper
82	100		
128	47.0	36.5	54.7
54	42.5	33.4	50.2







#24
Nitrobenzene
Concen: 41.537 ng
RT: 8.922 min Scan# 992
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion: 77 Resp: 689231

Ion	Ratio	Lower	Upper
77	100		
123	48.7	39.2	58.8
65	12.2	9.8	14.6







#25
Isophorone
Concen: 41.648 ng
RT: 9.446 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion: 82 Resp: 1264480

Ion	Ratio	Lower	Upper
82	100		
95	6.9	5.4	8.2
138	18.8	14.4	21.6







#26
2-Nitrophenol
Concen: 43.376 ng
RT: 9.628 min Scan# 1112
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:139 Resp: 352229

Ion	Ratio	Lower	Upper
139	100		
109	28.2	21.8	32.6
65	45.3	36.2	54.2







#27
2,4-Dimethylphenol
Concen: 41.689 ng
RT: 9.687 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:122 Resp: 425062

Ion	Ratio	Lower	Upper
122	100		
107	116.0	93.7	140.5
121	59.3	48.2	72.2







#28
bis(2-Chloroethoxy)methane
Concen: 41.402 ng
RT: 9.922 min Scan# 1162
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion: 93 Resp: 849161

Ion	Ratio	Lower	Upper
93	100		
95	32.3	26.0	39.0
123	11.7	9.1	13.7







#29
2,4-Dichlorophenol
Concen: 42.461 ng
RT: 10.163 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:162 Resp: 590306

Ion	Ratio	Lower	Upper
162	100		
164	64.7	43.3	83.3
98	37.4	16.6	56.6







#30
1,2,4-Trichlorobenzene
Concen: 41.612 ng
RT: 10.369 min Scan# 1238
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:180 Resp: 619818

Ion	Ratio	Lower	Upper
180	100		
182	95.8	77.3	115.9
145	31.6	25.1	37.7







#31
Naphthalene
Concen: 41.386 ng
RT: 10.557 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:128 Resp: 2085090

Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.6	13.0
127	13.2	10.3	15.5







#32
Benzoic acid
Concen: 40.959 ng
RT: 9.834 min Scan# 1147
Delta R.T. -0.041 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:122 Resp: 486627

Ion	Ratio	Lower	Upper
122	100		
105	122.5	104.8	144.8
77	90.6	69.2	109.2







#33
4-Chloroaniline
Concen: 42.937 ng
RT: 10.669 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:127 Resp: 761817

Ion	Ratio	Lower	Upper
127	100		
129	32.1	26.1	39.1
65	27.1	22.1	33.1
92	18.4	15.0	22.4







#34
Hexachlorobutadiene
Concen: 42.007 ng
RT: 10.846 min Scan# 1319
Delta R.T. 0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:225 Resp: 367684

Ion	Ratio	Lower	Upper
225	100		
223	61.9	49.5	74.3
227	63.4	50.9	76.3







#35
Caprolactam
Concen: 42.152 ng
RT: 11.457 min Scan# 14
Delta R.T. -0.023 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:113 Resp: 216229

Ion	Ratio	Lower	Upper
113	100		
55	136.8	115.9	155.9
56	102.3	81.1	121.1







#36
4-Chloro-3-methylphenol
Concen: 41.737 ng
RT: 11.799 min Scan# 1481
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:107 Resp: 675857

Ion	Ratio	Lower	Upper
107	100		
144	26.2	20.5	30.7
142	81.6	63.4	95.2







#37
2-Methylnaphthalene
Concen: 41.403 ng
RT: 12.169 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:142 Resp: 1446647

Ion	Ratio	Lower	Upper
142	100		
141	88.1	70.3	105.5
115	34.4	26.9	40.3







#38
1-Methylnaphthalene
Concen: 41.210 ng
RT: 12.387 min Scan# 1581
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:142 Resp: 1408633

Ion	Ratio	Lower	Upper
142	100		
141	92.7	73.6	110.4
116	3.7	3.0	4.6







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.363 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:164 Resp: 600031

Ion	Ratio	Lower	Upper
164	100		
162	98.5	80.6	120.8
160	43.2	35.3	52.9







#40
1,2,4,5-Tetrachlorobenzene
Concen: 42.275 ng
RT: 12.540 min Scan# 1607
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:216 Resp: 697580

Ion	Ratio	Lower	Upper
216	100		
214	78.6	62.6	93.8
179	21.5	17.2	25.8
108	18.8	15.0	22.6







#41
Hexachlorocyclopentadiene
Concen: 43.639 ng
RT: 12.516 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:237 Resp: 254676

Ion	Ratio	Lower	Upper
237	100		
235	63.3	43.9	83.9
272	13.9	0.0	33.6







#42
2,4,6-Tribromophenol
Concen: 85.488 ng
RT: 15.875 min Scan# 2174
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:330 Resp: 709703

Ion	Ratio	Lower	Upper
330	100		
332	96.7	77.1	115.7
141	30.7	24.7	37.1







#43
2,4,6-Trichlorophenol
Concen: 43.329 ng
RT: 12.781 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:196 Resp: 475349

Ion	Ratio	Lower	Upper
196	100		
198	96.3	74.3	111.5
200	31.4	24.0	36.0







#44
2,4,5-Trichlorophenol
Concen: 42.828 ng
RT: 12.863 min Scan# 1662
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:196 Resp: 520298

Ion	Ratio	Lower	Upper
196	100		
198	95.0	77.3	115.9
97	47.4	40.4	60.6
132	28.1	23.4	35.0







#45
2-Fluorobiphenyl
Concen: 84.715 ng
RT: 12.975 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:172 Resp: 3319933

Ion	Ratio	Lower	Upper
172	100		
171	34.9	28.2	42.2
170	23.3	18.6	27.8







#46
1,1'-Biphenyl
Concen: 41.595 ng
RT: 13.187 min Scan# 1717
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:154 Resp: 1834570

Ion	Ratio	Lower	Upper
154	100		
153	39.9	20.3	60.3
76	14.0	0.0	33.5







#47
2-Chloronaphthalene
Concen: 41.786 ng
RT: 13.228 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:162 Resp: 1377424

Ion	Ratio	Lower	Upper
162	100		
127	39.9	31.9	47.9
164	33.0	26.2	39.4







#48
2-Nitroaniline
Concen: 42.475 ng
RT: 13.434 min Scan# 1759
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion: 65 Resp: 392742

Ion	Ratio	Lower	Upper
65	100		
92	77.0	62.3	93.5
138	123.6	97.2	145.8







#49
Acenaphthylene
Concen: 41.783 ng
RT: 14.081 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:152 Resp: 2168474

Ion	Ratio	Lower	Upper
152	100		
151	20.1	15.9	23.9
153	12.9	10.6	15.8







#50
Dimethylphthalate
Concen: 41.478 ng
RT: 13.816 min Scan# 1824
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:163 Resp: 1809923

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.5	5.3
164	10.2	8.0	12.0







#51
2,6-Dinitrotoluene
Concen: 42.990 ng
RT: 13.940 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:165 Resp: 398345

Ion	Ratio	Lower	Upper
165	100		
63	45.1	38.0	57.0
89	50.5	41.7	62.5







#52
Acenaphthene
Concen: 40.853 ng
RT: 14.428 min Scan# 1928
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:154 Resp: 1344151

Ion	Ratio	Lower	Upper
154	100		
153	112.6	89.8	134.6
152	53.9	42.6	63.8







#53
3-Nitroaniline
Concen: 43.931 ng
RT: 14.275 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:138 Resp: 427828

Ion	Ratio	Lower	Upper
138	100		
108	10.1	8.3	12.5
92	110.2	89.8	134.8







#54
2,4-Dinitrophenol
Concen: 42.485 ng
RT: 14.481 min Scan# 1937
Delta R.T. -0.011 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:184 Resp: 231882

Ion	Ratio	Lower	Upper
184	100		
63	56.3	42.6	64.0
154	58.5	47.2	70.8







#55
Dibenzofuran
Concen: 40.679 ng
RT: 14.763 min Scan# 1
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:168 Resp: 2187786

Ion	Ratio	Lower	Upper
168	100		
139	38.6	30.3	45.5
169	13.3	10.6	16.0







#56
4-Nitrophenol
Concen: 44.968 ng
RT: 14.592 min Scan# 1956
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:139 Resp: 378214

Ion	Ratio	Lower	Upper
139	100		
109	66.9	46.1	86.1
65	86.3	65.0	105.0







#57
2,4-Dinitrotoluene
Concen: 43.527 ng
RT: 14.734 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:165 Resp: 550175

Ion	Ratio	Lower	Upper
165	100		
63	36.1	30.4	45.6
89	65.2	54.6	82.0
182	3.0	2.2	3.2







#58
Fluorene
Concen: 41.750 ng
RT: 15.428 min Scan# 2098
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:166 Resp: 1738215

Ion	Ratio	Lower	Upper
166	100		
165	97.0	78.1	117.1
167	13.2	10.9	16.3







#59
2,3,4,6-Tetrachlorophenol
Concen: 41.995 ng
RT: 14.998 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:232 Resp: 470547

Ion	Ratio	Lower	Upper
232	100		
131	46.8	36.8	55.2
130	2.3	1.9	2.9
166	31.3	25.0	37.6







#60
Diethylphthalate
Concen: 41.570 ng
RT: 15.198 min Scan# 2059
Delta R.T. -0.023 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:149 Resp: 1869274

Ion	Ratio	Lower	Upper
149	100		
177	21.5	17.4	26.2
150	12.8	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 41.777 ng
RT: 15.422 min Scan# 2
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:204 Resp: 844629

Ion	Ratio	Lower	Upper
204	100		
206	32.7	26.5	39.7
141	62.6	49.4	74.0







#62
4-Nitroaniline
Concen: 43.114 ng
RT: 15.451 min Scan# 2102
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:138 Resp: 444482

Ion	Ratio	Lower	Upper
138	100		
92	49.9	27.8	67.8
108	72.1	48.9	88.9







#63
Azobenzene
Concen: 41.686 ng
RT: 15.722 min Scan# 21
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion: 77 Resp: 1722900

Ion	Ratio	Lower	Upper
77	100		
182	25.1	7.3	47.3
105	22.1	2.3	42.3
51	24.1	3.4	43.4







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.163 min Scan# 2393
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:188 Resp: 1178916

Ion	Ratio	Lower	Upper
188	100		
94	9.8	8.6	13.0
80	11.0	9.8	14.6







#65
4,6-Dinitro-2-methylphenol
Concen: 42.724 ng
RT: 15.510 min Scan# 21
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:198 Resp: 317548

Ion	Ratio	Lower	Upper
198	100		
51	35.0	18.4	58.4
105	43.4	26.5	66.5







#66
n-Nitrosodiphenylamine
Concen: 41.500 ng
RT: 15.645 min Scan# 2135
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:169 Resp: 1460282

Ion	Ratio	Lower	Upper
169	100		
168	66.1	53.0	79.4
167	35.8	28.9	43.3







#67
4-Bromophenyl-phenylether
Concen: 42.314 ng
RT: 16.334 min Scan# 2
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:248 Resp: 545579

Ion	Ratio	Lower	Upper
248	100		
250	99.3	78.2	117.4
141	76.1	57.7	86.5







#68
Hexachlorobenzene
Concen: 41.909 ng
RT: 16.451 min Scan# 2272
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:284 Resp: 642095

Ion	Ratio	Lower	Upper
284	100		
142	34.3	24.6	37.0
249	30.1	23.8	35.6







#69
Atrazine
Concen: 31.567 ng
RT: 16.622 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:200 Resp: 282908

Ion	Ratio	Lower	Upper
200	100		
173	25.4	7.5	47.5
215	48.7	28.8	68.8







#70
Pentachlorophenol
Concen: 42.900 ng
RT: 16.810 min Scan# 2333
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:266 Resp: 458529

Ion	Ratio	Lower	Upper
266	100		
268	65.3	51.0	76.6
264	63.0	49.8	74.6







#71
Phenanthrene
Concen: 40.615 ng
RT: 17.204 min Scan# 24
Delta R.T. -0.023 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:178 Resp: 2612052

Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.4	23.0
179	15.5	12.5	18.7







#72
Anthracene
Concen: 42.051 ng
RT: 17.298 min Scan# 2416
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:178 Resp: 2606508

Ion	Ratio	Lower	Upper
178	100		
176	18.6	14.7	22.1
179	15.4	12.4	18.6







#73
Carbazole
Concen: 42.351 ng
RT: 17.581 min Scan# 24
Delta R.T. -0.023 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:167 Resp: 2564955

Ion	Ratio	Lower	Upper
167	100		
166	21.8	17.1	25.7
139	13.1	10.6	16.0







#74
Di-n-butylphthalate
Concen: 44.539 ng
RT: 18.175 min Scan# 2565
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:149 Resp: 3360635

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.4	11.2
104	4.9	3.9	5.9







#75
Fluoranthene
Concen: 41.771 ng
RT: 19.298 min Scan# 2
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:202 Resp: 3195045

Ion	Ratio	Lower	Upper
202	100		
101	12.6	0.0	33.2
203	17.5	0.0	37.7







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.616 min Scan# 3150
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:240 Resp: 1288125

Ion	Ratio	Lower	Upper
240	100		
120	12.7	9.8	14.8
236	25.5	20.9	31.3







#77
Benzidine
Concen: 36.041 ng
RT: 19.492 min Scan# 2
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:184 Resp: 588487

Ion	Ratio	Lower	Upper
184	100		
185	14.3	11.6	17.4
183	11.3	9.2	13.8







#78
Pyrene
Concen: 40.022 ng
RT: 19.675 min Scan# 2820
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:202 Resp: 3276290

Ion	Ratio	Lower	Upper
202	100		
200	21.0	17.2	25.8
203	17.6	14.4	21.6







#79
Terphenyl-d14
Concen: 81.649 ng
RT: 19.892 min Scan# 2
Delta R.T. -0.029 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:244 Resp: 5217888

Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.6	8.4
122	10.9	9.4	14.0







#80
Butylbenzylphthalate
Concen: 43.107 ng
RT: 20.627 min Scan# 2982
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:149 Resp: 1511473

Ion	Ratio	Lower	Upper
149	100		
91	68.1	56.4	84.6
206	21.2	16.3	24.5







#81
Benzo(a)anthracene
Concen: 41.207 ng
RT: 21.598 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:228 Resp: 3299250

Ion	Ratio	Lower	Upper
228	100		
226	26.9	21.8	32.8
229	19.6	15.9	23.9







#82
3,3'-Dichlorobenzidine
Concen: 43.139 ng
RT: 21.516 min Scan# 3133
Delta R.T. -0.023 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:252 Resp: 1212290

Ion	Ratio	Lower	Upper
252	100		
254	64.4	51.7	77.5
126	11.8	8.8	13.2







#83
Chrysene
Concen: 40.888 ng
RT: 21.669 min Scan# 3
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:228 Resp: 3136332

Ion	Ratio	Lower	Upper
228	100		
226	30.0	24.2	36.2
229	19.8	15.6	23.4







#84
Bis(2-ethylhexyl)phthalate
Concen: 44.400 ng
RT: 21.539 min Scan# 3137
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:149 Resp: 2282222

Ion	Ratio	Lower	Upper
149	100		
167	27.9	22.4	33.6
279	5.2	4.0	6.0







#85
Di-n-octyl phthalate
Concen: 45.513 ng
RT: 22.827 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:149 Resp: 3803266

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	7.4	6.0	9.0







#86
Perylene-d12
Concen: 20.000 ng
RT: 24.986 min Scan# 3723
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:264 Resp: 1476872

Ion	Ratio	Lower	Upper
264	100		
260	23.6	18.9	28.3
265	22.1	17.8	26.6







#87
Indeno(1,2,3-cd)pyrene
Concen: 42.071 ng m
RT: 28.798 min Scan# 4
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:276 Resp: 4313548

Ion	Ratio	Lower	Upper
276	100		
138	16.8	14.5	21.7
277	20.9	20.2	30.2







#88
Benzo(b)fluoranthene
Concen: 41.288 ng
RT: 23.933 min Scan# 3544
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:252 Resp: 3654984

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.7	26.5
125	10.8	9.4	14.0







#89
Benzo(k)fluoranthene
Concen: 41.081 ng
RT: 23.998 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:252 Resp: 3512834

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	11.1	9.1	13.7







#90
Benzo(a)pyrene
Concen: 41.855 ng
RT: 24.833 min Scan# 3697
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:252 Resp: 3196130

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.6	26.4
125	11.9	10.1	15.1







#91
Dibenzo(a,h)anthracene
Concen: 41.902 ng
RT: 28.880 min Scan# 4
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:278 Resp: 3566000

Ion	Ratio	Lower	Upper
278	100		
139	16.8	12.6	19.0
279	24.0	18.5	27.7






#92
Benzo(g,h,i)perylene
Concen: 41.455 ng
RT: 29.992 min Scan# 4574
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:276 Resp: 3590938

Ion	Ratio	Lower	Upper
276	100		
277	23.4	18.9	28.3
138	21.2	18.7	28.1



Evaluate Continuing Calibration Report

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024283.D
Acq On : 14 Apr 2025 18:35
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Quant Time: Apr 15 04:54:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 15 04:48:42 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2		1,4-Dioxane	0.512	0.517	-1.0	95	0.00
3		Pyridine	1.351	1.349	0.1	91	0.00
4		n-Nitrosodimethylamine	0.448	0.451	-0.7	93	0.00
5	S	2-Fluorophenol	1.210	1.229	-1.6	92	0.00
6		Aniline	1.480	1.531	-3.4	91	0.00
7	S	Phenol-d6	1.656	1.678	-1.3	91	0.00
8		2-Chlorophenol	1.350	1.362	-0.9	92	0.00
9		Benzaldehyde	0.819	0.861	-5.1	97	0.00
10	C	Phenol	1.661	1.674	-0.8	90	0.00
11		bis(2-Chloroethyl)ether	1.339	1.335	0.3	90	0.00
12		1,3-Dichlorobenzene	1.454	1.465	-0.8	93	0.00
13	C	1,4-Dichlorobenzene	1.475	1.494	-1.3	94	0.00
14		1,2-Dichlorobenzene	1.424	1.421	0.2	93	0.00
15		Benzyl Alcohol	1.071	1.115	-4.1	91	0.00
16		2,2'-oxybis(1-Chloropropane	1.447	1.453	-0.4	90	0.00
17		2-Methylphenol	1.075	1.103	-2.6	91	0.00
18		Hexachloroethane	0.533	0.540	-1.3	94	0.00
19	P	n-Nitroso-di-n-propylamine	1.025	1.037	-1.2	90	0.00
20		3+4-Methylphenols	1.491	1.522	-2.1	90	0.00
21	I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
22		Acetophenone	0.495	0.515	-4.0	91	0.00
23	S	Nitrobenzene-d5	0.351	0.366	-4.3	91	0.00
24		Nitrobenzene	0.347	0.360	-3.7	90	0.00
25		Isophorone	0.635	0.661	-4.1	90	0.00
26	C	2-Nitrophenol	0.170	0.184	-8.2	92	0.00
27		2,4-Dimethylphenol	0.213	0.222	-4.2	90	0.00
28		bis(2-Chloroethoxy)methane	0.429	0.444	-3.5	90	0.00
29	C	2,4-Dichlorophenol	0.291	0.309	-6.2	91	0.00
30		1,2,4-Trichlorobenzene	0.311	0.324	-4.2	92	0.00
31		Naphthalene	1.054	1.090	-3.4	92	0.00
32		Benzoic acid	0.248	0.254	-2.4	96	-0.04
33		4-Chloroaniline	0.371	0.398	-7.3	92	0.00
34	C	Hexachlorobutadiene	0.183	0.192	-4.9	93	0.00
35		Caprolactam	0.107	0.113	-5.6	92	-0.02
36	C	4-Chloro-3-methylphenol	0.339	0.353	-4.1	90	0.00
37		2-Methylnaphthalene	0.731	0.756	-3.4	92	0.00
38		1-Methylnaphthalene	0.715	0.736	-2.9	92	0.00
39	I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
40		1,2,4,5-Tetrachlorobenzene	0.550	0.581	-5.6	92	0.00
41	P	Hexachlorocyclopentadiene	0.195	0.212	-8.7	89	-0.01
42	S	2,4,6-Tribromophenol	0.277	0.296	-6.9	93	-0.02
43	C	2,4,6-Trichlorophenol	0.366	0.396	-8.2	93	0.00
44		2,4,5-Trichlorophenol	0.405	0.434	-7.2	93	0.00
45	S	2-Fluorobiphenyl	1.306	1.383	-5.9	92	0.00
46		1,1'-Biphenyl	1.470	1.529	-4.0	91	-0.01
47		2-Chloronaphthalene	1.099	1.148	-4.5	92	0.00

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024283.D
Acq On : 14 Apr 2025 18:35
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Quant Time: Apr 15 04:54:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 15 04:48:42 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.308	0.327	-6.2	90	-0.01
49		Acenaphthylene	1.730	1.807	-4.5	91	0.00
50		Dimethylphthalate	1.454	1.508	-3.7	92	-0.02
51		2,6-Dinitrotoluene	0.309	0.332	-7.4	93	0.00
52	C	Acenaphthene	1.097	1.120	-2.1	91	-0.01
53		3-Nitroaniline	0.325	0.357	-9.8	92	0.00
54	P	2,4-Dinitrophenol	0.182	0.193	-6.0	95	-0.01
55		Dibenzofuran	1.793	1.823	-1.7	91	-0.02
56	P	4-Nitrophenol	0.280	0.315	-12.5	94	-0.01
57		2,4-Dinitrotoluene	0.421	0.458	-8.8	93	-0.01
58		Fluorene	1.388	1.448	-4.3	95	-0.01
59		2,3,4,6-Tetrachlorophenol	0.373	0.392	-5.1	92	0.00
60		Diethylphthalate	1.499	1.558	-3.9	92	-0.02
61		4-Chlorophenyl-phenylether	0.674	0.704	-4.5	94	-0.02
62		4-Nitroaniline	0.344	0.370	-7.6	92	-0.02
63		Azobenzene	1.378	1.436	-4.2	92	-0.02
64	I	Phenanthrene-d10	1.000	1.000	0.0	92	-0.02
65		4,6-Dinitro-2-methylphenol	0.126	0.135	-7.1	95	-0.02
66	c	n-Nitrosodiphenylamine	0.597	0.619	-3.7	91	-0.01
67		4-Bromophenyl-phenylether	0.219	0.231	-5.5	93	-0.01
68		Hexachlorobenzene	0.260	0.272	-4.6	93	-0.02
69		Atrazine	0.152	0.120	21.1	92	0.00
70	C	Pentachlorophenol	0.181	0.194	-7.2	92	-0.02
71		Phenanthrene	1.091	1.108	-1.6	91	-0.02
72		Anthracene	1.052	1.105	-5.0	92	-0.02
73		Carbazole	1.027	1.088	-5.9	93	-0.02
74		Di-n-butylphthalate	1.280	1.425	-11.3	94	-0.02
75	C	Fluoranthene	1.298	1.355	-4.4	93	-0.02
76	I	Chrysene-d12	1.000	1.000	0.0	95	0.00
77		Benzidine	0.254	0.228	10.2	50#	-0.02
78		Pyrene	1.271	1.272	-0.1	93	-0.01
79	S	Terphenyl-d14	0.992	1.013	-2.1	94	-0.03
80		Butylbenzylphthalate	0.544	0.587	-7.9	94	-0.02
81		Benzo(a)anthracene	1.243	1.281	-3.1	95	0.00
82		3,3'-Dichlorobenzidine	0.436	0.471	-8.0	92	-0.02
83		Chrysene	1.191	1.217	-2.2	96	-0.01
84		Bis(2-ethylhexyl)phthalate	0.798	0.886	-11.0	95	0.00
85	c	Di-n-octyl phthalate	1.297	1.476	-13.8	96	0.00
86	I	Perylene-d12	1.000	1.000	0.0	99	0.00
87		Indeno(1,2,3-cd)pyrene	1.388	1.460	-5.2	101	-0.02
88		Benzo(b)fluoranthene	1.199	1.237	-3.2	98	0.00
89		Benzo(k)fluoranthene	1.158	1.189	-2.7	98	0.00
90	C	Benzo(a)pyrene	1.034	1.082	-4.6	99	0.00
91		Dibenzo(a,h)anthracene	1.152	1.207	-4.8	100	0.00
92		Benzo(g,h,i)perylene	1.173	1.216	-3.7	99	0.00

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024283.D
Acq On : 14 Apr 2025 18:35
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Quant Time: Apr 15 04:54:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 15 04:48:42 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Evaluate Continuing Calibration Report

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024283.D
Acq On : 14 Apr 2025 18:35
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Quant Time: Apr 15 04:54:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 15 04:48:42 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	94	0.00
2		1,4-Dioxane	40.000	40.412	-1.0	95	0.00
3		Pyridine	40.000	39.955	0.1	91	0.00
4		n-Nitrosodimethylamine	40.000	40.205	-0.5	93	0.00
5	S	2-Fluorophenol	80.000	81.302	-1.6	92	0.00
6		Aniline	40.000	41.370	-3.4	91	0.00
7	S	Phenol-d6	80.000	81.063	-1.3	91	0.00
8		2-Chlorophenol	40.000	40.334	-0.8	92	0.00
9		Benzaldehyde	40.000	42.024	-5.1	97	0.00
10	C	Phenol	40.000	40.306	-0.8	90	0.00
11		bis(2-Chloroethyl)ether	40.000	39.902	0.2	90	0.00
12		1,3-Dichlorobenzene	40.000	40.298	-0.7	93	0.00
13	C	1,4-Dichlorobenzene	40.000	40.516	-1.3	94	0.00
14		1,2-Dichlorobenzene	40.000	39.898	0.3	93	0.00
15		Benzyl Alcohol	40.000	41.621	-4.1	91	0.00
16		2,2'-oxybis(1-Chloropropane	40.000	40.173	-0.4	90	0.00
17		2-Methylphenol	40.000	41.046	-2.6	91	0.00
18		Hexachloroethane	40.000	40.522	-1.3	94	0.00
19	P	n-Nitroso-di-n-propylamine	40.000	40.471	-1.2	90	0.00
20		3+4-Methylphenols	40.000	40.810	-2.0	90	0.00
21	I	Naphthalene-d8	20.000	20.000	0.0	91	0.00
22		Acetophenone	40.000	41.652	-4.1	91	0.00
23	S	Nitrobenzene-d5	80.000	83.531	-4.4	91	0.00
24		Nitrobenzene	40.000	41.537	-3.8	90	0.00
25		Isophorone	40.000	41.648	-4.1	90	0.00
26	C	2-Nitrophenol	40.000	43.376	-8.4	92	0.00
27		2,4-Dimethylphenol	40.000	41.689	-4.2	90	0.00
28		bis(2-Chloroethoxy)methane	40.000	41.402	-3.5	90	0.00
29	C	2,4-Dichlorophenol	40.000	42.461	-6.2	91	0.00
30		1,2,4-Trichlorobenzene	40.000	41.612	-4.0	92	0.00
31		Naphthalene	40.000	41.386	-3.5	92	0.00
32		Benzoic acid	40.000	40.959	-2.4	96	-0.04
33		4-Chloroaniline	40.000	42.937	-7.3	92	0.00
34	C	Hexachlorobutadiene	40.000	42.007	-5.0	93	0.00
35		Caprolactam	40.000	42.152	-5.4	92	-0.02
36	C	4-Chloro-3-methylphenol	40.000	41.737	-4.3	90	0.00
37		2-Methylnaphthalene	40.000	41.403	-3.5	92	0.00
38		1-Methylnaphthalene	40.000	41.210	-3.0	92	0.00
39	I	Acenaphthene-d10	20.000	20.000	0.0	91	0.00
40		1,2,4,5-Tetrachlorobenzene	40.000	42.275	-5.7	92	0.00
41	P	Hexachlorocyclopentadiene	40.000	43.639	-9.1	89	-0.01
42	S	2,4,6-Tribromophenol	80.000	85.488	-6.9	93	-0.02
43	C	2,4,6-Trichlorophenol	40.000	43.329	-8.3	93	0.00
44		2,4,5-Trichlorophenol	40.000	42.828	-7.1	93	0.00
45	S	2-Fluorobiphenyl	80.000	84.715	-5.9	92	0.00
46		1,1'-Biphenyl	40.000	41.595	-4.0	91	-0.01
47		2-Chloronaphthalene	40.000	41.786	-4.5	92	0.00

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024283.D
Acq On : 14 Apr 2025 18:35
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Quant Time: Apr 15 04:54:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 15 04:48:42 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	42.475	-6.2	90	-0.01
49		Acenaphthylene	40.000	41.783	-4.5	91	0.00
50		Dimethylphthalate	40.000	41.478	-3.7	92	-0.02
51		2,6-Dinitrotoluene	40.000	42.990	-7.5	93	0.00
52	C	Acenaphthene	40.000	40.853	-2.1	91	-0.01
53		3-Nitroaniline	40.000	43.931	-9.8	92	0.00
54	P	2,4-Dinitrophenol	40.000	42.485	-6.2	95	-0.01
55		Dibenzofuran	40.000	40.679	-1.7	91	-0.02
56	P	4-Nitrophenol	40.000	44.968	-12.4	94	-0.01
57		2,4-Dinitrotoluene	40.000	43.527	-8.8	93	-0.01
58		Fluorene	40.000	41.750	-4.4	95	-0.01
59		2,3,4,6-Tetrachlorophenol	40.000	41.995	-5.0	92	0.00
60		Diethylphthalate	40.000	41.570	-3.9	92	-0.02
61		4-Chlorophenyl-phenylether	40.000	41.777	-4.4	94	-0.02
62		4-Nitroaniline	40.000	43.114	-7.8	92	-0.02
63		Azobenzene	40.000	41.686	-4.2	92	-0.02
64	I	Phenanthrene-d10	20.000	20.000	0.0	92	-0.02
65		4,6-Dinitro-2-methylphenol	40.000	42.724	-6.8	95	-0.02
66	c	n-Nitrosodiphenylamine	40.000	41.500	-3.8	91	-0.01
67		4-Bromophenyl-phenylether	40.000	42.314	-5.8	93	-0.01
68		Hexachlorobenzene	40.000	41.909	-4.8	93	-0.02
69		Atrazine	40.000	31.567	21.1	92	0.00
70	C	Pentachlorophenol	40.000	42.900	-7.2	92	-0.02
71		Phenanthrene	40.000	40.615	-1.5	91	-0.02
72		Anthracene	40.000	42.051	-5.1	92	-0.02
73		Carbazole	40.000	42.351	-5.9	93	-0.02
74		Di-n-butylphthalate	40.000	44.539	-11.3	94	-0.02
75	C	Fluoranthene	40.000	41.771	-4.4	93	-0.02
76	I	Chrysene-d12	20.000	20.000	0.0	95	0.00
77		Benzidine	40.000	36.041	9.9	50	-0.02
78		Pyrene	40.000	40.022	-0.1	93	-0.01
79	S	Terphenyl-d14	80.000	81.649	-2.1	94	-0.03
80		Butylbenzylphthalate	40.000	43.107	-7.8	94	-0.02
81		Benzo(a)anthracene	40.000	41.207	-3.0	95	0.00
82		3,3'-Dichlorobenzidine	40.000	43.139	-7.8	92	-0.02
83		Chrysene	40.000	40.888	-2.2	96	-0.01
84		Bis(2-ethylhexyl)phthalate	40.000	44.400	-11.0	95	0.00
85	c	Di-n-octyl phthalate	40.000	45.513	-13.8	96	0.00
86	I	Perylene-d12	20.000	20.000	0.0	99	0.00
87		Indeno(1,2,3-cd)pyrene	40.000	42.071	-5.2	101	-0.02
88		Benzo(b)fluoranthene	40.000	41.288	-3.2	98	0.00
89		Benzo(k)fluoranthene	40.000	41.081	-2.7	98	0.00
90	C	Benzo(a)pyrene	40.000	41.855	-4.6	99	0.00
91		Dibenzo(a,h)anthracene	40.000	41.902	-4.8	100	0.00
92		Benzo(g,h,i)perylene	40.000	41.455	-3.6	99	0.00

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024283.D
Acq On : 14 Apr 2025 18:35
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Quant Time: Apr 15 04:54:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 15 04:48:42 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG No.: Q1901

Instrument ID: BNA_F Calibration Date/Time: 05/01/2025 10:17

Lab File ID: BF142250.D Init. Calib. Date(s): 04/30/2025 04/30/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:24 14:43

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.454	1.532		5.4	
2-Fluorophenol	1.252	1.250		-0.2	
Phenol-d6	1.512	1.575		4.2	
1,4-Dichlorobenzene	1.475	1.526		3.5	20.0
2-Methylphenol	1.060	1.103		4.1	
3+4-Methylphenols	1.326	1.404		5.9	
Nitrobenzene-d5	0.304	0.333		9.5	
Hexachloroethane	0.483	0.505		4.6	
Nitrobenzene	0.299	0.320		7.0	
Hexachlorobutadiene	0.175	0.186		6.3	20.0
2,4,6-Trichlorophenol	0.349	0.354		1.4	20.0
2-Fluorobiphenyl	1.418	1.409		-0.6	
2,4,5-Trichlorophenol	0.367	0.386		5.2	
2,4-Dinitrotoluene	0.239	0.289		20.9	
2,4,6-Tribromophenol	0.182	0.196		7.7	
Hexachlorobenzene	0.249	0.257		3.2	
Pentachlorophenol	0.132	0.137		3.8	20.0
Terphenyl-d14	1.375	1.422		3.4	

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142250.D
Acq On : 01 May 2025 10:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: May 01 11:46:27 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay

05/02/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	186578	20.000	ng	0.00
21) Naphthalene-d8	8.192	136	719586	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	381198	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	647816	20.000	ng	0.00
76) Chrysene-d12	14.074	240	384442	20.000	ng	0.00
86) Perylene-d12	15.562	264	349272	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	932561	79.825	ng	0.00
7) Phenol-d6	6.522	99	1175099	83.296	ng	0.00
23) Nitrobenzene-d5	7.475	82	958053	87.469	ng	0.00
42) 2,4,6-Tribromophenol	10.733	330	298805	85.983	ng	0.00
45) 2-Fluorobiphenyl	9.269	172	2147916	79.446	ng	0.00
79) Terphenyl-d14	13.021	244	2186580	82.722	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.699	88	216157	41.429	ng	98
3) Pyridine	3.463	79	571670	42.158	ng	98
4) n-Nitrosodimethylamine	3.410	42	297662	42.418	ng	97
6) Aniline	6.569	93	830257	41.579	ng	100
8) 2-Chlorophenol	6.687	128	475745	39.409	ng	98
9) Benzaldehyde	6.457	77	418699	42.464	ng	100
10) Phenol	6.540	94	617909m	40.684	ng	
11) bis(2-Chloroethyl)ether	6.640	93	494668	42.030	ng	99
12) 1,3-Dichlorobenzene	6.851	146	566821	41.505	ng	99
13) 1,4-Dichlorobenzene	6.928	146	569459	41.388	ng	100
14) 1,2-Dichlorobenzene	7.081	146	550181	41.885	ng	99
15) Benzyl Alcohol	7.045	79	411183	41.095	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.181	45	889834	41.870	ng	99
17) 2-Methylphenol	7.151	107	411493	41.610	ng	99
18) Hexachloroethane	7.422	117	188542	41.867	ng	94
19) n-Nitroso-di-n-propyla...	7.322	70	356048	40.842	ng	99
20) 3+4-Methylphenols	7.304	107	523765	42.351	ng	97
22) Acetophenone	7.316	105	690233	41.462	ng	98
24) Nitrobenzene	7.492	77	461069	42.902	ng	96
25) Isophorone	7.728	82	966817	42.059	ng	100
26) 2-Nitrophenol	7.804	139	140561	34.044	ng	99
27) 2,4-Dimethylphenol	7.834	122	485594	41.898	ng	100
28) bis(2-Chloroethoxy)met...	7.939	93	606286	41.820	ng	100
29) 2,4-Dichlorophenol	8.039	162	409683	42.280	ng	99
30) 1,2,4-Trichlorobenzene	8.134	180	453172	41.910	ng	99
31) Naphthalene	8.216	128	1482230	41.304	ng	100
32) Benzoic acid	7.928	122	168727m	34.514	ng	
33) 4-Chloroaniline	8.257	127	622889	41.683	ng	100
34) Hexachlorobutadiene	8.334	225	267370	42.382	ng	100
35) Caprolactam	8.622	113	129344	44.410	ng	98
36) 4-Chloro-3-methylphenol	8.728	107	431195	43.337	ng	99
37) 2-Methylnaphthalene	8.904	142	913767	41.414	ng	99
38) 1-Methylnaphthalene	9.004	142	950787	41.597	ng	99
40) 1,2,4,5-Tetrachloroben...	9.069	216	435999	40.641	ng	100
41) Hexachlorocyclopentadiene	9.057	237	241144	39.605	ng	99
43) 2,4,6-Trichlorophenol	9.175	196	269577	40.551	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142250.D
Acq On : 01 May 2025 10:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: May 01 11:46:27 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay

05/02/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.210	196	294308	42.108	ng		99
46) 1,1'-Biphenyl	9.369	154	1212716	40.922	ng		99
47) 2-Chloronaphthalene	9.392	162	896541	40.663	ng		99
48) 2-Nitroaniline	9.480	65	211351	37.508	ng		97
49) Acenaphthylene	9.810	152	1521943	40.906	ng		100
50) Dimethylphthalate	9.669	163	1026519	42.093	ng		100
51) 2,6-Dinitrotoluene	9.728	165	176156	38.527	ng		98
52) Acenaphthene	9.980	154	903201	41.576	ng		100
53) 3-Nitroaniline	9.892	138	225294	40.152	ng		100
54) 2,4-Dinitrophenol	9.992	184	42525	37.926	ng	#	1
55) Dibenzofuran	10.151	168	1350916	41.484	ng		99
56) 4-Nitrophenol	10.039	139	159868	41.706	ng		98
57) 2,4-Dinitrotoluene	10.128	165	220248	39.850	ng		99
58) Fluorene	10.498	166	1007989	40.709	ng		99
59) 2,3,4,6-Tetrachlorophenol	10.263	232	245919	42.475	ng		99
60) Diethylphthalate	10.369	149	1030685	42.799	ng		99
61) 4-Chlorophenyl-phenyle...	10.486	204	486030	41.333	ng		99
62) 4-Nitroaniline	10.504	138	215440	42.018	ng		98
63) Azobenzene	10.645	77	996298	41.895	ng		98
65) 4,6-Dinitro-2-methylph...	10.533	198	69294	35.905	ng		99
66) n-Nitrosodiphenylamine	10.604	169	914630	41.322	ng		100
67) 4-Bromophenyl-phenylether	10.980	248	304870	42.258	ng		99
68) Hexachlorobenzene	11.039	284	332960	41.207	ng		94
69) Atrazine	11.127	200	264074	45.262	ng		100
70) Pentachlorophenol	11.233	266	177109	41.491	ng		99
71) Phenanthrene	11.457	178	1446110	41.326	ng		100
72) Anthracene	11.510	178	1479001	41.339	ng		100
73) Carbazole	11.663	167	1352106	41.429	ng		100
74) Di-n-butylphthalate	11.998	149	1472584	44.864	ng		100
75) Fluoranthene	12.645	202	1485579	42.947	ng		100
77) Benzidine	12.769	184	754042	46.160	ng		99
78) Pyrene	12.874	202	1490017	42.016	ng		100
80) Butylbenzylphthalate	13.498	149	385725	40.417	ng		98
81) Benzo(a)anthracene	14.063	228	1059415	41.049	ng		99
82) 3,3'-Dichlorobenzidine	14.027	252	323049	42.615	ng		100
83) Chrysene	14.104	228	997135	42.445	ng		100
84) Bis(2-ethylhexyl)phtha...	14.057	149	477774	36.467	ng		100
85) Di-n-octyl phthalate	14.674	149	782817	33.845	ng		100
87) Indeno(1,2,3-cd)pyrene	17.074	276	1071010	41.864	ng		100
88) Benzo(b)fluoranthene	15.127	252	867766	39.998	ng		99
89) Benzo(k)fluoranthene	15.157	252	866064	43.643	ng		99
90) Benzo(a)pyrene	15.504	252	833612	41.969	ng		99
91) Dibenzo(a,h)anthracene	17.098	278	878980	42.333	ng		99
92) Benzo(g,h,i)perylene	17.533	276	873517	41.848	ng		100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142250.D
Acq On : 01 May 2025 10:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: May 01 11:46:27 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay

05/02/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:152 Resp: 18657

Ion	Ratio	Lower	Upper
152	100		
150	159.1	130.2	195.2
115	58.3	47.0	70.4



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#2
1,4-Dioxane
Concen: 41.429 ng
RT: 2.699 min Scan# 86
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion: 88 Resp: 216157

Ion	Ratio	Lower	Upper
88	100		
58	92.7	72.2	108.2
43	34.7	27.0	40.6






#3
Pyridine
Concen: 42.158 ng
RT: 3.463 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion: 79 Resp: 5716

Ion	Ratio	Lower	Upper
79	100		
52	87.4	71.5	107.3
51	42.2	34.5	51.7



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#4
n-Nitrosodimethylamine
Concen: 42.418 ng
RT: 3.410 min Scan# 207
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion: 42 Resp: 297662

Ion	Ratio	Lower	Upper
42	100		
74	108.6	89.5	134.3
44	8.5	6.6	9.8






#5
2-Fluorophenol
Concen: 79.825 ng
RT: 5.516 min Scan# 5
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:112 Resp: 93256

Ion	Ratio	Lower	Upper
112	100		
64	59.8	48.4	72.6
63	30.8	24.8	37.2



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025




#6
Aniline
Concen: 41.579 ng
RT: 6.569 min Scan# 744
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion: 93 Resp: 830257

Ion	Ratio	Lower	Upper
93	100		
66	40.1	32.1	48.1
65	20.7	16.6	24.8






#7
Phenol-d6
Concen: 83.296 ng
RT: 6.522 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion: 99 Resp: 1175099

Ion	Ratio	Lower	Upper
99	100		
42	21.0	16.3	24.5
71	31.9	25.0	37.4



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025




#8
2-Chlorophenol
Concen: 39.409 ng
RT: 6.687 min Scan# 764
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:128 Resp: 475745

Ion	Ratio	Lower	Upper
128	100		
130	32.3	13.1	53.1
64	49.9	28.6	68.6






#9
Benzaldehyde
Concen: 42.464 ng
RT: 6.457 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion: 77 Resp: 41869

Ion	Ratio	Lower	Upper
77	100		
105	98.6	79.1	119.1
106	95.6	76.0	116.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#10
Phenol
Concen: 40.684 ng m
RT: 6.540 min Scan# 739
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion: 94 Resp: 617909

Ion	Ratio	Lower	Upper
94	100		
65	28.8	8.9	48.9
66	37.5	18.7	58.7






#11
bis(2-Chloroethyl)ether
Concen: 42.030 ng
RT: 6.640 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion: 93 Resp: 49466

Ion	Ratio	Lower	Upper
93	100		
63	83.6	62.2	102.2
95	32.2	12.7	52.7



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#12
1,3-Dichlorobenzene
Concen: 41.505 ng
RT: 6.851 min Scan# 792
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:146 Resp: 566821

Ion	Ratio	Lower	Upper
146	100		
148	63.8	50.5	75.7
75	29.6	23.2	34.8






#13
1,4-Dichlorobenzene
Concen: 41.388 ng
RT: 6.928 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:146 Resp: 56945

Ion	Ratio	Lower	Upper
146	100		
148	63.1	50.6	76.0
111	42.0	33.4	50.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025




#14
1,2-Dichlorobenzene
Concen: 41.885 ng
RT: 7.081 min Scan# 831
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:146 Resp: 550181

Ion	Ratio	Lower	Upper
146	100		
148	63.9	50.5	75.7
111	43.0	34.9	52.3






#15
Benzyl Alcohol
Concen: 41.095 ng
RT: 7.045 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion: 79 Resp: 41118

Ion	Ratio	Lower	Upper
79	100		
108	81.7	64.5	96.7
77	62.1	49.2	73.8



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 41.870 ng
RT: 7.181 min Scan# 848
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion: 45 Resp: 889834

Ion	Ratio	Lower	Upper
45	100		
77	10.8	0.0	31.3
79	8.4	0.0	28.5






#17
2-Methylphenol
Concen: 41.610 ng
RT: 7.151 min Scan# 84
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:107 Resp: 41149

Ion	Ratio	Lower	Upper
107	100		
108	114.4	91.0	136.4
77	43.3	34.4	51.6
79	42.9	34.4	51.6



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#18
Hexachloroethane
Concen: 41.867 ng
RT: 7.422 min Scan# 889
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:117 Resp: 188542

Ion	Ratio	Lower	Upper
117	100		
119	94.5	77.1	115.7
201	90.2	79.4	119.2






#19
n-Nitroso-di-n-propylamine
Concen: 40.842 ng
RT: 7.322 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion: 70 Resp: 35604

Ion	Ratio	Lower	Upper
70	100		
42	66.0	53.4	80.2
101	9.7	7.7	11.5
130	21.3	17.2	25.8



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025




#20
3+4-Methylphenols
Concen: 42.351 ng
RT: 7.304 min Scan# 869
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:107 Resp: 523765

Ion	Ratio	Lower	Upper
107	100		
108	90.6	71.2	111.2
77	65.8	41.1	81.1
79	27.6	8.1	48.1




05/02/2025



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.192 min Scan# 1
Delta R.T. -0.001 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:136 Resp: 71958

Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.7	13.1
54	9.6	7.7	11.5
68	6.1	5.0	7.6



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#22
Acetophenone
Concen: 41.462 ng
RT: 7.316 min Scan# 871
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:105 Resp: 690233

Ion	Ratio	Lower	Upper
105	100		
71	7.2	5.7	8.5
51	36.0	30.1	45.1
120	23.7	19.0	28.4






#23
Nitrobenzene-d5
Concen: 87.469 ng
RT: 7.475 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion: 82 Resp: 95805

Ion	Ratio	Lower	Upper
82	100		
128	42.5	37.0	55.6
54	59.1	48.0	72.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025




#24
Nitrobenzene
Concen: 42.902 ng
RT: 7.492 min Scan# 901
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion: 77 Resp: 461069

Ion	Ratio	Lower	Upper
77	100		
123	45.1	38.6	58.0
65	13.5	11.5	17.3






#25
Isophorone
Concen: 42.059 ng
RT: 7.728 min Scan# 94
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion: 82 Resp: 96681

Ion	Ratio	Lower	Upper
82	100		
95	7.1	5.6	8.4
138	19.2	15.3	22.9



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#26
2-Nitrophenol
Concen: 34.044 ng
RT: 7.804 min Scan# 954
Delta R.T. 0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:139 Resp: 140561

Ion	Ratio	Lower	Upper
139	100		
109	27.7	21.8	32.8
65	52.3	41.1	61.7






#27
2,4-Dimethylphenol
Concen: 41.898 ng
RT: 7.834 min Scan# 9
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:122 Resp: 48559

Ion	Ratio	Lower	Upper
122	100		
107	107.9	86.0	129.0
121	57.8	46.3	69.5



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025




#28
bis(2-Chloroethoxy)methane
Concen: 41.820 ng
RT: 7.939 min Scan# 977
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion: 93 Resp: 606286

Ion	Ratio	Lower	Upper
93	100		
95	32.6	26.2	39.2
123	12.0	9.7	14.5






#29
2,4-Dichlorophenol
Concen: 42.280 ng
RT: 8.039 min Scan# 9
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:162 Resp: 40968

Ion	Ratio	Lower	Upper
162	100		
164	63.5	43.0	83.0
98	36.9	16.7	56.7



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025




#30
1,2,4-Trichlorobenzene
Concen: 41.910 ng
RT: 8.134 min Scan# 1010
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:180 Resp: 453172

Ion	Ratio	Lower	Upper
180	100		
182	93.3	75.7	113.5
145	31.3	25.4	38.0






#31
Naphthalene
Concen: 41.304 ng
RT: 8.216 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:128 Resp: 148223

Ion	Ratio	Lower	Upper
128	100		
129	11.3	9.0	13.4
127	13.2	10.6	15.8



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#32
Benzoic acid
Concen: 34.514 ng m
RT: 7.928 min Scan# 975
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:122 Resp: 168727

Ion	Ratio	Lower	Upper
122	100		
105	118.8	98.2	138.2
77	89.1	66.8	106.8






#33
4-Chloroaniline
Concen: 41.683 ng
RT: 8.257 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:127 Resp: 62288

Ion	Ratio	Lower	Upper
127	100		
129	32.6	26.1	39.1
65	31.6	25.5	38.3
92	19.3	15.4	23.2



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#34
Hexachlorobutadiene
Concen: 42.382 ng
RT: 8.334 min Scan# 1044
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:225 Resp: 267370

Ion	Ratio	Lower	Upper
225	100		
223	62.2	49.7	74.5
227	63.5	51.0	76.4






#35
Caprolactam
Concen: 44.410 ng
RT: 8.622 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:113 Resp: 12934

Ion	Ratio	Lower	Upper
113	100		
55	195.1	177.7	217.7
56	143.4	127.1	167.1



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#36
4-Chloro-3-methylphenol
Concen: 43.337 ng
RT: 8.728 min Scan# 1111
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:107 Resp: 431195

Ion	Ratio	Lower	Upper
107	100		
144	26.2	21.0	31.4
142	81.4	66.1	99.1






#37
2-Methylnaphthalene
Concen: 41.414 ng
RT: 8.904 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:142 Resp: 91376

Ion	Ratio	Lower	Upper
142	100		
141	87.6	70.5	105.7



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#38
1-Methylnaphthalene
Concen: 41.597 ng
RT: 9.004 min Scan# 1158
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:142 Resp: 950787

Ion	Ratio	Lower	Upper
142	100		
141	90.5	72.8	109.2






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.945 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:164 Resp: 38119

Ion	Ratio	Lower	Upper
164	100		
162	98.2	78.6	117.8
160	44.5	35.7	53.5



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 40.641 ng
RT: 9.069 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:216 Resp: 435999

Ion	Ratio	Lower	Upper
216	100		
214	79.7	63.8	95.6
179	21.1	16.7	25.1
108	19.0	15.6	23.4






#41
Hexachlorocyclopentadiene
Concen: 39.605 ng
RT: 9.057 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:237 Resp: 24114

Ion	Ratio	Lower	Upper
237	100		
235	62.7	43.5	83.5
272	12.6	0.0	32.9



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#42
2,4,6-Tribromophenol
Concen: 85.983 ng
RT: 10.733 min Scan# 1452
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:330 Resp: 298805

Ion	Ratio	Lower	Upper
330	100		
332	96.5	76.5	114.7
141	31.9	24.3	36.5






#43
2,4,6-Trichlorophenol
Concen: 40.551 ng
RT: 9.175 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:196 Resp: 26957

Ion	Ratio	Lower	Upper
196	100		
198	97.1	77.4	116.2
200	31.0	24.6	37.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025




#44
2,4,5-Trichlorophenol
Concen: 42.108 ng
RT: 9.210 min Scan# 1193
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:196 Resp: 294308

Ion	Ratio	Lower	Upper
196	100		
198	95.7	75.2	112.8
97	50.9	39.8	59.8
132	29.8	23.8	35.8






#45
2-Fluorobiphenyl
Concen: 79.446 ng
RT: 9.269 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:172 Resp: 214791

Ion	Ratio	Lower	Upper
172	100		
171	35.8	29.0	43.4
170	24.1	19.4	29.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025




#46
1,1'-Biphenyl
Concen: 40.922 ng
RT: 9.369 min Scan# 1220
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:154 Resp: 1212716

Ion	Ratio	Lower	Upper
154	100		
153	41.0	20.5	60.5
76	12.6	0.0	32.8






#47
2-Chloronaphthalene
Concen: 40.663 ng
RT: 9.392 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:162 Resp: 89654

Ion	Ratio	Lower	Upper
162	100		
127	41.5	32.6	49.0
164	33.3	26.1	39.1



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#48
2-Nitroaniline
Concen: 37.508 ng
RT: 9.480 min Scan# 1239
Delta R.T. -0.007 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion: 65 Resp: 211351

Ion	Ratio	Lower	Upper
65	100		
92	66.4	54.3	81.5
138	106.6	88.6	133.0






#49
Acenaphthylene
Concen: 40.906 ng
RT: 9.810 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:152 Resp: 152194

Ion	Ratio	Lower	Upper
152	100		
151	20.2	16.2	24.2
153	13.5	10.7	16.1



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#50
Dimethylphthalate
Concen: 42.093 ng
RT: 9.669 min Scan# 1271
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:163 Resp: 1026519

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.2	4.8
164	10.4	8.2	12.2




05/02/2025



#51
2,6-Dinitrotoluene
Concen: 38.527 ng
RT: 9.728 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:165 Resp: 17615

Ion	Ratio	Lower	Upper
165	100		
63	61.3	48.5	72.7
89	56.7	43.9	65.9



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#52
Acenaphthene
Concen: 41.576 ng
RT: 9.980 min Scan# 1324
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:154 Resp: 903201

Ion	Ratio	Lower	Upper
154	100		
153	110.3	88.6	133.0
152	52.6	42.1	63.1






#53
3-Nitroaniline
Concen: 40.152 ng
RT: 9.892 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:138 Resp: 22529

Ion	Ratio	Lower	Upper
138	100		
108	14.5	11.7	17.5
92	113.6	91.2	136.8



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#54
2,4-Dinitrophenol
Concen: 37.926 ng
RT: 9.992 min Scan# 1326
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:184 Resp: 42525

Ion	Ratio	Lower	Upper
184	100		
63	136.6	69.2	103.8#
154	883.4	271.4	407.2#






#55
Dibenzofuran
Concen: 41.484 ng
RT: 10.151 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:168 Resp: 135091

Ion	Ratio	Lower	Upper
168	100		
139	37.2	29.4	44.0
169	13.7	11.0	16.4



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#56
4-Nitrophenol
Concen: 41.706 ng
RT: 10.039 min Scan# 1334
Delta R.T. 0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:139 Resp: 159868

Ion	Ratio	Lower	Upper
139	100		
109	67.4	44.9	84.9
65	93.9	74.7	114.7






#57
2,4-Dinitrotoluene
Concen: 39.850 ng
RT: 10.128 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:165 Resp: 22024

Ion	Ratio	Lower	Upper
165	100		
63	48.8	38.8	58.2
89	72.1	57.1	85.7
182	2.5	2.2	3.2



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#58
Fluorene
Concen: 40.709 ng
RT: 10.498 min Scan# 1412
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:166 Resp: 1007989

Ion	Ratio	Lower	Upper
166	100		
165	93.4	73.9	110.9
167	13.7	10.8	16.2






#59
2,3,4,6-Tetrachlorophenol
Concen: 42.475 ng
RT: 10.263 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:232 Resp: 24591

Ion	Ratio	Lower	Upper
232	100		
131	46.0	35.9	53.9
130	2.1	1.8	2.6
166	29.4	24.0	36.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#60
Diethylphthalate
Concen: 42.799 ng
RT: 10.369 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:149 Resp: 1030685

Ion	Ratio	Lower	Upper
149	100		
177	21.7	17.6	26.4
150	12.5	10.2	15.2






#61
4-Chlorophenyl-phenylether
Concen: 41.333 ng
RT: 10.486 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:204 Resp: 48603

Ion	Ratio	Lower	Upper
204	100		
206	32.7	25.6	38.4
141	60.4	48.9	73.3



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#62
4-Nitroaniline
Concen: 42.018 ng
RT: 10.504 min Scan# 1413
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:138 Resp: 215440

Ion	Ratio	Lower	Upper
138	100		
92	50.8	29.2	69.2
108	70.7	49.1	89.1






#63
Azobenzene
Concen: 41.895 ng
RT: 10.645 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion: 77 Resp: 99629

Ion	Ratio	Lower	Upper
77	100		
182	25.2	7.3	47.3
105	22.3	2.5	42.5
51	35.1	14.4	54.4



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:188 Resp: 647816

Ion	Ratio	Lower	Upper
188	100		
94	9.9	8.2	12.2
80	10.7	8.6	12.8






#65
4,6-Dinitro-2-methylphenol
Concen: 35.905 ng
RT: 10.533 min Scan# 14
Delta R.T. -0.001 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:198 Resp: 6929

Ion	Ratio	Lower	Upper
198	100		
51	61.0	40.6	80.6
105	48.7	29.2	69.2



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025




#66
n-Nitrosodiphenylamine
Concen: 41.322 ng
RT: 10.604 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:169 Resp: 914630

Ion	Ratio	Lower	Upper
169	100		
168	66.2	53.0	79.6
167	36.5	28.8	43.2






#67
4-Bromophenyl-phenylether
Concen: 42.258 ng
RT: 10.980 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:248 Resp: 30487(

Ion	Ratio	Lower	Upper
248	100		
250	97.4	77.2	115.8
141	71.7	58.4	87.6



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025




#68
Hexachlorobenzene
Concen: 41.207 ng
RT: 11.039 min Scan# 1504
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:284 Resp: 332960

Ion	Ratio	Lower	Upper
284	100		
142	35.1	24.6	37.0
249	31.0	23.2	34.8






#69
Atrazine
Concen: 45.262 ng
RT: 11.127 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:200 Resp: 26407

Ion	Ratio	Lower	Upper
200	100		
173	28.4	8.6	48.6
215	48.7	28.5	68.5



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025




#70
Pentachlorophenol
Concen: 41.491 ng
RT: 11.233 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:266 Resp: 177109

Ion	Ratio	Lower	Upper
266	100		
268	63.2	50.0	75.0
264	61.4	49.3	73.9




05/02/2025



#71
Phenanthrene
Concen: 41.326 ng
RT: 11.457 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:178 Resp: 144611

Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.5	23.3
179	15.7	12.6	19.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025




#72
Anthracene
Concen: 41.339 ng
RT: 11.510 min Scan# 1584
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:178 Resp: 1479001

Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.3	22.9
179	15.8	12.7	19.1






#73
Carbazole
Concen: 41.429 ng
RT: 11.663 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:167 Resp: 135210

Ion	Ratio	Lower	Upper
167	100		
166	21.5	17.3	25.9
139	13.0	10.3	15.5



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#74
Di-n-butylphthalate
Concen: 44.864 ng
RT: 11.998 min Scan# 1667
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:149 Resp: 1472584

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.5	11.3
104	4.6	3.6	5.4






#75
Fluoranthene
Concen: 42.947 ng
RT: 12.645 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:202 Resp: 148557

Ion	Ratio	Lower	Upper
202	100		
101	12.3	0.0	32.4
203	17.8	0.0	37.5



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.074 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:240 Resp: 384442

Ion	Ratio	Lower	Upper
240	100		
120	12.5	10.1	15.1
236	25.7	20.5	30.7






#77
Benzidine
Concen: 46.160 ng
RT: 12.769 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:184 Resp: 75404

Ion	Ratio	Lower	Upper
184	100		
185	15.0	11.8	17.6
183	10.9	8.9	13.3



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025




#78
Pyrene
Concen: 42.016 ng
RT: 12.874 min Scan# 1816
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:202 Resp: 1490017

Ion	Ratio	Lower	Upper
202	100		
200	20.1	16.1	24.1
203	18.0	14.1	21.1






#79
Terphenyl-d14
Concen: 82.722 ng
RT: 13.021 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:244 Resp: 218658

Ion	Ratio	Lower	Upper
244	100		
212	6.7	5.3	7.9
122	11.3	9.1	13.7



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#80
Butylbenzylphthalate
Concen: 40.417 ng
RT: 13.498 min Scan# 1922
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:149 Resp: 385725

Ion	Ratio	Lower	Upper
149	100		
91	66.6	54.4	81.6
206	20.5	15.9	23.9






#81
Benzo(a)anthracene
Concen: 41.049 ng
RT: 14.063 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:228 Resp: 105941

Ion	Ratio	Lower	Upper
228	100		
226	27.1	21.9	32.9
229	19.6	16.0	24.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#82
3,3'-Dichlorobenzidine
Concen: 42.615 ng
RT: 14.027 min Scan# 2012
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:252 Resp: 323049

Ion	Ratio	Lower	Upper
252	100		
254	63.9	51.0	76.6
126	12.0	9.4	14.2






#83
Chrysene
Concen: 42.445 ng
RT: 14.104 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:228 Resp: 99713

Ion	Ratio	Lower	Upper
228	100		
226	30.1	24.0	36.0
229	19.8	15.7	23.5



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 36.467 ng
RT: 14.057 min Scan# 2017
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:149 Resp: 477774

Ion	Ratio	Lower	Upper
149	100		
167	27.1	21.7	32.5
279	3.7	3.0	4.4




05/02/2025



#85
Di-n-octyl phthalate
Concen: 33.845 ng
RT: 14.674 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:149 Resp: 78281

Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.1	1.7
43	11.7	9.3	13.9



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 15.562 min Scan# 2273
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:264 Resp: 349272

Ion	Ratio	Lower	Upper
264	100		
260	23.5	18.4	27.6
265	21.2	17.3	25.9






#87
Indeno(1,2,3-cd)pyrene
Concen: 41.864 ng
RT: 17.074 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:276 Resp: 107101

Ion	Ratio	Lower	Upper
276	100		
138	28.3	22.6	33.8
277	25.2	20.1	30.1



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025




#88
Benzo(b)fluoranthene
Concen: 39.998 ng
RT: 15.127 min Scan# 2199
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:252 Resp: 867766

Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.7	26.5
125	10.5	9.0	13.4






#89
Benzo(k)fluoranthene
Concen: 43.643 ng
RT: 15.157 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:252 Resp: 86606

Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.8	26.8
125	10.9	9.0	13.6



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut Upadhyay


05/02/2025




#90
Benzo(a)pyrene
Concen: 41.969 ng
RT: 15.504 min Scan# 2263
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:252 Resp: 833612

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.6	26.4
125	11.7	9.5	14.3






#91
Dibenzo(a,h)anthracene
Concen: 42.333 ng
RT: 17.098 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:278 Resp: 87898

Ion	Ratio	Lower	Upper
278	100		
139	18.3	15.2	22.8
279	23.5	19.0	28.6



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025



#92
Benzo(g,h,i)perylene
Concen: 41.848 ng
RT: 17.533 min Scan# 2608
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:276 Resp: 873517

Ion	Ratio	Lower	Upper
276	100		
277	23.4	18.7	28.1
138	24.3	19.4	29.0



Sample Results: **BF142250.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142250.D
Acq On : 01 May 2025 10:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 01 11:46:27 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2		1,4-Dioxane	0.559	0.579	-3.6	101	0.00
3		Pyridine	1.454	1.532	-5.4	104	0.00
4		n-Nitrosodimethylamine	0.752	0.798	-6.1	103	0.00
5	S	2-Fluorophenol	1.252	1.250	0.2	99	0.00
6		Aniline	2.140	2.225	-4.0	103	0.00
7	S	Phenol-d6	1.512	1.575	-4.2	102	0.00
8		2-Chlorophenol	1.294	1.275	1.5	96	0.00
9		Benzaldehyde	1.057	1.122	-6.1	104	0.00
10	C	Phenol	1.628	1.656	-1.7	100	0.00
11		bis(2-Chloroethyl)ether	1.262	1.326	-5.1	104	0.00
12		1,3-Dichlorobenzene	1.464	1.519	-3.8	103	0.00
13	C	1,4-Dichlorobenzene	1.475	1.526	-3.5	102	0.00
14		1,2-Dichlorobenzene	1.408	1.474	-4.7	104	0.00
15		Benzyl Alcohol	1.073	1.102	-2.7	99	0.00
16		2,2'-oxybis(1-Chloropropane	2.278	2.385	-4.7	104	0.00
17		2-Methylphenol	1.060	1.103	-4.1	102	0.00
18		Hexachloroethane	0.483	0.505	-4.6	103	0.00
19	P	n-Nitroso-di-n-propylamine	0.934	0.954	-2.1	102	0.00
20		3+4-Methylphenols	1.326	1.404	-5.9	104	0.00
21	I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
22		Acetophenone	0.463	0.480	-3.7	104	0.00
23	S	Nitrobenzene-d5	0.304	0.333	-9.5	100	0.00
24		Nitrobenzene	0.299	0.320	-7.0	100	0.00
25		Isophorone	0.639	0.672	-5.2	104	0.00
26	C	2-Nitrophenol	0.102	0.098	3.9	86	0.00
27		2,4-Dimethylphenol	0.322	0.337	-4.7	102	0.00
28		bis(2-Chloroethoxy)methane	0.403	0.421	-4.5	103	0.00
29	C	2,4-Dichlorophenol	0.269	0.285	-5.9	100	0.00
30		1,2,4-Trichlorobenzene	0.301	0.315	-4.7	103	0.00
31		Naphthalene	0.997	1.030	-3.3	103	0.00
32		Benzoic acid	0.125	0.117	6.4	92	0.00
33		4-Chloroaniline	0.415	0.433	-4.3	103	0.00
34	C	Hexachlorobutadiene	0.175	0.186	-6.3	103	0.00
35		Caprolactam	0.081	0.090	-11.1	106	0.00
36	C	4-Chloro-3-methylphenol	0.277	0.300	-8.3	103	0.00
37		2-Methylnaphthalene	0.613	0.635	-3.6	104	0.00
38		1-Methylnaphthalene	0.635	0.661	-4.1	103	0.00
39	I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
40		1,2,4,5-Tetrachlorobenzene	0.563	0.572	-1.6	102	0.00
41	P	Hexachlorocyclopentadiene	0.319	0.316	0.9	92	0.00
42	S	2,4,6-Tribromophenol	0.182	0.196	-7.7	99	0.00
43	C	2,4,6-Trichlorophenol	0.349	0.354	-1.4	98	0.00
44		2,4,5-Trichlorophenol	0.367	0.386	-5.2	98	0.00
45	S	2-Fluorobiphenyl	1.418	1.409	0.6	103	0.00
46		1,1'-Biphenyl	1.555	1.591	-2.3	103	0.00
47		2-Chloronaphthalene	1.157	1.176	-1.6	102	0.00

Sample Results: **BF142250.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142250.D
Acq On : 01 May 2025 10:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 01 11:46:27 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.249	0.277	-11.2	97	0.00
49		Acenaphthylene	1.952	1.996	-2.3	103	0.00
50		Dimethylphthalate	1.280	1.346	-5.2	105	0.00
51		2,6-Dinitrotoluene	0.204	0.231	-13.2	98	0.00
52	C	Acenaphthene	1.140	1.185	-3.9	104	0.00
53		3-Nitroaniline	0.257	0.296	-15.2	102	0.00
54	P	2,4-Dinitrophenol	0.058	0.056	3.4	101	0.00
55		Dibenzofuran	1.709	1.772	-3.7	104	0.00
56	P	4-Nitrophenol	0.201	0.210	-4.5	100	0.00
57		2,4-Dinitrotoluene	0.239	0.289	-20.9	104	0.00
58		Fluorene	1.299	1.322	-1.8	103	0.00
59		2,3,4,6-Tetrachlorophenol	0.304	0.323	-6.3	99	0.00
60		Diethylphthalate	1.263	1.352	-7.0	106	0.00
61		4-Chlorophenyl-phenylether	0.617	0.638	-3.4	104	0.00
62		4-Nitroaniline	0.238	0.283	-18.9	107	0.00
63		Azobenzene	1.248	1.307	-4.7	105	0.00
64	I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
65		4,6-Dinitro-2-methylphenol	0.054	0.053	1.9	102	0.00
66	c	n-Nitrosodiphenylamine	0.683	0.706	-3.4	104	0.00
67		4-Bromophenyl-phenylether	0.223	0.235	-5.4	105	0.00
68		Hexachlorobenzene	0.249	0.257	-3.2	104	0.00
69		Atrazine	0.180	0.204	-13.3	112	0.00
70	C	Pentachlorophenol	0.132	0.137	-3.8	100	0.00
71		Phenanthrene	1.080	1.116	-3.3	105	0.00
72		Anthracene	1.105	1.142	-3.3	106	0.00
73		Carbazole	1.008	1.044	-3.6	106	0.00
74		Di-n-butylphthalate	1.013	1.137	-12.2	112	0.00
75	C	Fluoranthene	1.068	1.147	-7.4	111	0.00
76	I	Chrysene-d12	1.000	1.000	0.0	116	0.00
77		Benzidine	0.850	0.981	-15.4	118	0.00
78		Pyrene	1.845	1.938	-5.0	111	0.00
79	S	Terphenyl-d14	1.375	1.422	-3.4	111	0.00
80		Butylbenzylphthalate	0.425	0.502	-18.1	115	0.00
81		Benzo(a)anthracene	1.343	1.378	-2.6	111	0.00
82		3,3'-Dichlorobenzidine	0.394	0.420	-6.6	110	0.00
83		Chrysene	1.222	1.297	-6.1	122	0.00
84		Bis(2-ethylhexyl)phthalate	0.604	0.621	-2.8	104	0.00
85	c	Di-n-octyl phthalate	1.133	1.018	10.2	94	0.00
86	I	Perylene-d12	1.000	1.000	0.0	106	0.00
87		Indeno(1,2,3-cd)pyrene	1.465	1.533	-4.6	101	0.00
88		Benzo(b)fluoranthene	1.242	1.242	0.0	101	0.00
89		Benzo(k)fluoranthene	1.136	1.240	-9.2	115	0.00
90	C	Benzo(a)pyrene	1.137	1.193	-4.9	105	0.00
91		Dibenzo(a,h)anthracene	1.189	1.258	-5.8	102	0.00
92		Benzo(g,h,i)perylene	1.195	1.250	-4.6	102	0.00

Sample Results: **BF142250.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142250.D
Acq On : 01 May 2025 10:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 01 11:46:27 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Sample Results: **BF142250.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142250.D
Acq On : 01 May 2025 10:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 01 11:46:27 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2		1,4-Dioxane	40.000	41.429	-3.6	101	0.00
3		Pyridine	40.000	42.158	-5.4	104	0.00
4		n-Nitrosodimethylamine	40.000	42.418	-6.0	103	0.00
5	S	2-Fluorophenol	80.000	79.825	0.2	99	0.00
6		Aniline	40.000	41.579	-3.9	103	0.00
7	S	Phenol-d6	80.000	83.296	-4.1	102	0.00
8		2-Chlorophenol	40.000	39.409	1.5	96	0.00
9		Benzaldehyde	40.000	42.464	-6.2	104	0.00
10	C	Phenol	40.000	40.684	-1.7	100	0.00
11		bis(2-Chloroethyl)ether	40.000	42.030	-5.1	104	0.00
12		1,3-Dichlorobenzene	40.000	41.505	-3.8	103	0.00
13	C	1,4-Dichlorobenzene	40.000	41.388	-3.5	102	0.00
14		1,2-Dichlorobenzene	40.000	41.885	-4.7	104	0.00
15		Benzyl Alcohol	40.000	41.095	-2.7	99	0.00
16		2,2'-oxybis(1-Chloropropane	40.000	41.870	-4.7	104	0.00
17		2-Methylphenol	40.000	41.610	-4.0	102	0.00
18		Hexachloroethane	40.000	41.867	-4.7	103	0.00
19	P	n-Nitroso-di-n-propylamine	40.000	40.842	-2.1	102	0.00
20		3+4-Methylphenols	40.000	42.351	-5.9	104	0.00
21	I	Naphthalene-d8	20.000	20.000	0.0	102	0.00
22		Acetophenone	40.000	41.462	-3.7	104	0.00
23	S	Nitrobenzene-d5	80.000	87.469	-9.3	100	0.00
24		Nitrobenzene	40.000	42.902	-7.3	100	0.00
25		Isophorone	40.000	42.059	-5.1	104	0.00
26	C	2-Nitrophenol	40.000	34.044	14.9	86	0.00
27		2,4-Dimethylphenol	40.000	41.898	-4.7	102	0.00
28		bis(2-Chloroethoxy)methane	40.000	41.820	-4.6	103	0.00
29	C	2,4-Dichlorophenol	40.000	42.280	-5.7	100	0.00
30		1,2,4-Trichlorobenzene	40.000	41.910	-4.8	103	0.00
31		Naphthalene	40.000	41.304	-3.3	103	0.00
32		Benzoic acid	40.000	34.514	13.7	92	0.00
33		4-Chloroaniline	40.000	41.683	-4.2	103	0.00
34	C	Hexachlorobutadiene	40.000	42.382	-6.0	103	0.00
35		Caprolactam	40.000	44.410	-11.0	106	0.00
36	C	4-Chloro-3-methylphenol	40.000	43.337	-8.3	103	0.00
37		2-Methylnaphthalene	40.000	41.414	-3.5	104	0.00
38		1-Methylnaphthalene	40.000	41.597	-4.0	103	0.00
39	I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40		1,2,4,5-Tetrachlorobenzene	40.000	40.641	-1.6	102	0.00
41	P	Hexachlorocyclopentadiene	40.000	39.605	1.0	92	0.00
42	S	2,4,6-Tribromophenol	80.000	85.983	-7.5	99	0.00
43	C	2,4,6-Trichlorophenol	40.000	40.551	-1.4	98	0.00
44		2,4,5-Trichlorophenol	40.000	42.108	-5.3	98	0.00
45	S	2-Fluorobiphenyl	80.000	79.446	0.7	103	0.00
46		1,1'-Biphenyl	40.000	40.922	-2.3	103	0.00
47		2-Chloronaphthalene	40.000	40.663	-1.7	102	0.00

Sample Results: **BF142250.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142250.D
Acq On : 01 May 2025 10:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 01 11:46:27 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	37.508	6.2	97	0.00
49		Acenaphthylene	40.000	40.906	-2.3	103	0.00
50		Dimethylphthalate	40.000	42.093	-5.2	105	0.00
51		2,6-Dinitrotoluene	40.000	38.527	3.7	98	0.00
52	C	Acenaphthene	40.000	41.576	-3.9	104	0.00
53		3-Nitroaniline	40.000	40.152	-0.4	102	0.00
54	P	2,4-Dinitrophenol	40.000	37.926	5.2	101	0.00
55		Dibenzofuran	40.000	41.484	-3.7	104	0.00
56	P	4-Nitrophenol	40.000	41.706	-4.3	100	0.00
57		2,4-Dinitrotoluene	40.000	39.850	0.4	104	0.00
58		Fluorene	40.000	40.709	-1.8	103	0.00
59		2,3,4,6-Tetrachlorophenol	40.000	42.475	-6.2	99	0.00
60		Diethylphthalate	40.000	42.799	-7.0	106	0.00
61		4-Chlorophenyl-phenylether	40.000	41.333	-3.3	104	0.00
62		4-Nitroaniline	40.000	42.018	-5.0	107	0.00
63		Azobenzene	40.000	41.895	-4.7	105	0.00
64	I	Phenanthrene-d10	20.000	20.000	0.0	105	0.00
65		4,6-Dinitro-2-methylphenol	40.000	35.905	10.2	102	0.00
66	c	n-Nitrosodiphenylamine	40.000	41.322	-3.3	104	0.00
67		4-Bromophenyl-phenylether	40.000	42.258	-5.6	105	0.00
68		Hexachlorobenzene	40.000	41.207	-3.0	104	0.00
69		Atrazine	40.000	45.262	-13.2	112	0.00
70	C	Pentachlorophenol	40.000	41.491	-3.7	100	0.00
71		Phenanthrene	40.000	41.326	-3.3	105	0.00
72		Anthracene	40.000	41.339	-3.3	106	0.00
73		Carbazole	40.000	41.429	-3.6	106	0.00
74		Di-n-butylphthalate	40.000	44.864	-12.2	112	0.00
75	C	Fluoranthene	40.000	42.947	-7.4	111	0.00
76	I	Chrysene-d12	20.000	20.000	0.0	116	0.00
77		Benzidine	40.000	46.160	-15.4	118	0.00
78		Pyrene	40.000	42.016	-5.0	111	0.00
79	S	Terphenyl-d14	80.000	82.722	-3.4	111	0.00
80		Butylbenzylphthalate	40.000	40.417	-1.0	115	0.00
81		Benzo(a)anthracene	40.000	41.049	-2.6	111	0.00
82		3,3'-Dichlorobenzidine	40.000	42.615	-6.5	110	0.00
83		Chrysene	40.000	42.445	-6.1	122	0.00
84		Bis(2-ethylhexyl)phthalate	40.000	36.467	8.8	104	0.00
85	c	Di-n-octyl phthalate	40.000	33.845	15.4	94	0.00
86	I	Perylene-d12	20.000	20.000	0.0	106	0.00
87		Indeno(1,2,3-cd)pyrene	40.000	41.864	-4.7	101	0.00
88		Benzo(b)fluoranthene	40.000	39.998	0.0	101	0.00
89		Benzo(k)fluoranthene	40.000	43.643	-9.1	115	0.00
90	C	Benzo(a)pyrene	40.000	41.969	-4.9	105	0.00
91		Dibenzo(a,h)anthracene	40.000	42.333	-5.8	102	0.00
92		Benzo(g,h,i)perylene	40.000	41.848	-4.6	102	0.00

Sample Results: **BF142250.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142250.D
Acq On : 01 May 2025 10:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 01 11:46:27 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Raw Data: **BF142274.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG No.: Q1901

Instrument ID: BNA_F Calibration Date/Time: 05/02/2025 10:39

Lab File ID: BF142274.D Init. Calib. Date(s): 04/30/2025 04/30/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:24 14:43

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.454	1.384		-4.8	
2-Fluorophenol	1.252	1.195		-4.6	
Phenol-d6	1.512	1.457		-3.6	
1,4-Dichlorobenzene	1.475	1.364		-7.5	20.0
2-Methylphenol	1.060	1.005		-5.2	
3+4-Methylphenols	1.326	1.295		-2.3	
Nitrobenzene-d5	0.304	0.332		9.2	
Hexachloroethane	0.483	0.473		-2.1	
Nitrobenzene	0.299	0.315		5.4	
Hexachlorobutadiene	0.175	0.164		-6.3	20.0
2,4,6-Trichlorophenol	0.349	0.360		3.2	20.0
2-Fluorobiphenyl	1.418	1.293		-8.8	
2,4,5-Trichlorophenol	0.367	0.363		-1.1	
2,4-Dinitrotoluene	0.239	0.314		31.4	
2,4,6-Tribromophenol	0.182	0.190		4.4	
Hexachlorobenzene	0.249	0.232		-6.8	
Pentachlorophenol	0.132	0.137		3.8	20.0
Terphenyl-d14	1.375	1.344		-2.3	

All other compounds must meet a minimum RRF of 0.010.

Sample Results: **BF142274.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142274.D
Acq On : 02 May 2025 10:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: May 02 11:50:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	189457	20.000	ng	0.00
21) Naphthalene-d8	8.192	136	747235	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	389025	20.000	ng	0.00
64) Phenanthrene-d10	11.427	188	652685	20.000	ng	0.00
76) Chrysene-d12	14.068	240	354752	20.000	ng	0.00
86) Perylene-d12	15.557	264	327562	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	905917	76.366	ng	0.00
7) Phenol-d6	6.522	99	1104282	77.087	ng	0.00
23) Nitrobenzene-d5	7.469	82	992335	87.247	ng	0.00
42) 2,4,6-Tribromophenol	10.733	330	295525	83.329	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	2012359	72.935	ng	0.00
79) Terphenyl-d14	13.016	244	1907599	78.207	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.687	88	199149	37.589	ng	97
3) Pyridine	3.457	79	524389	38.083	ng	99
4) n-Nitrosodimethylamine	3.404	42	276556	38.811	ng	99
6) Aniline	6.569	93	768076	37.880	ng	99
8) 2-Chlorophenol	6.687	128	480593	39.206	ng	98
9) Benzaldehyde	6.457	77	378814	37.835	ng	99
10) Phenol	6.540	94	585536m	37.967	ng	
11) bis(2-Chloroethyl)ether	6.640	93	452236	37.841	ng	99
12) 1,3-Dichlorobenzene	6.851	146	524346	37.811	ng	100
13) 1,4-Dichlorobenzene	6.928	146	516930	36.999	ng	99
14) 1,2-Dichlorobenzene	7.081	146	505354	37.888	ng	99
15) Benzyl Alcohol	7.045	79	393158	38.697	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.181	45	817791	37.895	ng	99
17) 2-Methylphenol	7.151	107	380905	37.931	ng	99
18) Hexachloroethane	7.422	117	179209	39.190	ng	96
19) n-Nitroso-di-n-propyla...	7.316	70	330837	37.373	ng	97
20) 3+4-Methylphenols	7.304	107	490537	39.062	ng	94
22) Acetophenone	7.316	105	628918	36.381	ng	98
24) Nitrobenzene	7.487	77	470320	42.143	ng	99
25) Isophorone	7.728	82	878449	36.800	ng	99
26) 2-Nitrophenol	7.804	139	201999	43.995	ng	97
27) 2,4-Dimethylphenol	7.834	122	458385	38.087	ng	100
28) bis(2-Chloroethoxy)met...	7.934	93	552958	36.730	ng	100
29) 2,4-Dichlorophenol	8.039	162	389539	38.714	ng	99
30) 1,2,4-Trichlorobenzene	8.128	180	414934	36.954	ng	100
31) Naphthalene	8.210	128	1363971	36.602	ng	100
32) Benzoic acid	7.928	122	233233m	42.364	ng	
33) 4-Chloroaniline	8.257	127	579952	37.373	ng	99
34) Hexachlorobutadiene	8.328	225	245743	37.512	ng	98
35) Caprolactam	8.622	113	120939	39.988	ng	98
36) 4-Chloro-3-methylphenol	8.728	107	397404	38.463	ng	98
37) 2-Methylnaphthalene	8.904	142	837448	36.551	ng	100
38) 1-Methylnaphthalene	9.004	142	871131	36.702	ng	100
40) 1,2,4,5-Tetrachloroben...	9.069	216	405798	37.065	ng	99
41) Hexachlorocyclopentadiene	9.057	237	254671	40.985	ng	99
43) 2,4,6-Trichlorophenol	9.175	196	280415	41.332	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142274.D
Acq On : 02 May 2025 10:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: May 02 11:50:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.210	196	282545	39.611	ng	98
46) 1,1'-Biphenyl	9.363	154	1118632	36.988	ng	99
47) 2-Chloronaphthalene	9.392	162	833002	37.021	ng	100
48) 2-Nitroaniline	9.481	65	245382	42.014	ng	97
49) Acenaphthylene	9.810	152	1407244	37.062	ng	99
50) Dimethylphthalate	9.663	163	931138	37.413	ng	100
51) 2,6-Dinitrotoluene	9.722	165	191021	40.686	ng	98
52) Acenaphthene	9.981	154	832658	37.558	ng	99
53) 3-Nitroaniline	9.892	138	232079	40.499	ng	97
54) 2,4-Dinitrophenol	9.992	184	66816	50.599	ng	91
55) Dibenzofuran	10.151	168	1230121	37.015	ng	100
56) 4-Nitrophenol	10.033	139	176430	45.101	ng	98
57) 2,4-Dinitrotoluene	10.122	165	244640	42.928	ng	100
58) Fluorene	10.492	166	931222	36.852	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.263	232	240492	40.702	ng	99
60) Diethylphthalate	10.363	149	912963	37.148	ng	100
61) 4-Chlorophenyl-phenyle...	10.486	204	441273	36.772	ng	97
62) 4-Nitroaniline	10.504	138	220163	42.072	ng	98
63) Azobenzene	10.645	77	895687	36.907	ng	99
65) 4,6-Dinitro-2-methylph...	10.528	198	97288	45.858	ng	97
66) n-Nitrosodiphenylamine	10.604	169	829293	37.187	ng	100
67) 4-Bromophenyl-phenylether	10.975	248	277516	38.179	ng	100
68) Hexachlorobenzene	11.039	284	302639	37.175	ng	98
69) Atrazine	11.128	200	228706	38.907	ng	99
70) Pentachlorophenol	11.227	266	178739	41.560	ng	98
71) Phenanthrene	11.457	178	1313668	37.261	ng	100
72) Anthracene	11.510	178	1349761	37.446	ng	99
73) Carbazole	11.657	167	1245984	37.893	ng	99
74) Di-n-butylphthalate	11.992	149	1235690	37.366	ng	99
75) Fluoranthene	12.645	202	1316477	37.775	ng	99
77) Benzidine	12.763	184	629233	41.743	ng	100
78) Pyrene	12.874	202	1327219	40.557	ng	100
80) Butylbenzylphthalate	13.492	149	318246	36.531	ng	98
81) Benzo(a)anthracene	14.057	228	917653	38.532	ng	99
82) 3,3'-Dichlorobenzidine	14.021	252	264696	37.840	ng	99
83) Chrysene	14.098	228	786998	36.304	ng	99
84) Bis(2-ethylhexyl)phtha...	14.051	149	344949	29.284	ng	100
85) Di-n-octyl phthalate	14.668	149	576205	28.225	ng	100
87) Indeno(1,2,3-cd)pyrene	17.068	276	957230	39.896	ng	100
88) Benzo(b)fluoranthene	15.121	252	702872	34.545	ng	99
89) Benzo(k)fluoranthene	15.151	252	719163	38.642	ng	99
90) Benzo(a)pyrene	15.498	252	703822	37.783	ng	100
91) Dibenzo(a,h)anthracene	17.086	278	781717	40.144	ng	99
92) Benzo(g,h,i)perylene	17.521	276	796796	40.703	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142274.D
Acq On : 02 May 2025 10:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1


Quant Time: May 02 11:50:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:152 Resp: 18945


Ion	Ratio	Lower	Upper
152	100		
150	162.5	130.2	195.2
115	59.4	47.0	70.4



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#2
1,4-Dioxane
Concen: 37.589 ng
RT: 2.687 min Scan# 84
Delta R.T. -0.012 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion: 88 Resp: 199149

Ion	Ratio	Lower	Upper
88	100		
58	92.8	72.2	108.2
43	35.2	27.0	40.6






#3
Pyridine
Concen: 38.083 ng
RT: 3.457 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion: 79 Resp: 52438


Ion	Ratio	Lower	Upper
79	100		
52	88.7	71.5	107.3
51	43.0	34.5	51.7



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#4
n-Nitrosodimethylamine
Concen: 38.811 ng
RT: 3.404 min Scan# 206
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion: 42 Resp: 276556

Ion	Ratio	Lower	Upper
42	100		
74	112.7	89.5	134.3
44	7.7	6.6	9.8





#5
2-Fluorophenol
Concen: 76.366 ng
RT: 5.516 min Scan# 5
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39


Tgt Ion:112 Resp: 90591


Ion	Ratio	Lower	Upper
112	100		
64	59.5	48.4	72.6
63	30.8	24.8	37.2

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#6
Aniline
Concen: 37.880 ng
RT: 6.569 min Scan# 744
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion: 93 Resp: 768076

Ion	Ratio	Lower	Upper
93	100		
66	39.8	32.1	48.1
65	20.4	16.6	24.8






#7
Phenol-d6
Concen: 77.087 ng
RT: 6.522 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion: 99 Resp: 110428


Ion	Ratio	Lower	Upper
99	100		
42	21.5	16.3	24.5
71	32.0	25.0	37.4



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#8
2-Chlorophenol
Concen: 39.206 ng
RT: 6.687 min Scan# 764
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:128 Resp: 480593

Ion	Ratio	Lower	Upper
128	100		
130	32.5	13.1	53.1
64	49.8	28.6	68.6






#9
Benzaldehyde
Concen: 37.835 ng
RT: 6.457 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion: 77 Resp: 37881


Ion	Ratio	Lower	Upper
77	100		
105	99.9	79.1	119.1
106	95.6	76.0	116.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#10
Phenol
Concen: 37.967 ng m
RT: 6.540 min Scan# 739
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion: 94 Resp: 585536

Ion	Ratio	Lower	Upper
94	100		
65	28.8	8.9	48.9
66	38.2	18.7	58.7






#11
bis(2-Chloroethyl)ether
Concen: 37.841 ng
RT: 6.640 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion: 93 Resp: 45223


Ion	Ratio	Lower	Upper
93	100		
63	83.2	62.2	102.2
95	32.1	12.7	52.7



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#12
1,3-Dichlorobenzene
Concen: 37.811 ng
RT: 6.851 min Scan# 792
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:146 Resp: 524346

Ion	Ratio	Lower	Upper
146	100		
148	63.1	50.5	75.7
75	29.1	23.2	34.8






#13
1,4-Dichlorobenzene
Concen: 36.999 ng
RT: 6.928 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:146 Resp: 51693


Ion	Ratio	Lower	Upper
146	100		
148	64.5	50.6	76.0
111	42.4	33.4	50.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#14
1,2-Dichlorobenzene
Concen: 37.888 ng
RT: 7.081 min Scan# 831
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:146 Resp: 505354

Ion	Ratio	Lower	Upper
146	100		
148	63.9	50.5	75.7
111	42.9	34.9	52.3





#15
Benzyl Alcohol
Concen: 38.697 ng
RT: 7.045 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39


Tgt Ion: 79 Resp: 39315


Ion	Ratio	Lower	Upper
79	100		
108	81.8	64.5	96.7
77	61.8	49.2	73.8

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#16
2,2'-oxybis(1-Chloropropane)
Concen: 37.895 ng
RT: 7.181 min Scan# 848
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion: 45 Resp: 817791

Ion	Ratio	Lower	Upper
45	100		
77	10.6	0.0	31.3
79	8.3	0.0	28.5






#17
2-Methylphenol
Concen: 37.931 ng
RT: 7.151 min Scan# 84
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:107 Resp: 38090


Ion	Ratio	Lower	Upper
107	100		
108	114.8	91.0	136.4
77	43.4	34.4	51.6
79	42.5	34.4	51.6



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#18
Hexachloroethane
Concen: 39.190 ng
RT: 7.422 min Scan# 889
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:117 Resp: 179209

Ion	Ratio	Lower	Upper
117	100		
119	96.8	77.1	115.7
201	92.0	79.4	119.2






#19
n-Nitroso-di-n-propylamine
Concen: 37.373 ng
RT: 7.316 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion: 70 Resp: 33083


Ion	Ratio	Lower	Upper
70	100		
42	70.2	53.4	80.2
101	9.7	7.7	11.5
130	21.2	17.2	25.8



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#20
3+4-Methylphenols
Concen: 39.062 ng
RT: 7.304 min Scan# 869
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:107 Resp: 490537

Ion	Ratio	Lower	Upper
107	100		
108	90.4	71.2	111.2
77	73.0	41.1	81.1
79	27.9	8.1	48.1






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.192 min Scan# 1
Delta R.T. -0.001 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:136 Resp: 74723


Ion	Ratio	Lower	Upper
136	100		
137	10.7	8.7	13.1
54	9.1	7.7	11.5
68	6.0	5.0	7.6



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#22
Acetophenone
Concen: 36.381 ng
RT: 7.316 min Scan# 871
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:105 Resp: 628918

Ion	Ratio	Lower	Upper
105	100		
71	7.8	5.7	8.5
51	35.6	30.1	45.1
120	24.0	19.0	28.4






#23
Nitrobenzene-d5
Concen: 87.247 ng
RT: 7.469 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion: 82 Resp: 99233


Ion	Ratio	Lower	Upper
82	100		
128	46.4	37.0	55.6
54	60.8	48.0	72.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#24
Nitrobenzene
Concen: 42.143 ng
RT: 7.487 min Scan# 900
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion: 77 Resp: 470320

Ion	Ratio	Lower	Upper
77	100		
123	47.0	38.6	58.0
65	14.5	11.5	17.3





#25
Isophorone
Concen: 36.800 ng
RT: 7.728 min Scan# 94
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39


Tgt Ion: 82 Resp: 87844


Ion	Ratio	Lower	Upper
82	100		
95	7.1	5.6	8.4
138	19.7	15.3	22.9

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#26
2-Nitrophenol
Concen: 43.995 ng
RT: 7.804 min Scan# 954
Delta R.T. 0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:139 Resp: 201999

Ion	Ratio	Lower	Upper
139	100		
109	26.2	21.8	32.8
65	48.9	41.1	61.7






#27
2,4-Dimethylphenol
Concen: 38.087 ng
RT: 7.834 min Scan# 9
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:122 Resp: 45838


Ion	Ratio	Lower	Upper
122	100		
107	107.6	86.0	129.0
121	58.0	46.3	69.5



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#28
bis(2-Chloroethoxy)methane
Concen: 36.730 ng
RT: 7.934 min Scan# 976
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion: 93 Resp: 552958

Ion	Ratio	Lower	Upper
93	100		
95	32.7	26.2	39.2
123	12.4	9.7	14.5






#29
2,4-Dichlorophenol
Concen: 38.714 ng
RT: 8.039 min Scan# 9
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:162 Resp: 38953


Ion	Ratio	Lower	Upper
162	100		
164	64.3	43.0	83.0
98	36.3	16.7	56.7



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#30
1,2,4-Trichlorobenzene
Concen: 36.954 ng
RT: 8.128 min Scan# 1009
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:180 Resp: 414934

Ion	Ratio	Lower	Upper
180	100		
182	94.7	75.7	113.5
145	32.0	25.4	38.0





#31
Naphthalene
Concen: 36.602 ng
RT: 8.210 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39


Tgt Ion:128 Resp: 136397


Ion	Ratio	Lower	Upper
128	100		
129	11.3	9.0	13.4
127	13.0	10.6	15.8

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#32
Benzoic acid
Concen: 42.364 ng m
RT: 7.928 min Scan# 975
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:122 Resp: 233233

Ion	Ratio	Lower	Upper
122	100		
105	118.1	98.2	138.2
77	89.8	66.8	106.8






#33
4-Chloroaniline
Concen: 37.373 ng
RT: 8.257 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:127 Resp: 57995


Ion	Ratio	Lower	Upper
127	100		
129	32.3	26.1	39.1
65	31.1	25.5	38.3
92	19.1	15.4	23.2



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#34
Hexachlorobutadiene
Concen: 37.512 ng
RT: 8.328 min Scan# 1043
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:225 Resp: 245743

Ion	Ratio	Lower	Upper
225	100		
223	62.8	49.7	74.5
227	62.0	51.0	76.4






#35
Caprolactam
Concen: 39.988 ng
RT: 8.622 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:113 Resp: 12093


Ion	Ratio	Lower	Upper
113	100		
55	194.6	177.7	217.7
56	143.4	127.1	167.1



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#36
4-Chloro-3-methylphenol
Concen: 38.463 ng
RT: 8.728 min Scan# 1111
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:107 Resp: 397404

Ion	Ratio	Lower	Upper
107	100		
144	26.9	21.0	31.4
142	84.0	66.1	99.1






#37
2-Methylnaphthalene
Concen: 36.551 ng
RT: 8.904 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:142 Resp: 83744


Ion	Ratio	Lower	Upper
142	100		
141	88.5	70.5	105.7



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#38
1-Methylnaphthalene
Concen: 36.702 ng
RT: 9.004 min Scan# 1158
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:142 Resp: 871131

Ion	Ratio	Lower	Upper
142	100		
141	91.0	72.8	109.2






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.945 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:164 Resp: 38902


Ion	Ratio	Lower	Upper
164	100		
162	98.2	78.6	117.8
160	44.4	35.7	53.5



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 37.065 ng
RT: 9.069 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:216 Resp: 405798

Ion	Ratio	Lower	Upper
216	100		
214	79.9	63.8	95.6
179	21.4	16.7	25.1
108	18.8	15.6	23.4






#41
Hexachlorocyclopentadiene
Concen: 40.985 ng
RT: 9.057 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:237 Resp: 25467


Ion	Ratio	Lower	Upper
237	100		
235	62.7	43.5	83.5
272	13.3	0.0	32.9



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#42
2,4,6-Tribromophenol
Concen: 83.329 ng
RT: 10.733 min Scan# 1452
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:330 Resp: 295525

Ion	Ratio	Lower	Upper
330	100		
332	96.1	76.5	114.7
141	31.6	24.3	36.5






#43
2,4,6-Trichlorophenol
Concen: 41.332 ng
RT: 9.175 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:196 Resp: 28041


Ion	Ratio	Lower	Upper
196	100		
198	96.2	77.4	116.2
200	31.1	24.6	37.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#44
2,4,5-Trichlorophenol
Concen: 39.611 ng
RT: 9.210 min Scan# 1193
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:196 Resp: 282545

Ion	Ratio	Lower	Upper
196	100		
198	95.4	75.2	112.8
97	48.3	39.8	59.8
132	28.9	23.8	35.8






#45
2-Fluorobiphenyl
Concen: 72.935 ng
RT: 9.263 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:172 Resp: 201235


Ion	Ratio	Lower	Upper
172	100		
171	35.8	29.0	43.4
170	23.7	19.4	29.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#46
1,1'-Biphenyl
Concen: 36.988 ng
RT: 9.363 min Scan# 1219
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:154 Resp: 1118632

Ion	Ratio	Lower	Upper
154	100		
153	39.7	20.5	60.5
76	13.1	0.0	32.8






#47
2-Chloronaphthalene
Concen: 37.021 ng
RT: 9.392 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:162 Resp: 83300


Ion	Ratio	Lower	Upper
162	100		
127	40.6	32.6	49.0
164	32.9	26.1	39.1



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#48
2-Nitroaniline
Concen: 42.014 ng
RT: 9.481 min Scan# 1239
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion: 65 Resp: 245382

Ion	Ratio	Lower	Upper
65	100		
92	66.8	54.3	81.5
138	107.4	88.6	133.0






#49
Acenaphthylene
Concen: 37.062 ng
RT: 9.810 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:152 Resp: 140724


Ion	Ratio	Lower	Upper
152	100		
151	19.8	16.2	24.2
153	13.3	10.7	16.1



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#50
Dimethylphthalate
Concen: 37.413 ng
RT: 9.663 min Scan# 1270
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:163 Resp: 931138

Ion	Ratio	Lower	Upper
163	100		
194	4.0	3.2	4.8
164	10.3	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 40.686 ng
RT: 9.722 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39


Tgt Ion:165 Resp: 19102


Ion	Ratio	Lower	Upper
165	100		
63	63.3	48.5	72.7
89	55.3	43.9	65.9

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#52
Acenaphthene
Concen: 37.558 ng
RT: 9.981 min Scan# 1324
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:154 Resp: 832658

Ion	Ratio	Lower	Upper
154	100		
153	110.1	88.6	133.0
152	52.1	42.1	63.1





#53
3-Nitroaniline
Concen: 40.499 ng
RT: 9.892 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39


Tgt Ion:138 Resp: 23207


Ion	Ratio	Lower	Upper
138	100		
108	11.7	11.7	17.5
92	111.1	91.2	136.8

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#54
2,4-Dinitrophenol
Concen: 50.599 ng
RT: 9.992 min Scan# 1326
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:184 Resp: 66816

Ion	Ratio	Lower	Upper
184	100		
63	90.7	69.2	103.8
154	361.3	271.4	407.2






#55
Dibenzofuran
Concen: 37.015 ng
RT: 10.151 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:168 Resp: 123012


Ion	Ratio	Lower	Upper
168	100		
139	36.5	29.4	44.0
169	13.4	11.0	16.4



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#56
4-Nitrophenol
Concen: 45.101 ng
RT: 10.033 min Scan# 1333
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:139 Resp: 176430

Ion	Ratio	Lower	Upper
139	100		
109	66.5	44.9	84.9
65	97.0	74.7	114.7






#57
2,4-Dinitrotoluene
Concen: 42.928 ng
RT: 10.122 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:165 Resp: 24464


Ion	Ratio	Lower	Upper
165	100		
63	48.6	38.8	58.2
89	70.9	57.1	85.7
182	2.7	2.2	3.2



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#58
Fluorene
Concen: 36.852 ng
RT: 10.492 min Scan# 1411
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:166 Resp: 931222

Ion	Ratio	Lower	Upper
166	100		
165	93.3	73.9	110.9
167	13.5	10.8	16.2






#59
2,3,4,6-Tetrachlorophenol
Concen: 40.702 ng
RT: 10.263 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:232 Resp: 24049


Ion	Ratio	Lower	Upper
232	100		
131	45.3	35.9	53.9
130	2.2	1.8	2.6
166	29.5	24.0	36.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#60
Diethylphthalate
Concen: 37.148 ng
RT: 10.363 min Scan# 1389
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:149 Resp: 912963

Ion	Ratio	Lower	Upper
149	100		
177	21.7	17.6	26.4
150	12.6	10.2	15.2





#61
4-Chlorophenyl-phenylether
Concen: 36.772 ng
RT: 10.486 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39


Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:204 Resp: 44127

Ion	Ratio	Lower	Upper
204	100		
206	32.9	25.6	38.4
141	58.7	48.9	73.3

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#62
4-Nitroaniline
Concen: 42.072 ng
RT: 10.504 min Scan# 1413
Delta R.T. 0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:138 Resp: 220163

Ion	Ratio	Lower	Upper
138	100		
92	48.3	29.2	69.2
108	67.7	49.1	89.1





#63
Azobenzene
Concen: 36.907 ng
RT: 10.645 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39


Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 89568

Ion	Ratio	Lower	Upper
77	100		
182	26.3	7.3	47.3
105	22.2	2.5	42.5
51	35.0	14.4	54.4

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.427 min Scan# 1570
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:188 Resp: 652685

Ion	Ratio	Lower	Upper
188	100		
94	10.5	8.2	12.2
80	11.1	8.6	12.8





#65
4,6-Dinitro-2-methylphenol
Concen: 45.858 ng
RT: 10.528 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39


Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:198 Resp: 9728

Ion	Ratio	Lower	Upper
198	100		
51	63.1	40.6	80.6
105	47.7	29.2	69.2

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#66
n-Nitrosodiphenylamine
Concen: 37.187 ng
RT: 10.604 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:169 Resp: 829293

Ion	Ratio	Lower	Upper
169	100		
168	66.1	53.0	79.6
167	36.0	28.8	43.2






#67
4-Bromophenyl-phenylether
Concen: 38.179 ng
RT: 10.975 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:248 Resp: 27751(


Ion	Ratio	Lower	Upper
248	100		
250	96.1	77.2	115.8
141	73.4	58.4	87.6



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#68
Hexachlorobenzene
Concen: 37.175 ng
RT: 11.039 min Scan# 1504
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:284 Resp: 302639

Ion	Ratio	Lower	Upper
284	100		
142	32.8	24.6	37.0
249	29.7	23.2	34.8






#69
Atrazine
Concen: 38.907 ng
RT: 11.128 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:200 Resp: 22870


Ion	Ratio	Lower	Upper
200	100		
173	27.5	8.6	48.6
215	49.1	28.5	68.5



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#70
Pentachlorophenol
Concen: 41.560 ng
RT: 11.227 min Scan# 1536
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:266 Resp: 178739

Ion	Ratio	Lower	Upper
266	100		
268	63.4	50.0	75.0
264	63.0	49.3	73.9






#71
Phenanthrene
Concen: 37.261 ng
RT: 11.457 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:178 Resp: 131366


Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.5	23.3
179	15.6	12.6	19.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#72
Anthracene
Concen: 37.446 ng
RT: 11.510 min Scan# 1584
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:178 Resp: 1349761

Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.3	22.9
179	15.5	12.7	19.1






#73
Carbazole
Concen: 37.893 ng
RT: 11.657 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:167 Resp: 124598


Ion	Ratio	Lower	Upper
167	100		
166	21.3	17.3	25.9
139	12.8	10.3	15.5



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#74
Di-n-butylphthalate
Concen: 37.366 ng
RT: 11.992 min Scan# 1666
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:149 Resp: 1235690

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.5	11.3
104	4.5	3.6	5.4





#75
Fluoranthene
Concen: 37.775 ng
RT: 12.645 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39


Tgt Ion:202 Resp: 131647


Ion	Ratio	Lower	Upper
202	100		
101	11.2	0.0	32.4
203	17.6	0.0	37.5

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.068 min Scan# 2019
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:240 Resp: 354752

Ion	Ratio	Lower	Upper
240	100		
120	12.8	10.1	15.1
236	25.7	20.5	30.7






#77
Benzidine
Concen: 41.743 ng
RT: 12.763 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:184 Resp: 62923


Ion	Ratio	Lower	Upper
184	100		
185	14.6	11.8	17.6
183	11.1	8.9	13.3



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#78
Pyrene
Concen: 40.557 ng
RT: 12.874 min Scan# 1816
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:202 Resp: 1327219

Ion	Ratio	Lower	Upper
202	100		
200	20.0	16.1	24.1
203	17.6	14.1	21.1






#79
Terphenyl-d14
Concen: 78.207 ng
RT: 13.016 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:244 Resp: 190759


Ion	Ratio	Lower	Upper
244	100		
212	6.7	5.3	7.9
122	11.4	9.1	13.7



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#80
Butylbenzylphthalate
Concen: 36.531 ng
RT: 13.492 min Scan# 1921
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:149 Resp: 318246

Ion	Ratio	Lower	Upper
149	100		
91	66.3	54.4	81.6
206	20.4	15.9	23.9





#81
Benzo(a)anthracene
Concen: 38.532 ng
RT: 14.057 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39


Tgt Ion:228 Resp: 91765


Ion	Ratio	Lower	Upper
228	100		
226	26.8	21.9	32.9
229	19.7	16.0	24.0

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#82
3,3'-Dichlorobenzidine
Concen: 37.840 ng
RT: 14.021 min Scan# 2011
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:252 Resp: 264696

Ion	Ratio	Lower	Upper
252	100		
254	63.1	51.0	76.6
126	12.1	9.4	14.2






#83
Chrysene
Concen: 36.304 ng
RT: 14.098 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:228 Resp: 78699


Ion	Ratio	Lower	Upper
228	100		
226	30.3	24.0	36.0
229	19.9	15.7	23.5



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 29.284 ng
RT: 14.051 min Scan# 2016
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:149 Resp: 344949

Ion	Ratio	Lower	Upper
149	100		
167	27.0	21.7	32.5
279	3.8	3.0	4.4





#85
Di-n-octyl phthalate
Concen: 28.225 ng
RT: 14.668 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39


Tgt Ion:149 Resp: 57620


Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.1	1.7
43	11.6	9.3	13.9

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.557 min Scan# 2272
Delta R.T. -0.012 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:264 Resp: 327562

Ion	Ratio	Lower	Upper
264	100		
260	23.5	18.4	27.6
265	21.8	17.3	25.9






#87
Indeno(1,2,3-cd)pyrene
Concen: 39.896 ng
RT: 17.068 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:276 Resp: 95723


Ion	Ratio	Lower	Upper
276	100		
138	28.0	22.6	33.8
277	24.9	20.1	30.1



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#88
Benzo(b)fluoranthene
Concen: 34.545 ng
RT: 15.121 min Scan# 2198
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:252 Resp: 702872

Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.7	26.5
125	10.5	9.0	13.4






#89
Benzo(k)fluoranthene
Concen: 38.642 ng
RT: 15.151 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:252 Resp: 71916


Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.8	26.8
125	10.5	9.0	13.6



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#90
Benzo(a)pyrene
Concen: 37.783 ng
RT: 15.498 min Scan# 2262
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:252 Resp: 703822

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.6	26.4
125	11.5	9.5	14.3






#91
Dibenzo(a,h)anthracene
Concen: 40.144 ng
RT: 17.086 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:278 Resp: 78171


Ion	Ratio	Lower	Upper
278	100		
139	18.2	15.2	22.8
279	23.9	19.0	28.6



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#92
Benzo(g,h,i)perylene
Concen: 40.703 ng
RT: 17.521 min Scan# 2606
Delta R.T. -0.018 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:276 Resp: 796796

Ion	Ratio	Lower	Upper
276	100		
277	23.5	18.7	28.1
138	24.5	19.4	29.0



Sample Results: **BF142274.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142274.D
Acq On : 02 May 2025 10:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 02 11:50:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2		1,4-Dioxane	0.559	0.526	5.9	93	-0.01
3		Pyridine	1.454	1.384	4.8	96	0.00
4		n-Nitrosodimethylamine	0.752	0.730	2.9	95	0.00
5	S	2-Fluorophenol	1.252	1.195	4.6	96	0.00
6		Aniline	2.140	2.027	5.3	95	0.00
7	S	Phenol-d6	1.512	1.457	3.6	96	0.00
8		2-Chlorophenol	1.294	1.268	2.0	97	0.00
9		Benzaldehyde	1.057	1.000	5.4	94	0.00
10	C	Phenol	1.628	1.545	5.1	95	0.00
11		bis(2-Chloroethyl)ether	1.262	1.194	5.4	95	0.00
12		1,3-Dichlorobenzene	1.464	1.384	5.5	95	0.00
13	C	1,4-Dichlorobenzene	1.475	1.364	7.5	93	0.00
14		1,2-Dichlorobenzene	1.408	1.334	5.3	95	0.00
15		Benzyl Alcohol	1.073	1.038	3.3	95	0.00
16		2,2'-oxybis(1-Chloropropane	2.278	2.158	5.3	95	0.00
17		2-Methylphenol	1.060	1.005	5.2	94	0.00
18		Hexachloroethane	0.483	0.473	2.1	97	0.00
19	P	n-Nitroso-di-n-propylamine	0.934	0.873	6.5	95	0.00
20		3+4-Methylphenols	1.326	1.295	2.3	97	0.00
21	I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
22		Acetophenone	0.463	0.421	9.1	95	0.00
23	S	Nitrobenzene-d5	0.304	0.332	-9.2	103	0.00
24		Nitrobenzene	0.299	0.315	-5.4	102	0.00
25		Isophorone	0.639	0.588	8.0	94	0.00
26	C	2-Nitrophenol	0.102	0.135	-32.4#	124	0.00
27		2,4-Dimethylphenol	0.322	0.307	4.7	96	0.00
28		bis(2-Chloroethoxy)methane	0.403	0.370	8.2	94	0.00
29	C	2,4-Dichlorophenol	0.269	0.261	3.0	95	0.00
30		1,2,4-Trichlorobenzene	0.301	0.278	7.6	94	0.00
31		Naphthalene	0.997	0.913	8.4	95	0.00
32		Benzoic acid	0.125	0.156	-24.8	127	0.00
33		4-Chloroaniline	0.415	0.388	6.5	96	0.00
34	C	Hexachlorobutadiene	0.175	0.164	6.3	94	0.00
35		Caprolactam	0.081	0.081	0.0	99	0.00
36	C	4-Chloro-3-methylphenol	0.277	0.266	4.0	95	0.00
37		2-Methylnaphthalene	0.613	0.560	8.6	95	0.00
38		1-Methylnaphthalene	0.635	0.583	8.2	95	0.00
39	I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
40		1,2,4,5-Tetrachlorobenzene	0.563	0.522	7.3	95	0.00
41	P	Hexachlorocyclopentadiene	0.319	0.327	-2.5	98	0.00
42	S	2,4,6-Tribromophenol	0.182	0.190	-4.4	98	0.00
43	C	2,4,6-Trichlorophenol	0.349	0.360	-3.2	102	0.00
44		2,4,5-Trichlorophenol	0.367	0.363	1.1	94	0.00
45	S	2-Fluorobiphenyl	1.418	1.293	8.8	96	0.00
46		1,1'-Biphenyl	1.555	1.438	7.5	95	0.00
47		2-Chloronaphthalene	1.157	1.071	7.4	95	0.00

Sample Results: **BF142274.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142274.D
Acq On : 02 May 2025 10:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 02 11:50:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.249	0.315	-26.5#	112	0.00
49		Acenaphthylene	1.952	1.809	7.3	95	0.00
50		Dimethylphthalate	1.280	1.197	6.5	95	0.00
51		2,6-Dinitrotoluene	0.204	0.246	-20.6	107	0.00
52	C	Acenaphthene	1.140	1.070	6.1	96	0.00
53		3-Nitroaniline	0.257	0.298	-16.0	105	0.00
54	P	2,4-Dinitrophenol	0.058	0.086	-48.3#	158#	0.00
55		Dibenzofuran	1.709	1.581	7.5	95	0.00
56	P	4-Nitrophenol	0.201	0.227	-12.9	110	0.00
57		2,4-Dinitrotoluene	0.239	0.314	-31.4#	116	0.00
58		Fluorene	1.299	1.197	7.9	95	0.00
59		2,3,4,6-Tetrachlorophenol	0.304	0.309	-1.6	97	0.00
60		Diethylphthalate	1.263	1.173	7.1	94	0.00
61		4-Chlorophenyl-phenylether	0.617	0.567	8.1	94	0.00
62		4-Nitroaniline	0.238	0.283	-18.9	109	0.00
63		Azobenzene	1.248	1.151	7.8	94	0.00
64	I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
65		4,6-Dinitro-2-methylphenol	0.054	0.075	-38.9#	144	0.00
66	c	n-Nitrosodiphenylamine	0.683	0.635	7.0	95	0.00
67		4-Bromophenyl-phenylether	0.223	0.213	4.5	95	0.00
68		Hexachlorobenzene	0.249	0.232	6.8	94	0.00
69		Atrazine	0.180	0.175	2.8	97	0.00
70	C	Pentachlorophenol	0.132	0.137	-3.8	101	0.00
71		Phenanthrene	1.080	1.006	6.9	96	0.00
72		Anthracene	1.105	1.034	6.4	97	0.00
73		Carbazole	1.008	0.955	5.3	98	0.00
74		Di-n-butylphthalate	1.013	0.947	6.5	94	0.00
75	C	Fluoranthene	1.068	1.009	5.5	98	0.00
76	I	Chrysene-d12	1.000	1.000	0.0	107	0.00
77		Benzidine	0.850	0.887	-4.4	99	0.00
78		Pyrene	1.845	1.871	-1.4	99	0.00
79	S	Terphenyl-d14	1.375	1.344	2.3	97	0.00
80		Butylbenzylphthalate	0.425	0.449	-5.6	95	0.00
81		Benzo(a)anthracene	1.343	1.293	3.7	96	0.00
82		3,3'-Dichlorobenzidine	0.394	0.373	5.3	90	0.00
83		Chrysene	1.222	1.109	9.2	96	0.00
84		Bis(2-ethylhexyl)phthalate	0.604	0.486	19.5	75	0.00
85	c	Di-n-octyl phthalate	1.133	0.812	28.3#	69	-0.01
86	I	Perylene-d12	1.000	1.000	0.0	100	-0.01
87		Indeno(1,2,3-cd)pyrene	1.465	1.461	0.3	90	-0.01
88		Benzo(b)fluoranthene	1.242	1.073	13.6	82	0.00
89		Benzo(k)fluoranthene	1.136	1.098	3.3	96	0.00
90	C	Benzo(a)pyrene	1.137	1.074	5.5	89	0.00
91		Dibenzo(a,h)anthracene	1.189	1.193	-0.3	91	-0.01
92		Benzo(g,h,i)perylene	1.195	1.216	-1.8	93	-0.02

Sample Results: **BF142274.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142274.D
Acq On : 02 May 2025 10:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 02 11:50:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 2

Sample Results: **BF142274.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142274.D
Acq On : 02 May 2025 10:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 02 11:50:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	105	0.00
2		1,4-Dioxane	40.000	37.589	6.0	93	-0.01
3		Pyridine	40.000	38.083	4.8	96	0.00
4		n-Nitrosodimethylamine	40.000	38.811	3.0	95	0.00
5	S	2-Fluorophenol	80.000	76.366	4.5	96	0.00
6		Aniline	40.000	37.880	5.3	95	0.00
7	S	Phenol-d6	80.000	77.087	3.6	96	0.00
8		2-Chlorophenol	40.000	39.206	2.0	97	0.00
9		Benzaldehyde	40.000	37.835	5.4	94	0.00
10	C	Phenol	40.000	37.967	5.1	95	0.00
11		bis(2-Chloroethyl)ether	40.000	37.841	5.4	95	0.00
12		1,3-Dichlorobenzene	40.000	37.811	5.5	95	0.00
13	C	1,4-Dichlorobenzene	40.000	36.999	7.5	93	0.00
14		1,2-Dichlorobenzene	40.000	37.888	5.3	95	0.00
15		Benzyl Alcohol	40.000	38.697	3.3	95	0.00
16		2,2'-oxybis(1-Chloropropane	40.000	37.895	5.3	95	0.00
17		2-Methylphenol	40.000	37.931	5.2	94	0.00
18		Hexachloroethane	40.000	39.190	2.0	97	0.00
19	P	n-Nitroso-di-n-propylamine	40.000	37.373	6.6	95	0.00
20		3+4-Methylphenols	40.000	39.062	2.3	97	0.00
21	I	Naphthalene-d8	20.000	20.000	0.0	106	0.00
22		Acetophenone	40.000	36.381	9.0	95	0.00
23	S	Nitrobenzene-d5	80.000	87.247	-9.1	103	0.00
24		Nitrobenzene	40.000	42.143	-5.4	102	0.00
25		Isophorone	40.000	36.800	8.0	94	0.00
26	C	2-Nitrophenol	40.000	43.995	-10.0	124	0.00
27		2,4-Dimethylphenol	40.000	38.087	4.8	96	0.00
28		bis(2-Chloroethoxy)methane	40.000	36.730	8.2	94	0.00
29	C	2,4-Dichlorophenol	40.000	38.714	3.2	95	0.00
30		1,2,4-Trichlorobenzene	40.000	36.954	7.6	94	0.00
31		Naphthalene	40.000	36.602	8.5	95	0.00
32		Benzoic acid	40.000	42.364	-5.9	127	0.00
33		4-Chloroaniline	40.000	37.373	6.6	96	0.00
34	C	Hexachlorobutadiene	40.000	37.512	6.2	94	0.00
35		Caprolactam	40.000	39.988	0.0	99	0.00
36	C	4-Chloro-3-methylphenol	40.000	38.463	3.8	95	0.00
37		2-Methylnaphthalene	40.000	36.551	8.6	95	0.00
38		1-Methylnaphthalene	40.000	36.702	8.2	95	0.00
39	I	Acenaphthene-d10	20.000	20.000	0.0	105	0.00
40		1,2,4,5-Tetrachlorobenzene	40.000	37.065	7.3	95	0.00
41	P	Hexachlorocyclopentadiene	40.000	40.985	-2.5	98	0.00
42	S	2,4,6-Tribromophenol	80.000	83.329	-4.2	98	0.00
43	C	2,4,6-Trichlorophenol	40.000	41.332	-3.3	102	0.00
44		2,4,5-Trichlorophenol	40.000	39.611	1.0	94	0.00
45	S	2-Fluorobiphenyl	80.000	72.935	8.8	96	0.00
46		1,1'-Biphenyl	40.000	36.988	7.5	95	0.00
47		2-Chloronaphthalene	40.000	37.021	7.4	95	0.00

Sample Results: **BF142274.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142274.D
Acq On : 02 May 2025 10:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 02 11:50:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	42.014	-5.0	112	0.00
49		Acenaphthylene	40.000	37.062	7.3	95	0.00
50		Dimethylphthalate	40.000	37.413	6.5	95	0.00
51		2,6-Dinitrotoluene	40.000	40.686	-1.7	107	0.00
52	C	Acenaphthene	40.000	37.558	6.1	96	0.00
53		3-Nitroaniline	40.000	40.499	-1.2	105	0.00
54	P	2,4-Dinitrophenol	40.000	50.599	-26.5#	158	0.00
55		Dibenzofuran	40.000	37.015	7.5	95	0.00
56	P	4-Nitrophenol	40.000	45.101	-12.8	110	0.00
57		2,4-Dinitrotoluene	40.000	42.928	-7.3	116	0.00
58		Fluorene	40.000	36.852	7.9	95	0.00
59		2,3,4,6-Tetrachlorophenol	40.000	40.702	-1.8	97	0.00
60		Diethylphthalate	40.000	37.148	7.1	94	0.00
61		4-Chlorophenyl-phenylether	40.000	36.772	8.1	94	0.00
62		4-Nitroaniline	40.000	42.072	-5.2	109	0.00
63		Azobenzene	40.000	36.907	7.7	94	0.00
64	I	Phenanthrene-d10	20.000	20.000	0.0	106	0.00
65		4,6-Dinitro-2-methylphenol	40.000	45.858	-14.6	144	0.00
66	c	n-Nitrosodiphenylamine	40.000	37.187	7.0	95	0.00
67		4-Bromophenyl-phenylether	40.000	38.179	4.6	95	0.00
68		Hexachlorobenzene	40.000	37.175	7.1	94	0.00
69		Atrazine	40.000	38.907	2.7	97	0.00
70	C	Pentachlorophenol	40.000	41.560	-3.9	101	0.00
71		Phenanthrene	40.000	37.261	6.8	96	0.00
72		Anthracene	40.000	37.446	6.4	97	0.00
73		Carbazole	40.000	37.893	5.3	98	0.00
74		Di-n-butylphthalate	40.000	37.366	6.6	94	0.00
75	C	Fluoranthene	40.000	37.775	5.6	98	0.00
76	I	Chrysene-d12	20.000	20.000	0.0	107	0.00
77		Benzidine	40.000	41.743	-4.4	99	0.00
78		Pyrene	40.000	40.557	-1.4	99	0.00
79	S	Terphenyl-d14	80.000	78.207	2.2	97	0.00
80		Butylbenzylphthalate	40.000	36.531	8.7	95	0.00
81		Benzo(a)anthracene	40.000	38.532	3.7	96	0.00
82		3,3'-Dichlorobenzidine	40.000	37.840	5.4	90	0.00
83		Chrysene	40.000	36.304	9.2	96	0.00
84		Bis(2-ethylhexyl)phthalate	40.000	29.284	26.8#	75	0.00
85	c	Di-n-octyl phthalate	40.000	28.225	29.4#	69	-0.01
86	I	Perylene-d12	20.000	20.000	0.0	100	-0.01
87		Indeno(1,2,3-cd)pyrene	40.000	39.896	0.3	90	-0.01
88		Benzo(b)fluoranthene	40.000	34.545	13.6	82	0.00
89		Benzo(k)fluoranthene	40.000	38.642	3.4	96	0.00
90	C	Benzo(a)pyrene	40.000	37.783	5.5	89	0.00
91		Dibenzo(a,h)anthracene	40.000	40.144	-0.4	91	-0.01
92		Benzo(g,h,i)perylene	40.000	40.703	-1.8	93	-0.02

Sample Results: **BF142274.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142274.D
Acq On : 02 May 2025 10:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 02 11:50:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 1

Raw Data: **BP024522.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG No.: Q1901

Instrument ID: BNA_P Calibration Date/Time: 05/05/2025 12:38

Lab File ID: BP024522.D Init. Calib. Date(s): 04/14/2025 04/14/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:06 17:13

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.351	1.195		-11.5	
2-Fluorophenol	1.210	1.171		-3.2	
Phenol-d6	1.656	1.555		-6.1	
1,4-Dichlorobenzene	1.475	1.397		-5.3	20.0
2-Methylphenol	1.075	1.030		-4.2	
3+4-Methylphenols	1.491	1.423		-4.6	
Nitrobenzene-d5	0.351	0.347		-1.1	
Hexachloroethane	0.533	0.505		-5.3	
Nitrobenzene	0.347	0.332		-4.3	
Hexachlorobutadiene	0.183	0.193		5.5	20.0
2,4,6-Trichlorophenol	0.366	0.371		1.4	20.0
2-Fluorobiphenyl	1.306	1.264		-3.2	
2,4,5-Trichlorophenol	0.405	0.407		0.5	
2,4-Dinitrotoluene	0.421	0.437		3.8	
2,4,6-Tribromophenol	0.277	0.323		16.6	
Hexachlorobenzene	0.260	0.274		5.4	
Pentachlorophenol	0.181	0.195		7.7	20.0
Terphenyl-d14	0.992	0.960		-3.2	

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050525\
Data File : BP024522.D
Acq On : 05 May 2025 12:38
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Quant Time: May 05 13:05:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.705	152	132437	20.000	ng		-0.02
21) Naphthalene-d8	10.475	136	539349	20.000	ng		-0.02
39) Acenaphthene-d10	14.340	164	358648	20.000	ng		-0.01
64) Phenanthrene-d10	17.151	188	715330	20.000	ng		0.00
76) Chrysene-d12	21.604	240	882732	20.000	ng	#	0.01
86) Perylene-d12	24.927	264	1072388	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.328	112	620420	77.458	ng		-0.02
7) Phenol-d6	6.899	99	823894	75.121	ng		-0.01
23) Nitrobenzene-d5	8.852	82	747857	79.065	ng		-0.02
42) 2,4,6-Tribromophenol	15.869	330	462897	93.287	ng		0.00
45) 2-Fluorobiphenyl	12.951	172	1813216	77.408	ng		-0.01
79) Terphenyl-d14	19.875	244	3389636	77.399	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.276	88	121212	35.777	ng		96
3) Pyridine	3.676	79	316441	35.372	ng		94
4) n-Nitrosodimethylamine	3.581	42	126596	42.639	ng		78
6) Aniline	7.046	93	366900	37.437	ng		98
8) 2-Chlorophenol	7.281	128	348215	38.938	ng		98
9) Benzaldehyde	6.864	77	227897	42.007	ng		95
10) Phenol	6.922	94	403296	36.656	ng		92
11) bis(2-Chloroethyl)ether	7.140	93	306380	34.563	ng		100
12) 1,3-Dichlorobenzene	7.599	146	367864	38.210	ng		99
13) 1,4-Dichlorobenzene	7.740	146	369994	37.877	ng		100
14) 1,2-Dichlorobenzene	8.058	146	361092	38.282	ng		99
15) Benzyl Alcohol	7.946	79	285104	40.197	ng		97
16) 2,2'-oxybis(1-Chloropr...	8.234	45	331133	34.559	ng		99
17) 2-Methylphenol	8.152	107	272914	38.345	ng		98
18) Hexachloroethane	8.769	117	133738	37.884	ng		96
19) n-Nitroso-di-n-propyla...	8.505	70	248576	36.635	ng		95
20) 3+4-Methylphenols	8.475	107	377025	38.178	ng		97
22) Acetophenone	8.522	105	525529	39.368	ng	#	98
24) Nitrobenzene	8.899	77	358077	38.268	ng		97
25) Isophorone	9.416	82	655624	38.294	ng		98
26) 2-Nitrophenol	9.599	139	196984	43.018	ng		97
27) 2,4-Dimethylphenol	9.663	122	228268	39.701	ng		99
28) bis(2-Chloroethoxy)met...	9.893	93	411315	35.563	ng		99
29) 2,4-Dichlorophenol	10.140	162	320136	40.835	ng		98
30) 1,2,4-Trichlorobenzene	10.340	180	332331	39.565	ng		99
31) Naphthalene	10.528	128	1096383	38.590	ng		99
32) Benzoic acid	9.816	122	269510	40.228	ng		97
33) 4-Chloroaniline	10.640	127	403971	40.376	ng		99
34) Hexachlorobutadiene	10.810	225	207808	42.102	ng		99
35) Caprolactam	11.434	113	123328	42.634	ng		97
36) 4-Chloro-3-methylphenol	11.775	107	379400	41.549	ng		99
37) 2-Methylnaphthalene	12.140	142	768396	38.998	ng		98
38) 1-Methylnaphthalene	12.363	142	749001	38.858	ng		100
40) 1,2,4,5-Tetrachloroben...	12.510	216	386220	39.159	ng		100
41) Hexachlorocyclopentadiene	12.487	237	126536	36.275	ng		100
43) 2,4,6-Trichlorophenol	12.757	196	265918	40.552	ng		99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050525\
Data File : BP024522.D
Acq On : 05 May 2025 12:38
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Quant Time: May 05 13:05:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	12.840	196	291771	40.181	ng		99
46) 1,1'-Biphenyl	13.157	154	996822	37.812	ng		98
47) 2-Chloronaphthalene	13.204	162	751741	38.154	ng		99
48) 2-Nitroaniline	13.410	65	232134	42.002	ng		92
49) Acenaphthylene	14.057	152	1209138	38.978	ng		100
50) Dimethylphthalate	13.793	163	1025676	39.326	ng		100
51) 2,6-Dinitrotoluene	13.916	165	223409	40.338	ng		96
52) Acenaphthene	14.404	154	745486	37.907	ng		98
53) 3-Nitroaniline	14.257	138	238177	40.918	ng		99
54) 2,4-Dinitrophenol	14.469	184	135885	41.652	ng		94
55) Dibenzofuran	14.746	168	1215274	37.805	ng		97
56) 4-Nitrophenol	14.598	139	204781	40.734	ng		87
57) 2,4-Dinitrotoluene	14.716	165	313164	41.451	ng		99
58) Fluorene	15.404	166	957243	38.466	ng		100
59) 2,3,4,6-Tetrachlorophenol	14.987	232	270913	40.451	ng		99
60) Diethylphthalate	15.187	149	1061854	39.507	ng		100
61) 4-Chlorophenyl-phenyle...	15.404	204	472625	39.111	ng		99
62) 4-Nitroaniline	15.440	138	264096	42.858	ng		88
63) Azobenzene	15.704	77	909080	36.799	ng		96
65) 4,6-Dinitro-2-methylph...	15.504	198	186357	41.322	ng		95
66) n-Nitrosodiphenylamine	15.628	169	817880	38.306	ng		99
67) 4-Bromophenyl-phenylether	16.322	248	317430	40.574	ng		97
68) Hexachlorobenzene	16.440	284	391978	42.165	ng		99
69) Atrazine	16.616	200	216677	39.846	ng		97
70) Pentachlorophenol	16.798	266	279598	43.112	ng		99
71) Phenanthrene	17.198	178	1501631	38.480	ng		99
72) Anthracene	17.292	178	1493914	39.721	ng		99
73) Carbazole	17.575	167	1492224	40.606	ng		99
74) Di-n-butylphthalate	18.169	149	1940609	42.387	ng		100
75) Fluoranthene	19.286	202	1924412	41.464	ng		96
77) Benzidine	19.481	184	658085	58.813	ng		99
78) Pyrene	19.663	202	2017997	35.972	ng		99
80) Butylbenzylphthalate	20.616	149	962897	40.073	ng		94
81) Benzo(a)anthracene	21.580	228	2118786	38.617	ng		99
82) 3,3'-Dichlorobenzidine	21.504	252	834874	43.353	ng		99
83) Chrysene	21.645	228	2034193	38.698	ng		99
84) Bis(2-ethylhexyl)phtha...	21.510	149	1436161	40.771	ng		100
85) Di-n-octyl phthalate	22.786	149	2491117	43.501	ng		100
87) Indeno(1,2,3-cd)pyrene	28.721	276	2911665	39.109	ng	#	96
88) Benzo(b)fluoranthene	23.892	252	2425230	37.729	ng	#	97
89) Benzo(k)fluoranthene	23.957	252	2349778	37.844	ng	#	97
90) Benzo(a)pyrene	24.774	252	2146872	38.719	ng		98
91) Dibenzo(a,h)anthracene	28.786	278	2420922	39.176	ng		97
92) Benzo(g,h,i)perylene	29.909	276	2470173	39.272	ng	#	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050525\
Data File : BP024522.D
Acq On : 05 May 2025 12:38
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Quant Time: May 05 13:05:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.705 min Scan# 7
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:152 Resp: 132437

Ion	Ratio	Lower	Upper
152	100		
150	159.3	124.9	187.3
115	62.1	47.7	71.5







#2
1,4-Dioxane
Concen: 35.777 ng
RT: 3.276 min Scan# 32
Delta R.T. -0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion: 88 Resp: 121212

Ion	Ratio	Lower	Upper
88	100		
58	66.8	51.4	77.2
43	24.8	17.1	25.7







#3
Pyridine
Concen: 35.372 ng
RT: 3.676 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 316441

Ion	Ratio	Lower	Upper
79	100		
52	60.8	45.4	68.2
51	32.5	22.6	33.8







#4
n-Nitrosodimethylamine
Concen: 42.639 ng
RT: 3.581 min Scan# 84
Delta R.T. -0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion: 42 Resp: 126596

Ion	Ratio	Lower	Upper
42	100		
74	143.0	139.5	209.3
44	6.9	5.8	8.8







#5
2-Fluorophenol
Concen: 77.458 ng
RT: 5.328 min Scan# 3
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:112 Resp: 620420

Ion	Ratio	Lower	Upper
112	100		
64	54.5	41.1	61.7
63	29.1	20.6	30.8







#6
Aniline
Concen: 37.437 ng
RT: 7.046 min Scan# 673
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion: 93 Resp: 366900

Ion	Ratio	Lower	Upper
93	100		
66	37.5	29.1	43.7
65	19.2	14.7	22.1







#7
Phenol-d6
Concen: 75.121 ng
RT: 6.899 min Scan# 64
Delta R.T. -0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 99 Resp: 823894

Ion	Ratio	Lower	Upper
99	100		
42	15.7	10.6	16.0
71	33.8	23.7	35.5







#8
2-Chlorophenol
Concen: 38.938 ng
RT: 7.281 min Scan# 713
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:128 Resp: 348215

Ion	Ratio	Lower	Upper
128	100		
130	31.7	12.9	52.9
64	41.8	22.7	62.7







#9
Benzaldehyde
Concen: 42.007 ng
RT: 6.864 min Scan# 64
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 227897

Ion	Ratio	Lower	Upper
77	100		
105	103.6	79.2	119.2
106	101.1	76.2	116.2







#10
Phenol
Concen: 36.656 ng
RT: 6.922 min Scan# 652
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion: 94 Resp: 403296

Ion	Ratio	Lower	Upper
94	100		
65	27.4	4.5	44.5
66	38.1	12.4	52.4







#11
bis(2-Chloroethyl)ether
Concen: 34.563 ng
RT: 7.140 min Scan# 6
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 93 Resp: 306380

Ion	Ratio	Lower	Upper
93	100		
63	68.5	48.6	88.6
95	32.6	12.3	52.3







#12
1,3-Dichlorobenzene
Concen: 38.210 ng
RT: 7.599 min Scan# 767
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:146 Resp: 367864

Ion	Ratio	Lower	Upper
146	100		
148	65.1	51.4	77.0
75	28.9	22.9	34.3







#13
1,4-Dichlorobenzene
Concen: 37.877 ng
RT: 7.740 min Scan# 7
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:146 Resp: 369994

Ion	Ratio	Lower	Upper
146	100		
148	63.9	51.1	76.7
111	41.1	32.6	48.8







#14
1,2-Dichlorobenzene
Concen: 38.282 ng
RT: 8.058 min Scan# 845
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:146 Resp: 361092

Ion	Ratio	Lower	Upper
146	100		
148	64.7	51.0	76.4
111	42.7	33.8	50.8







#15
Benzyl Alcohol
Concen: 40.197 ng
RT: 7.946 min Scan# 8
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 285104

Ion	Ratio	Lower	Upper
79	100		
108	77.5	65.1	97.7
77	63.1	50.7	76.1







#16
2,2'-oxybis(1-Chloropropane)
Concen: 34.559 ng
RT: 8.234 min Scan# 875
Delta R.T. -0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion: 45 Resp: 331133

Ion	Ratio	Lower	Upper
45	100		
77	17.1	0.0	37.7
79	13.6	0.0	33.5







#17
2-Methylphenol
Concen: 38.345 ng
RT: 8.152 min Scan# 8
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:107 Resp: 272914

Ion	Ratio	Lower	Upper
107	100		
108	111.4	89.8	134.6
77	45.7	34.2	51.4
79	45.3	33.8	50.8







#18
Hexachloroethane
Concen: 37.884 ng
RT: 8.769 min Scan# 966
Delta R.T. -0.023 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:117 Resp: 133738

Ion	Ratio	Lower	Upper
117	100		
119	95.3	77.8	116.6
201	100.7	75.9	113.9







#19
n-Nitroso-di-n-propylamine
Concen: 36.635 ng
RT: 8.505 min Scan# 9
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 70 Resp: 248576

Ion	Ratio	Lower	Upper
70	100		
42	51.2	37.3	55.9
101	10.6	8.2	12.4
130	24.7	18.6	28.0







#20
3+4-Methylphenols
Concen: 38.178 ng
RT: 8.475 min Scan# 916
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:107 Resp: 377025

Ion	Ratio	Lower	Upper
107	100		
108	86.7	69.8	109.8
77	32.9	11.5	51.5
79	28.1	7.8	47.8







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.475 min Scan# 1
Delta R.T. -0.023 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:136 Resp: 539349

Ion	Ratio	Lower	Upper
136	100		
137	10.6	9.2	13.8
54	6.9	6.0	9.0
68	5.6	5.5	8.3







#22
Acetophenone
Concen: 39.368 ng
RT: 8.522 min Scan# 924
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:105 Resp: 525529

Ion	Ratio	Lower	Upper
105	100		
71	1.5	1.7	2.5#
51	22.7	16.9	25.3
120	23.0	18.6	28.0







#23
Nitrobenzene-d5
Concen: 79.065 ng
RT: 8.852 min Scan# 9
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 747857

Ion	Ratio	Lower	Upper
82	100		
128	46.4	36.5	54.7
54	46.0	33.4	50.2







#24
Nitrobenzene
Concen: 38.268 ng
RT: 8.899 min Scan# 988
Delta R.T. -0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion: 77 Resp: 358077

Ion	Ratio	Lower	Upper
77	100		
123	51.1	39.2	58.8
65	13.2	9.8	14.6







#25
Isophorone
Concen: 38.294 ng
RT: 9.416 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 655624

Ion	Ratio	Lower	Upper
82	100		
95	7.7	5.4	8.2
138	18.9	14.4	21.6







#26
2-Nitrophenol
Concen: 43.018 ng
RT: 9.599 min Scan# 1107
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:139 Resp: 196984

Ion	Ratio	Lower	Upper
139	100		
109	30.4	21.8	32.6
65	46.2	36.2	54.2







#27
2,4-Dimethylphenol
Concen: 39.701 ng
RT: 9.663 min Scan# 11
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:122 Resp: 228268

Ion	Ratio	Lower	Upper
122	100		
107	116.9	93.7	140.5
121	58.7	48.2	72.2







#28
bis(2-Chloroethoxy)methane
Concen: 35.563 ng
RT: 9.893 min Scan# 1157
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion: 93 Resp: 411315

Ion	Ratio	Lower	Upper
93	100		
95	32.2	26.0	39.0
123	12.4	9.1	13.7







#29
2,4-Dichlorophenol
Concen: 40.835 ng
RT: 10.140 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:162 Resp: 320136

Ion	Ratio	Lower	Upper
162	100		
164	65.0	43.3	83.3
98	37.2	16.6	56.6







#30
1,2,4-Trichlorobenzene
Concen: 39.565 ng
RT: 10.340 min Scan# 1233
Delta R.T. -0.023 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:180 Resp: 332331

Ion	Ratio	Lower	Upper
180	100		
182	96.0	77.3	115.9
145	32.7	25.1	37.7







#31
Naphthalene
Concen: 38.590 ng
RT: 10.528 min Scan# 1
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:128 Resp: 1096383

Ion	Ratio	Lower	Upper
128	100		
129	10.7	8.6	13.0
127	13.3	10.3	15.5







#32
Benzoic acid
Concen: 40.228 ng
RT: 9.816 min Scan# 1144
Delta R.T. -0.006 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:122 Resp: 269510

Ion	Ratio	Lower	Upper
122	100		
105	127.6	104.8	144.8
77	92.6	69.2	109.2







#33
4-Chloroaniline
Concen: 40.376 ng
RT: 10.640 min Scan# 1
Delta R.T. -0.023 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:127 Resp: 403971

Ion	Ratio	Lower	Upper
127	100		
129	32.3	26.1	39.1
65	28.8	22.1	33.1
92	19.5	15.0	22.4







#34
Hexachlorobutadiene
Concen: 42.102 ng
RT: 10.810 min Scan# 1313
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:225 Resp: 207808

Ion	Ratio	Lower	Upper
225	100		
223	62.8	49.5	74.3
227	64.4	50.9	76.3







#35
Caprolactam
Concen: 42.634 ng
RT: 11.434 min Scan# 14
Delta R.T. -0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:113 Resp: 123328

Ion	Ratio	Lower	Upper
113	100		
55	137.8	115.9	155.9
56	106.8	81.1	121.1







#36
4-Chloro-3-methylphenol
Concen: 41.549 ng
RT: 11.775 min Scan# 1477
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:107 Resp: 379400

Ion	Ratio	Lower	Upper
107	100		
144	25.5	20.5	30.7
142	78.7	63.4	95.2







#37
2-Methylnaphthalene
Concen: 38.998 ng
RT: 12.140 min Scan# 1
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:142 Resp: 768396

Ion	Ratio	Lower	Upper
142	100		
141	89.0	70.3	105.5
115	36.2	26.9	40.3







#38
1-Methylnaphthalene
Concen: 38.858 ng
RT: 12.363 min Scan# 1577
Delta R.T. -0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:142 Resp: 749001

Ion	Ratio	Lower	Upper
142	100		
141	92.0	73.6	110.4
116	3.9	3.0	4.6







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.340 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:164 Resp: 358648

Ion	Ratio	Lower	Upper
164	100		
162	103.2	80.6	120.8
160	44.3	35.3	52.9







#40
1,2,4,5-Tetrachlorobenzene
Concen: 39.159 ng
RT: 12.510 min Scan# 1602
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:216 Resp: 386220

Ion	Ratio	Lower	Upper
216	100		
214	77.9	62.6	93.8
179	21.9	17.2	25.8
108	18.6	15.0	22.6







#41
Hexachlorocyclopentadiene
Concen: 36.275 ng
RT: 12.487 min Scan# 1
Delta R.T. -0.023 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:237 Resp: 126536

Ion	Ratio	Lower	Upper
237	100		
235	63.9	43.9	83.9
272	13.1	0.0	33.6







#42
2,4,6-Tribromophenol
Concen: 93.287 ng
RT: 15.869 min Scan# 2173
Delta R.T. 0.006 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:330 Resp: 462897

Ion	Ratio	Lower	Upper
330	100		
332	97.1	77.1	115.7
141	28.3	24.7	37.1







#43
2,4,6-Trichlorophenol
Concen: 40.552 ng
RT: 12.757 min Scan# 1
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:196 Resp: 265918

Ion	Ratio	Lower	Upper
196	100		
198	94.3	74.3	111.5
200	30.3	24.0	36.0







#44
2,4,5-Trichlorophenol
Concen: 40.181 ng
RT: 12.840 min Scan# 1658
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:196 Resp: 291771

Ion	Ratio	Lower	Upper
196	100		
198	96.0	77.3	115.9
97	50.0	40.4	60.6
132	29.4	23.4	35.0







#45
2-Fluorobiphenyl
Concen: 77.408 ng
RT: 12.951 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:172 Resp: 1813216

Ion	Ratio	Lower	Upper
172	100		
171	35.8	28.2	42.2
170	23.4	18.6	27.8







#46
1,1'-Biphenyl
Concen: 37.812 ng
RT: 13.157 min Scan# 1712
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:154 Resp: 996822

Ion	Ratio	Lower	Upper
154	100		
153	41.5	20.3	60.3
76	11.9	0.0	33.5







#47
2-Chloronaphthalene
Concen: 38.154 ng
RT: 13.204 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:162 Resp: 751741

Ion	Ratio	Lower	Upper
162	100		
127	41.2	31.9	47.9
164	32.4	26.2	39.4







#48
2-Nitroaniline
Concen: 42.002 ng
RT: 13.410 min Scan# 1755
Delta R.T. -0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion: 65 Resp: 232134

Ion	Ratio	Lower	Upper
65	100		
92	70.6	62.3	93.5
138	113.3	97.2	145.8







#49
Acenaphthylene
Concen: 38.978 ng
RT: 14.057 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:152 Resp: 1209138

Ion	Ratio	Lower	Upper
152	100		
151	20.0	15.9	23.9
153	12.9	10.6	15.8







#50
Dimethylphthalate
Concen: 39.326 ng
RT: 13.793 min Scan# 1820
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:163 Resp: 1025676

Ion	Ratio	Lower	Upper
163	100		
194	4.2	3.5	5.3
164	10.1	8.0	12.0







#51
2,6-Dinitrotoluene
Concen: 40.338 ng
RT: 13.916 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 223409

Ion	Ratio	Lower	Upper
165	100		
63	51.1	38.0	57.0
89	54.5	41.7	62.5







#52
Acenaphthene
Concen: 37.907 ng
RT: 14.404 min Scan# 1924
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:154 Resp: 745486

Ion	Ratio	Lower	Upper
154	100		
153	113.7	89.8	134.6
152	55.0	42.6	63.8







#53
3-Nitroaniline
Concen: 40.918 ng
RT: 14.257 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:138 Resp: 238177

Ion	Ratio	Lower	Upper
138	100		
108	11.3	8.3	12.5
92	110.9	89.8	134.8







#54
2,4-Dinitrophenol
Concen: 41.652 ng
RT: 14.469 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:184 Resp: 135885

Ion	Ratio	Lower	Upper
184	100		
63	56.5	42.6	64.0
154	63.8	47.2	70.8







#55
Dibenzofuran
Concen: 37.805 ng
RT: 14.746 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:168 Resp: 1215274

Ion	Ratio	Lower	Upper
168	100		
139	40.5	30.3	45.5
169	13.3	10.6	16.0







#56
4-Nitrophenol
Concen: 40.734 ng
RT: 14.598 min Scan# 1957
Delta R.T. 0.006 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:139 Resp: 204781

Ion	Ratio	Lower	Upper
139	100		
109	80.5	46.1	86.1
65	92.7	65.0	105.0







#57
2,4-Dinitrotoluene
Concen: 41.451 ng
RT: 14.716 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 313164

Ion	Ratio	Lower	Upper
165	100		
63	39.5	30.4	45.6
89	68.1	54.6	82.0
182	2.8	2.2	3.2







#58
Fluorene
Concen: 38.466 ng
RT: 15.404 min Scan# 2094
Delta R.T. -0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:166 Resp: 957243

Ion	Ratio	Lower	Upper
166	100		
165	97.8	78.1	117.1
167	13.5	10.9	16.3







#59
2,3,4,6-Tetrachlorophenol
Concen: 40.451 ng
RT: 14.987 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:232 Resp: 270913

Ion	Ratio	Lower	Upper
232	100		
131	47.1	36.8	55.2
130	2.5	1.9	2.9
166	30.9	25.0	37.6







#60
Diethylphthalate
Concen: 39.507 ng
RT: 15.187 min Scan# 2057
Delta R.T. -0.006 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:149 Resp: 1061854

Ion	Ratio	Lower	Upper
149	100		
177	21.8	17.4	26.2
150	12.5	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 39.111 ng
RT: 15.404 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:204 Resp: 472625

Ion	Ratio	Lower	Upper
204	100		
206	32.4	26.5	39.7
141	62.4	49.4	74.0







#62
4-Nitroaniline
Concen: 42.858 ng
RT: 15.440 min Scan# 2100
Delta R.T. 0.000 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:138 Resp: 264096

Ion	Ratio	Lower	Upper
138	100		
92	50.3	27.8	67.8
108	83.0	48.9	88.9







#63
Azobenzene
Concen: 36.799 ng
RT: 15.704 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 909080

Ion	Ratio	Lower	Upper
77	100		
182	26.5	7.3	47.3
105	22.1	2.3	42.3
51	28.4	3.4	43.4







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.151 min Scan# 2391
Delta R.T. 0.006 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:188 Resp: 715330

Ion	Ratio	Lower	Upper
188	100		
94	8.8	8.6	13.0
80	10.0	9.8	14.6







#65
4,6-Dinitro-2-methylphenol
Concen: 41.322 ng
RT: 15.504 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:198 Resp: 186357

Ion	Ratio	Lower	Upper
198	100		
51	40.2	18.4	58.4
105	42.6	26.5	66.5







#66
n-Nitrosodiphenylamine
Concen: 38.306 ng
RT: 15.628 min Scan# 2132
Delta R.T. 0.000 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:169 Resp: 817880

Ion	Ratio	Lower	Upper
169	100		
168	67.5	53.0	79.4
167	35.8	28.9	43.3







#67
4-Bromophenyl-phenylether
Concen: 40.574 ng
RT: 16.322 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:248 Resp: 317430

Ion	Ratio	Lower	Upper
248	100		
250	98.4	78.2	117.4
141	76.6	57.7	86.5







#68
Hexachlorobenzene
Concen: 42.165 ng
RT: 16.440 min Scan# 2270
Delta R.T. 0.000 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:284 Resp: 391978

Ion	Ratio	Lower	Upper
284	100		
142	32.1	24.6	37.0
249	30.0	23.8	35.6







#69
Atrazine
Concen: 39.846 ng
RT: 16.616 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:200 Resp: 216677

Ion	Ratio	Lower	Upper
200	100		
173	25.3	7.5	47.5
215	47.4	28.8	68.8







#70
Pentachlorophenol
Concen: 43.112 ng
RT: 16.798 min Scan# 2331
Delta R.T. 0.000 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:266 Resp: 279598

Ion	Ratio	Lower	Upper
266	100		
268	65.2	51.0	76.6
264	62.8	49.8	74.6







#71
Phenanthrene
Concen: 38.480 ng
RT: 17.198 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:178 Resp: 1501631

Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.4	23.0
179	15.1	12.5	18.7







#72
Anthracene
Concen: 39.721 ng
RT: 17.292 min Scan# 2415
Delta R.T. 0.006 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:178 Resp: 1493914

Ion	Ratio	Lower	Upper
178	100		
176	18.6	14.7	22.1
179	15.2	12.4	18.6







#73
Carbazole
Concen: 40.606 ng
RT: 17.575 min Scan# 24
Delta R.T. 0.006 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:167 Resp: 1492224

Ion	Ratio	Lower	Upper
167	100		
166	21.8	17.1	25.7
139	13.5	10.6	16.0







#74
Di-n-butylphthalate
Concen: 42.387 ng
RT: 18.169 min Scan# 2564
Delta R.T. 0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:149 Resp: 1940609

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.4	11.2
104	5.1	3.9	5.9







#75
Fluoranthene
Concen: 41.464 ng
RT: 19.286 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:202 Resp: 1924412

Ion	Ratio	Lower	Upper
202	100		
101	10.2	0.0	33.2
203	17.0	0.0	37.7







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.604 min Scan# 3148
Delta R.T. 0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:240 Resp: 882732

Ion	Ratio	Lower	Upper
240	100		
120	9.7	9.8	14.8#
236	25.6	20.9	31.3







#77
Benzidine
Concen: 58.813 ng
RT: 19.481 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:184 Resp: 658085

Ion	Ratio	Lower	Upper
184	100		
185	14.0	11.6	17.4
183	12.2	9.2	13.8







#78
Pyrene
Concen: 35.972 ng
RT: 19.663 min Scan# 2818
Delta R.T. 0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:202 Resp: 2017997

Ion	Ratio	Lower	Upper
202	100		
200	21.2	17.2	25.8
203	17.5	14.4	21.6







#79
Terphenyl-d14
Concen: 77.399 ng
RT: 19.875 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:244 Resp: 3389636

Ion	Ratio	Lower	Upper
244	100		
212	7.1	5.6	8.4
122	9.9	9.4	14.0







#80
Butylbenzylphthalate
Concen: 40.073 ng
RT: 20.616 min Scan# 2980
Delta R.T. 0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:149 Resp: 962897

Ion	Ratio	Lower	Upper
149	100		
91	64.6	56.4	84.6
206	21.7	16.3	24.5







#81
Benzo(a)anthracene
Concen: 38.617 ng
RT: 21.580 min Scan# 31
Delta R.T. 0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:228 Resp: 2118786

Ion	Ratio	Lower	Upper
228	100		
226	27.0	21.8	32.8
229	19.7	15.9	23.9







#82
3,3'-Dichlorobenzidine
Concen: 43.353 ng
RT: 21.504 min Scan# 3131
Delta R.T. 0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:252 Resp: 834874

Ion	Ratio	Lower	Upper
252	100		
254	65.0	51.7	77.5
126	9.9	8.8	13.2







#83
Chrysene
Concen: 38.698 ng
RT: 21.645 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:228 Resp: 2034193

Ion	Ratio	Lower	Upper
228	100		
226	29.7	24.2	36.2
229	19.4	15.6	23.4







#84
Bis(2-ethylhexyl)phthalate
Concen: 40.771 ng
RT: 21.510 min Scan# 3132
Delta R.T. 0.000 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:149 Resp: 1436161

Ion	Ratio	Lower	Upper
149	100		
167	27.9	22.4	33.6
279	5.5	4.0	6.0







#85
Di-n-octyl phthalate
Concen: 43.501 ng
RT: 22.786 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:149 Resp: 2491117

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	7.5	6.0	9.0







#86
Perylene-d12
Concen: 20.000 ng
RT: 24.927 min Scan# 3713
Delta R.T. -0.006 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:264 Resp: 1072388

Ion	Ratio	Lower	Upper
264	100		
260	23.4	18.9	28.3
265	22.2	17.8	26.6







#87
Indeno(1,2,3-cd)pyrene
Concen: 39.109 ng
RT: 28.721 min Scan# 4
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:276 Resp: 2911665

Ion	Ratio	Lower	Upper
276	100		
138	22.0	14.5	21.7#
277	25.6	20.2	30.2







#88
Benzo(b)fluoranthene
Concen: 37.729 ng
RT: 23.892 min Scan# 3537
Delta R.T. 0.006 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:252 Resp: 2425230

Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.7	26.5
125	8.8	9.4	14.0#







#89
Benzo(k)fluoranthene
Concen: 37.844 ng
RT: 23.957 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:252 Resp: 2349778

Ion	Ratio	Lower	Upper
252	100		
253	21.5	17.7	26.5
125	8.7	9.1	13.7#







#90
Benzo(a)pyrene
Concen: 38.719 ng
RT: 24.774 min Scan# 3687
Delta R.T. -0.006 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:252 Resp: 2146872

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.6	26.4
125	10.1	10.1	15.1







#91
Dibenzo(a,h)anthracene
Concen: 39.176 ng
RT: 28.786 min Scan# 4
Delta R.T. -0.017 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 2420922

Ion	Ratio	Lower	Upper
278	100		
139	13.9	12.6	19.0
279	24.2	18.5	27.7






#92
Benzo(g,h,i)perylene
Concen: 39.272 ng
RT: 29.909 min Scan# 4560
Delta R.T. 0.012 min
Lab File: BP024522.D
Acq: 05 May 2025 12:38

Tgt Ion:276 Resp: 2470173

Ion	Ratio	Lower	Upper
276	100		
277	23.8	18.9	28.3
138	17.6	18.7	28.1#



Sample Results: **BP024522.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050525\
Data File : BP024522.D
Acq On : 05 May 2025 12:38
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 05 13:05:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	54	-0.02
2		1,4-Dioxane	0.512	0.458	10.5	48#	-0.01
3		Pyridine	1.351	1.195	11.5	46#	-0.01
4		n-Nitrosodimethylamine	0.448	0.478	-6.7	56	-0.01
5	S	2-Fluorophenol	1.210	1.171	3.2	50	-0.02
6		Aniline	1.480	1.385	6.4	47#	-0.02
7	S	Phenol-d6	1.656	1.555	6.1	48#	-0.01
8		2-Chlorophenol	1.350	1.315	2.6	51	-0.02
9		Benzaldehyde	0.819	0.860	-5.0	55	-0.02
10	C	Phenol	1.661	1.523	8.3	47#	-0.02
11		bis(2-Chloroethyl)ether	1.339	1.157	13.6	44#	-0.02
12		1,3-Dichlorobenzene	1.454	1.389	4.5	50	-0.02
13	C	1,4-Dichlorobenzene	1.475	1.397	5.3	50	-0.02
14		1,2-Dichlorobenzene	1.424	1.363	4.3	51	-0.02
15		Benzyl Alcohol	1.071	1.076	-0.5	50#	-0.02
16		2,2'-oxybis(1-Chloropropane	1.447	1.250	13.6	44#	-0.01
17		2-Methylphenol	1.075	1.030	4.2	49#	-0.02
18		Hexachloroethane	0.533	0.505	5.3	50	-0.02
19	P	n-Nitroso-di-n-propylamine	1.025	0.938	8.5	47#	-0.02
20		3+4-Methylphenols	1.491	1.423	4.6	48#	-0.02
21	I	Naphthalene-d8	1.000	1.000	0.0	52	-0.02
22		Acetophenone	0.495	0.487	1.6	48#	-0.02
23	S	Nitrobenzene-d5	0.351	0.347	1.1	49#	-0.02
24		Nitrobenzene	0.347	0.332	4.3	47#	-0.01
25		Isophorone	0.635	0.608	4.3	47#	-0.01
26	C	2-Nitrophenol	0.170	0.183	-7.6	51	-0.02
27		2,4-Dimethylphenol	0.213	0.212	0.5	48#	-0.02
28		bis(2-Chloroethoxy)methane	0.429	0.381	11.2	43#	-0.02
29	C	2,4-Dichlorophenol	0.291	0.297	-2.1	50#	-0.01
30		1,2,4-Trichlorobenzene	0.311	0.308	1.0	50#	-0.02
31		Naphthalene	1.054	1.016	3.6	49#	-0.02
32		Benzoic acid	0.248	0.250	-0.8	53	0.00
33		4-Chloroaniline	0.371	0.374	-0.8	49#	-0.02
34	C	Hexachlorobutadiene	0.183	0.193	-5.5	53	-0.02
35		Caprolactam	0.107	0.114	-6.5	53	-0.01
36	C	4-Chloro-3-methylphenol	0.339	0.352	-3.8	51	-0.02
37		2-Methylnaphthalene	0.731	0.712	2.6	49#	-0.02
38		1-Methylnaphthalene	0.715	0.694	2.9	49#	-0.01
39	I	Acenaphthene-d10	1.000	1.000	0.0	55	-0.01
40		1,2,4,5-Tetrachlorobenzene	0.550	0.538	2.2	51	-0.02
41	P	Hexachlorocyclopentadiene	0.195	0.176	9.7	44#	-0.02
42	S	2,4,6-Tribromophenol	0.277	0.323	-16.6	61	0.00
43	C	2,4,6-Trichlorophenol	0.366	0.371	-1.4	52	-0.02
44		2,4,5-Trichlorophenol	0.405	0.407	-0.5	52	-0.02
45	S	2-Fluorobiphenyl	1.306	1.264	3.2	50	-0.01
46		1,1'-Biphenyl	1.470	1.390	5.4	50#	-0.02
47		2-Chloronaphthalene	1.099	1.048	4.6	50	-0.01

Sample Results: **BP024522.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050525\
Data File : BP024522.D
Acq On : 05 May 2025 12:38
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 05 13:05:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.308	0.324	-5.2	53	-0.01
49		Acenaphthylene	1.730	1.686	2.5	51	-0.01
50		Dimethylphthalate	1.454	1.430	1.7	52	-0.02
51		2,6-Dinitrotoluene	0.309	0.311	-0.6	52	-0.01
52	C	Acenaphthene	1.097	1.039	5.3	50	-0.02
53		3-Nitroaniline	0.325	0.332	-2.2	51	0.00
54	P	2,4-Dinitrophenol	0.182	0.189	-3.8	56	0.00
55		Dibenzofuran	1.793	1.694	5.5	50	-0.01
56	P	4-Nitrophenol	0.280	0.285	-1.8	51	0.00
57		2,4-Dinitrotoluene	0.421	0.437	-3.8	53	0.00
58		Fluorene	1.388	1.335	3.8	52	-0.01
59		2,3,4,6-Tetrachlorophenol	0.373	0.378	-1.3	53	0.00
60		Diethylphthalate	1.499	1.480	1.3	52	0.00
61		4-Chlorophenyl-phenylether	0.674	0.659	2.2	53	0.00
62		4-Nitroaniline	0.344	0.368	-7.0	54	0.00
63		Azobenzene	1.378	1.267	8.1	48#	0.00
64	I	Phenanthrene-d10	1.000	1.000	0.0	56	0.00
65		4,6-Dinitro-2-methylphenol	0.126	0.130	-3.2	56	0.00
66	c	n-Nitrosodiphenylamine	0.597	0.572	4.2	51	0.00
67		4-Bromophenyl-phenylether	0.219	0.222	-1.4	54	0.00
68		Hexachlorobenzene	0.260	0.274	-5.4	57	0.00
69		Atrazine	0.152	0.151	0.7	70	0.00
70	C	Pentachlorophenol	0.181	0.195	-7.7	56	0.00
71		Phenanthrene	1.091	1.050	3.8	52	0.00
72		Anthracene	1.052	1.044	0.8	53	0.00
73		Carbazole	1.027	1.043	-1.6	54	0.00
74		Di-n-butylphthalate	1.280	1.356	-5.9	54	0.01
75	C	Fluoranthene	1.298	1.345	-3.6	56	0.00
76	I	Chrysene-d12	1.000	1.000	0.0	65	0.01
77		Benzidine	0.254	0.373	-46.9#	56	0.00
78		Pyrene	1.271	1.143	10.1	57	0.01
79	S	Terphenyl-d14	0.992	0.960	3.2	61	0.00
80		Butylbenzylphthalate	0.544	0.545	-0.2	60	0.01
81		Benzo(a)anthracene	1.243	1.200	3.5	61	0.01
82		3,3'-Dichlorobenzidine	0.436	0.473	-8.5	64	0.01
83		Chrysene	1.191	1.152	3.3	62	0.00
84		Bis(2-ethylhexyl)phthalate	0.798	0.813	-1.9	60	0.00
85	c	Di-n-octyl phthalate	1.297	1.411	-8.8	63	0.00
86	I	Perylene-d12	1.000	1.000	0.0	72	0.00
87		Indeno(1,2,3-cd)pyrene	1.388	1.358	2.2	68	-0.02
88		Benzo(b)fluoranthene	1.199	1.131	5.7	65	0.00
89		Benzo(k)fluoranthene	1.158	1.096	5.4	65	0.00
90	C	Benzo(a)pyrene	1.034	1.001	3.2	66	0.00
91		Dibenzo(a,h)anthracene	1.152	1.129	2.0	68	-0.02
92		Benzo(g,h,i)perylene	1.173	1.152	1.8	68	0.01

Sample Results: **BP024522.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050525\
Data File : BP024522.D
Acq On : 05 May 2025 12:38
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 05 13:05:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Sample Results: **BP024522.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050525\
Data File : BP024522.D
Acq On : 05 May 2025 12:38
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 05 13:05:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	54	-0.02
2		1,4-Dioxane	40.000	35.777	10.6	48	-0.01
3		Pyridine	40.000	35.372	11.6	46	-0.01
4		n-Nitrosodimethylamine	40.000	42.639	-6.6	56	-0.01
5	S	2-Fluorophenol	80.000	77.458	3.2	50	-0.02
6		Aniline	40.000	37.437	6.4	47	-0.02
7	S	Phenol-d6	80.000	75.121	6.1	48	-0.01
8		2-Chlorophenol	40.000	38.938	2.7	51	-0.02
9		Benzaldehyde	40.000	42.007	-5.0	55	-0.02
10	C	Phenol	40.000	36.656	8.4	47	-0.02
11		bis(2-Chloroethyl)ether	40.000	34.563	13.6	44	-0.02
12		1,3-Dichlorobenzene	40.000	38.210	4.5	50	-0.02
13	C	1,4-Dichlorobenzene	40.000	37.877	5.3	50	-0.02
14		1,2-Dichlorobenzene	40.000	38.282	4.3	51	-0.02
15		Benzyl Alcohol	40.000	40.197	-0.5	50	-0.02
16		2,2'-oxybis(1-Chloropropane	40.000	34.559	13.6	44	-0.01
17		2-Methylphenol	40.000	38.345	4.1	49	-0.02
18		Hexachloroethane	40.000	37.884	5.3	50	-0.02
19	P	n-Nitroso-di-n-propylamine	40.000	36.635	8.4	47	-0.02
20		3+4-Methylphenols	40.000	38.178	4.6	48	-0.02
21	I	Naphthalene-d8	20.000	20.000	0.0	52	-0.02
22		Acetophenone	40.000	39.368	1.6	48	-0.02
23	S	Nitrobenzene-d5	80.000	79.065	1.2	49	-0.02
24		Nitrobenzene	40.000	38.268	4.3	47	-0.01
25		Isophorone	40.000	38.294	4.3	47	-0.01
26	C	2-Nitrophenol	40.000	43.018	-7.5	51	-0.02
27		2,4-Dimethylphenol	40.000	39.701	0.7	48	-0.02
28		bis(2-Chloroethoxy)methane	40.000	35.563	11.1	43	-0.02
29	C	2,4-Dichlorophenol	40.000	40.835	-2.1	50	-0.01
30		1,2,4-Trichlorobenzene	40.000	39.565	1.1	50	-0.02
31		Naphthalene	40.000	38.590	3.5	49	-0.02
32		Benzoic acid	40.000	40.228	-0.6	53	0.00
33		4-Chloroaniline	40.000	40.376	-0.9	49	-0.02
34	C	Hexachlorobutadiene	40.000	42.102	-5.3	53	-0.02
35		Caprolactam	40.000	42.634	-6.6	53	-0.01
36	C	4-Chloro-3-methylphenol	40.000	41.549	-3.9	51	-0.02
37		2-Methylnaphthalene	40.000	38.998	2.5	49	-0.02
38		1-Methylnaphthalene	40.000	38.858	2.9	49	-0.01
39	I	Acenaphthene-d10	20.000	20.000	0.0	55	-0.01
40		1,2,4,5-Tetrachlorobenzene	40.000	39.159	2.1	51	-0.02
41	P	Hexachlorocyclopentadiene	40.000	36.275	9.3	44	-0.02
42	S	2,4,6-Tribromophenol	80.000	93.287	-16.6	61	0.00
43	C	2,4,6-Trichlorophenol	40.000	40.552	-1.4	52	-0.02
44		2,4,5-Trichlorophenol	40.000	40.181	-0.5	52	-0.02
45	S	2-Fluorobiphenyl	80.000	77.408	3.2	50	-0.01
46		1,1'-Biphenyl	40.000	37.812	5.5	50	-0.02
47		2-Chloronaphthalene	40.000	38.154	4.6	50	-0.01

Sample Results: **BP024522.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050525\
Data File : BP024522.D
Acq On : 05 May 2025 12:38
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 05 13:05:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	42.002	-5.0	53	-0.01
49		Acenaphthylene	40.000	38.978	2.6	51	-0.01
50		Dimethylphthalate	40.000	39.326	1.7	52	-0.02
51		2,6-Dinitrotoluene	40.000	40.338	-0.8	52	-0.01
52	C	Acenaphthene	40.000	37.907	5.2	50	-0.02
53		3-Nitroaniline	40.000	40.918	-2.3	51	0.00
54	P	2,4-Dinitrophenol	40.000	41.652	-4.1	56	0.00
55		Dibenzofuran	40.000	37.805	5.5	50	-0.01
56	P	4-Nitrophenol	40.000	40.734	-1.8	51	0.00
57		2,4-Dinitrotoluene	40.000	41.451	-3.6	53	0.00
58		Fluorene	40.000	38.466	3.8	52	-0.01
59		2,3,4,6-Tetrachlorophenol	40.000	40.451	-1.1	53	0.00
60		Diethylphthalate	40.000	39.507	1.2	52	0.00
61		4-Chlorophenyl-phenylether	40.000	39.111	2.2	53	0.00
62		4-Nitroaniline	40.000	42.858	-7.1	54	0.00
63		Azobenzene	40.000	36.799	8.0	48	0.00
64	I	Phenanthrene-d10	20.000	20.000	0.0	56	0.00
65		4,6-Dinitro-2-methylphenol	40.000	41.322	-3.3	56	0.00
66	c	n-Nitrosodiphenylamine	40.000	38.306	4.2	51	0.00
67		4-Bromophenyl-phenylether	40.000	40.574	-1.4	54	0.00
68		Hexachlorobenzene	40.000	42.165	-5.4	57	0.00
69		Atrazine	40.000	39.846	0.4	70	0.00
70	C	Pentachlorophenol	40.000	43.112	-7.8	56	0.00
71		Phenanthrene	40.000	38.480	3.8	52	0.00
72		Anthracene	40.000	39.721	0.7	53	0.00
73		Carbazole	40.000	40.606	-1.5	54	0.00
74		Di-n-butylphthalate	40.000	42.387	-6.0	54	0.01
75	C	Fluoranthene	40.000	41.464	-3.7	56	0.00
76	I	Chrysene-d12	20.000	20.000	0.0	65	0.01
77		Benzidine	40.000	58.813	-47.0#	56	0.00
78		Pyrene	40.000	35.972	10.1	57	0.01
79	S	Terphenyl-d14	80.000	77.399	3.3	61	0.00
80		Butylbenzylphthalate	40.000	40.073	-0.2	60	0.01
81		Benzo(a)anthracene	40.000	38.617	3.5	61	0.01
82		3,3'-Dichlorobenzidine	40.000	43.353	-8.4	64	0.01
83		Chrysene	40.000	38.698	3.3	62	0.00
84		Bis(2-ethylhexyl)phthalate	40.000	40.771	-1.9	60	0.00
85	c	Di-n-octyl phthalate	40.000	43.501	-8.8	63	0.00
86	I	Perylene-d12	20.000	20.000	0.0	72	0.00
87		Indeno(1,2,3-cd)pyrene	40.000	39.109	2.2	68	-0.02
88		Benzo(b)fluoranthene	40.000	37.729	5.7	65	0.00
89		Benzo(k)fluoranthene	40.000	37.844	5.4	65	0.00
90	C	Benzo(a)pyrene	40.000	38.719	3.2	66	0.00
91		Dibenzo(a,h)anthracene	40.000	39.176	2.1	68	-0.02
92		Benzo(g,h,i)perylene	40.000	39.272	1.8	68	0.01

Sample Results: **BP024522.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050525\
Data File : BP024522.D
Acq On : 05 May 2025 12:38
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 05 13:05:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

QC SAMPLE DATA


Sample Results: **BF142238.D**


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142238.D
Acq On : 30 Apr 2025 10:55
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Apr 30 16:00:01 2025





AutoFind: Scans 1614, 1615, 1616; Background Corrected with Scan 1607


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.6	86007	PASS
68	69	0.00	2	1.8	1316	PASS
69	198	0.00	100	33.1	71903	PASS
70	69	0.00	2	0.5	377	PASS
127	198	10	80	45.1	97928	PASS
197	198	0.00	2	0.8	1760	PASS
198	198	100	100	100.0	217131	PASS
199	198	5	9	6.5	14095	PASS
275	198	10	60	27.6	60019	PASS
365	198	1	100	3.7	7976	PASS
441	198	0.01	100	19.0	41171	PASS
442	442	50	100	100.0	263808	PASS
443	442	15	24	19.9	52467	PASS


Sample Results: **BF142238.D**


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142238.D
Acq On : 30 Apr 2025 10:55
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Apr 30 16:04:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration







TIC: BF142238.D\data.ms

(70) Pentachlorophenol (C)

11.233min (+ 0.000) 69575.24 ng


response 180630


Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.50	63.39
264.00	61.60	62.57
0.00	0.00	0.00


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142238.D
Acq On : 30 Apr 2025 10:55
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Apr 30 16:04:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration







TIC: BF142238.D\data.ms

(77) Benzidine

12.774min (+ 0.006) 97416.22 ng

response 1440496

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.25
183.00	11.10	11.10
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
4/30/2025	BNA_F	BF142238.D
Compound Name	**Response**	**Retention Time**
DDT	378546	13.598
DDD	2841	13.333
DDE	0	12.904
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
2841	381387	0.74

Instrument :
BNA_F
ClientSampleId :
DFTPP


Sample Results: **BF142249.D**


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142249.D
Acq On : 01 May 2025 09:48
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Apr 30 16:00:01 2025





AutoFind: Scans 1613, 1614, 1615; Background Corrected with Scan 1607

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.2	103579	PASS
68	69	0.00	2	1.6	1339	PASS
69	198	0.00	100	35.5	83059	PASS
70	69	0.00	2	0.5	413	PASS
127	198	10	80	47.7	111579	PASS
197	198	0.00	2	0.7	1704	PASS
198	198	100	100	100.0	234155	PASS
199	198	5	9	6.7	15632	PASS
275	198	10	60	26.7	62555	PASS
365	198	1	100	3.4	8074	PASS
441	198	0.01	100	17.1	40008	PASS
442	442	50	100	100.0	263019	PASS
443	442	15	24	19.1	50208	PASS

DDT Breakdown

Date	Instrument Name	DFTPP Data File
5/1/2025	BNA_F	BF142249.D
Compound Name	**Response**	**Retention Time**
DDT	427368	13.598
DDD	6172	13.327
DDE	0	12.904
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
6172	433540	1.42


Instrument :
BNA_F
ClientSampleId :
DFTPP


Sample Results: **BF142249.D**


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142249.D
Acq On : 01 May 2025 09:48
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: May 01 12:00:18 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration







TIC: BF142249.D\data.ms

(70) Pentachlorophenol (C)

11.233min (+ 0.000) 113470.87 ng


response 170474


Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.50	65.06
264.00	61.60	64.75
0.00	0.00	0.00


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142249.D
Acq On : 01 May 2025 09:48
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: May 01 12:00:18 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration







TIC: BF142249.D\data.ms

(77) Benzidine

12.774min (+ 0.006) 156734.40 ng

response 1611689

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.68
183.00	11.10	11.08
0.00	0.00	0.00


Sample Results: **BF142273.D**


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142273.D
Acq On : 02 May 2025 10:11
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Apr 30 16:00:01 2025





AutoFind: Scans 1613, 1614, 1615; Background Corrected with Scan 1606

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.1	103072	PASS
68	69	0.00	2	1.6	1298	PASS
69	198	0.00	100	33.5	82011	PASS
70	69	0.00	2	0.6	473	PASS
127	198	10	80	45.6	111691	PASS
197	198	0.00	2	0.8	2015	PASS
198	198	100	100	100.0	244928	PASS
199	198	5	9	6.8	16632	PASS
275	198	10	60	28.7	70309	PASS
365	198	1	100	3.7	9025	PASS
441	198	0.01	100	18.7	45787	PASS
442	442	50	100	100.0	296981	PASS
443	442	15	24	19.3	57432	PASS

DDT Breakdown

Date	Instrument Name	DFTPP Data File
5/2/2025	BNA_F	BF142273.D
Compound Name	**Response**	**Retention Time**
DDT	646019	13.592
DDD	7521	13.322
DDE	403	12.951
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
7924	653943	1.21


Instrument :
BNA_F
ClientSampleId :
DFTPP


Sample Results: **BF142273.D**


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142273.D
Acq On : 02 May 2025 10:11
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: May 02 11:49:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration







TIC: BF142273.D\data.ms

(70) Pentachlorophenol (C)

11.228min (-0.005) 200113.05 ng

response 342837


Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.50	63.76
264.00	61.60	64.00
0.00	0.00	0.00


Sample Results: **BF142273.D**


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142273.D
Acq On : 02 May 2025 10:11
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: May 02 11:49:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration







TIC: BF142273.D\data.ms

(77) Benzidine

12.769min (-0.000) 120826.54 ng

response 1673715

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.73
183.00	11.10	11.31
0.00	0.00	0.00


Sample Results: **BP024274.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024274.D
Acq On : 14 Apr 2025 10:25
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Apr 15 04:48:42 2025



AutoFind: Scans 2474, 2475, 2476; Background Corrected with Scan 2463

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.4	129796	PASS
68	69	0.00	2	1.8	2815	PASS
69	198	0.00	100	36.0	153442	PASS
70	69	0.00	2	0.5	836	PASS
127	198	10	80	48.5	206858	PASS
197	198	0.00	2	0.5	2122	PASS
198	198	100	100	100.0	426527	PASS
199	198	5	9	6.9	29549	PASS
275	198	10	60	29.1	124199	PASS
365	198	1	100	4.1	17663	PASS
441	198	0.01	100	15.2	64862	PASS
442	442	100	100	100.0	422528	PASS
443	442	15	24	19.4	82090	PASS

DDT Breakdown

Date	Instrument Name	DFTPP Data File
4/14/2025	BNA_P	BP024274.D
Compound Name	**Response**	**Retention Time**
DDT	1715151	20.798
DDD	26002	20.386
DDE	1572	19.81
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
27574	1742725	1.58


Instrument :
BNA_P
ClientSampleId :
DFTPP


Sample Results: **BP024274.D**


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024274.D
Acq On : 14 Apr 2025 10:25
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Apr 14 15:07:47 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 15:04:09 2025
Response via : Initial Calibration







TIC: BP024274.D\data.ms

(70) Pentachlorophenol (C)

16.822min (+ 0.012) 3117222.42 ng

response 611847


Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.80	65.61
264.00	62.20	62.81
0.00	0.00	0.00


Sample Results: **BP024274.D**


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024274.D
Acq On : 14 Apr 2025 10:25
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Apr 14 15:07:47 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 15:04:09 2025
Response via : Initial Calibration







TIC: BP024274.D\data.ms

(77) Benzidine

19.504min (+ 0.006) 7415471.13 ng

response 2354373

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.50	14.60
183.00	11.50	11.64
0.00	0.00	0.00


Sample Results: **BP024521.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050525\
Data File : BP024521.D
Acq On : 05 May 2025 11:57
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Apr 18 12:04:48 2025



AutoFind: Scans 2470, 2471, 2472; Background Corrected with Scan 2457

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.2	94677	PASS
68	69	0.00	2	1.1	1105	PASS
69	198	0.00	100	31.3	105085	PASS
70	69	0.00	2	0.6	601	PASS
127	198	10	80	45.5	153112	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	336159	PASS
199	198	5	9	6.7	22574	PASS
275	198	10	60	33.1	111207	PASS
365	198	1	100	5.3	17800	PASS
441	198	0.01	100	21.1	71014	PASS
442	442	100	100	100.0	455698	PASS
443	442	15	24	19.2	87372	PASS

DDT Breakdown

Date	Instrument Name	DFTPP Data File
5/5/2025	BNA_P	BP024503.D
Compound Name	**Response**	**Retention Time**
DDT	1294320	20.763
DDD	38430	20.316
DDE	2470	19.775
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
40900	1335220	3.06


Instrument :
BNA_P
ClientSampleId :
DFTPP


Sample Results: **BP024521.D**


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050525\
Data File : BP024521.D
Acq On : 05 May 2025 11:57
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: May 05 13:04:49 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration







TIC: BP024521.D\data.ms

(70) Pentachlorophenol (C)

16.798min (+ 0.000) 40429.31 ng

response 499235


Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.80	64.03
264.00	62.20	61.82
0.00	0.00	0.00


Sample Results: **BP024521.D**


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050525\
Data File : BP024521.D
Acq On : 05 May 2025 11:57
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: May 05 13:04:49 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration







TIC: BP024521.D\data.ms

(77) Benzidine

19.475min (-0.006) 48141.36 ng

response 2700279

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.50	14.03
183.00	11.50	12.14
0.00	0.00	0.00

Raw Data: **BF142254.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	
Client Sample ID:	PB167810BL	SDG No.:	Q1901
Lab Sample ID:	PB167810BL	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142254.D	1	04/30/25 13:15	05/01/25 12:10	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	1.30	U	1.30	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.53	U	0.53	5.00	ug/L
95-48-7	2-Methylphenol	1.10	U	1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	1.10	U	1.10	10.0	ug/L
67-72-1	Hexachloroethane	0.65	U	0.65	5.00	ug/L
98-95-3	Nitrobenzene	0.76	U	0.76	5.00	ug/L
87-68-3	Hexachlorobutadiene	0.54	U	0.54	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	0.51	U	0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	0.62	U	0.62	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	1.20	U	1.20	5.00	ug/L
118-74-1	Hexachlorobenzene	0.52	U	0.52	5.00	ug/L
87-86-5	Pentachlorophenol	1.60	U	1.60	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	110		15 (10) - 110 (139)	73%	SPK: 150
13127-88-3	Phenol-d6	109		15 (10) - 110 (134)	73%	SPK: 150
4165-60-0	Nitrobenzene-d5	83.3		30 (49) - 130 (133)	83%	SPK: 100
321-60-8	2-Fluorobiphenyl	69.1		30 (52) - 130 (132)	69%	SPK: 100
118-79-6	2,4,6-Tribromophenol	124		15 (44) - 110 (137)	83%	SPK: 150
1718-51-0	Terphenyl-d14	63.3		30 (48) - 130 (125)	63%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	224000	6.904			
1146-65-2	Naphthalene-d8	858000	8.187			
15067-26-2	Acenaphthene-d10	450000	9.945			
1517-22-2	Phenanthrene-d10	795000	11.428			
1719-03-5	Chrysene-d12	568000	14.069			
1520-96-3	Perylene-d12	405000	15.563			

Raw Data: **BF142254.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025			Date Received:	
Client Sample ID:	PB167810BL			SDG No.:	Q1901
Lab Sample ID:	PB167810BL			Matrix:	TCLP
Analytical Method:	SW8270			% Solid:	0
Sample Wt/Vol:	1000	Units: mL		Final Vol:	1000 uL
Soil Aliquot Vol:		uL		Test:	TCLP BNA
Extraction Type :		Decanted : N		Level :	LOW
Injection Volume :		GPC Factor : 1.0		GPC Cleanup : N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142254.D	1	04/30/25 13:15	05/01/25 12:10	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142254.D
Acq On : 01 May 2025 12:10
Operator : RC/JU
Sample : PB167810BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB167810BL

Quant Time: May 01 12:47:29 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.904	152	223997	20.000	ng	0.00
21) Naphthalene-d8	8.187	136	857825	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	449926	20.000	ng	0.00
64) Phenanthrene-d10	11.428	188	794711	20.000	ng	0.00
76) Chrysene-d12	14.069	240	568207	20.000	ng	0.00
86) Perylene-d12	15.563	264	404816	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.528	112	1546121	110.236	ng	0.01
7) Phenol-d6	6.522	99	1853180	109.417	ng	0.00
23) Nitrobenzene-d5	7.469	82	1087156	83.261	ng	0.00
42) 2,4,6-Tribromophenol	10.733	330	508946	124.082	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	2205083	69.102	ng	0.00
79) Terphenyl-d14	13.022	244	2472921	63.298	ng	0.00
Target Compounds						Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed


Sample Results: **BF142254.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142254.D
Acq On : 01 May 2025 12:10
Operator : RC/JU
Sample : PB167810BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB167810BL

Quant Time: May 01 12:47:29 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.904 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF142254.D
Acq: 01 May 2025 12:10

Instrument :
BNA_F
ClientSampleId :
PB167810BL

Tgt Ion:152 Resp: 223997

Ion	Ratio	Lower	Upper
152	100		
150	156.1	130.2	195.2
115	59.2	47.0	70.4







#5
2-Fluorophenol
Concen: 110.236 ng
RT: 5.528 min Scan# 567
Delta R.T. 0.012 min
Lab File: BF142254.D
Acq: 01 May 2025 12:10

Tgt Ion:112 Resp: 1546121

Ion	Ratio	Lower	Upper
112	100		
64	60.1	48.4	72.6
63	31.1	24.8	37.2







#7
Phenol-d6
Concen: 109.417 ng
RT: 6.522 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF142254.D
Acq: 01 May 2025 12:10

Instrument :
BNA_F
ClientSampleId :
PB167810BL

Tgt Ion: 99 Resp: 1853180

Ion	Ratio	Lower	Upper
99	100		
42	21.1	16.3	24.5
71	32.5	25.0	37.4







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.187 min Scan# 1019
Delta R.T. -0.006 min
Lab File: BF142254.D
Acq: 01 May 2025 12:10

Tgt Ion:136 Resp: 857825

Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.7	13.1
54	9.6	7.7	11.5
68	6.4	5.0	7.6







#23
Nitrobenzene-d5
Concen: 83.261 ng
RT: 7.469 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF142254.D
Acq: 01 May 2025 12:10

Instrument :
BNA_F
ClientSampleId :
PB167810BL

Tgt Ion: 82 Resp: 1087156

Ion	Ratio	Lower	Upper
82	100		
128	46.4	37.0	55.6
54	59.8	48.0	72.0







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.945 min Scan# 1318
Delta R.T. -0.006 min
Lab File: BF142254.D
Acq: 01 May 2025 12:10

Tgt Ion:164 Resp: 449926

Ion	Ratio	Lower	Upper
164	100		
162	100.6	78.6	117.8
160	45.1	35.7	53.5







#42
2,4,6-Tribromophenol
Concen: 124.082 ng
RT: 10.733 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF142254.D
Acq: 01 May 2025 12:10

Instrument :
BNA_F
ClientSampleId :
PB167810BL

Tgt Ion:330 Resp: 508946

Ion	Ratio	Lower	Upper
330	100		
332	96.7	76.5	114.7
141	29.6	24.3	36.5







#45
2-Fluorobiphenyl
Concen: 69.102 ng
RT: 9.263 min Scan# 1202
Delta R.T. -0.006 min
Lab File: BF142254.D
Acq: 01 May 2025 12:10

Tgt Ion:172 Resp: 2205083

Ion	Ratio	Lower	Upper
172	100		
171	36.1	29.0	43.4
170	23.9	19.4	29.0







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.428 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142254.D
Acq: 01 May 2025 12:10

Instrument :
BNA_F
ClientSampleId :
PB167810BL

Tgt Ion:188 Resp: 794711

Ion	Ratio	Lower	Upper
188	100		
94	10.8	8.2	12.2
80	11.1	8.6	12.8







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.069 min Scan# 2019
Delta R.T. -0.006 min
Lab File: BF142254.D
Acq: 01 May 2025 12:10

Tgt Ion:240 Resp: 568207

Ion	Ratio	Lower	Upper
240	100		
120	12.6	10.1	15.1
236	25.7	20.5	30.7







#79
Terphenyl-d14
Concen: 63.298 ng
RT: 13.022 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142254.D
Acq: 01 May 2025 12:10

Instrument :
BNA_F
ClientSampleId :
PB167810BL

Tgt Ion:244 Resp: 2472921

Ion	Ratio	Lower	Upper
244	100		
212	6.7	5.3	7.9
122	10.9	9.1	13.7






#86
Perylene-d12
Concen: 20.000 ng
RT: 15.563 min Scan# 2273
Delta R.T. -0.006 min
Lab File: BF142254.D
Acq: 01 May 2025 12:10

Tgt Ion:264 Resp: 404816

Ion	Ratio	Lower	Upper
264	100		
260	22.8	18.4	27.6
265	21.1	17.3	25.9



Raw Data: **BF142255.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	
Client Sample ID:	PB167810BS	SDG No.:	Q1901
Lab Sample ID:	PB167810BS	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142255.D	1	04/30/25 13:15	05/01/25 12:39	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	37.2		1.30	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	39.6		0.53	5.00	ug/L
95-48-7	2-Methylphenol	41.0		1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	40.8		1.10	10.0	ug/L
67-72-1	Hexachloroethane	41.8		0.65	5.00	ug/L
98-95-3	Nitrobenzene	45.8		0.76	5.00	ug/L
87-68-3	Hexachlorobutadiene	41.5		0.54	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	43.4		0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	44.9		0.62	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	49.8		1.20	5.00	ug/L
118-74-1	Hexachlorobenzene	41.6		0.52	5.00	ug/L
87-86-5	Pentachlorophenol	93.4	E	1.60	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	111		15 (10) - 110 (139)	74%	SPK: 150
13127-88-3	Phenol-d6	113		15 (10) - 110 (134)	75%	SPK: 150
4165-60-0	Nitrobenzene-d5	82.4		30 (49) - 130 (133)	82%	SPK: 100
321-60-8	2-Fluorobiphenyl	68.5		30 (52) - 130 (132)	69%	SPK: 100
118-79-6	2,4,6-Tribromophenol	132		15 (44) - 110 (137)	88%	SPK: 150
1718-51-0	Terphenyl-d14	74.1		30 (48) - 130 (125)	74%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	229000	6.91			
1146-65-2	Naphthalene-d8	893000	8.192			
15067-26-2	Acenaphthene-d10	472000	9.945			
1517-22-2	Phenanthrene-d10	805000	11.433			
1719-03-5	Chrysene-d12	464000	14.074			
1520-96-3	Perylene-d12	447000	15.568			

Raw Data: **BF142255.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:			
Project:	Amtrak Sawtooth Bridges 2025			Date Received:			
Client Sample ID:	PB167810BS			SDG No.:	Q1901		
Lab Sample ID:	PB167810BS			Matrix:	TCLP		
Analytical Method:	SW8270			% Solid:	0		
Sample Wt/Vol:	1000	Units:	mL	Final Vol:	1000		uL
Soil Aliquot Vol:			uL	Test:	TCLP BNA		
Extraction Type :			Decanted : N	Level :	LOW		
Injection Volume :			GPC Factor : 1.0	GPC Cleanup :	N	PH :	
Prep Method :	SW3510C						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142255.D	1	04/30/25 13:15	05/01/25 12:39	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142255.D
Acq On : 01 May 2025 12:39
Operator : RC/JU
Sample : PB167810BS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Quant Time: May 01 13:05:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	228833	20.000	ng	0.00
21) Naphthalene-d8	8.192	136	892782	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	472284	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	804874	20.000	ng	0.00
76) Chrysene-d12	14.074	240	464382	20.000	ng	0.00
86) Perylene-d12	15.568	264	446826	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.534	112	1593182	111.191	ng	0.02
7) Phenol-d6	6.528	99	1946714	112.510	ng	0.00
23) Nitrobenzene-d5	7.469	82	1119930	82.412	ng	0.00
42) 2,4,6-Tribromophenol	10.739	330	568456	132.029	ng	0.00
45) 2-Fluorobiphenyl	9.269	172	2294519	68.501	ng	0.00
79) Terphenyl-d14	13.021	244	2366082	74.104	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.798	88	217494	33.987	ng	98
3) Pyridine	3.551	79	618200	37.171	ng	100
4) n-Nitrosodimethylamine	3.504	42	358234	41.623	ng	99
6) Aniline	6.569	93	471322	19.245	ng	100
8) 2-Chlorophenol	6.692	128	631635	42.661	ng	99
9) Benzaldehyde	6.457	77	305773	25.285	ng	98
10) Phenol	6.545	94	755191m	40.542	ng	
11) bis(2-Chloroethyl)ether	6.639	93	588988	40.804	ng	99
12) 1,3-Dichlorobenzene	6.851	146	659968	39.402	ng	98
13) 1,4-Dichlorobenzene	6.928	146	667827	39.574	ng	99
14) 1,2-Dichlorobenzene	7.075	146	650967	40.407	ng	99
15) Benzyl Alcohol	7.045	79	515024	41.969	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.181	45	1072194	41.135	ng	100
17) 2-Methylphenol	7.151	107	497856	41.046	ng	99
18) Hexachloroethane	7.422	117	231050	41.832	ng	97
19) n-Nitroso-di-n-propyla...	7.322	70	427592	39.992	ng	99
20) 3+4-Methylphenols	7.304	107	618253	40.760	ng	92
22) Acetophenone	7.316	105	826671	40.025	ng	99
24) Nitrobenzene	7.492	77	610784	45.807	ng	99
25) Isophorone	7.728	82	1177064	41.271	ng	99
26) 2-Nitrophenol	7.804	139	256947	46.237	ng	99
27) 2,4-Dimethylphenol	7.839	122	602302	41.886	ng	100
28) bis(2-Chloroethoxy)met...	7.939	93	734241	40.821	ng	99
29) 2,4-Dichlorophenol	8.039	162	518664	43.143	ng	99
30) 1,2,4-Trichlorobenzene	8.128	180	544203	40.565	ng	99
31) Naphthalene	8.216	128	1769061	39.733	ng	100
32) Benzoic acid	7.945	122	321202m	47.181	ng	
33) 4-Chloroaniline	8.251	127	82158	4.431	ng	99
34) Hexachlorobutadiene	8.328	225	324756	41.492	ng	99
35) Caprolactam	8.622	113	163644m	45.287	ng	
36) 4-Chloro-3-methylphenol	8.733	107	523494	42.406	ng	99
37) 2-Methylnaphthalene	8.904	142	1090096	39.822	ng	99
38) 1-Methylnaphthalene	9.004	142	1143793	40.333	ng	100
40) 1,2,4,5-Tetrachloroben...	9.069	216	524634	39.471	ng	99
41) Hexachlorocyclopentadiene	9.057	237	668416	88.607	ng	100
43) 2,4,6-Trichlorophenol	9.175	196	357651	43.423	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142255.D
Acq On : 01 May 2025 12:39
Operator : RC/JU
Sample : PB167810BS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Quant Time: May 01 13:05:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.216	196	388570	44.872	ng		98
46) 1,1'-Biphenyl	9.369	154	1458088	39.713	ng		100
47) 2-Chloronaphthalene	9.392	162	1093139	40.017	ng		100
48) 2-Nitroaniline	9.486	65	321433	44.957	ng		99
49) Acenaphthylene	9.810	152	1847757	40.085	ng		100
50) Dimethylphthalate	9.669	163	1268421	41.981	ng		100
51) 2,6-Dinitrotoluene	9.727	165	263569	45.692	ng		98
52) Acenaphthene	9.980	154	1164359	43.261	ng		100
53) 3-Nitroaniline	9.892	138	98911	16.287	ng		95
54) 2,4-Dinitrophenol	9.998	184	196654	90.602	ng	#	18
55) Dibenzofuran	10.157	168	1606504	39.818	ng		100
56) 4-Nitrophenol	10.045	139	484293	101.975	ng		98
57) 2,4-Dinitrotoluene	10.133	165	350987	49.818	ng		98
58) Fluorene	10.498	166	1221943	39.832	ng		100
59) 2,3,4,6-Tetrachlorophenol	10.269	232	330098	46.019	ng		99
60) Diethylphthalate	10.369	149	1267747	42.490	ng		100
61) 4-Chlorophenyl-phenyle...	10.486	204	585254	40.172	ng		99
62) 4-Nitroaniline	10.510	138	286425	44.888	ng		99
63) Azobenzene	10.651	77	1208231	41.009	ng		99
65) 4,6-Dinitro-2-methylph...	10.539	198	126239	47.699	ng		95
66) n-Nitrosodiphenylamine	10.610	169	1104532	40.164	ng		99
67) 4-Bromophenyl-phenylether	10.980	248	373384	41.655	ng		98
68) Hexachlorobenzene	11.045	284	417991	41.636	ng		100
69) Atrazine	11.133	200	332195	45.827	ng		98
70) Pentachlorophenol	11.233	266	495448	93.418	ng		98
71) Phenanthrene	11.463	178	1740817	40.041	ng		100
72) Anthracene	11.510	178	1786049	40.180	ng		100
73) Carbazole	11.663	167	1677441	41.368	ng		100
74) Di-n-butylphthalate	11.998	149	1839790	45.113	ng		100
75) Fluoranthene	12.645	202	1819275	42.331	ng		99
77) Benzidine	12.763	184	347980	17.635	ng		98
78) Pyrene	12.880	202	1803569	42.103	ng		99
80) Butylbenzylphthalate	13.498	149	532587	45.667	ng		99
81) Benzo(a)anthracene	14.062	228	1330282	42.672	ng		99
82) 3,3'-Dichlorobenzidine	14.027	252	74664	8.154	ng		98
83) Chrysene	14.104	228	1152993	40.631	ng		100
84) Bis(2-ethylhexyl)phtha...	14.057	149	632694	39.646	ng		99
85) Di-n-octyl phthalate	14.674	149	1144596	39.690	ng		100
87) Indeno(1,2,3-cd)pyrene	17.080	276	1374940	42.010	ng		100
88) Benzo(b)fluoranthene	15.127	252	1207610	43.510	ng		99
89) Benzo(k)fluoranthene	15.156	252	1003805	39.540	ng		99
90) Benzo(a)pyrene	15.504	252	1073003	42.227	ng		100
91) Dibenzo(a,h)anthracene	17.098	278	1112811	41.894	ng		99
92) Benzo(g,h,i)perylene	17.539	276	1122234	42.026	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142255.D
Acq On : 01 May 2025 12:39
Operator : RC/JU
Sample : PB167810BS
Misc :
ALS Vial : 7 Sample Multiplier: 1


Quant Time: May 01 13:05:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:152 Resp: 22883


Ion	Ratio	Lower	Upper
152	100		
150	162.7	130.2	195.2
115	58.2	47.0	70.4



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#2
1,4-Dioxane
Concen: 33.987 ng
RT: 2.798 min Scan# 103
Delta R.T. 0.100 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion: 88 Resp: 217494

Ion	Ratio	Lower	Upper
88	100		
58	92.0	72.2	108.2
43	35.3	27.0	40.6






#3
Pyridine
Concen: 37.171 ng
RT: 3.551 min Scan# 2
Delta R.T. 0.088 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion: 79 Resp: 61820


Ion	Ratio	Lower	Upper
79	100		
52	88.9	71.5	107.3
51	42.9	34.5	51.7



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#4
n-Nitrosodimethylamine
Concen: 41.623 ng
RT: 3.504 min Scan# 223
Delta R.T. 0.094 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion: 42 Resp: 358234

Ion	Ratio	Lower	Upper
42	100		
74	111.4	89.5	134.3
44	7.3	6.6	9.8





#5
2-Fluorophenol
Concen: 111.191 ng
RT: 5.534 min Scan# 5
Delta R.T. 0.017 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion:112 Resp: 159318


Ion	Ratio	Lower	Upper
112	100		
64	59.3	48.4	72.6
63	30.6	24.8	37.2

Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025







#6
Aniline
Concen: 19.245 ng
RT: 6.569 min Scan# 744
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion: 93 Resp: 471322

Ion	Ratio	Lower	Upper
93	100		
66	39.9	32.1	48.1
65	20.4	16.6	24.8





#7
Phenol-d6
Concen: 112.510 ng
RT: 6.528 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion: 99 Resp: 194671


Ion	Ratio	Lower	Upper
99	100		
42	21.0	16.3	24.5
71	32.4	25.0	37.4

Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025







#8
2-Chlorophenol
Concen: 42.661 ng
RT: 6.692 min Scan# 765
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:128 Resp: 631635

Ion	Ratio	Lower	Upper
128	100		
130	32.9	13.1	53.1
64	47.6	28.6	68.6





#9
Benzaldehyde
Concen: 25.285 ng
RT: 6.457 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion: 77 Resp: 30577


Ion	Ratio	Lower	Upper
77	100		
105	101.1	79.1	119.1
106	97.8	76.0	116.0

Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025







#10
Phenol
Concen: 40.542 ng m
RT: 6.545 min Scan# 740
Delta R.T. 0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion: 94 Resp: 755191

Ion	Ratio	Lower	Upper
94	100		
65	29.0	8.9	48.9
66	39.0	18.7	58.7





#11
bis(2-Chloroethyl)ether
Concen: 40.804 ng
RT: 6.639 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion: 93 Resp: 58898


Ion	Ratio	Lower	Upper
93	100		
63	83.8	62.2	102.2
95	33.0	12.7	52.7

Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025







#12
1,3-Dichlorobenzene
Concen: 39.402 ng
RT: 6.851 min Scan# 792
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:146 Resp: 659968

Ion	Ratio	Lower	Upper
146	100		
148	64.6	50.5	75.7
75	29.9	23.2	34.8






#13
1,4-Dichlorobenzene
Concen: 39.574 ng
RT: 6.928 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:146 Resp: 66782


Ion	Ratio	Lower	Upper
146	100		
148	64.2	50.6	76.0
111	41.9	33.4	50.0



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#14
1,2-Dichlorobenzene
Concen: 40.407 ng
RT: 7.075 min Scan# 830
Delta R.T. -0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:146 Resp: 650967

Ion	Ratio	Lower	Upper
146	100		
148	62.9	50.5	75.7
111	44.3	34.9	52.3





#15
Benzyl Alcohol
Concen: 41.969 ng
RT: 7.045 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion: 79 Resp: 51502


Ion	Ratio	Lower	Upper
79	100		
108	81.8	64.5	96.7
77	62.7	49.2	73.8

Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025







#16
2,2'-oxybis(1-Chloropropane)
Concen: 41.135 ng
RT: 7.181 min Scan# 848
Delta R.T. -0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion: 45 Resp: 1072194

Ion	Ratio	Lower	Upper
45	100		
77	11.1	0.0	31.3
79	8.4	0.0	28.5






#17
2-Methylphenol
Concen: 41.046 ng
RT: 7.151 min Scan# 84
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:107 Resp: 49785


Ion	Ratio	Lower	Upper
107	100		
108	113.7	91.0	136.4
77	43.7	34.4	51.6
79	43.8	34.4	51.6



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#18
Hexachloroethane
Concen: 41.832 ng
RT: 7.422 min Scan# 889
Delta R.T. -0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:117 Resp: 231050

Ion	Ratio	Lower	Upper
117	100		
119	96.8	77.1	115.7
201	93.2	79.4	119.2






#19
n-Nitroso-di-n-propylamine
Concen: 39.992 ng
RT: 7.322 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion: 70 Resp: 42759


Ion	Ratio	Lower	Upper
70	100		
42	67.8	53.4	80.2
101	9.9	7.7	11.5
130	22.1	17.2	25.8



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#20
3+4-Methylphenols
Concen: 40.760 ng
RT: 7.304 min Scan# 869
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:107 Resp: 618253

Ion	Ratio	Lower	Upper
107	100		
108	91.9	71.2	111.2
77	77.5	41.1	81.1
79	28.2	8.1	48.1






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.192 min Scan# 1
Delta R.T. -0.001 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:136 Resp: 89278


Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.7	13.1
54	9.4	7.7	11.5
68	6.1	5.0	7.6



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#22
Acetophenone
Concen: 40.025 ng
RT: 7.316 min Scan# 871
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:105 Resp: 826671

Ion	Ratio	Lower	Upper
105	100		
71	7.5	5.7	8.5
51	37.2	30.1	45.1
120	23.5	19.0	28.4





#23
Nitrobenzene-d5
Concen: 82.412 ng
RT: 7.469 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion: 82 Resp: 111993


Ion	Ratio	Lower	Upper
82	100		
128	45.8	37.0	55.6
54	60.7	48.0	72.0

Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025







#24
Nitrobenzene
Concen: 45.807 ng
RT: 7.492 min Scan# 901
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion: 77 Resp: 610784

Ion	Ratio	Lower	Upper
77	100		
123	49.2	38.6	58.0
65	14.4	11.5	17.3





#25
Isophorone
Concen: 41.271 ng
RT: 7.728 min Scan# 94
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion: 82 Resp: 117706


Ion	Ratio	Lower	Upper
82	100		
95	7.1	5.6	8.4
138	19.4	15.3	22.9

Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025







#26
2-Nitrophenol
Concen: 46.237 ng
RT: 7.804 min Scan# 954
Delta R.T. 0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:139 Resp: 256947

Ion	Ratio	Lower	Upper
139	100		
109	27.6	21.8	32.8
65	50.4	41.1	61.7






#27
2,4-Dimethylphenol
Concen: 41.886 ng
RT: 7.839 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:122 Resp: 60230


Ion	Ratio	Lower	Upper
122	100		
107	107.6	86.0	129.0
121	57.3	46.3	69.5



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#28
bis(2-Chloroethoxy)methane
Concen: 40.821 ng
RT: 7.939 min Scan# 977
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion: 93 Resp: 734241

Ion	Ratio	Lower	Upper
93	100		
95	32.8	26.2	39.2
123	12.9	9.7	14.5





#29
2,4-Dichlorophenol
Concen: 43.143 ng
RT: 8.039 min Scan# 9
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion:162 Resp: 51866


Ion	Ratio	Lower	Upper
162	100		
164	63.7	43.0	83.0
98	37.4	16.7	56.7

Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025







#30
1,2,4-Trichlorobenzene
Concen: 40.565 ng
RT: 8.128 min Scan# 1009
Delta R.T. -0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:180 Resp: 544203

Ion	Ratio	Lower	Upper
180	100		
182	94.4	75.7	113.5
145	32.5	25.4	38.0





#31
Naphthalene
Concen: 39.733 ng
RT: 8.216 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Instrument :
BNA_F
ClientSampleId :
PB167810BS


Tgt Ion:128 Resp: 176906

Ion	Ratio	Lower	Upper
128	100		
129	11.4	9.0	13.4
127	13.2	10.6	15.8

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025







#32
Benzoic acid
Concen: 47.181 ng m
RT: 7.945 min Scan# 978
Delta R.T. 0.011 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:122 Resp: 321202

Ion	Ratio	Lower	Upper
122	100		
105	121.8	98.2	138.2
77	90.5	66.8	106.8






#33
4-Chloroaniline
Concen: 4.431 ng
RT: 8.251 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:127 Resp: 8215


Ion	Ratio	Lower	Upper
127	100		
129	32.0	26.1	39.1
65	32.6	25.5	38.3
92	18.3	15.4	23.2



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#34
Hexachlorobutadiene
Concen: 41.492 ng
RT: 8.328 min Scan# 1043
Delta R.T. -0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:225 Resp: 324756

Ion	Ratio	Lower	Upper
225	100		
223	62.7	49.7	74.5
227	63.2	51.0	76.4






#35
Caprolactam
Concen: 45.287 ng m
RT: 8.622 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:113 Resp: 16364


Ion	Ratio	Lower	Upper
113	100		
55	204.8	177.7	217.7
56	148.3	127.1	167.1



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#36
4-Chloro-3-methylphenol
Concen: 42.406 ng
RT: 8.733 min Scan# 1112
Delta R.T. 0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:107 Resp: 523494

Ion	Ratio	Lower	Upper
107	100		
144	26.8	21.0	31.4
142	83.5	66.1	99.1





#37
2-Methylnaphthalene
Concen: 39.822 ng
RT: 8.904 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion:142 Resp: 109009


Ion	Ratio	Lower	Upper
142	100		
141	88.9	70.5	105.7

Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025







#38
1-Methylnaphthalene
Concen: 40.333 ng
RT: 9.004 min Scan# 1158
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:142 Resp: 1143793

Ion	Ratio	Lower	Upper
142	100		
141	91.4	72.8	109.2






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.945 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:164 Resp: 47228


Ion	Ratio	Lower	Upper
164	100		
162	100.2	78.6	117.8
160	44.6	35.7	53.5



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 39.471 ng
RT: 9.069 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:216 Resp: 524634

Ion	Ratio	Lower	Upper
216	100		
214	79.4	63.8	95.6
179	20.8	16.7	25.1
108	20.9	15.6	23.4






#41
Hexachlorocyclopentadiene
Concen: 88.607 ng
RT: 9.057 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:237 Resp: 66841


Ion	Ratio	Lower	Upper
237	100		
235	63.6	43.5	83.5
272	13.1	0.0	32.9



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#42
2,4,6-Tribromophenol
Concen: 132.029 ng
RT: 10.739 min Scan# 1453
Delta R.T. 0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:330 Resp: 568456

Ion	Ratio	Lower	Upper
330	100		
332	96.0	76.5	114.7
141	29.1	24.3	36.5






#43
2,4,6-Trichlorophenol
Concen: 43.423 ng
RT: 9.175 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:196 Resp: 35765


Ion	Ratio	Lower	Upper
196	100		
198	95.6	77.4	116.2
200	30.7	24.6	37.0



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#44
2,4,5-Trichlorophenol
Concen: 44.872 ng
RT: 9.216 min Scan# 1194
Delta R.T. 0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:196 Resp: 388570

Ion	Ratio	Lower	Upper
196	100		
198	96.0	75.2	112.8
97	48.5	39.8	59.8
132	28.8	23.8	35.8






#45
2-Fluorobiphenyl
Concen: 68.501 ng
RT: 9.269 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:172 Resp: 229451


Ion	Ratio	Lower	Upper
172	100		
171	35.8	29.0	43.4
170	24.0	19.4	29.0



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#46
1,1'-Biphenyl
Concen: 39.713 ng
RT: 9.369 min Scan# 1220
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:154 Resp: 1458088

Ion	Ratio	Lower	Upper
154	100		
153	40.6	20.5	60.5
76	12.8	0.0	32.8






#47
2-Chloronaphthalene
Concen: 40.017 ng
RT: 9.392 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:162 Resp: 109313


Ion	Ratio	Lower	Upper
162	100		
127	40.9	32.6	49.0
164	32.7	26.1	39.1



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#48
2-Nitroaniline
Concen: 44.957 ng
RT: 9.486 min Scan# 1240
Delta R.T. -0.001 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion: 65 Resp: 321433

Ion	Ratio	Lower	Upper
65	100		
92	67.2	54.3	81.5
138	109.3	88.6	133.0






#49
Acenaphthylene
Concen: 40.085 ng
RT: 9.810 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:152 Resp: 184775


Ion	Ratio	Lower	Upper
152	100		
151	20.1	16.2	24.2
153	13.6	10.7	16.1



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#50
Dimethylphthalate
Concen: 41.981 ng
RT: 9.669 min Scan# 1271
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:163 Resp: 1268421

Ion	Ratio	Lower	Upper
163	100		
194	4.0	3.2	4.8
164	10.3	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 45.692 ng
RT: 9.727 min Scan# 1
Delta R.T. -0.001 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion:165 Resp: 26356


Ion	Ratio	Lower	Upper
165	100		
63	59.7	48.5	72.7
89	53.2	43.9	65.9

Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025







#52
Acenaphthene
Concen: 43.261 ng
RT: 9.980 min Scan# 1324
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:154 Resp: 1164359

Ion	Ratio	Lower	Upper
154	100		
153	110.8	88.6	133.0
152	52.8	42.1	63.1





#53
3-Nitroaniline
Concen: 16.287 ng
RT: 9.892 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion:138 Resp: 9891


Ion	Ratio	Lower	Upper
138	100		
108	12.4	11.7	17.5
92	108.3	91.2	136.8

Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025







#54
2,4-Dinitrophenol
Concen: 90.602 ng
RT: 9.998 min Scan# 1327
Delta R.T. 0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:184 Resp: 196654

Ion	Ratio	Lower	Upper
184	100		
63	76.2	69.2	103.8
154	127.3	271.4	407.2#






#55
Dibenzofuran
Concen: 39.818 ng
RT: 10.157 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:168 Resp: 160650


Ion	Ratio	Lower	Upper
168	100		
139	36.8	29.4	44.0
169	13.8	11.0	16.4



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#56
4-Nitrophenol
Concen: 101.975 ng
RT: 10.045 min Scan# 1335
Delta R.T. 0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:139 Resp: 484293

Ion	Ratio	Lower	Upper
139	100		
109	66.2	44.9	84.9
65	97.6	74.7	114.7






#57
2,4-Dinitrotoluene
Concen: 49.818 ng
RT: 10.133 min Scan# 1
Delta R.T. 0.005 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:165 Resp: 35098


Ion	Ratio	Lower	Upper
165	100		
63	48.1	38.8	58.2
89	69.0	57.1	85.7
182	2.7	2.2	3.2



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#58
Fluorene
Concen: 39.832 ng
RT: 10.498 min Scan# 1412
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:166 Resp: 1221943

Ion	Ratio	Lower	Upper
166	100		
165	92.5	73.9	110.9
167	13.5	10.8	16.2






#59
2,3,4,6-Tetrachlorophenol
Concen: 46.019 ng
RT: 10.269 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:232 Resp: 33009


Ion	Ratio	Lower	Upper
232	100		
131	44.4	35.9	53.9
130	2.1	1.8	2.6
166	28.8	24.0	36.0



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#60
Diethylphthalate
Concen: 42.490 ng
RT: 10.369 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:149 Resp: 1267747

Ion	Ratio	Lower	Upper
149	100		
177	21.8	17.6	26.4
150	12.8	10.2	15.2





#61
4-Chlorophenyl-phenylether
Concen: 40.172 ng
RT: 10.486 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Instrument :
BNA_F
ClientSampleId :
PB167810BS


Tgt Ion:204 Resp: 58525

Ion	Ratio	Lower	Upper
204	100		
206	32.3	25.6	38.4
141	59.8	48.9	73.3

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025







#62
4-Nitroaniline
Concen: 44.888 ng
RT: 10.510 min Scan# 1414
Delta R.T. 0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:138 Resp: 286425

Ion	Ratio	Lower	Upper
138	100		
92	49.2	29.2	69.2
108	67.6	49.1	89.1





#63
Azobenzene
Concen: 41.009 ng
RT: 10.651 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Instrument :
BNA_F
ClientSampleId :
PB167810BS


Tgt Ion: 77 Resp: 120823

Ion	Ratio	Lower	Upper
77	100		
182	27.8	7.3	47.3
105	22.7	2.5	42.5
51	34.8	14.4	54.4

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:188 Resp: 804874

Ion	Ratio	Lower	Upper
188	100		
94	10.3	8.2	12.2
80	10.9	8.6	12.8





#65
4,6-Dinitro-2-methylphenol
Concen: 47.699 ng
RT: 10.539 min Scan# 14
Delta R.T. 0.005 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion:198 Resp: 12623


Ion	Ratio	Lower	Upper
198	100		
51	55.6	40.6	80.6
105	46.4	29.2	69.2

Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025







#66
n-Nitrosodiphenylamine
Concen: 40.164 ng
RT: 10.610 min Scan# 1431
Delta R.T. 0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:169 Resp: 1104532

Ion	Ratio	Lower	Upper
169	100		
168	65.8	53.0	79.6
167	36.5	28.8	43.2






#67
4-Bromophenyl-phenylether
Concen: 41.655 ng
RT: 10.980 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:248 Resp: 37338


Ion	Ratio	Lower	Upper
248	100		
250	97.9	77.2	115.8
141	71.4	58.4	87.6



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#68
Hexachlorobenzene
Concen: 41.636 ng
RT: 11.045 min Scan# 1505
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:284 Resp: 417991

Ion	Ratio	Lower	Upper
284	100		
142	30.8	24.6	37.0
249	29.2	23.2	34.8






#69
Atrazine
Concen: 45.827 ng
RT: 11.133 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:200 Resp: 33219


Ion	Ratio	Lower	Upper
200	100		
173	28.5	8.6	48.6
215	50.7	28.5	68.5



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#70
Pentachlorophenol
Concen: 93.418 ng
RT: 11.233 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:266 Resp: 495448

Ion	Ratio	Lower	Upper
266	100		
268	63.8	50.0	75.0
264	63.3	49.3	73.9





#71
Phenanthrene
Concen: 40.041 ng
RT: 11.463 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion:178 Resp: 174081


Ion	Ratio	Lower	Upper
178	100		
176	19.6	15.5	23.3
179	15.5	12.6	19.0

Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025







#72
Anthracene
Concen: 40.180 ng
RT: 11.510 min Scan# 1584
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:178 Resp: 1786049

Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.3	22.9
179	15.9	12.7	19.1






#73
Carbazole
Concen: 41.368 ng
RT: 11.663 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:167 Resp: 167744


Ion	Ratio	Lower	Upper
167	100		
166	21.5	17.3	25.9
139	12.9	10.3	15.5



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#74
Di-n-butylphthalate
Concen: 45.113 ng
RT: 11.998 min Scan# 1667
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:149 Resp: 1839790

Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.5	11.3
104	4.5	3.6	5.4






#75
Fluoranthene
Concen: 42.331 ng
RT: 12.645 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:202 Resp: 181927


Ion	Ratio	Lower	Upper
202	100		
101	12.3	0.0	32.4
203	17.9	0.0	37.5



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.074 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:240 Resp: 464382

Ion	Ratio	Lower	Upper
240	100		
120	12.4	10.1	15.1
236	25.9	20.5	30.7






#77
Benzidine
Concen: 17.635 ng
RT: 12.763 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:184 Resp: 34798


Ion	Ratio	Lower	Upper
184	100		
185	15.8	11.8	17.6
183	10.5	8.9	13.3



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#78
Pyrene
Concen: 42.103 ng
RT: 12.880 min Scan# 1817
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:202 Resp: 1803569

Ion	Ratio	Lower	Upper
202	100		
200	20.3	16.1	24.1
203	17.9	14.1	21.1






#79
Terphenyl-d14
Concen: 74.104 ng
RT: 13.021 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:244 Resp: 236608


Ion	Ratio	Lower	Upper
244	100		
212	6.7	5.3	7.9
122	11.1	9.1	13.7



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#80
Butylbenzylphthalate
Concen: 45.667 ng
RT: 13.498 min Scan# 1922
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:149 Resp: 532587

Ion	Ratio	Lower	Upper
149	100		
91	68.7	54.4	81.6
206	20.9	15.9	23.9





#81
Benzo(a)anthracene
Concen: 42.672 ng
RT: 14.062 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion:228 Resp: 133028


Ion	Ratio	Lower	Upper
228	100		
226	27.0	21.9	32.9
229	20.2	16.0	24.0

Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025







#82
3,3'-Dichlorobenzidine
Concen: 8.154 ng
RT: 14.027 min Scan# 2012
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:252 Resp: 74664

Ion	Ratio	Lower	Upper
252	100		
254	65.3	51.0	76.6
126	11.3	9.4	14.2






#83
Chrysene
Concen: 40.631 ng
RT: 14.104 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:228 Resp: 115299


Ion	Ratio	Lower	Upper
228	100		
226	29.9	24.0	36.0
229	19.5	15.7	23.5



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 39.646 ng
RT: 14.057 min Scan# 2017
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:149 Resp: 632694

Ion	Ratio	Lower	Upper
149	100		
167	26.8	21.7	32.5
279	4.0	3.0	4.4






#85
Di-n-octyl phthalate
Concen: 39.690 ng
RT: 14.674 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:149 Resp: 114459


Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.1	1.7
43	11.5	9.3	13.9



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 15.568 min Scan# 2274
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:264 Resp: 446826

Ion	Ratio	Lower	Upper
264	100		
260	23.6	18.4	27.6
265	21.3	17.3	25.9





#87
Indeno(1,2,3-cd)pyrene
Concen: 42.010 ng
RT: 17.080 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion:276 Resp: 137494


Ion	Ratio	Lower	Upper
276	100		
138	28.3	22.6	33.8
277	25.1	20.1	30.1

Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025







#88
Benzo(b)fluoranthene
Concen: 43.510 ng
RT: 15.127 min Scan# 2199
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:252 Resp: 1207610

Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.7	26.5
125	10.7	9.0	13.4






#89
Benzo(k)fluoranthene
Concen: 39.540 ng
RT: 15.156 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:252 Resp: 100380


Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.8	26.8
125	11.0	9.0	13.6



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025




#90
Benzo(a)pyrene
Concen: 42.227 ng
RT: 15.504 min Scan# 2263
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:252 Resp: 1073003

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.6	26.4
125	11.8	9.5	14.3






#91
Dibenzo(a,h)anthracene
Concen: 41.894 ng
RT: 17.098 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:278 Resp: 111281


Ion	Ratio	Lower	Upper
278	100		
139	19.3	15.2	22.8
279	24.1	19.0	28.6



Instrument :
BNA_F
ClientSampleId :
PB167810BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/02/2025
Supervised By :Jagrut Upadhyay 05/02/2025



#92
Benzo(g,h,i)perylene
Concen: 42.026 ng
RT: 17.539 min Scan# 2609
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:276 Resp: 1122234

Ion	Ratio	Lower	Upper
276	100		
277	23.8	18.7	28.1
138	24.4	19.4	29.0



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MS	SDG No.:	Q1901
Lab Sample ID:	Q1901-08MS	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142277.D	1	04/30/25 13:15	05/02/25 12:27	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	270		12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	340		5.30	50.0	ug/L
95-48-7	2-Methylphenol	360		11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	360		11.0	100	ug/L
67-72-1	Hexachloroethane	350		6.50	50.0	ug/L
98-95-3	Nitrobenzene	440		7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	380		5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	470		5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	460		6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	500		12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	440		5.20	50.0	ug/L
87-86-5	Pentachlorophenol	940	E	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	96.6		15 (10) - 110 (139)	64%	SPK: 150
13127-88-3	Phenol-d6	89.8		15 (10) - 110 (134)	60%	SPK: 150
4165-60-0	Nitrobenzene-d5	84.0		30 (49) - 130 (133)	84%	SPK: 100
321-60-8	2-Fluorobiphenyl	70.7		30 (52) - 130 (132)	71%	SPK: 100
118-79-6	2,4,6-Tribromophenol	126		15 (44) - 110 (137)	84%	SPK: 150
1718-51-0	Terphenyl-d14	70.4		30 (48) - 130 (125)	70%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	213000	6.91			
1146-65-2	Naphthalene-d8	808000	8.193			
15067-26-2	Acenaphthene-d10	402000	9.945			
1517-22-2	Phenanthrene-d10	630000	11.433			
1719-03-5	Chrysene-d12	370000	14.069			
1520-96-3	Perylene-d12	379000	15.557			

Raw Data: **BF142277.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MS	SDG No.:	Q1901
Lab Sample ID:	Q1901-08MS	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142277.D	1	04/30/25 13:15	05/02/25 12:27	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Sample Results: **BF142277.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142277.D
Acq On : 02 May 2025 12:27
Operator : RC/JU
Sample : Q1901-08MS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Quant Time: May 02 13:05:51 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	212656	20.000	ng	0.00
21) Naphthalene-d8	8.193	136	808361	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	402059	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	629997	20.000	ng	0.00
76) Chrysene-d12	14.069	240	370460	20.000	ng	0.00
86) Perylene-d12	15.557	264	379220	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.528	112	1285671	96.555	ng	0.01
7) Phenol-d6	6.528	99	1443545	89.776	ng	0.00
23) Nitrobenzene-d5	7.469	82	1033716	84.012	ng	0.00
42) 2,4,6-Tribromophenol	10.734	330	462379	126.149	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	2016863	70.728	ng	0.00
79) Terphenyl-d14	13.016	244	1792164	70.359	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.822	88	169976	28.583	ng	# 84
3) Pyridine	3.546	79	415695	26.896	ng	99
4) n-Nitrosodimethylamine	3.481	42	278483	34.818	ng	99
6) Aniline	6.569	93	185208	8.138	ng	100
8) 2-Chlorophenol	6.693	128	531955	38.662	ng	99
9) Benzaldehyde	6.457	77	175465	15.613	ng	99
10) Phenol	6.540	94	537381m	31.043	ng	
11) bis(2-Chloroethyl)ether	6.646	93	502092	37.430	ng	99
12) 1,3-Dichlorobenzene	6.851	146	511679	32.873	ng	99
13) 1,4-Dichlorobenzene	6.928	146	527665	33.647	ng	99
14) 1,2-Dichlorobenzene	7.081	146	509008	33.999	ng	99
15) Benzyl Alcohol	7.046	79	445258	39.044	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.181	45	916352	37.830	ng	99
17) 2-Methylphenol	7.151	107	403260	35.777	ng	100
18) Hexachloroethane	7.422	117	180036	35.075	ng	97
19) n-Nitroso-di-n-propyla...	7.316	70	379242	38.168	ng	98
20) 3+4-Methylphenols	7.304	107	508536	36.077	ng	91
22) Acetophenone	7.316	105	714285	38.195	ng	97
24) Nitrobenzene	7.487	77	532026	44.068	ng	100
25) Isophorone	7.728	82	1017548	39.404	ng	99
26) 2-Nitrophenol	7.804	139	249952	48.946	ng	98
27) 2,4-Dimethylphenol	7.834	122	465938	35.787	ng	100
28) bis(2-Chloroethoxy)met...	7.934	93	653302	40.114	ng	100
29) 2,4-Dichlorophenol	8.040	162	450542	41.390	ng	99
30) 1,2,4-Trichlorobenzene	8.128	180	460388	37.902	ng	100
31) Naphthalene	8.210	128	1515067	37.582	ng	100
32) Benzoic acid	7.922	122	217267	37.981	ng	97
33) 4-Chloroaniline	8.251	127	79798	4.754	ng	99
34) Hexachlorobutadiene	8.328	225	267842	37.794	ng	99
35) Caprolactam	8.616	113	114486m	34.992	ng	
36) 4-Chloro-3-methylphenol	8.728	107	450003	40.260	ng	99
37) 2-Methylnaphthalene	8.898	142	965025	38.934	ng	99
38) 1-Methylnaphthalene	8.998	142	1001608	39.008	ng	99
40) 1,2,4,5-Tetrachloroben...	9.069	216	466320	41.212	ng	99
41) Hexachlorocyclopentadiene	9.057	237	570339	88.811	ng	100
43) 2,4,6-Trichlorophenol	9.175	196	326161	46.517	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142277.D
Acq On : 02 May 2025 12:27
Operator : RC/JU
Sample : Q1901-08MS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Quant Time: May 02 13:05:51 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.210	196	336279	45.616	ng		99
46) 1,1'-Biphenyl	9.363	154	1292053	41.337	ng		99
47) 2-Chloronaphthalene	9.392	162	960343	41.296	ng		99
48) 2-Nitroaniline	9.481	65	294596	48.035	ng		98
49) Acenaphthylene	9.810	152	1605889	40.923	ng		100
50) Dimethylphthalate	9.663	163	1104923	42.957	ng		100
51) 2,6-Dinitrotoluene	9.722	165	230556	46.839	ng		99
52) Acenaphthene	9.981	154	1000886	43.682	ng		99
53) 3-Nitroaniline	9.886	138	93454	17.726	ng		98
54) 2,4-Dinitrophenol	9.992	184	156192	86.643	ng	#	52
55) Dibenzofuran	10.151	168	1420255	41.350	ng		100
56) 4-Nitrophenol	10.039	139	369705	91.443	ng		99
57) 2,4-Dinitrotoluene	10.128	165	301103	50.164	ng		96
58) Fluorene	10.492	166	1070926	41.007	ng		99
59) 2,3,4,6-Tetrachlorophenol	10.263	232	273427	44.776	ng		99
60) Diethylphthalate	10.363	149	1082961	42.636	ng		100
61) 4-Chlorophenyl-phenyle...	10.486	204	522609	42.138	ng		96
62) 4-Nitroaniline	10.504	138	228504	42.238	ng		97
63) Azobenzene	10.645	77	1037185	41.352	ng		99
65) 4,6-Dinitro-2-methylph...	10.534	198	106866	50.723	ng		90
66) n-Nitrosodiphenylamine	10.604	169	918104	42.652	ng		99
67) 4-Bromophenyl-phenylether	10.975	248	319645	45.559	ng		98
68) Hexachlorobenzene	11.039	284	349088	44.425	ng		98
69) Atrazine	11.128	200	274854	48.442	ng		99
70) Pentachlorophenol	11.233	266	390510	94.071	ng		98
71) Phenanthrene	11.457	178	1459064	42.876	ng		100
72) Anthracene	11.510	178	1476623	42.440	ng		99
73) Carbazole	11.657	167	1322084	41.655	ng		100
74) Di-n-butylphthalate	11.992	149	1508741	47.265	ng		100
75) Fluoranthene	12.645	202	1362829	40.513	ng		98
77) Benzidine	12.763	184	40553	2.576	ng		97
78) Pyrene	12.875	202	1353270	39.600	ng		100
80) Butylbenzylphthalate	13.492	149	486120	51.715	ng		99
81) Benzo(a)anthracene	14.057	228	1077111	43.310	ng		100
82) 3,3'-Dichlorobenzidine	14.022	252	61349	8.398	ng		98
83) Chrysene	14.098	228	946310	41.802	ng		100
84) Bis(2-ethylhexyl)phtha...	14.051	149	628324	48.509	ng		99
85) Di-n-octyl phthalate	14.669	149	1009277	43.232	ng		99
87) Indeno(1,2,3-cd)pyrene	17.068	276	1154061	41.548	ng		100
88) Benzo(b)fluoranthene	15.121	252	1089899	46.269	ng		99
89) Benzo(k)fluoranthene	15.151	252	857957	39.820	ng		99
90) Benzo(a)pyrene	15.498	252	936845	43.441	ng		99
91) Dibenzo(a,h)anthracene	17.086	278	933265	41.398	ng		100
92) Benzo(g,h,i)perylene	17.521	276	936160	41.307	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Sample Results: **BF142277.D**


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142277.D
Acq On : 02 May 2025 12:27
Operator : RC/JU
Sample : Q1901-08MS
Misc :
ALS Vial : 5 Sample Multiplier: 1


Quant Time: May 02 13:05:51 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:152 Resp: 21265


Ion	Ratio	Lower	Upper
152	100		
150	159.0	130.2	195.2
115	58.1	47.0	70.4



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#2
1,4-Dioxane
Concen: 28.583 ng
RT: 2.822 min Scan# 107
Delta R.T. 0.124 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion: 88 Resp: 169976

Ion	Ratio	Lower	Upper
88	100		
58	90.1	72.2	108.2
43	0.0	27.0	40.6#






#3
Pyridine
Concen: 26.896 ng
RT: 3.546 min Scan# 2
Delta R.T. 0.082 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion: 79 Resp: 41569


Ion	Ratio	Lower	Upper
79	100		
52	88.4	71.5	107.3
51	43.7	34.5	51.7



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#4
n-Nitrosodimethylamine
Concen: 34.818 ng
RT: 3.481 min Scan# 219
Delta R.T. 0.071 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion: 42 Resp: 278483

Ion	Ratio	Lower	Upper
42	100		
74	111.1	89.5	134.3
44	7.8	6.6	9.8





#5
2-Fluorophenol
Concen: 96.555 ng
RT: 5.528 min Scan# 5
Delta R.T. 0.012 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Tgt Ion:112 Resp: 128567


Ion	Ratio	Lower	Upper
112	100		
64	60.6	48.4	72.6
63	31.2	24.8	37.2

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#6
Aniline
Concen: 8.138 ng
RT: 6.569 min Scan# 744
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion: 93 Resp: 185208

Ion	Ratio	Lower	Upper
93	100		
66	40.4	32.1	48.1
65	20.7	16.6	24.8





#7
Phenol-d6
Concen: 89.776 ng
RT: 6.528 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Tgt Ion: 99 Resp: 144354


Ion	Ratio	Lower	Upper
99	100		
42	20.9	16.3	24.5
71	31.8	25.0	37.4

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#8
2-Chlorophenol
Concen: 38.662 ng
RT: 6.693 min Scan# 765
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:128 Resp: 531955

Ion	Ratio	Lower	Upper
128	100		
130	32.7	13.1	53.1
64	47.9	28.6	68.6






#9
Benzaldehyde
Concen: 15.613 ng
RT: 6.457 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion: 77 Resp: 17546


Ion	Ratio	Lower	Upper
77	100		
105	100.9	79.1	119.1
106	96.2	76.0	116.0



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#10
Phenol
Concen: 31.043 ng m
RT: 6.540 min Scan# 739
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion: 94 Resp: 537381

Ion	Ratio	Lower	Upper
94	100		
65	29.4	8.9	48.9
66	40.3	18.7	58.7





#11
bis(2-Chloroethyl)ether
Concen: 37.430 ng
RT: 6.646 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Tgt Ion: 93 Resp: 50209


Ion	Ratio	Lower	Upper
93	100		
63	81.5	62.2	102.2
95	32.2	12.7	52.7

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#12
1,3-Dichlorobenzene
Concen: 32.873 ng
RT: 6.851 min Scan# 792
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:146 Resp: 511679

Ion	Ratio	Lower	Upper
146	100		
148	64.1	50.5	75.7
75	28.9	23.2	34.8






#13
1,4-Dichlorobenzene
Concen: 33.647 ng
RT: 6.928 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:146 Resp: 52766


Ion	Ratio	Lower	Upper
146	100		
148	63.0	50.6	76.0
111	41.0	33.4	50.0



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#14
1,2-Dichlorobenzene
Concen: 33.999 ng
RT: 7.081 min Scan# 831
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:146 Resp: 509008

Ion	Ratio	Lower	Upper
146	100		
148	64.1	50.5	75.7
111	43.0	34.9	52.3





#15
Benzyl Alcohol
Concen: 39.044 ng
RT: 7.046 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Tgt Ion: 79 Resp: 44525


Ion	Ratio	Lower	Upper
79	100		
108	81.1	64.5	96.7
77	62.0	49.2	73.8

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#16
2,2'-oxybis(1-Chloropropane)
Concen: 37.830 ng
RT: 7.181 min Scan# 848
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion: 45 Resp: 916352

Ion	Ratio	Lower	Upper
45	100		
77	10.8	0.0	31.3
79	8.3	0.0	28.5






#17
2-Methylphenol
Concen: 35.777 ng
RT: 7.151 min Scan# 84
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:107 Resp: 40326


Ion	Ratio	Lower	Upper
107	100		
108	113.7	91.0	136.4
77	43.0	34.4	51.6
79	43.0	34.4	51.6



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#18
Hexachloroethane
Concen: 35.075 ng
RT: 7.422 min Scan# 889
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:117 Resp: 180036

Ion	Ratio	Lower	Upper
117	100		
119	96.1	77.1	115.7
201	94.4	79.4	119.2





#19
n-Nitroso-di-n-propylamine
Concen: 38.168 ng
RT: 7.316 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS


Tgt Ion: 70 Resp: 37924

Ion	Ratio	Lower	Upper
70	100		
42	69.0	53.4	80.2
101	9.9	7.7	11.5
130	21.4	17.2	25.8

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#20
3+4-Methylphenols
Concen: 36.077 ng
RT: 7.304 min Scan# 869
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:107 Resp: 508536

Ion	Ratio	Lower	Upper
107	100		
108	90.6	71.2	111.2
77	80.7	41.1	81.1
79	28.2	8.1	48.1






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.193 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:136 Resp: 80836


Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.7	13.1
54	9.2	7.7	11.5
68	6.2	5.0	7.6



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#22
Acetophenone
Concen: 38.195 ng
RT: 7.316 min Scan# 871
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:105 Resp: 714285

Ion	Ratio	Lower	Upper
105	100		
71	8.5	5.7	8.5
51	35.0	30.1	45.1
120	23.7	19.0	28.4





#23
Nitrobenzene-d5
Concen: 84.012 ng
RT: 7.469 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Tgt Ion: 82 Resp: 103371


Ion	Ratio	Lower	Upper
82	100		
128	46.1	37.0	55.6
54	60.7	48.0	72.0

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#24
Nitrobenzene
Concen: 44.068 ng
RT: 7.487 min Scan# 900
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion: 77 Resp: 532026

Ion	Ratio	Lower	Upper
77	100		
123	48.3	38.6	58.0
65	14.6	11.5	17.3





#25
Isophorone
Concen: 39.404 ng
RT: 7.728 min Scan# 94
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS


Tgt Ion: 82 Resp: 101754

Ion	Ratio	Lower	Upper
82	100		
95	7.1	5.6	8.4
138	19.9	15.3	22.9

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#26
2-Nitrophenol
Concen: 48.946 ng
RT: 7.804 min Scan# 954
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:139 Resp: 249952

Ion	Ratio	Lower	Upper
139	100		
109	27.3	21.8	32.8
65	49.3	41.1	61.7





#27
2,4-Dimethylphenol
Concen: 35.787 ng
RT: 7.834 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Tgt Ion:122 Resp: 46593


Ion	Ratio	Lower	Upper
122	100		
107	107.3	86.0	129.0
121	57.6	46.3	69.5

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#28
bis(2-Chloroethoxy)methane
Concen: 40.114 ng
RT: 7.934 min Scan# 976
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion: 93 Resp: 653302

Ion	Ratio	Lower	Upper
93	100		
95	32.7	26.2	39.2
123	12.5	9.7	14.5





#29
2,4-Dichlorophenol
Concen: 41.390 ng
RT: 8.040 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Tgt Ion:162 Resp: 45054


Ion	Ratio	Lower	Upper
162	100		
164	63.6	43.0	83.0
98	36.1	16.7	56.7

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#30
1,2,4-Trichlorobenzene
Concen: 37.902 ng
RT: 8.128 min Scan# 1009
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:180 Resp: 460388

Ion	Ratio	Lower	Upper
180	100		
182	94.7	75.7	113.5
145	31.6	25.4	38.0





#31
Naphthalene
Concen: 37.582 ng
RT: 8.210 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Tgt Ion:128 Resp: 151506


Ion	Ratio	Lower	Upper
128	100		
129	11.3	9.0	13.4
127	13.2	10.6	15.8

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#32
Benzoic acid
Concen: 37.981 ng
RT: 7.922 min Scan# 974
Delta R.T. -0.012 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:122 Resp: 217267

Ion	Ratio	Lower	Upper
122	100		
105	116.5	98.2	138.2
77	91.2	66.8	106.8






#33
4-Chloroaniline
Concen: 4.754 ng
RT: 8.251 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:127 Resp: 7979


Ion	Ratio	Lower	Upper
127	100		
129	31.7	26.1	39.1
65	31.0	25.5	38.3
92	19.6	15.4	23.2



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#34
Hexachlorobutadiene
Concen: 37.794 ng
RT: 8.328 min Scan# 1043
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:225 Resp: 267842

Ion	Ratio	Lower	Upper
225	100		
223	63.2	49.7	74.5
227	63.5	51.0	76.4





#35
Caprolactam
Concen: 34.992 ng m
RT: 8.616 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Tgt Ion:113 Resp: 11448


Ion	Ratio	Lower	Upper
113	100		
55	188.8	177.7	217.7
56	140.2	127.1	167.1

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#36
4-Chloro-3-methylphenol
Concen: 40.260 ng
RT: 8.728 min Scan# 1111
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:107 Resp: 450003

Ion	Ratio	Lower	Upper
107	100		
144	26.8	21.0	31.4
142	83.8	66.1	99.1





#37
2-Methylnaphthalene
Concen: 38.934 ng
RT: 8.898 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Tgt Ion:142 Resp: 96502


Ion	Ratio	Lower	Upper
142	100		
141	87.2	70.5	105.7

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#38
1-Methylnaphthalene
Concen: 39.008 ng
RT: 8.998 min Scan# 1157
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:142 Resp: 1001608

Ion	Ratio	Lower	Upper
142	100		
141	89.9	72.8	109.2






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.945 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:164 Resp: 40205


Ion	Ratio	Lower	Upper
164	100		
162	99.8	78.6	117.8
160	44.9	35.7	53.5



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 41.212 ng
RT: 9.069 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:216 Resp: 466320

Ion	Ratio	Lower	Upper
216	100		
214	78.7	63.8	95.6
179	20.9	16.7	25.1
108	20.5	15.6	23.4





#41
Hexachlorocyclopentadiene
Concen: 88.811 ng
RT: 9.057 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS


Tgt Ion:237 Resp: 57033

Ion	Ratio	Lower	Upper
237	100		
235	63.2	43.5	83.5
272	13.2	0.0	32.9

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#42
2,4,6-Tribromophenol
Concen: 126.149 ng
RT: 10.734 min Scan# 1452
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:330 Resp: 462379

Ion	Ratio	Lower	Upper
330	100		
332	96.5	76.5	114.7
141	29.6	24.3	36.5






#43
2,4,6-Trichlorophenol
Concen: 46.517 ng
RT: 9.175 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:196 Resp: 32616


Ion	Ratio	Lower	Upper
196	100		
198	94.4	77.4	116.2
200	30.6	24.6	37.0



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#44
2,4,5-Trichlorophenol
Concen: 45.616 ng
RT: 9.210 min Scan# 1193
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:196 Resp: 336279

Ion	Ratio	Lower	Upper
196	100		
198	95.4	75.2	112.8
97	49.4	39.8	59.8
132	29.9	23.8	35.8





#45
2-Fluorobiphenyl
Concen: 70.728 ng
RT: 9.263 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS


Tgt Ion:172 Resp: 201686

Ion	Ratio	Lower	Upper
172	100		
171	35.9	29.0	43.4
170	23.8	19.4	29.0

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#46
1,1'-Biphenyl
Concen: 41.337 ng
RT: 9.363 min Scan# 1219
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:154 Resp: 1292053

Ion	Ratio	Lower	Upper
154	100		
153	39.8	20.5	60.5
76	13.1	0.0	32.8






#47
2-Chloronaphthalene
Concen: 41.296 ng
RT: 9.392 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:162 Resp: 96034


Ion	Ratio	Lower	Upper
162	100		
127	40.9	32.6	49.0
164	33.3	26.1	39.1



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#48
2-Nitroaniline
Concen: 48.035 ng
RT: 9.481 min Scan# 1239
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion: 65 Resp: 294596

Ion	Ratio	Lower	Upper
65	100		
92	65.4	54.3	81.5
138	108.7	88.6	133.0





#49
Acenaphthylene
Concen: 40.923 ng
RT: 9.810 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Tgt Ion:152 Resp: 160588


Ion	Ratio	Lower	Upper
152	100		
151	20.3	16.2	24.2
153	13.3	10.7	16.1

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#50
Dimethylphthalate
Concen: 42.957 ng
RT: 9.663 min Scan# 1270
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:163 Resp: 1104923

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.2	4.8
164	10.0	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 46.839 ng
RT: 9.722 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Tgt Ion:165 Resp: 23055


Ion	Ratio	Lower	Upper
165	100		
63	62.0	48.5	72.7
89	54.3	43.9	65.9

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#52
Acenaphthene
Concen: 43.682 ng
RT: 9.981 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:154 Resp: 1000886

Ion	Ratio	Lower	Upper
154	100		
153	109.9	88.6	133.0
152	52.4	42.1	63.1





#53
3-Nitroaniline
Concen: 17.726 ng
RT: 9.886 min Scan# 1
Delta R.T. -0.005 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Tgt Ion:138 Resp: 9345


Ion	Ratio	Lower	Upper
138	100		
108	11.7	11.7	17.5
92	113.0	91.2	136.8

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#54
2,4-Dinitrophenol
Concen: 86.643 ng
RT: 9.992 min Scan# 1326
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:184 Resp: 156192

Ion	Ratio	Lower	Upper
184	100		
63	80.1	69.2	103.8
154	214.0	271.4	407.2#






#55
Dibenzofuran
Concen: 41.350 ng
RT: 10.151 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:168 Resp: 142025


Ion	Ratio	Lower	Upper
168	100		
139	36.7	29.4	44.0
169	13.5	11.0	16.4



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#56
4-Nitrophenol
Concen: 91.443 ng
RT: 10.039 min Scan# 1334
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:139 Resp: 369705

Ion	Ratio	Lower	Upper
139	100		
109	65.6	44.9	84.9
65	96.2	74.7	114.7






#57
2,4-Dinitrotoluene
Concen: 50.164 ng
RT: 10.128 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:165 Resp: 30110


Ion	Ratio	Lower	Upper
165	100		
63	46.0	38.8	58.2
89	67.2	57.1	85.7
182	2.5	2.2	3.2



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#58
Fluorene
Concen: 41.007 ng
RT: 10.492 min Scan# 1411
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:166 Resp: 1070926

Ion	Ratio	Lower	Upper
166	100		
165	93.0	73.9	110.9
167	13.7	10.8	16.2






#59
2,3,4,6-Tetrachlorophenol
Concen: 44.776 ng
RT: 10.263 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:232 Resp: 27342


Ion	Ratio	Lower	Upper
232	100		
131	45.5	35.9	53.9
130	2.1	1.8	2.6
166	29.5	24.0	36.0



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#60
Diethylphthalate
Concen: 42.636 ng
RT: 10.363 min Scan# 1389
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:149 Resp: 1082961

Ion	Ratio	Lower	Upper
149	100		
177	21.9	17.6	26.4
150	12.7	10.2	15.2





#61
4-Chlorophenyl-phenylether
Concen: 42.138 ng
RT: 10.486 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS


Tgt Ion:204 Resp: 52260

Ion	Ratio	Lower	Upper
204	100		
206	32.1	25.6	38.4
141	56.9	48.9	73.3

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#62
4-Nitroaniline
Concen: 42.238 ng
RT: 10.504 min Scan# 1413
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:138 Resp: 228504

Ion	Ratio	Lower	Upper
138	100		
92	47.9	29.2	69.2
108	66.2	49.1	89.1





#63
Azobenzene
Concen: 41.352 ng
RT: 10.645 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS


Tgt Ion: 77 Resp: 103718

Ion	Ratio	Lower	Upper
77	100		
182	26.8	7.3	47.3
105	22.5	2.5	42.5
51	35.0	14.4	54.4

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:188 Resp: 629997

Ion	Ratio	Lower	Upper
188	100		
94	9.8	8.2	12.2
80	10.1	8.6	12.8





#65
4,6-Dinitro-2-methylphenol
Concen: 50.723 ng
RT: 10.534 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS


Tgt Ion:198 Resp: 10686

Ion	Ratio	Lower	Upper
198	100		
51	52.8	40.6	80.6
105	42.9	29.2	69.2

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#66
n-Nitrosodiphenylamine
Concen: 42.652 ng
RT: 10.604 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:169 Resp: 918104

Ion	Ratio	Lower	Upper
169	100		
168	65.8	53.0	79.6
167	35.6	28.8	43.2






#67
4-Bromophenyl-phenylether
Concen: 45.559 ng
RT: 10.975 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:248 Resp: 31964


Ion	Ratio	Lower	Upper
248	100		
250	98.6	77.2	115.8
141	73.8	58.4	87.6



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#68
Hexachlorobenzene
Concen: 44.425 ng
RT: 11.039 min Scan# 1504
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:284 Resp: 349088

Ion	Ratio	Lower	Upper
284	100		
142	32.4	24.6	37.0
249	29.3	23.2	34.8






#69
Atrazine
Concen: 48.442 ng
RT: 11.128 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:200 Resp: 27485


Ion	Ratio	Lower	Upper
200	100		
173	28.4	8.6	48.6
215	49.6	28.5	68.5



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#70
Pentachlorophenol
Concen: 94.071 ng
RT: 11.233 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:266 Resp: 390510

Ion	Ratio	Lower	Upper
266	100		
268	63.6	50.0	75.0
264	62.9	49.3	73.9






#71
Phenanthrene
Concen: 42.876 ng
RT: 11.457 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:178 Resp: 145906


Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.5	23.3
179	15.9	12.6	19.0



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#72
Anthracene
Concen: 42.440 ng
RT: 11.510 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:178 Resp: 1476623

Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.3	22.9
179	15.6	12.7	19.1






#73
Carbazole
Concen: 41.655 ng
RT: 11.657 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:167 Resp: 132208


Ion	Ratio	Lower	Upper
167	100		
166	21.5	17.3	25.9
139	12.8	10.3	15.5



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#74
Di-n-butylphthalate
Concen: 47.265 ng
RT: 11.992 min Scan# 1666
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:149 Resp: 1508741

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.5	11.3
104	4.5	3.6	5.4






#75
Fluoranthene
Concen: 40.513 ng
RT: 12.645 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:202 Resp: 136282


Ion	Ratio	Lower	Upper
202	100		
101	11.3	0.0	32.4
203	17.8	0.0	37.5



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.069 min Scan# 2019
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:240 Resp: 370460

Ion	Ratio	Lower	Upper
240	100		
120	12.3	10.1	15.1
236	25.2	20.5	30.7






#77
Benzidine
Concen: 2.576 ng
RT: 12.763 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:184 Resp: 4055


Ion	Ratio	Lower	Upper
184	100		
185	15.8	11.8	17.6
183	10.0	8.9	13.3



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#78
Pyrene
Concen: 39.600 ng
RT: 12.875 min Scan# 1816
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:202 Resp: 1353270

Ion	Ratio	Lower	Upper
202	100		
200	20.0	16.1	24.1
203	17.7	14.1	21.1






#79
Terphenyl-d14
Concen: 70.359 ng
RT: 13.016 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:244 Resp: 179216


Ion	Ratio	Lower	Upper
244	100		
212	6.5	5.3	7.9
122	11.2	9.1	13.7



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#80
Butylbenzylphthalate
Concen: 51.715 ng
RT: 13.492 min Scan# 1921
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:149 Resp: 486120

Ion	Ratio	Lower	Upper
149	100		
91	68.1	54.4	81.6
206	20.8	15.9	23.9





#81
Benzo(a)anthracene
Concen: 43.310 ng
RT: 14.057 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS


Tgt Ion:228 Resp: 107711

Ion	Ratio	Lower	Upper
228	100		
226	27.5	21.9	32.9
229	20.1	16.0	24.0

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#82
3,3'-Dichlorobenzidine
Concen: 8.398 ng
RT: 14.022 min Scan# 2011
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:252 Resp: 61349

Ion	Ratio	Lower	Upper
252	100		
254	65.6	51.0	76.6
126	12.2	9.4	14.2






#83
Chrysene
Concen: 41.802 ng
RT: 14.098 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:228 Resp: 94631


Ion	Ratio	Lower	Upper
228	100		
226	30.1	24.0	36.0
229	19.7	15.7	23.5



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 48.509 ng
RT: 14.051 min Scan# 2016
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:149 Resp: 628324

Ion	Ratio	Lower	Upper
149	100		
167	27.7	21.7	32.5
279	4.1	3.0	4.4





#85
Di-n-octyl phthalate
Concen: 43.232 ng
RT: 14.669 min Scan# 2
Delta R.T. -0.011 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Tgt Ion:149 Resp: 100927


Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.1	1.7
43	11.3	9.3	13.9

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.557 min Scan# 2272
Delta R.T. -0.012 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:264 Resp: 379220

Ion	Ratio	Lower	Upper
264	100		
260	23.2	18.4	27.6
265	21.4	17.3	25.9






#87
Indeno(1,2,3-cd)pyrene
Concen: 41.548 ng
RT: 17.068 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:276 Resp: 115406


Ion	Ratio	Lower	Upper
276	100		
138	28.1	22.6	33.8
277	25.0	20.1	30.1



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#88
Benzo(b)fluoranthene
Concen: 46.269 ng
RT: 15.121 min Scan# 2198
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:252 Resp: 1089899

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.7	26.5
125	10.3	9.0	13.4






#89
Benzo(k)fluoranthene
Concen: 39.820 ng
RT: 15.151 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:252 Resp: 85795


Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.8	26.8
125	10.7	9.0	13.6



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#90
Benzo(a)pyrene
Concen: 43.441 ng
RT: 15.498 min Scan# 2262
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:252 Resp: 936845

Ion	Ratio	Lower	Upper
252	100		
253	22.4	17.6	26.4
125	11.4	9.5	14.3





#91
Dibenzo(a,h)anthracene
Concen: 41.398 ng
RT: 17.086 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS


Tgt Ion:278 Resp: 93326

Ion	Ratio	Lower	Upper
278	100		
139	18.6	15.2	22.8
279	23.8	19.0	28.6

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025






#92
Benzo(g,h,i)perylene
Concen: 41.307 ng
RT: 17.521 min Scan# 2606
Delta R.T. -0.018 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:276 Resp: 936160

Ion	Ratio	Lower	Upper
276	100		
277	24.1	18.7	28.1
138	24.6	19.4	29.0



Raw Data: **BF142278.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MSD	SDG No.:	Q1901
Lab Sample ID:	Q1901-08MSD	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142278.D	1	04/30/25 13:15	05/02/25 12:55	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	280		12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	340		5.30	50.0	ug/L
95-48-7	2-Methylphenol	370		11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	370		11.0	100	ug/L
67-72-1	Hexachloroethane	370		6.50	50.0	ug/L
98-95-3	Nitrobenzene	460		7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	390		5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	480		5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	470		6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	530		12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	450		5.20	50.0	ug/L
87-86-5	Pentachlorophenol	970	E	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	99.4		15 (10) - 110 (139)	66%	SPK: 150
13127-88-3	Phenol-d6	92.8		15 (10) - 110 (134)	62%	SPK: 150
4165-60-0	Nitrobenzene-d5	87.8		30 (49) - 130 (133)	88%	SPK: 100
321-60-8	2-Fluorobiphenyl	72.3		30 (52) - 130 (132)	72%	SPK: 100
118-79-6	2,4,6-Tribromophenol	133		15 (44) - 110 (137)	88%	SPK: 150
1718-51-0	Terphenyl-d14	72.8		30 (48) - 130 (125)	73%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	217000	6.91			
1146-65-2	Naphthalene-d8	833000	8.193			
15067-26-2	Acenaphthene-d10	419000	9.945			
1517-22-2	Phenanthrene-d10	684000	11.433			
1719-03-5	Chrysene-d12	400000	14.069			
1520-96-3	Perylene-d12	401000	15.563			

Raw Data: **BF142278.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025			Date Received:	04/28/25
Client Sample ID:	B-167-SB01MSD			SDG No.:	Q1901
Lab Sample ID:	Q1901-08MSD			Matrix:	TCLP
Analytical Method:	SW8270			% Solid:	0
Sample Wt/Vol:	100	Units:	mL	Final Vol:	1000 uL
Soil Aliquot Vol:			uL	Test:	TCLP BNA
Extraction Type :			Decanted : N	Level :	LOW
Injection Volume :			GPC Factor : 1.0	GPC Cleanup : N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142278.D	1	04/30/25 13:15	05/02/25 12:55	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Sample Results: **BF142278.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142278.D
Acq On : 02 May 2025 12:55
Operator : RC/JU
Sample : Q1901-08MSD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: May 02 13:41:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	217028	20.000	ng	0.00
21) Naphthalene-d8	8.193	136	832936	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	419407	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	684254	20.000	ng	0.00
76) Chrysene-d12	14.069	240	400079	20.000	ng	0.00
86) Perylene-d12	15.563	264	401381	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.528	112	1351308	99.440	ng	0.01
7) Phenol-d6	6.528	99	1523396	92.834	ng	0.00
23) Nitrobenzene-d5	7.469	82	1112775	87.769	ng	0.00
42) 2,4,6-Tribromophenol	10.734	330	506639	132.507	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	2151303	72.323	ng	0.00
79) Terphenyl-d14	13.016	244	2003750	72.842	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.828	88	174625	28.773	ng	# 82
3) Pyridine	3.546	79	441222	27.973	ng	100
4) n-Nitrosodimethylamine	3.487	42	294114	36.032	ng	99
6) Aniline	6.569	93	197398	8.499	ng	98
8) 2-Chlorophenol	6.693	128	562936	40.089	ng	99
9) Benzaldehyde	6.457	77	183950	16.038	ng	99
10) Phenol	6.540	94	571858m	32.370	ng	
11) bis(2-Chloroethyl)ether	6.646	93	532809	38.919	ng	99
12) 1,3-Dichlorobenzene	6.851	146	542334	34.140	ng	99
13) 1,4-Dichlorobenzene	6.928	146	548211	34.253	ng	99
14) 1,2-Dichlorobenzene	7.081	146	540561	35.379	ng	99
15) Benzyl Alcohol	7.046	79	478246	41.092	ng	100
16) 2,2'-oxybis(1-Chloropr...	7.181	45	960997	38.874	ng	100
17) 2-Methylphenol	7.151	107	427556	37.168	ng	99
18) Hexachloroethane	7.422	117	192593	36.766	ng	96
19) n-Nitroso-di-n-propyla...	7.322	70	400784	39.523	ng	99
20) 3+4-Methylphenols	7.304	107	538223	37.414	ng	92
22) Acetophenone	7.316	105	748030	38.819	ng	97
24) Nitrobenzene	7.487	77	571114	45.910	ng	100
25) Isophorone	7.728	82	1075681	40.426	ng	99
26) 2-Nitrophenol	7.804	139	277477	51.929	ng	98
27) 2,4-Dimethylphenol	7.834	122	495291	36.919	ng	100
28) bis(2-Chloroethoxy)met...	7.934	93	685392	40.843	ng	99
29) 2,4-Dichlorophenol	8.040	162	483547	43.112	ng	99
30) 1,2,4-Trichlorobenzene	8.128	180	482951	38.586	ng	99
31) Naphthalene	8.210	128	1609412	38.745	ng	100
32) Benzoic acid	7.928	122	263320	42.769	ng	97
33) 4-Chloroaniline	8.251	127	72531	4.193	ng	99
34) Hexachlorobutadiene	8.328	225	287976	39.436	ng	99
35) Caprolactam	8.616	113	125300m	37.167	ng	
36) 4-Chloro-3-methylphenol	8.728	107	484815	42.095	ng	99
37) 2-Methylnaphthalene	8.904	142	1012785	39.656	ng	100
38) 1-Methylnaphthalene	8.998	142	1065609	40.276	ng	99
40) 1,2,4,5-Tetrachloroben...	9.069	216	493589	41.817	ng	99
41) Hexachlorocyclopentadiene	9.057	237	611613	91.299	ng	99
43) 2,4,6-Trichlorophenol	9.175	196	351147	48.009	ng	99

Sample Results: **BF142278.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142278.D
Acq On : 02 May 2025 12:55
Operator : RC/JU
Sample : Q1901-08MSD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: May 02 13:41:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.210	196	361688	47.034	ng		99
46) 1,1'-Biphenyl	9.363	154	1368430	41.970	ng		100
47) 2-Chloronaphthalene	9.392	162	1020867	42.083	ng		99
48) 2-Nitroaniline	9.481	65	319718	49.782	ng		98
49) Acenaphthylene	9.810	152	1710079	41.776	ng		100
50) Dimethylphthalate	9.663	163	1199810	44.716	ng		100
51) 2,6-Dinitrotoluene	9.722	165	253926	49.228	ng		99
52) Acenaphthene	9.981	154	1087078	45.482	ng		98
53) 3-Nitroaniline	9.886	138	99043	17.958	ng		98
54) 2,4-Dinitrophenol	9.998	184	184644	93.898	ng	#	1
55) Dibenzofuran	10.151	168	1522541	42.495	ng		100
56) 4-Nitrophenol	10.045	139	407185	96.548	ng		99
57) 2,4-Dinitrotoluene	10.128	165	336232	53.345	ng		96
58) Fluorene	10.492	166	1144424	42.008	ng		99
59) 2,3,4,6-Tetrachlorophenol	10.263	232	303708	47.678	ng		99
60) Diethylphthalate	10.369	149	1182771	44.640	ng		99
61) 4-Chlorophenyl-phenyle...	10.486	204	556055	42.980	ng		97
62) 4-Nitroaniline	10.504	138	257712	45.444	ng		98
63) Azobenzene	10.645	77	1119990	42.806	ng		99
65) 4,6-Dinitro-2-methylph...	10.534	198	125145	53.859	ng		94
66) n-Nitrosodiphenylamine	10.604	169	983989	42.088	ng		99
67) 4-Bromophenyl-phenylether	10.975	248	347963	45.662	ng		99
68) Hexachlorobenzene	11.039	284	382262	44.789	ng		97
69) Atrazine	11.128	200	312642	50.733	ng		99
70) Pentachlorophenol	11.233	266	437319	96.993	ng		99
71) Phenanthrene	11.457	178	1591587	43.062	ng		100
72) Anthracene	11.510	178	1599092	42.316	ng		100
73) Carbazole	11.657	167	1465976	42.526	ng		100
74) Di-n-butylphthalate	11.992	149	1677931	48.397	ng		100
75) Fluoranthene	12.645	202	1520599	41.619	ng		98
77) Benzidine	12.763	184	44894	2.641	ng		97
78) Pyrene	12.875	202	1524967	41.321	ng		100
80) Butylbenzylphthalate	13.492	149	558469	54.776	ng		99
81) Benzo(a)anthracene	14.063	228	1194007	44.456	ng		100
82) 3,3'-Dichlorobenzidine	14.022	252	70246	8.904	ng		99
83) Chrysene	14.098	228	1076915	44.049	ng		99
84) Bis(2-ethylhexyl)phtha...	14.051	149	722253	51.410	ng		99
85) Di-n-octyl phthalate	14.669	149	1156306	45.494	ng		99
87) Indeno(1,2,3-cd)pyrene	17.068	276	1219041	41.464	ng		100
88) Benzo(b)fluoranthene	15.121	252	1168207	46.856	ng		99
89) Benzo(k)fluoranthene	15.151	252	987950	43.322	ng		100
90) Benzo(a)pyrene	15.498	252	1023759	44.850	ng		100
91) Dibenzo(a,h)anthracene	17.092	278	988821	41.441	ng		99
92) Benzo(g,h,i)perylene	17.527	276	978093	40.775	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142278.D
Acq On : 02 May 2025 12:55
Operator : RC/JU
Sample : Q1901-08MSD
Misc :
ALS Vial : 6 Sample Multiplier: 1


Quant Time: May 02 13:41:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:152 Resp: 21702


Ion	Ratio	Lower	Upper
152	100		
150	159.7	130.2	195.2
115	57.6	47.0	70.4



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#2
1,4-Dioxane
Concen: 28.773 ng
RT: 2.828 min Scan# 108
Delta R.T. 0.129 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion: 88 Resp: 174625

Ion	Ratio	Lower	Upper
88	100		
58	92.4	72.2	108.2
43	0.0	27.0	40.6#





#3
Pyridine
Concen: 27.973 ng
RT: 3.546 min Scan# 2
Delta R.T. 0.082 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion: 79 Resp: 44122


Ion	Ratio	Lower	Upper
79	100		
52	89.1	71.5	107.3
51	42.8	34.5	51.7

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#4
n-Nitrosodimethylamine
Concen: 36.032 ng
RT: 3.487 min Scan# 220
Delta R.T. 0.076 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion: 42 Resp: 294114

Ion	Ratio	Lower	Upper
42	100		
74	112.6	89.5	134.3
44	7.7	6.6	9.8





#5
2-Fluorophenol
Concen: 99.440 ng
RT: 5.528 min Scan# 5
Delta R.T. 0.012 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion:112 Resp: 135130


Ion	Ratio	Lower	Upper
112	100		
64	61.0	48.4	72.6
63	31.8	24.8	37.2

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#6
Aniline
Concen: 8.499 ng
RT: 6.569 min Scan# 744
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion: 93 Resp: 197398

Ion	Ratio	Lower	Upper
93	100		
66	41.3	32.1	48.1
65	21.0	16.6	24.8






#7
Phenol-d6
Concen: 92.834 ng
RT: 6.528 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion: 99 Resp: 152339


Ion	Ratio	Lower	Upper
99	100		
42	21.0	16.3	24.5
71	32.1	25.0	37.4



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#8
2-Chlorophenol
Concen: 40.089 ng
RT: 6.693 min Scan# 765
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:128 Resp: 562936

Ion	Ratio	Lower	Upper
128	100		
130	32.6	13.1	53.1
64	48.3	28.6	68.6






#9
Benzaldehyde
Concen: 16.038 ng
RT: 6.457 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion: 77 Resp: 18395


Ion	Ratio	Lower	Upper
77	100		
105	99.6	79.1	119.1
106	96.6	76.0	116.0



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#10
Phenol
Concen: 32.370 ng m
RT: 6.540 min Scan# 739
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion: 94 Resp: 571858

Ion	Ratio	Lower	Upper
94	100		
65	29.7	8.9	48.9
66	41.7	18.7	58.7





#11
bis(2-Chloroethyl)ether
Concen: 38.919 ng
RT: 6.646 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion: 93 Resp: 53280


Ion	Ratio	Lower	Upper
93	100		
63	81.3	62.2	102.2
95	31.8	12.7	52.7

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#12
1,3-Dichlorobenzene
Concen: 34.140 ng
RT: 6.851 min Scan# 792
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:146 Resp: 542334

Ion	Ratio	Lower	Upper
146	100		
148	63.9	50.5	75.7
75	28.8	23.2	34.8






#13
1,4-Dichlorobenzene
Concen: 34.253 ng
RT: 6.928 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:146 Resp: 54821


Ion	Ratio	Lower	Upper
146	100		
148	64.0	50.6	76.0
111	42.1	33.4	50.0



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#14
1,2-Dichlorobenzene
Concen: 35.379 ng
RT: 7.081 min Scan# 831
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:146 Resp: 540561

Ion	Ratio	Lower	Upper
146	100		
148	63.6	50.5	75.7
111	42.7	34.9	52.3





#15
Benzyl Alcohol
Concen: 41.092 ng
RT: 7.046 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion: 79 Resp: 47824


Ion	Ratio	Lower	Upper
79	100		
108	80.7	64.5	96.7
77	62.2	49.2	73.8

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#16
2,2'-oxybis(1-Chloropropane)
Concen: 38.874 ng
RT: 7.181 min Scan# 848
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion: 45 Resp: 960997

Ion	Ratio	Lower	Upper
45	100		
77	11.0	0.0	31.3
79	8.4	0.0	28.5






#17
2-Methylphenol
Concen: 37.168 ng
RT: 7.151 min Scan# 84
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:107 Resp: 42755


Ion	Ratio	Lower	Upper
107	100		
108	114.4	91.0	136.4
77	43.5	34.4	51.6
79	42.8	34.4	51.6



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#18
Hexachloroethane
Concen: 36.766 ng
RT: 7.422 min Scan# 889
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:117 Resp: 192593

Ion	Ratio	Lower	Upper
117	100		
119	94.7	77.1	115.7
201	93.5	79.4	119.2






#19
n-Nitroso-di-n-propylamine
Concen: 39.523 ng
RT: 7.322 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion: 70 Resp: 40078


Ion	Ratio	Lower	Upper
70	100		
42	67.2	53.4	80.2
101	9.9	7.7	11.5
130	23.0	17.2	25.8



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#20
3+4-Methylphenols
Concen: 37.414 ng
RT: 7.304 min Scan# 869
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:107 Resp: 538223

Ion	Ratio	Lower	Upper
107	100		
108	91.5	71.2	111.2
77	78.2	41.1	81.1
79	28.1	8.1	48.1






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.193 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:136 Resp: 83293


Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.7	13.1
54	9.3	7.7	11.5
68	6.0	5.0	7.6



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#22
Acetophenone
Concen: 38.819 ng
RT: 7.316 min Scan# 871
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:105 Resp: 748030

Ion	Ratio	Lower	Upper
105	100		
71	8.1	5.7	8.5
51	35.4	30.1	45.1
120	24.0	19.0	28.4





#23
Nitrobenzene-d5
Concen: 87.769 ng
RT: 7.469 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion: 82 Resp: 111277


Ion	Ratio	Lower	Upper
82	100		
128	46.9	37.0	55.6
54	60.6	48.0	72.0

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#24
Nitrobenzene
Concen: 45.910 ng
RT: 7.487 min Scan# 900
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion: 77 Resp: 571114

Ion	Ratio	Lower	Upper
77	100		
123	48.0	38.6	58.0
65	14.6	11.5	17.3





#25
Isophorone
Concen: 40.426 ng
RT: 7.728 min Scan# 94
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion: 82 Resp: 107568


Ion	Ratio	Lower	Upper
82	100		
95	7.2	5.6	8.4
138	19.8	15.3	22.9

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#26
2-Nitrophenol
Concen: 51.929 ng
RT: 7.804 min Scan# 954
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:139 Resp: 277477

Ion	Ratio	Lower	Upper
139	100		
109	27.3	21.8	32.8
65	48.8	41.1	61.7





#27
2,4-Dimethylphenol
Concen: 36.919 ng
RT: 7.834 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion:122 Resp: 49529


Ion	Ratio	Lower	Upper
122	100		
107	107.9	86.0	129.0
121	58.0	46.3	69.5

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#28
bis(2-Chloroethoxy)methane
Concen: 40.843 ng
RT: 7.934 min Scan# 976
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion: 93 Resp: 685392

Ion	Ratio	Lower	Upper
93	100		
95	33.0	26.2	39.2
123	13.0	9.7	14.5






#29
2,4-Dichlorophenol
Concen: 43.112 ng
RT: 8.040 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:162 Resp: 48354


Ion	Ratio	Lower	Upper
162	100		
164	64.0	43.0	83.0
98	35.6	16.7	56.7



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#30
1,2,4-Trichlorobenzene
Concen: 38.586 ng
RT: 8.128 min Scan# 1009
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:180 Resp: 482951

Ion	Ratio	Lower	Upper
180	100		
182	95.4	75.7	113.5
145	32.3	25.4	38.0





#31
Naphthalene
Concen: 38.745 ng
RT: 8.210 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion:128 Resp: 160941


Ion	Ratio	Lower	Upper
128	100		
129	11.5	9.0	13.4
127	13.3	10.6	15.8

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#32
Benzoic acid
Concen: 42.769 ng
RT: 7.928 min Scan# 975
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:122 Resp: 263320

Ion	Ratio	Lower	Upper
122	100		
105	119.9	98.2	138.2
77	91.8	66.8	106.8






#33
4-Chloroaniline
Concen: 4.193 ng
RT: 8.251 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:127 Resp: 7253


Ion	Ratio	Lower	Upper
127	100		
129	31.9	26.1	39.1
65	31.0	25.5	38.3
92	19.2	15.4	23.2



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#34
Hexachlorobutadiene
Concen: 39.436 ng
RT: 8.328 min Scan# 1043
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:225 Resp: 287976

Ion	Ratio	Lower	Upper
225	100		
223	62.7	49.7	74.5
227	62.9	51.0	76.4





#35
Caprolactam
Concen: 37.167 ng m
RT: 8.616 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion:113 Resp: 12530


Ion	Ratio	Lower	Upper
113	100		
55	197.1	177.7	217.7
56	142.9	127.1	167.1

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#36
4-Chloro-3-methylphenol
Concen: 42.095 ng
RT: 8.728 min Scan# 1111
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:107 Resp: 484815

Ion	Ratio	Lower	Upper
107	100		
144	26.8	21.0	31.4
142	83.9	66.1	99.1





#37
2-Methylnaphthalene
Concen: 39.656 ng
RT: 8.904 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion:142 Resp: 101278


Ion	Ratio	Lower	Upper
142	100		
141	88.1	70.5	105.7

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#38
1-Methylnaphthalene
Concen: 40.276 ng
RT: 8.998 min Scan# 1157
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:142 Resp: 1065609

Ion	Ratio	Lower	Upper
142	100		
141	90.2	72.8	109.2






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.945 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:164 Resp: 41940


Ion	Ratio	Lower	Upper
164	100		
162	98.5	78.6	117.8
160	44.5	35.7	53.5



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 41.817 ng
RT: 9.069 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:216 Resp: 493589

Ion	Ratio	Lower	Upper
216	100		
214	79.0	63.8	95.6
179	21.2	16.7	25.1
108	20.6	15.6	23.4





#41
Hexachlorocyclopentadiene
Concen: 91.299 ng
RT: 9.057 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD


Tgt Ion:237 Resp: 61161

Ion	Ratio	Lower	Upper
237	100		
235	63.0	43.5	83.5
272	13.1	0.0	32.9

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#42
2,4,6-Tribromophenol
Concen: 132.507 ng
RT: 10.734 min Scan# 1452
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:330 Resp: 506639

Ion	Ratio	Lower	Upper
330	100		
332	97.3	76.5	114.7
141	29.5	24.3	36.5






#43
2,4,6-Trichlorophenol
Concen: 48.009 ng
RT: 9.175 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:196 Resp: 35114


Ion	Ratio	Lower	Upper
196	100		
198	95.1	77.4	116.2
200	30.8	24.6	37.0



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#44
2,4,5-Trichlorophenol
Concen: 47.034 ng
RT: 9.210 min Scan# 1193
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:196 Resp: 361688

Ion	Ratio	Lower	Upper
196	100		
198	94.8	75.2	112.8
97	50.1	39.8	59.8
132	30.5	23.8	35.8






#45
2-Fluorobiphenyl
Concen: 72.323 ng
RT: 9.263 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:172 Resp: 215130


Ion	Ratio	Lower	Upper
172	100		
171	35.5	29.0	43.4
170	23.8	19.4	29.0



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#46
1,1'-Biphenyl
Concen: 41.970 ng
RT: 9.363 min Scan# 1219
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:154 Resp: 1368430

Ion	Ratio	Lower	Upper
154	100		
153	40.4	20.5	60.5
76	12.9	0.0	32.8






#47
2-Chloronaphthalene
Concen: 42.083 ng
RT: 9.392 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:162 Resp: 102086


Ion	Ratio	Lower	Upper
162	100		
127	40.5	32.6	49.0
164	33.1	26.1	39.1



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#48
2-Nitroaniline
Concen: 49.782 ng
RT: 9.481 min Scan# 1239
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion: 65 Resp: 319718

Ion	Ratio	Lower	Upper
65	100		
92	67.2	54.3	81.5
138	108.2	88.6	133.0





#49
Acenaphthylene
Concen: 41.776 ng
RT: 9.810 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion:152 Resp: 171007


Ion	Ratio	Lower	Upper
152	100		
151	20.4	16.2	24.2
153	13.3	10.7	16.1

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#50
Dimethylphthalate
Concen: 44.716 ng
RT: 9.663 min Scan# 1270
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:163 Resp: 1199810

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.2	4.8
164	10.4	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 49.228 ng
RT: 9.722 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion:165 Resp: 25392


Ion	Ratio	Lower	Upper
165	100		
63	60.1	48.5	72.7
89	53.9	43.9	65.9

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#52
Acenaphthene
Concen: 45.482 ng
RT: 9.981 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:154 Resp: 1087078

Ion	Ratio	Lower	Upper
154	100		
153	108.5	88.6	133.0
152	52.1	42.1	63.1





#53
3-Nitroaniline
Concen: 17.958 ng
RT: 9.886 min Scan# 1
Delta R.T. -0.005 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion:138 Resp: 9904


Ion	Ratio	Lower	Upper
138	100		
108	12.6	11.7	17.5
92	112.3	91.2	136.8

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#54
2,4-Dinitrophenol
Concen: 93.898 ng
RT: 9.998 min Scan# 1327
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:184 Resp: 184644

Ion	Ratio	Lower	Upper
184	100		
63	65.1	69.2	103.8#
154	79.6	271.4	407.2#






#55
Dibenzofuran
Concen: 42.495 ng
RT: 10.151 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:168 Resp: 152254


Ion	Ratio	Lower	Upper
168	100		
139	36.5	29.4	44.0
169	13.6	11.0	16.4



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#56
4-Nitrophenol
Concen: 96.548 ng
RT: 10.045 min Scan# 1335
Delta R.T. 0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:139 Resp: 407185

Ion	Ratio	Lower	Upper
139	100		
109	64.3	44.9	84.9
65	93.1	74.7	114.7






#57
2,4-Dinitrotoluene
Concen: 53.345 ng
RT: 10.128 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:165 Resp: 33623


Ion	Ratio	Lower	Upper
165	100		
63	46.4	38.8	58.2
89	66.8	57.1	85.7
182	2.8	2.2	3.2



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#58
Fluorene
Concen: 42.008 ng
RT: 10.492 min Scan# 1411
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:166 Resp: 1144424

Ion	Ratio	Lower	Upper
166	100		
165	93.0	73.9	110.9
167	13.8	10.8	16.2






#59
2,3,4,6-Tetrachlorophenol
Concen: 47.678 ng
RT: 10.263 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:232 Resp: 30370


Ion	Ratio	Lower	Upper
232	100		
131	44.5	35.9	53.9
130	2.1	1.8	2.6
166	29.4	24.0	36.0



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#60
Diethylphthalate
Concen: 44.640 ng
RT: 10.369 min Scan# 1390
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:149 Resp: 1182771

Ion	Ratio	Lower	Upper
149	100		
177	22.3	17.6	26.4
150	12.5	10.2	15.2





#61
4-Chlorophenyl-phenylether
Concen: 42.980 ng
RT: 10.486 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD


Tgt Ion:204 Resp: 55605

Ion	Ratio	Lower	Upper
204	100		
206	32.9	25.6	38.4
141	58.0	48.9	73.3

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#62
4-Nitroaniline
Concen: 45.444 ng
RT: 10.504 min Scan# 1413
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:138 Resp: 257712

Ion	Ratio	Lower	Upper
138	100		
92	48.3	29.2	69.2
108	66.7	49.1	89.1






#63
Azobenzene
Concen: 42.806 ng
RT: 10.645 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion: 77 Resp: 111999


Ion	Ratio	Lower	Upper
77	100		
182	27.0	7.3	47.3
105	22.6	2.5	42.5
51	34.9	14.4	54.4



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:188 Resp: 684254

Ion	Ratio	Lower	Upper
188	100		
94	9.7	8.2	12.2
80	10.1	8.6	12.8





#65
4,6-Dinitro-2-methylphenol
Concen: 53.859 ng
RT: 10.534 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion:198 Resp: 12514


Ion	Ratio	Lower	Upper
198	100		
51	55.4	40.6	80.6
105	45.3	29.2	69.2

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#66
n-Nitrosodiphenylamine
Concen: 42.088 ng
RT: 10.604 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:169 Resp: 983989

Ion	Ratio	Lower	Upper
169	100		
168	66.7	53.0	79.6
167	36.4	28.8	43.2






#67
4-Bromophenyl-phenylether
Concen: 45.662 ng
RT: 10.975 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:248 Resp: 34796


Ion	Ratio	Lower	Upper
248	100		
250	98.4	77.2	115.8
141	73.4	58.4	87.6



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#68
Hexachlorobenzene
Concen: 44.789 ng
RT: 11.039 min Scan# 1504
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:284 Resp: 382262

Ion	Ratio	Lower	Upper
284	100		
142	32.4	24.6	37.0
249	30.4	23.2	34.8





#69
Atrazine
Concen: 50.733 ng
RT: 11.128 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD


Tgt Ion:200 Resp: 31264

Ion	Ratio	Lower	Upper
200	100		
173	28.0	8.6	48.6
215	48.8	28.5	68.5

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#70
Pentachlorophenol
Concen: 96.993 ng
RT: 11.233 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:266 Resp: 437319

Ion	Ratio	Lower	Upper
266	100		
268	63.5	50.0	75.0
264	62.5	49.3	73.9






#71
Phenanthrene
Concen: 43.062 ng
RT: 11.457 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:178 Resp: 159158


Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.5	23.3
179	15.7	12.6	19.0



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#72
Anthracene
Concen: 42.316 ng
RT: 11.510 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:178 Resp: 1599092

Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.3	22.9
179	15.7	12.7	19.1






#73
Carbazole
Concen: 42.526 ng
RT: 11.657 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:167 Resp: 146597


Ion	Ratio	Lower	Upper
167	100		
166	21.5	17.3	25.9
139	13.1	10.3	15.5



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#74
Di-n-butylphthalate
Concen: 48.397 ng
RT: 11.992 min Scan# 1666
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:149 Resp: 1677931

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.5	11.3
104	4.5	3.6	5.4






#75
Fluoranthene
Concen: 41.619 ng
RT: 12.645 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:202 Resp: 152059


Ion	Ratio	Lower	Upper
202	100		
101	11.2	0.0	32.4
203	18.0	0.0	37.5



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.069 min Scan# 2019
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:240 Resp: 400079

Ion	Ratio	Lower	Upper
240	100		
120	12.6	10.1	15.1
236	25.2	20.5	30.7





#77
Benzidine
Concen: 2.641 ng
RT: 12.763 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion:184 Resp: 4489


Ion	Ratio	Lower	Upper
184	100		
185	16.7	11.8	17.6
183	11.2	8.9	13.3

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#78
Pyrene
Concen: 41.321 ng
RT: 12.875 min Scan# 1816
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:202 Resp: 1524967

Ion	Ratio	Lower	Upper
202	100		
200	20.2	16.1	24.1
203	17.6	14.1	21.1






#79
Terphenyl-d14
Concen: 72.842 ng
RT: 13.016 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:244 Resp: 200375


Ion	Ratio	Lower	Upper
244	100		
212	6.6	5.3	7.9
122	11.4	9.1	13.7



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#80
Butylbenzylphthalate
Concen: 54.776 ng
RT: 13.492 min Scan# 1921
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:149 Resp: 558469

Ion	Ratio	Lower	Upper
149	100		
91	68.6	54.4	81.6
206	20.7	15.9	23.9






#81
Benzo(a)anthracene
Concen: 44.456 ng
RT: 14.063 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:228 Resp: 119400


Ion	Ratio	Lower	Upper
228	100		
226	27.2	21.9	32.9
229	19.9	16.0	24.0



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#82
3,3'-Dichlorobenzidine
Concen: 8.904 ng
RT: 14.022 min Scan# 2011
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:252 Resp: 70246

Ion	Ratio	Lower	Upper
252	100		
254	64.2	51.0	76.6
126	12.1	9.4	14.2





#83
Chrysene
Concen: 44.049 ng
RT: 14.098 min Scan# 2(
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD


Tgt Ion:228 Resp: 107691

Ion	Ratio	Lower	Upper
228	100		
226	30.4	24.0	36.0
229	20.0	15.7	23.5

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025







#84
Bis(2-ethylhexyl)phthalate
Concen: 51.410 ng
RT: 14.051 min Scan# 2016
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:149 Resp: 722253

Ion	Ratio	Lower	Upper
149	100		
167	27.6	21.7	32.5
279	4.1	3.0	4.4






#85
Di-n-octyl phthalate
Concen: 45.494 ng
RT: 14.669 min Scan# 2
Delta R.T. -0.011 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:149 Resp: 115630


Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.1	1.7
43	11.3	9.3	13.9



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 15.563 min Scan# 2273
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:264 Resp: 401381

Ion	Ratio	Lower	Upper
264	100		
260	23.1	18.4	27.6
265	21.7	17.3	25.9






#87
Indeno(1,2,3-cd)pyrene
Concen: 41.464 ng
RT: 17.068 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:276 Resp: 121904


Ion	Ratio	Lower	Upper
276	100		
138	27.9	22.6	33.8
277	25.0	20.1	30.1



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#88
Benzo(b)fluoranthene
Concen: 46.856 ng
RT: 15.121 min Scan# 2198
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:252 Resp: 1168207

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.7	26.5
125	10.5	9.0	13.4






#89
Benzo(k)fluoranthene
Concen: 43.322 ng
RT: 15.151 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:252 Resp: 98795


Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.8	26.8
125	11.1	9.0	13.6



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#90
Benzo(a)pyrene
Concen: 44.850 ng
RT: 15.498 min Scan# 2262
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:252 Resp: 1023759

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.6	26.4
125	11.7	9.5	14.3






#91
Dibenzo(a,h)anthracene
Concen: 41.441 ng
RT: 17.092 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:278 Resp: 98882


Ion	Ratio	Lower	Upper
278	100		
139	18.0	15.2	22.8
279	24.0	19.0	28.6



Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#92
Benzo(g,h,i)perylene
Concen: 40.775 ng
RT: 17.527 min Scan# 2607
Delta R.T. -0.012 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:276 Resp: 978093

Ion	Ratio	Lower	Upper
276	100		
277	23.0	18.7	28.1
138	23.9	19.4	29.0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BF043025	Instrument	BNA_f

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDICC005	BF142240.D	2,3,4,6-Tetrachlorophenol	Rahul	5/1/2025 11:26:17 AM	Jagrut	5/1/2025 1:17:43 PM	Peak Integrated by Software
SSTDICC020	BF142242.D	Benzoic acid	Rahul	5/1/2025 11:26:23 AM	Jagrut	5/1/2025 1:17:45 PM	Peak Integrated by Software
SSTDICCC040	BF142243.D	Benzoic acid	Rahul	5/1/2025 11:26:25 AM	Jagrut	5/1/2025 1:17:48 PM	Peak Integrated by Software
SSTDICC050	BF142244.D	Phenol	Rahul	5/1/2025 11:26:27 AM	Jagrut	5/1/2025 1:17:50 PM	Peak Integrated by Software
SSTDICC060	BF142245.D	Phenol	Rahul	5/1/2025 11:26:30 AM	Jagrut	5/1/2025 1:17:52 PM	Peak Integrated by Software
SSTDICC080	BF142246.D	Benzoic acid	Rahul	5/1/2025 11:26:33 AM	Jagrut	5/1/2025 1:17:55 PM	Peak Integrated by Software
SSTDICC080	BF142246.D	Phenol	Rahul	5/1/2025 11:26:33 AM	Jagrut	5/1/2025 1:17:55 PM	Peak Integrated by Software
SSTDICV040	BF142247.D	Benzoic acid	Rahul	5/1/2025 11:26:36 AM	Jagrut	5/1/2025 1:17:57 PM	Peak Integrated by Software
SSTDICV040	BF142247.D	Phenol	Rahul	5/1/2025 11:26:36 AM	Jagrut	5/1/2025 1:17:57 PM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	bf050125	Instrument	BNA_f

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDCCC040	BF142250.D	Benzoic acid	Rahul	5/2/2025 10:45:14 AM	Jagrut	5/2/2025 4:47:49 PM	Peak Integrated by Software
SSTDCCC040	BF142250.D	Phenol	Rahul	5/2/2025 10:45:14 AM	Jagrut	5/2/2025 4:47:49 PM	Peak Integrated by Software
PB167810BS	BF142255.D	Benzoic acid	Rahul	5/2/2025 10:45:25 AM	Jagrut	5/2/2025 4:47:56 PM	Peak Integrated by Software
PB167810BS	BF142255.D	Caprolactam	Rahul	5/2/2025 10:45:25 AM	Jagrut	5/2/2025 4:47:56 PM	Peak Integrated by Software
PB167810BS	BF142255.D	Phenol	Rahul	5/2/2025 10:45:25 AM	Jagrut	5/2/2025 4:47:56 PM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	bf050225	Instrument	BNA_f

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDCCC040	BF142274.D	Benzoic acid	Rahul	5/5/2025 10:14:38 AM	Jagrut	5/5/2025 11:51:58 AM	Peak Integrated by Software
SSTDCCC040	BF142274.D	Phenol	Rahul	5/5/2025 10:14:38 AM	Jagrut	5/5/2025 11:51:58 AM	Peak Integrated by Software
Q1901-08MS	BF142277.D	Caprolactam	Rahul	5/5/2025 10:14:41 AM	Jagrut	5/5/2025 11:52:00 AM	Peak Integrated by Software
Q1901-08MS	BF142277.D	Phenol	Rahul	5/5/2025 10:14:41 AM	Jagrut	5/5/2025 11:52:00 AM	Peak Integrated by Software
Q1901-08MSD	BF142278.D	Caprolactam	Rahul	5/5/2025 10:14:44 AM	Jagrut	5/5/2025 11:52:03 AM	Peak Integrated by Software
Q1901-08MSD	BF142278.D	Phenol	Rahul	5/5/2025 10:14:44 AM	Jagrut	5/5/2025 11:52:03 AM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	bp041425	Instrument	BNA_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDICC005	BP024276.D	Benzaldehyde	Rahul	4/15/2025 9:45:41 AM	Jagrut	4/15/2025 10:52:38 AM	Peak Integrated by Software
SSTDICC005	BP024276.D	Benzidine	Rahul	4/15/2025 9:45:41 AM	Jagrut	4/15/2025 10:52:38 AM	Peak Integrated by Software
SSTDICC010	BP024277.D	Benzaldehyde	Rahul	4/15/2025 9:46:12 AM	Jagrut	4/15/2025 10:52:41 AM	Peak Integrated by Software
SSTDICC010	BP024277.D	Benzoic acid	Rahul	4/15/2025 9:46:12 AM	Jagrut	4/15/2025 10:52:41 AM	Peak Integrated by Software
SSTDICC020	BP024278.D	Benzoic acid	Rahul	4/15/2025 9:46:28 AM	Jagrut	4/15/2025 10:52:44 AM	Peak Integrated by Software
SSTDICC050	BP024280.D	Indeno(1,2,3-cd)pyrene	Rahul	4/15/2025 9:47:02 AM	Jagrut	4/15/2025 10:52:46 AM	Peak Integrated by Software
SSTDICC050	BP024280.D	Pyridine	Rahul	4/15/2025 9:47:02 AM	Jagrut	4/15/2025 10:52:46 AM	Peak Integrated by Software
SSTDICC060	BP024281.D	Benzoic acid	Rahul	4/15/2025 9:47:31 AM	Jagrut	4/15/2025 10:52:48 AM	Peak Integrated by Software
SSTDICC060	BP024281.D	Indeno(1,2,3-cd)pyrene	Rahul	4/15/2025 9:47:31 AM	Jagrut	4/15/2025 10:52:48 AM	Peak Integrated by Software
SSTDICC060	BP024281.D	Pyridine	Rahul	4/15/2025 9:47:31 AM	Jagrut	4/15/2025 10:52:48 AM	Peak Integrated by Software
SSTDICC080	BP024282.D	Benzidine	Rahul	4/15/2025 9:48:04 AM	Jagrut	4/15/2025 10:52:51 AM	Peak Integrated by Software
SSTDICC080	BP024282.D	Benzoic acid	Rahul	4/15/2025 9:48:04 AM	Jagrut	4/15/2025 10:52:51 AM	Peak Integrated by Software
SSTDICC080	BP024282.D	Pyridine	Rahul	4/15/2025 9:48:04 AM	Jagrut	4/15/2025 10:52:51 AM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	bp041425	Instrument	BNA_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDICV040	BP024283.D	Indeno(1,2,3-cd)pyrene	Rahul	4/15/2025 9:50:13 AM	Jagrut	4/15/2025 10:52:54 AM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BP050525	Instrument	BNA_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDCCC040	BP024538.D	Benzidine	Rahul	5/6/2025 9:57:47 AM	Jagrut	5/6/2025 10:50:38 AM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF043025

Review By	Rahul	Review On	5/1/2025 1:20:20 PM				
Supervise By	Jagrut	Supervise On	5/1/2025 1:23:08 PM				
SubDirectory	BF043025	HP Acquire Method	BNA_F		HP Processing Method	BF043025	

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12661,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF142238.D	30 Apr 2025 10:55	RC/JU	Ok
2	SSTDICC2.5	BF142239.D	30 Apr 2025 11:24	RC/JU	Ok
3	SSTDICC005	BF142240.D	30 Apr 2025 11:52	RC/JU	Ok,M
4	SSTDICC010	BF142241.D	30 Apr 2025 12:20	RC/JU	Ok
5	SSTDICC020	BF142242.D	30 Apr 2025 12:49	RC/JU	Ok,M
6	SSTDICCC040	BF142243.D	30 Apr 2025 13:17	RC/JU	Ok,M
7	SSTDICC050	BF142244.D	30 Apr 2025 13:46	RC/JU	Ok,M
8	SSTDICC060	BF142245.D	30 Apr 2025 14:15	RC/JU	Ok,M
9	SSTDICC080	BF142246.D	30 Apr 2025 14:43	RC/JU	Ok,M
10	SSTDICV040	BF142247.D	30 Apr 2025 16:17	RC/JU	Ok,M
11	PB167726BL	BF142248.D	30 Apr 2025 17:15	RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF050125

Review By	Rahul	Review On	5/2/2025 10:46:38 AM				
Supervise By	Jagrut	Supervise On	5/2/2025 4:49:12 PM				
SubDirectory	BF050125	HP Acquire Method	BNA_F		HP Processing Method	BF043025	

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12661,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF142249.D	01 May 2025 09:48	RC/JU	Ok
2	SSTDCCC040	BF142250.D	01 May 2025 10:17	RC/JU	Ok,M
3	PB167798BL	BF142251.D	01 May 2025 10:45	RC/JU	Ok
4	PB167798BS	BF142252.D	01 May 2025 11:13	RC/JU	Ok,M
5	PB167798BSD	BF142253.D	01 May 2025 11:42	RC/JU	Ok,M
6	PB167810BL	BF142254.D	01 May 2025 12:10	RC/JU	Ok
7	PB167810BS	BF142255.D	01 May 2025 12:39	RC/JU	Ok,M
8	PB167774TB	BF142256.D	01 May 2025 13:07	RC/JU	Ok
9	Q1890-01	BF142257.D	01 May 2025 13:40	RC/JU	Ok
10	Q1920-01	BF142258.D	01 May 2025 14:09	RC/JU	Ok,M
11	Q1914-14	BF142259.D	01 May 2025 14:37	RC/JU	Ok
12	Q1913-04	BF142260.D	01 May 2025 15:06	RC/JU	Ok
13	Q1915-01	BF142261.D	01 May 2025 15:34	RC/JU	Ok
14	Q1903-01	BF142262.D	01 May 2025 16:03	RC/JU	Ok
15	Q1903-02	BF142263.D	01 May 2025 16:31	RC/JU	Ok
16	Q1903-03	BF142264.D	01 May 2025 17:00	RC/JU	Ok
17	Q1906-01	BF142265.D	01 May 2025 17:29	RC/JU	Ok
18	Q1906-13	BF142266.D	01 May 2025 17:57	RC/JU	Ok
19	Q1901-06	BF142267.D	01 May 2025 18:26	RC/JU	Ok
20	Q1901-02	BF142268.D	01 May 2025 18:55	RC/JU	Ok,M
21	Q1901-02MS	BF142269.D	01 May 2025 19:24	RC/JU	Ok,M

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF050125

Review By	Rahul	Review On	5/2/2025 10:46:38 AM				
Supervise By	Jagrut	Supervise On	5/2/2025 4:49:12 PM				
SubDirectory	BF050125	HP Acquire Method	BNA_F		HP Processing Method	BF043025	

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12661,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	Q1901-02MSD	BF142270.D	01 May 2025 19:52	RC/JU	Ok,M
23	Q1901-04	BF142271.D	01 May 2025 20:20	RC/JU	Ok
24	Q1901-05	BF142272.D	01 May 2025 20:49	RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF050225

Review By	Rahul	Review On	5/5/2025 10:18:55 AM				
Supervise By	Jagrut	Supervise On	5/5/2025 11:52:39 AM				
SubDirectory	BF050225	HP Acquire Method	BNA_F		HP Processing Method	BF043025	

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12662,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF142273.D	02 May 2025 10:11	RC/JU	Ok
2	SSTDCCC040	BF142274.D	02 May 2025 10:39	RC/JU	Ok,M
3	PB167774TB	BF142275.D	02 May 2025 11:07	RC/JU	Not Ok
4	Q1901-08	BF142276.D	02 May 2025 11:58	RC/JU	Ok
5	Q1901-08MS	BF142277.D	02 May 2025 12:27	RC/JU	Ok,M
6	Q1901-08MSD	BF142278.D	02 May 2025 12:55	RC/JU	Ok,M
7	Q1901-09	BF142279.D	02 May 2025 13:24	RC/JU	Ok
8	Q1901-10	BF142280.D	02 May 2025 13:53	RC/JU	ReRun
9	Q1901-11	BF142281.D	02 May 2025 14:22	RC/JU	Ok
10	Q1907-02	BF142282.D	02 May 2025 14:51	RC/JU	ReRun
11	Q1916-04	BF142283.D	02 May 2025 16:20	RC/JU	Ok
12	Q1916-04MS	BF142284.D	02 May 2025 16:49	RC/JU	Ok,M
13	Q1916-04MSD	BF142285.D	02 May 2025 17:18	RC/JU	Ok,M
14	Q1932-02	BF142286.D	02 May 2025 17:47	RC/JU	Ok
15	Q1932-04	BF142287.D	02 May 2025 18:16	RC/JU	Ok
16	Q1932-06	BF142288.D	02 May 2025 18:46	RC/JU	ReRun
17	Q1932-08	BF142289.D	02 May 2025 19:15	RC/JU	Ok
18	Q1917-04	BF142290.D	02 May 2025 19:44	RC/JU	Ok
19	Q1922-04	BF142291.D	02 May 2025 20:13	RC/JU	Ok
20	Q1922-08	BF142292.D	02 May 2025 20:42	RC/JU	Ok
21	Q1929-03	BF142293.D	02 May 2025 21:11	RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP041425

Review By	Rahul	Review On	4/15/2025 9:55:40 AM		
Supervise By	Jagrut	Supervise On	4/15/2025 10:53:04 AM		
SubDirectory	BP041425	HP Acquire Method	BNA_P	HP Processing Method	bp041425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12659,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BP024274.D	14 Apr 2025 10:25	RC/JU	Ok
2	SSTDICC2.5	BP024275.D	14 Apr 2025 11:06	RC/JU	Ok
3	SSTDICC005	BP024276.D	14 Apr 2025 11:47	RC/JU	Ok,M
4	SSTDICC010	BP024277.D	14 Apr 2025 12:27	RC/JU	Ok,M
5	SSTDICC020	BP024278.D	14 Apr 2025 13:08	RC/JU	Ok,M
6	SSTDICCC040	BP024279.D	14 Apr 2025 13:49	RC/JU	Ok
7	SSTDICC050	BP024280.D	14 Apr 2025 15:10	RC/JU	Ok,M
8	SSTDICC060	BP024281.D	14 Apr 2025 16:32	RC/JU	Ok,M
9	SSTDICC080	BP024282.D	14 Apr 2025 17:13	RC/JU	Ok,M
10	SSTDICV040	BP024283.D	14 Apr 2025 18:35	RC/JU	Ok,M
11	PB167488TB	BP024284.D	14 Apr 2025 19:16	RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP050525

Review By	Rahul	Review On	5/6/2025 9:58:11 AM				
Supervise By	Jagrut	Supervise On	5/6/2025 10:50:58 AM				
SubDirectory	BP050525	HP Acquire Method	BNA_P		HP Processing Method	bp041425	

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12662,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BP024521.D	05 May 2025 11:57	RC/JU	Ok
2	SSTDCCC040	BP024522.D	05 May 2025 12:38	RC/JU	Ok
3	PB167836BL	BP024523.D	05 May 2025 13:18	RC/JU	Ok
4	PB167836BS	BP024524.D	05 May 2025 13:59	RC/JU	Ok
5	Q1883-05RX	BP024525.D	05 May 2025 14:43	RC/JU	Confirms
6	Q1928-01MS	BP024526.D	05 May 2025 15:24	RC/JU	Ok
7	Q1928-01MSD	BP024527.D	05 May 2025 16:05	RC/JU	Ok,M
8	Q1935-01	BP024528.D	05 May 2025 16:45	RC/JU	Ok
9	Q1901-10RE	BP024529.D	05 May 2025 17:26	RC/JU	Confirms
10	Q1907-02RE	BP024530.D	05 May 2025 18:07	RC/JU	Confirms
11	Q1932-06RE	BP024531.D	05 May 2025 18:47	RC/JU	Confirms
12	Q1929-07	BP024532.D	05 May 2025 19:28	RC/JU	Ok
13	Q1929-11	BP024533.D	05 May 2025 20:09	RC/JU	Ok
14	Q1935-04	BP024534.D	05 May 2025 20:49	RC/JU	Ok
15	Q1935-08	BP024535.D	05 May 2025 21:30	RC/JU	Ok
16	Q1928-04	BP024536.D	05 May 2025 22:11	RC/JU	Ok
17	Q1946-01	BP024537.D	05 May 2025 22:52	RC/JU	Ok
18	SSTDCCC040	BP024538.D	05 May 2025 23:32	RC/JU	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF043025

Review By	Rahul	Review On	5/1/2025 1:20:20 PM		
Supervise By	Jagrut	Supervise On	5/1/2025 1:23:08 PM		
SubDirectory	BF043025	HP Acquire Method	BNA_F	HP Processing Method	BF043025

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12661,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142238.D	30 Apr 2025 10:55		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF142239.D	30 Apr 2025 11:24		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF142240.D	30 Apr 2025 11:52	Compound#32,54,56,65,70,85 removed from 5 ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BF142241.D	30 Apr 2025 12:20		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF142242.D	30 Apr 2025 12:49		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BF142243.D	30 Apr 2025 13:17	Compound#32,48,51,53,57,62,65,80,84,85 Kept on LR	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BF142244.D	30 Apr 2025 13:46	Compound#26,54 Kept on QR	RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BF142245.D	30 Apr 2025 14:15		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BF142246.D	30 Apr 2025 14:43		RC/JU	Ok,M
10	SSTDICV040	ICVBF043025	BF142247.D	30 Apr 2025 16:17		RC/JU	Ok,M
11	PB167726BL	PB167726BL	BF142248.D	30 Apr 2025 17:15		RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF050125

Review By	Rahul	Review On	5/2/2025 10:46:38 AM		
Supervise By	Jagrut	Supervise On	5/2/2025 4:49:12 PM		
SubDirectory	BF050125	HP Acquire Method	BNA_F	HP Processing Method	BF043025

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12661,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142249.D	01 May 2025 09:48		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142250.D	01 May 2025 10:17		RC/JU	Ok,M
3	PB167798BL	PB167798BL	BF142251.D	01 May 2025 10:45		RC/JU	Ok
4	PB167798BS	PB167798BS	BF142252.D	01 May 2025 11:13		RC/JU	Ok,M
5	PB167798BSD	PB167798BSD	BF142253.D	01 May 2025 11:42		RC/JU	Ok,M
6	PB167810BL	PB167810BL	BF142254.D	01 May 2025 12:10		RC/JU	Ok
7	PB167810BS	PB167810BS	BF142255.D	01 May 2025 12:39		RC/JU	Ok,M
8	PB167774TB	PB167774TB	BF142256.D	01 May 2025 13:07		RC/JU	Ok
9	Q1890-01	TW-WTS-07	BF142257.D	01 May 2025 13:40		RC/JU	Ok
10	Q1920-01	291	BF142258.D	01 May 2025 14:09		RC/JU	Ok,M
11	Q1914-14	FB04282025	BF142259.D	01 May 2025 14:37		RC/JU	Ok
12	Q1913-04	WC-13-A-202504	BF142260.D	01 May 2025 15:06		RC/JU	Ok
13	Q1915-01	WC-04282025	BF142261.D	01 May 2025 15:34		RC/JU	Ok
14	Q1903-01	COMP-4	BF142262.D	01 May 2025 16:03		RC/JU	Ok
15	Q1903-02	COMP-5	BF142263.D	01 May 2025 16:31		RC/JU	Ok
16	Q1903-03	COMP-6	BF142264.D	01 May 2025 17:00		RC/JU	Ok
17	Q1906-01	WC-4	BF142265.D	01 May 2025 17:29		RC/JU	Ok
18	Q1906-13	WC-7	BF142266.D	01 May 2025 17:57		RC/JU	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF050125

Review By	Rahul	Review On	5/2/2025 10:46:38 AM		
Supervise By	Jagrut	Supervise On	5/2/2025 4:49:12 PM		
SubDirectory	BF050125	HP Acquire Method	BNA_F	HP Processing Method	BF043025

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12661,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	Q1901-06	FB04262025	BF142267.D	01 May 2025 18:26		RC/JU	Ok
20	Q1901-02	B-167-SB01	BF142268.D	01 May 2025 18:55		RC/JU	Ok,M
21	Q1901-02MS	B-167-SB01MS	BF142269.D	01 May 2025 19:24		RC/JU	Ok,M
22	Q1901-02MSD	B-167-SB01MSD	BF142270.D	01 May 2025 19:52		RC/JU	Ok,M
23	Q1901-04	B-167-SB02	BF142271.D	01 May 2025 20:20		RC/JU	Ok
24	Q1901-05	B-170-SB02	BF142272.D	01 May 2025 20:49		RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF050225

Review By	Rahul	Review On	5/5/2025 10:18:55 AM		
Supervise By	Jagrut	Supervise On	5/5/2025 11:52:39 AM		
SubDirectory	BF050225	HP Acquire Method	BNA_F	HP Processing Method	BF043025

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12662,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142273.D	02 May 2025 10:11		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142274.D	02 May 2025 10:39		RC/JU	Ok,M
3	PB167774TB	PB167774TB	BF142275.D	02 May 2025 11:07	Analyzed for contamination check	RC/JU	Not Ok
4	Q1901-08	B-167-SB01	BF142276.D	02 May 2025 11:58		RC/JU	Ok
5	Q1901-08MS	B-167-SB01MS	BF142277.D	02 May 2025 12:27		RC/JU	Ok,M
6	Q1901-08MSD	B-167-SB01MSD	BF142278.D	02 May 2025 12:55		RC/JU	Ok,M
7	Q1901-09	B-170-SB01	BF142279.D	02 May 2025 13:24		RC/JU	Ok
8	Q1901-10	B-167-SB02	BF142280.D	02 May 2025 13:53	Internal Standard Fail	RC/JU	ReRun
9	Q1901-11	B-170-SB02	BF142281.D	02 May 2025 14:22		RC/JU	Ok
10	Q1907-02	CO-8R-WC	BF142282.D	02 May 2025 14:51	Internal Standard and surrogates Fail	RC/JU	ReRun
11	Q1916-04	WC-12	BF142283.D	02 May 2025 16:20		RC/JU	Ok
12	Q1916-04MS	WC-12MS	BF142284.D	02 May 2025 16:49		RC/JU	Ok,M
13	Q1916-04MSD	WC-12MSD	BF142285.D	02 May 2025 17:18		RC/JU	Ok,M
14	Q1932-02	COMP-4	BF142286.D	02 May 2025 17:47		RC/JU	Ok
15	Q1932-04	COMP-5	BF142287.D	02 May 2025 18:16		RC/JU	Ok
16	Q1932-06	COMP-6	BF142288.D	02 May 2025 18:46	Surrogate Fail	RC/JU	ReRun
17	Q1932-08	COMP-7	BF142289.D	02 May 2025 19:15		RC/JU	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF050225

Review By	Rahul	Review On	5/5/2025 10:18:55 AM		
Supervise By	Jagrut	Supervise On	5/5/2025 11:52:39 AM		
SubDirectory	BF050225	HP Acquire Method	BNA_F	HP Processing Method	BF043025

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12662,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	Q1917-04	MH-JJ	BF142290.D	02 May 2025 19:44		RC/JU	Ok
19	Q1922-04	MH-R	BF142291.D	02 May 2025 20:13		RC/JU	Ok
20	Q1922-08	MH-S	BF142292.D	02 May 2025 20:42		RC/JU	Ok
21	Q1929-03	WC-A4-02-C	BF142293.D	02 May 2025 21:11		RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP041425

Review By	Rahul	Review On	4/15/2025 9:55:40 AM		
Supervise By	Jagrut	Supervise On	4/15/2025 10:53:04 AM		
SubDirectory	BP041425	HP Acquire Method	BNA_P	HP Processing Method	bp041425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12659,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024274.D	14 Apr 2025 10:25		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP024275.D	14 Apr 2025 11:06		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP024276.D	14 Apr 2025 11:47	Compound #32,54,65 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BP024277.D	14 Apr 2025 12:27		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BP024278.D	14 Apr 2025 13:08		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BP024279.D	14 Apr 2025 13:49	This calibration is fail for Com#77	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BP024280.D	14 Apr 2025 15:10	This calibration is good for 8270E, 8270 DOD and 625.1 methods except Compound#77	RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BP024281.D	14 Apr 2025 16:32	Compound #69 Removed from 60PPM	RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BP024282.D	14 Apr 2025 17:13	Compound #69 Removed from 80PPM	RC/JU	Ok,M
10	SSTDICV040	ICVBP041425	BP024283.D	14 Apr 2025 18:35		RC/JU	Ok,M
11	PB167488TB	PB167488TB	BP024284.D	14 Apr 2025 19:16		RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP050525

Review By	Rahul	Review On	5/6/2025 9:58:11 AM			
Supervise By	Jagrut	Supervise On	5/6/2025 10:50:58 AM			
SubDirectory	BP050525	HP Acquire Method	BNA_P	HP Processing Method		bp041425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12662,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024521.D	05 May 2025 11:57		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP024522.D	05 May 2025 12:38		RC/JU	Ok
3	PB167836BL	PB167836BL	BP024523.D	05 May 2025 13:18		RC/JU	Ok
4	PB167836BS	PB167836BS	BP024524.D	05 May 2025 13:59		RC/JU	Ok
5	Q1883-05RX	OU4-PCS-TC-29-04232	BP024525.D	05 May 2025 14:43	Surrogate Fail	RC/JU	Confirms
6	Q1928-01MS	MH-QMS	BP024526.D	05 May 2025 15:24		RC/JU	Ok
7	Q1928-01MSD	MH-QMSD	BP024527.D	05 May 2025 16:05		RC/JU	Ok,M
8	Q1935-01	MH-NN	BP024528.D	05 May 2025 16:45		RC/JU	Ok
9	Q1901-10RE	B-167-SB02RE	BP024529.D	05 May 2025 17:26	Internal Standard and surrogate Fail	RC/JU	Confirms
10	Q1907-02RE	CO-8R-WCRE	BP024530.D	05 May 2025 18:07	Internal Standard and surrogate Fail	RC/JU	Confirms
11	Q1932-06RE	COMP-6RE	BP024531.D	05 May 2025 18:47	Surrogate Fail	RC/JU	Confirms
12	Q1929-07	WC-A1-03-C	BP024532.D	05 May 2025 19:28		RC/JU	Ok
13	Q1929-11	WC-A1-04-C	BP024533.D	05 May 2025 20:09		RC/JU	Ok
14	Q1935-04	MH-NN	BP024534.D	05 May 2025 20:49		RC/JU	Ok
15	Q1935-08	MH-MM	BP024535.D	05 May 2025 21:30		RC/JU	Ok
16	Q1928-04	MH-Q	BP024536.D	05 May 2025 22:11		RC/JU	Ok
17	Q1946-01	BB-CONCRETE	BP024537.D	05 May 2025 22:52		RC/JU	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP050525

Review By	Rahul	Review On	5/6/2025 9:58:11 AM		
Supervise By	Jagrut	Supervise On	5/6/2025 10:50:58 AM		
SubDirectory	BP050525	HP Acquire Method	BNA_P	HP Processing Method	bp041425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12662,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	SSTDCCC040	SSTDCCC040EC	BP024538.D	05 May 2025 23:32		RC/JU	Ok,M

M : Manual Integration

Alliance TECHNICAL GROUP

TCLP EXTRACTION LOGPAGE

PB167774

SOP ID : M1311-TCLP-15

SDG No : N/A

Weigh By : JP

Balance ID : WC SC-7

pH Meter ID : WC PH METER-1

Extraction By : JP

Filter By : JP

Pippete ID : WC

Tumbler ID : T-1 / T-2

TCLP Filter ID : 115525

Start Prep Date : 04/29/2025 **Time :** 17:20

End Prep Date : 04/30/2025 **Time :** 11:30

Combination Ratio : 20

ZHE Cleaning Batch : ~~N/A~~ 10

Initial Room Temperature: 23 °C

Final Room Temperature: 22 °C

TCLP Technician Signature : JP

Supervisor By : JZ

Standared Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
TCLP-FLUID-1	N/A	WP112795
HCL-TCLP,1N	N/A	WP112797
HNO3-TCLP,1N	N/A	WP112799
pH Strips	N/A	W1931,W1934,W3171,W3172
pH Strips	W1940,W1941,W1942	W3166,W1938,W1939,
1 Liter Amber	N/A	90924-08
120ml Plastic bottle	N/A	2738
1:1 HNO3	N/A	MP84041

Extraction Conformance/Non-Conformance Comments:

Matrix spikes are added after fiftration and before preservation. TUMBLER T-1 /T-2 checked,30 rpm. Particle size reduction is not required. q1894-01,02 and 03 we receive limited volume so no fluid determation. Q1901-08 IS USED FOR MS-MSD.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
04/30/25 13:00	JP Lab Room	SLG. RJ/ GH
	Preparation Group	Analysis Group MTN,JG

Alliance TECHNICAL GROUP

TCLP EXTRACTION LOGPAGE

PB167774

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB167774TB	LEB774	21	N/A	2000	N/A	N/A	N/A	4.93	1.0	T-2
Q1894-01	PAINTED-GROUND-PIPE--1	01	25.00	500	N/A	N/A	N/A	5.5	1.0	T-1
Q1894-02	PAINTED-GROUND-PIPE-2	02	35.00	700	N/A	N/A	N/A	5.6	1.5	T-1
Q1894-03	PAINTED-GROUND-PIPE-3	03	34.00	680	N/A	N/A	N/A	5.6	1.5	T-1
Q1901-08	B-167-SB01	04	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-1
Q1901-09	B-170-SB01	05	100.03	2000	N/A	N/A	N/A	7.0	1.5	T-1
Q1901-10	B-167-SB02	06	100.04	2000	N/A	N/A	N/A	5.8	1.0	T-1
Q1901-11	B-170-SB02	07	100.02	2000	N/A	N/A	N/A	5.6	1.5	T-1
Q1904-02	VNJ-210	08	100.03	2000	N/A	N/A	N/A	6.0	1.0	T-1
Q1905-04	MH-G	09	100.02	2000	N/A	N/A	N/A	3.0	1.0	T-1
Q1905-08	MH-H	10	100.01	2000	N/A	N/A	N/A	3.0	1.5	T-1
Q1906-04	WC-4	11	100.02	2000	N/A	N/A	N/A	5.5	1.0	T-2
Q1906-08	WC-5	12	100.03	2000	N/A	N/A	N/A	5.6	1.5	T-2
Q1906-12	WC-6	13	100.02	2000	N/A	N/A	N/A	5.0	1.0	T-2
Q1906-16	WC-7	14	100.03	2000	N/A	N/A	N/A	5.0	1.5	T-2
Q1907-02	CO-8R-WC	15	100.02	2000	N/A	N/A	N/A	5.8	1.5	T-2
Q1912-04	MH-E	16	100.03	2000	N/A	N/A	N/A	4.5	1.0	T-2
Q1912-08	MH-F	17	100.02	2000	N/A	N/A	N/A	4.0	1.5	T-2
Q1913-01	WC-12-S-202504	18	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-2
Q1913-03	WC-13-S-202504	19	100.03	2000	N/A	N/A	N/A	6.2	1.5	T-2
Q1915-01	WC-04282025	20	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-2

Alliance TECHNICAL GROUP

TCLP Solid Determination

PB167774

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
PB167774TB	LEB774	N/A	N/A	N/A	N/A	N/A	N/A
Q1894-01	PAINTED-GROUND-PIPE--1	N/A	N/A	N/A	N/A	100	N/A
Q1894-02	PAINTED-GROUND-PIPE-2	N/A	N/A	N/A	N/A	100	N/A
Q1894-03	PAINTED-GROUND-PIPE-3	N/A	N/A	N/A	N/A	100	N/A
Q1901-08	B-167-SB01	N/A	N/A	N/A	N/A	100	N/A
Q1901-09	B-170-SB01	N/A	N/A	N/A	N/A	100	N/A
Q1901-10	B-167-SB02	N/A	N/A	N/A	N/A	100	N/A
Q1901-11	B-170-SB02	N/A	N/A	N/A	N/A	100	N/A
Q1904-02	VNJ-210	N/A	N/A	N/A	N/A	100	N/A
Q1905-04	MH-G	N/A	N/A	N/A	N/A	100	N/A
Q1905-08	MH-H	N/A	N/A	N/A	N/A	100	N/A
Q1906-04	WC-4	N/A	N/A	N/A	N/A	100	N/A
Q1906-08	WC-5	N/A	N/A	N/A	N/A	100	N/A
Q1906-12	WC-6	N/A	N/A	N/A	N/A	100	N/A
Q1906-16	WC-7	N/A	N/A	N/A	N/A	100	N/A
Q1907-02	CO-8R-WC	N/A	N/A	N/A	N/A	100	N/A
Q1912-04	MH-E	N/A	N/A	N/A	N/A	100	N/A
Q1912-08	MH-F	N/A	N/A	N/A	N/A	100	N/A
Q1913-01	WC-12-S-202504	N/A	N/A	N/A	N/A	100	N/A
Q1913-03	WC-13-S-202504	N/A	N/A	N/A	N/A	100	N/A
Q1915-01	WC-04282025	N/A	N/A	N/A	N/A	100	N/A

Alliance TECHNICAL GROUP

TCLP Fluid Determination

PB167774

Hot Block ID : WC S-1 / WC S-2

Thermometer ID : FLASHPOINT

SampleID	ClientID	Sample Weight (g)	Volume DI Water (mL)	PH after 5 min stir	PH after 10 min stir	Extraction Fluid 1 or 2	pH Extraction Fluid
PB167774TB	LEB774	N/A	N/A	N/A	N/A	N/A	N/A
Q1894-01	PAINTED-GROUND-PIPE--1	N/A	N/A	N/A	N/A	N/A	N/A
Q1894-02	PAINTED-GROUND-PIPE-2	N/A	N/A	N/A	N/A	N/A	N/A
Q1894-03	PAINTED-GROUND-PIPE-3	N/A	N/A	N/A	N/A	N/A	N/A
Q1901-08	B-167-SB01	5.02	96.5	8.4	3.0	#1	4.93
Q1901-09	B-170-SB01	5.01	96.5	9.0	4.0	#1	4.93
Q1901-10	B-167-SB02	5.02	96.5	7.0	3.5	#1	4.93
Q1901-11	B-170-SB02	5.03	96.5	8.2	3.5	#1	4.93
Q1904-02	VNJ-210	5.04	96.5	8.4	3.0	#1	4.93
Q1905-04	MH-G	5.03	96.5	5.6	1.5	#1	4.93
Q1905-08	MH-H	5.02	96.5	6.0	2.0	#1	4.93
Q1906-04	WC-4	5.03	96.5	7.6	2.5	#1	4.93
Q1906-08	WC-5	5.02	96.5	7.0	2.5	#1	4.93
Q1906-12	WC-6	5.01	96.5	6.6	2.0	#1	4.93
Q1906-16	WC-7	5.02	96.5	6.6	2.0	#1	4.93
Q1907-02	CO-8R-WC	5.02	96.5	8.4	3.0	#1	4.93
Q1912-04	MH-E	5.01	96.5	6.6	2.0	#1	4.93
Q1912-08	MH-F	5.02	96.5	6.4	2.0	#1	4.93
Q1913-01	WC-12-S-202504	5.02	96.5	8.2	3.5	#1	4.93
Q1913-03	WC-13-S-202504	5.01	96.5	8.0	3.0	#1	4.93
Q1915-01	WC-04282025	5.02	96.5	8.4	3.5	#1	4.93

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TCLP Q1907 **WorkList ID :** 189188 **Department :** TCLP Extraction **Date :** 04-29-2025 07:58:37

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1913-01	WC-12-S-202504	Solid	TCLP Extraction	Cool 4 deg C	AECO02	L21	04/29/2025	1311
Q1913-03	WC-13-S-202504	Solid	TCLP Extraction	Cool 4 deg C	AECO02	L21	04/29/2025	1311
Q1915-01	WC-04282025	Solid	TCLP Extraction	Cool 4 deg C	AECO02	L21	04/29/2025	1311
Q1901-08	B-167-SB01	Solid	TCLP Extraction	Cool 4 deg C	CAMP02	L41	04/28/2025	1311
Q1901-09	B-170-SB01	Solid	TCLP Extraction	Cool 4 deg C	PORT06	L51	04/26/2025	1311
Q1901-10	B-167-SB02	Solid	TCLP Extraction	Cool 4 deg C	PORT06	L51	04/26/2025	1311
Q1901-11	B-170-SB02	Solid	TCLP Extraction	Cool 4 deg C	PORT06	L51	04/26/2025	1311
Q1894-01	PAINTED-GROUND-PIPE--1	Solid	TCLP Extraction	Cool 4 deg C	PORT06	L51	04/26/2025	1311
Q1894-02	PAINTED-GROUND-PIPE-2	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L31	04/28/2025	1311
Q1894-03	PAINTED-GROUND-PIPE-3	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L31	04/28/2025	1311
Q1904-02	VNJ-210	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L31	04/28/2025	1311
Q1905-04	MH-G	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L41	04/28/2025	1311
Q1905-08	MH-H	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L51	04/28/2025	1311
Q1906-04	WC-4	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L51	04/28/2025	1311
Q1906-08	WC-5	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L41	04/28/2025	1311
Q1906-12	WC-6	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L41	04/28/2025	1311
Q1906-16	WC-7	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L41	04/28/2025	1311
Q1912-04	MH-E	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L41	04/28/2025	1311
Q1912-08	MH-F	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L51	04/29/2025	1311
Q1907-02	CO-8R-WC	Solid	TCLP Extraction	Cool 4 deg C	WALS01		04/28/2025	1311

Date/Time 04/29/25 17:00

Raw Sample Received by: [illegible] (WSC)

Raw Sample Relinquished by: JD (SM)

Date/Time 04/29/25 19:30

Raw Sample Received by: JD (SM)

Raw Sample Relinquished by: [illegible] (WSC)

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PB167810

SOP ID:	M3510C,3580A-Extraction SVOC-20				
Clean Up SOP #:	N/A			**Extraction Start Date :**	04/30/2025
Matrix :	Water			**Extraction Start Time :**	13:15
Weigh By:	N/A	**Extraction By:**	RJ	**Extraction End Date :**	04/30/2025
Balance check:	N/A	**Filter By:**	RJ	**Extraction End Time :**	18:25
Balance ID:	N/A	**pH Meter ID:**	N/A	**Concentration By:**	EH
pH Strip Lot#:	E3880	**Hood ID:**	4,5,6,7	**Supervisor By :**	RUPESH

Extraction Method: [√] Seperatory Funnel [ ] Continious Liquid/Liquid [ ] Sonication [ ] Waste Dilution [ ] Soxhlet

Standared Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	50/100 PPM	SP6752
Surrogate	1.0ML	100/150 PPM	SP6754
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3926
Baked Na2SO4	N/A	EP2607
10N NaoH	N/A	EP2559
H2SO4 1:1	N/A	EP2865
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

1.5 ML Vial lot# 2210443. pH Adjusted<2 with 1:1 H2SO4 &>11 with 10 N NaOH.

KD Bath ID:	WATER BATH-1,2	**Envap ID:**	NEVAP-02
KD Bath Temperature:	60 °C	**Envap Temperature:**	40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
4/30/25	RS (Extr Lab)	RC/SVOC
18:10	**Preparation Group**	**Analysis Group**

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PrepBatch ID : PB167810

Analytical Method: M3510C,3580A-Extraction SVOC-20 **Concentration Date:** 04/30/2025

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB167774TB	PB167774TB	TCLP BNA	100	6	RUPESH	ritesh	1			SEP-1
PB167805TB	PB167805TB	TCLP BNA	100	6	RUPESH	ritesh	1			2
PB167810BL	PB167810BL	TCLP BNA	1000	6	RUPESH	ritesh	1			3
PB167810BS	PB167810BS	TCLP BNA	1000	6	RUPESH	ritesh	1			4
Q1901-08	B-167-SB01	TCLP BNA	100	6	RUPESH	ritesh	1	A		5
Q1901-08MS	B-167-SB01MS	TCLP BNA	100	6	RUPESH	ritesh	1	A		6
Q1901-08MSD	B-167-SB01MSD	TCLP BNA	100	6	RUPESH	ritesh	1	A		7
Q1901-09	B-170-SB01	TCLP BNA	100	6	RUPESH	ritesh	1	A		8
Q1901-10	B-167-SB02	TCLP BNA	100	6	RUPESH	ritesh	1	A		9
Q1901-11	B-170-SB02	TCLP BNA	100	6	RUPESH	ritesh	1	A		10
Q1904-02	VNJ-210	TCLP BNA	100	6	RUPESH	ritesh	1	A		11
Q1905-04	MH-G	TCLP BNA	100	6	RUPESH	ritesh	1	A		12
Q1905-08	MH-H	TCLP BNA	100	6	RUPESH	ritesh	1	A		13
Q1906-04	WC-4	TCLP BNA	100	6	RUPESH	ritesh	1	A		14
Q1906-08	WC-5	TCLP BNA	100	6	RUPESH	ritesh	1	A		15
Q1906-12	WC-6	TCLP BNA	100	6	RUPESH	ritesh	1	A		16
Q1906-16	WC-7	TCLP BNA	100	6	RUPESH	ritesh	1	A		SEP-1
Q1907-02	CO-8R-WC	TCLP BNA	100	6	RUPESH	ritesh	1	A		2
Q1912-04	MH-E	TCLP BNA	100	6	RUPESH	ritesh	1	A		3
Q1912-08	MH-F	TCLP BNA	100	6	RUPESH	ritesh	1	A		4
Q1913-01	WC-12-S-202504	TCLP BNA	100	6	RUPESH	ritesh	1	A		5
Q1913-02	WC-12-A-202504	TCLP BNA	100	6	RUPESH	ritesh	1	A		6
Q1913-03	WC-13-S-202504	TCLP BNA	100	6	RUPESH	ritesh	1	A		7
Q1913-04	WC-13-A-202504	TCLP BNA	100	6	RUPESH	ritesh	1	A		8
Q1915-01	WC-04282025	TCLP BNA	100	6	RUPESH	ritesh	1	A		9

RS
4/30

* **Extracts relinquished on the same date as received.**

Alliance TECHNICAL GROUP

TCLP EXTRACTION LOGPAGE

PB167774

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB167774TB	LEB774	21	N/A	2000	N/A	N/A	N/A	4.93	1.0	T-2
Q1894-01	PAINTED-GROUND-PIPE--1	01	25.00	500	N/A	N/A	N/A	5.5	1.0	T-1
Q1894-02	PAINTED-GROUND-PIPE-2	02	35.00	700	N/A	N/A	N/A	5.6	1.5	T-1
Q1894-03	PAINTED-GROUND-PIPE-3	03	34.00	680	N/A	N/A	N/A	5.6	1.5	T-1
Q1901-08	B-167-SB01	04	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-1
Q1901-09	B-170-SB01	05	100.03	2000	N/A	N/A	N/A	7.0	1.5	T-1
Q1901-10	B-167-SB02	06	100.04	2000	N/A	N/A	N/A	5.8	1.0	T-1
Q1901-11	B-170-SB02	07	100.02	2000	N/A	N/A	N/A	5.6	1.5	T-1
Q1904-02	VNJ-210	08	100.03	2000	N/A	N/A	N/A	6.0	1.0	T-1
Q1905-04	MH-G	09	100.02	2000	N/A	N/A	N/A	3.0	1.0	T-1
Q1905-08	MH-H	10	100.01	2000	N/A	N/A	N/A	3.0	1.5	T-1
Q1906-04	WC-4	11	100.02	2000	N/A	N/A	N/A	5.5	1.0	T-2
Q1906-08	WC-5	12	100.03	2000	N/A	N/A	N/A	5.6	1.5	T-2
Q1906-12	WC-6	13	100.02	2000	N/A	N/A	N/A	5.0	1.0	T-2
Q1906-16	WC-7	14	100.03	2000	N/A	N/A	N/A	5.0	1.5	T-2
Q1907-02	CO-8R-WC	15	100.02	2000	N/A	N/A	N/A	5.8	1.5	T-2
Q1912-04	MH-E	16	100.03	2000	N/A	N/A	N/A	4.5	1.0	T-2
Q1912-08	MH-F	17	100.02	2000	N/A	N/A	N/A	4.0	1.5	T-2
Q1913-01	WC-12-S-202504	18	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-2
Q1913-03	WC-13-S-202504	19	100.03	2000	N/A	N/A	N/A	6.2	1.5	T-2
Q1915-01	WC-04282025	20	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-2

04/30/25
23:00

Alliance TECHNICAL GROUP

TCLP EXTRACTION LOGPAGE

PB167805

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB167805TB	LEB805	N/A	N/A	N/A	N/A	N/A	N/A	4.93	1.0	N/A
Q1913-02	WC-12-A-202504	N/A	N/A	N/A	N/A	N/A	N/A	12.0	1.5	N/A
Q1913-04	WC-13-A-202504	N/A	N/A	N/A	N/A	N/A	N/A	12.0	1.5	N/A

04/30/25
13:00



SHIPPING DOCUMENTS

Alliance TECHNICAL GROUP

25-27
30-32

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 · Fax (908) 789-8922
www.chemtech.net

ALLIANCE PROJECT NO.
QUOTE NO. Q1901
COC Number 2046751

CLIENT INFORMATION

REPORT TO BE SENT TO:
COMPANY: Gannett Fleming
ADDRESS: 1010 Adams Ave
CITY Audubon STATE: PA ZIP: 19403
ATTENTION: Joe Krupsky
PHONE: 610-301-8342 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Amtrak replacement of SB
PROJECT NO.: 95000878 LOCATION: Kearny NJ
PROJECT MANAGER: Joe Krupsky
e-mail: JaQuC@RamSys.com
PHONE: 610-310-8342 FAX:

CLIENT BILLING INFORMATION

BILL TO: Alliance PO#:
ADDRESS: 284 Sheffield
CITY Mountainside STATE: NJ ZIP: 07092
ATTENTION: Samantha Beasley PHONE: 908-728-3488

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): 10 DAYS*
EDD: 10 DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC + Raw Data) ☐ NYS ASP A ☐ NYS ASP B ☐ Other
☐ EDD FORMAT Bem EDD

ANALYSIS

1: TCL-VOLATILES
2: PLABS
3: CrVI (soil)
4: TCL-SVOCs BNA
5: TAL Metals
6: EPH
7: TCLP (Liq)
8: RCRA (Liq)
9:

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE COMP	SAMPLE TYPE GRAB	SAMPLE COLLECTION DATE	SAMPLE COLLECTION TIME	# OF BOTTLES	PRESERVATIVES 1	2	3	4	5	6	7	8	9	COMMENTS
1.	B-170-SB00	S		X	4/26/25	900	4	X									
2.	B-167-SB01	S		X		1000	7	X	X	X	X		X	X	X	X	
3.	B-170-SB01	S		X		1015	7	X	X	X	X		X	X	X	X	
4.	B-167-SB02	S		X		1030	8	X	X	X	X		X	X	X	X	
5.	B-170-SB02	S		X		1145	7	X	X	X	X		X	X	X	X	
6.	FB042625	DI W				1215	11	X	X	X	X	X	X	X	X	X	pH 1-3
7.	TB042625	DIW				NA	2	X									
8.																	
9.																	
10.																	

PRESERVATIVES: ← Specify Preservatives — A-HCl D-NaOH B-HN03 E-ICE C-H2SO4 F-OTHER

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

	DATE/TIME	RECEIVED BY:
1. RELINQUISHED BY SAMPLER:	4/26/25 1315	1. 4-28-25 0700
2. RELINQUISHED BY SAMPLER:	DATE/TIME:	2. RECEIVED BY:
3. RELINQUISHED BY SAMPLER:	DATE/TIME:	3. RECEIVED BY:

Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 3.0 °C
Comments: Trip Blank # ("Adjust Factor + 1"

Page _____ of

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete ☐ YES ☐ NO

Copyright © 2024

WHITE - ALLIANCE COPY FOR RETURN TO CLIENT YELLOW - ALLIANCE COPY PINK - SAMPLER COPY

284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

QA Control Code: A2070148

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1901 PORT06
Client Name : Portal Partners Tri-Venture
Client Contact : Joseph Krupansky
Invoice Name : Portal Partners Tri-Venture
Invoice Contact : Joseph Krupansky

Order Date : 4/28/2025 10:24:00 AM
Project Name : Amtrak Sawtooth Bridges 2
Receive DateTime : 4/28/2025 7:00:00 AM
Purchase Order :

Project Mgr :
Report Type : NJ Reduced
EDD Type : EXCEL NJCLEANUP
Hard Copy Date :
Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1901-01	B-170-SB00	Solid	04/26/2025	09:00					
					VOC-TCLVOA-10		8260D		10 Bus. Days
Q1901-02	B-167-SB01	Solid	04/26/2025	10:00					
					VOC-TCLVOA-10		8260D		10 Bus. Days
Q1901-03	B-170-SB01	Solid	04/26/2025	10:15					
					VOC-TCLVOA-10		8260D		10 Bus. Days
Q1901-04	B-167-SB02	Solid	04/26/2025	10:30					
					VOC-TCLVOA-10		8260D		10 Bus. Days
Q1901-05	B-170-SB02	Solid	04/26/2025	11:45					
					VOC-TCLVOA-10		8260D		10 Bus. Days
Q1901-06	FB04262025	Water	04/26/2025	12:15					
					VOC-TCLVOA-10		8260-Low		10 Bus. Days
Q1901-07	TB04262025	Water	04/26/2025	00:00					
					VOC-TCLVOA-10		8260-Low		10 Bus. Days

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1901 PORT06
Client Name : Portal Partners Tri-Venture
Client Contact : Joseph Krupansky
Invoice Name : Portal Partners Tri-Venture
Invoice Contact : Joseph Krupansky

Order Date : 4/28/2025 10:24:00 AM
Project Name : Amtrak Sawtooth Bridges 2
Receive DateTime : 4/28/2025 7:00:00 AM
Purchase Order :

Project Mgr :
Report Type : NJ Reduced
EDD Type : EXCEL NJCLEANUP
Hard Copy Date :
Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES

Relinguished By : CP
Date / Time : 4/28/25 11:45

Received By : JC
Date / Time : 4/28/25 1145

Storage Area : VOA Refridgerator Room