

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

DataFile: BF142255.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB167810BS	Pyridine	50	37.2	ug/L	74				20 (29)	160 (97)	
	1,4-Dichlorobenzene	50	39.6	ug/L	79				70 (76)	130 (103)	
	2-Methylphenol	50	41.0	ug/L	82				70 (69)	130 (109)	
	3+4-Methylphenols	50	40.8	ug/L	82				20 (67)	160 (106)	
	Hexachloroethane	50	41.8	ug/L	84				20 (76)	160 (118)	
	Nitrobenzene	50	45.8	ug/L	92				70 (58)	130 (106)	
	Hexachlorobutadiene	50	41.5	ug/L	83				70 (69)	130 (101)	
	2,4,6-Trichlorophenol	50	43.4	ug/L	87				70 (61)	130 (110)	
	2,4,5-Trichlorophenol	50	44.9	ug/L	90				70 (70)	130 (106)	
	2,4-Dinitrotoluene	50	49.8	ug/L	100				70 (60)	130 (115)	
	Hexachlorobenzene	50	41.6	ug/L	83				70 (73)	130 (106)	
	Pentachlorophenol	100	93.4	ug/L	93				20 (47)	160 (114)	