

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PORT06  
 Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG No.: Q1901  
 Instrument ID: MSVOA\_X Calibration Date(s): 04/01/2025 04/01/2025  
 Heated Purge: (Y/N) N Calibration Time(s): 17:06 19:02  
 GC Column: DB-624UI ID: 0.18 (mm)

| LAB FILE ID:                   |        | RRF001 = VX045525.D |        | RRF005 = VX045526.D |        | RRF020 = VX045527.D |       |       |  |
|--------------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|--|
|                                |        | RRF050 = VX045528.D |        | RRF100 = VX045529.D |        | RRF150 = VX045530.D |       |       |  |
| COMPOUND                       | RRF001 | RRF005              | RRF020 | RRF050              | RRF100 | RRF150              | RRF   | % RSD |  |
| Dichlorodifluoromethane        | 0.615  | 0.695               | 0.700  | 0.820               | 0.837  | 0.820               | 0.748 | 12.1  |  |
| Chloromethane                  | 0.764  | 0.734               | 0.777  | 0.784               | 0.815  | 0.734               | 0.768 | 4.1   |  |
| Vinyl Chloride                 | 0.671  | 0.662               | 0.701  | 0.716               | 0.738  | 0.730               | 0.703 | 4.4   |  |
| Bromomethane                   |        | 0.342               | 0.327  | 0.330               | 0.341  | 0.327               | 0.333 | 2.2   |  |
| Chloroethane                   | 0.378  | 0.397               | 0.390  | 0.398               | 0.355  | 0.319               | 0.373 | 8.3   |  |
| Trichlorofluoromethane         | 1.051  | 0.999               | 1.075  | 1.086               | 1.089  | 0.989               | 1.048 | 4.2   |  |
| 1,1,2-Trichlorotrifluoroethane | 0.575  | 0.599               | 0.635  | 0.621               | 0.621  | 0.634               | 0.614 | 3.7   |  |
| 1,1-Dichloroethene             | 0.563  | 0.588               | 0.612  | 0.604               | 0.618  | 0.616               | 0.600 | 3.5   |  |
| Acetone                        | 0.400  | 0.369               | 0.387  | 0.378               | 0.365  | 0.355               | 0.375 | 4.3   |  |
| Carbon Disulfide               | 1.334  | 1.327               | 1.434  | 1.553               | 1.604  | 1.642               | 1.483 | 9.2   |  |
| Methyl tert-butyl Ether        | 1.915  | 1.964               | 2.169  | 2.118               | 2.217  | 2.216               | 2.100 | 6.2   |  |
| Methyl Acetate                 | 0.901  | 0.863               | 0.864  | 0.857               | 0.855  | 0.846               | 0.864 | 2.2   |  |
| Methylene Chloride             | 0.692  | 0.695               | 0.726  | 0.709               | 0.705  | 0.690               | 0.703 | 2     |  |
| trans-1,2-Dichloroethene       | 0.574  | 0.594               | 0.636  | 0.624               | 0.621  | 0.630               | 0.613 | 3.9   |  |
| 1,1-Dichloroethane             | 1.211  | 1.240               | 1.318  | 1.269               | 1.283  | 1.292               | 1.269 | 3     |  |
| Cyclohexane                    |        | 1.044               | 1.148  | 1.123               | 1.145  | 1.149               | 1.122 | 4     |  |
| 2-Butanone                     | 0.474  | 0.537               | 0.582  | 0.586               | 0.571  | 0.548               | 0.550 | 7.5   |  |
| Carbon Tetrachloride           | 0.450  | 0.497               | 0.521  | 0.536               | 0.545  | 0.556               | 0.518 | 7.5   |  |
| cis-1,2-Dichloroethene         | 0.762  | 0.696               | 0.760  | 0.751               | 0.754  | 0.760               | 0.747 | 3.4   |  |
| Bromochloromethane             | 0.671  | 0.608               | 0.617  | 0.596               | 0.610  | 0.576               | 0.613 | 5.2   |  |
| Chloroform                     | 1.244  | 1.309               | 1.348  | 1.315               | 1.309  | 1.309               | 1.306 | 2.6   |  |
| 1,1,1-Trichloroethane          | 1.028  | 1.052               | 1.120  | 1.109               | 1.153  | 1.167               | 1.105 | 5     |  |
| Methylcyclohexane              | 0.519  | 0.527               | 0.596  | 0.622               | 0.628  | 0.632               | 0.587 | 8.8   |  |
| Benzene                        | 1.414  | 1.416               | 1.519  | 1.481               | 1.483  | 1.465               | 1.463 | 2.8   |  |
| 1,2-Dichloroethane             | 0.533  | 0.585               | 0.649  | 0.622               | 0.620  | 0.617               | 0.604 | 6.7   |  |
| Trichloroethene                | 0.349  | 0.322               | 0.356  | 0.351               | 0.351  | 0.356               | 0.348 | 3.7   |  |
| 1,2-Dichloropropane            | 0.309  | 0.365               | 0.388  | 0.376               | 0.379  | 0.374               | 0.365 | 7.8   |  |
| Bromodichloromethane           | 0.521  | 0.519               | 0.576  | 0.572               | 0.587  | 0.583               | 0.560 | 5.6   |  |
| 4-Methyl-2-Pentanone           | 0.499  | 0.578               | 0.661  | 0.663               | 0.647  | 0.589               | 0.606 | 10.6  |  |
| Toluene                        | 0.817  | 0.866               | 0.939  | 0.910               | 0.905  | 0.875               | 0.885 | 4.8   |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
 All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PORT06  
Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG No.: Q1901  
Instrument ID: MSVOA\_X Calibration Date(s): 04/01/2025 04/01/2025  
Heated Purge: (Y/N) N Calibration Time(s): 17:06 19:02  
GC Column: DB-624UI ID: 0.18 (mm)

| LAB FILE ID:                | RRF001 = VX045525.D | RRF005 = VX045526.D | RRF020 = VX045527.D | RRF050 = VX045528.D | RRF100 = VX045529.D | RRF150 = VX045530.D |       |       |
|-----------------------------|---------------------|---------------------|---------------------|---------------------|---------------------|---------------------|-------|-------|
| COMPOUND                    | RRF001              | RRF005              | RRF020              | RRF050              | RRF100              | RRF150              | RRF   | % RSD |
| t-1,3-Dichloropropene       | 0.316               | 0.400               | 0.465               | 0.508               | 0.555               | 0.558               | 0.467 | 20.3  |
| cis-1,3-Dichloropropene     | 0.371               | 0.474               | 0.546               | 0.568               | 0.597               | 0.600               | 0.526 | 16.9  |
| 1,1,2-Trichloroethane       | 0.337               | 0.346               | 0.376               | 0.359               | 0.358               | 0.340               | 0.353 | 4.1   |
| 2-Hexanone                  | 0.363               | 0.429               | 0.488               | 0.492               | 0.484               | 0.439               | 0.449 | 11.1  |
| Dibromochloromethane        | 0.313               | 0.352               | 0.404               | 0.412               | 0.421               | 0.405               | 0.385 | 11.1  |
| 1,2-Dibromoethane           | 0.294               | 0.351               | 0.380               | 0.373               | 0.380               | 0.364               | 0.357 | 9.1   |
| Tetrachloroethene           | 0.346               | 0.373               | 0.371               | 0.347               | 0.333               | 0.347               | 0.353 | 4.5   |
| Chlorobenzene               | 0.951               | 1.054               | 1.123               | 1.084               | 1.086               | 1.112               | 1.068 | 5.8   |
| Ethyl Benzene               | 1.608               | 1.819               | 2.002               | 2.007               | 1.993               | 2.053               | 1.914 | 8.9   |
| m/p-Xylenes                 | 0.594               | 0.669               | 0.732               | 0.728               | 0.723               | 0.729               | 0.696 | 7.9   |
| o-Xylene                    | 0.562               | 0.676               | 0.725               | 0.719               | 0.716               | 0.714               | 0.686 | 9.2   |
| Styrene                     | 0.897               | 1.043               | 1.202               | 1.216               | 1.230               | 1.202               | 1.132 | 11.8  |
| Bromoform                   | 0.220               | 0.252               | 0.272               | 0.296               | 0.307               | 0.315               | 0.277 | 13.1  |
| Isopropylbenzene            | 3.581               | 3.850               | 4.224               | 4.181               | 4.043               | 4.151               | 4.005 | 6.2   |
| 1,1,2,2-Tetrachloroethane   | 1.457               | 1.428               | 1.461               | 1.384               | 1.354               | 1.358               | 1.407 | 3.4   |
| 1,3-Dichlorobenzene         | 1.658               | 1.605               | 1.726               | 1.699               | 1.714               | 1.706               | 1.684 | 2.7   |
| 1,4-Dichlorobenzene         | 1.671               | 1.724               | 1.768               | 1.703               | 1.674               | 1.706               | 1.708 | 2.1   |
| 1,2-Dichlorobenzene         | 1.644               | 1.645               | 1.750               | 1.678               | 1.665               | 1.667               | 1.675 | 2.3   |
| 1,2-Dibromo-3-Chloropropane | 0.196               | 0.260               | 0.312               | 0.316               | 0.333               | 0.349               | 0.294 | 19.4  |
| 1,2,4-Trichlorobenzene      | 0.727               | 0.844               | 0.948               | 0.947               | 1.045               | 1.083               | 0.932 | 14.1  |
| 1,2,3-Trichlorobenzene      | 0.782               | 0.916               | 0.999               | 1.000               | 1.063               | 1.097               | 0.976 | 11.6  |
| 1,2-Dichloroethane-d4       |                     | 0.962               | 0.900               | 0.868               | 0.904               | 0.937               | 0.914 | 3.9   |
| Dibromofluoromethane        |                     | 0.372               | 0.342               | 0.345               | 0.353               | 0.362               | 0.355 | 3.5   |
| Toluene-d8                  |                     | 1.257               | 1.233               | 1.214               | 1.230               | 1.257               | 1.238 | 1.5   |
| 4-Bromofluorobenzene        |                     | 0.413               | 0.448               | 0.453               | 0.481               | 0.460               | 0.451 | 5.5   |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.