

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901
 Lab File ID: VY022070.D Date Analyzed: 04/30/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:20
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	233640	13.346				
UPPER LIMIT	467280	13.846				
LOWER LIMIT	116820	12.846				
EPA SAMPLE NO.						
B-170-SB00	231898	13.35				
B-167-SB01	176600	13.35				
B-170-SB01	256467	13.35				
B-167-SB02	233488	13.35				
B-170-SB02	150773	13.35				
VY0430SBL01	186896	13.35				
VY0430SBS01	222589	13.35				
VY0430SBSD01	218185	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.