

Hit Summary Sheet
SW-846

SDG No.: Q1901
Client: Portal Partners Tri-Venture

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	B-167-SB02							
Q1901-04	B-167-SB02	SOIL	Acetone	210		8.10	42.6	ug/Kg
Q1901-04	B-167-SB02	SOIL	Carbon Disulfide	8.30	J	1.80	8.50	ug/Kg
Q1901-04	B-167-SB02	SOIL	2-Butanone	39.9	J	11.1	42.6	ug/Kg
			Total Voc :			258		
			Total Concentration:			258		
Client ID:	B-170-SB02							
Q1901-05	B-170-SB02	SOIL	Carbon Disulfide	12.5		1.30	6.00	ug/Kg
Q1901-05	B-170-SB02	SOIL	Benzene	1.60	J	0.95	6.00	ug/Kg
Q1901-05	B-170-SB02	SOIL	m/p-Xylenes	2.20	J	1.50	12.1	ug/Kg
Q1901-05	B-170-SB02	SOIL	o-Xylene	1.70	J	0.99	6.00	ug/Kg
Q1901-05	B-170-SB02	SOIL	Isopropylbenzene	12.2		0.94	6.00	ug/Kg
			Total Voc :			30.2		
Q1901-05	B-170-SB02	SOIL	1,2,4-Trimethylbenzene	* 1.50	J	0.77	6.00	ug/Kg
			Total Tics :			1.50		
			Total Concentration:			31.7		
Client ID:	FB04262025							
Q1901-06	FB04262025	Water	Acetone	4.70	J	1.50	5.00	ug/L
			Total Voc :			4.70		
			Total Concentration:			4.70		
Client ID:	TB04262025							
Q1901-07	TB04262025	Water	Acetone	1.70	J	1.50	5.00	ug/L
			Total Voc :			1.70		
			Total Concentration:			1.70		



SAMPLE DATA

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-170-SB00	SDG No.:	Q1901
Lab Sample ID:	Q1901-01	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	92.6
Sample Wt/Vol:	6.67 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022105.D	1		05/01/25 14:15	VY050125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	0.92	U	0.92	4.00	ug/Kg
74-87-3	Chloromethane	0.92	U	0.92	4.00	ug/Kg
75-01-4	Vinyl Chloride	0.64	U	0.64	4.00	ug/Kg
74-83-9	Bromomethane	0.87	U	0.87	4.00	ug/Kg
75-00-3	Chloroethane	1.00	U	1.00	4.00	ug/Kg
75-69-4	Trichlorofluoromethane	0.98	U	0.98	4.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	0.86	U	0.86	4.00	ug/Kg
75-35-4	1,1-Dichloroethene	0.81	U	0.81	4.00	ug/Kg
67-64-1	Acetone	3.80	U	3.80	20.2	ug/Kg
75-15-0	Carbon Disulfide	0.86	U	0.86	4.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.59	U	0.59	4.00	ug/Kg
79-20-9	Methyl Acetate	1.20	U	1.20	4.00	ug/Kg
75-09-2	Methylene Chloride	2.90	U	2.90	8.10	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.70	U	0.70	4.00	ug/Kg
75-34-3	1,1-Dichloroethane	0.65	U	0.65	4.00	ug/Kg
110-82-7	Cyclohexane	0.64	U	0.64	4.00	ug/Kg
78-93-3	2-Butanone	5.30	U	5.30	20.2	ug/Kg
56-23-5	Carbon Tetrachloride	0.79	U	0.79	4.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.61	U	0.61	4.00	ug/Kg
74-97-5	Bromochloromethane	0.93	U	0.93	4.00	ug/Kg
67-66-3	Chloroform	0.68	U	0.68	4.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.75	U	0.75	4.00	ug/Kg
108-87-2	Methylcyclohexane	0.74	U	0.74	4.00	ug/Kg
71-43-2	Benzene	0.64	U	0.64	4.00	ug/Kg
107-06-2	1,2-Dichloroethane	0.64	U	0.64	4.00	ug/Kg
79-01-6	Trichloroethene	0.66	U	0.66	4.00	ug/Kg
78-87-5	1,2-Dichloropropane	0.74	U	0.74	4.00	ug/Kg
75-27-4	Bromodichloromethane	0.63	U	0.63	4.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	2.90	U	2.90	20.2	ug/Kg
108-88-3	Toluene	0.63	U	0.63	4.00	ug/Kg

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	04/26/25	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	04/28/25	
Client Sample ID:	B-170-SB00		SDG No.:	Q1901	
Lab Sample ID:	Q1901-01		Matrix:	SOIL	
Analytical Method:	SW8260		% Solid:	92.6	
Sample Wt/Vol:	6.67	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022105.D	1		05/01/25 14:15	VY050125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.53	U	0.53	4.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.50	4.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.74	U	0.74	4.00	ug/Kg
591-78-6	2-Hexanone	3.00	U	3.00	20.2	ug/Kg
124-48-1	Dibromochloromethane	0.70	U	0.70	4.00	ug/Kg
106-93-4	1,2-Dibromoethane	0.71	U	0.71	4.00	ug/Kg
127-18-4	Tetrachloroethene	0.85	U	0.85	4.00	ug/Kg
108-90-7	Chlorobenzene	0.74	U	0.74	4.00	ug/Kg
100-41-4	Ethyl Benzene	0.54	U	0.54	4.00	ug/Kg
179601-23-1	m/p-Xylenes	1.00	U	1.00	8.10	ug/Kg
95-47-6	o-Xylene	0.66	U	0.66	4.00	ug/Kg
100-42-5	Styrene	0.57	U	0.57	4.00	ug/Kg
75-25-2	Bromoform	0.70	U	0.70	4.00	ug/Kg
98-82-8	Isopropylbenzene	0.63	U	0.63	4.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	0.98	U	0.98	4.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.40	U	1.40	4.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.30	U	1.30	4.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.20	U	1.20	4.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.50	U	1.50	4.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	2.40	U	2.40	4.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	2.60	U	2.60	4.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.5		70 (63) - 130 (155)	107%	SPK: 50
1868-53-7	Dibromofluoromethane	50.8		70 (70) - 130 (134)	102%	SPK: 50
2037-26-5	Toluene-d8	48.2		70 (74) - 130 (123)	96%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.9		70 (38) - 130 (136)	110%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	310000	7.707			
540-36-3	1,4-Difluorobenzene	585000	8.616			
3114-55-4	Chlorobenzene-d5	519000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	196000	13.347			

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-170-SB00	SDG No.:	Q1901
Lab Sample ID:	Q1901-01	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	92.6
Sample Wt/Vol:	6.67 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022105.D	1		05/01/25 14:15	VY050125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-167-SB01	SDG No.:	Q1901
Lab Sample ID:	Q1901-02	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	82.6
Sample Wt/Vol:	7.83 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022087.D	1		04/30/25 17:05	VY043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	0.88	U	0.88	3.90	ug/Kg
74-87-3	Chloromethane	0.88	U	0.88	3.90	ug/Kg
75-01-4	Vinyl Chloride	0.61	U	0.61	3.90	ug/Kg
74-83-9	Bromomethane	0.83	U	0.83	3.90	ug/Kg
75-00-3	Chloroethane	0.97	U	0.97	3.90	ug/Kg
75-69-4	Trichlorofluoromethane	0.94	U	0.94	3.90	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	0.82	U	0.82	3.90	ug/Kg
75-35-4	1,1-Dichloroethene	0.77	U	0.77	3.90	ug/Kg
67-64-1	Acetone	3.70	U	3.70	19.3	ug/Kg
75-15-0	Carbon Disulfide	0.82	U	0.82	3.90	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.56	U	0.56	3.90	ug/Kg
79-20-9	Methyl Acetate	1.20	U	1.20	3.90	ug/Kg
75-09-2	Methylene Chloride	2.70	U	2.70	7.70	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.66	U	0.66	3.90	ug/Kg
75-34-3	1,1-Dichloroethane	0.62	U	0.62	3.90	ug/Kg
110-82-7	Cyclohexane	0.61	U	0.61	3.90	ug/Kg
78-93-3	2-Butanone	5.10	U	5.10	19.3	ug/Kg
56-23-5	Carbon Tetrachloride	0.75	U	0.75	3.90	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.58	U	0.58	3.90	ug/Kg
74-97-5	Bromochloromethane	0.89	U	0.89	3.90	ug/Kg
67-66-3	Chloroform	0.65	U	0.65	3.90	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.72	U	0.72	3.90	ug/Kg
108-87-2	Methylcyclohexane	0.70	U	0.70	3.90	ug/Kg
71-43-2	Benzene	0.61	U	0.61	3.90	ug/Kg
107-06-2	1,2-Dichloroethane	0.61	U	0.61	3.90	ug/Kg
79-01-6	Trichloroethene	0.63	U	0.63	3.90	ug/Kg
78-87-5	1,2-Dichloropropane	0.70	U	0.70	3.90	ug/Kg
75-27-4	Bromodichloromethane	0.60	U	0.60	3.90	ug/Kg
108-10-1	4-Methyl-2-Pentanone	2.80	U	2.80	19.3	ug/Kg
108-88-3	Toluene	0.60	U	0.60	3.90	ug/Kg

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	04/26/25	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	04/28/25	
Client Sample ID:	B-167-SB01		SDG No.:	Q1901	
Lab Sample ID:	Q1901-02		Matrix:	SOIL	
Analytical Method:	SW8260		% Solid:	82.6	
Sample Wt/Vol:	7.83	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022087.D	1		04/30/25 17:05	VY043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.50	3.90	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.48	U	0.48	3.90	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.71	U	0.71	3.90	ug/Kg
591-78-6	2-Hexanone	2.90	U	2.90	19.3	ug/Kg
124-48-1	Dibromochloromethane	0.67	U	0.67	3.90	ug/Kg
106-93-4	1,2-Dibromoethane	0.68	U	0.68	3.90	ug/Kg
127-18-4	Tetrachloroethene	0.81	U	0.81	3.90	ug/Kg
108-90-7	Chlorobenzene	0.70	U	0.70	3.90	ug/Kg
100-41-4	Ethyl Benzene	0.52	U	0.52	3.90	ug/Kg
179601-23-1	m/p-Xylenes	0.96	U	0.96	7.70	ug/Kg
95-47-6	o-Xylene	0.63	U	0.63	3.90	ug/Kg
100-42-5	Styrene	0.55	U	0.55	3.90	ug/Kg
75-25-2	Bromoform	0.66	U	0.66	3.90	ug/Kg
98-82-8	Isopropylbenzene	0.60	U	0.60	3.90	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	0.94	U	0.94	3.90	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.30	U	1.30	3.90	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.20	U	1.20	3.90	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.10	U	1.10	3.90	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.40	U	1.40	3.90	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	2.30	U	2.30	3.90	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	2.50	U	2.50	3.90	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.3		70 (63) - 130 (155)	107%	SPK: 50
1868-53-7	Dibromofluoromethane	51.1		70 (70) - 130 (134)	102%	SPK: 50
2037-26-5	Toluene-d8	47.7		70 (74) - 130 (123)	95%	SPK: 50
460-00-4	4-Bromofluorobenzene	36.1		70 (38) - 130 (136)	72%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	348000	7.707			
540-36-3	1,4-Difluorobenzene	652000	8.61			
3114-55-4	Chlorobenzene-d5	556000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	177000	13.346			

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-167-SB01	SDG No.:	Q1901
Lab Sample ID:	Q1901-02	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	82.6
Sample Wt/Vol:	7.83 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022087.D	1		04/30/25 17:05	VY043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-170-SB01	SDG No.:	Q1901
Lab Sample ID:	Q1901-03	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	83
Sample Wt/Vol:	7.35 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022106.D	1		05/01/25 14:38	VY050125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	0.93	U	0.93	4.10	ug/Kg
74-87-3	Chloromethane	0.93	U	0.93	4.10	ug/Kg
75-01-4	Vinyl Chloride	0.65	U	0.65	4.10	ug/Kg
74-83-9	Bromomethane	0.88	U	0.88	4.10	ug/Kg
75-00-3	Chloroethane	1.00	U	1.00	4.10	ug/Kg
75-69-4	Trichlorofluoromethane	0.99	U	0.99	4.10	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	0.87	U	0.87	4.10	ug/Kg
75-35-4	1,1-Dichloroethene	0.82	U	0.82	4.10	ug/Kg
67-64-1	Acetone	3.90	U	3.90	20.5	ug/Kg
75-15-0	Carbon Disulfide	0.87	U	0.87	4.10	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.60	U	0.60	4.10	ug/Kg
79-20-9	Methyl Acetate	1.30	U	1.30	4.10	ug/Kg
75-09-2	Methylene Chloride	2.90	U	2.90	8.20	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.70	U	0.70	4.10	ug/Kg
75-34-3	1,1-Dichloroethane	0.66	U	0.66	4.10	ug/Kg
110-82-7	Cyclohexane	0.65	U	0.65	4.10	ug/Kg
78-93-3	2-Butanone	5.40	U	5.40	20.5	ug/Kg
56-23-5	Carbon Tetrachloride	0.80	U	0.80	4.10	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.61	U	0.61	4.10	ug/Kg
74-97-5	Bromochloromethane	0.94	U	0.94	4.10	ug/Kg
67-66-3	Chloroform	0.69	U	0.69	4.10	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.76	U	0.76	4.10	ug/Kg
108-87-2	Methylcyclohexane	0.75	U	0.75	4.10	ug/Kg
71-43-2	Benzene	0.65	U	0.65	4.10	ug/Kg
107-06-2	1,2-Dichloroethane	0.65	U	0.65	4.10	ug/Kg
79-01-6	Trichloroethene	0.66	U	0.66	4.10	ug/Kg
78-87-5	1,2-Dichloropropane	0.75	U	0.75	4.10	ug/Kg
75-27-4	Bromodichloromethane	0.64	U	0.64	4.10	ug/Kg
108-10-1	4-Methyl-2-Pentanone	2.90	U	2.90	20.5	ug/Kg
108-88-3	Toluene	0.64	U	0.64	4.10	ug/Kg

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	04/26/25	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	04/28/25	
Client Sample ID:	B-170-SB01		SDG No.:	Q1901	
Lab Sample ID:	Q1901-03		Matrix:	SOIL	
Analytical Method:	SW8260		% Solid:	83	
Sample Wt/Vol:	7.35	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022106.D	1		05/01/25 14:38	VY050125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.53	U	0.53	4.10	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.51	U	0.51	4.10	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.75	U	0.75	4.10	ug/Kg
591-78-6	2-Hexanone	3.00	U	3.00	20.5	ug/Kg
124-48-1	Dibromochloromethane	0.71	U	0.71	4.10	ug/Kg
106-93-4	1,2-Dibromoethane	0.72	U	0.72	4.10	ug/Kg
127-18-4	Tetrachloroethene	0.86	U	0.86	4.10	ug/Kg
108-90-7	Chlorobenzene	0.75	U	0.75	4.10	ug/Kg
100-41-4	Ethyl Benzene	0.55	U	0.55	4.10	ug/Kg
179601-23-1	m/p-Xylenes	1.00	U	1.00	8.20	ug/Kg
95-47-6	o-Xylene	0.67	U	0.67	4.10	ug/Kg
100-42-5	Styrene	0.58	U	0.58	4.10	ug/Kg
75-25-2	Bromoform	0.70	U	0.70	4.10	ug/Kg
98-82-8	Isopropylbenzene	0.64	U	0.64	4.10	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	0.99	U	0.99	4.10	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.40	U	1.40	4.10	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.30	U	1.30	4.10	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.20	U	1.20	4.10	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.50	U	1.50	4.10	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	2.40	U	2.40	4.10	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	2.60	U	2.60	4.10	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	55.0		70 (63) - 130 (155)	110%	SPK: 50
1868-53-7	Dibromofluoromethane	51.4		70 (70) - 130 (134)	103%	SPK: 50
2037-26-5	Toluene-d8	47.9		70 (74) - 130 (123)	96%	SPK: 50
460-00-4	4-Bromofluorobenzene	56.4		70 (38) - 130 (136)	113%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	308000	7.713			
540-36-3	1,4-Difluorobenzene	584000	8.615			
3114-55-4	Chlorobenzene-d5	521000	11.42			
3855-82-1	1,4-Dichlorobenzene-d4	197000	13.346			

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-170-SB01	SDG No.:	Q1901
Lab Sample ID:	Q1901-03	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	83
Sample Wt/Vol:	7.35 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022106.D	1		05/01/25 14:38	VY050125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-167-SB02	SDG No.:	Q1901
Lab Sample ID:	Q1901-04	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	53.2
Sample Wt/Vol:	5.52 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022089.D	1		04/30/25 17:52	VY043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.90	U	1.90	8.50	ug/Kg
74-87-3	Chloromethane	1.90	U	1.90	8.50	ug/Kg
75-01-4	Vinyl Chloride	1.30	U	1.30	8.50	ug/Kg
74-83-9	Bromomethane	1.80	U	1.80	8.50	ug/Kg
75-00-3	Chloroethane	2.10	U	2.10	8.50	ug/Kg
75-69-4	Trichlorofluoromethane	2.10	U	2.10	8.50	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	1.80	U	1.80	8.50	ug/Kg
75-35-4	1,1-Dichloroethene	1.70	U	1.70	8.50	ug/Kg
67-64-1	Acetone	210		8.10	42.6	ug/Kg
75-15-0	Carbon Disulfide	8.30	J	1.80	8.50	ug/Kg
1634-04-4	Methyl tert-butyl Ether	1.20	U	1.20	8.50	ug/Kg
79-20-9	Methyl Acetate	2.60	U	2.60	8.50	ug/Kg
75-09-2	Methylene Chloride	6.00	U	6.00	17.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	1.50	U	1.50	8.50	ug/Kg
75-34-3	1,1-Dichloroethane	1.40	U	1.40	8.50	ug/Kg
110-82-7	Cyclohexane	1.30	U	1.30	8.50	ug/Kg
78-93-3	2-Butanone	39.9	J	11.1	42.6	ug/Kg
56-23-5	Carbon Tetrachloride	1.70	U	1.70	8.50	ug/Kg
156-59-2	cis-1,2-Dichloroethene	1.30	U	1.30	8.50	ug/Kg
74-97-5	Bromochloromethane	2.00	U	2.00	8.50	ug/Kg
67-66-3	Chloroform	1.40	U	1.40	8.50	ug/Kg
71-55-6	1,1,1-Trichloroethane	1.60	U	1.60	8.50	ug/Kg
108-87-2	Methylcyclohexane	1.50	U	1.50	8.50	ug/Kg
71-43-2	Benzene	1.30	U	1.30	8.50	ug/Kg
107-06-2	1,2-Dichloroethane	1.30	U	1.30	8.50	ug/Kg
79-01-6	Trichloroethene	1.40	U	1.40	8.50	ug/Kg
78-87-5	1,2-Dichloropropane	1.50	U	1.50	8.50	ug/Kg
75-27-4	Bromodichloromethane	1.30	U	1.30	8.50	ug/Kg
108-10-1	4-Methyl-2-Pentanone	6.10	U	6.10	42.6	ug/Kg
108-88-3	Toluene	1.30	U	1.30	8.50	ug/Kg

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-167-SB02	SDG No.:	Q1901
Lab Sample ID:	Q1901-04	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	53.2
Sample Wt/Vol:	5.52 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022089.D	1		04/30/25 17:52	VY043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	1.10	U	1.10	8.50	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	1.10	U	1.10	8.50	ug/Kg
79-00-5	1,1,2-Trichloroethane	1.60	U	1.60	8.50	ug/Kg
591-78-6	2-Hexanone	6.30	U	6.30	42.6	ug/Kg
124-48-1	Dibromochloromethane	1.50	U	1.50	8.50	ug/Kg
106-93-4	1,2-Dibromoethane	1.50	U	1.50	8.50	ug/Kg
127-18-4	Tetrachloroethene	1.80	U	1.80	8.50	ug/Kg
108-90-7	Chlorobenzene	1.50	U	1.50	8.50	ug/Kg
100-41-4	Ethyl Benzene	1.10	U	1.10	8.50	ug/Kg
179601-23-1	m/p-Xylenes	2.10	U	2.10	17.0	ug/Kg
95-47-6	o-Xylene	1.40	U	1.40	8.50	ug/Kg
100-42-5	Styrene	1.20	U	1.20	8.50	ug/Kg
75-25-2	Bromoform	1.50	U	1.50	8.50	ug/Kg
98-82-8	Isopropylbenzene	1.30	U	1.30	8.50	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	2.10	U	2.10	8.50	ug/Kg
541-73-1	1,3-Dichlorobenzene	2.90	U	2.90	8.50	ug/Kg
106-46-7	1,4-Dichlorobenzene	2.70	U	2.70	8.50	ug/Kg
95-50-1	1,2-Dichlorobenzene	2.50	U	2.50	8.50	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	3.10	U	3.10	8.50	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	5.10	U	5.10	8.50	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	5.40	U	5.40	8.50	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.5		70 (63) - 130 (155)	107%	SPK: 50
1868-53-7	Dibromofluoromethane	51.0		70 (70) - 130 (134)	102%	SPK: 50
2037-26-5	Toluene-d8	47.7		70 (74) - 130 (123)	95%	SPK: 50
460-00-4	4-Bromofluorobenzene	39.3		70 (38) - 130 (136)	79%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	381000	7.707			
540-36-3	1,4-Difluorobenzene	713000	8.609			
3114-55-4	Chlorobenzene-d5	629000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	233000	13.346			

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	04/26/25	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	04/28/25	
Client Sample ID:	B-167-SB02		SDG No.:	Q1901	
Lab Sample ID:	Q1901-04		Matrix:	SOIL	
Analytical Method:	SW8260		% Solid:	53.2	
Sample Wt/Vol:	5.52	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022089.D	1		04/30/25 17:52	VY043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	04/26/25	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	04/28/25	
Client Sample ID:	B-170-SB02		SDG No.:	Q1901	
Lab Sample ID:	Q1901-05		Matrix:	SOIL	
Analytical Method:	SW8260		% Solid:	75.1	
Sample Wt/Vol:	5.52	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022090.D	1		04/30/25 18:16	VY043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.40	U	1.40	6.00	ug/Kg
74-87-3	Chloromethane	1.40	U	1.40	6.00	ug/Kg
75-01-4	Vinyl Chloride	0.95	U	0.95	6.00	ug/Kg
74-83-9	Bromomethane	1.30	U	1.30	6.00	ug/Kg
75-00-3	Chloroethane	1.50	U	1.50	6.00	ug/Kg
75-69-4	Trichlorofluoromethane	1.50	U	1.50	6.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	1.30	U	1.30	6.00	ug/Kg
75-35-4	1,1-Dichloroethene	1.20	U	1.20	6.00	ug/Kg
67-64-1	Acetone	5.70	U	5.70	30.2	ug/Kg
75-15-0	Carbon Disulfide	12.5		1.30	6.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.88	U	0.88	6.00	ug/Kg
79-20-9	Methyl Acetate	1.90	U	1.90	6.00	ug/Kg
75-09-2	Methylene Chloride	4.30	U	4.30	12.1	ug/Kg
156-60-5	trans-1,2-Dichloroethene	1.00	U	1.00	6.00	ug/Kg
75-34-3	1,1-Dichloroethane	0.96	U	0.96	6.00	ug/Kg
110-82-7	Cyclohexane	0.95	U	0.95	6.00	ug/Kg
78-93-3	2-Butanone	7.90	U	7.90	30.2	ug/Kg
56-23-5	Carbon Tetrachloride	1.20	U	1.20	6.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.90	U	0.90	6.00	ug/Kg
74-97-5	Bromochloromethane	1.40	U	1.40	6.00	ug/Kg
67-66-3	Chloroform	1.00	U	1.00	6.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	1.10	U	1.10	6.00	ug/Kg
108-87-2	Methylcyclohexane	1.10	U	1.10	6.00	ug/Kg
71-43-2	Benzene	1.60	J	0.95	6.00	ug/Kg
107-06-2	1,2-Dichloroethane	0.95	U	0.95	6.00	ug/Kg
79-01-6	Trichloroethene	0.98	U	0.98	6.00	ug/Kg
78-87-5	1,2-Dichloropropane	1.10	U	1.10	6.00	ug/Kg
75-27-4	Bromodichloromethane	0.94	U	0.94	6.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	4.30	U	4.30	30.2	ug/Kg
108-88-3	Toluene	0.94	U	0.94	6.00	ug/Kg

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	B-170-SB02	SDG No.:	Q1901
Lab Sample ID:	Q1901-05	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	75.1
Sample Wt/Vol:	5.52 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022090.D	1		04/30/25 18:16	VY043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.78	U	0.78	6.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.75	U	0.75	6.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	1.10	U	1.10	6.00	ug/Kg
591-78-6	2-Hexanone	4.50	U	4.50	30.2	ug/Kg
124-48-1	Dibromochloromethane	1.00	U	1.00	6.00	ug/Kg
106-93-4	1,2-Dibromoethane	1.10	U	1.10	6.00	ug/Kg
127-18-4	Tetrachloroethene	1.30	U	1.30	6.00	ug/Kg
108-90-7	Chlorobenzene	1.10	U	1.10	6.00	ug/Kg
100-41-4	Ethyl Benzene	0.81	U	0.81	6.00	ug/Kg
179601-23-1	m/p-Xylenes	2.20	J	1.50	12.1	ug/Kg
95-47-6	o-Xylene	1.70	J	0.99	6.00	ug/Kg
100-42-5	Styrene	0.86	U	0.86	6.00	ug/Kg
75-25-2	Bromoform	1.00	U	1.00	6.00	ug/Kg
98-82-8	Isopropylbenzene	12.2		0.94	6.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.50	U	1.50	6.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	2.10	U	2.10	6.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.90	U	1.90	6.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.70	U	1.70	6.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	2.20	U	2.20	6.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	3.60	U	3.60	6.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	3.80	U	3.80	6.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.6		70 (63) - 130 (155)	99%	SPK: 50
1868-53-7	Dibromofluoromethane	49.8		70 (70) - 130 (134)	100%	SPK: 50
2037-26-5	Toluene-d8	46.7		70 (74) - 130 (123)	93%	SPK: 50
460-00-4	4-Bromofluorobenzene	33.0	*	70 (38) - 130 (136)	66%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	385000	7.707			
540-36-3	1,4-Difluorobenzene	710000	8.61			
3114-55-4	Chlorobenzene-d5	564000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	151000	13.347			

TENTATIVE IDENTIFIED COMPOUNDS

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	04/26/25	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	04/28/25	
Client Sample ID:	B-170-SB02		SDG No.:	Q1901	
Lab Sample ID:	Q1901-05		Matrix:	SOIL	
Analytical Method:	SW8260		% Solid:	75.1	
Sample Wt/Vol:	5.52	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022090.D	1		04/30/25 18:16	VY043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
95-63-6	1,2,4-Trimethylbenzene	1.50	J		13.0	ug/Kg

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	04/26/25	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	04/28/25	
Client Sample ID:	FB04262025		SDG No.:	Q1901	
Lab Sample ID:	Q1901-06		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045979.D	1		04/29/25 13:58	VX042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	4.70	J	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	UQ	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	04/26/25	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	04/28/25	
Client Sample ID:	FB04262025		SDG No.:	Q1901	
Lab Sample ID:	Q1901-06		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045979.D	1		04/29/25 13:58	VX042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	U	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	U	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.6		70 (74) - 130 (125)	109%	SPK: 50
1868-53-7	Dibromofluoromethane	51.6		70 (75) - 130 (124)	103%	SPK: 50
2037-26-5	Toluene-d8	50.5		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.4		70 (77) - 130 (121)	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	64500	5.544			
540-36-3	1,4-Difluorobenzene	128000	6.757			
3114-55-4	Chlorobenzene-d5	116000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	48400	12.018			

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	FB04262025	SDG No.:	Q1901
Lab Sample ID:	Q1901-06	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	DB-624UI	ID :	0.18
Prep Method :		Level :	LOW

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045979.D	1		04/29/25 13:58	VX042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/26/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/28/25
Client Sample ID:	TB04262025	SDG No.:	Q1901
Lab Sample ID:	Q1901-07	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045978.D	1		04/29/25 13:35	VX042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	1.70	J	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	UQ	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	04/26/25	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	04/28/25	
Client Sample ID:	TB04262025		SDG No.:	Q1901	
Lab Sample ID:	Q1901-07		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045978.D	1		04/29/25 13:35	VX042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	U	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	U	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.2		70 (74) - 130 (125)	108%	SPK: 50
1868-53-7	Dibromofluoromethane	51.9		70 (75) - 130 (124)	104%	SPK: 50
2037-26-5	Toluene-d8	50.8		70 (86) - 130 (113)	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.9		70 (77) - 130 (121)	108%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	67300	5.544			
540-36-3	1,4-Difluorobenzene	131000	6.757			
3114-55-4	Chlorobenzene-d5	123000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	53800	12.018			



QC SUMMARY

Surrogate Summary

SDG No.: **Q1901**

Client: **Portal Partners Tri-Venture**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1901-01	B-170-SB00	1,2-Dichloroethane-d4	50	53.5	107	70 (63)	130 (155)
		Dibromofluoromethane	50	50.8	102	70 (70)	130 (134)
		Toluene-d8	50	48.2	96	70 (74)	130 (123)
		4-Bromofluorobenzene	50	54.9	110	70 (38)	130 (136)
Q1901-02	B-167-SB01	1,2-Dichloroethane-d4	50	53.3	107	70 (63)	130 (155)
		Dibromofluoromethane	50	51.1	102	70 (70)	130 (134)
		Toluene-d8	50	47.7	95	70 (74)	130 (123)
		4-Bromofluorobenzene	50	36.1	72	70 (38)	130 (136)
Q1901-03	B-170-SB01	1,2-Dichloroethane-d4	50	55.0	110	70 (63)	130 (155)
		Dibromofluoromethane	50	51.4	103	70 (70)	130 (134)
		Toluene-d8	50	47.9	96	70 (74)	130 (123)
		4-Bromofluorobenzene	50	56.5	113	70 (38)	130 (136)
Q1901-04	B-167-SB02	1,2-Dichloroethane-d4	50	53.5	107	70 (63)	130 (155)
		Dibromofluoromethane	50	51.0	102	70 (70)	130 (134)
		Toluene-d8	50	47.7	95	70 (74)	130 (123)
		4-Bromofluorobenzene	50	39.4	79	70 (38)	130 (136)
Q1901-05	B-170-SB02	1,2-Dichloroethane-d4	50	49.6	99	70 (63)	130 (155)
		Dibromofluoromethane	50	49.8	100	70 (70)	130 (134)
		Toluene-d8	50	46.7	93	70 (74)	130 (123)
		4-Bromofluorobenzene	50	33.0	66 *	70 (38)	130 (136)
VY0430SBL01	VY0430SBL01	1,2-Dichloroethane-d4	50	52.3	105	70 (63)	130 (155)
		Dibromofluoromethane	50	51.8	104	70 (70)	130 (134)
		Toluene-d8	50	48.2	96	70 (74)	130 (123)
		4-Bromofluorobenzene	50	53.3	107	70 (38)	130 (136)
VY0430SBS01	VY0430SBS01	1,2-Dichloroethane-d4	50	49.5	99	70 (63)	130 (155)
		Dibromofluoromethane	50	52.6	105	70 (70)	130 (134)
		Toluene-d8	50	52.7	105	70 (74)	130 (123)
		4-Bromofluorobenzene	50	51.8	104	70 (38)	130 (136)
VY0430SBSD01	VY0430SBSD01	1,2-Dichloroethane-d4	50	49.0	98	70 (63)	130 (155)
		Dibromofluoromethane	50	50.8	102	70 (70)	130 (134)
		Toluene-d8	50	51.3	103	70 (74)	130 (123)
		4-Bromofluorobenzene	50	50.7	101	70 (38)	130 (136)
VY0501SBL01	VY0501SBL01	1,2-Dichloroethane-d4	50	49.2	98	70 (63)	130 (155)
		Dibromofluoromethane	50	50.6	101	70 (70)	130 (134)
		Toluene-d8	50	47.8	96	70 (74)	130 (123)
		4-Bromofluorobenzene	50	53.1	106	70 (38)	130 (136)
VY0501SBS01	VY0501SBS01	1,2-Dichloroethane-d4	50	47.5	95	70 (63)	130 (155)
		Dibromofluoromethane	50	49.5	99	70 (70)	130 (134)
		Toluene-d8	50	50.1	100	70 (74)	130 (123)
		4-Bromofluorobenzene	50	49.3	99	70 (38)	130 (136)

Surrogate Summary

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1901-06	FB04262025	1,2-Dichloroethane-d4	50	54.6	109	70 (74)	130 (125)
		Dibromofluoromethane	50	51.6	103	70 (75)	130 (124)
		Toluene-d8	50	50.5	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.4	101	70 (77)	130 (121)
Q1901-07	TB04262025	1,2-Dichloroethane-d4	50	54.2	108	70 (74)	130 (125)
		Dibromofluoromethane	50	51.9	104	70 (75)	130 (124)
		Toluene-d8	50	50.8	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.9	108	70 (77)	130 (121)
VX0429WBL01	VX0429WBL01	1,2-Dichloroethane-d4	50	55.1	110	70 (74)	130 (125)
		Dibromofluoromethane	50	51.6	103	70 (75)	130 (124)
		Toluene-d8	50	49.7	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.9	100	70 (77)	130 (121)
VX0429WBS01	VX0429WBS01	1,2-Dichloroethane-d4	50	53.9	108	70 (74)	130 (125)
		Dibromofluoromethane	50	55.4	111	70 (75)	130 (124)
		Toluene-d8	50	53.0	106	70 (86)	130 (113)
		4-Bromofluorobenzene	50	56.4	113	70 (77)	130 (121)
VX0429WBSD0	VX0429WBSD01	1,2-Dichloroethane-d4	50	54.3	109	70 (74)	130 (125)
		Dibromofluoromethane	50	54.8	110	70 (75)	130 (124)
		Toluene-d8	50	51.7	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	55.4	111	70 (77)	130 (121)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: SW8260-Low

Datafile : VX045975.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0429WBS01	Dichlorodifluoromethane	20	17.4	ug/L	87			40 (69)	160 (116)	
	Chloromethane	20	16.3	ug/L	81			40 (65)	160 (116)	
	Vinyl chloride	20	16.9	ug/L	85			70 (65)	130 (117)	
	Bromomethane	20	17.5	ug/L	88			40 (58)	160 (125)	
	Chloroethane	20	20.4	ug/L	102			40 (56)	160 (128)	
	Trichlorofluoromethane	20	19.4	ug/L	97			40 (73)	160 (115)	
	1,1,2-Trichlorotrifluoroethane	20	20.4	ug/L	102			70 (80)	130 (112)	
	1,1-Dichloroethene	20	18.5	ug/L	93			70 (74)	130 (110)	
	Acetone	100	97.8	ug/L	98			40 (60)	160 (125)	
	Carbon disulfide	20	13.4	ug/L	67			40 (64)	160 (112)	
	Methyl tert-butyl Ether	20	21.1	ug/L	106			70 (78)	130 (114)	
	Methyl Acetate	20	28.4	ug/L	142		*	70 (67)	130 (125)	
	Methylene Chloride	20	19.2	ug/L	96			70 (72)	130 (114)	
	trans-1,2-Dichloroethene	20	17.8	ug/L	89			70 (75)	130 (108)	
	1,1-Dichloroethane	20	20.1	ug/L	101			70 (78)	130 (112)	
	Cyclohexane	20	18.0	ug/L	90			70 (75)	130 (110)	
	2-Butanone	100	100	ug/L	100			40 (65)	160 (122)	
	Carbon Tetrachloride	20	21.2	ug/L	106			70 (77)	130 (113)	
	cis-1,2-Dichloroethene	20	19.9	ug/L	100			70 (77)	130 (110)	
	Bromochloromethane	20	22.0	ug/L	110			70 (70)	130 (124)	
	Chloroform	20	21.1	ug/L	106			70 (79)	130 (113)	
	1,1,1-Trichloroethane	20	20.6	ug/L	103			70 (80)	130 (108)	
	Methylcyclohexane	20	18.8	ug/L	94			70 (72)	130 (115)	
	Benzene	20	19.7	ug/L	99			70 (82)	130 (109)	
	1,2-Dichloroethane	20	21.5	ug/L	108			70 (80)	130 (115)	
	Trichloroethene	20	19.9	ug/L	100			70 (77)	130 (113)	
	1,2-Dichloropropane	20	20.9	ug/L	104			70 (83)	130 (111)	
	Bromodichloromethane	20	21.4	ug/L	107			70 (83)	130 (110)	
	4-Methyl-2-Pentanone	100	110	ug/L	110			40 (74)	160 (118)	
	Toluene	20	20.5	ug/L	103			70 (82)	130 (110)	
	t-1,3-Dichloropropene	20	19.4	ug/L	97			70 (79)	130 (110)	
	cis-1,3-Dichloropropene	20	21.5	ug/L	108			70 (82)	130 (110)	
	1,1,2-Trichloroethane	20	21.9	ug/L	110			70 (83)	130 (112)	
	2-Hexanone	100	110	ug/L	110			40 (73)	160 (117)	
	Dibromochloromethane	20	21.1	ug/L	106			70 (82)	130 (110)	
	1,2-Dibromoethane	20	21.7	ug/L	109			70 (81)	130 (110)	
	Tetrachloroethene	20	20.0	ug/L	100			70 (67)	130 (123)	
	Chlorobenzene	20	20.8	ug/L	104			70 (82)	130 (109)	
	Ethyl Benzene	20	20.9	ug/L	104			70 (83)	130 (109)	
	m/p-Xylenes	40	41.7	ug/L	104			70 (82)	130 (110)	
	o-Xylene	20	20.9	ug/L	104			70 (83)	130 (109)	
	Styrene	20	21.4	ug/L	107			70 (80)	130 (111)	
	Bromoform	20	21.2	ug/L	106			70 (79)	130 (109)	
	Isopropylbenzene	20	20.4	ug/L	102			70 (83)	130 (112)	
	1,1,2,2-Tetrachloroethane	20	20.9	ug/L	104			70 (76)	130 (118)	
	1,3-Dichlorobenzene	20	20.1	ug/L	101			70 (82)	130 (108)	
	1,4-Dichlorobenzene	20	19.8	ug/L	99			70 (82)	130 (107)	
	1,2-Dichlorobenzene	20	21.2	ug/L	106			70 (82)	130 (109)	

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: SW8260-Low

Datafile : VX045975.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0429WBS01	1,2-Dibromo-3-Chloropropane	20	21.5	ug/L	108			40 (68)	160 (112)	
	1,2,4-Trichlorobenzene	20	19.5	ug/L	98			70 (75)	130 (113)	
	1,2,3-Trichlorobenzene	20	20.1	ug/L	101			70 (76)	130 (114)	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: SW8260-Low

Datafile : VX045976.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0429WBSD01	Dichlorodifluoromethane	20	17.3	ug/L	86	1		40 (69)	160 (116)	20 (20)
	Chloromethane	20	16.4	ug/L	82	1		40 (65)	160 (116)	20 (20)
	Vinyl chloride	20	16.9	ug/L	85	0		70 (65)	130 (117)	20 (20)
	Bromomethane	20	17.3	ug/L	86	2		40 (58)	160 (125)	20 (20)
	Chloroethane	20	18.9	ug/L	95	7		40 (56)	160 (128)	20 (20)
	Trichlorofluoromethane	20	19.5	ug/L	98	1		40 (73)	160 (115)	20 (20)
	1,1,2-Trichlorotrifluoroethane	20	21.3	ug/L	106	4		70 (80)	130 (112)	20 (20)
	1,1-Dichloroethene	20	19.2	ug/L	96	3		70 (74)	130 (110)	20 (20)
	Acetone	100	100	ug/L	100	2		40 (60)	160 (125)	20 (20)
	Carbon disulfide	20	13.4	ug/L	67	0		40 (64)	160 (112)	20 (20)
	Methyl tert-butyl Ether	20	21.9	ug/L	110	4		70 (78)	130 (114)	20 (20)
	Methyl Acetate	20	30.0	ug/L	150	5	*	70 (67)	130 (125)	20 (20)
	Methylene Chloride	20	20.4	ug/L	102	6		70 (72)	130 (114)	20 (20)
	trans-1,2-Dichloroethene	20	18.6	ug/L	93	4		70 (75)	130 (108)	20 (20)
	1,1-Dichloroethane	20	20.0	ug/L	100	1		70 (78)	130 (112)	20 (20)
	Cyclohexane	20	18.4	ug/L	92	2		70 (75)	130 (110)	20 (20)
	2-Butanone	100	110	ug/L	110	10		40 (65)	160 (122)	20 (20)
	Carbon Tetrachloride	20	21.2	ug/L	106	0		70 (77)	130 (113)	20 (20)
	cis-1,2-Dichloroethene	20	20.5	ug/L	103	3		70 (77)	130 (110)	20 (20)
	Bromochloromethane	20	22.9	ug/L	115	4		70 (70)	130 (124)	20 (20)
	Chloroform	20	21.5	ug/L	108	2		70 (79)	130 (113)	20 (20)
	1,1,1-Trichloroethane	20	20.9	ug/L	104	1		70 (80)	130 (108)	20 (20)
	Methylcyclohexane	20	19.9	ug/L	100	6		70 (72)	130 (115)	20 (20)
	Benzene	20	19.8	ug/L	99	0		70 (82)	130 (109)	20 (20)
	1,2-Dichloroethane	20	22.2	ug/L	111	3		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	19.6	ug/L	98	2		70 (77)	130 (113)	20 (20)
	1,2-Dichloropropane	20	21.7	ug/L	109	5		70 (83)	130 (111)	20 (20)
	Bromodichloromethane	20	21.4	ug/L	107	0		70 (83)	130 (110)	20 (20)
	4-Methyl-2-Pentanone	100	110	ug/L	110	0		40 (74)	160 (118)	20 (20)
	Toluene	20	20.1	ug/L	101	2		70 (82)	130 (110)	20 (20)
	t-1,3-Dichloropropene	20	20.3	ug/L	102	5		70 (79)	130 (110)	20 (20)
	cis-1,3-Dichloropropene	20	21.3	ug/L	106	2		70 (82)	130 (110)	20 (20)
	1,1,2-Trichloroethane	20	21.6	ug/L	108	2		70 (83)	130 (112)	20 (20)
	2-Hexanone	100	110	ug/L	110	0		40 (73)	160 (117)	20 (20)
	Dibromochloromethane	20	22.1	ug/L	111	5		70 (82)	130 (110)	20 (20)
	1,2-Dibromoethane	20	22.1	ug/L	111	2		70 (81)	130 (110)	20 (20)
	Tetrachloroethene	20	19.7	ug/L	99	1		70 (67)	130 (123)	20 (20)
	Chlorobenzene	20	21.5	ug/L	108	4		70 (82)	130 (109)	20 (20)
	Ethyl Benzene	20	21.2	ug/L	106	2		70 (83)	130 (109)	20 (20)
	m/p-Xylenes	40	42.6	ug/L	106	2		70 (82)	130 (110)	20 (20)
	o-Xylene	20	21.5	ug/L	108	4		70 (83)	130 (109)	20 (20)
	Styrene	20	21.7	ug/L	109	2		70 (80)	130 (111)	20 (20)
	Bromoform	20	21.2	ug/L	106	0		70 (79)	130 (109)	20 (20)
	Isopropylbenzene	20	19.9	ug/L	100	2		70 (83)	130 (112)	20 (20)
	1,1,2,2-Tetrachloroethane	20	21.0	ug/L	105	1		70 (76)	130 (118)	20 (20)
	1,3-Dichlorobenzene	20	20.2	ug/L	101	0		70 (82)	130 (108)	20 (20)
	1,4-Dichlorobenzene	20	20.0	ug/L	100	1		70 (82)	130 (107)	20 (20)
	1,2-Dichlorobenzene	20	21.2	ug/L	106	0		70 (82)	130 (109)	20 (20)

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: SW8260-Low

Datafile : VX045976.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0429WBSD01	1,2-Dibromo-3-Chloropropane	20	22.3	ug/L	112	4		40 (68)	160 (112)	20 (20)
	1,2,4-Trichlorobenzene	20	20.9	ug/L	104	6		70 (75)	130 (113)	20 (20)
	1,2,3-Trichlorobenzene	20	20.3	ug/L	102	1		70 (76)	130 (114)	20 (20)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: SW8260D

Datafile : VY022072.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0430SBS01	Dichlorodifluoromethane	20	19.3	ug/Kg	97			40 (64)	160 (136)	
	Chloromethane	20	22.2	ug/Kg	111			40 (70)	160 (130)	
	Vinyl chloride	20	20.7	ug/Kg	104			70 (72)	130 (129)	
	Bromomethane	20	22.8	ug/Kg	114			40 (58)	160 (141)	
	Chloroethane	20	21.2	ug/Kg	106			40 (69)	160 (130)	
	Trichlorofluoromethane	20	21.8	ug/Kg	109			40 (69)	160 (134)	
	1,1,2-Trichlorotrifluoroethane	20	20.4	ug/Kg	102			70 (81)	130 (123)	
	1,1-Dichloroethene	20	19.7	ug/Kg	99			70 (79)	130 (121)	
	Acetone	100	94.5	ug/Kg	95			40 (60)	160 (131)	
	Carbon disulfide	20	19.9	ug/Kg	100			40 (45)	160 (154)	
	Methyl tert-butyl Ether	20	19.0	ug/Kg	95			70 (77)	130 (129)	
	Methyl Acetate	20	20.3	ug/Kg	102			70 (69)	130 (149)	
	Methylene Chloride	20	20.0	ug/Kg	100			70 (56)	130 (174)	
	trans-1,2-Dichloroethene	20	19.5	ug/Kg	98			70 (80)	130 (123)	
	1,1-Dichloroethane	20	19.9	ug/Kg	100			70 (82)	130 (123)	
	Cyclohexane	20	18.9	ug/Kg	95			70 (76)	130 (122)	
	2-Butanone	100	92.7	ug/Kg	93			40 (69)	160 (131)	
	Carbon Tetrachloride	20	21.3	ug/Kg	106			70 (76)	130 (129)	
	cis-1,2-Dichloroethene	20	19.5	ug/Kg	98			70 (82)	130 (123)	
	Bromochloromethane	20	20.0	ug/Kg	100			70 (80)	130 (127)	
	Chloroform	20	20.4	ug/Kg	102			70 (82)	130 (125)	
	1,1,1-Trichloroethane	20	20.1	ug/Kg	101			70 (80)	130 (126)	
	Methylcyclohexane	20	19.2	ug/Kg	96			70 (77)	130 (123)	
	Benzene	20	20.6	ug/Kg	103			70 (84)	130 (121)	
	1,2-Dichloroethane	20	21.2	ug/Kg	106			70 (81)	130 (126)	
	Trichloroethene	20	20.9	ug/Kg	104			70 (83)	130 (122)	
	1,2-Dichloropropane	20	20.3	ug/Kg	102			70 (83)	130 (122)	
	Bromodichloromethane	20	20.9	ug/Kg	104			70 (82)	130 (123)	
	4-Methyl-2-Pentanone	100	100	ug/Kg	100			40 (70)	160 (135)	
	Toluene	20	21.0	ug/Kg	105			70 (83)	130 (122)	
	t-1,3-Dichloropropene	20	20.6	ug/Kg	103			70 (78)	130 (124)	
	cis-1,3-Dichloropropene	20	20.6	ug/Kg	103			70 (81)	130 (122)	
	1,1,2-Trichloroethane	20	20.9	ug/Kg	104			70 (82)	130 (125)	
	2-Hexanone	100	99.9	ug/Kg	100			40 (66)	160 (138)	
	Dibromochloromethane	20	21.1	ug/Kg	106			70 (79)	130 (125)	
	1,2-Dibromoethane	20	20.6	ug/Kg	103			70 (80)	130 (125)	
	Tetrachloroethene	20	21.1	ug/Kg	106			70 (83)	130 (125)	
	Chlorobenzene	20	20.1	ug/Kg	101			70 (84)	130 (122)	
	Ethyl Benzene	20	19.3	ug/Kg	97			70 (82)	130 (124)	
	m/p-Xylenes	40	40.6	ug/Kg	102			70 (83)	130 (124)	
	o-Xylene	20	19.7	ug/Kg	99			70 (83)	130 (123)	
	Styrene	20	20.0	ug/Kg	100			70 (82)	130 (124)	
	Bromoform	20	20.9	ug/Kg	104			70 (75)	130 (127)	
	Isopropylbenzene	20	18.8	ug/Kg	94			70 (82)	130 (124)	
	1,1,2,2-Tetrachloroethane	20	19.7	ug/Kg	99			70 (77)	130 (127)	
	1,3-Dichlorobenzene	20	19.9	ug/Kg	100			70 (83)	130 (122)	
	1,4-Dichlorobenzene	20	19.7	ug/Kg	99			70 (84)	130 (121)	
	1,2-Dichlorobenzene	20	20.1	ug/Kg	101			70 (83)	130 (124)	

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: SW8260D

Datafile : VY022072.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0430SBS01	1,2-Dibromo-3-Chloropropane	20	18.2	ug/Kg	91			40 (66)	160 (134)	
	1,2,4-Trichlorobenzene	20	19.2	ug/Kg	96			70 (78)	130 (127)	
	1,2,3-Trichlorobenzene	20	19.7	ug/Kg	99			70 (70)	130 (137)	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: SW8260D

Datafile : VY022073.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0430SBSD01	Dichlorodifluoromethane	20	20.3	ug/Kg	102	5		40 (64)	160 (136)	30 (20)
	Chloromethane	20	19.5	ug/Kg	98	12		40 (70)	160 (130)	30 (20)
	Vinyl chloride	20	19.7	ug/Kg	99	5		70 (72)	130 (129)	30 (20)
	Bromomethane	20	19.5	ug/Kg	98	15		40 (58)	160 (141)	30 (20)
	Chloroethane	20	19.5	ug/Kg	98	8		40 (69)	160 (130)	30 (20)
	Trichlorofluoromethane	20	20.4	ug/Kg	102	7		40 (69)	160 (134)	30 (20)
	1,1,2-Trichlorotrifluoroethane	20	20.4	ug/Kg	102	0		70 (81)	130 (123)	30 (20)
	1,1-Dichloroethene	20	20.2	ug/Kg	101	2		70 (79)	130 (121)	30 (20)
	Acetone	100	89.4	ug/Kg	89	7		40 (60)	160 (131)	30 (20)
	Carbon disulfide	20	19.7	ug/Kg	99	1		40 (45)	160 (154)	30 (20)
	Methyl tert-butyl Ether	20	18.6	ug/Kg	93	2		70 (77)	130 (129)	30 (20)
	Methyl Acetate	20	21.5	ug/Kg	108	6		70 (69)	130 (149)	30 (20)
	Methylene Chloride	20	20.1	ug/Kg	101	1		70 (56)	130 (174)	30 (20)
	trans-1,2-Dichloroethene	20	20.0	ug/Kg	100	2		70 (80)	130 (123)	30 (20)
	1,1-Dichloroethane	20	19.8	ug/Kg	99	1		70 (82)	130 (123)	30 (20)
	Cyclohexane	20	18.9	ug/Kg	95	0		70 (76)	130 (122)	30 (20)
	2-Butanone	100	91.4	ug/Kg	91	2		40 (69)	160 (131)	30 (20)
	Carbon Tetrachloride	20	20.7	ug/Kg	104	2		70 (76)	130 (129)	30 (20)
	cis-1,2-Dichloroethene	20	19.7	ug/Kg	99	1		70 (82)	130 (123)	30 (20)
	Bromochloromethane	20	20.0	ug/Kg	100	0		70 (80)	130 (127)	30 (20)
	Chloroform	20	20.1	ug/Kg	101	1		70 (82)	130 (125)	30 (20)
	1,1,1-Trichloroethane	20	20.2	ug/Kg	101	0		70 (80)	130 (126)	30 (20)
	Methylcyclohexane	20	19.0	ug/Kg	95	1		70 (77)	130 (123)	30 (20)
	Benzene	20	20.5	ug/Kg	103	0		70 (84)	130 (121)	30 (20)
	1,2-Dichloroethane	20	20.5	ug/Kg	103	3		70 (81)	130 (126)	30 (20)
	Trichloroethene	20	20.5	ug/Kg	103	1		70 (83)	130 (122)	30 (20)
	1,2-Dichloropropane	20	20.4	ug/Kg	102	0		70 (83)	130 (122)	30 (20)
	Bromodichloromethane	20	20.6	ug/Kg	103	1		70 (82)	130 (123)	30 (20)
	4-Methyl-2-Pentanone	100	98.5	ug/Kg	99	1		40 (70)	160 (135)	30 (20)
	Toluene	20	20.5	ug/Kg	103	2		70 (83)	130 (122)	30 (20)
	t-1,3-Dichloropropene	20	20.2	ug/Kg	101	2		70 (78)	130 (124)	30 (20)
	cis-1,3-Dichloropropene	20	20.1	ug/Kg	101	2		70 (81)	130 (122)	30 (20)
	1,1,2-Trichloroethane	20	20.6	ug/Kg	103	1		70 (82)	130 (125)	30 (20)
	2-Hexanone	100	94.3	ug/Kg	94	6		40 (66)	160 (138)	30 (20)
	Dibromochloromethane	20	20.2	ug/Kg	101	5		70 (79)	130 (125)	30 (20)
	1,2-Dibromoethane	20	20.1	ug/Kg	101	2		70 (80)	130 (125)	30 (20)
	Tetrachloroethene	20	21.3	ug/Kg	106	0		70 (83)	130 (125)	30 (20)
	Chlorobenzene	20	20.3	ug/Kg	102	1		70 (84)	130 (122)	30 (20)
	Ethyl Benzene	20	19.4	ug/Kg	97	0		70 (82)	130 (124)	30 (20)
	m/p-Xylenes	40	40.3	ug/Kg	101	1		70 (83)	130 (124)	30 (20)
	o-Xylene	20	19.4	ug/Kg	97	2		70 (83)	130 (123)	30 (20)
	Styrene	20	20.3	ug/Kg	102	2		70 (82)	130 (124)	30 (20)
	Bromoform	20	20.3	ug/Kg	102	2		70 (75)	130 (127)	30 (20)
	Isopropylbenzene	20	19.3	ug/Kg	97	3		70 (82)	130 (124)	30 (20)
	1,1,2,2-Tetrachloroethane	20	19.4	ug/Kg	97	2		70 (77)	130 (127)	30 (20)
	1,3-Dichlorobenzene	20	20.0	ug/Kg	100	0		70 (83)	130 (122)	30 (20)
	1,4-Dichlorobenzene	20	20.0	ug/Kg	100	1		70 (84)	130 (121)	30 (20)
	1,2-Dichlorobenzene	20	20.4	ug/Kg	102	1		70 (83)	130 (124)	30 (20)

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: SW8260D

Datafile : VY022073.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0430SBSD01	1,2-Dibromo-3-Chloropropane	20	18.4	ug/Kg	92	1		40 (66)	160 (134)	30 (20)
	1,2,4-Trichlorobenzene	20	18.8	ug/Kg	94	2		70 (78)	130 (127)	30 (20)
	1,2,3-Trichlorobenzene	20	18.9	ug/Kg	95	4		70 (70)	130 (137)	30 (20)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: SW8260D

Datafile : VY022097.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0501SBS01	Dichlorodifluoromethane	20	19.2	ug/Kg	96			40 (64)	160 (136)	
	Chloromethane	20	18.7	ug/Kg	94			40 (70)	160 (130)	
	Vinyl chloride	20	18.7	ug/Kg	94			70 (72)	130 (129)	
	Bromomethane	20	18.6	ug/Kg	93			40 (58)	160 (141)	
	Chloroethane	20	18.9	ug/Kg	95			40 (69)	160 (130)	
	Trichlorofluoromethane	20	19.9	ug/Kg	100			40 (69)	160 (134)	
	1,1,2-Trichlorotrifluoroethane	20	19.8	ug/Kg	99			70 (81)	130 (123)	
	1,1-Dichloroethene	20	19.9	ug/Kg	100			70 (79)	130 (121)	
	Acetone	100	91.2	ug/Kg	91			40 (60)	160 (131)	
	Carbon disulfide	20	19.5	ug/Kg	98			40 (45)	160 (154)	
	Methyl tert-butyl Ether	20	18.0	ug/Kg	90			70 (77)	130 (129)	
	Methyl Acetate	20	18.7	ug/Kg	94			70 (69)	130 (149)	
	Methylene Chloride	20	19.3	ug/Kg	97			70 (56)	130 (174)	
	trans-1,2-Dichloroethene	20	19.7	ug/Kg	99			70 (80)	130 (123)	
	1,1-Dichloroethane	20	19.8	ug/Kg	99			70 (82)	130 (123)	
	Cyclohexane	20	18.8	ug/Kg	94			70 (76)	130 (122)	
	2-Butanone	100	86.7	ug/Kg	87			40 (69)	160 (131)	
	Carbon Tetrachloride	20	20.6	ug/Kg	103			70 (76)	130 (129)	
	cis-1,2-Dichloroethene	20	19.8	ug/Kg	99			70 (82)	130 (123)	
	Bromochloromethane	20	18.9	ug/Kg	95			70 (80)	130 (127)	
	Chloroform	20	20.2	ug/Kg	101			70 (82)	130 (125)	
	1,1,1-Trichloroethane	20	20.1	ug/Kg	101			70 (80)	130 (126)	
	Methylcyclohexane	20	19.0	ug/Kg	95			70 (77)	130 (123)	
	Benzene	20	20.2	ug/Kg	101			70 (84)	130 (121)	
	1,2-Dichloroethane	20	19.9	ug/Kg	100			70 (81)	130 (126)	
	Trichloroethene	20	20.3	ug/Kg	102			70 (83)	130 (122)	
	1,2-Dichloropropane	20	20.1	ug/Kg	101			70 (83)	130 (122)	
	Bromodichloromethane	20	20.6	ug/Kg	103			70 (82)	130 (123)	
	4-Methyl-2-Pentanone	100	88.1	ug/Kg	88			40 (70)	160 (135)	
	Toluene	20	20.6	ug/Kg	103			70 (83)	130 (122)	
	t-1,3-Dichloropropene	20	18.9	ug/Kg	95			70 (78)	130 (124)	
	cis-1,3-Dichloropropene	20	19.9	ug/Kg	100			70 (81)	130 (122)	
	1,1,2-Trichloroethane	20	20.2	ug/Kg	101			70 (82)	130 (125)	
	2-Hexanone	100	84.7	ug/Kg	85			40 (66)	160 (138)	
	Dibromochloromethane	20	19.8	ug/Kg	99			70 (79)	130 (125)	
	1,2-Dibromoethane	20	19.9	ug/Kg	100			70 (80)	130 (125)	
	Tetrachloroethene	20	20.9	ug/Kg	104			70 (83)	130 (125)	
	Chlorobenzene	20	19.8	ug/Kg	99			70 (84)	130 (122)	
	Ethyl Benzene	20	19.5	ug/Kg	98			70 (82)	130 (124)	
	m/p-Xylenes	40	40.5	ug/Kg	101			70 (83)	130 (124)	
	o-Xylene	20	19.6	ug/Kg	98			70 (83)	130 (123)	
	Styrene	20	20.1	ug/Kg	101			70 (82)	130 (124)	
	Bromoform	20	19.6	ug/Kg	98			70 (75)	130 (127)	
	Isopropylbenzene	20	19.2	ug/Kg	96			70 (82)	130 (124)	
	1,1,2,2-Tetrachloroethane	20	18.4	ug/Kg	92			70 (77)	130 (127)	
	1,3-Dichlorobenzene	20	19.9	ug/Kg	100			70 (83)	130 (122)	
	1,4-Dichlorobenzene	20	19.6	ug/Kg	98			70 (84)	130 (121)	
	1,2-Dichlorobenzene	20	19.4	ug/Kg	97			70 (83)	130 (124)	

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1901

Client: Portal Partners Tri-Venture

Analytical Method: SW8260D

Datafile : VY022097.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0501SBS01	1,2-Dibromo-3-Chloropropane	20	17.7	ug/Kg	89			40 (66)	160 (134)	
	1,2,4-Trichlorobenzene	20	19.0	ug/Kg	95			70 (78)	130 (127)	
	1,2,3-Trichlorobenzene	20	18.8	ug/Kg	94			70 (70)	130 (137)	

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0429WBL01

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM Case No.: Q1901

SAS No.: Q1901 SDG NO.: Q1901

Lab File ID: VX045974.D

Lab Sample ID: VX0429WBL01

Date Analyzed: 04/29/2025

Time Analyzed: 11:58

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0429WBS01	VX0429WBS01	VX045975.D	04/29/2025
VX0429WBSD01	VX0429WBSD01	VX045976.D	04/29/2025
TB04262025	Q1901-07	VX045978.D	04/29/2025
FB04262025	Q1901-06	VX045979.D	04/29/2025

COMMENTS: _____

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VY0430SBL01

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM Case No.: Q1901

SAS No.: Q1901 SDG NO.: Q1901

Lab File ID: VY022071.D

Lab Sample ID: VY0430SBL01

Date Analyzed: 04/30/2025

Time Analyzed: 09:57

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) Y

Instrument ID: MSVOA_Y

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VY0430SBS01	VY0430SBS01	VY022072.D	04/30/2025
VY0430SBSD01	VY0430SBSD01	VY022073.D	04/30/2025
B-167-SB01	Q1901-02	VY022087.D	04/30/2025
B-167-SB02	Q1901-04	VY022089.D	04/30/2025
B-170-SB02	Q1901-05	VY022090.D	04/30/2025

COMMENTS: _____

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VY0501SBL01

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM Case No.: Q1901

SAS No.: Q1901 SDG NO.: Q1901

Lab File ID: VY022096.D

Lab Sample ID: VY0501SBL01

Date Analyzed: 05/01/2025

Time Analyzed: 10:11

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) Y

Instrument ID: MSVOA_Y

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VY0501SBS01	VY0501SBS01	VY022097.D	05/01/2025
B-170-SB00	Q1901-01	VY022105.D	05/01/2025
B-170-SB01	Q1901-03	VY022106.D	05/01/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901
Lab File ID: VX045524.D BFB Injection Date: 04/01/2025
Instrument ID: MSVOA_X BFB Injection Time: 16:15
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.6
75	30.0 - 60.0% of mass 95	58.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	0.8 (1.2) 1
174	50.0 - 100.0% of mass 95	66.8
175	5.0 - 9.0% of mass 174	5.2 (7.8) 1
176	95.0 - 101.0% of mass 174	65.3 (97.8) 1
177	5.0 - 9.0% of mass 176	4.4 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX045525.D	04/01/2025	17:06
VSTDIC005	VSTDIC005	VX045526.D	04/01/2025	17:29
VSTDIC020	VSTDIC020	VX045527.D	04/01/2025	17:52
VSTDIC050	VSTDIC050	VX045528.D	04/01/2025	18:15
VSTDIC100	VSTDIC100	VX045529.D	04/01/2025	18:38
VSTDIC150	VSTDIC150	VX045530.D	04/01/2025	19:02

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901
Lab File ID: VX045971.D BFB Injection Date: 04/29/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:35
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.5
75	30.0 - 60.0% of mass 95	57.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.8 (1.2) 1
174	50.0 - 100.0% of mass 95	66.5
175	5.0 - 9.0% of mass 174	4.8 (7.3) 1
176	95.0 - 101.0% of mass 174	64.2 (96.5) 1
177	5.0 - 9.0% of mass 176	4.5 (7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX045972.D	04/29/2025	11:06
VX0429WBL01	VX0429WBL01	VX045974.D	04/29/2025	11:58
VX0429WBS01	VX0429WBS01	VX045975.D	04/29/2025	12:21
VX0429WBSD01	VX0429WBSD01	VX045976.D	04/29/2025	12:49
TB04262025	Q1901-07	VX045978.D	04/29/2025	13:35
FB04262025	Q1901-06	VX045979.D	04/29/2025	13:58

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901
Lab File ID: VY021952.D BFB Injection Date: 04/22/2025
Instrument ID: MSVOA_Y BFB Injection Time: 11:33
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	16.9
75	30.0 - 60.0% of mass 95	49.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	1.5 (1.7) 1
174	50.0 - 100.0% of mass 95	88.4
175	5.0 - 9.0% of mass 174	7.1 (8) 1
176	95.0 - 101.0% of mass 174	84.9 (96.1) 1
177	5.0 - 9.0% of mass 176	5.5 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VY021953.D	04/22/2025	13:39
VSTDIC010	VSTDIC010	VY021954.D	04/22/2025	14:44
VSTDIC020	VSTDIC020	VY021955.D	04/22/2025	15:07
VSTDIC050	VSTDIC050	VY021956.D	04/22/2025	15:29
VSTDIC100	VSTDIC100	VY021957.D	04/22/2025	15:52
VSTDIC150	VSTDIC150	VY021958.D	04/22/2025	16:15

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901
Lab File ID: VY022069.D BFB Injection Date: 04/30/2025
Instrument ID: MSVOA_Y BFB Injection Time: 08:50
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	15.8
75	30.0 - 60.0% of mass 95	48.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	1.6 (1.8) 1
174	50.0 - 100.0% of mass 95	90.5
175	5.0 - 9.0% of mass 174	7 (7.8) 1
176	95.0 - 101.0% of mass 174	87.9 (97.2) 1
177	5.0 - 9.0% of mass 176	5.4 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY022070.D	04/30/2025	09:20
VY0430SBL01	VY0430SBL01	VY022071.D	04/30/2025	09:57
VY0430SBS01	VY0430SBS01	VY022072.D	04/30/2025	10:30
VY0430SBSD01	VY0430SBSD01	VY022073.D	04/30/2025	10:52
B-167-SB01	Q1901-02	VY022087.D	04/30/2025	17:05
B-167-SB02	Q1901-04	VY022089.D	04/30/2025	17:52
B-170-SB02	Q1901-05	VY022090.D	04/30/2025	18:16

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901
Lab File ID: VY022094.D BFB Injection Date: 05/01/2025
Instrument ID: MSVOA_Y BFB Injection Time: 08:20
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	15.8
75	30.0 - 60.0% of mass 95	47.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	1.5 (1.7) 1
174	50.0 - 100.0% of mass 95	87.7
175	5.0 - 9.0% of mass 174	6.9 (7.8) 1
176	95.0 - 101.0% of mass 174	83.4 (95.1) 1
177	5.0 - 9.0% of mass 176	5.6 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY022095.D	05/01/2025	08:50
VY0501SBL01	VY0501SBL01	VY022096.D	05/01/2025	10:11
VY0501SBS01	VY0501SBS01	VY022097.D	05/01/2025	10:42
B-170-SB00	Q1901-01	VY022105.D	05/01/2025	14:15
B-170-SB01	Q1901-03	VY022106.D	05/01/2025	14:38

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901
 Lab File ID: VX045972.D Date Analyzed: 04/29/2025
 Instrument ID: MSVOA_X Time Analyzed: 11:06
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	79737	5.54	136469	6.76	117675	10.05
UPPER LIMIT	159474	6.044	272938	7.257	235350	10.549
LOWER LIMIT	39868.5	5.044	68234.5	6.257	58837.5	9.549
EPA SAMPLE NO.						
FB04262025	64509	5.54	127954	6.76	115874	10.05
TB04262025	67274	5.54	131446	6.76	123412	10.05
VX0429WBL01	63902	5.55	127024	6.76	116648	10.05
VX0429WBS01	80632	5.55	140487	6.76	122568	10.05
VX0429WBSD01	77095	5.54	135079	6.76	118384	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901
Lab File ID: VX045972.D Date Analyzed: 04/29/2025
Instrument ID: MSVOA_X Time Analyzed: 11:06
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	59672	12.018				
UPPER LIMIT	119344	12.518				
LOWER LIMIT	29836	11.518				
EPA SAMPLE NO.						
FB04262025	48370	12.02				
TB04262025	53777	12.02				
VX0429WBL01	47876	12.02				
VX0429WBS01	59236	12.02				
VX0429WBSD01	57470	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901
Lab File ID: VY022070.D Date Analyzed: 04/30/2025
Instrument ID: MSVOA_Y Time Analyzed: 09:20
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	299393	7.71	438020	8.62	420486	11.41
UPPER LIMIT	598786	8.207	876040	9.116	840972	11.914
LOWER LIMIT	149697	7.207	219010	8.116	210243	10.914
EPA SAMPLE NO.						
B-167-SB01	348376	7.71	652384	8.61	556238	11.41
B-167-SB02	381091	7.71	712829	8.61	629182	11.41
B-170-SB02	384809	7.71	709625	8.61	563942	11.41
VY0430SBL01	311598	7.71	580894	8.62	511027	11.41
VY0430SBS01	286487	7.71	433702	8.62	405299	11.41
VY0430SBSD01	289313	7.71	439988	8.61	406945	11.41

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901
 Lab File ID: VY022070.D Date Analyzed: 04/30/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:20
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	233640	13.346				
UPPER LIMIT	467280	13.846				
LOWER LIMIT	116820	12.846				
EPA SAMPLE NO.						
B-167-SB01	176600	13.35				
B-167-SB02	233488	13.35				
B-170-SB02	150773	13.35				
VY0430SBL01	186896	13.35				
VY0430SBS01	222589	13.35				
VY0430SBSD01	218185	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901
Lab File ID: VY022095.D Date Analyzed: 05/01/2025
Instrument ID: MSVOA_Y Time Analyzed: 08:50
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	300031	7.71	433614	8.62	410078	11.42
UPPER LIMIT	600062	8.207	867228	9.116	820156	11.92
LOWER LIMIT	150016	7.207	216807	8.116	205039	10.92
EPA SAMPLE NO.						
B-170-SB00	310481	7.71	585270	8.62	519369	11.41
B-170-SB01	307701	7.71	584162	8.62	521346	11.42
VY0501SBL01	317996	7.71	585580	8.62	508953	11.42
VY0501SBS01	294485	7.71	452802	8.62	414235	11.41

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901
 Lab File ID: VY022095.D Date Analyzed: 05/01/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:50
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	227184	13.347				
UPPER LIMIT	454368	13.847				
LOWER LIMIT	113592	12.847				
EPA SAMPLE NO.						
B-170-SB00	196391	13.35				
B-170-SB01	197125	13.35				
VY0501SBL01	180066	13.35				
VY0501SBS01	225406	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	
Client Sample ID:	VX0429WBL01	SDG No.:	Q1901
Lab Sample ID:	VX0429WBL01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045974.D	1		04/29/25 11:58	VX042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	1.50	U	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	
Client Sample ID:	VX0429WBL01		SDG No.:	Q1901
Lab Sample ID:	VX0429WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045974.D	1		04/29/25 11:58	VX042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	U	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	U	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	55.1		70 (74) - 130 (125)	110%	SPK: 50
1868-53-7	Dibromofluoromethane	51.6		70 (75) - 130 (124)	103%	SPK: 50
2037-26-5	Toluene-d8	49.7		70 (86) - 130 (113)	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.9		70 (77) - 130 (121)	100%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	63900	5.55			
540-36-3	1,4-Difluorobenzene	127000	6.757			
3114-55-4	Chlorobenzene-d5	117000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	47900	12.018			

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	
Client Sample ID:	VX0429WBL01		SDG No.:	Q1901
Lab Sample ID:	VX0429WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045974.D	1		04/29/25 11:58	VX042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	
Client Sample ID:	VY0430SBL01	SDG No.:	Q1901
Lab Sample ID:	VY0430SBL01	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022071.D	1		04/30/25 09:57	VY043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.10	U	1.10	5.00	ug/Kg
74-87-3	Chloromethane	1.10	U	1.10	5.00	ug/Kg
75-01-4	Vinyl Chloride	0.79	U	0.79	5.00	ug/Kg
74-83-9	Bromomethane	1.10	U	1.10	5.00	ug/Kg
75-00-3	Chloroethane	1.30	U	1.30	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	1.20	U	1.20	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	1.10	U	1.10	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	1.00	U	1.00	5.00	ug/Kg
67-64-1	Acetone	4.70	U	4.70	25.0	ug/Kg
75-15-0	Carbon Disulfide	1.10	U	1.10	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.73	U	0.73	5.00	ug/Kg
79-20-9	Methyl Acetate	1.50	U	1.50	5.00	ug/Kg
75-09-2	Methylene Chloride	3.50	U	3.50	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.86	U	0.86	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	0.80	U	0.80	5.00	ug/Kg
110-82-7	Cyclohexane	0.79	U	0.79	5.00	ug/Kg
78-93-3	2-Butanone	6.50	U	6.50	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	0.97	U	0.97	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.75	5.00	ug/Kg
74-97-5	Bromochloromethane	1.20	U	1.20	5.00	ug/Kg
67-66-3	Chloroform	0.84	U	0.84	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.93	U	0.93	5.00	ug/Kg
108-87-2	Methylcyclohexane	0.91	U	0.91	5.00	ug/Kg
71-43-2	Benzene	0.79	U	0.79	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	0.79	U	0.79	5.00	ug/Kg
79-01-6	Trichloroethene	0.81	U	0.81	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	0.91	U	0.91	5.00	ug/Kg
75-27-4	Bromodichloromethane	0.78	U	0.78	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.60	U	3.60	25.0	ug/Kg
108-88-3	Toluene	0.78	U	0.78	5.00	ug/Kg

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	
Client Sample ID:	VY0430SBL01	SDG No.:	Q1901
Lab Sample ID:	VY0430SBL01	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022071.D	1		04/30/25 09:57	VY043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.65	U	0.65	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.62	U	0.62	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.92	U	0.92	5.00	ug/Kg
591-78-6	2-Hexanone	3.70	U	3.70	25.0	ug/Kg
124-48-1	Dibromochloromethane	0.87	U	0.87	5.00	ug/Kg
106-93-4	1,2-Dibromoethane	0.88	U	0.88	5.00	ug/Kg
127-18-4	Tetrachloroethene	1.10	U	1.10	5.00	ug/Kg
108-90-7	Chlorobenzene	0.91	U	0.91	5.00	ug/Kg
100-41-4	Ethyl Benzene	0.67	U	0.67	5.00	ug/Kg
179601-23-1	m/p-Xylenes	1.20	U	1.20	10.0	ug/Kg
95-47-6	o-Xylene	0.82	U	0.82	5.00	ug/Kg
100-42-5	Styrene	0.71	U	0.71	5.00	ug/Kg
75-25-2	Bromoform	0.86	U	0.86	5.00	ug/Kg
98-82-8	Isopropylbenzene	0.78	U	0.78	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.20	U	1.20	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.70	U	1.70	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.60	U	1.60	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.50	U	1.50	5.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.80	U	1.80	5.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	3.00	U	3.00	5.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	3.20	U	3.20	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.3		70 (63) - 130 (155)	105%	SPK: 50
1868-53-7	Dibromofluoromethane	51.8		70 (70) - 130 (134)	104%	SPK: 50
2037-26-5	Toluene-d8	48.2		70 (74) - 130 (123)	96%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.3		70 (38) - 130 (136)	107%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	312000	7.707			
540-36-3	1,4-Difluorobenzene	581000	8.616			
3114-55-4	Chlorobenzene-d5	511000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	187000	13.347			

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	
Client Sample ID:	VY0430SBL01		SDG No.:	Q1901
Lab Sample ID:	VY0430SBL01		Matrix:	SOIL
Analytical Method:	SW8260		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022071.D	1		04/30/25 09:57	VY043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	
Client Sample ID:	VY0501SBL01	SDG No.:	Q1901
Lab Sample ID:	VY0501SBL01	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	100
Sample Wt/Vol:	5	Units:	g
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022096.D	1		05/01/25 10:11	VY050125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.10	U	1.10	5.00	ug/Kg
74-87-3	Chloromethane	1.10	U	1.10	5.00	ug/Kg
75-01-4	Vinyl Chloride	0.79	U	0.79	5.00	ug/Kg
74-83-9	Bromomethane	1.10	U	1.10	5.00	ug/Kg
75-00-3	Chloroethane	1.30	U	1.30	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	1.20	U	1.20	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	1.10	U	1.10	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	1.00	U	1.00	5.00	ug/Kg
67-64-1	Acetone	4.70	U	4.70	25.0	ug/Kg
75-15-0	Carbon Disulfide	1.10	U	1.10	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.73	U	0.73	5.00	ug/Kg
79-20-9	Methyl Acetate	1.50	U	1.50	5.00	ug/Kg
75-09-2	Methylene Chloride	3.50	U	3.50	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.86	U	0.86	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	0.80	U	0.80	5.00	ug/Kg
110-82-7	Cyclohexane	0.79	U	0.79	5.00	ug/Kg
78-93-3	2-Butanone	6.50	U	6.50	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	0.97	U	0.97	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.75	5.00	ug/Kg
74-97-5	Bromochloromethane	1.20	U	1.20	5.00	ug/Kg
67-66-3	Chloroform	0.84	U	0.84	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.93	U	0.93	5.00	ug/Kg
108-87-2	Methylcyclohexane	0.91	U	0.91	5.00	ug/Kg
71-43-2	Benzene	0.79	U	0.79	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	0.79	U	0.79	5.00	ug/Kg
79-01-6	Trichloroethene	0.81	U	0.81	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	0.91	U	0.91	5.00	ug/Kg
75-27-4	Bromodichloromethane	0.78	U	0.78	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.60	U	3.60	25.0	ug/Kg
108-88-3	Toluene	0.78	U	0.78	5.00	ug/Kg

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	
Client Sample ID:	VY0501SBL01		SDG No.:	Q1901
Lab Sample ID:	VY0501SBL01		Matrix:	SOIL
Analytical Method:	SW8260		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022096.D	1		05/01/25 10:11	VY050125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.65	U	0.65	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.62	U	0.62	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.92	U	0.92	5.00	ug/Kg
591-78-6	2-Hexanone	3.70	U	3.70	25.0	ug/Kg
124-48-1	Dibromochloromethane	0.87	U	0.87	5.00	ug/Kg
106-93-4	1,2-Dibromoethane	0.88	U	0.88	5.00	ug/Kg
127-18-4	Tetrachloroethene	1.10	U	1.10	5.00	ug/Kg
108-90-7	Chlorobenzene	0.91	U	0.91	5.00	ug/Kg
100-41-4	Ethyl Benzene	0.67	U	0.67	5.00	ug/Kg
179601-23-1	m/p-Xylenes	1.20	U	1.20	10.0	ug/Kg
95-47-6	o-Xylene	0.82	U	0.82	5.00	ug/Kg
100-42-5	Styrene	0.71	U	0.71	5.00	ug/Kg
75-25-2	Bromoform	0.86	U	0.86	5.00	ug/Kg
98-82-8	Isopropylbenzene	0.78	U	0.78	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.20	U	1.20	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.70	U	1.70	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.60	U	1.60	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.50	U	1.50	5.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.80	U	1.80	5.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	3.00	U	3.00	5.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	3.20	U	3.20	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.2		70 (63) - 130 (155)	98%	SPK: 50
1868-53-7	Dibromofluoromethane	50.6		70 (70) - 130 (134)	101%	SPK: 50
2037-26-5	Toluene-d8	47.8		70 (74) - 130 (123)	96%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.1		70 (38) - 130 (136)	106%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	318000	7.707			
540-36-3	1,4-Difluorobenzene	586000	8.615			
3114-55-4	Chlorobenzene-d5	509000	11.42			
3855-82-1	1,4-Dichlorobenzene-d4	180000	13.346			

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	
Client Sample ID:	VY0501SBL01		SDG No.:	Q1901
Lab Sample ID:	VY0501SBL01		Matrix:	SOIL
Analytical Method:	SW8260		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022096.D	1		05/01/25 10:11	VY050125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	
Client Sample ID:	VX0429WBS01	SDG No.:	Q1901
Lab Sample ID:	VX0429WBS01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045975.D	1		04/29/25 12:21	VX042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	17.4		0.22	1.00	ug/L
74-87-3	Chloromethane	16.3		0.32	1.00	ug/L
75-01-4	Vinyl Chloride	16.9		0.26	1.00	ug/L
74-83-9	Bromomethane	17.5		1.40	5.00	ug/L
75-00-3	Chloroethane	20.4		0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.4		0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	20.4		0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.5		0.23	1.00	ug/L
67-64-1	Acetone	97.8		1.50	5.00	ug/L
75-15-0	Carbon Disulfide	13.4		0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	21.1		0.16	1.00	ug/L
79-20-9	Methyl Acetate	28.4		0.27	1.00	ug/L
75-09-2	Methylene Chloride	19.2		0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	17.8		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	20.1		0.23	1.00	ug/L
110-82-7	Cyclohexane	18.0		1.50	5.00	ug/L
78-93-3	2-Butanone	100		0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	21.2		0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.9		0.19	1.00	ug/L
74-97-5	Bromochloromethane	22.0		0.22	1.00	ug/L
67-66-3	Chloroform	21.1		0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	20.6		0.20	1.00	ug/L
108-87-2	Methylcyclohexane	18.8		0.16	1.00	ug/L
71-43-2	Benzene	19.7		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	21.5		0.22	1.00	ug/L
79-01-6	Trichloroethene	19.9		0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	20.9		0.20	1.00	ug/L
75-27-4	Bromodichloromethane	21.4		0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.68	5.00	ug/L
108-88-3	Toluene	20.5		0.14	1.00	ug/L

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	
Client Sample ID:	VX0429WBS01		SDG No.:	Q1901
Lab Sample ID:	VX0429WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045975.D	1		04/29/25 12:21	VX042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	19.4		0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	21.5		0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	21.9		0.21	1.00	ug/L
591-78-6	2-Hexanone	110		0.89	5.00	ug/L
124-48-1	Dibromochloromethane	21.1		0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	21.7		0.15	1.00	ug/L
127-18-4	Tetrachloroethene	20.0		0.23	1.00	ug/L
108-90-7	Chlorobenzene	20.8		0.12	1.00	ug/L
100-41-4	Ethyl Benzene	20.9		0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	41.7		0.24	2.00	ug/L
95-47-6	o-Xylene	20.9		0.12	1.00	ug/L
100-42-5	Styrene	21.4		0.15	1.00	ug/L
75-25-2	Bromoform	21.2		0.19	1.00	ug/L
98-82-8	Isopropylbenzene	20.4		0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	20.9		0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	20.1		0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.8		0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	21.2		0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	21.5		0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	19.5		0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	20.1		0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.9		70 (74) - 130 (125)	108%	SPK: 50
1868-53-7	Dibromofluoromethane	55.4		70 (75) - 130 (124)	111%	SPK: 50
2037-26-5	Toluene-d8	53.0		70 (86) - 130 (113)	106%	SPK: 50
460-00-4	4-Bromofluorobenzene	56.3		70 (77) - 130 (121)	113%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	80600	5.55			
540-36-3	1,4-Difluorobenzene	140000	6.757			
3114-55-4	Chlorobenzene-d5	123000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	59200	12.018			

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	
Client Sample ID:	VX0429WBS01		SDG No.:	Q1901
Lab Sample ID:	VX0429WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045975.D	1		04/29/25 12:21	VX042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	
Client Sample ID:	VY0430SBS01	SDG No.:	Q1901
Lab Sample ID:	VY0430SBS01	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	100
Sample Wt/Vol:	5	Units:	g
Soil Aliquot Vol:			uL
GC Column:	RXI-624	Test:	VOC-TCLVOA-10
Prep Method :	ID : 0.25	Level :	LOW

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022072.D	1		04/30/25 10:30	VY043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	19.3		1.10	5.00	ug/Kg
74-87-3	Chloromethane	22.2		1.10	5.00	ug/Kg
75-01-4	Vinyl Chloride	20.7		0.79	5.00	ug/Kg
74-83-9	Bromomethane	22.8		1.10	5.00	ug/Kg
75-00-3	Chloroethane	21.2		1.30	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	21.8		1.20	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	20.4		1.10	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	19.7		1.00	5.00	ug/Kg
67-64-1	Acetone	94.5		4.70	25.0	ug/Kg
75-15-0	Carbon Disulfide	19.9		1.10	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	19.0		0.73	5.00	ug/Kg
79-20-9	Methyl Acetate	20.3		1.50	5.00	ug/Kg
75-09-2	Methylene Chloride	20.0		3.50	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	19.5		0.86	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	19.9		0.80	5.00	ug/Kg
110-82-7	Cyclohexane	18.9		0.79	5.00	ug/Kg
78-93-3	2-Butanone	92.7		6.50	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	21.3		0.97	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	19.5		0.75	5.00	ug/Kg
74-97-5	Bromochloromethane	20.0		1.20	5.00	ug/Kg
67-66-3	Chloroform	20.4		0.84	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	20.1		0.93	5.00	ug/Kg
108-87-2	Methylcyclohexane	19.2		0.91	5.00	ug/Kg
71-43-2	Benzene	20.6		0.79	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	21.2		0.79	5.00	ug/Kg
79-01-6	Trichloroethene	20.9		0.81	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	20.3		0.91	5.00	ug/Kg
75-27-4	Bromodichloromethane	20.9		0.78	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	100		3.60	25.0	ug/Kg
108-88-3	Toluene	21.0		0.78	5.00	ug/Kg

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	
Client Sample ID:	VY0430SBS01	SDG No.:	Q1901
Lab Sample ID:	VY0430SBS01	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	100
Sample Wt/Vol:	5	Units:	g
Soil Aliquot Vol:			uL
GC Column:	RXI-624	Test:	VOC-TCLVOA-10
Prep Method :	ID : 0.25	Level :	LOW

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022072.D	1		04/30/25 10:30	VY043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	20.6		0.65	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	20.6		0.62	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	20.9		0.92	5.00	ug/Kg
591-78-6	2-Hexanone	99.9		3.70	25.0	ug/Kg
124-48-1	Dibromochloromethane	21.1		0.87	5.00	ug/Kg
106-93-4	1,2-Dibromoethane	20.6		0.88	5.00	ug/Kg
127-18-4	Tetrachloroethene	21.1		1.10	5.00	ug/Kg
108-90-7	Chlorobenzene	20.1		0.91	5.00	ug/Kg
100-41-4	Ethyl Benzene	19.3		0.67	5.00	ug/Kg
179601-23-1	m/p-Xylenes	40.6		1.20	10.0	ug/Kg
95-47-6	o-Xylene	19.7		0.82	5.00	ug/Kg
100-42-5	Styrene	20.0		0.71	5.00	ug/Kg
75-25-2	Bromoform	20.9		0.86	5.00	ug/Kg
98-82-8	Isopropylbenzene	18.8		0.78	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	19.7		1.20	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	19.9		1.70	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	19.7		1.60	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	20.1		1.50	5.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	18.2		1.80	5.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	19.2		3.00	5.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	19.7		3.20	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.5		70 (63) - 130 (155)	99%	SPK: 50
1868-53-7	Dibromofluoromethane	52.6		70 (70) - 130 (134)	105%	SPK: 50
2037-26-5	Toluene-d8	52.7		70 (74) - 130 (123)	105%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.8		70 (38) - 130 (136)	104%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	286000	7.707			
540-36-3	1,4-Difluorobenzene	434000	8.616			
3114-55-4	Chlorobenzene-d5	405000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	223000	13.347			

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	
Client Sample ID:	VY0430SBS01		SDG No.:	Q1901
Lab Sample ID:	VY0430SBS01		Matrix:	SOIL
Analytical Method:	SW8260		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022072.D	1		04/30/25 10:30	VY043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	
Client Sample ID:	VY0501SBS01	SDG No.:	Q1901
Lab Sample ID:	VY0501SBS01	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022097.D	1		05/01/25 10:42	VY050125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	19.2		1.10	5.00	ug/Kg
74-87-3	Chloromethane	18.7		1.10	5.00	ug/Kg
75-01-4	Vinyl Chloride	18.7		0.79	5.00	ug/Kg
74-83-9	Bromomethane	18.6		1.10	5.00	ug/Kg
75-00-3	Chloroethane	18.9		1.30	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	19.9		1.20	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	19.8		1.10	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	19.9		1.00	5.00	ug/Kg
67-64-1	Acetone	91.2		4.70	25.0	ug/Kg
75-15-0	Carbon Disulfide	19.5		1.10	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	18.0		0.73	5.00	ug/Kg
79-20-9	Methyl Acetate	18.7		1.50	5.00	ug/Kg
75-09-2	Methylene Chloride	19.3		3.50	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	19.7		0.86	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	19.8		0.80	5.00	ug/Kg
110-82-7	Cyclohexane	18.8		0.79	5.00	ug/Kg
78-93-3	2-Butanone	86.7		6.50	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	20.6		0.97	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	19.8		0.75	5.00	ug/Kg
74-97-5	Bromochloromethane	18.9		1.20	5.00	ug/Kg
67-66-3	Chloroform	20.2		0.84	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	20.1		0.93	5.00	ug/Kg
108-87-2	Methylcyclohexane	19.0		0.91	5.00	ug/Kg
71-43-2	Benzene	20.2		0.79	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	19.9		0.79	5.00	ug/Kg
79-01-6	Trichloroethene	20.3		0.81	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	20.1		0.91	5.00	ug/Kg
75-27-4	Bromodichloromethane	20.6		0.78	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	88.1		3.60	25.0	ug/Kg
108-88-3	Toluene	20.6		0.78	5.00	ug/Kg

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	18.9		0.65	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	19.9		0.62	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	20.2		0.92	5.00	ug/Kg
591-78-6	2-Hexanone	84.7		3.70	25.0	ug/Kg
124-48-1	Dibromochloromethane	19.8		0.87	5.00	ug/Kg
106-93-4	1,2-Dibromoethane	19.9		0.88	5.00	ug/Kg
127-18-4	Tetrachloroethene	20.9		1.10	5.00	ug/Kg
108-90-7	Chlorobenzene	19.8		0.91	5.00	ug/Kg
100-41-4	Ethyl Benzene	19.5		0.67	5.00	ug/Kg
179601-23-1	m/p-Xylenes	40.5		1.20	10.0	ug/Kg
95-47-6	o-Xylene	19.6		0.82	5.00	ug/Kg
100-42-5	Styrene	20.1		0.71	5.00	ug/Kg
75-25-2	Bromoform	19.6		0.86	5.00	ug/Kg
98-82-8	Isopropylbenzene	19.2		0.78	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	18.4		1.20	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	19.9		1.70	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	19.6		1.60	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	19.4		1.50	5.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	17.7		1.80	5.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	19.0		3.00	5.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	18.8		3.20	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	47.5		70 (63) - 130 (155)	95%	SPK: 50
1868-53-7	Dibromofluoromethane	49.5		70 (70) - 130 (134)	99%	SPK: 50
2037-26-5	Toluene-d8	50.1		70 (74) - 130 (123)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.3		70 (38) - 130 (136)	99%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	294000	7.707			
540-36-3	1,4-Difluorobenzene	453000	8.616			
3114-55-4	Chlorobenzene-d5	414000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	225000	13.346			

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	
Client Sample ID:	VY0501SBS01		SDG No.:	Q1901
Lab Sample ID:	VY0501SBS01		Matrix:	SOIL
Analytical Method:	SW8260		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022097.D	1		05/01/25 10:42	VY050125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	
Client Sample ID:	VX0429WBSD01	SDG No.:	Q1901
Lab Sample ID:	VX0429WBSD01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045976.D	1		04/29/25 12:49	VX042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	17.3		0.22	1.00	ug/L
74-87-3	Chloromethane	16.4		0.32	1.00	ug/L
75-01-4	Vinyl Chloride	16.9		0.26	1.00	ug/L
74-83-9	Bromomethane	17.3		1.40	5.00	ug/L
75-00-3	Chloroethane	18.9		0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.5		0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	21.3		0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.2		0.23	1.00	ug/L
67-64-1	Acetone	100		1.50	5.00	ug/L
75-15-0	Carbon Disulfide	13.4		0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	21.9		0.16	1.00	ug/L
79-20-9	Methyl Acetate	30.0		0.27	1.00	ug/L
75-09-2	Methylene Chloride	20.4		0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.6		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	20.0		0.23	1.00	ug/L
110-82-7	Cyclohexane	18.4		1.50	5.00	ug/L
78-93-3	2-Butanone	110		0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	21.2		0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	20.5		0.19	1.00	ug/L
74-97-5	Bromochloromethane	22.9		0.22	1.00	ug/L
67-66-3	Chloroform	21.5		0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	20.9		0.20	1.00	ug/L
108-87-2	Methylcyclohexane	19.9		0.16	1.00	ug/L
71-43-2	Benzene	19.8		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	22.2		0.22	1.00	ug/L
79-01-6	Trichloroethene	19.6		0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	21.7		0.20	1.00	ug/L
75-27-4	Bromodichloromethane	21.4		0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.68	5.00	ug/L
108-88-3	Toluene	20.1		0.14	1.00	ug/L

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025			Date Received:	
Client Sample ID:	VX0429WBSD01			SDG No.:	Q1901
Lab Sample ID:	VX0429WBSD01			Matrix:	Water
Analytical Method:	SW8260			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID :	0.18	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045976.D	1		04/29/25 12:49	VX042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	20.3		0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	21.3		0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	21.6		0.21	1.00	ug/L
591-78-6	2-Hexanone	110		0.89	5.00	ug/L
124-48-1	Dibromochloromethane	22.1		0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	22.1		0.15	1.00	ug/L
127-18-4	Tetrachloroethene	19.7		0.23	1.00	ug/L
108-90-7	Chlorobenzene	21.5		0.12	1.00	ug/L
100-41-4	Ethyl Benzene	21.2		0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	42.6		0.24	2.00	ug/L
95-47-6	o-Xylene	21.5		0.12	1.00	ug/L
100-42-5	Styrene	21.7		0.15	1.00	ug/L
75-25-2	Bromoform	21.2		0.19	1.00	ug/L
98-82-8	Isopropylbenzene	19.9		0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	21.0		0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	20.2		0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	20.0		0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	21.2		0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	22.3		0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	20.9		0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	20.3		0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.3		70 (74) - 130 (125)	109%	SPK: 50
1868-53-7	Dibromofluoromethane	54.8		70 (75) - 130 (124)	110%	SPK: 50
2037-26-5	Toluene-d8	51.7		70 (86) - 130 (113)	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	55.4		70 (77) - 130 (121)	111%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	77100	5.544			
540-36-3	1,4-Difluorobenzene	135000	6.757			
3114-55-4	Chlorobenzene-d5	118000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	57500	12.018			

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	
Client Sample ID:	VX0429WBSD01		SDG No.:	Q1901
Lab Sample ID:	VX0429WBSD01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045976.D	1		04/29/25 12:49	VX042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	
Client Sample ID:	VY0430SBSD01	SDG No.:	Q1901
Lab Sample ID:	VY0430SBSD01	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022073.D	1		04/30/25 10:52	VY043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	20.3		1.10	5.00	ug/Kg
74-87-3	Chloromethane	19.5		1.10	5.00	ug/Kg
75-01-4	Vinyl Chloride	19.7		0.79	5.00	ug/Kg
74-83-9	Bromomethane	19.5		1.10	5.00	ug/Kg
75-00-3	Chloroethane	19.5		1.30	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	20.4		1.20	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	20.4		1.10	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	20.2		1.00	5.00	ug/Kg
67-64-1	Acetone	89.4		4.70	25.0	ug/Kg
75-15-0	Carbon Disulfide	19.7		1.10	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	18.6		0.73	5.00	ug/Kg
79-20-9	Methyl Acetate	21.5		1.50	5.00	ug/Kg
75-09-2	Methylene Chloride	20.1		3.50	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	20.0		0.86	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	19.8		0.80	5.00	ug/Kg
110-82-7	Cyclohexane	18.9		0.79	5.00	ug/Kg
78-93-3	2-Butanone	91.4		6.50	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	20.7		0.97	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	19.7		0.75	5.00	ug/Kg
74-97-5	Bromochloromethane	20.0		1.20	5.00	ug/Kg
67-66-3	Chloroform	20.1		0.84	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	20.2		0.93	5.00	ug/Kg
108-87-2	Methylcyclohexane	19.0		0.91	5.00	ug/Kg
71-43-2	Benzene	20.5		0.79	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	20.5		0.79	5.00	ug/Kg
79-01-6	Trichloroethene	20.5		0.81	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	20.4		0.91	5.00	ug/Kg
75-27-4	Bromodichloromethane	20.6		0.78	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	98.5		3.60	25.0	ug/Kg
108-88-3	Toluene	20.5		0.78	5.00	ug/Kg

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	
Client Sample ID:	VY0430SBSD01	SDG No.:	Q1901	
Lab Sample ID:	VY0430SBSD01	Matrix:	SOIL	
Analytical Method:	SW8260	% Solid:	100	
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022073.D	1		04/30/25 10:52	VY043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	20.2		0.65	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	20.1		0.62	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	20.6		0.92	5.00	ug/Kg
591-78-6	2-Hexanone	94.3		3.70	25.0	ug/Kg
124-48-1	Dibromochloromethane	20.2		0.87	5.00	ug/Kg
106-93-4	1,2-Dibromoethane	20.1		0.88	5.00	ug/Kg
127-18-4	Tetrachloroethene	21.3		1.10	5.00	ug/Kg
108-90-7	Chlorobenzene	20.3		0.91	5.00	ug/Kg
100-41-4	Ethyl Benzene	19.4		0.67	5.00	ug/Kg
179601-23-1	m/p-Xylenes	40.3		1.20	10.0	ug/Kg
95-47-6	o-Xylene	19.4		0.82	5.00	ug/Kg
100-42-5	Styrene	20.3		0.71	5.00	ug/Kg
75-25-2	Bromoform	20.3		0.86	5.00	ug/Kg
98-82-8	Isopropylbenzene	19.3		0.78	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	19.4		1.20	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	20.0		1.70	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	20.0		1.60	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	20.4		1.50	5.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	18.4		1.80	5.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	18.8		3.00	5.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	18.9		3.20	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.0		70 (63) - 130 (155)	98%	SPK: 50
1868-53-7	Dibromofluoromethane	50.8		70 (70) - 130 (134)	102%	SPK: 50
2037-26-5	Toluene-d8	51.3		70 (74) - 130 (123)	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.7		70 (38) - 130 (136)	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	289000	7.707			
540-36-3	1,4-Difluorobenzene	440000	8.61			
3114-55-4	Chlorobenzene-d5	407000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	218000	13.347			

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	
Client Sample ID:	VY0430SBSD01		SDG No.:	Q1901
Lab Sample ID:	VY0430SBSD01		Matrix:	SOIL
Analytical Method:	SW8260		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022073.D	1		04/30/25 10:52	VY043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG No.: Q1901

Instrument ID: MSVOA_X Calibration Date(s): 04/01/2025 04/01/2025

Heated Purge: (Y/N) N Calibration Time(s): 17:06 19:02

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF001 = VX045525.D		RRF005 = VX045526.D		RRF020 = VX045527.D		RRF050 = VX045528.D		RRF100 = VX045529.D		RRF150 = VX045530.D	
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD					
Dichlorodifluoromethane	0.615	0.695	0.700	0.820	0.837	0.820	0.748	12.1					
Chloromethane	0.764	0.734	0.777	0.784	0.815	0.734	0.768	4.1					
Vinyl Chloride	0.671	0.662	0.701	0.716	0.738	0.730	0.703	4.4					
Bromomethane		0.342	0.327	0.330	0.341	0.327	0.333	2.2					
Chloroethane	0.378	0.397	0.390	0.398	0.355	0.319	0.373	8.3					
Trichlorofluoromethane	1.051	0.999	1.075	1.086	1.089	0.989	1.048	4.2					
1,1,2-Trichlorotrifluoroethane	0.575	0.599	0.635	0.621	0.621	0.634	0.614	3.7					
1,1-Dichloroethene	0.563	0.588	0.612	0.604	0.618	0.616	0.600	3.5					
Acetone	0.400	0.369	0.387	0.378	0.365	0.355	0.375	4.3					
Carbon Disulfide	1.334	1.327	1.434	1.553	1.604	1.642	1.483	9.2					
Methyl tert-butyl Ether	1.915	1.964	2.169	2.118	2.217	2.216	2.100	6.2					
Methyl Acetate	0.901	0.863	0.864	0.857	0.855	0.846	0.864	2.2					
Methylene Chloride	0.692	0.695	0.726	0.709	0.705	0.690	0.703	2					
trans-1,2-Dichloroethene	0.574	0.594	0.636	0.624	0.621	0.630	0.613	3.9					
1,1-Dichloroethane	1.211	1.240	1.318	1.269	1.283	1.292	1.269	3					
Cyclohexane		1.044	1.148	1.123	1.145	1.149	1.122	4					
2-Butanone	0.474	0.537	0.582	0.586	0.571	0.548	0.550	7.5					
Carbon Tetrachloride	0.450	0.497	0.521	0.536	0.545	0.556	0.518	7.5					
cis-1,2-Dichloroethene	0.762	0.696	0.760	0.751	0.754	0.760	0.747	3.4					
Bromochloromethane	0.671	0.608	0.617	0.596	0.610	0.576	0.613	5.2					
Chloroform	1.244	1.309	1.348	1.315	1.309	1.309	1.306	2.6					
1,1,1-Trichloroethane	1.028	1.052	1.120	1.109	1.153	1.167	1.105	5					
Methylcyclohexane	0.519	0.527	0.596	0.622	0.628	0.632	0.587	8.8					
Benzene	1.414	1.416	1.519	1.481	1.483	1.465	1.463	2.8					
1,2-Dichloroethane	0.533	0.585	0.649	0.622	0.620	0.617	0.604	6.7					
Trichloroethene	0.349	0.322	0.356	0.351	0.351	0.356	0.348	3.7					
1,2-Dichloropropane	0.309	0.365	0.388	0.376	0.379	0.374	0.365	7.8					
Bromodichloromethane	0.521	0.519	0.576	0.572	0.587	0.583	0.560	5.6					
4-Methyl-2-Pentanone	0.499	0.578	0.661	0.663	0.647	0.589	0.606	10.6					
Toluene	0.817	0.866	0.939	0.910	0.905	0.875	0.885	4.8					

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>PORT06</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1901</u>
Instrument ID:	<u>MSVOA_X</u>	SAS No.:	<u>Q1901</u>
Heated Purge: (Y/N)	<u>N</u>	SDG No.:	<u>Q1901</u>
GC Column:	<u>DB-624UI</u>	Calibration Date(s):	<u>04/01/2025</u>
ID:	<u>0.18 (mm)</u>	Calibration Time(s):	<u>17:06</u>
			<u>19:02</u>

LAB FILE ID:		RRF001 = VX045525.D		RRF005 = VX045526.D		RRF020 = VX045527.D		
		RRF050 = VX045528.D		RRF100 = VX045529.D		RRF150 = VX045530.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
t-1,3-Dichloropropene	0.316	0.400	0.465	0.508	0.555	0.558	0.467	20.3
cis-1,3-Dichloropropene	0.371	0.474	0.546	0.568	0.597	0.600	0.526	16.9
1,1,2-Trichloroethane	0.337	0.346	0.376	0.359	0.358	0.340	0.353	4.1
2-Hexanone	0.363	0.429	0.488	0.492	0.484	0.439	0.449	11.1
Dibromochloromethane	0.313	0.352	0.404	0.412	0.421	0.405	0.385	11.1
1,2-Dibromoethane	0.294	0.351	0.380	0.373	0.380	0.364	0.357	9.1
Tetrachloroethene	0.346	0.373	0.371	0.347	0.333	0.347	0.353	4.5
Chlorobenzene	0.951	1.054	1.123	1.084	1.086	1.112	1.068	5.8
Ethyl Benzene	1.608	1.819	2.002	2.007	1.993	2.053	1.914	8.9
m/p-Xylenes	0.594	0.669	0.732	0.728	0.723	0.729	0.696	7.9
o-Xylene	0.562	0.676	0.725	0.719	0.716	0.714	0.686	9.2
Styrene	0.897	1.043	1.202	1.216	1.230	1.202	1.132	11.8
Bromoform	0.220	0.252	0.272	0.296	0.307	0.315	0.277	13.1
Isopropylbenzene	3.581	3.850	4.224	4.181	4.043	4.151	4.005	6.2
1,1,2,2-Tetrachloroethane	1.457	1.428	1.461	1.384	1.354	1.358	1.407	3.4
1,3-Dichlorobenzene	1.658	1.605	1.726	1.699	1.714	1.706	1.684	2.7
1,4-Dichlorobenzene	1.671	1.724	1.768	1.703	1.674	1.706	1.708	2.1
1,2-Dichlorobenzene	1.644	1.645	1.750	1.678	1.665	1.667	1.675	2.3
1,2-Dibromo-3-Chloropropane	0.196	0.260	0.312	0.316	0.333	0.349	0.294	19.4
1,2,4-Trichlorobenzene	0.727	0.844	0.948	0.947	1.045	1.083	0.932	14.1
1,2,3-Trichlorobenzene	0.782	0.916	0.999	1.000	1.063	1.097	0.976	11.6
1,2-Dichloroethane-d4		0.962	0.900	0.868	0.904	0.937	0.914	3.9
Dibromofluoromethane		0.372	0.342	0.345	0.353	0.362	0.355	3.5
Toluene-d8		1.257	1.233	1.214	1.230	1.257	1.238	1.5
4-Bromofluorobenzene		0.413	0.448	0.453	0.481	0.460	0.451	5.5

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG No.: Q1901

Instrument ID: MSVOA_Y Calibration Date(s): 04/22/2025 04/22/2025

Heated Purge: (Y/N) Y Calibration Time(s): 13:39 16:15

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VY021953.D		RRF010 = VY021954.D		RRF020 = VY021955.D		RRF050 = VY021956.D		RRF100 = VY021957.D		RRF150 = VY021958.D	
COMPOUND		RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD				
Dichlorodifluoromethane		0.469	0.477	0.405	0.393	0.408	0.405	0.426	8.7				
Chloromethane		0.704	0.670	0.636	0.602	0.529	0.498	0.607	13.2				
Vinyl Chloride		0.829	0.830	0.758	0.730	0.680	0.647	0.745	10.1				
Bromomethane		0.782	0.697	0.674	0.596	0.554	0.550	0.642	14.3				
Chloroethane		0.589	0.550	0.524	0.494	0.460	0.429	0.508	11.6				
Trichlorofluoromethane		1.048	1.075	0.987	0.960	0.921	0.907	0.983	6.9				
1,1,2-Trichlorotrifluoroethane		0.604	0.591	0.530	0.489	0.497	0.488	0.533	9.8				
1,1-Dichloroethene		0.525	0.532	0.483	0.471	0.487	0.484	0.497	5				
Acetone		0.092	0.084	0.066	0.063	0.058	0.057	0.070	20.7				
Carbon Disulfide		1.736	1.737	1.569	1.493	1.516	1.459	1.585	7.7				
Methyl tert-butyl Ether		1.098	1.067	1.102	1.133	1.141	1.134	1.113	2.6				
Methyl Acetate		0.214	0.200	0.231	0.223	0.208	0.219	0.216	5.2				
Methylene Chloride		0.759	0.625	0.574	0.533	0.518	0.494	0.584	16.7				
trans-1,2-Dichloroethene		0.619	0.578	0.533	0.532	0.546	0.532	0.557	6.3				
1,1-Dichloroethane		0.987	0.943	0.893	0.853	0.855	0.826	0.893	6.9				
Cyclohexane		0.836	0.857	0.738	0.708	0.730	0.711	0.763	8.6				
2-Butanone		0.111	0.107	0.108	0.106	0.099	0.100	0.105	4.5				
Carbon Tetrachloride		0.540	0.563	0.512	0.507	0.535	0.523	0.530	3.9				
cis-1,2-Dichloroethene		0.626	0.638	0.604	0.612	0.633	0.621	0.622	2.1				
Bromochloromethane		0.376	0.362	0.365	0.343	0.321	0.314	0.347	7.2				
Chloroform		1.084	1.046	0.974	0.956	0.958	0.925	0.990	6.2				
1,1,1-Trichloroethane		0.972	0.944	0.896	0.853	0.865	0.846	0.896	5.8				
Methylcyclohexane		0.525	0.569	0.519	0.549	0.600	0.589	0.558	6				
Benzene		1.423	1.439	1.351	1.337	1.415	1.358	1.387	3.1				
1,2-Dichloroethane		0.347	0.367	0.350	0.335	0.336	0.325	0.343	4.3				
Trichloroethene		0.402	0.405	0.369	0.367	0.387	0.375	0.384	4.3				
1,2-Dichloropropane		0.320	0.317	0.302	0.300	0.308	0.294	0.307	3.4				
Bromodichloromethane		0.479	0.504	0.464	0.470	0.490	0.471	0.480	3.1				
4-Methyl-2-Pentanone		0.144	0.149	0.167	0.172	0.166	0.165	0.160	6.9				
Toluene		0.838	0.915	0.875	0.894	0.952	0.913	0.898	4.4				

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>PORT06</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1901</u>
Instrument ID:	<u>MSVOA_Y</u>	SAS No.:	<u>Q1901</u>
Heated Purge: (Y/N)	<u>Y</u>	SDG No.:	<u>Q1901</u>
GC Column:	<u>RXI-624</u>	Calibration Date(s):	<u>04/22/2025</u>
ID:	<u>0.25 (mm)</u>	Calibration Time(s):	<u>13:39</u>
			<u>16:15</u>

LAB FILE ID:	RRF005 = VY021953.D	RRF010 = VY021954.D	RRF020 = VY021955.D	RRF050 = VY021956.D	RRF100 = VY021957.D	RRF150 = VY021958.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
t-1,3-Dichloropropene	0.383	0.408	0.399	0.417	0.437	0.425	0.411	4.7
cis-1,3-Dichloropropene	0.466	0.488	0.483	0.484	0.514	0.500	0.489	3.3
1,1,2-Trichloroethane	0.250	0.261	0.261	0.256	0.258	0.252	0.256	1.8
2-Hexanone	0.093	0.096	0.109	0.115	0.111	0.111	0.106	8.5
Dibromochloromethane	0.345	0.355	0.354	0.352	0.367	0.357	0.355	2
1,2-Dibromoethane	0.238	0.243	0.243	0.241	0.247	0.240	0.242	1.3
Tetrachloroethene	0.500	0.516	0.462	0.455	0.466	0.430	0.472	6.6
Chlorobenzene	1.188	1.152	1.055	1.083	1.139	1.092	1.118	4.5
Ethyl Benzene	1.693	1.840	1.718	1.835	1.990	1.898	1.829	6.1
m/p-Xylenes	0.664	0.720	0.699	0.734	0.797	0.754	0.728	6.3
o-Xylene	0.599	0.639	0.635	0.689	0.747	0.716	0.671	8.3
Styrene	0.950	1.080	1.078	1.169	1.272	1.209	1.126	10.2
Bromoform	0.234	0.221	0.224	0.229	0.236	0.229	0.229	2.5
Isopropylbenzene	3.061	3.316	3.102	3.325	3.639	3.599	3.341	7.2
1,1,2,2-Tetrachloroethane	0.601	0.562	0.555	0.548	0.555	0.558	0.563	3.4
1,3-Dichlorobenzene	1.730	1.784	1.595	1.647	1.789	1.742	1.714	4.5
1,4-Dichlorobenzene	1.821	1.726	1.605	1.620	1.733	1.682	1.698	4.7
1,2-Dichlorobenzene	1.523	1.522	1.434	1.452	1.537	1.505	1.495	2.8
1,2-Dibromo-3-Chloropropane	0.097	0.091	0.094	0.084	0.085	0.087	0.090	5.9
1,2,4-Trichlorobenzene	0.747	0.800	0.785	0.823	0.970	0.959	0.847	11.1
1,2,3-Trichlorobenzene	0.671	0.673	0.684	0.707	0.829	0.825	0.731	10.2
1,2-Dichloroethane-d4	0.525	0.477	0.459	0.466	0.418	0.427	0.462	8.3
Dibromofluoromethane	0.335	0.341	0.330	0.330	0.319	0.327	0.330	2.3
Toluene-d8	1.220	1.260	1.211	1.276	1.244	1.261	1.245	2
4-Bromofluorobenzene	0.408	0.429	0.401	0.426	0.423	0.428	0.419	2.8

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG No.: Q1901

Instrument ID: MSVOA_X Calibration Date/Time: 04/29/2025 11:06

Lab File ID: VX045972.D Init. Calib. Date(s): 04/01/2025 04/01/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 17:06 19:02

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.748	0.712		-4.81	20
Chloromethane	0.768	0.720	0.1	-6.25	20
Vinyl Chloride	0.703	0.677		-3.7	20
Bromomethane	0.333	0.325		-2.4	20
Chloroethane	0.373	0.394		5.63	20
Trichlorofluoromethane	1.048	1.127		7.54	20
1,1,2-Trichlorotrifluoroethane	0.614	0.689		12.22	20
1,1-Dichloroethene	0.600	0.621		3.5	20
Acetone	0.375	0.380		1.33	20
Carbon Disulfide	1.483	1.385		-6.61	20
Methyl tert-butyl Ether	2.100	2.373		13	20
Methyl Acetate	0.864	1.136		31.48	20
Methylene Chloride	0.703	0.722		2.7	20
trans-1,2-Dichloroethene	0.613	0.644		5.06	20
1,1-Dichloroethane	1.269	1.349	0.1	6.3	20
Cyclohexane	1.122	1.160		3.39	20
2-Butanone	0.550	0.579		5.27	20
Carbon Tetrachloride	0.518	0.608		17.38	20
cis-1,2-Dichloroethene	0.747	0.782		4.68	20
Bromochloromethane	0.613	0.597		-2.61	20
Chloroform	1.306	1.431		9.57	20
1,1,1-Trichloroethane	1.105	1.243		12.49	20
Methylcyclohexane	0.587	0.680		15.84	20
Benzene	1.463	1.586		8.41	20
1,2-Dichloroethane	0.604	0.686		13.58	20
Trichloroethene	0.348	0.388		11.49	20
1,2-Dichloropropane	0.365	0.412		12.88	20
Bromodichloromethane	0.560	0.647		15.54	20
4-Methyl-2-Pentanone	0.606	0.699		15.35	20
Toluene	0.885	0.989		11.75	20
t-1,3-Dichloropropene	0.467	0.593		26.98	20
cis-1,3-Dichloropropene	0.526	0.649		23.38	20
1,1,2-Trichloroethane	0.353	0.398		12.75	20
2-Hexanone	0.449	0.521		16.04	20
Dibromochloromethane	0.385	0.460		19.48	20
1,2-Dibromoethane	0.357	0.410		14.85	20
Tetrachloroethene	0.353	0.405		14.73	20
Chlorobenzene	1.068	1.251	0.3	17.14	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG No.: Q1901

Instrument ID: MSVOA_X Calibration Date/Time: 04/29/2025 11:06

Lab File ID: VX045972.D Init. Calib. Date(s): 04/01/2025 04/01/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 17:06 19:02

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.914	2.266		18.39	20
m/p-Xylenes	0.696	0.830		19.25	20
o-Xylene	0.686	0.806		17.49	20
Styrene	1.132	1.379		21.82	20
Bromoform	0.277	0.351	0.1	26.72	20
Isopropylbenzene	4.005	4.316		7.76	20
1,1,2,2-Tetrachloroethane	1.407	1.407	0.3	0	20
1,3-Dichlorobenzene	1.684	1.834		8.91	20
1,4-Dichlorobenzene	1.708	1.848		8.2	20
1,2-Dichlorobenzene	1.675	1.817		8.48	20
1,2-Dibromo-3-Chloropropane	0.294	0.337		14.63	20
1,2,4-Trichlorobenzene	0.932	1.076		15.45	20
1,2,3-Trichlorobenzene	0.976	1.071		9.73	20
1,2-Dichloroethane-d4	0.914	0.918		0.44	20
Dibromofluoromethane	0.355	0.375		5.63	20
Toluene-d8	1.238	1.242		0.32	20
4-Bromofluorobenzene	0.451	0.499		10.64	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG No.: Q1901

Instrument ID: MSVOA_Y Calibration Date/Time: 04/30/2025 09:20

Lab File ID: VY022070.D Init. Calib. Date(s): 04/22/2025 04/22/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:39 16:15

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.426	0.361		-15.26	20
Chloromethane	0.607	0.597	0.1	-1.65	20
Vinyl Chloride	0.745	0.723		-2.95	20
Bromomethane	0.642	0.627		-2.34	20
Chloroethane	0.508	0.502		-1.18	20
Trichlorofluoromethane	0.983	0.946		-3.76	20
1,1,2-Trichlorotrifluoroethane	0.533	0.496		-6.94	20
1,1-Dichloroethene	0.497	0.459		-7.65	20
Acetone	0.070	0.061		-12.86	20
Carbon Disulfide	1.585	1.443		-8.96	20
Methyl tert-butyl Ether	1.113	1.095		-1.62	20
Methyl Acetate	0.216	0.257		18.98	20
Methylene Chloride	0.584	0.519		-11.13	20
trans-1,2-Dichloroethene	0.557	0.526		-5.57	20
1,1-Dichloroethane	0.893	0.845	0.1	-5.38	20
Cyclohexane	0.763	0.696		-8.78	20
2-Butanone	0.105	0.102		-2.86	20
Carbon Tetrachloride	0.530	0.543		2.45	20
cis-1,2-Dichloroethene	0.622	0.600		-3.54	20
Bromochloromethane	0.347	0.344		-0.87	20
Chloroform	0.990	0.980		-1.01	20
1,1,1-Trichloroethane	0.896	0.866		-3.35	20
Methylcyclohexane	0.558	0.556		-0.36	20
Benzene	1.387	1.433		3.32	20
1,2-Dichloroethane	0.343	0.359		4.66	20
Trichloroethene	0.384	0.385		0.26	20
1,2-Dichloropropane	0.307	0.321		4.56	20
Bromodichloromethane	0.480	0.508		5.83	20
4-Methyl-2-Pentanone	0.160	0.181		13.13	20
Toluene	0.898	0.944		5.12	20
t-1,3-Dichloropropene	0.411	0.436		6.08	20
cis-1,3-Dichloropropene	0.489	0.515		5.32	20
1,1,2-Trichloroethane	0.256	0.278		8.59	20
2-Hexanone	0.106	0.121		14.15	20
Dibromochloromethane	0.355	0.394		10.99	20
1,2-Dibromoethane	0.242	0.266		9.92	20
Tetrachloroethene	0.472	0.424		-10.17	20
Chlorobenzene	1.118	1.125	0.3	0.63	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG No.: Q1901

Instrument ID: MSVOA_Y Calibration Date/Time: 04/30/2025 09:20

Lab File ID: VY022070.D Init. Calib. Date(s): 04/22/2025 04/22/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:39 16:15

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.829	1.863		1.86	20
m/p-Xylenes	0.728	0.766		5.22	20
o-Xylene	0.671	0.703		4.77	20
Styrene	1.126	1.213		7.73	20
Bromoform	0.229	0.245	0.1	6.99	20
Isopropylbenzene	3.341	3.262		-2.37	20
1,1,2,2-Tetrachloroethane	0.563	0.583	0.3	3.55	20
1,3-Dichlorobenzene	1.714	1.713		-0.06	20
1,4-Dichlorobenzene	1.698	1.666		-1.88	20
1,2-Dichlorobenzene	1.495	1.499		0.27	20
1,2-Dibromo-3-Chloropropane	0.090	0.085		-5.56	20
1,2,4-Trichlorobenzene	0.847	0.852		0.59	20
1,2,3-Trichlorobenzene	0.731	0.750		2.6	20
1,2-Dichloroethane-d4	0.462	0.441		-4.55	20
Dibromofluoromethane	0.330	0.337		2.12	20
Toluene-d8	1.245	1.278		2.65	20
4-Bromofluorobenzene	0.419	0.436		4.06	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG No.: Q1901

Instrument ID: MSVOA_Y Calibration Date/Time: 05/01/2025 08:50

Lab File ID: VY022095.D Init. Calib. Date(s): 04/22/2025 04/22/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:39 16:15

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.426	0.359		-15.73	20
Chloromethane	0.607	0.542	0.1	-10.71	20
Vinyl Chloride	0.745	0.677		-9.13	20
Bromomethane	0.642	0.511		-20.41	20
Chloroethane	0.508	0.454		-10.63	20
Trichlorofluoromethane	0.983	0.891		-9.36	20
1,1,2-Trichlorotrifluoroethane	0.533	0.484		-9.19	20
1,1-Dichloroethene	0.497	0.441		-11.27	20
Acetone	0.070	0.062		-11.43	20
Carbon Disulfide	1.585	1.372		-13.44	20
Methyl tert-butyl Ether	1.113	1.031		-7.37	20
Methyl Acetate	0.216	0.226		4.63	20
Methylene Chloride	0.584	0.488		-16.44	20
trans-1,2-Dichloroethene	0.557	0.508		-8.8	20
1,1-Dichloroethane	0.893	0.788	0.1	-11.76	20
Cyclohexane	0.763	0.665		-12.84	20
2-Butanone	0.105	0.100		-4.76	20
Carbon Tetrachloride	0.530	0.528		-0.38	20
cis-1,2-Dichloroethene	0.622	0.570		-8.36	20
Bromochloromethane	0.347	0.340		-2.02	20
Chloroform	0.990	0.919		-7.17	20
1,1,1-Trichloroethane	0.896	0.816		-8.93	20
Methylcyclohexane	0.558	0.550		-1.43	20
Benzene	1.387	1.370		-1.23	20
1,2-Dichloroethane	0.343	0.348		1.46	20
Trichloroethene	0.384	0.385		0.26	20
1,2-Dichloropropane	0.307	0.304		-0.98	20
Bromodichloromethane	0.480	0.484		0.83	20
4-Methyl-2-Pentanone	0.160	0.176		10	20
Toluene	0.898	0.925		3.01	20
t-1,3-Dichloropropene	0.411	0.416		1.22	20
cis-1,3-Dichloropropene	0.489	0.484		-1.02	20
1,1,2-Trichloroethane	0.256	0.267		4.3	20
2-Hexanone	0.106	0.118		11.32	20
Dibromochloromethane	0.355	0.371		4.51	20
1,2-Dibromoethane	0.242	0.251		3.72	20
Tetrachloroethene	0.472	0.481		1.91	20
Chlorobenzene	1.118	1.078	0.3	-3.58	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG No.: Q1901

Instrument ID: MSVOA_Y Calibration Date/Time: 05/01/2025 08:50

Lab File ID: VY022095.D Init. Calib. Date(s): 04/22/2025 04/22/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:39 16:15

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.829	1.811		-0.98	20
m/p-Xylenes	0.728	0.745		2.34	20
o-Xylene	0.671	0.676		0.75	20
Styrene	1.126	1.172		4.09	20
Bromoform	0.229	0.238	0.1	3.93	20
Isopropylbenzene	3.341	3.204		-4.1	20
1,1,2,2-Tetrachloroethane	0.563	0.530	0.3	-5.86	20
1,3-Dichlorobenzene	1.714	1.634		-4.67	20
1,4-Dichlorobenzene	1.698	1.606		-5.42	20
1,2-Dichlorobenzene	1.495	1.427		-4.55	20
1,2-Dibromo-3-Chloropropane	0.090	0.085		-5.56	20
1,2,4-Trichlorobenzene	0.847	0.802		-5.31	20
1,2,3-Trichlorobenzene	0.731	0.697		-4.65	20
1,2-Dichloroethane-d4	0.462	0.431		-6.71	20
Dibromofluoromethane	0.330	0.335		1.51	20
Toluene-d8	1.245	1.270		2.01	20
4-Bromofluorobenzene	0.419	0.427		1.91	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



SAMPLE RAW DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
 Data File : VY022105.D
 Acq On : 01 May 2025 14:15
 Operator : SY/MD
 Sample : Q1901-01
 Misc : 6.67g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-170-SB00

A
B
C
D
E
F
G
H
I
J

Quant Time: May 02 01:30:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	310481	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	585270	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	519369	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	196391	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	153526	53.533	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.060%
35) Dibromofluoromethane	7.634	113	196359	50.783	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.560%
50) Toluene-d8	10.109	98	702240	48.174	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.340%
62) 4-Bromofluorobenzene	12.408	95	269562	54.932	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	109.860%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
Data File : VY022105.D
Acq On : 01 May 2025 14:15
Operator : SY/MD
Sample : Q1901-01
Misc : 6.67g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-170-SB00

Quant Time: May 02 01:30:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 982
Delta R.T. 0.000 min
Lab File: VY022105.D
Acq: 01 May 2025 14:15




Instrument :
MSVOA_Y
ClientSampleId :
B-170-SB00


Tgt Ion:168 Resp: 310481
Ion Ratio Lower Upper
168 100
99 54.3 43.1 64.7



#33
1,2-Dichloroethane-d4
Concen: 53.533 ug/l
RT: 8.061 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VY022105.D
Acq: 01 May 2025 14:15

Tgt Ion: 65 Resp: 153526
Ion Ratio Lower Upper
65 100
67 52.9 0.0 105.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VY022105.D

Acq: 01 May 2025 14:15

Instrument :

MSVOA_Y

ClientSampleId :

B-170-SB00





Tgt Ion:114 Resp: 585270

Ion Ratio Lower Upper

114 100

63 17.1 0.0 35.4

88 14.4 0.0 28.2



#35

Dibromofluoromethane

Concen: 50.783 ug/l

RT: 7.634 min Scan# 970

Delta R.T. 0.000 min

Lab File: VY022105.D

Acq: 01 May 2025 14:15




Tgt Ion:113 Resp: 196359


Ion Ratio Lower Upper

113 100

111 102.3 81.8 122.8

192 19.9 15.6 23.4









#50
Toluene-d8
Concen: 48.174 ug/l
RT: 10.109 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY022105.D
Acq: 01 May 2025 14:15




Instrument :
MSVOA_Y
ClientSampleId :
B-170-SB00


Tgt Ion: 98 Resp: 702240
Ion Ratio Lower Upper
98 100
100 64.0 51.6 77.4



#62
4-Bromofluorobenzene
Concen: 54.932 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. 0.006 min
Lab File: VY022105.D
Acq: 01 May 2025 14:15

Tgt Ion: 95 Resp: 269562
Ion Ratio Lower Upper
95 100
174 89.5 0.0 179.0
176 86.2 0.0 171.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.414 min Scan# 1590

Delta R.T. 0.000 min

Lab File: VY022105.D

Acq: 01 May 2025 14:15

Instrument :

MSVOA_Y

ClientSampleId :

B-170-SB00





Tgt Ion:117 Resp: 519369

Ion Ratio Lower Upper

117 100

82 52.1 42.1 63.1

119 31.6 25.7 38.5



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.347 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VY022105.D

Acq: 01 May 2025 14:15




Tgt Ion:152 Resp: 196391

Ion Ratio Lower Upper

152 100

115 56.3 28.0 84.0

150 158.2 0.0 345.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
Data File : VY022105.D
Acq On : 01 May 2025 14:15
Operator : SY/MD
Sample : Q1901-01
Misc : 6.67g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-170-SB00

A
B
C
D
E
F
G
H
I
J

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M

Title : SW846 8260

Signal : TIC: VY022105.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	4.610	465	474	484	rVB5	7040	22569	1.27%	0.250%
2	7.634	959	970	976	rBV	266775	648695	36.62%	7.195%
3	7.707	976	982	996	rVB	397528	893839	50.45%	9.914%
4	8.061	1030	1040	1051	rBV	196876	433782	24.49%	4.811%
5	8.616	1122	1131	1145	rBV	657004	1283538	72.45%	14.236%
6	10.109	1368	1376	1390	rBV	1017914	1771561	100.00%	19.648%
7	11.414	1581	1590	1603	rBV	905873	1482936	83.71%	16.447%
8	12.408	1745	1753	1766	rBV	897673	1360181	76.78%	15.086%
9	13.347	1900	1907	1916	rBV	754380	1119179	63.17%	12.413%




Sum of corrected areas: 9016280

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
Data File : VY022105.D
Acq On : 01 May 2025 14:15
Operator : SY/MD
Sample : Q1901-01
Misc : 6.67g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-170-SB00

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
Data File : VY022105.D
Acq On : 01 May 2025 14:15
Operator : SY/MD
Sample : Q1901-01
Misc : 6.67g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-170-SB00

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
Data File : VY022105.D
Acq On : 01 May 2025 14:15
Operator : SY/MD
Sample : Q1901-01
Misc : 6.67g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-170-SB00

A

B

C

D

E

F

G

H

I

J

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
 Data File : VY022087.D
 Acq On : 30 Apr 2025 17:05
 Operator : SY/MD
 Sample : Q1901-02
 Misc : 7.83g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-167-SB01

A
B
C
D
E
F
G
H
I
J

Quant Time: May 01 01:38:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	348376	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	652384	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	556238	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	176600	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	171544	53.310	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 106.620%	
35) Dibromofluoromethane	7.628	113	220356	51.127	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.260%	
50) Toluene-d8	10.103	98	774744	47.680	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 95.360%	
62) 4-Bromofluorobenzene	12.402	95	197352	36.079	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 72.160%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
Data File : VY022087.D
Acq On : 30 Apr 2025 17:05
Operator : SY/MD
Sample : Q1901-02
Misc : 7.83g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-167-SB01

Quant Time: May 01 01:38:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 982
Delta R.T. 0.000 min
Lab File: VY022087.D
Acq: 30 Apr 2025 17:05




Instrument :
MSVOA_Y
ClientSampleId :
B-167-SB01


Tgt Ion:168 Resp: 348376
Ion Ratio Lower Upper
168 100
99 53.2 43.1 64.7



#33
1,2-Dichloroethane-d4
Concen: 53.310 ug/l
RT: 8.061 min Scan# 1040
Delta R.T. -0.000 min
Lab File: VY022087.D
Acq: 30 Apr 2025 17:05

Tgt Ion: 65 Resp: 171544
Ion Ratio Lower Upper
65 100
67 54.4 0.0 105.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.610 min Scan# 1131

Delta R.T. -0.006 min

Lab File: VY022087.D

Acq: 30 Apr 2025 17:05

Instrument :

MSVOA_Y

ClientSampleId :

B-167-SB01





Tgt Ion:114 Resp: 652384

Ion Ratio Lower Upper

114 100

63 17.3 0.0 35.4

88 15.0 0.0 28.2



#35

Dibromofluoromethane

Concen: 51.127 ug/l

RT: 7.628 min Scan# 969

Delta R.T. -0.006 min

Lab File: VY022087.D

Acq: 30 Apr 2025 17:05




Tgt Ion:113 Resp: 220356


Ion Ratio Lower Upper

113 100

111 101.6 81.8 122.8

192 20.5 15.6 23.4









#50
Toluene-d8
Concen: 47.680 ug/l
RT: 10.103 min Scan# 1375
Delta R.T. -0.000 min
Lab File: VY022087.D
Acq: 30 Apr 2025 17:05




Instrument :
MSVOA_Y
ClientSampleId :
B-167-SB01


Tgt Ion: 98 Resp: 774744
Ion Ratio Lower Upper
98 100
100 64.3 51.6 77.4



#62
4-Bromofluorobenzene
Concen: 36.079 ug/l
RT: 12.402 min Scan# 1752
Delta R.T. -0.000 min
Lab File: VY022087.D
Acq: 30 Apr 2025 17:05

Tgt Ion: 95 Resp: 197352
Ion Ratio Lower Upper
95 100
174 89.1 0.0 179.0
176 85.9 0.0 171.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.414 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY022087.D

Acq: 30 Apr 2025 17:05

Instrument :

MSVOA_Y

ClientSampleId :

B-167-SB01





Tgt Ion:117 Resp: 556238

Ion Ratio Lower Upper

117 100

82 50.8 42.1 63.1

119 31.7 25.7 38.5



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.346 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VY022087.D

Acq: 30 Apr 2025 17:05




Tgt Ion:152 Resp: 176600

Ion Ratio Lower Upper

152 100

115 57.1 28.0 84.0

150 156.7 0.0 345.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
 Data File : VY022087.D
 Acq On : 30 Apr 2025 17:05
 Operator : SY/MD
 Sample : Q1901-02
 Misc : 7.83g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-167-SB01

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M

Title : SW846 8260

Signal : TIC: VY022087.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	4.610	466	474	486	rBV3	10797	33974	1.73%	0.360%
2	7.628	960	969	975	rBV	303091	717981	36.65%	7.598%
3	7.707	975	982	995	rVB	430331	1006973	51.40%	10.656%
4	8.055	1029	1039	1053	rBV	220335	507154	25.89%	5.367%
5	8.610	1121	1130	1145	rBV	738952	1430705	73.03%	15.141%
6	10.103	1366	1375	1384	rBV	1162637	1959042	100.00%	20.732%
7	11.414	1581	1590	1603	rBV	989968	1589638	81.14%	16.822%
8	12.408	1745	1753	1764	rBV	637371	1033090	52.73%	10.933%
9	12.584	1772	1782	1783	rBV5	10844	22801	1.16%	0.241%
10	13.346	1900	1907	1914	rBV	661155	1006788	51.39%	10.654%
11	13.883	1990	1995	2001	rVB	37935	61829	3.16%	0.654%
12	14.535	2091	2102	2116	rBV2	26000	79525	4.06%	0.842%




Sum of corrected areas: 9449500

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
Data File : VY022087.D
Acq On : 30 Apr 2025 17:05
Operator : SY/MD
Sample : Q1901-02
Misc : 7.83g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-167-SB01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
Data File : VY022087.D
Acq On : 30 Apr 2025 17:05
Operator : SY/MD
Sample : Q1901-02
Misc : 7.83g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-167-SB01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

A
B
C
D
E
F
G
H
I
J

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
Data File : VY022087.D
Acq On : 30 Apr 2025 17:05
Operator : SY/MD
Sample : Q1901-02
Misc : 7.83g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-167-SB01

A

B

C

D

E

F

G

H

I

J

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
Data File : VY022106.D
Acq On : 01 May 2025 14:38
Operator : SY/MD
Sample : Q1901-03
Misc : 7.35g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-170-SB01

A
B
C
D
E
F
G
H
I
J

Quant Time: May 02 01:30:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	307701	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	584162	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	521346	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	197125	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	156205	54.960	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.920%
35) Dibromofluoromethane	7.634	113	198196	51.356	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.720%
50) Toluene-d8	10.109	98	696993	47.905	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.800%
62) 4-Bromofluorobenzene	12.407	95	276478	56.448	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	112.900%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
Data File : VY022106.D
Acq On : 01 May 2025 14:38
Operator : SY/MD
Sample : Q1901-03
Misc : 7.35g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-170-SB01

Quant Time: May 02 01:30:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 99
Delta R.T. 0.006 min
Lab File: VY022106.D
Acq: 01 May 2025 14:38




Instrument :
MSVOA_Y
ClientSampleId :
B-170-SB01


Tgt Ion:168 Resp: 307701
Ion Ratio Lower Upper
168 100
99 51.8 43.1 64.7



#33
1,2-Dichloroethane-d4
Concen: 54.960 ug/l
RT: 8.061 min Scan# 1040
Delta R.T. -0.000 min
Lab File: VY022106.D
Acq: 01 May 2025 14:38

Tgt Ion: 65 Resp: 156205
Ion Ratio Lower Upper
65 100
67 53.9 0.0 105.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.615 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VY022106.D

Acq: 01 May 2025 14:38

Instrument :

MSVOA_Y

ClientSampleId :

B-170-SB01





Tgt Ion:114 Resp: 584162

Ion Ratio Lower Upper

114 100

63 16.9 0.0 35.4

88 14.6 0.0 28.2



#35

Dibromofluoromethane

Concen: 51.356 ug/l

RT: 7.634 min Scan# 970

Delta R.T. -0.000 min

Lab File: VY022106.D

Acq: 01 May 2025 14:38




Tgt Ion:113 Resp: 198196


Ion Ratio Lower Upper

113 100

111 102.7 81.8 122.8

192 20.8 15.6 23.4





#50

Toluene-d8

Concen: 47.905 ug/l

RT: 10.109 min Scan# 1

Delta R.T. 0.006 min

Lab File: VY022106.D

Acq: 01 May 2025 14:38

Instrument :

MSVOA_Y

ClientSampleId :





B-170-SB01

Tgt Ion: 98 Resp: 696993

Ion Ratio Lower Upper

98 100

100 64.3 51.6 77.4



#62

4-Bromofluorobenzene

Concen: 56.448 ug/l

RT: 12.407 min Scan# 1753

Delta R.T. 0.006 min

Lab File: VY022106.D

Acq: 01 May 2025 14:38




Tgt Ion: 95 Resp: 276478


Ion Ratio Lower Upper

95 100

174 89.2 0.0 179.0

176 85.3 0.0 171.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.420 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY022106.D

Acq: 01 May 2025 14:38

Instrument :

MSVOA_Y

ClientSampleId :

B-170-SB01





Tgt Ion:117 Resp: 521346

Ion Ratio Lower Upper

117 100

82 49.5 42.1 63.1

119 32.4 25.7 38.5



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.346 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VY022106.D

Acq: 01 May 2025 14:38




Tgt Ion:152 Resp: 197125

Ion Ratio Lower Upper

152 100

115 56.6 28.0 84.0

150 156.6 0.0 345.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
 Data File : VY022106.D
 Acq On : 01 May 2025 14:38
 Operator : SY/MD
 Sample : Q1901-03
 Misc : 7.35g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-170-SB01

A
B
C
D
E
F
G
H
I
J

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M

Title : SW846 8260

Signal : TIC: VY022106.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	2.537	121	134	135	rBV10	7895	24776	1.40%	0.273%
2	4.616	464	475	484	rBV3	7008	21743	1.23%	0.240%
3	7.634	960	970	976	rBV	273875	653719	37.01%	7.204%
4	7.707	976	982	996	rVB	387917	888752	50.32%	9.794%
5	8.061	1030	1040	1053	rBV	196710	442842	25.07%	4.880%
6	8.615	1122	1131	1142	rBV	648317	1273587	72.11%	14.035%
7	10.109	1366	1376	1394	rBV	1048811	1766285	100.00%	19.464%
8	11.414	1582	1590	1604	rBV	900719	1484318	84.04%	16.357%
9	12.407	1744	1753	1765	rBV	925787	1393935	78.92%	15.361%
10	13.346	1900	1907	1917	rBV	761599	1124448	63.66%	12.391%




Sum of corrected areas: 9074405

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
Data File : VY022106.D
Acq On : 01 May 2025 14:38
Operator : SY/MD
Sample : Q1901-03
Misc : 7.35g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-170-SB01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
Data File : VY022106.D
Acq On : 01 May 2025 14:38
Operator : SY/MD
Sample : Q1901-03
Misc : 7.35g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-170-SB01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

A
B
C
D
E
F
G
H
I
J

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
Data File : VY022106.D
Acq On : 01 May 2025 14:38
Operator : SY/MD
Sample : Q1901-03
Misc : 7.35g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-170-SB01

A

B

C

D

E

F

G

H

I

J

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
 Data File : VY022089.D
 Acq On : 30 Apr 2025 17:52
 Operator : SY/MD
 Sample : Q1901-04
 Misc : 5.52g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-167-SB02

Quant Time: May 01 01:39:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.707	168	381091	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	712829	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	629182	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	233488	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	188495	53.549	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.100%
35) Dibromofluoromethane	7.628	113	240173	50.999	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.000%
50) Toluene-d8	10.103	98	847633	47.743	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.480%
62) 4-Bromofluorobenzene	12.401	95	235165	39.347	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	78.700%
Target Compounds						
						Qvalue
16) Acetone	3.866	43	62048	123.893	ug/l	98
17) Carbon Disulfide	4.104	76	59112	4.894	ug/l	99
18) Methyl Acetate	4.385	43	2398	1.458	ug/l #	72
25) 2-Butanone	6.890	43	18807	23.439	ug/l	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
Data File : VY022089.D
Acq On : 30 Apr 2025 17:52
Operator : SY/MD
Sample : Q1901-04
Misc : 5.52g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-167-SB02

Quant Time: May 01 01:39:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY022089.D
Acq: 30 Apr 2025 17:52




Instrument :
MSVOA_Y
ClientSampleId :
B-167-SB02


Tgt Ion:168 Resp: 381091
Ion Ratio Lower Upper
168 100
99 52.3 43.1 64.7



#16
Acetone
Concen: 123.893 ug/l
RT: 3.866 min Scan# 352
Delta R.T. -0.013 min
Lab File: VY022089.D
Acq: 30 Apr 2025 17:52

Tgt Ion: 43 Resp: 62048
Ion Ratio Lower Upper
43 100
58 32.8 27.0 40.4









#17
Carbon Disulfide
Concen: 4.894 ug/l
RT: 4.104 min Scan# 391
Delta R.T. 0.006 min
Lab File: VY022089.D
Acq: 30 Apr 2025 17:52




Instrument :
MSVOA_Y
ClientSampleId :
B-167-SB02


Tgt Ion: 76 Resp: 59112
Ion Ratio Lower Upper
76 100
78 9.1 7.6 11.4



#18
Methyl Acetate
Concen: 1.458 ug/l
RT: 4.385 min Scan# 437
Delta R.T. -0.000 min
Lab File: VY022089.D
Acq: 30 Apr 2025 17:52

Tgt Ion: 43 Resp: 2398
Ion Ratio Lower Upper
43 100
74 15.2 24.2 36.4#





#25

2-Butanone

Concen: 23.439 ug/l

RT: 6.890 min Scan# 84

Delta R.T. -0.006 min

Lab File: VY022089.D

Acq: 30 Apr 2025 17:52

Instrument :

MSVOA_Y

ClientSampleId :





B-167-SB02

Tgt Ion: 43 Resp: 18807

Ion Ratio Lower Upper

43 100

72 30.3 27.0 40.4



#33

1,2-Dichloroethane-d4

Concen: 53.549 ug/l

RT: 8.055 min Scan# 1039

Delta R.T. -0.006 min

Lab File: VY022089.D




Acq: 30 Apr 2025 17:52


Tgt Ion: 65 Resp: 188495

Ion Ratio Lower Upper

65 100

67 54.0 0.0 105.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.609 min Scan# 1131

Delta R.T. -0.006 min

Lab File: VY022089.D

Acq: 30 Apr 2025 17:52

Instrument :

MSVOA_Y

ClientSampleId :

B-167-SB02





Tgt Ion:114 Resp: 712829

Ion Ratio Lower Upper

114 100

63 18.0 0.0 35.4

88 14.7 0.0 28.2



#35

Dibromofluoromethane

Concen: 50.999 ug/l

RT: 7.628 min Scan# 969

Delta R.T. -0.006 min

Lab File: VY022089.D

Acq: 30 Apr 2025 17:52




Tgt Ion:113 Resp: 240173


Ion Ratio Lower Upper

113 100

111 101.7 81.8 122.8

192 19.9 15.6 23.4









#50
Toluene-d8
Concen: 47.743 ug/l
RT: 10.103 min Scan# 1375
Delta R.T. -0.000 min
Lab File: VY022089.D
Acq: 30 Apr 2025 17:52




Instrument :
MSVOA_Y
ClientSampleId :
B-167-SB02


Tgt Ion: 98 Resp: 847633
Ion Ratio Lower Upper
98 100
100 64.8 51.6 77.4



#62
4-Bromofluorobenzene
Concen: 39.347 ug/l
RT: 12.401 min Scan# 1752
Delta R.T. -0.000 min
Lab File: VY022089.D
Acq: 30 Apr 2025 17:52

Tgt Ion: 95 Resp: 235165
Ion Ratio Lower Upper
95 100
174 89.7 0.0 179.0
176 85.7 0.0 171.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.414 min Scan# 117

Delta R.T. -0.000 min

Lab File: VY022089.D

Acq: 30 Apr 2025 17:52

Instrument :

MSVOA_Y

ClientSampleId :

B-167-SB02





Tgt Ion:117 Resp: 629182

Ion Ratio Lower Upper

117 100

82 51.2 42.1 63.1

119 32.8 25.7 38.5



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.346 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VY022089.D

Acq: 30 Apr 2025 17:52




Tgt Ion:152 Resp: 233488

Ion Ratio Lower Upper

152 100

115 55.7 28.0 84.0

150 154.3 0.0 345.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
 Data File : VY022089.D
 Acq On : 30 Apr 2025 17:52
 Operator : SY/MD
 Sample : Q1901-04
 Misc : 5.52g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-167-SB02

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M

Title : SW846 8260

Signal : TIC: VY022089.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.824	12	17	20	rBV	26571	39947	1.86%	0.367%
2	2.482	118	125	133	rBV10	8194	24622	1.15%	0.226%
3	2.635	144	150	159	rVB	13267	28351	1.32%	0.261%
4	3.866	339	352	364	rBV2	35738	104429	4.87%	0.960%
5	4.104	379	391	405	rBV	28435	81522	3.80%	0.749%
6	4.610	464	474	486	rBV4	11267	36513	1.70%	0.336%
7	6.896	841	849	861	rBV2	12268	34253	1.60%	0.315%
8	7.628	959	969	975	rBV2	334318	784778	36.61%	7.215%
9	7.707	975	982	996	rVB	481283	1101628	51.39%	10.128%
10	8.061	1030	1040	1050	rBV	235746	535711	24.99%	4.925%
11	8.609	1120	1130	1143	rBV	795260	1558696	72.72%	14.329%
12	10.103	1367	1375	1383	rBV	1317102	2143539	100.00%	19.706%
13	11.414	1581	1590	1603	rBV	1131706	1797586	83.86%	16.526%
14	12.408	1745	1753	1765	rBV	761764	1238894	57.80%	11.389%
15	13.346	1899	1907	1917	rVB	871260	1311773	61.20%	12.059%
16	13.883	1990	1995	2001	rVB2	35199	55328	2.58%	0.509%




Sum of corrected areas: 10877570

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
Data File : VY022089.D
Acq On : 30 Apr 2025 17:52
Operator : SY/MD
Sample : Q1901-04
Misc : 5.52g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-167-SB02

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
Data File : VY022089.D
Acq On : 30 Apr 2025 17:52
Operator : SY/MD
Sample : Q1901-04
Misc : 5.52g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-167-SB02

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
Data File : VY022089.D
Acq On : 30 Apr 2025 17:52
Operator : SY/MD
Sample : Q1901-04
Misc : 5.52g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-167-SB02

A

B

C

D

E

F

G

H

I

J

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
 Data File : VY022090.D
 Acq On : 30 Apr 2025 18:16
 Operator : SY/MD
 Sample : Q1901-05
 Misc : 5.52g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-170-SB02

A
B
C
D
E
F
G
H
I
J

Quant Time: May 01 01:39:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.707	168	384809	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	709625	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	563942	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	150773	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	176326	49.608	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.220%
35) Dibromofluoromethane	7.628	113	233310	49.766	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.540%
50) Toluene-d8	10.103	98	825206	46.689	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.380%
62) 4-Bromofluorobenzene	12.402	95	196123	32.962	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	65.920%
Target Compounds						
					Qvalue	
17) Carbon Disulfide	4.098	76	126155	10.343	ug/l	98
40) Benzene	8.079	78	25594	1.300	ug/l	97
68) m/p-Xylenes	11.627	106	14972	1.824	ug/l	96
69) o-Xylene	11.957	106	10754	1.421	ug/l	96
73) Isopropylbenzene	12.255	105	101940	10.120	ug/l	99
84) 1,2,4-Trimethylbenzene	13.042	105	10477	1.262	ug/l	88


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
Data File : VY022090.D
Acq On : 30 Apr 2025 18:16
Operator : SY/MD
Sample : Q1901-05
Misc : 5.52g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-170-SB02

Quant Time: May 01 01:39:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 982
Delta R.T. 0.000 min
Lab File: VY022090.D
Acq: 30 Apr 2025 18:16




Instrument :
MSVOA_Y
ClientSampleId :
B-170-SB02


Tgt Ion:168 Resp: 384809
Ion Ratio Lower Upper
168 100
99 52.7 43.1 64.7



#17
Carbon Disulfide
Concen: 10.343 ug/l
RT: 4.098 min Scan# 390
Delta R.T. 0.000 min
Lab File: VY022090.D
Acq: 30 Apr 2025 18:16

Tgt Ion: 76 Resp: 126155
Ion Ratio Lower Upper
76 100
78 8.7 7.6 11.4





#33

1,2-Dichloroethane-d4

Concen: 49.608 ug/l

RT: 8.061 min Scan# 110

Delta R.T. -0.000 min

Lab File: VY022090.D

Acq: 30 Apr 2025 18:16

Instrument :

MSVOA_Y

ClientSampleId :





B-170-SB02

Tgt Ion: 65 Resp: 176326

Ion Ratio Lower Upper

65 100

67 54.2 0.0 105.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.610 min Scan# 1130

Delta R.T. -0.006 min

Lab File: VY022090.D

Acq: 30 Apr 2025 18:16




Tgt Ion: 114 Resp: 709625


Ion Ratio Lower Upper

114 100

63 17.3 0.0 35.4

88 15.0 0.0 28.2





#35

Dibromofluoromethane

Concen: 49.766 ug/l

RT: 7.628 min Scan# 91

Delta R.T. -0.006 min

Lab File: VY022090.D

Acq: 30 Apr 2025 18:16

Instrument :

MSVOA_Y

ClientSampleId :

B-170-SB02





Tgt Ion:113 Resp: 233310

Ion Ratio Lower Upper

113 100

111 101.8 81.8 122.8

192 19.9 15.6 23.4



#40

Benzene

Concen: 1.300 ug/l

RT: 8.079 min Scan# 1043

Delta R.T. -0.000 min

Lab File: VY022090.D




Acq: 30 Apr 2025 18:16


Tgt Ion: 78 Resp: 25594

Ion Ratio Lower Upper

78 100

77 25.4 19.3 28.9









#50
Toluene-d8
Concen: 46.689 ug/l
RT: 10.103 min Scan# 1375
Delta R.T. -0.000 min
Lab File: VY022090.D
Acq: 30 Apr 2025 18:16




Instrument :
MSVOA_Y
ClientSampleId :
B-170-SB02


Tgt Ion: 98 Resp: 825206
Ion Ratio Lower Upper
98 100
100 64.0 51.6 77.4



#62
4-Bromofluorobenzene
Concen: 32.962 ug/l
RT: 12.402 min Scan# 1752
Delta R.T. -0.000 min
Lab File: VY022090.D
Acq: 30 Apr 2025 18:16

Tgt Ion: 95 Resp: 196123
Ion Ratio Lower Upper
95 100
174 88.1 0.0 179.0
176 86.9 0.0 171.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.414 min Scan# 1590

Delta R.T. -0.000 min

Lab File: VY022090.D

Acq: 30 Apr 2025 18:16

Instrument :

MSVOA_Y

ClientSampleId :

B-170-SB02





Tgt Ion:117 Resp: 563942

Ion Ratio Lower Upper

117 100

82 52.5 42.1 63.1

119 33.3 25.7 38.5



#68

m/p-Xylenes

Concen: 1.824 ug/l

RT: 11.627 min Scan# 1625

Delta R.T. -0.000 min

Lab File: VY022090.D




Acq: 30 Apr 2025 18:16


Tgt Ion:106 Resp: 14972

Ion Ratio Lower Upper

106 100

91 200.6 156.2 234.2





#69

o-Xylene

Concen: 1.421 ug/l

RT: 11.957 min Scan# 1

Delta R.T. 0.006 min

Lab File: VY022090.D

Acq: 30 Apr 2025 18:16

Instrument :

MSVOA_Y

ClientSampleId :





B-170-SB02

Tgt Ion:106 Resp: 10754

Ion Ratio Lower Upper

106 100

91 200.2 103.0 308.9



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.347 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VY022090.D

Acq: 30 Apr 2025 18:16




Tgt Ion:152 Resp: 150773


Ion Ratio Lower Upper

152 100

115 56.3 28.0 84.0

150 158.4 0.0 345.6









#73
Isopropylbenzene
Concen: 10.120 ug/l
RT: 12.255 min Scan# 1728
Delta R.T. -0.000 min
Lab File: VY022090.D
Acq: 30 Apr 2025 18:16




Instrument :
MSVOA_Y
ClientSampleId :
B-170-SB02

Tgt Ion:105 Resp: 101940
Ion Ratio Lower Upper
105 100
120 27.5 13.5 40.5



#84
1,2,4-Trimethylbenzene
Concen: 1.262 ug/l
RT: 13.042 min Scan# 1857
Delta R.T. 0.000 min
Lab File: VY022090.D
Acq: 30 Apr 2025 18:16

Tgt Ion:105 Resp: 10477
Ion Ratio Lower Upper
105 100
120 54.6 23.4 70.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
 Data File : VY022090.D
 Acq On : 30 Apr 2025 18:16
 Operator : SY/MD
 Sample : Q1901-05
 Misc : 5.52g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-170-SB02

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Title : SW846 8260

Signal : TIC: VY022090.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	4.098	379	390	404	rBV	58790	162890	7.86%	1.557%
2	4.604	462	473	487	rBV2	14528	47906	2.31%	0.458%
3	7.634	958	970	975	rBV	315823	762259	36.76%	7.287%
4	7.707	975	982	993	rVB	480223	1107079	53.39%	10.583%
5	8.061	1030	1040	1053	rBV2	233211	554610	26.75%	5.302%
6	8.610	1122	1130	1143	rBV	800530	1550988	74.80%	14.826%
7	10.103	1368	1375	1382	rBV	1256679	2073621	100.00%	19.822%
8	10.164	1382	1385	1398	rVB	23397	45119	2.18%	0.431%
9	11.414	1582	1590	1602	rBV	1004276	1642238	79.20%	15.699%
10	11.627	1617	1625	1634	rVB	51745	100894	4.87%	0.964%
11	11.950	1672	1678	1686	rVB2	38480	63628	3.07%	0.608%
12	12.255	1721	1728	1736	rBV	164363	259914	12.53%	2.485%
13	12.402	1744	1752	1762	rBV	625201	1028788	49.61%	9.834%
14	13.042	1852	1857	1862	rVB	20750	32057	1.55%	0.306%
15	13.347	1900	1907	1918	rBV	564738	889239	42.88%	8.500%
16	13.548	1932	1940	1945	rBV2	16611	26148	1.26%	0.250%
17	13.883	1989	1995	2001	rBV2	51425	85381	4.12%	0.816%
18	15.145	2192	2202	2209	rBV	16275	28335	1.37%	0.271%




Sum of corrected areas: 10461094

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
Data File : VY022090.D
Acq On : 30 Apr 2025 18:16
Operator : SY/MD
Sample : Q1901-05
Misc : 5.52g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-170-SB02

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
Data File : VY022090.D
Acq On : 30 Apr 2025 18:16
Operator : SY/MD
Sample : Q1901-05
Misc : 5.52g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-170-SB02

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
Data File : VY022090.D
Acq On : 30 Apr 2025 18:16
Operator : SY/MD
Sample : Q1901-05
Misc : 5.52g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-170-SB02

A

B

C

D

E

F

G

H

I

J

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
 Data File : VX045979.D
 Acq On : 29 Apr 2025 13:58
 Operator : JC/MD
 Sample : Q1901-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB04262025

A
B
C
D
E
F
G
H
I
J

Quant Time: Apr 30 01:38:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.544	168	64509	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	127954	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	115874	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48370	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64442	54.626	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 109.260%			
35) Dibromofluoromethane	5.385	113	46820	51.573	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.140%			
50) Toluene-d8	8.647	98	159888	50.459	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.920%			
62) 4-Bromofluorobenzene	11.079	95	58165	50.394	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 100.780%			
Target Compounds					Qvalue	
16) Acetone	2.380	43	2285	4.717	ug/l	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
Data File : VX045979.D
Acq On : 29 Apr 2025 13:58
Operator : JC/MD
Sample : Q1901-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB04262025

Quant Time: Apr 30 01:38:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX045979.D
Acq: 29 Apr 2025 13:58




Instrument :
MSVOA_X
ClientSampleId :
FB04262025


Tgt Ion:168 Resp: 64509
Ion Ratio Lower Upper
168 100
99 70.6 52.3 78.5



#16
Acetone
Concen: 4.717 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX045979.D
Acq: 29 Apr 2025 13:58

Tgt Ion: 43 Resp: 2285
Ion Ratio Lower Upper
43 100
58 24.8 22.3 33.5





#33

1,2-Dichloroethane-d4

Concen: 54.626 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX045979.D

Acq: 29 Apr 2025 13:58

Instrument :

MSVOA_X

ClientSampleId :





FB04262025

Tgt Ion: 65 Resp: 64442

Ion Ratio Lower Upper

65 100

67 48.7 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX045979.D

Acq: 29 Apr 2025 13:58




Tgt Ion: 114 Resp: 127954


Ion Ratio Lower Upper

114 100

63 23.5 0.0 46.8

88 16.7 0.0 35.4





#35

Dibromofluoromethane

Concen: 51.573 ug/l

RT: 5.385 min Scan# 704

Delta R.T. 0.006 min

Lab File: VX045979.D


Acq: 29 Apr 2025 13:58

Instrument :

MSVOA_X

ClientSampleId :

FB04262025






Tgt Ion:113 Resp: 46820

Ion Ratio Lower Upper

113 100

111 100.7 81.8 122.6

192 15.9 13.8 20.6



#50

Toluene-d8

Concen: 50.459 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX045979.D




Acq: 29 Apr 2025 13:58


Tgt Ion: 98 Resp: 159888

Ion Ratio Lower Upper

98 100

100 64.6 52.2 78.4





#62

4-Bromofluorobenzene

Concen: 50.394 ug/l

RT: 11.079 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX045979.D

Acq: 29 Apr 2025 13:58

Instrument :

MSVOA_X

ClientSampleId :

FB04262025





Tgt Ion: 95 Resp: 58165

Ion Ratio Lower Upper

95 100

174 66.7 0.0 135.8

176 64.7 0.0 131.4



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.049 min Scan# 1470

Delta R.T. 0.000 min

Lab File: VX045979.D

Acq: 29 Apr 2025 13:58




Tgt Ion: 117 Resp: 115874


Ion Ratio Lower Upper

117 100

82 62.0 49.2 73.8

119 31.4 25.1 37.7





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX045979.D


Acq: 29 Apr 2025 13:58

Instrument :

MSVOA_X

ClientSampleId :

FB04262025





Tgt Ion:152 Resp: 48370

Ion Ratio Lower Upper

152 100

115 65.3 46.9 140.7

150 159.8 0.0 349.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
 Data File : VX045979.D
 Acq On : 29 Apr 2025 13:58
 Operator : JC/MD
 Sample : Q1901-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB04262025

A
B
C
D
E
F
G
H
I
J

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M

Title : SW846 8260

Signal : TIC: VX045979.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.575	71	80	81	rBV3	4097	8657	1.93%	0.368%
2	1.599	81	84	89	rVB2	5445	8862	1.98%	0.377%
3	5.379	693	704	721	rBV2	52966	157443	35.14%	6.694%
4	5.544	721	731	747	rVV	75321	216016	48.21%	9.185%
5	5.952	789	798	814	rBV	62439	167024	37.28%	7.102%
6	6.757	919	930	946	rBV	131566	322972	72.08%	13.732%
7	8.647	1233	1240	1256	rBV	279471	448060	100.00%	19.051%
8	10.049	1465	1470	1488	rBV	280871	399593	89.18%	16.990%
9	11.079	1634	1639	1652	rBV	228889	289322	64.57%	12.301%
10	12.018	1788	1793	1807	rBV	265146	333976	74.54%	14.200%




Sum of corrected areas: 2351925

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
Data File : VX045979.D
Acq On : 29 Apr 2025 13:58
Operator : JC/MD
Sample : Q1901-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB04262025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
Data File : VX045979.D
Acq On : 29 Apr 2025 13:58
Operator : JC/MD
Sample : Q1901-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB04262025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
Data File : VX045979.D
Acq On : 29 Apr 2025 13:58
Operator : JC/MD
Sample : Q1901-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB04262025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc
------------------	----	---------	-------	----------	---	----	------	------

A
B
C
D
E
F
G
H
I
J

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
Data File : VX045978.D
Acq On : 29 Apr 2025 13:35
Operator : JC/MD
Sample : Q1901-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB04262025

A
B
C
D
E
F
G
H
I
J

Quant Time: Apr 30 01:38:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.544	168	67274	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	131446	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	123412	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53777	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66629	54.158	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.320%	
35) Dibromofluoromethane	5.379	113	48413	51.911	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.820%	
50) Toluene-d8	8.647	98	165475	50.834	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.660%	
62) 4-Bromofluorobenzene	11.079	95	63879	53.875	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 107.740%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	854	1.690	ug/l	91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
Data File : VX045978.D
Acq On : 29 Apr 2025 13:35
Operator : JC/MD
Sample : Q1901-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB04262025

Quant Time: Apr 30 01:38:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX045978.D
Acq: 29 Apr 2025 13:35




Instrument :
MSVOA_X
ClientSampleId :
TB04262025


Tgt Ion:168 Resp: 67274
Ion Ratio Lower Upper
168 100
99 76.0 52.3 78.5



#16
Acetone
Concen: 1.690 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX045978.D
Acq: 29 Apr 2025 13:35

Tgt Ion: 43 Resp: 854
Ion Ratio Lower Upper
43 100
58 32.7 22.3 33.5





#33

1,2-Dichloroethane-d4

Concen: 54.158 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX045978.D

Acq: 29 Apr 2025 13:35

Instrument :

MSVOA_X

ClientSampleId :





TB04262025

Tgt Ion: 65 Resp: 66629

Ion Ratio Lower Upper

65 100

67 48.3 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX045978.D

Acq: 29 Apr 2025 13:35




Tgt Ion: 114 Resp: 131446


Ion Ratio Lower Upper

114 100

63 23.7 0.0 46.8

88 17.5 0.0 35.4





#35

Dibromofluoromethane

Concen: 51.911 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.000 min

Lab File: VX045978.D


Acq: 29 Apr 2025 13:35

Instrument :

MSVOA_X

ClientSampleId :

TB04262025






Tgt Ion:113 Resp: 48413

Ion Ratio Lower Upper

113 100

111 101.4 81.8 122.6

192 16.2 13.8 20.6



#50

Toluene-d8

Concen: 50.834 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX045978.D




Acq: 29 Apr 2025 13:35


Tgt Ion: 98 Resp: 165475

Ion Ratio Lower Upper

98 100

100 64.7 52.2 78.4





#62

4-Bromofluorobenzene

Concen: 53.875 ug/l

RT: 11.079 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX045978.D

Acq: 29 Apr 2025 13:35

Instrument :

MSVOA_X

ClientSampleId :

TB04262025





Tgt Ion: 95 Resp: 63879

Ion Ratio Lower Upper

95 100

174 66.4 0.0 135.8

176 64.0 0.0 131.4



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.049 min Scan# 1470

Delta R.T. -0.000 min

Lab File: VX045978.D

Acq: 29 Apr 2025 13:35




Tgt Ion: 117 Resp: 123412


Ion Ratio Lower Upper

117 100

82 61.8 49.2 73.8

119 31.9 25.1 37.7





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX045978.D


Acq: 29 Apr 2025 13:35

Instrument :

MSVOA_X

ClientSampleId :

TB04262025





Tgt Ion:152 Resp: 53777

Ion Ratio Lower Upper

152 100

115 63.8 46.9 140.7

150 155.3 0.0 349.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
Data File : VX045978.D
Acq On : 29 Apr 2025 13:35
Operator : JC/MD
Sample : Q1901-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB04262025

A
B
C
D
E
F
G
H
I
J

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs : 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Title : SW846 8260

Signal : TIC: VX045978.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.587	69	82	83	rBV3	5015	13453	2.89%	0.542%
2	1.611	83	86	88	rVB	4904	5340	1.15%	0.215%
3	2.544	235	239	251	rVB	3493	6601	1.42%	0.266%
4	5.379	694	704	721	rBV	55481	165258	35.52%	6.654%
5	5.544	721	731	744	rBV	77631	222460	47.82%	8.957%
6	5.952	789	798	812	rBV	65089	172328	37.04%	6.938%
7	6.757	921	930	947	rBV	136842	336873	72.41%	13.563%
8	8.647	1234	1240	1254	rBV	294795	465244	100.00%	18.732%
9	10.049	1465	1470	1490	rBV	295471	421063	90.50%	16.953%
10	11.079	1634	1639	1652	rBV	243881	309987	66.63%	12.481%
11	12.018	1788	1793	1810	rBV	288514	365134	78.48%	14.701%




Sum of corrected areas: 2483741

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
Data File : VX045978.D
Acq On : 29 Apr 2025 13:35
Operator : JC/MD
Sample : Q1901-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB04262025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
Data File : VX045978.D
Acq On : 29 Apr 2025 13:35
Operator : JC/MD
Sample : Q1901-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB04262025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
Data File : VX045978.D
Acq On : 29 Apr 2025 13:35
Operator : JC/MD
Sample : Q1901-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB04262025

A

B

C

D

E

F

G

H

I

J

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
 Data File : VX045974.D
 Acq On : 29 Apr 2025 11:58
 Operator : JC/MD
 Sample : VX0429WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0429WBL01

A
B
C
D
E
F
G
H
I
J

Quant Time: Apr 30 01:35:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	63902	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	127024	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	116648	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	47876	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64358	55.073	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.140%
35) Dibromofluoromethane	5.379	113	46535	51.635	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.260%
50) Toluene-d8	8.647	98	156429	49.728	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.460%
62) 4-Bromofluorobenzene	11.079	95	57215	49.934	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.860%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
Data File : VX045974.D
Acq On : 29 Apr 2025 11:58
Operator : JC/MD
Sample : VX0429WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0429WBL01

Quant Time: Apr 30 01:35:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 732
Delta R.T. -0.000 min
Lab File: VX045974.D
Acq: 29 Apr 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VX0429WBL01


Tgt Ion:168 Resp: 63902
Ion Ratio Lower Upper
168 100
99 69.0 52.3 78.5



#33
1,2-Dichloroethane-d4
Concen: 55.073 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.000 min
Lab File: VX045974.D
Acq: 29 Apr 2025 11:58

Tgt Ion: 65 Resp: 64358
Ion Ratio Lower Upper
65 100
67 48.7 0.0 99.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 91

Delta R.T. -0.000 min

Lab File: VX045974.D

Acq: 29 Apr 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VX0429WBL01





Tgt Ion:114 Resp: 127024

Ion Ratio Lower Upper

114 100

63 23.6 0.0 46.8

88 16.3 0.0 35.4



#35

Dibromofluoromethane

Concen: 51.635 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.000 min

Lab File: VX045974.D

Acq: 29 Apr 2025 11:58




Tgt Ion:113 Resp: 46535


Ion Ratio Lower Upper

113 100

111 102.7 81.8 122.6

192 16.7 13.8 20.6





#50

Toluene-d8

Concen: 49.728 ug/l

RT: 8.647 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX045974.D

Acq: 29 Apr 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :





VX0429WBL01

Tgt Ion: 98 Resp: 156429

Ion Ratio Lower Upper

98 100

100 66.1 52.2 78.4



#62

4-Bromofluorobenzene

Concen: 49.934 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. -0.000 min

Lab File: VX045974.D

Acq: 29 Apr 2025 11:58




Tgt Ion: 95 Resp: 57215


Ion Ratio Lower Upper

95 100

174 69.2 0.0 135.8

176 66.8 0.0 131.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.049 min Scan# 1470

Delta R.T. -0.000 min

Lab File: VX045974.D

Acq: 29 Apr 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VX0429WBL01





Tgt Ion:117 Resp: 116648

Ion Ratio Lower Upper

117 100

82 62.1 49.2 73.8

119 31.1 25.1 37.7



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX045974.D

Acq: 29 Apr 2025 11:58




Tgt Ion:152 Resp: 47876

Ion Ratio Lower Upper

152 100

115 63.2 46.9 140.7

150 155.7 0.0 349.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
Data File : VX045974.D
Acq On : 29 Apr 2025 11:58
Operator : JC/MD
Sample : VX0429WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0429WBL01

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M

Title : SW846 8260

Signal : TIC: VX045974.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.587	70	82	83	rBV4	5005	12957	2.94%	0.557%
2	1.605	83	85	88	rVB2	4539	4821	1.09%	0.207%
3	5.379	693	704	721	rBV	52386	158935	36.02%	6.827%
4	5.544	721	731	745	rBV	73483	211614	47.96%	9.089%
5	5.952	787	798	813	rBV	60594	167974	38.07%	7.215%
6	6.757	921	930	946	rBV2	132130	322044	72.99%	13.833%
7	8.647	1233	1240	1253	rBV	279678	441191	100.00%	18.951%
8	10.049	1465	1470	1489	rBV	282498	397007	89.99%	17.053%
9	11.079	1634	1639	1651	rBV	224816	286625	64.97%	12.311%
10	12.018	1788	1793	1808	rBV	256856	324949	73.65%	13.958%


Sum of corrected areas: 2328117

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
Data File : VX045974.D
Acq On : 29 Apr 2025 11:58
Operator : JC/MD
Sample : VX0429WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0429WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
Data File : VX045974.D
Acq On : 29 Apr 2025 11:58
Operator : JC/MD
Sample : VX0429WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0429WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
Data File : VX045974.D
Acq On : 29 Apr 2025 11:58
Operator : JC/MD
Sample : VX0429WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0429WBL01

A

B

C

D

E

F

G

H

I

J

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
 Data File : VY022071.D
 Acq On : 30 Apr 2025 09:57
 Operator : SY/MD
 Sample : VY0430SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0430SBL01

A
B
C
D
E
F
G
H
I
J

Quant Time: May 01 01:31:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	311598	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	580894	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	511027	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	186896	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	150607	52.327	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.660%
35) Dibromofluoromethane	7.634	113	198874	51.821	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.640%
50) Toluene-d8	10.109	98	697047	48.178	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.360%
62) 4-Bromofluorobenzene	12.408	95	259683	53.317	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	106.640%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
Data File : VY022071.D
Acq On : 30 Apr 2025 09:57
Operator : SY/MD
Sample : VY0430SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0430SBL01

Quant Time: May 01 01:31:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 982
Delta R.T. 0.000 min
Lab File: VY022071.D
Acq: 30 Apr 2025 09:57




Instrument :
MSVOA_Y
ClientSampleId :
VY0430SBL01


Tgt Ion:168 Resp: 311598
Ion Ratio Lower Upper
168 100
99 51.8 43.1 64.7



#33
1,2-Dichloroethane-d4
Concen: 52.327 ug/l
RT: 8.061 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VY022071.D
Acq: 30 Apr 2025 09:57

Tgt Ion: 65 Resp: 150607
Ion Ratio Lower Upper
65 100
67 53.4 0.0 105.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VY022071.D

Acq: 30 Apr 2025 09:57

Instrument :

MSVOA_Y

ClientSampleId :

VY0430SBL01





Tgt Ion:114 Resp: 580894

Ion Ratio Lower Upper

114 100

63 16.8 0.0 35.4

88 14.7 0.0 28.2



#35

Dibromofluoromethane

Concen: 51.821 ug/l

RT: 7.634 min Scan# 970

Delta R.T. 0.000 min

Lab File: VY022071.D

Acq: 30 Apr 2025 09:57




Tgt Ion:113 Resp: 198874


Ion Ratio Lower Upper

113 100

111 102.6 81.8 122.8

192 20.5 15.6 23.4









#50
Toluene-d8
Concen: 48.178 ug/l
RT: 10.109 min Scan# 1375
Delta R.T. 0.006 min
Lab File: VY022071.D
Acq: 30 Apr 2025 09:57




Instrument :
MSVOA_Y
ClientSampleId :
VY0430SBL01


Tgt Ion: 98 Resp: 697047
Ion Ratio Lower Upper
98 100
100 63.8 51.6 77.4



#62
4-Bromofluorobenzene
Concen: 53.317 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. 0.006 min
Lab File: VY022071.D
Acq: 30 Apr 2025 09:57

Tgt Ion: 95 Resp: 259683
Ion Ratio Lower Upper
95 100
174 91.2 0.0 179.0
176 87.8 0.0 171.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.414 min Scan# 1590

Delta R.T. 0.000 min

Lab File: VY022071.D


Acq: 30 Apr 2025 09:57

Instrument :

MSVOA_Y

ClientSampleId :

VY0430SBL01






Tgt Ion:117 Resp: 511027

Ion Ratio Lower Upper

117 100

82 52.2 42.1 63.1

119 31.6 25.7 38.5



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.347 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VY022071.D

Acq: 30 Apr 2025 09:57




Tgt Ion:152 Resp: 186896

Ion Ratio Lower Upper

152 100

115 55.6 28.0 84.0

150 155.1 0.0 345.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
Data File : VY022071.D
Acq On : 30 Apr 2025 09:57
Operator : SY/MD
Sample : VY0430SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0430SBL01

A
B
C
D
E
F
G
H
I
J

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M

Title : SW846 8260

Signal : TIC: VY022071.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	7.634	959	970	976	rBV	275568	662472	37.68%	7.480%
2	7.707	976	982	996	rVB	395565	894886	50.90%	10.104%
3	8.061	1030	1040	1051	rBV2	190974	427512	24.31%	4.827%
4	8.616	1122	1131	1144	rBV	648749	1273883	72.45%	14.384%
5	10.109	1367	1376	1395	rBV	1027561	1758262	100.00%	19.853%
6	11.414	1583	1590	1603	rBV	899099	1457691	82.91%	16.459%
7	12.408	1745	1753	1762	rBV	891073	1325837	75.41%	14.970%
8	13.347	1900	1907	1915	rBV	712938	1055872	60.05%	11.922%


Sum of corrected areas: 8856415

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
Data File : VY022071.D
Acq On : 30 Apr 2025 09:57
Operator : SY/MD
Sample : VY0430SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0430SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
Data File : VY022071.D
Acq On : 30 Apr 2025 09:57
Operator : SY/MD
Sample : VY0430SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0430SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
Data File : VY022071.D
Acq On : 30 Apr 2025 09:57
Operator : SY/MD
Sample : VY0430SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0430SBL01

A

B

C

D

E

F

G

H

I

J

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST0.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
 Data File : VY022096.D
 Acq On : 01 May 2025 10:11
 Operator : SY/MD
 Sample : VY0501SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0501SBL01

A
B
C
D
E
F
G
H
I
J

Quant Time: May 01 12:53:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	317996	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	585580	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	508953	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	180066	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.060	65	144517	49.201	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.400%
35) Dibromofluoromethane	7.634	113	195715	50.590	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.180%
50) Toluene-d8	10.109	98	696910	47.783	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.560%
62) 4-Bromofluorobenzene	12.407	95	260745	53.107	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	106.220%


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
Data File : VY022096.D
Acq On : 01 May 2025 10:11
Operator : SY/MD
Sample : VY0501SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0501SBL01

Quant Time: May 01 12:53:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 982
Delta R.T. -0.000 min
Lab File: VY022096.D
Acq: 01 May 2025 10:11




Instrument :
MSVOA_Y
ClientSampleId :
VY0501SBL01


Tgt Ion:168 Resp: 317996
Ion Ratio Lower Upper
168 100
99 52.4 43.1 64.7



#33
1,2-Dichloroethane-d4
Concen: 49.201 ug/l
RT: 8.060 min Scan# 1040
Delta R.T. -0.000 min
Lab File: VY022096.D
Acq: 01 May 2025 10:11

Tgt Ion: 65 Resp: 144517
Ion Ratio Lower Upper
65 100
67 53.4 0.0 105.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.615 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VY022096.D

Acq: 01 May 2025 10:11

Instrument :

MSVOA_Y

ClientSampleId :

VY0501SBL01





Tgt Ion:114 Resp: 585580

Ion Ratio Lower Upper

114 100

63 17.5 0.0 35.4

88 14.7 0.0 28.2



#35

Dibromofluoromethane

Concen: 50.590 ug/l

RT: 7.634 min Scan# 970

Delta R.T. -0.000 min

Lab File: VY022096.D

Acq: 01 May 2025 10:11




Tgt Ion:113 Resp: 195715


Ion Ratio Lower Upper

113 100

111 101.2 81.8 122.8

192 20.6 15.6 23.4





#50

Toluene-d8

Concen: 47.783 ug/l

RT: 10.109 min Scan# 1375

Delta R.T. 0.006 min

Lab File: VY022096.D

Acq: 01 May 2025 10:11

Instrument :

MSVOA_Y

ClientSampleId :





VY0501SBL01

Tgt Ion: 98 Resp: 696910

Ion Ratio Lower Upper

98 100

100 64.2 51.6 77.4



#62

4-Bromofluorobenzene

Concen: 53.107 ug/l

RT: 12.407 min Scan# 1753

Delta R.T. 0.006 min

Lab File: VY022096.D

Acq: 01 May 2025 10:11




Tgt Ion: 95 Resp: 260745


Ion Ratio Lower Upper

95 100

174 91.9 0.0 179.0

176 88.4 0.0 171.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.420 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY022096.D

Acq: 01 May 2025 10:11

Instrument :

MSVOA_Y

ClientSampleId :

VY0501SBL01





Tgt Ion:117 Resp: 508953

Ion Ratio Lower Upper

117 100

82 50.4 42.1 63.1

119 31.8 25.7 38.5



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.346 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VY022096.D

Acq: 01 May 2025 10:11




Tgt Ion:152 Resp: 180066

Ion Ratio Lower Upper

152 100

115 55.9 28.0 84.0

150 156.0 0.0 345.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
Data File : VY022096.D
Acq On : 01 May 2025 10:11
Operator : SY/MD
Sample : VY0501SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0501SBL01

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M

Title : SW846 8260

Signal : TIC: VY022096.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	2.507	120	129	130	rBV7	7844	17809	1.02%	0.202%
2	7.634	960	970	976	rBV	269609	642245	36.64%	7.286%
3	7.707	976	982	996	rVB	395838	906459	51.71%	10.283%
4	8.060	1031	1040	1053	rBV	186572	414158	23.63%	4.698%
5	8.615	1122	1131	1147	rBV	653108	1282212	73.15%	14.546%
6	10.109	1368	1376	1384	rBV	1041890	1752899	100.00%	19.886%
7	11.420	1582	1591	1603	rBV	891567	1442394	82.29%	16.363%
8	12.407	1746	1753	1768	rBV	873676	1332394	76.01%	15.116%
9	13.346	1901	1907	1922	rVB	682693	1024172	58.43%	11.619%


Sum of corrected areas: 8814742

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
Data File : VY022096.D
Acq On : 01 May 2025 10:11
Operator : SY/MD
Sample : VY0501SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0501SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
Data File : VY022096.D
Acq On : 01 May 2025 10:11
Operator : SY/MD
Sample : VY0501SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0501SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
Data File : VY022096.D
Acq On : 01 May 2025 10:11
Operator : SY/MD
Sample : VY0501SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0501SBL01

A

B

C

D

E

F

G

H

I

J

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
 Data File : VX045975.D
 Acq On : 29 Apr 2025 12:21
 Operator : JC/MD
 Sample : VX0429WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0429WBS01

Quant Time: Apr 30 01:36:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 04/30/2025
 Supervised By : Semsettin Yesilyurt 04/30/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	80632	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	140487	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	122568	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	59236	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.946	65	79501	53.915	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.840%	
35) Dibromofluoromethane	5.379	113	55216	55.396	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.800%	
50) Toluene-d8	8.647	98	184446	53.016	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.040%	
62) 4-Bromofluorobenzene	11.079	95	71406	56.347	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 112.700%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.167	85	20933	17.360	ug/l	99
3) Chloromethane	1.301	50	20239	16.337	ug/l	96
4) Vinyl Chloride	1.374	62	19140	16.883	ug/l	99
5) Bromomethane	1.593	94	9415	17.515	ug/l	100
6) Chloroethane	1.673	64	12245	20.358	ug/l	98
7) Trichlorofluoromethane	1.880	101	32749	19.371	ug/l	98
8) Diethyl Ether	2.130	74	11050	19.469	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.319	101	20234	20.425	ug/l	98
10) Methyl Iodide	2.447	142	22380	18.112	ug/l	96
11) Tert butyl alcohol	2.971	59	20596	103.932	ug/l	98
12) 1,1-Dichloroethene	2.313	96	17859	18.450	ug/l	96
13) Acrolein	2.233	56	30163	110.531	ug/l	99
14) Allyl chloride	2.654	41	35730	19.454	ug/l	99
15) Acrylonitrile	3.063	53	61950	99.034	ug/l	98
16) Acetone	2.380	43	59194	97.762	ug/l	97
17) Carbon Disulfide	2.508	76	32070	13.414	ug/l	99
18) Methyl Acetate	2.703	43	39576	28.389	ug/l	98
19) Methyl tert-butyl Ether	3.111	73	71602	21.145	ug/l	98
20) Methylene Chloride	2.782	84	21785	19.218	ug/l	99
21) trans-1,2-Dichloroethene	3.087	96	17645	17.843	ug/l	93
22) Diisopropyl ether	3.758	45	75771	20.943	ug/l #	87
23) Vinyl Acetate	3.721	43	305248	97.653	ug/l	99
24) 1,1-Dichloroethane	3.605	63	41051	20.063	ug/l	97
25) 2-Butanone	4.556	43	89868	101.395	ug/l	96
26) 2,2-Dichloropropane	4.471	77	31094	23.521	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	23917	19.851	ug/l	98
28) Bromochloromethane	4.891	49	21747	21.994	ug/l	97
29) Tetrahydrofuran	5.007	42	55579	96.722	ug/l	98
30) Chloroform	5.087	83	44363	21.072	ug/l	97
31) Cyclohexane	5.465	56	32630	18.039	ug/l	95
32) 1,1,1-Trichloroethane	5.379	97	36667	20.583	ug/l	99
36) 1,1-Dichloropropene	5.690	75	26951	20.026	ug/l	99
37) Ethyl Acetate	4.709	43	32815	19.348	ug/l	99
38) Carbon Tetrachloride	5.672	117	30835	21.200	ug/l	91
39) Methylcyclohexane	7.379	83	31049	18.821	ug/l	99
40) Benzene	6.032	78	81137	19.740	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
 Data File : VX045975.D
 Acq On : 29 Apr 2025 12:21
 Operator : JC/MD
 Sample : VX0429WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0429WBS01

Quant Time: Apr 30 01:36:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :John Carlone 04/30/2025
 Supervised By :Semsettin Yesilyurt 04/30/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	19414	21.695	ug/l	98
42) 1,2-Dichloroethane	6.080	62	36439	21.464	ug/l	100
43) Isopropyl Acetate	6.342	43	52611	20.467	ug/l	99
44) Trichloroethene	7.123	130	19439	19.898	ug/l	99
45) 1,2-Dichloropropane	7.428	63	21438	20.904	ug/l	100
46) Dibromomethane	7.580	93	16625	21.118	ug/l	99
47) Bromodichloromethane	7.818	83	33664	21.412	ug/l	99
48) Methyl methacrylate	7.690	41	28494	21.469	ug/l	98
49) 1,4-Dioxane	7.659	88	9791	407.741	ug/l	98
51) 4-Methyl-2-Pentanone	8.574	43	180040	105.654	ug/l	99
52) Toluene	8.714	92	50978	20.497	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	27141	19.419	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	31830	21.544	ug/l	97
55) 1,1,2-Trichloroethane	9.147	97	21739	21.934	ug/l	98
56) Ethyl methacrylate	9.116	69	33190	21.639	ug/l	99
57) 1,3-Dichloropropane	9.305	76	36368	21.073	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	86646	111.669	ug/l	99
59) 2-Hexanone	9.427	43	135946	107.732	ug/l	98
60) Dibromochloromethane	9.519	129	22822	21.122	ug/l	100
61) 1,2-Dibromoethane	9.610	107	21796	21.725	ug/l	97
64) Tetrachloroethene	9.269	164	17310	20.003	ug/l	96
65) Chlorobenzene	10.080	112	54355	20.753	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	19764	21.736	ug/l	100
67) Ethyl Benzene	10.189	91	97814	20.852	ug/l	97
68) m/p-Xylenes	10.299	106	71200	41.733	ug/l	97
69) o-Xylene	10.640	106	35169	20.926	ug/l	97
70) Styrene	10.653	104	59455	21.434	ug/l	98
71) Bromoform	10.799	173	14381	21.181	ug/l #	100
73) Isopropylbenzene	10.957	105	96809	20.403	ug/l	98
74) N-amyl acetate	10.842	43	45523	20.057	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.207	83	34807	20.882	ug/l	98
76) 1,2,3-Trichloropropane	11.238	75	28492m	19.719	ug/l	
77) Bromobenzene	11.195	156	21147	19.289	ug/l	99
78) n-propylbenzene	11.299	91	108978	19.940	ug/l	100
79) 2-Chlorotoluene	11.360	91	69847	19.991	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	80136	20.424	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	7544	17.720	ug/l	93
82) 4-Chlorotoluene	11.451	91	78882	20.249	ug/l	100
83) tert-Butylbenzene	11.713	119	79835	20.552	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	80856	20.532	ug/l	100
85) sec-Butylbenzene	11.890	105	99135	20.778	ug/l	100
86) p-Isopropyltoluene	12.006	119	80992	20.591	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	40055	20.071	ug/l	100
88) 1,4-Dichlorobenzene	12.037	146	40055	19.798	ug/l	97
89) n-Butylbenzene	12.329	91	67884	19.902	ug/l	99
90) Hexachloroethane	12.536	117	13846	20.482	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	42151	21.245	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	7487	21.478	ug/l	93
93) 1,2,4-Trichlorobenzene	13.585	180	21510	19.477	ug/l	99
94) Hexachlorobutadiene	13.719	225	10079	21.179	ug/l	98
95) Naphthalene	13.774	128	79412	19.322	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	23277	20.128	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
Data File : VX045975.D
Acq On : 29 Apr 2025 12:21
Operator : JC/MD
Sample : VX0429WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Apr 30 01:36:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VX0429WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/30/2025
Supervised By :Semsettin Yesilyurt 04/30/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
 Data File : VX045975.D
 Acq On : 29 Apr 2025 12:21
 Operator : JC/MD
 Sample : VX0429WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0429WBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 04/30/2025
 Supervised By :Semsettin Yesilyurt 04/30/2025

Quant Time: Apr 30 01:36:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
 Data File : VY022072.D
 Acq On : 30 Apr 2025 10:30
 Operator : SY/MD
 Sample : VY0430SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0430SBS01

Quant Time: May 01 01:31:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 05/05/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.707	168	286487	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	433702	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	405299	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	222589	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	131013	49.509	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	99.020%		
35) Dibromofluoromethane	7.628	113	150835	52.643	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	105.280%		
50) Toluene-d8	10.103	98	568923	52.668	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	105.340%		
62) 4-Bromofluorobenzene	12.402	95	188289	51.779	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	103.560%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.861	85	47203	19.336	ug/l	98
3) Chloromethane	2.062	50	77084	22.176	ug/l	98
4) Vinyl Chloride	2.202	62	88437	20.704	ug/l	98
5) Bromomethane	2.586	94	83723	22.751	ug/l	97
6) Chloroethane	2.727	64	61631	21.186	ug/l	100
7) Trichlorofluoromethane	3.050	101	122851	21.811	ug/l	97
8) Diethyl Ether	3.452	74	28746	19.608	ug/l	99
9) 1,1,2-Trichlorotrifluo...	3.806	101	62287	20.385	ug/l	99
10) Methyl Iodide	4.001	142	60748	17.172	ug/l	99
11) Tert butyl alcohol	4.885	59	13923	95.838	ug/l #	87
12) 1,1-Dichloroethene	3.787	96	56168	19.722	ug/l	97
13) Acrolein	3.653	56	22527	86.396	ug/l	99
14) Allyl chloride	4.379	41	62659	18.895	ug/l	99
15) Acrylonitrile	5.055	53	51344	95.667	ug/l	99
16) Acetone	3.879	43	37345	94.535	ug/l	95
17) Carbon Disulfide	4.098	76	180355	19.861	ug/l	98
18) Methyl Acetate	4.385	43	25031	20.250	ug/l	99
19) Methyl tert-butyl Ether	5.116	73	121034	18.988	ug/l	100
20) Methylene Chloride	4.610	84	67082	20.048	ug/l	95
21) trans-1,2-Dichloroethene	5.104	96	62226	19.509	ug/l	97
22) Diisopropyl ether	6.013	45	144165	19.535	ug/l	96
23) Vinyl Acetate	5.958	43	381862	96.013	ug/l	99
24) 1,1-Dichloroethane	5.909	63	101810	19.903	ug/l	95
25) 2-Butanone	6.890	43	55933	92.728	ug/l	97
26) 2,2-Dichloropropane	6.878	77	91436	20.394	ug/l	98
27) cis-1,2-Dichloroethene	6.890	96	69433	19.469	ug/l	99
28) Bromochloromethane	7.244	49	39832	20.047	ug/l	96
29) Tetrahydrofuran	7.262	42	37976	96.936	ug/l	99
30) Chloroform	7.415	83	115625	20.378	ug/l	99
31) Cyclohexane	7.701	56	82489	18.857	ug/l	98
32) 1,1,1-Trichloroethane	7.610	97	103303	20.124	ug/l	99
36) 1,1-Dichloropropene	7.829	75	77745	20.200	ug/l	99
37) Ethyl Acetate	6.982	43	28394	20.541	ug/l	97
38) Carbon Tetrachloride	7.817	117	97877	21.282	ug/l	96
39) Methylcyclohexane	9.103	83	92916	19.184	ug/l	94
40) Benzene	8.079	78	247402	20.563	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
 Data File : VY022072.D
 Acq On : 30 Apr 2025 10:30
 Operator : SY/MD
 Sample : VY0430SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0430SBS01

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 05/05/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 01 01:31:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.232	41	15307	20.257	ug/l #	94
42) 1,2-Dichloroethane	8.159	62	63278	21.243	ug/l	99
43) Isopropyl Acetate	8.195	43	51820	19.521	ug/l	99
44) Trichloroethene	8.860	130	69720	20.921	ug/l	96
45) 1,2-Dichloropropane	9.140	63	54103	20.334	ug/l	96
46) Dibromomethane	9.225	93	34989	21.008	ug/l	99
47) Bromodichloromethane	9.420	83	87089	20.931	ug/l	96
48) Methyl methacrylate	9.219	41	23285	18.974	ug/l	94
49) 1,4-Dioxane	9.231	88	6982	393.308	ug/l	91
51) 4-Methyl-2-Pentanone	10.000	43	140074	100.658	ug/l	100
52) Toluene	10.170	92	163564	21.002	ug/l	100
53) t-1,3-Dichloropropene	10.390	75	73442	20.578	ug/l	98
54) cis-1,3-Dichloropropene	9.853	75	87328	20.585	ug/l	96
55) 1,1,2-Trichloroethane	10.573	97	46477	20.891	ug/l	98
56) Ethyl methacrylate	10.439	69	47660	18.733	ug/l	97
57) 1,3-Dichloropropane	10.713	76	73654	20.651	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.713	63	125085	100.556	ug/l	98
59) 2-Hexanone	10.762	43	91664	99.867	ug/l	98
60) Dibromochloromethane	10.908	129	65090	21.132	ug/l	99
61) 1,2-Dibromoethane	11.012	107	43234	20.584	ug/l	99
64) Tetrachloroethene	10.646	164	80599	21.080	ug/l	96
65) Chlorobenzene	11.438	112	182280	20.111	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.512	131	66284	20.773	ug/l	100
67) Ethyl Benzene	11.518	91	285528	19.258	ug/l	100
68) m/p-Xylenes	11.627	106	239665	40.623	ug/l	97
69) o-Xylene	11.957	106	106876	19.654	ug/l	99
70) Styrene	11.969	104	182708	20.012	ug/l	100
71) Bromoform	12.127	173	38719	20.871	ug/l #	99
73) Isopropylbenzene	12.255	105	279562	18.799	ug/l	99
74) N-amyl acetate	12.072	43	45327	17.740	ug/l	96
75) 1,1,2,2-Tetrachloroethane	12.505	83	49468	19.730	ug/l	99
76) 1,2,3-Trichloropropane	12.554	75	35538m	18.640	ug/l	
77) Bromobenzene	12.530	156	73867	19.657	ug/l	97
78) n-propylbenzene	12.597	91	342875	19.361	ug/l	100
79) 2-Chlorotoluene	12.676	91	200281	19.446	ug/l	99
80) 1,3,5-Trimethylbenzene	12.737	105	235676	19.234	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.304	75	14374	18.155	ug/l	94
82) 4-Chlorotoluene	12.774	91	212735	19.870	ug/l	100
83) tert-Butylbenzene	12.999	119	205879	18.735	ug/l	99
84) 1,2,4-Trimethylbenzene	13.042	105	241750	19.725	ug/l	98
85) sec-Butylbenzene	13.176	105	309315	19.235	ug/l	99
86) p-Isopropyltoluene	13.292	119	260507	19.078	ug/l	99
87) 1,3-Dichlorobenzene	13.286	146	152160	19.937	ug/l	99
88) 1,4-Dichlorobenzene	13.365	146	148875	19.695	ug/l	99
89) n-Butylbenzene	13.615	91	225950	18.513	ug/l	99
90) Hexachloroethane	13.883	117	58174	19.391	ug/l	98
91) 1,2-Dichlorobenzene	13.658	146	134059	20.139	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.273	75	7240	18.150	ug/l	97
93) 1,2,4-Trichlorobenzene	14.919	180	72337	19.176	ug/l	99
94) Hexachlorobutadiene	15.023	225	47088	20.289	ug/l	100
95) Naphthalene	15.145	128	109185	17.424	ug/l	99
96) 1,2,3-Trichlorobenzene	15.328	180	64048	19.670	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
Data File : VY022072.D
Acq On : 30 Apr 2025 10:30
Operator : SY/MD
Sample : VY0430SBS01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0430SBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/05/2025
Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 01 01:31:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_Y
ClientSampleId :
VY0430SBS01

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 05/05/2025
Supervised By :Mahesh Dadoda 05/05/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
 Data File : VY022097.D
 Acq On : 01 May 2025 10:42
 Operator : SY/MD
 Sample : VY0501SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0501SBS01

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 05/05/2025
 Supervised By : Mahesh Dadoda 05/05/2025

Quant Time: May 01 12:53:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.707	168	294485	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	452802	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	414235	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	225406	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	129149	47.479	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	94.960%		
35) Dibromofluoromethane	7.634	113	148223	49.549	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	99.100%		
50) Toluene-d8	10.109	98	564747	50.076	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	100.160%		
62) 4-Bromofluorobenzene	12.408	95	187006	49.257	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	98.520%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.867	85	48273	19.238	ug/l	99
3) Chloromethane	2.068	50	66754	18.683	ug/l	99
4) Vinyl Chloride	2.202	62	82174	18.716	ug/l	94
5) Bromomethane	2.592	94	70373	18.604	ug/l	97
6) Chloroethane	2.732	64	56478	18.887	ug/l	99
7) Trichlorofluoromethane	3.056	101	115218	19.900	ug/l	98
8) Diethyl Ether	3.452	74	28039	18.606	ug/l	99
9) 1,1,2-Trichlorotrifluo...	3.812	101	62306	19.838	ug/l	97
10) Methyl Iodide	4.001	142	67604	18.591	ug/l	100
11) Tert butyl alcohol	4.866	59	12138	81.282	ug/l	98
12) 1,1-Dichloroethene	3.787	96	58265	19.903	ug/l	98
13) Acrolein	3.653	56	17856	66.621	ug/l	97
14) Allyl chloride	4.379	41	63510	18.632	ug/l	97
15) Acrylonitrile	5.055	53	48849	88.546	ug/l	98
16) Acetone	3.866	43	37306	91.214	ug/l	99
17) Carbon Disulfide	4.104	76	182459	19.547	ug/l	98
18) Methyl Acetate	4.379	43	23736	18.681	ug/l	95
19) Methyl tert-butyl Ether	5.116	73	117798	17.978	ug/l	99
20) Methylene Chloride	4.610	84	66341	19.288	ug/l	97
21) trans-1,2-Dichloroethene	5.110	96	64503	19.674	ug/l	95
22) Diisopropyl ether	6.012	45	147267	19.413	ug/l	96
23) Vinyl Acetate	5.958	43	372032	91.000	ug/l	99
24) 1,1-Dichloroethane	5.915	63	104275	19.831	ug/l #	96
25) 2-Butanone	6.896	43	53781	86.739	ug/l	98
26) 2,2-Dichloropropane	6.884	77	93707	20.333	ug/l	99
27) cis-1,2-Dichloroethene	6.890	96	72723	19.837	ug/l	99
28) Bromochloromethane	7.244	49	38682	18.940	ug/l	97
29) Tetrahydrofuran	7.256	42	34870	86.590	ug/l	99
30) Chloroform	7.421	83	118101	20.249	ug/l	98
31) Cyclohexane	7.701	56	84609	18.817	ug/l #	92
32) 1,1,1-Trichloroethane	7.616	97	106076	20.103	ug/l	99
36) 1,1-Dichloropropene	7.835	75	79939	19.894	ug/l	100
37) Ethyl Acetate	6.982	43	25337	17.556	ug/l	97
38) Carbon Tetrachloride	7.817	117	99102	20.639	ug/l	98
39) Methylcyclohexane	9.109	83	95859	18.956	ug/l	96
40) Benzene	8.079	78	254338	20.248	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
 Data File : VY022097.D
 Acq On : 01 May 2025 10:42
 Operator : SY/MD
 Sample : VY0501SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0501SBS01

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 05/05/2025
 Supervised By : Mahesh Dadoda 05/05/2025

Quant Time: May 01 12:53:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.213	41	12694	16.090	ug/l #	84
42) 1,2-Dichloroethane	8.158	62	61949	19.920	ug/l	99
43) Isopropyl Acetate	8.195	43	48886	17.639	ug/l	99
44) Trichloroethene	8.866	130	70695	20.319	ug/l	97
45) 1,2-Dichloropropane	9.140	63	55919	20.131	ug/l	98
46) Dibromomethane	9.231	93	34103	19.612	ug/l	98
47) Bromodichloromethane	9.420	83	89355	20.570	ug/l	97
48) Methyl methacrylate	9.219	41	21714	16.948	ug/l	93
49) 1,4-Dioxane	9.231	88	6896	372.078	ug/l	95
51) 4-Methyl-2-Pentanone	9.999	43	127974	88.084	ug/l	99
52) Toluene	10.170	92	167187	20.562	ug/l	99
53) t-1,3-Dichloropropene	10.396	75	70469	18.912	ug/l	98
54) cis-1,3-Dichloropropene	9.853	75	88193	19.912	ug/l	98
55) 1,1,2-Trichloroethane	10.573	97	46804	20.150	ug/l	96
56) Ethyl methacrylate	10.438	69	46419	17.476	ug/l	97
57) 1,3-Dichloropropane	10.719	76	73345	19.697	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.713	63	112618	86.715	ug/l	100
59) 2-Hexanone	10.762	43	81146	84.678	ug/l	98
60) Dibromochloromethane	10.908	129	63759	19.827	ug/l	98
61) 1,2-Dibromoethane	11.018	107	43603	19.884	ug/l	96
64) Tetrachloroethene	10.646	164	81769	20.925	ug/l	98
65) Chlorobenzene	11.444	112	183689	19.830	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.517	131	65069	19.952	ug/l	97
67) Ethyl Benzene	11.517	91	295862	19.525	ug/l	99
68) m/p-Xylenes	11.627	106	244316	40.518	ug/l	99
69) o-Xylene	11.956	106	109092	19.629	ug/l	99
70) Styrene	11.969	104	187388	20.081	ug/l	99
71) Bromoform	12.133	173	37191	19.615	ug/l #	98
73) Isopropylbenzene	12.255	105	288742	19.173	ug/l	100
74) N-amyl acetate	12.072	43	41858	16.178	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.505	83	46736	18.407	ug/l	97
76) 1,2,3-Trichloropropane	12.554	75	35111m	18.186	ug/l	
77) Bromobenzene	12.530	156	73542	19.326	ug/l	98
78) n-propylbenzene	12.597	91	349898	19.510	ug/l	100
79) 2-Chlorotoluene	12.682	91	203324	19.494	ug/l	100
80) 1,3,5-Trimethylbenzene	12.737	105	243534	19.627	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.298	75	13702	17.090	ug/l	96
82) 4-Chlorotoluene	12.779	91	211849	19.539	ug/l	98
83) tert-Butylbenzene	12.999	119	211608	19.015	ug/l	99
84) 1,2,4-Trimethylbenzene	13.042	105	242538	19.542	ug/l	98
85) sec-Butylbenzene	13.176	105	316527	19.438	ug/l	99
86) p-Isopropyltoluene	13.292	119	268577	19.423	ug/l	99
87) 1,3-Dichlorobenzene	13.285	146	153561	19.869	ug/l	99
88) 1,4-Dichlorobenzene	13.371	146	150028	19.600	ug/l	99
89) n-Butylbenzene	13.621	91	230584	18.656	ug/l	99
90) Hexachloroethane	13.877	117	59920	19.723	ug/l	99
91) 1,2-Dichlorobenzene	13.657	146	130714	19.391	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.273	75	7153	17.708	ug/l	89
93) 1,2,4-Trichlorobenzene	14.919	180	72623	19.012	ug/l	99
94) Hexachlorobutadiene	15.023	225	47671	20.283	ug/l	97
95) Naphthalene	15.145	128	104223	16.425	ug/l	99
96) 1,2,3-Trichlorobenzene	15.328	180	61917	18.778	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
Data File : VY022097.D
Acq On : 01 May 2025 10:42
Operator : SY/MD
Sample : VY0501SBS01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0501SBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/05/2025
Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 01 12:53:46 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
 Data File : VY022097.D
 Acq On : 01 May 2025 10:42
 Operator : SY/MD
 Sample : VY0501SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0501SBS01

Quant Time: May 01 12:53:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 05/05/2025
 Supervised By :Mahesh Dadoda 05/05/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
 Data File : VX045976.D
 Acq On : 29 Apr 2025 12:49
 Operator : JC/MD
 Sample : VX0429WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0429WBSD01

Manual Integrations APPROVED

Reviewed By : John Carlone 04/30/2025
 Supervised By : Semsettin Yesilyurt 04/30/2025

Quant Time: Apr 30 01:37:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	77095	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	135079	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	118384	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	57470	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.946	65	76515	54.271	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.540%	
35) Dibromofluoromethane	5.379	113	52478	54.757	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.520%	
50) Toluene-d8	8.647	98	172943	51.700	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.400%	
62) 4-Bromofluorobenzene	11.079	95	67496	55.394	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 110.780%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.166	85	19940	17.295	ug/l	98
3) Chloromethane	1.307	50	19464	16.433	ug/l	97
4) Vinyl Chloride	1.374	62	18281	16.865	ug/l	97
5) Bromomethane	1.593	94	8917	17.349	ug/l	97
6) Chloroethane	1.672	64	10885	18.927	ug/l	98
7) Trichlorofluoromethane	1.880	101	31444	19.453	ug/l	99
8) Diethyl Ether	2.130	74	11373	20.957	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.319	101	20162	21.286	ug/l	96
10) Methyl Iodide	2.447	142	21226	17.966	ug/l	95
11) Tert butyl alcohol	2.965	59	21757	114.828	ug/l	100
12) 1,1-Dichloroethene	2.313	96	17753	19.182	ug/l	97
13) Acrolein	2.233	56	29845	114.384	ug/l	99
14) Allyl chloride	2.660	41	34613	19.710	ug/l	100
15) Acrylonitrile	3.062	53	63530	106.219	ug/l	99
16) Acetone	2.380	43	59734	103.180	ug/l	98
17) Carbon Disulfide	2.508	76	30724	13.441	ug/l	98
18) Methyl Acetate	2.703	43	39952	29.973	ug/l	98
19) Methyl tert-butyl Ether	3.111	73	70922	21.905	ug/l	100
20) Methylene Chloride	2.782	84	22066	20.359	ug/l	99
21) trans-1,2-Dichloroethene	3.087	96	17617	18.632	ug/l	95
22) Diisopropyl ether	3.757	45	74574	21.558	ug/l	94
23) Vinyl Acetate	3.715	43	306061	102.405	ug/l	99
24) 1,1-Dichloroethane	3.599	63	39043	19.957	ug/l	99
25) 2-Butanone	4.556	43	92556	109.219	ug/l	98
26) 2,2-Dichloropropane	4.465	77	29266	23.154	ug/l	100
27) cis-1,2-Dichloroethene	4.477	96	23658	20.537	ug/l	97
28) Bromochloromethane	4.891	49	21694	22.947	ug/l	96
29) Tetrahydrofuran	5.007	42	57573	104.788	ug/l	99
30) Chloroform	5.086	83	43275	21.498	ug/l	97
31) Cyclohexane	5.458	56	31835	18.407	ug/l	94
32) 1,1,1-Trichloroethane	5.373	97	35637	20.922	ug/l	99
36) 1,1-Dichloropropene	5.684	75	25072	19.375	ug/l	100
37) Ethyl Acetate	4.708	43	32920	20.187	ug/l	99
38) Carbon Tetrachloride	5.672	117	29700	21.237	ug/l	92
39) Methylcyclohexane	7.373	83	31598	19.921	ug/l	99
40) Benzene	6.031	78	78335	19.821	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
 Data File : VX045976.D
 Acq On : 29 Apr 2025 12:49
 Operator : JC/MD
 Sample : VX0429WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0429WBSD01

Manual Integrations APPROVED

Reviewed By :John Carlone 04/30/2025
 Supervised By :Semsettin Yesilyurt 04/30/2025

Quant Time: Apr 30 01:37:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	19612	22.794	ug/l	98
42) 1,2-Dichloroethane	6.080	62	36209	22.183	ug/l	100
43) Isopropyl Acetate	6.336	43	54185	21.923	ug/l	99
44) Trichloroethene	7.123	130	18371	19.557	ug/l	96
45) 1,2-Dichloropropane	7.421	63	21416	21.719	ug/l	100
46) Dibromomethane	7.580	93	16618	21.954	ug/l	99
47) Bromodichloromethane	7.818	83	32301	21.368	ug/l	97
48) Methyl methacrylate	7.690	41	27818	21.799	ug/l	100
49) 1,4-Dioxane	7.659	88	10150	439.614	ug/l	98
51) 4-Methyl-2-Pentanone	8.568	43	187368	114.356	ug/l	98
52) Toluene	8.714	92	48009	20.076	ug/l	98
53) t-1,3-Dichloropropene	8.976	75	27378	20.260	ug/l	93
54) cis-1,3-Dichloropropene	8.360	75	30230	21.280	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	20571	21.587	ug/l	96
56) Ethyl methacrylate	9.116	69	32392	21.964	ug/l	98
57) 1,3-Dichloropropane	9.305	76	36964	22.275	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	85871	115.101	ug/l	99
59) 2-Hexanone	9.427	43	137555	113.371	ug/l	98
60) Dibromochloromethane	9.519	129	22935	22.077	ug/l	96
61) 1,2-Dibromoethane	9.604	107	21310	22.091	ug/l	99
64) Tetrachloroethene	9.269	164	16478	19.715	ug/l	96
65) Chlorobenzene	10.073	112	54296	21.464	ug/l	96
66) 1,1,1,2-Tetrachloroethane	10.159	131	19338	22.019	ug/l	100
67) Ethyl Benzene	10.189	91	96091	21.209	ug/l	97
68) m/p-Xylenes	10.299	106	70144	42.567	ug/l	100
69) o-Xylene	10.640	106	34876	21.485	ug/l	100
70) Styrene	10.653	104	58240	21.738	ug/l	99
71) Bromoform	10.799	173	13887	21.177	ug/l #	96
73) Isopropylbenzene	10.957	105	91610	19.901	ug/l	99
74) N-amyl acetate	10.842	43	45116	20.488	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.207	83	34033	21.045	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	29260m	20.873	ug/l	
77) Bromobenzene	11.195	156	20879	19.630	ug/l	96
78) n-propylbenzene	11.299	91	106699	20.123	ug/l	99
79) 2-Chlorotoluene	11.360	91	66357	19.576	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	79294	20.831	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	7415	17.952	ug/l	93
82) 4-Chlorotoluene	11.451	91	77140	20.411	ug/l	99
83) tert-Butylbenzene	11.713	119	78461	20.819	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	78597	20.572	ug/l	99
85) sec-Butylbenzene	11.890	105	98401	21.258	ug/l	99
86) p-Isopropyltoluene	12.006	119	79322	20.786	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	39159	20.225	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	39249	19.996	ug/l	96
89) n-Butylbenzene	12.329	91	66533	20.105	ug/l	100
90) Hexachloroethane	12.536	117	13828	21.084	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	40899	21.247	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	7537	22.286	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	22349	20.858	ug/l	98
94) Hexachlorobutadiene	13.719	225	9997	21.652	ug/l	96
95) Naphthalene	13.774	128	82571	20.708	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	22765	20.290	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
Data File : VX045976.D
Acq On : 29 Apr 2025 12:49
Operator : JC/MD
Sample : VX0429WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0429WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/30/2025
Supervised By :Semsettin Yesilyurt 04/30/2025

Quant Time: Apr 30 01:37:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VX0429WBSD01

Manual Integrations APPROVED

Reviewed By :John Carlone 04/30/2025
Supervised By :Semsettin Yesilyurt 04/30/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
 Data File : VY022073.D
 Acq On : 30 Apr 2025 10:52
 Operator : SY/MD
 Sample : VY0430SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0430SBS01

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 05/05/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 01 01:32:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	289313	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	439988	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	406945	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	218185	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.061	65	130907	48.986	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.980%
35) Dibromofluoromethane	7.634	113	147653	50.796	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.600%
50) Toluene-d8	10.103	98	562631	51.341	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.680%
62) 4-Bromofluorobenzene	12.402	95	187095	50.715	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	101.440%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.861	85	49931	20.254	ug/l	99
3) Chloromethane	2.068	50	68325	19.465	ug/l	99
4) Vinyl Chloride	2.202	62	84965	19.697	ug/l	100
5) Bromomethane	2.586	94	72512	19.512	ug/l	94
6) Chloroethane	2.727	64	57263	19.492	ug/l	100
7) Trichlorofluoromethane	3.050	101	116170	20.423	ug/l	97
8) Diethyl Ether	3.452	74	27804	18.780	ug/l	98
9) 1,1,2-Trichlorotrifluo...	3.818	101	62953	20.402	ug/l	100
10) Methyl Iodide	3.995	142	67087	18.779	ug/l	100
11) Tert butyl alcohol	4.860	59	14050	95.768	ug/l #	70
12) 1,1-Dichloroethene	3.787	96	58062	20.188	ug/l	95
13) Acrolein	3.653	56	22474	85.350	ug/l	98
14) Allyl chloride	4.379	41	63213	18.876	ug/l	98
15) Acrylonitrile	5.055	53	51846	95.658	ug/l	100
16) Acetone	3.873	43	36077	89.420	ug/l	92
17) Carbon Disulfide	4.104	76	180959	19.733	ug/l	99
18) Methyl Acetate	4.385	43	26790	21.461	ug/l	99
19) Methyl tert-butyl Ether	5.110	73	119952	18.634	ug/l	98
20) Methylene Chloride	4.610	84	67948	20.108	ug/l	97
21) trans-1,2-Dichloroethene	5.110	96	64406	19.996	ug/l	97
22) Diisopropyl ether	6.019	45	146890	19.710	ug/l	98
23) Vinyl Acetate	5.958	43	371342	92.455	ug/l	99
24) 1,1-Dichloroethane	5.915	63	102207	19.785	ug/l	95
25) 2-Butanone	6.890	43	55654	91.364	ug/l	100
26) 2,2-Dichloropropane	6.884	77	91186	20.140	ug/l	99
27) cis-1,2-Dichloroethene	6.890	96	70781	19.653	ug/l	98
28) Bromochloromethane	7.244	49	40106	19.988	ug/l	98
29) Tetrahydrofuran	7.262	42	36921	93.322	ug/l	98
30) Chloroform	7.421	83	115178	20.101	ug/l	98
31) Cyclohexane	7.701	56	83695	18.946	ug/l #	91
32) 1,1,1-Trichloroethane	7.616	97	104622	20.181	ug/l	99
36) 1,1-Dichloropropene	7.829	75	78332	20.062	ug/l	99
37) Ethyl Acetate	6.988	43	27342m	19.497	ug/l	
38) Carbon Tetrachloride	7.817	117	96787	20.744	ug/l	96
39) Methylcyclohexane	9.103	83	93388	19.006	ug/l	94
40) Benzene	8.079	78	250157	20.495	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
 Data File : VY022073.D
 Acq On : 30 Apr 2025 10:52
 Operator : SY/MD
 Sample : VY0430SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0430SBS01

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 05/05/2025
 Supervised By : Mahesh Dadoda 05/05/2025

Quant Time: May 01 01:32:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.214	41	15478	20.191	ug/l #	95
42) 1,2-Dichloroethane	8.158	62	61936	20.495	ug/l	99
43) Isopropyl Acetate	8.195	43	51000	18.937	ug/l	98
44) Trichloroethene	8.860	130	69297	20.497	ug/l	93
45) 1,2-Dichloropropane	9.140	63	54935	20.352	ug/l	98
46) Dibromomethane	9.231	93	35003	20.716	ug/l	99
47) Bromodichloromethane	9.420	83	87120	20.639	ug/l	99
48) Methyl methacrylate	9.219	41	23107	18.560	ug/l	95
49) 1,4-Dioxane	9.225	88	7458	414.120	ug/l	93
51) 4-Methyl-2-Pentanone	9.994	43	139092	98.525	ug/l	98
52) Toluene	10.170	92	162291	20.541	ug/l	99
53) t-1,3-Dichloropropene	10.390	75	73102	20.190	ug/l	98
54) cis-1,3-Dichloropropene	9.853	75	86456	20.088	ug/l	97
55) 1,1,2-Trichloroethane	10.567	97	46575	20.636	ug/l	96
56) Ethyl methacrylate	10.439	69	47773	18.509	ug/l	98
57) 1,3-Dichloropropane	10.713	76	73864	20.414	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.707	63	124335	98.525	ug/l	99
59) 2-Hexanone	10.762	43	87825	94.317	ug/l	97
60) Dibromochloromethane	10.908	129	63230	20.235	ug/l	98
61) 1,2-Dibromoethane	11.012	107	42794	20.084	ug/l	98
64) Tetrachloroethene	10.646	164	81751	21.295	ug/l	97
65) Chlorobenzene	11.438	112	184325	20.255	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.512	131	66577	20.780	ug/l	99
67) Ethyl Benzene	11.518	91	289472	19.445	ug/l	100
68) m/p-Xylenes	11.627	106	238544	40.269	ug/l	99
69) o-Xylene	11.951	106	105935	19.403	ug/l	99
70) Styrene	11.969	104	185895	20.278	ug/l	98
71) Bromoform	12.133	173	37727	20.254	ug/l #	99
73) Isopropylbenzene	12.255	105	280709	19.257	ug/l	100
74) N-amyl acetate	12.066	43	44208	17.652	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.505	83	47576	19.358	ug/l	98
76) 1,2,3-Trichloropropane	12.554	75	33877m	18.127	ug/l	
77) Bromobenzene	12.530	156	73436	19.937	ug/l	98
78) n-propylbenzene	12.591	91	339221	19.541	ug/l	100
79) 2-Chlorotoluene	12.676	91	200769	19.887	ug/l	100
80) 1,3,5-Trimethylbenzene	12.737	105	239215	19.917	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.298	75	14890	19.187	ug/l	97
82) 4-Chlorotoluene	12.774	91	212766	20.274	ug/l	100
83) tert-Butylbenzene	12.993	119	209440	19.443	ug/l	100
84) 1,2,4-Trimethylbenzene	13.042	105	238623	19.863	ug/l	98
85) sec-Butylbenzene	13.176	105	313905	19.915	ug/l	100
86) p-Isopropyltoluene	13.292	119	262861	19.639	ug/l	99
87) 1,3-Dichlorobenzene	13.286	146	149261	19.952	ug/l	99
88) 1,4-Dichlorobenzene	13.365	146	148377	20.026	ug/l	99
89) n-Butylbenzene	13.615	91	223801	18.707	ug/l	99
90) Hexachloroethane	13.877	117	58123	19.765	ug/l	99
91) 1,2-Dichlorobenzene	13.658	146	132919	20.371	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.267	75	7198	18.409	ug/l	92
93) 1,2,4-Trichlorobenzene	14.919	180	69519	18.801	ug/l	97
94) Hexachlorobutadiene	15.023	225	46146	20.284	ug/l	98
95) Naphthalene	15.145	128	103951	16.924	ug/l	99
96) 1,2,3-Trichlorobenzene	15.328	180	60230	18.870	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
Data File : VY022073.D
Acq On : 30 Apr 2025 10:52
Operator : SY/MD
Sample : VY0430SBSD01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0430SBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/05/2025
Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 01 01:32:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY043025\
Data File : VY022073.D
Acq On : 30 Apr 2025 10:52
Operator : SY/MD
Sample : VY04305B5D01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0430SBSD01

Manual Integrations

Reviewed By :Semsettin Yesilyurt 05/05/2025
Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 01 01:32:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration



Manual Integration Report

Sequence:	VX040225	Instrument	MSVOA_x
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC001	VX045525.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:20 PM	MMDadoda	4/2/2025 2:16:05 PM	Peak Integrated by Software
VSTDICC001	VX045525.D	1,4-Dichlorobenzene	Amit	4/2/2025 2:14:20 PM	MMDadoda	4/2/2025 2:16:05 PM	Peak Integrated by Software
VSTDICC001	VX045525.D	1,4-Dioxane	Amit	4/2/2025 2:14:20 PM	MMDadoda	4/2/2025 2:16:05 PM	Peak Integrated by Software
VSTDICC001	VX045525.D	Ethyl methacrylate	Amit	4/2/2025 2:14:20 PM	MMDadoda	4/2/2025 2:16:05 PM	Peak Integrated by Software
VSTDICC001	VX045525.D	Methacrylonitrile	Amit	4/2/2025 2:14:20 PM	MMDadoda	4/2/2025 2:16:05 PM	Peak Integrated by Software
VSTDICC005	VX045526.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:22 PM	MMDadoda	4/2/2025 2:16:07 PM	Peak Integrated by Software
VSTDICC020	VX045527.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:25 PM	MMDadoda	4/2/2025 2:16:09 PM	Peak Integrated by Software
VSTDICCC050	VX045528.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:26 PM	MMDadoda	4/2/2025 2:16:11 PM	Peak Integrated by Software
VSTDICC100	VX045529.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:28 PM	MMDadoda	4/2/2025 2:16:15 PM	Peak Integrated by Software
VSTDICC150	VX045530.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:30 PM	MMDadoda	4/2/2025 2:16:46 PM	Peak Integrated by Software
VSTDICV050	VX045532.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:32 PM	MMDadoda	4/2/2025 2:17:50 PM	Peak Integrated by Software
VSTDCCC050	VX045534.D	1,2,3-Trichloropropane	JOHN	4/3/2025 10:04:10 AM	MMDadoda	4/3/2025 1:12:02 PM	Peak Integrated by Software
VSTDCCC050	VX045557.D	1,2,3-Trichloropropane	JOHN	4/3/2025 10:04:58 AM	MMDadoda	4/3/2025 1:12:08 PM	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	VX040225	Instrument	MSVOA_x
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
-----------	---------	-----------	-----------	-----------	---------------	---------------	--------

Manual Integration Report

Sequence:

VX042925

Instrument

MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VX045972.D	1,2,3-Trichloropropane	JOHN	4/30/2025 9:32:49 AM	Sam	4/30/2025 2:23:15 PM	Peak Integrated by Software
VX0429WBS01	VX045975.D	1,2,3-Trichloropropane	JOHN	4/30/2025 9:32:53 AM	Sam	4/30/2025 2:23:17 PM	Peak Integrated by Software
VX0429WBSD01	VX045976.D	1,2,3-Trichloropropane	JOHN	4/30/2025 9:32:57 AM	Sam	4/30/2025 2:23:20 PM	Peak Integrated by Software
VSTDCCC050	VX045983.D	1,2,3-Trichloropropane	JOHN	4/30/2025 9:33:01 AM	Sam	4/30/2025 2:23:25 PM	Peak Integrated by Software

Manual Integration Report

Sequence:	VY042225	Instrument	MSVOA_y
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC005	VY021953.D	1,2,3-Trichloropropane	SAM	4/23/2025 7:34:48 AM	MMDadoda	4/23/2025 1:28:59 PM	Peak Integrated by Software
VSTDICC005	VY021953.D	Ethyl Acetate	SAM	4/23/2025 7:34:48 AM	MMDadoda	4/23/2025 1:28:59 PM	Peak Integrated by Software
VSTDICC005	VY021953.D	Methacrylonitrile	SAM	4/23/2025 7:34:48 AM	MMDadoda	4/23/2025 1:28:59 PM	Peak Integrated by Software
VSTDICC005	VY021953.D	Tert butyl alcohol	SAM	4/23/2025 7:34:48 AM	MMDadoda	4/23/2025 1:28:59 PM	Peak Integrated by Software
VSTDICC010	VY021954.D	1,2,3-Trichloropropane	SAM	4/23/2025 7:34:50 AM	MMDadoda	4/23/2025 1:29:03 PM	Peak Integrated by Software
VSTDICC010	VY021954.D	Ethyl Acetate	SAM	4/23/2025 7:34:50 AM	MMDadoda	4/23/2025 1:29:03 PM	Peak Integrated by Software
VSTDICC010	VY021954.D	Methacrylonitrile	SAM	4/23/2025 7:34:50 AM	MMDadoda	4/23/2025 1:29:03 PM	Peak Integrated by Software
VSTDICC020	VY021955.D	1,2,3-Trichloropropane	SAM	4/23/2025 7:34:51 AM	MMDadoda	4/23/2025 1:29:07 PM	Peak Integrated by Software
VSTDICC020	VY021955.D	Methacrylonitrile	SAM	4/23/2025 7:34:51 AM	MMDadoda	4/23/2025 1:29:07 PM	Peak Integrated by Software
VSTDICCC050	VY021956.D	1,2,3-Trichloropropane	SAM	4/23/2025 7:34:53 AM	MMDadoda	4/23/2025 1:29:11 PM	Peak Integrated by Software
VSTDICC100	VY021957.D	1,2,3-Trichloropropane	SAM	4/23/2025 7:34:56 AM	MMDadoda	4/23/2025 1:29:15 PM	Peak Integrated by Software
VSTDICC150	VY021958.D	1,2,3-Trichloropropane	SAM	4/23/2025 7:34:57 AM	MMDadoda	4/23/2025 1:29:19 PM	Peak Integrated by Software
VSTDICV050	VY021960.D	1,2,3-Trichloropropane	SAM	4/23/2025 7:34:59 AM	MMDadoda	4/23/2025 1:29:23 PM	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

VY042225

Instrument

MSVOA_y

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:	VY043025	Instrument	MSVOA_y
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VY022070.D	1,2,3-Trichloropropane	sam	5/5/2025 4:23:25 AM	MMDadoda	5/5/2025 2:54:38 PM	Peak Integrated by Software
VY0430SBS01	VY022072.D	1,2,3-Trichloropropane	sam	5/5/2025 4:23:30 AM	MMDadoda	5/5/2025 2:54:41 PM	Peak Integrated by Software
VY0430SBSD0 1	VY022073.D	1,2,3-Trichloropropane	sam	5/5/2025 4:23:37 AM	MMDadoda	5/5/2025 2:54:44 PM	Peak Integrated by Software
VY0430SBSD0 1	VY022073.D	Ethyl Acetate	sam	5/5/2025 4:23:37 AM	MMDadoda	5/5/2025 2:54:44 PM	Peak Integrated by Software
VSTDCCC050	VY022093.D	1,2,3-Trichloropropane	sam	5/5/2025 4:23:47 AM	MMDadoda	5/5/2025 2:54:48 PM	Peak Integrated by Software

Manual Integration Report

Sequence:	vy050125	Instrument	MSVOA_y
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VY022095.D	1,2,3-Trichloropropane	Sam	5/5/2025 2:51:05 PM	MMDadoda	5/5/2025 2:57:20 PM	Peak Integrated by Software
VY0501SBS01	VY022097.D	1,2,3-Trichloropropane	sam	5/5/2025 4:29:49 AM	MMDadoda	5/5/2025 2:57:21 PM	Peak Integrated by Software
VSTDCCC050	VY022117.D	1,2,3-Trichloropropane	sam	5/5/2025 4:30:16 AM	MMDadoda	5/5/2025 2:57:56 PM	Peak Integrated by Software

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX040225

Review By	John Carlone	Review On	4/2/2025 9:42:46 AM
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:39 PM
SubDirectory	VX040225	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133559,VP133563		
Initial Calibration Stds	VP133569,VP133570,VP133571,VP133572,VP133573,VP133574		
CCC	VP133564,VP133565		
Internal Standard/PEM			
ICV/I.BLK	VP133575		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX045524.D	01 Apr 2025 16:15	JC/MD	Ok
2	VSTDIC001	VX045525.D	01 Apr 2025 17:06	JC/MD	Ok,M
3	VSTDIC005	VX045526.D	01 Apr 2025 17:29	JC/MD	Ok,M
4	VSTDIC020	VX045527.D	01 Apr 2025 17:52	JC/MD	Ok,M
5	VSTDIC050	VX045528.D	01 Apr 2025 18:15	JC/MD	Ok,M
6	VSTDIC100	VX045529.D	01 Apr 2025 18:38	JC/MD	Ok,M
7	VSTDIC150	VX045530.D	01 Apr 2025 19:02	JC/MD	Ok,M
8	IBLK	VX045531.D	01 Apr 2025 19:25	JC/MD	Ok
9	VSTDICV050	VX045532.D	01 Apr 2025 19:48	JC/MD	Ok,M
10	BFB	VX045533.D	02 Apr 2025 09:30	JC/MD	Ok
11	VSTDIC050	VX045534.D	02 Apr 2025 10:02	JC/MD	Ok,M
12	VX0402MBL01	VX045535.D	02 Apr 2025 10:30	JC/MD	Ok
13	VX0402WBL01	VX045536.D	02 Apr 2025 10:53	JC/MD	Ok
14	VX0402WBS01	VX045537.D	02 Apr 2025 11:16	JC/MD	Ok,M
15	VX0402WBSD01	VX045538.D	02 Apr 2025 11:44	JC/MD	Ok,M
16	Q1697-01	VX045539.D	02 Apr 2025 12:07	JC/MD	Dilution
17	Q1697-02	VX045540.D	02 Apr 2025 12:30	JC/MD	Dilution
18	IBLK	VX045541.D	02 Apr 2025 12:54	JC/MD	Ok
19	Q1697-01DL	VX045542.D	02 Apr 2025 13:21	JC/MD	Ok
20	Q1697-02DL	VX045543.D	02 Apr 2025 13:44	JC/MD	Ok
21	Q1697-03	VX045544.D	02 Apr 2025 14:07	JC/MD	Not Ok

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QCBatch ID # VX040225

Review By	John Carlone	Review On	4/2/2025 9:42:46 AM
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:39 PM
SubDirectory	VX040225	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133559,VP133563		
Initial Calibration Stds	VP133569,VP133570,VP133571,VP133572,VP133573,VP133574		
CCC	VP133564,VP133565		
Internal Standard/PEM			
ICV/I.BLK	VP133575		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q1697-04	VX045545.D	02 Apr 2025 14:31	JC/MD	Not Ok
23	Q1697-05	VX045546.D	02 Apr 2025 14:54	JC/MD	Not Ok
24	Q1697-07	VX045547.D	02 Apr 2025 15:18	JC/MD	Not Ok
25	IBLK	VX045548.D	02 Apr 2025 15:41	JC/MD	Ok
26	Q1697-03	VX045549.D	02 Apr 2025 16:04	JC/MD	Ok
27	Q1697-04	VX045550.D	02 Apr 2025 16:27	JC/MD	Ok
28	Q1697-05	VX045551.D	02 Apr 2025 16:51	JC/MD	Ok
29	Q1697-06	VX045552.D	02 Apr 2025 17:14	JC/MD	ReRun
30	Q1623-01	VX045553.D	02 Apr 2025 17:37	JC/MD	Ok
31	Q1623-02	VX045554.D	02 Apr 2025 18:01	JC/MD	Ok
32	Q1623-03	VX045555.D	02 Apr 2025 18:24	JC/MD	Ok
33	Q1623-04	VX045556.D	02 Apr 2025 18:47	JC/MD	Ok
34	VSTDCCC050	VX045557.D	02 Apr 2025 19:10	JC/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX042925

Review By	John Carlone	Review On	4/30/2025 9:47:21 AM
Supervise By	Semsettin Yesilyurt	Supervise On	4/30/2025 2:23:33 PM
SubDirectory	VX042925	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133784		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133785,VP133786		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX045971.D	29 Apr 2025 09:35	JC/MD	Ok
2	VSTDCCC050	VX045972.D	29 Apr 2025 11:06	JC/MD	Ok,M
3	VX0429MBL01	VX045973.D	29 Apr 2025 11:35	JC/MD	Ok
4	VX0429WBL01	VX045974.D	29 Apr 2025 11:58	JC/MD	Ok
5	VX0429WBS01	VX045975.D	29 Apr 2025 12:21	JC/MD	Ok,M
6	VX0429WBSD01	VX045976.D	29 Apr 2025 12:49	JC/MD	Ok,M
7	IBLK	VX045977.D	29 Apr 2025 13:12	JC/MD	Ok
8	Q1901-07	VX045978.D	29 Apr 2025 13:35	JC/MD	Ok
9	Q1901-06	VX045979.D	29 Apr 2025 13:58	JC/MD	Ok
10	Q1896-03	VX045980.D	29 Apr 2025 14:21	JC/MD	Ok
11	IBLK	VX045981.D	29 Apr 2025 14:45	JC/MD	Ok
12	IBLK	VX045982.D	29 Apr 2025 15:08	JC/MD	Ok
13	VSTDCCC050	VX045983.D	29 Apr 2025 16:15	JC/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QC Batch ID # VY042225

Review By	Semsettin Yesilyurt	Review On	4/23/2025 7:35:07 AM		
Supervise By	Maresh Dadoda	Supervise On	4/23/2025 1:28:52 PM		
SubDirectory	VY042225	HP Acquire Method	MSVOA_Y	HP Processing Method	82y042225s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP133718			
Initial Calibration Stds		VP133719,VP133720,VP133721,VP133722,VP133723,VP133724			
CCC					
Internal Standard/PEM		VP131783			
ICV/I.BLK		VP133725			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY021952.D	22 Apr 2025 11:33	SY/MD	Ok
2	VSTDIC005	VY021953.D	22 Apr 2025 13:39	SY/MD	Ok,M
3	VSTDIC010	VY021954.D	22 Apr 2025 14:44	SY/MD	Ok,M
4	VSTDIC020	VY021955.D	22 Apr 2025 15:07	SY/MD	Ok,M
5	VSTDIC050	VY021956.D	22 Apr 2025 15:29	SY/MD	Ok,M
6	VSTDIC100	VY021957.D	22 Apr 2025 15:52	SY/MD	Ok,M
7	VSTDIC150	VY021958.D	22 Apr 2025 16:15	SY/MD	Ok,M
8	IBLK	VY021959.D	22 Apr 2025 16:38	SY/MD	Ok
9	VSTDICV050	VY021960.D	22 Apr 2025 17:01	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QCBatch ID # VY043025

Review By	Semsettin Yesilyurt	Review On	5/5/2025 4:24:20 AM
Supervise By	Maresh Dadoda	Supervise On	5/5/2025 2:55:11 PM
SubDirectory	VY043025	HP Acquire Method	MSVOA_Y
		HP Processing Method	82y042225s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133787		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133788,VP133789 VP131783		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY022069.D	30 Apr 2025 08:50	SY/MD	Ok
2	VSTDCCC050	VY022070.D	30 Apr 2025 09:20	SY/MD	Ok,M
3	VY0430SBL01	VY022071.D	30 Apr 2025 09:57	SY/MD	Ok
4	VY0430SBS01	VY022072.D	30 Apr 2025 10:30	SY/MD	Ok,M
5	VY0430SBS01	VY022073.D	30 Apr 2025 10:52	SY/MD	Ok,M
6	Q1911-01	VY022074.D	30 Apr 2025 11:37	SY/MD	ReRun
7	Q1905-07	VY022075.D	30 Apr 2025 12:00	SY/MD	Ok
8	Q1910-01	VY022076.D	30 Apr 2025 12:24	SY/MD	Ok
9	Q1912-03	VY022077.D	30 Apr 2025 12:47	SY/MD	Ok
10	Q1912-07	VY022078.D	30 Apr 2025 13:11	SY/MD	Ok
11	Q1883-05	VY022079.D	30 Apr 2025 13:34	SY/MD	Ok
12	Q1883-07	VY022080.D	30 Apr 2025 13:58	SY/MD	Ok
13	Q1883-09	VY022081.D	30 Apr 2025 14:21	SY/MD	Ok
14	Q1883-11	VY022082.D	30 Apr 2025 14:44	SY/MD	ReRun
15	Q1883-13	VY022083.D	30 Apr 2025 15:11	SY/MD	Not Ok
16	Q1883-15	VY022084.D	30 Apr 2025 15:55	SY/MD	Ok
17	Q1907-01	VY022085.D	30 Apr 2025 16:18	SY/MD	Ok
18	Q1883-11	VY022086.D	30 Apr 2025 16:42	SY/MD	Ok
19	Q1901-02	VY022087.D	30 Apr 2025 17:05	SY/MD	Ok
20	Q1883-13	VY022088.D	30 Apr 2025 17:29	SY/MD	Ok
21	Q1901-04	VY022089.D	30 Apr 2025 17:52	SY/MD	Ok

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QCBatch ID # VY043025

Review By	Semsettin Yesilyurt	Review On	5/5/2025 4:24:20 AM		
Supervise By	Mahesh Dadoda	Supervise On	5/5/2025 2:55:11 PM		
SubDirectory	VY043025	HP Acquire Method	MSVOA_Y	HP Processing Method	82y042225s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133787			
CCC		VP133788,VP133789			
Internal Standard/PEM		VP131783			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	Q1901-05	VY022090.D	30 Apr 2025 18:16	SY/MD	Ok
23	Q1916-03	VY022091.D	30 Apr 2025 18:39	SY/MD	ReRun
24	Q1917-03	VY022092.D	30 Apr 2025 19:03	SY/MD	Ok
25	VSTDCCC050	VY022093.D	30 Apr 2025 19:25	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QCBatch ID # VY050125

Review By	Semsettin Yesilyurt	Review On	5/5/2025 2:51:02 PM		
Supervise By	Maresh Dadoda	Supervise On	5/5/2025 2:57:16 PM		
SubDirectory	VY050125	HP Acquire Method	MSVOA_Y	HP Processing Method	82y042225s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133796			
CCC		VP133797,VP133798			
Internal Standard/PEM		VP131783			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY022094.D	01 May 2025 08:20	SY/MD	Ok
2	VSTDCCC050	VY022095.D	01 May 2025 08:50	SY/MD	Ok,M
3	VY0501SBL01	VY022096.D	01 May 2025 10:11	SY/MD	Ok
4	VY0501SBS01	VY022097.D	01 May 2025 10:42	SY/MD	Ok,M
5	VY0501SBSD01	VY022098.D	01 May 2025 11:04	SY/MD	Ok,M
6	Q1920-01	VY022099.D	01 May 2025 11:54	SY/MD	Ok
7	Q1911-01RE	VY022100.D	01 May 2025 12:17	SY/MD	Confirms
8	Q1916-03	VY022101.D	01 May 2025 12:41	SY/MD	Ok
9	Q1923-01	VY022102.D	01 May 2025 13:04	SY/MD	ReRun
10	Q1922-03	VY022103.D	01 May 2025 13:28	SY/MD	Ok
11	Q1922-07	VY022104.D	01 May 2025 13:51	SY/MD	Ok
12	Q1901-01	VY022105.D	01 May 2025 14:15	SY/MD	Ok
13	Q1901-03	VY022106.D	01 May 2025 14:38	SY/MD	Ok
14	Q1914-01	VY022107.D	01 May 2025 15:02	SY/MD	Ok
15	Q1914-02	VY022108.D	01 May 2025 15:28	SY/MD	Ok
16	Q1914-03	VY022109.D	01 May 2025 16:04	SY/MD	Ok
17	Q1914-04	VY022110.D	01 May 2025 16:27	SY/MD	Ok
18	Q1914-05	VY022111.D	01 May 2025 16:49	SY/MD	Ok
19	Q1914-06	VY022112.D	01 May 2025 17:16	SY/MD	Dilution
20	Q1914-07	VY022113.D	01 May 2025 17:37	SY/MD	Dilution
21	Q1914-09	VY022114.D	01 May 2025 18:00	SY/MD	Dilution

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QCBatch ID # VY050125

Review By	Semsettin Yesilyurt	Review On	5/5/2025 2:51:02 PM		
Supervise By	Mahesh Dadoda	Supervise On	5/5/2025 2:57:16 PM		
SubDirectory	VY050125	HP Acquire Method	MSVOA_Y	HP Processing Method	82y042225s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133796			
CCC		VP133797,VP133798			
Internal Standard/PEM		VP131783			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	Q1914-10MS	VY022115.D	01 May 2025 18:23	SY/MD	Ok
23	Q1914-11MSD	VY022116.D	01 May 2025 18:46	SY/MD	Ok,M
24	VSTDCCC050	VY022117.D	01 May 2025 19:08	SY/MD	Ok,M

M : Manual Integration

A
B
C
D
E
F
G
H
I
J

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX040225

Review By	John Carlone	Review On	4/2/2025 9:42:46 AM
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:39 PM
SubDirectory	VX040225	HP Acquire Method	HP Processing Method 82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133559,VP133563
Initial Calibration Stds	VP133569,VP133570,VP133571,VP133572,VP133573,VP133574
CCC	VP133564,VP133565
Internal Standard/PEM	
ICV/I.BLK	VP133575
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045524.D	01 Apr 2025 16:15		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX045525.D	01 Apr 2025 17:06	%D failed for Comp. #53 in 01PPB	JC/MD	Ok,M
3	VSTDIC005	VSTDIC005	VX045526.D	01 Apr 2025 17:29		JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX045527.D	01 Apr 2025 17:52		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX045528.D	01 Apr 2025 18:15		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX045529.D	01 Apr 2025 18:38		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX045530.D	01 Apr 2025 19:02		JC/MD	Ok,M
8	IBLK	IBLK	VX045531.D	01 Apr 2025 19:25		JC/MD	Ok
9	VSTDICV050	ICVVX040225	VX045532.D	01 Apr 2025 19:48		JC/MD	Ok,M
10	BFB	BFB	VX045533.D	02 Apr 2025 09:30		JC/MD	Ok
11	VSTDIC050	VSTDIC050	VX045534.D	02 Apr 2025 10:02	pH#Lot#V12668	JC/MD	Ok,M
12	VX0402MBL01	VX0402MBL01	VX045535.D	02 Apr 2025 10:30		JC/MD	Ok
13	VX0402WBL01	VX0402WBL01	VX045536.D	02 Apr 2025 10:53		JC/MD	Ok
14	VX0402WBS01	VX0402WBS01	VX045537.D	02 Apr 2025 11:16		JC/MD	Ok,M
15	VX0402WBSD01	VX0402WBSD01	VX045538.D	02 Apr 2025 11:44		JC/MD	Ok,M
16	Q1697-01	MW-19B-72-040125	VX045539.D	02 Apr 2025 12:07	vial A pH<2 Need 200X	JC/MD	Dilution
17	Q1697-02	IW-01-55-040125	VX045540.D	02 Apr 2025 12:30	vial A pH<2 Need 10X	JC/MD	Dilution

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX040225

Review By	John Carlone	Review On	4/2/2025 9:42:46 AM
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:39 PM
SubDirectory	VX040225	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133559,VP133563		
Initial Calibration Stds	VP133569,VP133570,VP133571,VP133572,VP133573,VP133574		
CCC	VP133564,VP133565		
Internal Standard/PEM			
ICV/I.BLK	VP133575		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	IBLK	IBLK	VX045541.D	02 Apr 2025 12:54		JC/MD	Ok
19	Q1697-01DL	MW-19B-72-040125DL	VX045542.D	02 Apr 2025 13:21	vial B pH<2	JC/MD	Ok
20	Q1697-02DL	IW-01-55-040125DL	VX045543.D	02 Apr 2025 13:44	vial B pH<2	JC/MD	Ok
21	Q1697-03	IW-02-55-040125	VX045544.D	02 Apr 2025 14:07	Need lower dilution	JC/MD	Not Ok
22	Q1697-04	IW-02-55-040125-FD	VX045545.D	02 Apr 2025 14:31	Need lower dilution	JC/MD	Not Ok
23	Q1697-05	IW-03-55-040125	VX045546.D	02 Apr 2025 14:54	Need lower dilution	JC/MD	Not Ok
24	Q1697-07	MW-19B-72-040125	VX045547.D	02 Apr 2025 15:18	Not On Login	JC/MD	Not Ok
25	IBLK	IBLK	VX045548.D	02 Apr 2025 15:41		JC/MD	Ok
26	Q1697-03	IW-02-55-040125	VX045549.D	02 Apr 2025 16:04	vial A pH<2	JC/MD	Ok
27	Q1697-04	IW-02-55-040125-FD	VX045550.D	02 Apr 2025 16:27	vial A pH<2	JC/MD	Ok
28	Q1697-05	IW-03-55-040125	VX045551.D	02 Apr 2025 16:51	vial A pH<2	JC/MD	Ok
29	Q1697-06	TB-01-040125	VX045552.D	02 Apr 2025 17:14	vial A pH<2 TB;Hit of comp.#44	JC/MD	ReRun
30	Q1623-01	Storage-Blank-SOIL-RE	VX045553.D	02 Apr 2025 17:37	vial A pH<2	JC/MD	Ok
31	Q1623-02	Storage-Blank-WATER	VX045554.D	02 Apr 2025 18:01	vial A pH<2	JC/MD	Ok
32	Q1623-03	Storage-Blank-WATER	VX045555.D	02 Apr 2025 18:24	vial A pH<2	JC/MD	Ok
33	Q1623-04	Storage-Blank-SAMPLE	VX045556.D	02 Apr 2025 18:47	vial A pH<2	JC/MD	Ok
34	VSTDCCC050	VSTDCCC050EC	VX045557.D	02 Apr 2025 19:10		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX042925

Review By	John Carlone	Review On	4/30/2025 9:47:21 AM
Supervise By	Semsettin Yesilyurt	Supervise On	4/30/2025 2:23:33 PM
SubDirectory	VX042925	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133784		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133785,VP133786		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045971.D	29 Apr 2025 09:35		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX045972.D	29 Apr 2025 11:06	pH#Lot#V12668	JC/MD	Ok,M
3	VX0429MBL01	VX0429MBL01	VX045973.D	29 Apr 2025 11:35		JC/MD	Ok
4	VX0429WBL01	VX0429WBL01	VX045974.D	29 Apr 2025 11:58		JC/MD	Ok
5	VX0429WBS01	VX0429WBS01	VX045975.D	29 Apr 2025 12:21		JC/MD	Ok,M
6	VX0429WBSD01	VX0429WBSD01	VX045976.D	29 Apr 2025 12:49		JC/MD	Ok,M
7	IBLK	IBLK	VX045977.D	29 Apr 2025 13:12		JC/MD	Ok
8	Q1901-07	TB04262025	VX045978.D	29 Apr 2025 13:35	vial A pH<2 TB	JC/MD	Ok
9	Q1901-06	FB04262025	VX045979.D	29 Apr 2025 13:58	vial A pH<2 FB	JC/MD	Ok
10	Q1896-03	295-BERGEN-FRAC	VX045980.D	29 Apr 2025 14:21	vial A pH<2	JC/MD	Ok
11	IBLK	IBLK	VX045981.D	29 Apr 2025 14:45		JC/MD	Ok
12	IBLK	IBLK	VX045982.D	29 Apr 2025 15:08		JC/MD	Ok
13	VSTDCCC050	VSTDCCC050EC	VX045983.D	29 Apr 2025 16:15		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY042225

Review By	Semsettin Yesilyurt	Review On	4/23/2025 7:35:07 AM
Supervise By	Maresh Dadoda	Supervise On	4/23/2025 1:28:52 PM
SubDirectory	VY042225	HP Acquire Method	MSVOA_Y HP Processing Method 82y042225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133718 VP133719,VP133720,VP133721,VP133722,VP133723,VP133724
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131783 VP133725

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021952.D	22 Apr 2025 11:33		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY021953.D	22 Apr 2025 13:39		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY021954.D	22 Apr 2025 14:44	Com.#16 is on Linear Regression	SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY021955.D	22 Apr 2025 15:07		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY021956.D	22 Apr 2025 15:29		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY021957.D	22 Apr 2025 15:52		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY021958.D	22 Apr 2025 16:15		SY/MD	Ok,M
8	IBLK	IBLK	VY021959.D	22 Apr 2025 16:38		SY/MD	Ok
9	VSTDICV050	ICVVY042225	VY021960.D	22 Apr 2025 17:01		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY043025

Review By	Semsettin Yesilyurt	Review On	5/5/2025 4:24:20 AM
Supervise By	Maresh Dadoda	Supervise On	5/5/2025 2:55:11 PM
SubDirectory	VY043025	HP Acquire Method	MSVOA_Y HP Processing Method 82y042225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133787
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133788,VP133789 VP131783

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022069.D	30 Apr 2025 08:50		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022070.D	30 Apr 2025 09:20		SY/MD	Ok,M
3	VY0430SBL01	VY0430SBL01	VY022071.D	30 Apr 2025 09:57		SY/MD	Ok
4	VY0430SBS01	VY0430SBS01	VY022072.D	30 Apr 2025 10:30		SY/MD	Ok,M
5	VY0430SBSD01	VY0430SBSD01	VY022073.D	30 Apr 2025 10:52		SY/MD	Ok,M
6	Q1911-01	EO-03-04292025	VY022074.D	30 Apr 2025 11:37	vial-A Internal Standard Fail	SY/MD	ReRun
7	Q1905-07	MH-H-VOC	VY022075.D	30 Apr 2025 12:00	vial-A	SY/MD	Ok
8	Q1910-01	MOO-25-0123-27	VY022076.D	30 Apr 2025 12:24	vial-A	SY/MD	Ok
9	Q1912-03	MH-E-VOC	VY022077.D	30 Apr 2025 12:47	vial-A	SY/MD	Ok
10	Q1912-07	MH-F-VOC	VY022078.D	30 Apr 2025 13:11	vial-A	SY/MD	Ok
11	Q1883-05	OU4-PCS-TC-29-04232	VY022079.D	30 Apr 2025 13:34	vial-B	SY/MD	Ok
12	Q1883-07	OU4-PCS-TC-30-04232	VY022080.D	30 Apr 2025 13:58	vial-B	SY/MD	Ok
13	Q1883-09	OU4-PCS-TC-31-04232	VY022081.D	30 Apr 2025 14:21	vial-B	SY/MD	Ok
14	Q1883-11	OU4-PCS-TC-32-04232	VY022082.D	30 Apr 2025 14:44	Internal Standard Fail; Surrogate fail	SY/MD	ReRun
15	Q1883-13	OU4-VSL-18-042325	VY022083.D	30 Apr 2025 15:11	NOT PURGE	SY/MD	Not Ok
16	Q1883-15	OU4-VSL-19-042325	VY022084.D	30 Apr 2025 15:55	vial-B	SY/MD	Ok
17	Q1907-01	CO-8R-WC	VY022085.D	30 Apr 2025 16:18	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY043025

Review By	Semsettin Yesilyurt	Review On	5/5/2025 4:24:20 AM		
Supervise By	Mahesh Dadoda	Supervise On	5/5/2025 2:55:11 PM		
SubDirectory	VY043025	HP Acquire Method	MSVOA_Y	HP Processing Method	82y042225s.m
STD. NAME	STD REF.#				
Tune/Reschk Initial Calibration Stds	VP133787				
CCC	VP133788,VP133789				
Internal Standard/PEM	VP131783				
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q1883-11	OU4-PCS-TC-32-04232	VY022086.D	30 Apr 2025 16:42	vial-B	SY/MD	Ok
19	Q1901-02	B-167-SB01	VY022087.D	30 Apr 2025 17:05	vial-A	SY/MD	Ok
20	Q1883-13	OU4-VSL-18-042325	VY022088.D	30 Apr 2025 17:29	vial-B	SY/MD	Ok
21	Q1901-04	B-167-SB02	VY022089.D	30 Apr 2025 17:52	vial-A	SY/MD	Ok
22	Q1901-05	B-170-SB02	VY022090.D	30 Apr 2025 18:16	vial-A	SY/MD	Ok
23	Q1916-03	WC-12-VOC	VY022091.D	30 Apr 2025 18:39	vial-A Internal Standard Fail	SY/MD	ReRun
24	Q1917-03	MH-JJ-VOC	VY022092.D	30 Apr 2025 19:03	vial-A	SY/MD	Ok
25	VSTDCCC050	VSTDCCC050EC	VY022093.D	30 Apr 2025 19:25		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY050125

Review By	Semsettin Yesilyurt	Review On	5/5/2025 2:51:02 PM		
Supervise By	Mahesh Dadoda	Supervise On	5/5/2025 2:57:16 PM		
SubDirectory	VY050125	HP Acquire Method	MSVOA_Y	HP Processing Method	82y042225s.m
STD. NAME	STD REF.#				
Tune/Reschk Initial Calibration Stds	VP133796				
CCC	VP133797,VP133798				
Internal Standard/PEM	VP131783				
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022094.D	01 May 2025 08:20		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022095.D	01 May 2025 08:50		SY/MD	Ok,M
3	VY0501SBL01	VY0501SBL01	VY022096.D	01 May 2025 10:11		SY/MD	Ok
4	VY0501SBS01	VY0501SBS01	VY022097.D	01 May 2025 10:42		SY/MD	Ok,M
5	VY0501SBSD01	VY0501SBSD01	VY022098.D	01 May 2025 11:04		SY/MD	Ok,M
6	Q1920-01	291	VY022099.D	01 May 2025 11:54	vial-A	SY/MD	Ok
7	Q1911-01RE	EO-03-04292025RE	VY022100.D	01 May 2025 12:17	vial-B Internal Standard Fail	SY/MD	Confirms
8	Q1916-03	WC-12-VOC	VY022101.D	01 May 2025 12:41	vial-B	SY/MD	Ok
9	Q1923-01	TR-05-04302025	VY022102.D	01 May 2025 13:04	vial-A Internal Standard Fail	SY/MD	ReRun
10	Q1922-03	MH-R-VOC	VY022103.D	01 May 2025 13:28	vial-A	SY/MD	Ok
11	Q1922-07	MH-2-VOC	VY022104.D	01 May 2025 13:51	vial-A	SY/MD	Ok
12	Q1901-01	B-170-SB00	VY022105.D	01 May 2025 14:15	vial-A	SY/MD	Ok
13	Q1901-03	B-170-SB01	VY022106.D	01 May 2025 14:38	vial-A	SY/MD	Ok
14	Q1914-01	SS-1	VY022107.D	01 May 2025 15:02	vial-A	SY/MD	Ok
15	Q1914-02	SS-91	VY022108.D	01 May 2025 15:28	vial-A	SY/MD	Ok
16	Q1914-03	SS-2	VY022109.D	01 May 2025 16:04	vial-A	SY/MD	Ok
17	Q1914-04	SS-3	VY022110.D	01 May 2025 16:27	vial-A	SY/MD	Ok
18	Q1914-05	SS-4	VY022111.D	01 May 2025 16:49	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY050125

Review By	Semsettin Yesilyurt	Review On	5/5/2025 2:51:02 PM		
Supervise By	Mahesh Dadoda	Supervise On	5/5/2025 2:57:16 PM		
SubDirectory	VY050125	HP Acquire Method	MSVOA_Y	HP Processing Method	82y042225s.m
STD. NAME	STD REF.#				
Tune/Reschk Initial Calibration Stds	VP133796				
CCC	VP133797,VP133798				
Internal Standard/PEM	VP131783				
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q1914-06	SS-5	VY022112.D	01 May 2025 17:16	vial-A Need MeOH	SY/MD	Dilution
20	Q1914-07	SS-6	VY022113.D	01 May 2025 17:37	vial-A Internal Standard Fail;Need MeOH	SY/MD	Dilution
21	Q1914-09	SS-8	VY022114.D	01 May 2025 18:00	vial-A Internal Standard Fail; Surrogate Fail;Need MeOH	SY/MD	Dilution
22	Q1914-10MS	SS-8MS	VY022115.D	01 May 2025 18:23	vial-A Surrogate fail	SY/MD	Ok
23	Q1914-11MSD	SS-8MSD	VY022116.D	01 May 2025 18:46	vial-A Surrogate fail	SY/MD	Ok,M
24	VSTDCCC050	VSTDCCC050EC	VY022117.D	01 May 2025 19:08		SY/MD	Ok,M

M : Manual Integration

LAB CHRONICLE

OrderID:	Q1901	OrderDate:	4/28/2025 10:24:00 AM
Client:	Portal Partners Tri-Venture	Project:	Amtrak Sawtooth Bridges 2025
Contact:	Joseph Krupansky	Location:	L51,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1901-01	B-170-SB00	SOIL	VOC-TCLVOA-10	8260D	04/26/25		04/30/25	04/28/25
			VOC-TCLVOA-10	8260D			05/01/25	
Q1901-02	B-167-SB01	SOIL	VOC-TCLVOA-10	8260D	04/26/25		04/30/25	04/28/25
Q1901-03	B-170-SB01	SOIL	VOC-TCLVOA-10	8260D	04/26/25		04/30/25	04/28/25
			VOC-TCLVOA-10	8260D			05/01/25	
Q1901-04	B-167-SB02	SOIL	VOC-TCLVOA-10	8260D	04/26/25		04/30/25	04/28/25
Q1901-05	B-170-SB02	SOIL	VOC-TCLVOA-10	8260D	04/26/25		04/30/25	04/28/25
Q1901-06	FB04262025	Water	VOC-TCLVOA-10	8260-Low	04/26/25		04/29/25	04/28/25
Q1901-07	TB04262025	Water	VOC-TCLVOA-10	8260-Low	04/26/25		04/29/25	04/28/25
Q1901-08	B-167-SB01	TCLP	TCLP VOA	8260D	04/26/25		04/30/25	04/28/25
Q1901-09	B-170-SB01	TCLP	TCLP VOA	8260D	04/26/25		04/30/25	04/28/25
Q1901-10	B-167-SB02	TCLP	TCLP VOA	8260D	04/26/25		04/30/25	04/28/25
Q1901-11	B-170-SB02	TCLP			04/26/25			04/28/25

LAB CHRONICLE

TCLP VOA	8260D	04/30/25
----------	-------	----------

A

B

C

D

E

F

G

H

I

J