

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901
 Lab File ID: VX045972.D Date Analyzed: 04/29/2025
 Instrument ID: MSVOA_X Time Analyzed: 11:06
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	79737	5.54	136469	6.76	117675	10.05
UPPER LIMIT	159474	6.044	272938	7.257	235350	10.549
LOWER LIMIT	39868.5	5.044	68234.5	6.257	58837.5	9.549
EPA SAMPLE NO.						
FB04262025	64509	5.54	127954	6.76	115874	10.05
TB04262025	67274	5.54	131446	6.76	123412	10.05
VX0429WBL01	63902	5.55	127024	6.76	116648	10.05
VX0429WBS01	80632	5.55	140487	6.76	122568	10.05
VX0429WBSD01	77095	5.54	135079	6.76	118384	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.