

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901
 Lab File ID: VY022095.D Date Analyzed: 05/01/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:50
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	300031	7.71	433614	8.62	410078	11.42
UPPER LIMIT	600062	8.207	867228	9.116	820156	11.92
LOWER LIMIT	150016	7.207	216807	8.116	205039	10.92
EPA SAMPLE NO.						
B-170-SB00	310481	7.71	585270	8.62	519369	11.41
B-170-SB01	307701	7.71	584162	8.62	521346	11.42
VY0501SBL01	317996	7.71	585580	8.62	508953	11.42
VY0501SBS01	294485	7.71	452802	8.62	414235	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.