

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901 SDG NO.: Q1901
 Lab File ID: VX045972.D Date Analyzed: 04/29/2025
 Instrument ID: MSVOA_X Time Analyzed: 11:06
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	59672	12.018				
UPPER LIMIT	119344	12.518				
LOWER LIMIT	29836	11.518				
EPA SAMPLE NO.						
FB04262025	48370	12.02				
TB04262025	53777	12.02				
VX0429WBL01	47876	12.02				
VX0429WBS01	59236	12.02				
VX0429WBSD01	57470	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.