

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder **SDG No.:** Q1903
Contract: POWE02 **Lab Code:** CHEM **Case No.:** Q1903 **SAS No.:** Q1903
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV109	Mercury	4.20	4.0	105	90 - 110	CV	04/30/2025	13:35	LB135611

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder **SDG No.:** Q1903
Contract: POWE02 **Lab Code:** CHEM **Case No.:** Q1903 **SAS No.:** Q1903
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV52	Mercury	5.02	5.0	100	90 - 110	CV	04/30/2025	13:40	LB135611
CCV53	Mercury	4.99	5.0	100	90 - 110	CV	04/30/2025	14:10	LB135611
CCV54	Mercury	5.02	5.0	100	90 - 110	CV	04/30/2025	14:38	LB135611
CCV55	Mercury	5.06	5.0	101	90 - 110	CV	04/30/2025	14:56	LB135611

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder SDG No.: Q1903
Contract: POWE02 Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2320	2500	93	90 - 110	P	04/30/2025	16:18	LB135614
	Antimony	973	1000	97	90 - 110	P	04/30/2025	16:18	LB135614
	Arsenic	986	1000	99	90 - 110	P	04/30/2025	16:18	LB135614
	Barium	537	520	103	90 - 110	P	04/30/2025	16:18	LB135614
	Beryllium	473	510	93	90 - 110	P	04/30/2025	16:18	LB135614
	Boron	2570	2500	103	90 - 110	P	04/30/2025	16:18	LB135614
	Cadmium	500	510	98	90 - 110	P	04/30/2025	16:18	LB135614
	Chromium	510	520	98	90 - 110	P	04/30/2025	16:18	LB135614
	Cobalt	487	520	94	90 - 110	P	04/30/2025	16:18	LB135614
	Copper	495	510	97	90 - 110	P	04/30/2025	16:18	LB135614
	Iron	9190	10000	92	90 - 110	P	04/30/2025	16:18	LB135614
	Lead	987	1000	99	90 - 110	P	04/30/2025	16:18	LB135614
	Manganese	468	520	90	90 - 110	P	04/30/2025	16:18	LB135614
	Molybdenum	2500	2500	100	90 - 110	P	04/30/2025	16:18	LB135614
	Nickel	526	530	99	90 - 110	P	04/30/2025	16:18	LB135614
	Selenium	1010	1000	101	90 - 110	P	04/30/2025	16:18	LB135614
	Silver	262	250	105	90 - 110	P	04/30/2025	16:18	LB135614
	Thallium	1080	1000	108	90 - 110	P	04/30/2025	16:18	LB135614
	Vanadium	455	500	91	90 - 110	P	04/30/2025	16:18	LB135614
	Zinc	969	1000	97	90 - 110	P	04/30/2025	16:18	LB135614

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder SDG No.: Q1903
Contract: POWE02 Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	95.4	100	95	80 - 120	P	04/30/2025	16:28	LB135614
	Antimony	55.5	50.0	111	80 - 120	P	04/30/2025	16:28	LB135614
	Arsenic	22.6	20.0	113	80 - 120	P	04/30/2025	16:28	LB135614
	Barium	93.5	100	94	80 - 120	P	04/30/2025	16:28	LB135614
	Beryllium	6.43	6.0	107	80 - 120	P	04/30/2025	16:28	LB135614
	Boron	105	100	105	80 - 120	P	04/30/2025	16:28	LB135614
	Cadmium	6.58	6.0	110	80 - 120	P	04/30/2025	16:28	LB135614
	Chromium	9.76	10.0	98	80 - 120	P	04/30/2025	16:28	LB135614
	Cobalt	30.9	30.0	103	80 - 120	P	04/30/2025	16:28	LB135614
	Copper	23.3	20.0	116	80 - 120	P	04/30/2025	16:28	LB135614
	Iron	92.8	100	93	80 - 120	P	04/30/2025	16:28	LB135614
	Lead	12.4	12.0	103	80 - 120	P	04/30/2025	16:28	LB135614
	Manganese	21.3	20.0	106	80 - 120	P	04/30/2025	16:28	LB135614
	Molybdenum	220	200	110	80 - 120	P	04/30/2025	16:28	LB135614
	Nickel	41.8	40.0	104	80 - 120	P	04/30/2025	16:28	LB135614
	Selenium	16.3	20.0	82	80 - 120	P	04/30/2025	16:28	LB135614
	Silver	10.7	10.0	107	80 - 120	P	04/30/2025	16:28	LB135614
	Thallium	45.6	40.0	114	80 - 120	P	04/30/2025	16:28	LB135614
	Vanadium	40.2	40.0	100	80 - 120	P	04/30/2025	16:28	LB135614
	Zinc	42.5	40.0	106	80 - 120	P	04/30/2025	16:28	LB135614

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder SDG No.: Q1903
Contract: POWE02 Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9840	10000	98	90 - 110	P	04/30/2025	17:27	LB135614
	Antimony	4970	5000	100	90 - 110	P	04/30/2025	17:27	LB135614
	Arsenic	5000	5000	100	90 - 110	P	04/30/2025	17:27	LB135614
	Barium	9810	10000	98	90 - 110	P	04/30/2025	17:27	LB135614
	Beryllium	239	250	96	90 - 110	P	04/30/2025	17:27	LB135614
	Boron	4700	5000	94	90 - 110	P	04/30/2025	17:27	LB135614
	Cadmium	2530	2500	101	90 - 110	P	04/30/2025	17:27	LB135614
	Chromium	1020	1000	102	90 - 110	P	04/30/2025	17:27	LB135614
	Cobalt	2510	2500	100	90 - 110	P	04/30/2025	17:27	LB135614
	Copper	1250	1250	100	90 - 110	P	04/30/2025	17:27	LB135614
	Iron	5130	5000	103	90 - 110	P	04/30/2025	17:27	LB135614
	Lead	5060	5000	101	90 - 110	P	04/30/2025	17:27	LB135614
	Manganese	2430	2500	97	90 - 110	P	04/30/2025	17:27	LB135614
	Molybdenum	5050	5000	101	90 - 110	P	04/30/2025	17:27	LB135614
	Nickel	2520	2500	101	90 - 110	P	04/30/2025	17:27	LB135614
	Selenium	5010	5000	100	90 - 110	P	04/30/2025	17:27	LB135614
	Silver	1250	1250	100	90 - 110	P	04/30/2025	17:27	LB135614
	Thallium	5240	5000	105	90 - 110	P	04/30/2025	17:27	LB135614
	Vanadium	2450	2500	98	90 - 110	P	04/30/2025	17:27	LB135614
	Zinc	2510	2500	100	90 - 110	P	04/30/2025	17:27	LB135614
CCV02	Aluminum	10000	10000	100	90 - 110	P	04/30/2025	18:30	LB135614
	Antimony	4770	5000	95	90 - 110	P	04/30/2025	18:30	LB135614
	Arsenic	4820	5000	96	90 - 110	P	04/30/2025	18:30	LB135614
	Barium	9940	10000	99	90 - 110	P	04/30/2025	18:30	LB135614
	Beryllium	261	250	104	90 - 110	P	04/30/2025	18:30	LB135614
	Boron	5070	5000	102	90 - 110	P	04/30/2025	18:30	LB135614
	Cadmium	2560	2500	102	90 - 110	P	04/30/2025	18:30	LB135614
	Chromium	1040	1000	104	90 - 110	P	04/30/2025	18:30	LB135614
	Cobalt	2510	2500	101	90 - 110	P	04/30/2025	18:30	LB135614
	Copper	1220	1250	97	90 - 110	P	04/30/2025	18:30	LB135614
	Iron	5020	5000	100	90 - 110	P	04/30/2025	18:30	LB135614
	Lead	5080	5000	102	90 - 110	P	04/30/2025	18:30	LB135614
	Manganese	2520	2500	101	90 - 110	P	04/30/2025	18:30	LB135614
	Molybdenum	5000	5000	100	90 - 110	P	04/30/2025	18:30	LB135614

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder SDG No.: Q1903
Contract: POWE02 Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Nickel	2520	2500	101	90 - 110	P	04/30/2025	18:30	LB135614
	Selenium	4770	5000	95	90 - 110	P	04/30/2025	18:30	LB135614
	Silver	1250	1250	100	90 - 110	P	04/30/2025	18:30	LB135614
	Thallium	5180	5000	104	90 - 110	P	04/30/2025	18:30	LB135614
	Vanadium	2540	2500	102	90 - 110	P	04/30/2025	18:30	LB135614
	Zinc	2490	2500	100	90 - 110	P	04/30/2025	18:30	LB135614
CCV03	Aluminum	10100	10000	101	90 - 110	P	04/30/2025	19:37	LB135614
	Antimony	5030	5000	100	90 - 110	P	04/30/2025	19:37	LB135614
	Arsenic	5060	5000	101	90 - 110	P	04/30/2025	19:37	LB135614
	Barium	9770	10000	98	90 - 110	P	04/30/2025	19:37	LB135614
	Beryllium	264	250	106	90 - 110	P	04/30/2025	19:37	LB135614
	Boron	5210	5000	104	90 - 110	P	04/30/2025	19:37	LB135614
	Cadmium	2570	2500	103	90 - 110	P	04/30/2025	19:37	LB135614
	Chromium	1040	1000	104	90 - 110	P	04/30/2025	19:37	LB135614
	Cobalt	2540	2500	102	90 - 110	P	04/30/2025	19:37	LB135614
	Copper	1260	1250	101	90 - 110	P	04/30/2025	19:37	LB135614
	Iron	5170	5000	103	90 - 110	P	04/30/2025	19:37	LB135614
	Lead	5110	5000	102	90 - 110	P	04/30/2025	19:37	LB135614
	Manganese	2480	2500	99	90 - 110	P	04/30/2025	19:37	LB135614
	Molybdenum	5100	5000	102	90 - 110	P	04/30/2025	19:37	LB135614
	Nickel	2550	2500	102	90 - 110	P	04/30/2025	19:37	LB135614
	Selenium	5080	5000	102	90 - 110	P	04/30/2025	19:37	LB135614
	Silver	1270	1250	102	90 - 110	P	04/30/2025	19:37	LB135614
	Thallium	5240	5000	105	90 - 110	P	04/30/2025	19:37	LB135614
	Vanadium	2510	2500	100	90 - 110	P	04/30/2025	19:37	LB135614
	Zinc	2540	2500	102	90 - 110	P	04/30/2025	19:37	LB135614
CCV04	Aluminum	9940	10000	99	90 - 110	P	04/30/2025	20:23	LB135614
	Antimony	4980	5000	100	90 - 110	P	04/30/2025	20:23	LB135614
	Arsenic	5000	5000	100	90 - 110	P	04/30/2025	20:23	LB135614
	Barium	9720	10000	97	90 - 110	P	04/30/2025	20:23	LB135614
	Beryllium	251	250	100	90 - 110	P	04/30/2025	20:23	LB135614
	Boron	4890	5000	98	90 - 110	P	04/30/2025	20:23	LB135614
	Cadmium	2550	2500	102	90 - 110	P	04/30/2025	20:23	LB135614
	Chromium	1050	1000	105	90 - 110	P	04/30/2025	20:23	LB135614

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder **SDG No.:** Q1903
Contract: POWE02 **Lab Code:** CHEM **Case No.:** Q1903 **SAS No.:** Q1903
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Cobalt	2520	2500	101	90 - 110	P	04/30/2025	20:23	LB135614
	Copper	1250	1250	100	90 - 110	P	04/30/2025	20:23	LB135614
	Iron	5490	5000	110	90 - 110	P	04/30/2025	20:23	LB135614
	Lead	5100	5000	102	90 - 110	P	04/30/2025	20:23	LB135614
	Manganese	2430	2500	97	90 - 110	P	04/30/2025	20:23	LB135614
	Molybdenum	5060	5000	101	90 - 110	P	04/30/2025	20:23	LB135614
	Nickel	2530	2500	101	90 - 110	P	04/30/2025	20:23	LB135614
	Selenium	4970	5000	100	90 - 110	P	04/30/2025	20:23	LB135614
	Silver	1270	1250	101	90 - 110	P	04/30/2025	20:23	LB135614
	Thallium	5140	5000	103	90 - 110	P	04/30/2025	20:23	LB135614
	Vanadium	2480	2500	99	90 - 110	P	04/30/2025	20:23	LB135614
	Zinc	2520	2500	101	90 - 110	P	04/30/2025	20:23	LB135614
CCV05	Aluminum	9810	10000	98	90 - 110	P	04/30/2025	21:08	LB135614
	Antimony	4960	5000	99	90 - 110	P	04/30/2025	21:08	LB135614
	Arsenic	4970	5000	99	90 - 110	P	04/30/2025	21:08	LB135614
	Barium	9680	10000	97	90 - 110	P	04/30/2025	21:08	LB135614
	Beryllium	252	250	101	90 - 110	P	04/30/2025	21:08	LB135614
	Boron	4890	5000	98	90 - 110	P	04/30/2025	21:08	LB135614
	Cadmium	2530	2500	101	90 - 110	P	04/30/2025	21:08	LB135614
	Chromium	1050	1000	105	90 - 110	P	04/30/2025	21:08	LB135614
	Cobalt	2510	2500	100	90 - 110	P	04/30/2025	21:08	LB135614
	Copper	1240	1250	99	90 - 110	P	04/30/2025	21:08	LB135614
	Iron	5060	5000	101	90 - 110	P	04/30/2025	21:08	LB135614
	Lead	5070	5000	101	90 - 110	P	04/30/2025	21:08	LB135614
	Manganese	2440	2500	98	90 - 110	P	04/30/2025	21:08	LB135614
	Molybdenum	5040	5000	101	90 - 110	P	04/30/2025	21:08	LB135614
	Nickel	2510	2500	100	90 - 110	P	04/30/2025	21:08	LB135614
	Selenium	4940	5000	99	90 - 110	P	04/30/2025	21:08	LB135614
	Silver	1260	1250	100	90 - 110	P	04/30/2025	21:08	LB135614
	Thallium	5160	5000	103	90 - 110	P	04/30/2025	21:08	LB135614
	Vanadium	2480	2500	99	90 - 110	P	04/30/2025	21:08	LB135614
	Zinc	2500	2500	100	90 - 110	P	04/30/2025	21:08	LB135614
CCV06	Aluminum	9680	10000	97	90 - 110	P	04/30/2025	21:54	LB135614
	Antimony	5010	5000	100	90 - 110	P	04/30/2025	21:54	LB135614

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder **SDG No.:** Q1903
Contract: POWE02 **Lab Code:** CHEM **Case No.:** Q1903 **SAS No.:** Q1903
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV06	Arsenic	5030	5000	100	90 - 110	P	04/30/2025	21:54	LB135614
	Barium	9680	10000	97	90 - 110	P	04/30/2025	21:54	LB135614
	Beryllium	238	250	95	90 - 110	P	04/30/2025	21:54	LB135614
	Boron	4700	5000	94	90 - 110	P	04/30/2025	21:54	LB135614
	Cadmium	2500	2500	100	90 - 110	P	04/30/2025	21:54	LB135614
	Chromium	1010	1000	101	90 - 110	P	04/30/2025	21:54	LB135614
	Cobalt	2490	2500	99	90 - 110	P	04/30/2025	21:54	LB135614
	Copper	1250	1250	100	90 - 110	P	04/30/2025	21:54	LB135614
	Iron	5070	5000	101	90 - 110	P	04/30/2025	21:54	LB135614
	Lead	5010	5000	100	90 - 110	P	04/30/2025	21:54	LB135614
	Manganese	2400	2500	96	90 - 110	P	04/30/2025	21:54	LB135614
	Molybdenum	5000	5000	100	90 - 110	P	04/30/2025	21:54	LB135614
	Nickel	2490	2500	100	90 - 110	P	04/30/2025	21:54	LB135614
	Selenium	5060	5000	101	90 - 110	P	04/30/2025	21:54	LB135614
	Silver	1230	1250	98	90 - 110	P	04/30/2025	21:54	LB135614
	Thallium	5110	5000	102	90 - 110	P	04/30/2025	21:54	LB135614
	Vanadium	2430	2500	97	90 - 110	P	04/30/2025	21:54	LB135614
	Zinc	2450	2500	98	90 - 110	P	04/30/2025	21:54	LB135614
CCV07	Aluminum	9820	10000	98	90 - 110	P	04/30/2025	22:40	LB135614
	Antimony	5220	5000	104	90 - 110	P	04/30/2025	22:40	LB135614
	Arsenic	5170	5000	103	90 - 110	P	04/30/2025	22:40	LB135614
	Barium	9890	10000	99	90 - 110	P	04/30/2025	22:40	LB135614
	Beryllium	238	250	95	90 - 110	P	04/30/2025	22:40	LB135614
	Boron	4770	5000	95	90 - 110	P	04/30/2025	22:40	LB135614
	Cadmium	2440	2500	98	90 - 110	P	04/30/2025	22:40	LB135614
	Chromium	995	1000	100	90 - 110	P	04/30/2025	22:40	LB135614
	Cobalt	2450	2500	98	90 - 110	P	04/30/2025	22:40	LB135614
	Copper	1290	1250	103	90 - 110	P	04/30/2025	22:40	LB135614
	Iron	5110	5000	102	90 - 110	P	04/30/2025	22:40	LB135614
	Lead	4900	5000	98	90 - 110	P	04/30/2025	22:40	LB135614
	Manganese	2400	2500	96	90 - 110	P	04/30/2025	22:40	LB135614
	Molybdenum	5090	5000	102	90 - 110	P	04/30/2025	22:40	LB135614
	Nickel	2460	2500	98	90 - 110	P	04/30/2025	22:40	LB135614
	Selenium	5260	5000	105	90 - 110	P	04/30/2025	22:40	LB135614

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder SDG No.: Q1903
Contract: POWE02 Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Silver	1240	1250	99	90 - 110	P	04/30/2025	22:40	LB135614
	Thallium	5030	5000	101	90 - 110	P	04/30/2025	22:40	LB135614
	Vanadium	2440	2500	98	90 - 110	P	04/30/2025	22:40	LB135614
	Zinc	2510	2500	100	90 - 110	P	04/30/2025	22:40	LB135614
CCV08	Aluminum	9980	10000	100	90 - 110	P	04/30/2025	23:05	LB135614
	Antimony	5150	5000	103	90 - 110	P	04/30/2025	23:05	LB135614
	Arsenic	5120	5000	102	90 - 110	P	04/30/2025	23:05	LB135614
	Barium	9730	10000	97	90 - 110	P	04/30/2025	23:05	LB135614
	Beryllium	246	250	98	90 - 110	P	04/30/2025	23:05	LB135614
	Boron	4920	5000	98	90 - 110	P	04/30/2025	23:05	LB135614
	Cadmium	2440	2500	98	90 - 110	P	04/30/2025	23:05	LB135614
	Chromium	989	1000	99	90 - 110	P	04/30/2025	23:05	LB135614
	Cobalt	2440	2500	98	90 - 110	P	04/30/2025	23:05	LB135614
	Copper	1280	1250	103	90 - 110	P	04/30/2025	23:05	LB135614
	Iron	5100	5000	102	90 - 110	P	04/30/2025	23:05	LB135614
	Lead	4880	5000	98	90 - 110	P	04/30/2025	23:05	LB135614
	Manganese	2380	2500	95	90 - 110	P	04/30/2025	23:05	LB135614
	Molybdenum	5060	5000	101	90 - 110	P	04/30/2025	23:05	LB135614
	Nickel	2450	2500	98	90 - 110	P	04/30/2025	23:05	LB135614
	Selenium	5200	5000	104	90 - 110	P	04/30/2025	23:05	LB135614
	Silver	1240	1250	100	90 - 110	P	04/30/2025	23:05	LB135614
	Thallium	5020	5000	100	90 - 110	P	04/30/2025	23:05	LB135614
	Vanadium	2420	2500	97	90 - 110	P	04/30/2025	23:05	LB135614
	Zinc	2510	2500	100	90 - 110	P	04/30/2025	23:05	LB135614

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder SDG No.: Q1903
Contract: POWE02 Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2400	2500	96	90 - 110	P	05/06/2025	14:47	LB135689
	Antimony	972	1000	97	90 - 110	P	05/06/2025	14:47	LB135689
	Arsenic	976	1000	98	90 - 110	P	05/06/2025	14:47	LB135689
	Barium	531	520	102	90 - 110	P	05/06/2025	14:47	LB135689
	Beryllium	552	510	108	90 - 110	P	05/06/2025	14:47	LB135689
	Boron	2410	2500	96	90 - 110	P	05/06/2025	14:47	LB135689
	Cadmium	498	510	98	90 - 110	P	05/06/2025	14:47	LB135689
	Chromium	498	520	96	90 - 110	P	05/06/2025	14:47	LB135689
	Cobalt	487	520	94	90 - 110	P	05/06/2025	14:47	LB135689
	Copper	500	510	98	90 - 110	P	05/06/2025	14:47	LB135689
	Iron	10500	10000	105	90 - 110	P	05/06/2025	14:47	LB135689
	Lead	975	1000	98	90 - 110	P	05/06/2025	14:47	LB135689
	Manganese	493	520	95	90 - 110	P	05/06/2025	14:47	LB135689
	Molybdenum	2440	2500	98	90 - 110	P	05/06/2025	14:47	LB135689
	Nickel	492	530	93	90 - 110	P	05/06/2025	14:47	LB135689
	Selenium	1000	1000	100	90 - 110	P	05/06/2025	14:47	LB135689
	Silver	232	250	93	90 - 110	P	05/06/2025	14:47	LB135689
	Thallium	1000	1000	100	90 - 110	P	05/06/2025	14:47	LB135689
	Vanadium	478	500	96	90 - 110	P	05/06/2025	14:47	LB135689
	Zinc	975	1000	98	90 - 110	P	05/06/2025	14:47	LB135689

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder SDG No.: Q1903
Contract: POWE02 Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	105	100	105	80 - 120	P	05/06/2025	15:22	LB135689
	Antimony	50.3	50.0	101	80 - 120	P	05/06/2025	15:22	LB135689
	Arsenic	18.2	20.0	91	80 - 120	P	05/06/2025	15:22	LB135689
	Barium	91.5	100	92	80 - 120	P	05/06/2025	15:22	LB135689
	Beryllium	5.90	6.0	98	80 - 120	P	05/06/2025	15:22	LB135689
	Boron	99.9	100	100	80 - 120	P	05/06/2025	15:22	LB135689
	Cadmium	5.98	6.0	100	80 - 120	P	05/06/2025	15:22	LB135689
	Chromium	9.31	10.0	93	80 - 120	P	05/06/2025	15:22	LB135689
	Cobalt	28.6	30.0	95	80 - 120	P	05/06/2025	15:22	LB135689
	Copper	21.1	20.0	106	80 - 120	P	05/06/2025	15:22	LB135689
	Iron	98.2	100	98	80 - 120	P	05/06/2025	15:22	LB135689
	Lead	11.1	12.0	92	80 - 120	P	05/06/2025	15:22	LB135689
	Manganese	19.6	20.0	98	80 - 120	P	05/06/2025	15:22	LB135689
	Molybdenum	202	200	101	80 - 120	P	05/06/2025	15:22	LB135689
	Nickel	38.9	40.0	97	80 - 120	P	05/06/2025	15:22	LB135689
	Selenium	19.1	20.0	95	80 - 120	P	05/06/2025	15:22	LB135689
	Silver	9.80	10.0	98	80 - 120	P	05/06/2025	15:22	LB135689
	Thallium	39.9	40.0	100	80 - 120	P	05/06/2025	15:22	LB135689
	Vanadium	38.0	40.0	95	80 - 120	P	05/06/2025	15:22	LB135689
	Zinc	40.4	40.0	101	80 - 120	P	05/06/2025	15:22	LB135689

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder SDG No.: Q1903
Contract: POWE02 Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9630	10000	96	90 - 110	P	05/06/2025	16:29	LB135689
	Antimony	4790	5000	96	90 - 110	P	05/06/2025	16:29	LB135689
	Arsenic	4780	5000	96	90 - 110	P	05/06/2025	16:29	LB135689
	Barium	9340	10000	93	90 - 110	P	05/06/2025	16:29	LB135689
	Beryllium	249	250	100	90 - 110	P	05/06/2025	16:29	LB135689
	Boron	4980	5000	100	90 - 110	P	05/06/2025	16:29	LB135689
	Cadmium	2390	2500	95	90 - 110	P	05/06/2025	16:29	LB135689
	Chromium	998	1000	100	90 - 110	P	05/06/2025	16:29	LB135689
	Cobalt	2390	2500	96	90 - 110	P	05/06/2025	16:29	LB135689
	Copper	1210	1250	97	90 - 110	P	05/06/2025	16:29	LB135689
	Iron	5040	5000	101	90 - 110	P	05/06/2025	16:29	LB135689
	Lead	5030	5000	101	90 - 110	P	05/06/2025	16:29	LB135689
	Manganese	2400	2500	96	90 - 110	P	05/06/2025	16:29	LB135689
	Molybdenum	4800	5000	96	90 - 110	P	05/06/2025	16:29	LB135689
	Nickel	2400	2500	96	90 - 110	P	05/06/2025	16:29	LB135689
	Selenium	4800	5000	96	90 - 110	P	05/06/2025	16:29	LB135689
	Silver	1240	1250	99	90 - 110	P	05/06/2025	16:29	LB135689
	Thallium	4670	5000	94	90 - 110	P	05/06/2025	16:29	LB135689
	Vanadium	2400	2500	96	90 - 110	P	05/06/2025	16:29	LB135689
	Zinc	2470	2500	99	90 - 110	P	05/06/2025	16:29	LB135689
CCV02	Aluminum	9480	10000	95	90 - 110	P	05/06/2025	17:15	LB135689
	Antimony	4750	5000	95	90 - 110	P	05/06/2025	17:15	LB135689
	Arsenic	4730	5000	95	90 - 110	P	05/06/2025	17:15	LB135689
	Barium	9710	10000	97	90 - 110	P	05/06/2025	17:15	LB135689
	Beryllium	237	250	95	90 - 110	P	05/06/2025	17:15	LB135689
	Boron	4750	5000	95	90 - 110	P	05/06/2025	17:15	LB135689
	Cadmium	2390	2500	96	90 - 110	P	05/06/2025	17:15	LB135689
	Chromium	962	1000	96	90 - 110	P	05/06/2025	17:15	LB135689
	Cobalt	2390	2500	96	90 - 110	P	05/06/2025	17:15	LB135689
	Copper	1200	1250	96	90 - 110	P	05/06/2025	17:15	LB135689
	Iron	5000	5000	100	90 - 110	P	05/06/2025	17:15	LB135689
	Lead	4890	5000	98	90 - 110	P	05/06/2025	17:15	LB135689
	Manganese	2410	2500	96	90 - 110	P	05/06/2025	17:15	LB135689
	Molybdenum	4810	5000	96	90 - 110	P	05/06/2025	17:15	LB135689

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder SDG No.: Q1903
Contract: POWE02 Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Nickel	2390	2500	96	90 - 110	P	05/06/2025	17:15	LB135689
	Selenium	4750	5000	95	90 - 110	P	05/06/2025	17:15	LB135689
	Silver	1200	1250	96	90 - 110	P	05/06/2025	17:15	LB135689
	Thallium	4770	5000	96	90 - 110	P	05/06/2025	17:15	LB135689
	Vanadium	2390	2500	95	90 - 110	P	05/06/2025	17:15	LB135689
	Zinc	2410	2500	96	90 - 110	P	05/06/2025	17:15	LB135689
CCV03	Aluminum	9460	10000	95	90 - 110	P	05/06/2025	18:08	LB135689
	Antimony	4590	5000	92	90 - 110	P	05/06/2025	18:08	LB135689
	Arsenic	4570	5000	92	90 - 110	P	05/06/2025	18:08	LB135689
	Barium	9450	10000	94	90 - 110	P	05/06/2025	18:08	LB135689
	Beryllium	240	250	96	90 - 110	P	05/06/2025	18:08	LB135689
	Boron	4780	5000	96	90 - 110	P	05/06/2025	18:08	LB135689
	Cadmium	2340	2500	94	90 - 110	P	05/06/2025	18:08	LB135689
	Chromium	954	1000	95	90 - 110	P	05/06/2025	18:08	LB135689
	Cobalt	2340	2500	94	90 - 110	P	05/06/2025	18:08	LB135689
	Copper	1170	1250	94	90 - 110	P	05/06/2025	18:08	LB135689
	Iron	4700	5000	94	90 - 110	P	05/06/2025	18:08	LB135689
	Lead	4690	5000	94	90 - 110	P	05/06/2025	18:08	LB135689
	Manganese	2390	2500	95	90 - 110	P	05/06/2025	18:08	LB135689
	Molybdenum	4720	5000	94	90 - 110	P	05/06/2025	18:08	LB135689
	Nickel	2340	2500	94	90 - 110	P	05/06/2025	18:08	LB135689
	Selenium	4550	5000	91	90 - 110	P	05/06/2025	18:08	LB135689
	Silver	1180	1250	94	90 - 110	P	05/06/2025	18:08	LB135689
	Thallium	4710	5000	94	90 - 110	P	05/06/2025	18:08	LB135689
	Vanadium	2360	2500	94	90 - 110	P	05/06/2025	18:08	LB135689
	Zinc	2370	2500	95	90 - 110	P	05/06/2025	18:08	LB135689
CCV04	Aluminum	9960	10000	100	90 - 110	P	05/06/2025	19:16	LB135689
	Antimony	4840	5000	97	90 - 110	P	05/06/2025	19:16	LB135689
	Arsenic	4840	5000	97	90 - 110	P	05/06/2025	19:16	LB135689
	Barium	9930	10000	99	90 - 110	P	05/06/2025	19:16	LB135689
	Beryllium	264	250	106	90 - 110	P	05/06/2025	19:16	LB135689
	Boron	5310	5000	106	90 - 110	P	05/06/2025	19:16	LB135689
	Cadmium	2470	2500	99	90 - 110	P	05/06/2025	19:16	LB135689
	Chromium	988	1000	99	90 - 110	P	05/06/2025	19:16	LB135689

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder SDG No.: Q1903
Contract: POWE02 Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Cobalt	2470	2500	99	90 - 110	P	05/06/2025	19:16	LB135689
	Copper	1240	1250	99	90 - 110	P	05/06/2025	19:16	LB135689
	Iron	4820	5000	96	90 - 110	P	05/06/2025	19:16	LB135689
	Lead	4970	5000	99	90 - 110	P	05/06/2025	19:16	LB135689
	Manganese	2530	2500	101	90 - 110	P	05/06/2025	19:16	LB135689
	Molybdenum	4920	5000	98	90 - 110	P	05/06/2025	19:16	LB135689
	Nickel	2470	2500	99	90 - 110	P	05/06/2025	19:16	LB135689
	Selenium	4820	5000	96	90 - 110	P	05/06/2025	19:16	LB135689
	Silver	1240	1250	99	90 - 110	P	05/06/2025	19:16	LB135689
	Thallium	4760	5000	95	90 - 110	P	05/06/2025	19:16	LB135689
	Vanadium	2470	2500	99	90 - 110	P	05/06/2025	19:16	LB135689
	Zinc	2490	2500	100	90 - 110	P	05/06/2025	19:16	LB135689
CCV05	Aluminum	10000	10000	100	90 - 110	P	05/06/2025	20:01	LB135689
	Antimony	4800	5000	96	90 - 110	P	05/06/2025	20:01	LB135689
	Arsenic	4780	5000	96	90 - 110	P	05/06/2025	20:01	LB135689
	Barium	9940	10000	99	90 - 110	P	05/06/2025	20:01	LB135689
	Beryllium	260	250	104	90 - 110	P	05/06/2025	20:01	LB135689
	Boron	5210	5000	104	90 - 110	P	05/06/2025	20:01	LB135689
	Cadmium	2460	2500	99	90 - 110	P	05/06/2025	20:01	LB135689
	Chromium	995	1000	100	90 - 110	P	05/06/2025	20:01	LB135689
	Cobalt	2450	2500	98	90 - 110	P	05/06/2025	20:01	LB135689
	Copper	1230	1250	98	90 - 110	P	05/06/2025	20:01	LB135689
	Iron	4860	5000	97	90 - 110	P	05/06/2025	20:01	LB135689
	Lead	4900	5000	98	90 - 110	P	05/06/2025	20:01	LB135689
	Manganese	2520	2500	101	90 - 110	P	05/06/2025	20:01	LB135689
	Molybdenum	4940	5000	99	90 - 110	P	05/06/2025	20:01	LB135689
	Nickel	2450	2500	98	90 - 110	P	05/06/2025	20:01	LB135689
	Selenium	4730	5000	95	90 - 110	P	05/06/2025	20:01	LB135689
	Silver	1230	1250	99	90 - 110	P	05/06/2025	20:01	LB135689
	Thallium	4800	5000	96	90 - 110	P	05/06/2025	20:01	LB135689
	Vanadium	2470	2500	99	90 - 110	P	05/06/2025	20:01	LB135689
	Zinc	2480	2500	99	90 - 110	P	05/06/2025	20:01	LB135689
CCV06	Aluminum	9830	10000	98	90 - 110	P	05/06/2025	20:31	LB135689
	Antimony	4880	5000	98	90 - 110	P	05/06/2025	20:31	LB135689

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder SDG No.: Q1903
Contract: POWE02 Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV06	Arsenic	4840	5000	97	90 - 110	P	05/06/2025	20:31	LB135689
	Barium	9850	10000	98	90 - 110	P	05/06/2025	20:31	LB135689
	Beryllium	241	250	96	90 - 110	P	05/06/2025	20:31	LB135689
	Boron	4830	5000	97	90 - 110	P	05/06/2025	20:31	LB135689
	Cadmium	2440	2500	98	90 - 110	P	05/06/2025	20:31	LB135689
	Chromium	975	1000	98	90 - 110	P	05/06/2025	20:31	LB135689
	Cobalt	2430	2500	97	90 - 110	P	05/06/2025	20:31	LB135689
	Copper	1240	1250	99	90 - 110	P	05/06/2025	20:31	LB135689
	Iron	4920	5000	98	90 - 110	P	05/06/2025	20:31	LB135689
	Lead	4870	5000	97	90 - 110	P	05/06/2025	20:31	LB135689
	Manganese	2440	2500	98	90 - 110	P	05/06/2025	20:31	LB135689
	Molybdenum	4950	5000	99	90 - 110	P	05/06/2025	20:31	LB135689
	Nickel	2430	2500	97	90 - 110	P	05/06/2025	20:31	LB135689
	Selenium	4820	5000	96	90 - 110	P	05/06/2025	20:31	LB135689
	Silver	1210	1250	97	90 - 110	P	05/06/2025	20:31	LB135689
	Thallium	4750	5000	95	90 - 110	P	05/06/2025	20:31	LB135689
	Vanadium	2430	2500	97	90 - 110	P	05/06/2025	20:31	LB135689
	Zinc	2440	2500	98	90 - 110	P	05/06/2025	20:31	LB135689