

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM **Case No.:** Q1903 **SAS No.:** Q1903 **SDG NO.:** Q1903

Continuing Calib Date: 04/29/2025 **Initial Calibration Date(s):** 04/10/2025 04/10/2025

Continuing Calib Time: 09:45 **Initial Calibration Time(s):** 09:36 17:52

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Tetrachloro-m-xylene	3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl	8.73	8.73	8.63	8.83	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM **Case No.:** Q1903 **SAS No.:** Q1903 **SDG NO.:** Q1903

Continuing Calib Date: 04/29/2025 **Initial Calibration Date(s):** 04/10/2025 04/10/2025

Continuing Calib Time: 09:45 **Initial Calibration Time(s):** 09:36 17:52

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Tetrachloro-m-xylene	3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl	8.68	8.68	8.58	8.78	0.00