

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

Continuing Calib Date: 04/30/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 03:25 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.58	5.59	5.49	5.69	0.01
Aroclor-1254-2 (2)	5.73	5.74	5.64	5.84	0.01
Aroclor-1254-3 (3)	6.14	6.14	6.04	6.24	0.00
Aroclor-1254-4 (4)	6.37	6.37	6.27	6.47	0.00
Aroclor-1254-5 (5)	6.79	6.79	6.69	6.89	0.00
Tetrachloro-m-xylene	3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl	8.73	8.73	8.63	8.83	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

Continuing Calib Date: 04/30/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 03:25 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.56	5.57	5.47	5.67	0.01
Aroclor-1254-2 (2)	5.71	5.71	5.61	5.81	0.00
Aroclor-1254-3 (3)	6.11	6.12	6.02	6.22	0.01
Aroclor-1254-4 (4)	6.34	6.34	6.24	6.44	0.00
Aroclor-1254-5 (5)	6.76	6.76	6.66	6.86	0.00
Tetrachloro-m-xylene	3.68	3.69	3.59	3.79	0.01
Decachlorobiphenyl	8.68	8.68	8.58	8.78	0.00