

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM **Case No.:** Q1903 **SAS No.:** Q1903 **SDG NO.:** Q1903

Continuing Calib Date: 04/29/2025 **Initial Calibration Date(s):** 04/22/2025 04/22/2025

Continuing Calib Time: 17:02 **Initial Calibration Time(s):** 10:29 17:49

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Tetrachloro-m-xylene	4.51	4.51	4.41	4.61	0.00
Decachlorobiphenyl	10.23	10.24	10.14	10.34	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM **Case No.:** Q1903 **SAS No.:** Q1903 **SDG NO.:** Q1903

Continuing Calib Date: 04/29/2025 **Initial Calibration Date(s):** 04/22/2025 04/22/2025

Continuing Calib Time: 17:02 **Initial Calibration Time(s):** 10:29 17:49

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Tetrachloro-m-xylene	3.81	3.81	3.71	3.91	0.00
Decachlorobiphenyl	8.84	8.85	8.75	8.95	0.01