

LAB CHRONICLE

OrderID:	Q1903	OrderDate:	4/28/2025 11:30:00 AM
Client:	Kleinfelder	Project:	Mitchell School
Contact:	Mark Warchol	Location:	L51,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1903-01	COMP-4	SOIL	PCB Group1	8082A	04/25/25	04/29/25	04/29/25	04/28/25
Q1903-02	COMP-5	SOIL	PCB Group1	8082A	04/25/25	04/29/25	04/29/25	04/28/25
Q1903-03	COMP-6	SOIL	PCB Group1	8082A	04/25/25	04/29/25	04/29/25	04/28/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1903

Order ID: Q1903

Client: Kleinfelder

Project ID: Mitchell School

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000

A
B
C
D
E
F
G



SAMPLE DATA

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/25/25
Project:	Mitchell School	Date Received:	04/28/25
Client Sample ID:	COMP-4	SDG No.:	Q1903
Lab Sample ID:	Q1903-01	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	83.3
Sample Wt/Vol:	30.02	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110858.D	1	04/29/25 08:35	04/29/25 14:12	PB167776

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.4	ug/kg
11097-69-1	Aroclor-1254	3.90	U	3.90	20.4	ug/kg
11096-82-5	Aroclor-1260	3.90	U	3.90	20.4	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.9		32 - 144	120%	SPK: 20
2051-24-3	Decachlorobiphenyl	20.1		32 - 175	101%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/25/25
Project:	Mitchell School	Date Received:	04/28/25
Client Sample ID:	COMP-5	SDG No.:	Q1903
Lab Sample ID:	Q1903-02	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	78.9
Sample Wt/Vol:	30.06	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110859.D	1	04/29/25 08:35	04/29/25 14:31	PB167776

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	5.00	U	5.00	21.5	ug/kg
11097-69-1	Aroclor-1254	4.10	U	4.10	21.5	ug/kg
11096-82-5	Aroclor-1260	4.10	U	4.10	21.5	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.8		32 - 144	109%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.6		32 - 175	88%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/25/25
Project:	Mitchell School	Date Received:	04/28/25
Client Sample ID:	COMP-6	SDG No.:	Q1903
Lab Sample ID:	Q1903-03	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	83.9
Sample Wt/Vol:	30.08	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110860.D	1	04/29/25 08:35	04/29/25 14:49	PB167776

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.2	ug/kg
11097-69-1	Aroclor-1254	3.80	U	3.80	20.2	ug/kg
11096-82-5	Aroclor-1260	3.80	U	3.80	20.2	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.2		32 - 144	116%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.7		32 - 175	99%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



QC SUMMARY

Surrogate Summary

SDG No.: Q1903

Client: Kleinfelder

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO110348.D	PIBLK-PO110348.D	Tetrachloro-m-xylene	1	20	18.1	91		60	140
		Decachlorobiphenyl	1	20	19.2	96		60	140
		Tetrachloro-m-xylene	2	20	18.2	91		60	140
		Decachlorobiphenyl	2	20	20.2	101		60	140
I.BLK-PO110850.D	PIBLK-PO110850.D	Tetrachloro-m-xylene	1	20	18.0	90		60	140
		Decachlorobiphenyl	1	20	16.9	85		60	140
		Tetrachloro-m-xylene	2	20	16.6	83		60	140
		Decachlorobiphenyl	2	20	16.7	84		60	140
Q1903-01	COMP-4	Tetrachloro-m-xylene	1	20	23.9	120		32	144
		Decachlorobiphenyl	1	20	20.1	101		32	175
		Tetrachloro-m-xylene	2	20	21.8	109		32	144
		Decachlorobiphenyl	2	20	19.9	100		32	175
Q1903-02	COMP-5	Tetrachloro-m-xylene	1	20	21.8	109		32	144
		Decachlorobiphenyl	1	20	17.6	88		32	175
		Tetrachloro-m-xylene	2	20	20.0	100		32	144
		Decachlorobiphenyl	2	20	17.5	88		32	175
Q1903-03	COMP-6	Tetrachloro-m-xylene	1	20	23.2	116		32	144
		Decachlorobiphenyl	1	20	19.7	99		32	175
		Tetrachloro-m-xylene	2	20	21.2	106		32	144
		Decachlorobiphenyl	2	20	19.2	96		32	175
I.BLK-PO110865.D	PIBLK-PO110865.D	Tetrachloro-m-xylene	1	20	18.4	92		60	140
		Decachlorobiphenyl	1	20	17.5	88		60	140
		Tetrachloro-m-xylene	2	20	16.9	84		60	140
		Decachlorobiphenyl	2	20	16.9	85		60	140
Q1904-01MS	VNJ-210MS	Tetrachloro-m-xylene	1	20	23.3	117		32	144
		Decachlorobiphenyl	1	20	19.1	96		32	175
		Tetrachloro-m-xylene	2	20	21.3	106		32	144
		Decachlorobiphenyl	2	20	19.1	95		32	175
Q1904-01MSD	VNJ-210MSD	Tetrachloro-m-xylene	1	20	22.9	114		32	144
		Decachlorobiphenyl	1	20	19.0	95		32	175
		Tetrachloro-m-xylene	2	20	20.8	104		32	144
		Decachlorobiphenyl	2	20	18.7	93		32	175
I.BLK-PO110880.D	PIBLK-PO110880.D	Tetrachloro-m-xylene	1	20	18.4	92		60	140
		Decachlorobiphenyl	1	20	16.6	83		60	140
		Tetrachloro-m-xylene	2	20	17.0	85		60	140
		Decachlorobiphenyl	2	20	16.7	83		60	140
I.BLK-PP071388.D	PIBLK-PP071388.D	Tetrachloro-m-xylene	1	20	17.1	85		60	140
		Decachlorobiphenyl	1	20	17.6	88		60	140
		Tetrachloro-m-xylene	2	20	17.2	86		60	140
		Decachlorobiphenyl	2	20	17.6	88		60	140
I.BLK-PP071603.D	PIBLK-PP071603.D	Tetrachloro-m-xylene	1	20	15.9	79		60	140

Surrogate Summary

SDG No.: Q1903

Client: Kleinfelder

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP071603.D	PIBLK-PP071603.D	Decachlorobiphenyl	1	20	17.3	86		60	140
		Tetrachloro-m-xylene	2	20	18.3	91		60	140
PB167776BL	PB167776BL	Decachlorobiphenyl	2	20	18.6	93		60	140
		Tetrachloro-m-xylene	1	20	18.7	94		32	144
		Decachlorobiphenyl	1	20	19.5	97		32	175
		Tetrachloro-m-xylene	2	20	19.9	100		32	144
		Decachlorobiphenyl	2	20	21.1	105		32	175
		Tetrachloro-m-xylene	1	20	19.5	97		32	144
PB167776BS	PB167776BS	Decachlorobiphenyl	1	20	20.9	105		32	175
		Tetrachloro-m-xylene	2	20	20.7	103		32	144
		Decachlorobiphenyl	2	20	21.3	106		32	175
		Tetrachloro-m-xylene	1	20	16.0	80		60	140
I.BLK-PP071618.D	PIBLK-PP071618.D	Decachlorobiphenyl	1	20	18.0	90		60	140
		Tetrachloro-m-xylene	2	20	17.3	86		60	140
		Decachlorobiphenyl	2	20	19.3	96		60	140
		Tetrachloro-m-xylene	1	20	16.0	80		60	140

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1903

Client: Kleinfelder

Analytical Method: 8082A

DataFile : PO110867.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	VNJ-210MS											
Q1904-01MS	AR1016	183.8	0	201	ug/kg	109				55	146	
	AR1260	183.8	0	187	ug/kg	102				31	146	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1903

Client: Kleinfelder

Analytical Method: 8082A

DataFile : PO110868.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q1904-01MSD	VNJ-210MSD											
	AR1016	183.7	0	199	ug/kg	108		1		55	146	20
	AR1260	183.7	0	183	ug/kg	100		2		31	146	20



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1903

Client: Kleinfelder

Analytical Method: 8082A Datafile : PP071608.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB167776BS	AR1016	166.5	156	ug/kg	94				71	120		
	AR1260	166.5	157	ug/kg	94				65	130		

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167776BL

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM Case No.: Q1903

SAS No.: Q1903 SDG NO.: Q1903

Lab Sample ID: PB167776BL

Lab File ID: PP071607.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 04/29/2025

Date Analyzed (1): 04/29/2025

Date Analyzed (2): 04/29/2025

Time Analyzed (1): 13:30

Time Analyzed (2): 13:30

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
COMP-4	Q1903-01	PO110858.D	04/29/2025	04/29/2025
COMP-5	Q1903-02	PO110859.D	04/29/2025	04/29/2025
COMP-6	Q1903-03	PO110860.D	04/29/2025	04/29/2025
VNJ-210MS	Q1904-01MS	PO110867.D	04/29/2025	04/29/2025
VNJ-210MSD	Q1904-01MSD	PO110868.D	04/29/2025	04/29/2025
PB167776BS	PB167776BS	PP071608.D	04/29/2025	04/29/2025

COMMENTS: _____



CALIBRATION SUMMARY

A
B
C
D
E
F
G

Calibration Times: 09:36 17:52

LAB FILE ID: RT 1000 = PO110349.D RT 750 = PO110350.D
RT 500 = PO110351.D RT 250 = PO110352.D RT 050 = PO110353.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

Instrument ID: ECD_O Calibration Date(s): 04/10/2025 04/10/2025

Calibration Times: 09:36 17:52

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PO110349.D</u>	RT 750 = <u>PO110350.D</u>
RT 500 = <u>PO110351.D</u>	RT 250 = <u>PO110352.D</u>	RT 050 = <u>PO110353.D</u>

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Aroclor-1016-1	(1)	4.77	4.77	4.77	4.77	4.77	4.77	4.67	4.87
Aroclor-1016-2	(2)	4.78	4.79	4.78	4.78	4.79	4.78	4.68	4.88
Aroclor-1016-3	(3)	4.96	4.96	4.96	4.96	4.96	4.96	4.86	5.06
Aroclor-1016-4	(4)	5.00	5.00	5.00	5.00	5.00	5.00	4.90	5.10
Aroclor-1016-5	(5)	5.22	5.22	5.22	5.21	5.22	5.22	5.12	5.32
Aroclor-1260-1	(1)	6.25	6.25	6.25	6.25	6.25	6.25	6.15	6.35
Aroclor-1260-2	(2)	6.43	6.43	6.43	6.43	6.43	6.43	6.33	6.53
Aroclor-1260-3	(3)	6.59	6.59	6.59	6.59	6.59	6.59	6.49	6.69
Aroclor-1260-4	(4)	7.06	7.06	7.06	7.06	7.06	7.06	6.96	7.16
Aroclor-1260-5	(5)	7.30	7.30	7.30	7.30	7.30	7.30	7.20	7.40
Decachlorobiphenyl		8.69	8.69	8.68	8.68	8.68	8.68	8.58	8.78
Tetrachloro-m-xylene		3.69	3.69	3.69	3.69	3.69	3.69	3.59	3.79
Aroclor-1254-1	(1)	5.57	5.57	5.57	5.57	5.57	5.57	5.47	5.67
Aroclor-1254-2	(2)	5.71	5.71	5.71	5.71	5.71	5.71	5.61	5.81
Aroclor-1254-3	(3)	6.12	6.12	6.12	6.12	6.12	6.12	6.02	6.22
Aroclor-1254-4	(4)	6.34	6.34	6.34	6.34	6.34	6.34	6.24	6.44
Aroclor-1254-5	(5)	6.76	6.76	6.76	6.76	6.76	6.76	6.66	6.86
Decachlorobiphenyl		8.68	8.68	8.68	8.68	8.68	8.68	8.58	8.78
Tetrachloro-m-xylene		3.69	3.69	3.69	3.68	3.69	3.69	3.59	3.79

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM **Case No.:** Q1903 **SAS No.:** Q1903 **SDG NO.:** Q1903

Instrument ID: ECD_O **Calibration Date(s):** 04/10/2025 04/10/2025
Calibration Times: 09:36 17:52

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 =	PO110349.D	CF 750 =	PO110350.D			
CF 500 =		PO110351.D	CF 250 =	PO110352.D	CF 050 =	PO110353.D		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	301539199	312147691	325366094	341425388	360239300	328143534	7
Aroclor-1016-2	(2)	424202495	441346851	452587054	472466840	484138100	454948268	5
Aroclor-1016-3	(3)	286857785	300177247	315792572	336640016	373170260	322527576	10
Aroclor-1016-4	(4)	227624644	238137328	247842320	261218376	270999820	249164498	7
Aroclor-1016-5	(5)	238190199	250655776	262830422	281583724	311798080	269011640	11
Aroclor-1260-1	(1)	429982139	447081743	462005040	490865312	532292280	472445303	9
Aroclor-1260-2	(2)	524274151	548276752	560001580	596374412	703195060	586424391	12
Aroclor-1260-3	(3)	446012749	466468717	482076654	511871688	557932460	492872454	9
Aroclor-1260-4	(4)	382143422	400404183	417623168	445374396	476569960	424423026	9
Aroclor-1260-5	(5)	1002895230	1031868237	1046125762	1065122080	1071346720	1043471606	3
Decachlorobiphenyl		7251274760	7526590267	7768647820	8211713920	8716546400	7894954633	7
Tetrachloro-m-xylene		8603197440	8796106920	8951138380	8737007280	8657843200	8749058644	2
Aroclor-1254-1	(1)	517202507	536393799	555537214	581626948	599185020	557989098	6
Aroclor-1254-2	(2)	450798901	468206104	477861830	506770124	524875880	485702568	6
Aroclor-1254-3	(3)	746169545	769125461	779743060	811121676	809948840	783221716	4
Aroclor-1254-4	(4)	460149954	473941677	490156778	505394424	508383480	487605263	4
Aroclor-1254-5	(5)	657635451	678198725	698380664	724566284	745718260	700899877	5
Decachlorobiphenyl		7612990370	7886723680	8142983840	8479059160	8719027400	8168156890	5
Tetrachloro-m-xylene		8665092200	8825966600	8709112240	8974669960	8357494600	8706467120	3

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM **Case No.:** Q1903 **SAS No.:** Q1903 **SDG NO.:** Q1903

Instrument ID: ECD_O

Calibration Date(s): 04/10/2025 04/10/2025

Calibration Times: 09:36 17:52

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 =	PO110349.D	CF 750 =	PO110350.D			
CF 500 =		PO110351.D	CF 250 =	PO110352.D	CF 050 =	PO110353.D		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	159213491	166153001	173161426	184455932	194563640	175509498	8
Aroclor-1016-2	(2)	235577328	244073360	249915638	259769828	267988640	251464959	5
Aroclor-1016-3	(3)	123754519	129314112	134837594	142342816	148305560	135710920	7
Aroclor-1016-4	(4)	100154968	106815488	112688448	121621132	130559280	114367863	11
Aroclor-1016-5	(5)	131736277	138471012	145266388	155239544	174656140	149073872	11
Aroclor-1260-1	(1)	221282381	231306280	240804372	256804492	279620260	245963557	9
Aroclor-1260-2	(2)	260045501	271771868	281572488	298549944	346814040	291750768	12
Aroclor-1260-3	(3)	243412958	253822376	261595986	277903536	317685740	270884119	11
Aroclor-1260-4	(4)	177247956	186943988	195746612	210913364	232348220	200640028	11
Aroclor-1260-5	(5)	430279301	445092429	452357666	469475868	497083160	458857685	6
Decachlorobiphenyl		1704536950	1791384040	1872045360	2025627840	2231357400	1924990318	11
Tetrachloro-m-xylene		4868987290	4975470093	5049711440	5093085920	4961663200	4989783589	2
Aroclor-1254-1	(1)	275528987	285381728	296386294	310788612	339409960	301499116	8
Aroclor-1254-2	(2)	239912515	249193828	258146844	274680708	301492780	264685335	9
Aroclor-1254-3	(3)	379979735	392440955	403528434	417422392	423708940	403416091	4
Aroclor-1254-4	(4)	215870098	222432417	229245508	238627400	242389020	229712889	5
Aroclor-1254-5	(5)	307722172	319481015	330728230	346867608	356659680	332291741	6
Decachlorobiphenyl		1672742580	1794748347	1901326780	2085791440	2132484800	1917418789	10
Tetrachloro-m-xylene		4865625930	4959437227	4872791400	5058078960	4761547200	4903496143	2

RETENTION TIMES OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

Instrument ID: ECD_P Calibration Date(s): 04/22/2025 04/22/2025

Calibration Times: 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PP071389.D RT 750 = PP071390.D

RT 500 = PP071391.D RT 250 = PP071392.D RT 050 = PP071393.D

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Aroclor-1016-1	(1)	5.66	5.67	5.67	5.67	5.67	5.67	5.57	5.77
Aroclor-1016-2	(2)	5.69	5.69	5.69	5.69	5.69	5.69	5.59	5.79
Aroclor-1016-3	(3)	5.75	5.75	5.75	5.75	5.75	5.75	5.65	5.85
Aroclor-1016-4	(4)	5.85	5.85	5.85	5.85	5.85	5.85	5.75	5.95
Aroclor-1016-5	(5)	6.14	6.14	6.15	6.14	6.14	6.14	6.04	6.24
Aroclor-1260-1	(1)	7.26	7.26	7.27	7.26	7.26	7.26	7.16	7.36
Aroclor-1260-2	(2)	7.51	7.51	7.52	7.52	7.52	7.52	7.42	7.62
Aroclor-1260-3	(3)	7.87	7.87	7.88	7.87	7.88	7.87	7.77	7.97
Aroclor-1260-4	(4)	8.10	8.10	8.10	8.10	8.10	8.10	8.00	8.20
Aroclor-1260-5	(5)	8.42	8.42	8.42	8.42	8.42	8.42	8.32	8.52
Decachlorobiphenyl		10.23	10.23	10.24	10.23	10.24	10.23	10.13	10.33
Tetrachloro-m-xylene		4.51	4.51	4.52	4.51	4.52	4.51	4.41	4.61
Aroclor-1254-1	(1)	6.52	6.52	6.52	6.52	6.52	6.52	6.42	6.62
Aroclor-1254-2	(2)	6.74	6.74	6.74	6.73	6.74	6.74	6.64	6.84
Aroclor-1254-3	(3)	7.10	7.10	7.10	7.10	7.10	7.10	7.00	7.20
Aroclor-1254-4	(4)	7.38	7.38	7.38	7.38	7.38	7.38	7.28	7.48
Aroclor-1254-5	(5)	7.80	7.80	7.80	7.80	7.80	7.80	7.70	7.90
Decachlorobiphenyl		10.24	10.24	10.24	10.23	10.24	10.24	10.14	10.34
Tetrachloro-m-xylene		4.52	4.52	4.52	4.51	4.51	4.51	4.41	4.61

A
B
C
D
E
F
G

Calibration Times: 10:29 17:49

LAB FILE ID: RT 1000 = PP071389.D RT 750 = PP071390.D
RT 500 = PP071391.D RT 250 = PP071392.D RT 050 = PP071393.D

[illegible]

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM **Case No.:** Q1903 **SAS No.:** Q1903 **SDG NO.:** Q1903

Instrument ID: ECD_P **Calibration Date(s):** 04/22/2025 04/22/2025
Calibration Times: 10:29 17:49

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 =	PP071389.D	CF 750 =	PP071390.D			
CF 500 =		PP071391.D	CF 250 =	PP071392.D	CF 050 =	PP071393.D		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	64709894	67652695	70083532	71128656	61674340	67049823	6
Aroclor-1016-2	(2)	98977911	102364897	107369074	108813892	95615840	102628323	5
Aroclor-1016-3	(3)	59905708	62127143	64296308	65849928	56056440	61647105	6
Aroclor-1016-4	(4)	50563747	52419789	53619076	54941428	44946700	51298148	8
Aroclor-1016-5	(5)	47772963	49344036	50670668	50888564	42515680	48238382	7
Aroclor-1260-1	(1)	90879847	94857525	98814986	99659744	95910940	96024608	4
Aroclor-1260-2	(2)	137163459	142386411	147498372	149887700	138450980	143077384	4
Aroclor-1260-3	(3)	110350751	114445481	117702804	126515524	91505860	112104084	12
Aroclor-1260-4	(4)	105570732	110780085	111211068	131260612	105501660	112864831	9
Aroclor-1260-5	(5)	220536830	229070229	228919678	255108180	199231240	226573231	9
Decachlorobiphenyl		1409816860	1467002307	1522583160	1539948480	1300700600	1448010281	7
Tetrachloro-m-xylene		1947183160	2005072840	2066603680	2100352680	1793966200	1982635712	6
Aroclor-1254-1	(1)	78574372	81246615	85577334	88598188	75759400	81951182	6
Aroclor-1254-2	(2)	121624435	125816099	132146328	137115708	118678260	127076166	6
Aroclor-1254-3	(3)	123432630	127799839	132981654	137623252	117036460	127774767	6
Aroclor-1254-4	(4)	112512019	116259037	121737436	124392856	112665620	117513394	5
Aroclor-1254-5	(5)	106399089	105922557	113785120	116153044	96391680	107730298	7
Decachlorobiphenyl		1410101810	1469202520	1526389740	1549706840	1328725400	1456825262	6
Tetrachloro-m-xylene		1888241120	1936990333	2012368600	2071565920	1722515200	1926336235	7

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM **Case No.:** Q1903 **SAS No.:** Q1903 **SDG NO.:** Q1903

Instrument ID: ECD_P **Calibration Date(s):** 04/22/2025 04/22/2025
Calibration Times: 10:29 17:49

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP071389.D</u>		CF 750 = <u>PP071390.D</u>				
CF 500 = <u>PP071391.D</u>		CF 250 = <u>PP071392.D</u>		CF 050 = <u>PP071393.D</u>				
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	48121828	49414308	53806128	56180596	60145020	53533576	9
Aroclor-1016-2	(2)	69736709	72096436	77213744	78520924	84250740	76363711	7
Aroclor-1016-3	(3)	38517048	40455577	43831736	45082616	43165080	42210411	6
Aroclor-1016-4	(4)	31206636	32814844	36123020	37095092	35190100	34485938	7
Aroclor-1016-5	(5)	39458687	41985771	45850608	46345748	48595640	44447291	8
Aroclor-1260-1	(1)	63219917	66123699	73727666	76350368	82074520	72299234	11
Aroclor-1260-2	(2)	77338708	77548395	89921718	91492956	101005300	87461415	12
Aroclor-1260-3	(3)	72062678	70944365	81698234	79344004	87888600	78387576	9
Aroclor-1260-4	(4)	57333188	56779915	65331750	68083052	68025140	63110609	9
Aroclor-1260-5	(5)	143809415	139161576	155842678	157512280	157560640	150777318	6
Decachlorobiphenyl		831147600	875337973	902319620	934998560	841460400	877052831	5
Tetrachloro-m-xylene		1380433570	1378340613	1415112020	1448341640	1420974200	1408640409	2
Aroclor-1254-1	(1)	81690839	87361264	89755404	97285536	100908960	91400401	8
Aroclor-1254-2	(2)	70045476	75469849	77317914	83154376	87573280	78712179	9
Aroclor-1254-3	(3)	109725744	114103376	120096818	124380164	124235440	118508308	5
Aroclor-1254-4	(4)	70169900	74068720	77810162	81877428	81901880	77165618	7
Aroclor-1254-5	(5)	92229111	96109599	104690794	107384400	106924600	101467701	7
Decachlorobiphenyl		846307500	875335867	916104900	955261880	868223800	892246789	5
Tetrachloro-m-xylene		1390553300	1501013893	1450901500	1518608160	1468503000	1465915971	3

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 09:24 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-2	(2)	4.80	4.80	4.70	4.90	0.00
Aroclor-1016-3	(3)	4.85	4.86	4.76	4.96	0.01
Aroclor-1016-4	(4)	4.97	4.98	4.88	5.08	0.01
Aroclor-1016-5	(5)	5.23	5.23	5.13	5.33	0.00
Aroclor-1260-1	(1)	6.27	6.27	6.17	6.37	0.00
Aroclor-1260-2	(2)	6.46	6.46	6.36	6.56	0.00
Aroclor-1260-3	(3)	6.82	6.83	6.73	6.93	0.01
Aroclor-1260-4	(4)	7.09	7.09	6.99	7.19	0.00
Aroclor-1260-5	(5)	7.33	7.33	7.23	7.43	0.00
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.73	8.73	8.63	8.83	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 09:24 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.76	4.77	4.67	4.87	0.01
Aroclor-1016-2	(2)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-3	(3)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-4	(4)	5.00	5.00	4.90	5.10	0.00
Aroclor-1016-5	(5)	5.21	5.22	5.12	5.32	0.01
Aroclor-1260-1	(1)	6.24	6.25	6.15	6.35	0.01
Aroclor-1260-2	(2)	6.43	6.43	6.33	6.53	0.00
Aroclor-1260-3	(3)	6.58	6.59	6.49	6.69	0.01
Aroclor-1260-4	(4)	7.05	7.06	6.96	7.16	0.01
Aroclor-1260-5	(5)	7.29	7.30	7.20	7.40	0.01
Tetrachloro-m-xylene		3.68	3.69	3.59	3.79	0.01
Decachlorobiphenyl		8.68	8.68	8.58	8.78	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL01 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO110846.D Time Analyzed: 09:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.775	4.679	4.879	554.440	500.000	10.9
Aroclor-1016-2	4.795	4.698	4.898	558.040	500.000	11.6
Aroclor-1016-3	4.851	4.755	4.955	542.460	500.000	8.5
Aroclor-1016-4	4.971	4.875	5.075	553.630	500.000	10.7
Aroclor-1016-5	5.228	5.132	5.332	543.180	500.000	8.6
Aroclor-1260-1	6.268	6.172	6.372	535.720	500.000	7.1
Aroclor-1260-2	6.457	6.360	6.560	519.830	500.000	4.0
Aroclor-1260-3	6.824	6.729	6.929	522.530	500.000	4.5
Aroclor-1260-4	7.085	6.988	7.188	513.390	500.000	2.7
Aroclor-1260-5	7.326	7.230	7.430	515.480	500.000	3.1
Decachlorobiphenyl	8.727	8.632	8.832	47.400	50.000	-5.2
Tetrachloro-m-xylene	3.686	3.588	3.788	56.210	50.000	12.4

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL01 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO110846.D Time Analyzed: 09:24

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.762	4.666	4.866	520.260	500.000	4.1
Aroclor-1016-2	4.781	4.684	4.884	519.530	500.000	3.9
Aroclor-1016-3	4.956	4.860	5.060	502.490	500.000	0.5
Aroclor-1016-4	4.998	4.902	5.102	493.190	500.000	-1.4
Aroclor-1016-5	5.211	5.115	5.315	490.700	500.000	-1.9
Aroclor-1260-1	6.242	6.146	6.346	492.670	500.000	-1.5
Aroclor-1260-2	6.430	6.334	6.534	484.910	500.000	-3.0
Aroclor-1260-3	6.582	6.487	6.687	480.540	500.000	-3.9
Aroclor-1260-4	7.053	6.957	7.157	480.890	500.000	-3.8
Aroclor-1260-5	7.294	7.197	7.397	482.240	500.000	-3.6
Decachlorobiphenyl	8.679	8.584	8.784	46.500	50.000	-7.0
Tetrachloro-m-xylene	3.683	3.586	3.786	53.130	50.000	6.3

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 15:58 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-2	(2)	4.80	4.80	4.70	4.90	0.00
Aroclor-1016-3	(3)	4.85	4.86	4.76	4.96	0.01
Aroclor-1016-4	(4)	4.97	4.98	4.88	5.08	0.01
Aroclor-1016-5	(5)	5.23	5.23	5.13	5.33	0.00
Aroclor-1260-1	(1)	6.27	6.27	6.17	6.37	0.00
Aroclor-1260-2	(2)	6.46	6.46	6.36	6.56	0.00
Aroclor-1260-3	(3)	6.83	6.83	6.73	6.93	0.01
Aroclor-1260-4	(4)	7.08	7.09	6.99	7.19	0.01
Aroclor-1260-5	(5)	7.33	7.33	7.23	7.43	0.00
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.73	8.73	8.63	8.83	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 15:58 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.76	4.77	4.67	4.87	0.01
Aroclor-1016-2	(2)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-3	(3)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-4	(4)	5.00	5.00	4.90	5.10	0.00
Aroclor-1016-5	(5)	5.21	5.22	5.12	5.32	0.01
Aroclor-1260-1	(1)	6.24	6.25	6.15	6.35	0.01
Aroclor-1260-2	(2)	6.43	6.43	6.33	6.53	0.00
Aroclor-1260-3	(3)	6.58	6.59	6.49	6.69	0.01
Aroclor-1260-4	(4)	7.05	7.06	6.96	7.16	0.01
Aroclor-1260-5	(5)	7.29	7.30	7.20	7.40	0.01
Tetrachloro-m-xylene		3.68	3.69	3.59	3.79	0.01
Decachlorobiphenyl		8.68	8.68	8.58	8.78	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL02 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO110861.D Time Analyzed: 15:58

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.776	4.679	4.879	567.510	500.000	13.5
Aroclor-1016-2	4.795	4.698	4.898	567.550	500.000	13.5
Aroclor-1016-3	4.851	4.755	4.955	552.160	500.000	10.4
Aroclor-1016-4	4.972	4.875	5.075	566.160	500.000	13.2
Aroclor-1016-5	5.229	5.132	5.332	568.900	500.000	13.8
Aroclor-1260-1	6.268	6.172	6.372	552.510	500.000	10.5
Aroclor-1260-2	6.457	6.360	6.560	540.900	500.000	8.2
Aroclor-1260-3	6.825	6.729	6.929	541.450	500.000	8.3
Aroclor-1260-4	7.084	6.988	7.188	532.720	500.000	6.5
Aroclor-1260-5	7.327	7.230	7.430	535.700	500.000	7.1
Decachlorobiphenyl	8.728	8.632	8.832	49.210	50.000	-1.6
Tetrachloro-m-xylene	3.686	3.588	3.788	57.430	50.000	14.9

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL02 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO110861.D Time Analyzed: 15:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.763	4.666	4.866	533.100	500.000	6.6
Aroclor-1016-2	4.781	4.684	4.884	531.290	500.000	6.3
Aroclor-1016-3	4.956	4.860	5.060	513.330	500.000	2.7
Aroclor-1016-4	4.998	4.902	5.102	496.070	500.000	-0.8
Aroclor-1016-5	5.211	5.115	5.315	510.550	500.000	2.1
Aroclor-1260-1	6.242	6.146	6.346	502.200	500.000	0.4
Aroclor-1260-2	6.429	6.334	6.534	495.650	500.000	-0.9
Aroclor-1260-3	6.582	6.487	6.687	487.880	500.000	-2.4
Aroclor-1260-4	7.053	6.957	7.157	485.750	500.000	-2.9
Aroclor-1260-5	7.294	7.197	7.397	488.550	500.000	-2.3
Decachlorobiphenyl	8.679	8.584	8.784	46.590	50.000	-6.8
Tetrachloro-m-xylene	3.683	3.586	3.786	54.630	50.000	9.3

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 22:00 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-2	(2)	4.80	4.80	4.70	4.90	0.00
Aroclor-1016-3	(3)	4.85	4.86	4.76	4.96	0.01
Aroclor-1016-4	(4)	4.97	4.98	4.88	5.08	0.01
Aroclor-1016-5	(5)	5.23	5.23	5.13	5.33	0.00
Aroclor-1260-1	(1)	6.27	6.27	6.17	6.37	0.00
Aroclor-1260-2	(2)	6.46	6.46	6.36	6.56	0.00
Aroclor-1260-3	(3)	6.83	6.83	6.73	6.93	0.01
Aroclor-1260-4	(4)	7.09	7.09	6.99	7.19	0.00
Aroclor-1260-5	(5)	7.33	7.33	7.23	7.43	0.00
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.73	8.73	8.63	8.83	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 22:00 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.76	4.77	4.67	4.87	0.01
Aroclor-1016-2	(2)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-3	(3)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-4	(4)	5.00	5.00	4.90	5.10	0.00
Aroclor-1016-5	(5)	5.21	5.22	5.12	5.32	0.01
Aroclor-1260-1	(1)	6.24	6.25	6.15	6.35	0.01
Aroclor-1260-2	(2)	6.43	6.43	6.33	6.53	0.00
Aroclor-1260-3	(3)	6.58	6.59	6.49	6.69	0.01
Aroclor-1260-4	(4)	7.05	7.06	6.96	7.16	0.01
Aroclor-1260-5	(5)	7.30	7.30	7.20	7.40	0.00
Tetrachloro-m-xylene		3.68	3.69	3.59	3.79	0.01
Decachlorobiphenyl		8.68	8.68	8.58	8.78	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL03 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO110876.D Time Analyzed: 22:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.776	4.679	4.879	557.860	500.000	11.6
Aroclor-1016-2	4.795	4.698	4.898	559.810	500.000	12.0
Aroclor-1016-3	4.851	4.755	4.955	538.720	500.000	7.7
Aroclor-1016-4	4.971	4.875	5.075	545.640	500.000	9.1
Aroclor-1016-5	5.229	5.132	5.332	551.180	500.000	10.2
Aroclor-1260-1	6.269	6.172	6.372	518.480	500.000	3.7
Aroclor-1260-2	6.457	6.360	6.560	510.870	500.000	2.2
Aroclor-1260-3	6.825	6.729	6.929	503.660	500.000	0.7
Aroclor-1260-4	7.085	6.988	7.188	490.870	500.000	-1.8
Aroclor-1260-5	7.327	7.230	7.430	489.900	500.000	-2.0
Decachlorobiphenyl	8.728	8.632	8.832	45.530	50.000	-8.9
Tetrachloro-m-xylene	3.686	3.588	3.788	57.090	50.000	14.2

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL03 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO110876.D Time Analyzed: 22:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.762	4.666	4.866	524.970	500.000	5.0
Aroclor-1016-2	4.781	4.684	4.884	531.970	500.000	6.4
Aroclor-1016-3	4.957	4.860	5.060	500.250	500.000	0.1
Aroclor-1016-4	4.998	4.902	5.102	481.510	500.000	-3.7
Aroclor-1016-5	5.211	5.115	5.315	505.850	500.000	1.2
Aroclor-1260-1	6.243	6.146	6.346	482.790	500.000	-3.4
Aroclor-1260-2	6.431	6.334	6.534	475.660	500.000	-4.9
Aroclor-1260-3	6.583	6.487	6.687	468.070	500.000	-6.4
Aroclor-1260-4	7.054	6.957	7.157	461.140	500.000	-7.8
Aroclor-1260-5	7.295	7.197	7.397	461.730	500.000	-7.7
Decachlorobiphenyl	8.680	8.584	8.784	44.710	50.000	-10.6
Tetrachloro-m-xylene	3.683	3.586	3.786	54.470	50.000	8.9

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 10:33 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.66	5.67	5.57	5.77	0.01
Aroclor-1016-2	(2)	5.69	5.69	5.59	5.79	0.01
Aroclor-1016-3	(3)	5.75	5.75	5.65	5.85	0.00
Aroclor-1016-4	(4)	5.84	5.85	5.75	5.95	0.01
Aroclor-1016-5	(5)	6.14	6.15	6.05	6.25	0.01
Aroclor-1260-1	(1)	7.26	7.27	7.17	7.37	0.01
Aroclor-1260-2	(2)	7.51	7.52	7.42	7.62	0.01
Aroclor-1260-3	(3)	7.87	7.88	7.78	7.98	0.01
Aroclor-1260-4	(4)	8.09	8.10	8.00	8.20	0.01
Aroclor-1260-5	(5)	8.41	8.42	8.32	8.52	0.01
Tetrachloro-m-xylene		4.51	4.52	4.42	4.62	0.01
Decachlorobiphenyl		10.23	10.24	10.14	10.34	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 10:33 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.89	4.90	4.80	5.00	0.01
Aroclor-1016-2	(2)	4.91	4.92	4.82	5.02	0.01
Aroclor-1016-3	(3)	5.09	5.09	4.99	5.19	0.00
Aroclor-1016-4	(4)	5.13	5.14	5.04	5.24	0.01
Aroclor-1016-5	(5)	5.34	5.35	5.25	5.45	0.01
Aroclor-1260-1	(1)	6.38	6.39	6.29	6.49	0.01
Aroclor-1260-2	(2)	6.57	6.57	6.47	6.67	0.00
Aroclor-1260-3	(3)	6.72	6.73	6.63	6.83	0.01
Aroclor-1260-4	(4)	7.19	7.20	7.10	7.30	0.01
Aroclor-1260-5	(5)	7.43	7.44	7.34	7.54	0.01
Tetrachloro-m-xylene		3.81	3.81	3.71	3.91	0.00
Decachlorobiphenyl		8.84	8.85	8.75	8.95	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL04 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PP071599.D Time Analyzed: 10:33

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.663	5.571	5.771	514.370	500.000	2.9
Aroclor-1016-2	5.685	5.593	5.793	504.710	500.000	0.9
Aroclor-1016-3	5.747	5.654	5.854	510.290	500.000	2.1
Aroclor-1016-4	5.844	5.752	5.952	520.150	500.000	4.0
Aroclor-1016-5	6.137	6.045	6.245	506.180	500.000	1.2
Aroclor-1260-1	7.256	7.165	7.365	529.010	500.000	5.8
Aroclor-1260-2	7.509	7.418	7.618	515.870	500.000	3.2
Aroclor-1260-3	7.868	7.777	7.977	531.680	500.000	6.3
Aroclor-1260-4	8.092	8.002	8.202	513.780	500.000	2.8
Aroclor-1260-5	8.412	8.322	8.522	529.980	500.000	6.0
Decachlorobiphenyl	10.225	10.137	10.337	54.140	50.000	8.3
Tetrachloro-m-xylene	4.510	4.417	4.617	50.740	50.000	1.5

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL04 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PP071599.D Time Analyzed: 10:33

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.892	4.798	4.998	525.310	500.000	5.1
Aroclor-1016-2	4.911	4.816	5.016	530.190	500.000	6.0
Aroclor-1016-3	5.088	4.993	5.193	543.770	500.000	8.8
Aroclor-1016-4	5.130	5.036	5.236	528.260	500.000	5.7
Aroclor-1016-5	5.344	5.250	5.450	571.750	500.000	14.4
Aroclor-1260-1	6.378	6.285	6.485	556.170	500.000	11.2
Aroclor-1260-2	6.566	6.473	6.673	550.380	500.000	10.1
Aroclor-1260-3	6.720	6.627	6.827	543.580	500.000	8.7
Aroclor-1260-4	7.190	7.098	7.298	548.770	500.000	9.8
Aroclor-1260-5	7.432	7.339	7.539	538.670	500.000	7.7
Decachlorobiphenyl	8.840	8.749	8.949	53.370	50.000	6.7
Tetrachloro-m-xylene	3.808	3.712	3.912	56.790	50.000	13.6

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 16:29 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.67	5.67	5.57	5.77	0.01
Aroclor-1016-2	(2)	5.69	5.69	5.59	5.79	0.00
Aroclor-1016-3	(3)	5.75	5.75	5.65	5.85	0.00
Aroclor-1016-4	(4)	5.85	5.85	5.75	5.95	0.00
Aroclor-1016-5	(5)	6.14	6.15	6.05	6.25	0.01
Aroclor-1260-1	(1)	7.26	7.27	7.17	7.37	0.01
Aroclor-1260-2	(2)	7.51	7.52	7.42	7.62	0.01
Aroclor-1260-3	(3)	7.87	7.88	7.78	7.98	0.01
Aroclor-1260-4	(4)	8.10	8.10	8.00	8.20	0.00
Aroclor-1260-5	(5)	8.42	8.42	8.32	8.52	0.01
Tetrachloro-m-xylene		4.51	4.52	4.42	4.62	0.01
Decachlorobiphenyl		10.23	10.24	10.14	10.34	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 16:29 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.90	4.90	4.80	5.00	0.01
Aroclor-1016-2	(2)	4.91	4.92	4.82	5.02	0.01
Aroclor-1016-3	(3)	5.09	5.09	4.99	5.19	0.00
Aroclor-1016-4	(4)	5.13	5.14	5.04	5.24	0.01
Aroclor-1016-5	(5)	5.35	5.35	5.25	5.45	0.00
Aroclor-1260-1	(1)	6.38	6.39	6.29	6.49	0.01
Aroclor-1260-2	(2)	6.57	6.57	6.47	6.67	0.00
Aroclor-1260-3	(3)	6.72	6.73	6.63	6.83	0.01
Aroclor-1260-4	(4)	7.19	7.20	7.10	7.30	0.01
Aroclor-1260-5	(5)	7.44	7.44	7.34	7.54	0.01
Tetrachloro-m-xylene		3.81	3.81	3.71	3.91	0.00
Decachlorobiphenyl		8.84	8.85	8.75	8.95	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL05 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PP071614.D Time Analyzed: 16:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.665	5.571	5.771	519.380	500.000	3.9
Aroclor-1016-2	5.688	5.593	5.793	509.990	500.000	2.0
Aroclor-1016-3	5.750	5.654	5.854	509.200	500.000	1.8
Aroclor-1016-4	5.847	5.752	5.952	513.170	500.000	2.6
Aroclor-1016-5	6.140	6.045	6.245	520.600	500.000	4.1
Aroclor-1260-1	7.259	7.165	7.365	536.190	500.000	7.2
Aroclor-1260-2	7.512	7.418	7.618	518.580	500.000	3.7
Aroclor-1260-3	7.870	7.777	7.977	532.700	500.000	6.5
Aroclor-1260-4	8.095	8.002	8.202	515.410	500.000	3.1
Aroclor-1260-5	8.415	8.322	8.522	534.510	500.000	6.9
Decachlorobiphenyl	10.230	10.137	10.337	54.370	50.000	8.7
Tetrachloro-m-xylene	4.512	4.417	4.617	51.110	50.000	2.2

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL05 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PP071614.D Time Analyzed: 16:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.895	4.798	4.998	517.640	500.000	3.5
Aroclor-1016-2	4.914	4.816	5.016	516.230	500.000	3.2
Aroclor-1016-3	5.090	4.993	5.193	534.980	500.000	7.0
Aroclor-1016-4	5.132	5.036	5.236	540.310	500.000	8.1
Aroclor-1016-5	5.347	5.250	5.450	579.870	500.000	16.0
Aroclor-1260-1	6.381	6.285	6.485	542.610	500.000	8.5
Aroclor-1260-2	6.569	6.473	6.673	540.120	500.000	8.0
Aroclor-1260-3	6.722	6.627	6.827	523.900	500.000	4.8
Aroclor-1260-4	7.193	7.098	7.298	538.880	500.000	7.8
Aroclor-1260-5	7.435	7.339	7.539	527.470	500.000	5.5
Decachlorobiphenyl	8.844	8.749	8.949	54.370	50.000	8.7
Tetrachloro-m-xylene	3.810	3.712	3.912	50.630	50.000	1.3

Analytical Sequence

Client: Kleinfelder	SDG No.: Q1903	
Project: Mitchell School	Instrument ID: ECD_O	
GC Column: ZB-MR1	ID: 0.32 (mm)	Inst. Calib. Date(s): 04/10/2025 04/10/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	04/10/2025	09:17	PO110348.D	8.73	3.69
AR1660ICC1000	AR1660ICC1000	04/10/2025	09:36	PO110349.D	8.73	3.69
AR1660ICC750	AR1660ICC750	04/10/2025	09:54	PO110350.D	8.73	3.69
AR1660ICC500	AR1660ICC500	04/10/2025	10:13	PO110351.D	8.73	3.69
AR1660ICC250	AR1660ICC250	04/10/2025	10:31	PO110352.D	8.73	3.69
AR1660ICC050	AR1660ICC050	04/10/2025	10:49	PO110353.D	8.73	3.69
AR1221ICC500	AR1221ICC500	04/10/2025	11:08	PO110354.D	8.73	3.69
AR1232ICC500	AR1232ICC500	04/10/2025	11:26	PO110355.D	8.73	3.69
AR1242ICC1000	AR1242ICC1000	04/10/2025	11:44	PO110356.D	8.73	3.69
AR1242ICC750	AR1242ICC750	04/10/2025	12:03	PO110357.D	8.73	3.69
AR1242ICC500	AR1242ICC500	04/10/2025	12:21	PO110358.D	8.73	3.69
AR1242ICC250	AR1242ICC250	04/10/2025	12:39	PO110359.D	8.73	3.69
AR1242ICC050	AR1242ICC050	04/10/2025	12:58	PO110360.D	8.73	3.69
AR1248ICC1000	AR1248ICC1000	04/10/2025	13:16	PO110361.D	8.73	3.69
AR1248ICC750	AR1248ICC750	04/10/2025	13:35	PO110362.D	8.73	3.69
AR1248ICC500	AR1248ICC500	04/10/2025	13:53	PO110363.D	8.73	3.69
AR1248ICC250	AR1248ICC250	04/10/2025	14:11	PO110364.D	8.73	3.69
AR1248ICC050	AR1248ICC050	04/10/2025	14:30	PO110365.D	8.73	3.69
AR1254ICC1000	AR1254ICC1000	04/10/2025	14:48	PO110366.D	8.73	3.69
AR1254ICC750	AR1254ICC750	04/10/2025	15:06	PO110367.D	8.73	3.69
AR1254ICC500	AR1254ICC500	04/10/2025	15:25	PO110368.D	8.73	3.69
AR1254ICC250	AR1254ICC250	04/10/2025	15:43	PO110369.D	8.73	3.69
AR1254ICC050	AR1254ICC050	04/10/2025	16:02	PO110370.D	8.73	3.69
AR1262ICC500	AR1262ICC500	04/10/2025	16:20	PO110371.D	8.73	3.69
AR1268ICC1000	AR1268ICC1000	04/10/2025	16:38	PO110372.D	8.73	3.69
AR1268ICC750	AR1268ICC750	04/10/2025	16:57	PO110373.D	8.73	3.69
AR1268ICC500	AR1268ICC500	04/10/2025	17:15	PO110374.D	8.73	3.69
AR1268ICC250	AR1268ICC250	04/10/2025	17:33	PO110375.D	8.73	3.69
AR1268ICC050	AR1268ICC050	04/10/2025	17:52	PO110376.D	8.73	3.69
AR1660CCC500	AR1660CCC500	04/29/2025	09:24	PO110846.D	8.73	3.69
IBLK	IBLK	04/29/2025	10:40	PO110850.D	8.73	3.69
COMP-4	Q1903-01	04/29/2025	14:12	PO110858.D	8.73	3.69
COMP-5	Q1903-02	04/29/2025	14:31	PO110859.D	8.73	3.69
COMP-6	Q1903-03	04/29/2025	14:49	PO110860.D	8.73	3.69
AR1660CCC500	AR1660CCC500	04/29/2025	15:58	PO110861.D	8.73	3.69
IBLK	IBLK	04/29/2025	17:29	PO110865.D	8.73	3.69
VNJ-210MS	Q1904-01MS	04/29/2025	18:05	PO110867.D	8.73	3.68
VNJ-210MSD	Q1904-01MSD	04/29/2025	18:23	PO110868.D	8.73	3.69
AR1660CCC500	AR1660CCC500	04/29/2025	22:00	PO110876.D	8.73	3.69
IBLK	IBLK	04/29/2025	23:31	PO110880.D	8.73	3.69
IBLK	IBLK	04/22/2025	10:13	PP071388.D	10.23	4.51
AR1660ICC1000	AR1660ICC1000	04/22/2025	10:29	PP071389.D	10.23	4.51

Analytical Sequence

AR1660ICC750	AR1660ICC750	04/22/2025	10:45	PP071390.D	10.23	4.51
AR1660ICC500	AR1660ICC500	04/22/2025	11:02	PP071391.D	10.24	4.52
AR1660ICC250	AR1660ICC250	04/22/2025	11:18	PP071392.D	10.23	4.51
AR1660ICC050	AR1660ICC050	04/22/2025	11:34	PP071393.D	10.24	4.52
AR1221ICC500	AR1221ICC500	04/22/2025	11:51	PP071394.D	10.23	4.51
AR1232ICC500	AR1232ICC500	04/22/2025	12:07	PP071395.D	10.23	4.52
AR1242ICC1000	AR1242ICC1000	04/22/2025	12:23	PP071396.D	10.23	4.51
AR1242ICC750	AR1242ICC750	04/22/2025	12:39	PP071397.D	10.24	4.52
AR1242ICC500	AR1242ICC500	04/22/2025	12:56	PP071398.D	10.24	4.51
AR1242ICC250	AR1242ICC250	04/22/2025	13:12	PP071399.D	10.23	4.51
AR1242ICC050	AR1242ICC050	04/22/2025	13:28	PP071400.D	10.24	4.52
AR1248ICC1000	AR1248ICC1000	04/22/2025	13:45	PP071401.D	10.24	4.52
AR1248ICC750	AR1248ICC750	04/22/2025	14:01	PP071402.D	10.24	4.52
AR1248ICC500	AR1248ICC500	04/22/2025	14:17	PP071403.D	10.24	4.51
AR1248ICC250	AR1248ICC250	04/22/2025	14:33	PP071404.D	10.24	4.51
AR1248ICC050	AR1248ICC050	04/22/2025	14:50	PP071405.D	10.24	4.51
AR1254ICC1000	AR1254ICC1000	04/22/2025	15:06	PP071406.D	10.24	4.52
AR1254ICC750	AR1254ICC750	04/22/2025	15:22	PP071407.D	10.24	4.52
AR1254ICC500	AR1254ICC500	04/22/2025	15:38	PP071408.D	10.24	4.52
AR1254ICC250	AR1254ICC250	04/22/2025	15:55	PP071409.D	10.23	4.51
AR1254ICC050	AR1254ICC050	04/22/2025	16:11	PP071410.D	10.24	4.51
AR1262ICC500	AR1262ICC500	04/22/2025	16:27	PP071411.D	10.24	4.52
AR1268ICC1000	AR1268ICC1000	04/22/2025	16:44	PP071412.D	10.24	4.51
AR1268ICC750	AR1268ICC750	04/22/2025	17:00	PP071413.D	10.23	4.52
AR1268ICC500	AR1268ICC500	04/22/2025	17:16	PP071414.D	10.24	4.52
AR1268ICC250	AR1268ICC250	04/22/2025	17:33	PP071415.D	10.24	4.52
AR1268ICC050	AR1268ICC050	04/22/2025	17:49	PP071416.D	10.24	4.51
AR1660CCC500	AR1660CCC500	04/29/2025	10:33	PP071599.D	10.23	4.51
IBLK	IBLK	04/29/2025	11:40	PP071603.D	10.23	4.51
PB167776BL	PB167776BL	04/29/2025	13:30	PP071607.D	10.23	4.51
PB167776BS	PB167776BS	04/29/2025	13:46	PP071608.D	10.23	4.51
AR1660CCC500	AR1660CCC500	04/29/2025	16:29	PP071614.D	10.23	4.51
IBLK	IBLK	04/29/2025	17:51	PP071618.D	10.23	4.51

Analytical Sequence

Client: Kleinfelder

SDG No.: Q1903

Project: Mitchell School

Instrument ID: ECD_O

GC Column: ZB-MR2

ID: 0.32 (mm)

Inst. Calib. Date(s): 04/10/2025

04/10/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	04/10/2025	09:17	PO110348.D	8.68	3.68
AR1660ICC1000	AR1660ICC1000	04/10/2025	09:36	PO110349.D	8.69	3.69
AR1660ICC750	AR1660ICC750	04/10/2025	09:54	PO110350.D	8.69	3.69
AR1660ICC500	AR1660ICC500	04/10/2025	10:13	PO110351.D	8.68	3.69
AR1660ICC250	AR1660ICC250	04/10/2025	10:31	PO110352.D	8.68	3.69
AR1660ICC050	AR1660ICC050	04/10/2025	10:49	PO110353.D	8.68	3.69
AR1221ICC500	AR1221ICC500	04/10/2025	11:08	PO110354.D	8.68	3.69
AR1232ICC500	AR1232ICC500	04/10/2025	11:26	PO110355.D	8.68	3.69
AR1242ICC1000	AR1242ICC1000	04/10/2025	11:44	PO110356.D	8.68	3.69
AR1242ICC750	AR1242ICC750	04/10/2025	12:03	PO110357.D	8.68	3.69
AR1242ICC500	AR1242ICC500	04/10/2025	12:21	PO110358.D	8.68	3.68
AR1242ICC250	AR1242ICC250	04/10/2025	12:39	PO110359.D	8.68	3.69
AR1242ICC050	AR1242ICC050	04/10/2025	12:58	PO110360.D	8.68	3.68
AR1248ICC1000	AR1248ICC1000	04/10/2025	13:16	PO110361.D	8.68	3.69
AR1248ICC750	AR1248ICC750	04/10/2025	13:35	PO110362.D	8.69	3.69
AR1248ICC500	AR1248ICC500	04/10/2025	13:53	PO110363.D	8.69	3.69
AR1248ICC250	AR1248ICC250	04/10/2025	14:11	PO110364.D	8.68	3.69
AR1248ICC050	AR1248ICC050	04/10/2025	14:30	PO110365.D	8.68	3.68
AR1254ICC1000	AR1254ICC1000	04/10/2025	14:48	PO110366.D	8.68	3.69
AR1254ICC750	AR1254ICC750	04/10/2025	15:06	PO110367.D	8.68	3.69
AR1254ICC500	AR1254ICC500	04/10/2025	15:25	PO110368.D	8.68	3.69
AR1254ICC250	AR1254ICC250	04/10/2025	15:43	PO110369.D	8.68	3.68
AR1254ICC050	AR1254ICC050	04/10/2025	16:02	PO110370.D	8.68	3.69
AR1262ICC500	AR1262ICC500	04/10/2025	16:20	PO110371.D	8.68	3.69
AR1268ICC1000	AR1268ICC1000	04/10/2025	16:38	PO110372.D	8.68	3.68
AR1268ICC750	AR1268ICC750	04/10/2025	16:57	PO110373.D	8.68	3.68
AR1268ICC500	AR1268ICC500	04/10/2025	17:15	PO110374.D	8.68	3.69
AR1268ICC250	AR1268ICC250	04/10/2025	17:33	PO110375.D	8.68	3.69
AR1268ICC050	AR1268ICC050	04/10/2025	17:52	PO110376.D	8.68	3.68
AR1660CCC500	AR1660CCC500	04/29/2025	09:24	PO110846.D	8.68	3.68
IBLK	IBLK	04/29/2025	10:40	PO110850.D	8.68	3.68
COMP-4	Q1903-01	04/29/2025	14:12	PO110858.D	8.68	3.68
COMP-5	Q1903-02	04/29/2025	14:31	PO110859.D	8.68	3.68
COMP-6	Q1903-03	04/29/2025	14:49	PO110860.D	8.68	3.68
AR1660CCC500	AR1660CCC500	04/29/2025	15:58	PO110861.D	8.68	3.68
IBLK	IBLK	04/29/2025	17:29	PO110865.D	8.68	3.68
VNJ-210MS	Q1904-01MS	04/29/2025	18:05	PO110867.D	8.68	3.68
VNJ-210MSD	Q1904-01MSD	04/29/2025	18:23	PO110868.D	8.68	3.68
AR1660CCC500	AR1660CCC500	04/29/2025	22:00	PO110876.D	8.68	3.68
IBLK	IBLK	04/29/2025	23:31	PO110880.D	8.68	3.68
IBLK	IBLK	04/22/2025	10:13	PP071388.D	8.85	3.81
AR1660ICC1000	AR1660ICC1000	04/22/2025	10:29	PP071389.D	8.85	3.81

Analytical Sequence

AR1660ICC750	AR1660ICC750	04/22/2025	10:45	PP071390.D	8.85	3.81
AR1660ICC500	AR1660ICC500	04/22/2025	11:02	PP071391.D	8.85	3.81
AR1660ICC250	AR1660ICC250	04/22/2025	11:18	PP071392.D	8.85	3.81
AR1660ICC050	AR1660ICC050	04/22/2025	11:34	PP071393.D	8.85	3.81
AR1221ICC500	AR1221ICC500	04/22/2025	11:51	PP071394.D	8.85	3.81
AR1232ICC500	AR1232ICC500	04/22/2025	12:07	PP071395.D	8.85	3.81
AR1242ICC1000	AR1242ICC1000	04/22/2025	12:23	PP071396.D	8.85	3.81
AR1242ICC750	AR1242ICC750	04/22/2025	12:39	PP071397.D	8.85	3.81
AR1242ICC500	AR1242ICC500	04/22/2025	12:56	PP071398.D	8.85	3.81
AR1242ICC250	AR1242ICC250	04/22/2025	13:12	PP071399.D	8.85	3.81
AR1242ICC050	AR1242ICC050	04/22/2025	13:28	PP071400.D	8.85	3.81
AR1248ICC1000	AR1248ICC1000	04/22/2025	13:45	PP071401.D	8.85	3.81
AR1248ICC750	AR1248ICC750	04/22/2025	14:01	PP071402.D	8.85	3.81
AR1248ICC500	AR1248ICC500	04/22/2025	14:17	PP071403.D	8.85	3.81
AR1248ICC250	AR1248ICC250	04/22/2025	14:33	PP071404.D	8.85	3.81
AR1248ICC050	AR1248ICC050	04/22/2025	14:50	PP071405.D	8.85	3.81
AR1254ICC1000	AR1254ICC1000	04/22/2025	15:06	PP071406.D	8.85	3.81
AR1254ICC750	AR1254ICC750	04/22/2025	15:22	PP071407.D	8.85	3.81
AR1254ICC500	AR1254ICC500	04/22/2025	15:38	PP071408.D	8.85	3.81
AR1254ICC250	AR1254ICC250	04/22/2025	15:55	PP071409.D	8.85	3.81
AR1254ICC050	AR1254ICC050	04/22/2025	16:11	PP071410.D	8.85	3.81
AR1262ICC500	AR1262ICC500	04/22/2025	16:27	PP071411.D	8.85	3.81
AR1268ICC1000	AR1268ICC1000	04/22/2025	16:44	PP071412.D	8.85	3.81
AR1268ICC750	AR1268ICC750	04/22/2025	17:00	PP071413.D	8.85	3.81
AR1268ICC500	AR1268ICC500	04/22/2025	17:16	PP071414.D	8.85	3.81
AR1268ICC250	AR1268ICC250	04/22/2025	17:33	PP071415.D	8.85	3.81
AR1268ICC050	AR1268ICC050	04/22/2025	17:49	PP071416.D	8.85	3.81
AR1660CCC500	AR1660CCC500	04/29/2025	10:33	PP071599.D	8.84	3.81
IBLK	IBLK	04/29/2025	11:40	PP071603.D	8.84	3.81
PB167776BL	PB167776BL	04/29/2025	13:30	PP071607.D	8.84	3.81
PB167776BS	PB167776BS	04/29/2025	13:46	PP071608.D	8.84	3.81
AR1660CCC500	AR1660CCC500	04/29/2025	16:29	PP071614.D	8.84	3.81
IBLK	IBLK	04/29/2025	17:51	PP071618.D	8.84	3.81



QC SAMPLE DATA

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	Mitchell School	Date Received:	
Client Sample ID:	PB167776BL	SDG No.:	Q1903
Lab Sample ID:	PB167776BL	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	100
Sample Wt/Vol:	30.01 Units: g	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071607.D	1	04/29/25 08:35	04/29/25 13:30	PB167776

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	3.90	U	3.90	17.0	ug/kg
11097-69-1	Aroclor-1254	3.20	U	3.20	17.0	ug/kg
11096-82-5	Aroclor-1260	3.20	U	3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.9		32 - 144	100%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.1		32 - 175	105%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/10/25
Project:	Mitchell School	Date Received:	04/10/25
Client Sample ID:	PIBLK-PO110348.D	SDG No.:	Q1903
Lab Sample ID:	I.BLK-PO110348.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	Injection Volume :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110348.D	1		04/10/25	PO041025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.1		60 - 140	91%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.2		60 - 140	96%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/29/25
Project:	Mitchell School	Date Received:	04/29/25
Client Sample ID:	PIBLK-PO110850.D	SDG No.:	Q1903
Lab Sample ID:	I.BLK-PO110850.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110850.D	1		04/29/25	PO042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.6		60 - 140	83%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.7		60 - 140	84%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/29/25
Project:	Mitchell School	Date Received:	04/29/25
Client Sample ID:	PIBLK-PO110865.D	SDG No.:	Q1903
Lab Sample ID:	I.BLK-PO110865.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	Injection Volume :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110865.D	1		04/29/25	PO042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.9		60 - 140	84%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.9		60 - 140	85%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/29/25
Project:	Mitchell School	Date Received:	04/29/25
Client Sample ID:	PIBLK-PO110880.D	SDG No.:	Q1903
Lab Sample ID:	I.BLK-PO110880.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110880.D	1		04/29/25	PO042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.0		60 - 140	85%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.6		60 - 140	83%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/22/25
Project:	Mitchell School	Date Received:	04/22/25
Client Sample ID:	PIBLK-PP071388.D	SDG No.:	Q1903
Lab Sample ID:	I.BLK-PP071388.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	Injection Volume :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071388.D	1		04/22/25	PP042225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.1		60 - 140	85%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.6		60 - 140	88%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/29/25
Project:	Mitchell School	Date Received:	04/29/25
Client Sample ID:	PIBLK-PP071603.D	SDG No.:	Q1903
Lab Sample ID:	I.BLK-PP071603.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071603.D	1		04/29/25	PP042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	15.9		60 - 140	79%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.3		60 - 140	86%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/29/25
Project:	Mitchell School	Date Received:	04/29/25
Client Sample ID:	PIBLK-PP071618.D	SDG No.:	Q1903
Lab Sample ID:	I.BLK-PP071618.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071618.D	1		04/29/25	PP042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.0		60 - 140	80%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.0		60 - 140	90%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	Mitchell School	Date Received:	
Client Sample ID:	PB167776BS	SDG No.:	Q1903
Lab Sample ID:	PB167776BS	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	100
Sample Wt/Vol:	30.03	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071608.D	1	04/29/25 08:35	04/29/25 13:46	PB167776

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	156		3.90	17.0	ug/kg
11097-69-1	Aroclor-1254	3.20	U	3.20	17.0	ug/kg
11096-82-5	Aroclor-1260	157		3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.7		32 - 144	103%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.3		32 - 175	106%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/28/25
Project:	Mitchell School	Date Received:	04/28/25
Client Sample ID:	VNJ-210MS	SDG No.:	Q1903
Lab Sample ID:	Q1904-01MS	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	90.6
Sample Wt/Vol:	30.03	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110867.D	1	04/29/25 08:35	04/29/25 18:05	PB167776

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	201		4.40	18.7	ug/kg
11097-69-1	Aroclor-1254	3.50	U	3.50	18.7	ug/kg
11096-82-5	Aroclor-1260	187		3.60	18.7	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.3		32 - 144	117%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.1		32 - 175	96%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/28/25
Project:	Mitchell School	Date Received:	04/28/25
Client Sample ID:	VNJ-210MSD	SDG No.:	Q1903
Lab Sample ID:	Q1904-01MSD	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	90.6
Sample Wt/Vol:	30.05	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110868.D	1	04/29/25 08:35	04/29/25 18:23	PB167776

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	199		4.40	18.7	ug/kg
11097-69-1	Aroclor-1254	3.50	U	3.50	18.7	ug/kg
11096-82-5	Aroclor-1260	183		3.60	18.7	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.9		32 - 144	114%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.0		32 - 175	95%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit