

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/18/2025 04/18/2025

Continuing Calib Time: 15:31 Initial Calibration Time(s): 13:56 14:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.07	9.08	8.98	9.18	0.01
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00
Aldrin	5.27	5.27	5.17	5.37	0.00
Dieldrin	6.35	6.35	6.25	6.45	0.00
4,4'-DDE	6.20	6.20	6.10	6.30	0.00
4,4'-DDD	6.71	6.71	6.61	6.81	0.00
4,4'-DDT	7.02	7.02	6.92	7.12	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/18/2025 04/18/2025

Continuing Calib Time: 15:31 Initial Calibration Time(s): 13:56 14:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	8.08	8.08	7.98	8.18	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
Aldrin	4.37	4.37	4.27	4.47	0.00
Dieldrin	5.52	5.52	5.42	5.62	0.00
4,4'-DDE	5.38	5.38	5.28	5.48	0.00
4,4'-DDD	5.93	5.93	5.83	6.03	0.00
4,4'-DDT	6.19	6.19	6.09	6.29	0.00