

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL01 Date Analyzed: 04/29/2025

Lab Sample No.: PSTDCCC050 Data File : PD088322.D Time Analyzed: 12:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.709	6.606	6.806	57.420	50.000	14.8
4,4'-DDE	6.200	6.097	6.297	53.620	50.000	7.2
4,4'-DDT	7.025	6.922	7.122	52.040	50.000	4.1
Aldrin	5.276	5.173	5.373	55.870	50.000	11.7
Decachlorobiphenyl	9.078	8.975	9.175	49.320	50.000	-1.4
Dieldrin	6.351	6.249	6.449	56.120	50.000	12.2
Tetrachloro-m-xylene	3.554	3.452	3.652	52.960	50.000	5.9

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL01 Date Analyzed: 04/29/2025

Lab Sample No.: PSTDCCC050 Data File : PD088322.D Time Analyzed: 12:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.932	5.834	6.034	50.260	50.000	0.5
4,4'-DDE	5.377	5.280	5.480	49.820	50.000	-0.4
4,4'-DDT	6.187	6.088	6.288	48.210	50.000	-3.6
Aldrin	4.371	4.273	4.473	51.590	50.000	3.2
Decachlorobiphenyl	8.076	7.977	8.177	45.680	50.000	-8.6
Dieldrin	5.516	5.417	5.617	50.020	50.000	0.0
Tetrachloro-m-xylene	2.881	2.783	2.983	49.290	50.000	-1.4