

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL02 Date Analyzed: 04/29/2025

Lab Sample No.: PSTDCCC050 Data File : PD088334.D Time Analyzed: 15:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.706	6.606	6.806	57.730	50.000	15.5
4,4'-DDE	6.196	6.097	6.297	54.720	50.000	9.4
4,4'-DDT	7.022	6.922	7.122	52.530	50.000	5.1
Aldrin	5.272	5.173	5.373	57.330	50.000	14.7
Decachlorobiphenyl	9.074	8.975	9.175	50.420	50.000	0.8
Dieldrin	6.348	6.249	6.449	57.220	50.000	14.4
Tetrachloro-m-xylene	3.550	3.452	3.652	54.590	50.000	9.2

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL02 Date Analyzed: 04/29/2025

Lab Sample No.: PSTDCCC050 Data File : PD088334.D Time Analyzed: 15:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.933	5.834	6.034	50.950	50.000	1.9
4,4'-DDE	5.377	5.280	5.480	50.580	50.000	1.2
4,4'-DDT	6.187	6.088	6.288	48.480	50.000	-3.0
Aldrin	4.371	4.273	4.473	52.110	50.000	4.2
Decachlorobiphenyl	8.076	7.977	8.177	45.600	50.000	-8.8
Dieldrin	5.516	5.417	5.617	50.770	50.000	1.5
Tetrachloro-m-xylene	2.882	2.783	2.983	50.730	50.000	1.5