

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL03 Date Analyzed: 04/29/2025

Lab Sample No.: PSTDCCC050 Data File : PD088344.D Time Analyzed: 17:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.706	6.606	6.806	57.550	50.000	15.1
4,4'-DDE	6.197	6.097	6.297	53.850	50.000	7.7
4,4'-DDT	7.023	6.922	7.122	51.450	50.000	2.9
Aldrin	5.273	5.173	5.373	57.020	50.000	14.0
Decachlorobiphenyl	9.076	8.975	9.175	48.910	50.000	-2.2
Dieldrin	6.349	6.249	6.449	56.170	50.000	12.3
Tetrachloro-m-xylene	3.550	3.452	3.652	54.930	50.000	9.9

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL03 Date Analyzed: 04/29/2025

Lab Sample No.: PSTDCCC050 Data File : PD088344.D Time Analyzed: 17:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.933	5.834	6.034	50.240	50.000	0.5
4,4'-DDE	5.377	5.280	5.480	49.810	50.000	-0.4
4,4'-DDT	6.187	6.088	6.288	46.040	50.000	-7.9
Aldrin	4.372	4.273	4.473	51.740	50.000	3.5
Decachlorobiphenyl	8.077	7.977	8.177	42.580	50.000	-14.8
Dieldrin	5.516	5.417	5.617	49.830	50.000	-0.3
Tetrachloro-m-xylene	2.882	2.783	2.983	50.630	50.000	1.3