

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
 Data File : PD088139.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2025 17:21
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICPD041825

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 03:53:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 03:51:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.552	2.883	98011498	705.8E6	50.556	50.375
28) SA Decachlor...	9.074	8.076	156.5E6	873.1E6	49.237	49.975
Target Compounds						
2) A alpha-BHC	4.001	3.396	221.4E6	1115.2E6	51.003	50.483
3) MA gamma-BHC...	4.332	3.732	211.2E6	1037.4E6	50.823	50.460
4) MA Heptachlor	4.931	4.086	202.0E6	1034.4E6	50.784	50.400
5) MB Aldrin	5.273	4.372	199.5E6	1010.7E6	50.990	50.536
6) B beta-BHC	4.516	4.028	80388581	446.0E6	50.911	50.415
7) B delta-BHC	4.765	4.265	207.4E6	1030.8E6	50.892	50.347
8) B Heptachlo...	5.693	4.876	177.5E6	911.0E6	50.911	50.388
9) A Endosulfan I	6.076	5.250	167.8E6	870.5E6	50.788	50.475
10) B gamma-Chl...	5.947	5.129	179.7E6	979.0E6	50.819	50.365
11) B alpha-Chl...	6.028	5.194	179.5E6	945.1E6	50.851	50.389
12) B 4,4'-DDE	6.197	5.379	162.4E6	948.3E6	50.132	50.246
13) MA Dieldrin	6.349	5.516	179.4E6	966.1E6	50.821	50.456
14) MA Endrin	6.576	5.792	148.1E6	879.3E6	50.462	50.348
15) B Endosulfa...	6.787	6.084	151.1E6	841.2E6	50.228	50.256
16) A 4,4'-DDD	6.706	5.933	124.5E6	794.4E6	49.895	50.303
17) MA 4,4'-DDT	7.022	6.187	137.1E6	844.1E6	49.781	50.406
18) B Endrin al...	6.916	6.262	112.7E6	640.6E6	50.151	50.415
19) B Endosulfa...	7.150	6.486	140.8E6	824.1E6	50.153	50.192
20) A Methoxychlor	7.494	6.758	73171542	444.8E6	49.850	50.322
21) B Endrin ke...	7.631	6.994	151.9E6	901.7E6	50.003	50.358
22) Mirex	8.115	7.189	112.8E6	707.4E6	50.149	50.451

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
 Data File : PD088139.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2025 17:21
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD041825

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 03:53:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 03:51:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

