

Surrogate Summary

SDG No.: Q1903

Client: Kleinfelder

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD088121.D	PIBLK-PD088121.D	Decachlorobiphenyl	1	20	22.9	115		43	140
		Tetrachloro-m-xylene	1	20	20.1	101		77	126
		Decachlorobiphenyl	2	20	22.7	113		43	140
		Tetrachloro-m-xylene	2	20	20.9	104		77	126
I.BLK-PD088320.D	PIBLK-PD088320.D	Decachlorobiphenyl	1	20	18.7	94		43	140
		Tetrachloro-m-xylene	1	20	18.7	93		77	126
		Decachlorobiphenyl	2	20	17.6	88		43	140
		Tetrachloro-m-xylene	2	20	18.0	90		77	126
PB167777BL	PB167777BL	Decachlorobiphenyl	1	20	18.0	90		20	144
		Tetrachloro-m-xylene	1	20	17.6	88		19	148
		Decachlorobiphenyl	2	20	17.0	85		20	144
		Tetrachloro-m-xylene	2	20	18.1	91		19	148
PB167777BS	PB167777BS	Decachlorobiphenyl	1	20	21.6	108		20	144
		Tetrachloro-m-xylene	1	20	20.8	104		19	148
		Decachlorobiphenyl	2	20	20.5	102		20	144
		Tetrachloro-m-xylene	2	20	21.6	108		19	148
Q1903-01	COMP-4	Decachlorobiphenyl	1	20	19.3	97		20	144
		Tetrachloro-m-xylene	1	20	20.0	100		19	148
		Decachlorobiphenyl	2	20	16.4	82		20	144
		Tetrachloro-m-xylene	2	20	20.0	100		19	148
Q1903-01MS	COMP-4MS	Decachlorobiphenyl	1	20	16.4	82		20	144
		Tetrachloro-m-xylene	1	20	16.7	84		19	148
		Decachlorobiphenyl	2	20	15.2	76		20	144
		Tetrachloro-m-xylene	2	20	17.4	87		19	148
Q1903-01MSD	COMP-4MSD	Decachlorobiphenyl	1	20	16.6	83		20	144
		Tetrachloro-m-xylene	1	20	16.8	84		19	148
		Decachlorobiphenyl	2	20	15.2	76		20	144
		Tetrachloro-m-xylene	2	20	17.5	88		19	148
I.BLK-PD088333.D	PIBLK-PD088333.D	Decachlorobiphenyl	1	20	19.3	97		43	140
		Tetrachloro-m-xylene	1	20	19.2	96		77	126
		Decachlorobiphenyl	2	20	17.9	90		43	140
		Tetrachloro-m-xylene	2	20	19.1	95		77	126
Q1903-02	COMP-5	Decachlorobiphenyl	1	20	16.7	84		20	144
		Tetrachloro-m-xylene	1	20	19.2	96		19	148
		Decachlorobiphenyl	2	20	15.3	76		20	144
		Tetrachloro-m-xylene	2	20	18.9	94		19	148
Q1903-03	COMP-6	Decachlorobiphenyl	1	20	18.5	93		20	144
		Tetrachloro-m-xylene	1	20	20.0	100		19	148
		Decachlorobiphenyl	2	20	17.3	86		20	144
		Tetrachloro-m-xylene	2	20	19.7	99		19	148
I.BLK-PD088343.D	PIBLK-PD088343.D	Decachlorobiphenyl	1	20	18.8	94		43	140

Surrogate Summary

SDG No.: Q1903

Client: Kleinfelder

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD088343.D	PIBLK-PD088343.D	Tetrachloro-m-xylene	1	20	19.3	96		77	126
		Decachlorobiphenyl	2	20	15.8	79		43	140
		Tetrachloro-m-xylene	2	20	18.9	94		77	126