

LAB CHRONICLE

OrderID:	Q1903	OrderDate:	4/28/2025 11:30:00 AM
Client:	Kleinfelder	Project:	Mitchell School
Contact:	Mark Warchol	Location:	L51,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1903-01	COMP-4	SOIL	SVOCMS Group1	8270E	04/25/25	04/29/25	05/01/25	04/28/25
Q1903-02	COMP-5	SOIL	SVOCMS Group1	8270E	04/25/25	04/29/25	05/01/25	04/28/25
Q1903-03	COMP-6	SOIL	SVOCMS Group1	8270E	04/25/25	04/29/25	05/01/25	04/28/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1903
Client: Kleinfelder

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :				0.000				
Total Svoc :					0.00			
Total Concentration:					0.00			



SAMPLE DATA

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/25/25
Project:	Mitchell School	Date Received:	04/28/25
Client Sample ID:	COMP-4	SDG No.:	Q1903
Lab Sample ID:	Q1903-01	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	83.3
Sample Wt/Vol:	30.04 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142262.D	1	04/29/25 09:30	05/01/25 16:03	PB167779

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	27.2	U	27.2	200	ug/Kg
86-73-7	Fluorene	30.3	U	30.3	200	ug/Kg
85-01-8	Phenanthrene	25.1	U	25.1	200	ug/Kg
120-12-7	Anthracene	39.9	U	39.9	200	ug/Kg
129-00-0	Pyrene	43.2	U	43.2	200	ug/Kg
56-55-3	Benzo(a)anthracene	27.6	U	27.6	200	ug/Kg
218-01-9	Chrysene	23.9	U	23.9	200	ug/Kg
205-99-2	Benzo(b)fluoranthene	22.8	U	22.8	200	ug/Kg
50-32-8	Benzo(a)pyrene	35.4	U	35.4	200	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	34.9	U	34.9	200	ug/Kg
191-24-2	Benzo(g,h,i)perylene	30.8	U	30.8	200	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	61.6		18 - 107	62%	SPK: 100
321-60-8	2-Fluorobiphenyl	53.8		20 - 109	54%	SPK: 100
1718-51-0	Terphenyl-d14	40.1		10 - 105	40%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	207000	6.904			
1146-65-2	Naphthalene-d8	782000	8.186			
15067-26-2	Acenaphthene-d10	383000	9.939			
1517-22-2	Phenanthrene-d10	552000	11.427			
1719-03-5	Chrysene-d12	438000	14.068			
1520-96-3	Perylene-d12	441000	15.557			

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/25/25
Project:	Mitchell School	Date Received:	04/28/25
Client Sample ID:	COMP-4	SDG No.:	Q1903
Lab Sample ID:	Q1903-01	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	83.3
Sample Wt/Vol:	30.04	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	SVOCMS Group1
	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	1.0	GPC Cleanup : N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142262.D	1	04/29/25 09:30	05/01/25 16:03	PB167779

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/25/25
Project:	Mitchell School	Date Received:	04/28/25
Client Sample ID:	COMP-5	SDG No.:	Q1903
Lab Sample ID:	Q1903-02	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	78.9
Sample Wt/Vol:	30.02 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142263.D	1	04/29/25 09:30	05/01/25 16:31	PB167779

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
------------	-----------	-------	-----------	-----	------------	-------------------

TARGETS

91-20-3	Naphthalene	28.8	U	28.8	220	ug/Kg
86-73-7	Fluorene	32.0	U	32.0	220	ug/Kg
85-01-8	Phenanthrene	26.5	U	26.5	220	ug/Kg
120-12-7	Anthracene	42.2	U	42.2	220	ug/Kg
129-00-0	Pyrene	45.6	U	45.6	220	ug/Kg
56-55-3	Benzo(a)anthracene	29.1	U	29.1	220	ug/Kg
218-01-9	Chrysene	25.2	U	25.2	220	ug/Kg
205-99-2	Benzo(b)fluoranthene	24.1	U	24.1	220	ug/Kg
50-32-8	Benzo(a)pyrene	37.4	U	37.4	220	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	36.9	U	36.9	220	ug/Kg
191-24-2	Benzo(g,h,i)perylene	32.6	U	32.6	220	ug/Kg

SURROGATES

4165-60-0	Nitrobenzene-d5	61.4		18 - 107	61%	SPK: 100
321-60-8	2-Fluorobiphenyl	53.9		20 - 109	54%	SPK: 100
1718-51-0	Terphenyl-d14	38.3		10 - 105	38%	SPK: 100

INTERNAL STANDARDS

3855-82-1	1,4-Dichlorobenzene-d4	214000	6.904
1146-65-2	Naphthalene-d8	813000	8.187
15067-26-2	Acenaphthene-d10	392000	9.939
1517-22-2	Phenanthrene-d10	556000	11.428
1719-03-5	Chrysene-d12	454000	14.069
1520-96-3	Perylene-d12	452000	15.557

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/25/25
Project:	Mitchell School	Date Received:	04/28/25
Client Sample ID:	COMP-5	SDG No.:	Q1903
Lab Sample ID:	Q1903-02	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	78.9
Sample Wt/Vol:	30.02	Units:	g
Soil Aliquot Vol:			uL
Extraction Type :		Decanted :	N
Injection Volume :		GPC Factor :	1.0
Prep Method :	SW3541	Final Vol:	1000
		Test:	SVOCMS Group1
		Level :	LOW
		GPC Cleanup :	N
		PH :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142263.D	1	04/29/25 09:30	05/01/25 16:31	PB167779

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/25/25
Project:	Mitchell School	Date Received:	04/28/25
Client Sample ID:	COMP-6	SDG No.:	Q1903
Lab Sample ID:	Q1903-03	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	83.9
Sample Wt/Vol:	30.03 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142264.D	1	04/29/25 09:30	05/01/25 17:00	PB167779

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	27.0	U	27.0	200	ug/Kg
86-73-7	Fluorene	30.1	U	30.1	200	ug/Kg
85-01-8	Phenanthrene	24.9	U	24.9	200	ug/Kg
120-12-7	Anthracene	39.7	U	39.7	200	ug/Kg
129-00-0	Pyrene	42.9	U	42.9	200	ug/Kg
56-55-3	Benzo(a)anthracene	27.4	U	27.4	200	ug/Kg
218-01-9	Chrysene	23.7	U	23.7	200	ug/Kg
205-99-2	Benzo(b)fluoranthene	22.6	U	22.6	200	ug/Kg
50-32-8	Benzo(a)pyrene	35.1	U	35.1	200	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	34.6	U	34.6	200	ug/Kg
191-24-2	Benzo(g,h,i)perylene	30.6	U	30.6	200	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	60.3		18 - 107	60%	SPK: 100
321-60-8	2-Fluorobiphenyl	55.6		20 - 109	56%	SPK: 100
1718-51-0	Terphenyl-d14	38.6		10 - 105	39%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	206000	6.904			
1146-65-2	Naphthalene-d8	784000	8.186			
15067-26-2	Acenaphthene-d10	381000	9.939			
1517-22-2	Phenanthrene-d10	530000	11.427			
1719-03-5	Chrysene-d12	453000	14.068			
1520-96-3	Perylene-d12	438000	15.557			

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/25/25
Project:	Mitchell School	Date Received:	04/28/25
Client Sample ID:	COMP-6	SDG No.:	Q1903
Lab Sample ID:	Q1903-03	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	83.9
Sample Wt/Vol:	30.03	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	SVOCMS Group1
	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	1.0	GPC Cleanup :
		N	PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142264.D	1	04/29/25 09:30	05/01/25 17:00	PB167779

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q1903

Client: Kleinfelder

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167779BL	PB167779BL	Nitrobenzene-d5	100	87.4	87		18	107
		2-Fluorobiphenyl	100	88.3	88		20	109
		Terphenyl-d14	100	106	106	*	10	105
PB167779BS	PB167779BS	Nitrobenzene-d5	100	86.4	86		18	107
		2-Fluorobiphenyl	100	87.2	87		20	109
		Terphenyl-d14	100	88.8	89		10	105
Q1901-02MS	B-167-SB01MS	Nitrobenzene-d5	100	68.1	68		18	107
		2-Fluorobiphenyl	100	63.1	63		20	109
		Terphenyl-d14	100	43.9	44		10	105
Q1901-02MSD	B-167-SB01MSD	Nitrobenzene-d5	100	66.9	67		18	107
		2-Fluorobiphenyl	100	61.8	62		20	109
		Terphenyl-d14	100	42.1	42		10	105
Q1903-01	COMP-4	Nitrobenzene-d5	100	61.6	62		18	107
		2-Fluorobiphenyl	100	53.8	54		20	109
		Terphenyl-d14	100	40.1	40		10	105
Q1903-02	COMP-5	Nitrobenzene-d5	100	61.4	61		18	107
		2-Fluorobiphenyl	100	53.9	54		20	109
		Terphenyl-d14	100	38.3	38		10	105
Q1903-03	COMP-6	Nitrobenzene-d5	100	60.3	60		18	107
		2-Fluorobiphenyl	100	55.6	56		20	109
		Terphenyl-d14	100	38.6	39		10	105

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1903

Client: Kleinfelder

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1901-02MS	Client Sample ID:	B-167-SB01MS					DataFile:	BF142269.D		
Naphthalene	2000	0	1800	ug/Kg	90				72	110	
Fluorene	2000	0	1700	ug/Kg	85				68	116	
Phenanthrene	2000	140	2000	ug/Kg	93				52	128	
Anthracene	2000	0	1900	ug/Kg	95				62	124	
Pyrene	2000	130	1400	ug/Kg	64				26	142	
Benzo(a)anthracene	2000	95.0	1900	ug/Kg	90				71	114	
Chrysene	2000	0	1900	ug/Kg	95				57	121	
Benzo(b)fluoranthene	2000	97.3	2000	ug/Kg	95				67	121	
Benzo(a)pyrene	2000	0	1900	ug/Kg	95				70	142	
Indeno(1,2,3-cd)pyrene	2000	0	1400	ug/Kg	70				40	129	
Benzo(g,h,i)perylene	2000	0	1300	ug/Kg	65				24	125	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1903

Client: Kleinfelder

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q1901-02MSD		Client Sample ID: B-167-SB01MSD		DataFile: BF142270.D							
Naphthalene	2000	0	1700	ug/Kg	85		6		72	110	20
Fluorene	2000	0	1700	ug/Kg	85		0		68	116	20
Phenanthrene	2000	140	1900	ug/Kg	88		6		52	128	20
Anthracene	2000	0	1800	ug/Kg	90		5		62	124	20
Pyrene	2000	130	1400	ug/Kg	64		0		26	142	20
Benzo(a)anthracene	2000	95.0	1800	ug/Kg	85		6		71	114	20
Chrysene	2000	0	1800	ug/Kg	90		5		57	121	20
Benzo(b)fluoranthene	2000	97.3	2000	ug/Kg	95		0		67	121	20
Benzo(a)pyrene	2000	0	1900	ug/Kg	95		0		70	142	20
Indeno(1,2,3-cd)pyrene	2000	0	1300	ug/Kg	65		7		40	129	20
Benzo(g,h,i)perylene	2000	0	1200	ug/Kg	60		8		24	125	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1903

Client: Kleinfelder

Analytical Method: 8270E DataFile: BM050064.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB167779BS	Naphthalene	1700	1500	ug/Kg	88				62	100	
	Fluorene	1700	1600	ug/Kg	94				61	101	
	Phenanthrene	1700	1600	ug/Kg	94				59	103	
	Anthracene	1700	1600	ug/Kg	94				61	105	
	Pyrene	1700	1600	ug/Kg	94				59	103	
	Benzo(a)anthracene	1700	1600	ug/Kg	94				60	102	
	Chrysene	1700	1500	ug/Kg	88				59	101	
	Benzo(b)fluoranthene	1700	1600	ug/Kg	94				62	109	
	Benzo(a)pyrene	1700	1600	ug/Kg	94				63	103	
	Indeno(1,2,3-cd)pyrene	1700	1600	ug/Kg	94				63	101	
	Benzo(g,h,i)perylene	1700	1600	ug/Kg	94				70	108	

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167779BL

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM Case No.: Q1903

SAS No.: Q1903 SDG NO.: Q1903

Lab File ID: BM050063.D

Lab Sample ID: PB167779BL

Instrument ID: BNA_M

Date Extracted: 04/29/2025

Matrix: (soil/water) SOIL

Date Analyzed: 05/01/2025

Level: (low/med) LOW

Time Analyzed: 14:04

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB167779BS	PB167779BS	BM050064.D	05/01/2025
B-167-SB01MS	Q1901-02MS	BF142269.D	05/01/2025
B-167-SB01MSD	Q1901-02MSD	BF142270.D	05/01/2025
COMP-4	Q1903-01	BF142262.D	05/01/2025
COMP-5	Q1903-02	BF142263.D	05/01/2025
COMP-6	Q1903-03	BF142264.D	05/01/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM

SAS No.: Q1903 SDG NO.: Q1903

Lab File ID: BF142238.D

DFTPP Injection Date: 04/30/2025

Instrument ID: BNA_F

DFTPP Injection Time: 10:55

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	32.6
68	Less than 2.0% of mass 69	0.5 (1.8) 1
69	Mass 69 relative abundance	27.3
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	37.1
197	Less than 2.0% of mass 198	0.7
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.3
275	10.0 - 60.0% of mass 198	22.8
365	Greater than 1% of mass 198	3
441	Present, but less than mass 443	15.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.9 (19.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF142239.D	04/30/2025	11:24
SSTDICC005	SSTDICC005	BF142240.D	04/30/2025	11:52
SSTDICC010	SSTDICC010	BF142241.D	04/30/2025	12:20
SSTDICC020	SSTDICC020	BF142242.D	04/30/2025	12:49
SSTDICCC040	SSTDICCC040	BF142243.D	04/30/2025	13:17
SSTDICC050	SSTDICC050	BF142244.D	04/30/2025	13:46
SSTDICC060	SSTDICC060	BF142245.D	04/30/2025	14:15
SSTDICC080	SSTDICC080	BF142246.D	04/30/2025	14:43

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM

SAS No.: Q1903 SDG NO.: Q1903

Lab File ID: BF142249.D

DFTPP Injection Date: 05/01/2025

Instrument ID: BNA_F

DFTPP Injection Time: 09:48

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	39.4
68	Less than 2.0% of mass 69	0.5 (1.6) 1
69	Mass 69 relative abundance	31.6
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	42.4
197	Less than 2.0% of mass 198	0.6
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.9
275	10.0 - 60.0% of mass 198	23.8
365	Greater than 1% of mass 198	3.1
441	Present, but less than mass 443	15.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.1 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142250.D	05/01/2025	10:17
COMP-4	Q1903-01	BF142262.D	05/01/2025	16:03
COMP-5	Q1903-02	BF142263.D	05/01/2025	16:31
COMP-6	Q1903-03	BF142264.D	05/01/2025	17:00
B-167-SB01MS	Q1901-02MS	BF142269.D	05/01/2025	19:24
B-167-SB01MSD	Q1901-02MSD	BF142270.D	05/01/2025	19:52

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM

SAS No.: Q1903 SDG NO.: Q1903

Lab File ID: BM050023.D

DFTPP Injection Date: 04/28/2025

Instrument ID: BNA_M

DFTPP Injection Time: 11:46

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	23.8
68	Less than 2.0% of mass 69	0.4 (1.3) 1
69	Mass 69 relative abundance	28.5
70	Less than 2.0% of mass 69	0.1 (0.4) 1
127	10.0 - 80.0% of mass 198	35.7
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	26.2
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	11.6
442	Greater than 50% of mass 198	74.7
443	15.0 - 24.0% of mass 442	14.4 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BM050024.D	04/28/2025	12:30
SSTDICC005	SSTDICC005	BM050025.D	04/28/2025	13:09
SSTDICC010	SSTDICC010	BM050026.D	04/28/2025	13:48
SSTDICC020	SSTDICC020	BM050027.D	04/28/2025	14:27
SSTDICCC040	SSTDICCC040	BM050028.D	04/28/2025	15:06
SSTDICC050	SSTDICC050	BM050029.D	04/28/2025	15:45
SSTDICC060	SSTDICC060	BM050030.D	04/28/2025	16:24
SSTDICC080	SSTDICC080	BM050031.D	04/28/2025	17:04

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM

SAS No.: Q1903

SDG NO.: Q1903

Lab File ID: BM050061.D

DFTPP Injection Date: 05/01/2025

Instrument ID: BNA_M

DFTPP Injection Time: 12:24

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	22.1
68	Less than 2.0% of mass 69	0.4 (1.5) 1
69	Mass 69 relative abundance	26.4
70	Less than 2.0% of mass 69	0.1 (0.4) 1
127	10.0 - 80.0% of mass 198	34.3
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	27
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	12.6
442	Greater than 50% of mass 198	81.4
443	15.0 - 24.0% of mass 442	15.5 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM050062.D	05/01/2025	13:25
PB167779BL	PB167779BL	BM050063.D	05/01/2025	14:04
PB167779BS	PB167779BS	BM050064.D	05/01/2025	14:43

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/01/2025

Lab File ID: BF142250.D Time Analyzed: 10:17

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	186578	6.91	719586	8.19	381198	9.95
UPPER LIMIT	373156	7.41	1439170	8.692	762396	10.445
LOWER LIMIT	93289	6.41	359793	7.692	190599	9.445
EPA SAMPLE NO.						
01 B-167-SB01MS	207729	6.91	776941	8.19	359640	9.95
02 B-167-SB01MSD	204570	6.91	762454	8.19	351872	9.95
03 COMP-4	206685	6.90	782390	8.19	383432	9.94
04 COMP-5	213887	6.90	812832	8.19	392078	9.94
05 COMP-6	206274	6.90	784232	8.19	381114	9.94

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/01/2025

Lab File ID: BF142250.D Time Analyzed: 10:17

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	647816	11.433	384442	14.074	349272	15.562
UPPER LIMIT	1295630	11.933	768884	14.574	698544	16.062
LOWER LIMIT	323908	10.933	192221	13.574	174636	15.062
EPA SAMPLE NO.						
01 B-167-SB01MS	495470	11.43	427878	14.07	472653	15.56
02 B-167-SB01MSD	496148	11.43	429879	14.07	473333	15.56
03 COMP-4	552303	11.43	438090	14.07	441251	15.56
04 COMP-5	556035	11.43	454222	14.07	451548	15.56
05 COMP-6	529947	11.43	453109	14.07	437720	15.56

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/01/2025

Lab File ID: BM050062.D Time Analyzed: 13:25

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	329367	7.757	1090200	10.55	636680	14.40
UPPER LIMIT	658734	8.257	2180400	11.045	1273360	14.898
LOWER LIMIT	164684	7.257	545100	10.045	318340	13.898
EPA SAMPLE NO.						
01 PB167779BL	362102	7.75	1207360	10.55	720688	14.40
02 PB167779BS	339452	7.75	1125330	10.55	664515	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG NO.: Q1903

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/01/2025

Lab File ID: BM050062.D Time Analyzed: 13:25

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1171490	17.145	1137060	21.386	1050480	24.374
UPPER LIMIT	2342980	17.645	2274120	21.886	2100960	24.874
LOWER LIMIT	585745	16.645	568530	20.886	525240	23.874
EPA SAMPLE NO.						
01 PB167779BL	1342350	17.15	1132570	21.39	1130010	24.37
02 PB167779BS	1227380	17.14	1219130	21.38	1104890	24.37

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	Mitchell School	Date Received:	
Client Sample ID:	PB167779BL	SDG No.:	Q1903
Lab Sample ID:	PB167779BL	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	100
Sample Wt/Vol:	30.01 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050063.D	1	04/29/25 09:30	05/01/25 14:04	PB167779

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	22.7	U	22.7	170	ug/Kg
86-73-7	Fluorene	25.3	U	25.3	170	ug/Kg
85-01-8	Phenanthrene	20.9	U	20.9	170	ug/Kg
120-12-7	Anthracene	33.3	U	33.3	170	ug/Kg
129-00-0	Pyrene	36.0	U	36.0	170	ug/Kg
56-55-3	Benzo(a)anthracene	23.0	U	23.0	170	ug/Kg
218-01-9	Chrysene	19.9	U	19.9	170	ug/Kg
205-99-2	Benzo(b)fluoranthene	19.0	U	19.0	170	ug/Kg
50-32-8	Benzo(a)pyrene	29.5	U	29.5	170	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	29.1	U	29.1	170	ug/Kg
191-24-2	Benzo(g,h,i)perylene	25.7	U	25.7	170	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	87.4		18 - 107	87%	SPK: 100
321-60-8	2-Fluorobiphenyl	88.3		20 - 109	88%	SPK: 100
1718-51-0	Terphenyl-d14	106	*	10 - 105	106%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	362000	7.751			
1146-65-2	Naphthalene-d8	1210000	10.545			
15067-26-2	Acenaphthene-d10	721000	14.398			
1517-22-2	Phenanthrene-d10	1340000	17.145			
1719-03-5	Chrysene-d12	1130000	21.386			
1520-96-3	Perylene-d12	1130000	24.374			

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	Mitchell School	Date Received:	
Client Sample ID:	PB167779BL	SDG No.:	Q1903
Lab Sample ID:	PB167779BL	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	100
Sample Wt/Vol:	30.01 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050063.D	1	04/29/25 09:30	05/01/25 14:04	PB167779

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	Mitchell School	Date Received:	
Client Sample ID:	PB167779BS	SDG No.:	Q1903
Lab Sample ID:	PB167779BS	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	100
Sample Wt/Vol:	30.03 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050064.D	1	04/29/25 09:30	05/01/25 14:43	PB167779

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
------------	-----------	-------	-----------	-----	------------	-------------------

TARGETS

91-20-3	Naphthalene	1500		22.7	170	ug/Kg
86-73-7	Fluorene	1600		25.3	170	ug/Kg
85-01-8	Phenanthrene	1600		20.9	170	ug/Kg
120-12-7	Anthracene	1600		33.3	170	ug/Kg
129-00-0	Pyrene	1600		36.0	170	ug/Kg
56-55-3	Benzo(a)anthracene	1600		23.0	170	ug/Kg
218-01-9	Chrysene	1500		19.9	170	ug/Kg
205-99-2	Benzo(b)fluoranthene	1600		19.0	170	ug/Kg
50-32-8	Benzo(a)pyrene	1600		29.5	170	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	1600		29.1	170	ug/Kg
191-24-2	Benzo(g,h,i)perylene	1600		25.7	170	ug/Kg

SURROGATES

4165-60-0	Nitrobenzene-d5	86.4		18 - 107	86%	SPK: 100
321-60-8	2-Fluorobiphenyl	87.2		20 - 109	87%	SPK: 100
1718-51-0	Terphenyl-d14	88.8		10 - 105	89%	SPK: 100

INTERNAL STANDARDS

3855-82-1	1,4-Dichlorobenzene-d4	339000	7.751
1146-65-2	Naphthalene-d8	1130000	10.545
15067-26-2	Acenaphthene-d10	665000	14.392
1517-22-2	Phenanthrene-d10	1230000	17.139
1719-03-5	Chrysene-d12	1220000	21.38
1520-96-3	Perylene-d12	1100000	24.374

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	Mitchell School	Date Received:	
Client Sample ID:	PB167779BS	SDG No.:	Q1903
Lab Sample ID:	PB167779BS	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	100
Sample Wt/Vol:	30.03 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:		Test:	SVOCMS Group1
Extraction Type :		Decanted :	N
Injection Volume :		Level :	LOW
		GPC Factor :	1.0
		GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050064.D	1	04/29/25 09:30	05/01/25 14:43	PB167779

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/26/25
Project:	Mitchell School	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MS	SDG No.:	Q1903
Lab Sample ID:	Q1901-02MS	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	82.6
Sample Wt/Vol:	30.08 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142269.D	1	04/29/25 09:30	05/01/25 19:24	PB167779

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
------------	-----------	-------	-----------	-----	------------	-------------------

TARGETS

91-20-3	Naphthalene	1800		27.4	210	ug/Kg
86-73-7	Fluorene	1700		30.5	210	ug/Kg
85-01-8	Phenanthrene	2000		25.2	210	ug/Kg
120-12-7	Anthracene	1900		40.2	210	ug/Kg
129-00-0	Pyrene	1400		43.5	210	ug/Kg
56-55-3	Benzo(a)anthracene	1900		27.8	210	ug/Kg
218-01-9	Chrysene	1900		24.0	210	ug/Kg
205-99-2	Benzo(b)fluoranthene	2000		22.9	210	ug/Kg
50-32-8	Benzo(a)pyrene	1900		35.6	210	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	1400		35.1	210	ug/Kg
191-24-2	Benzo(g,h,i)perylene	1300		31.0	210	ug/Kg

SURROGATES

4165-60-0	Nitrobenzene-d5	68.1		18 - 107	68%	SPK: 100
321-60-8	2-Fluorobiphenyl	63.1		20 - 109	63%	SPK: 100
1718-51-0	Terphenyl-d14	43.9		10 - 105	44%	SPK: 100

INTERNAL STANDARDS

3855-82-1	1,4-Dichlorobenzene-d4	208000	6.91
1146-65-2	Naphthalene-d8	777000	8.192
15067-26-2	Acenaphthene-d10	360000	9.945
1517-22-2	Phenanthrene-d10	495000	11.427
1719-03-5	Chrysene-d12	428000	14.074
1520-96-3	Perylene-d12	473000	15.562

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/26/25
Project:	Mitchell School	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MS	SDG No.:	Q1903
Lab Sample ID:	Q1901-02MS	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	82.6
Sample Wt/Vol:	30.08 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142269.D	1	04/29/25 09:30	05/01/25 19:24	PB167779

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/26/25
Project:	Mitchell School	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MSD	SDG No.:	Q1903
Lab Sample ID:	Q1901-02MSD	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	82.6
Sample Wt/Vol:	30.06 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142270.D	1	04/29/25 09:30	05/01/25 19:52	PB167779

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
------------	-----------	-------	-----------	-----	------------	-------------------

TARGETS

91-20-3	Naphthalene	1700		27.4	210	ug/Kg
86-73-7	Fluorene	1700		30.6	210	ug/Kg
85-01-8	Phenanthrene	1900		25.3	210	ug/Kg
120-12-7	Anthracene	1800		40.2	210	ug/Kg
129-00-0	Pyrene	1400		43.5	210	ug/Kg
56-55-3	Benzo(a)anthracene	1800		27.8	210	ug/Kg
218-01-9	Chrysene	1800		24.0	210	ug/Kg
205-99-2	Benzo(b)fluoranthene	2000		23.0	210	ug/Kg
50-32-8	Benzo(a)pyrene	1900		35.6	210	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	1300		35.2	210	ug/Kg
191-24-2	Benzo(g,h,i)perylene	1200		31.1	210	ug/Kg

SURROGATES

4165-60-0	Nitrobenzene-d5	66.9		18 - 107	67%	SPK: 100
321-60-8	2-Fluorobiphenyl	61.8		20 - 109	62%	SPK: 100
1718-51-0	Terphenyl-d14	42.1		10 - 105	42%	SPK: 100

INTERNAL STANDARDS

3855-82-1	1,4-Dichlorobenzene-d4	205000	6.91
1146-65-2	Naphthalene-d8	762000	8.187
15067-26-2	Acenaphthene-d10	352000	9.945
1517-22-2	Phenanthrene-d10	496000	11.428
1719-03-5	Chrysene-d12	430000	14.069
1520-96-3	Perylene-d12	473000	15.557

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/26/25
Project:	Mitchell School	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MSD	SDG No.:	Q1903
Lab Sample ID:	Q1901-02MSD	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	82.6
Sample Wt/Vol:	30.06 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142270.D	1	04/29/25 09:30	05/01/25 19:52	PB167779

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF043025.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Apr 30 16:00:01 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BF142239.D 5 =BF142240.D 10 =BF142241.D 20 =BF142242.D 40 =BF142243.D 50 =BF142244.D 60 =BF142245.D 80 =BF142246.D

Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD

1) I 1,4-Dichlorobenzen...	-----ISTD-----									
2) 1,4-Dioxane	0.563	0.534	0.561	0.592	0.580	0.555	0.530	0.559	4.03	
3) Pyridine	1.465	1.391	1.454	1.522	1.503	1.460	1.380	1.454	3.62	
4) n-Nitrosodimet...	0.747	0.699	0.728	0.807	0.789	0.766	0.731	0.752	4.95	
5) S 2-Fluorophenol	1.324	1.238	1.265	1.314	1.285	1.210	1.131	1.252	5.35	
6) Aniline	2.212	2.085	2.157	2.245	2.203	2.101	1.981	2.140	4.29	
7) S Phenol-d6	1.559	1.505	1.534	1.593	1.550	1.475	1.370	1.512	4.85	
8) 2-Chlorophenol	1.252	1.232	1.304	1.379	1.354	1.313	1.225	1.294	4.64	
9) Benzaldehyde	1.154	1.088	1.095	1.117	1.078	0.993	0.875	1.057	8.91	
10) C Phenol	1.679	1.588	1.657	1.715	1.687	1.603	1.468	1.628	5.15	
11) bis(2-Chloroet...	1.345	1.250	1.270	1.318	1.285	1.231	1.132	1.262	5.46	
12) 1,3-Dichlorobe...	1.571	1.485	1.485	1.533	1.478	1.396	1.299	1.464	6.18	
13) C 1,4-Dichlorobe...	1.610	1.486	1.495	1.546	1.488	1.404	1.295	1.475	6.86	
14) 1,2-Dichlorobe...	1.489	1.418	1.431	1.475	1.432	1.356	1.255	1.408	5.67	
15) Benzyl Alcohol	1.047	1.017	1.064	1.149	1.126	1.082	1.023	1.073	4.69	
16) 2,2'-oxybis(1-...	2.368	2.287	2.303	2.386	2.344	2.216	2.043	2.278	5.19	
17) 2-Methylphenol	1.085	1.029	1.068	1.121	1.084	1.050	0.983	1.060	4.19	
18) Hexachloroethane	0.468	0.466	0.481	0.511	0.514	0.489	0.451	0.483	4.84	
19) P n-Nitroso-di-n...	0.953	0.987	0.937	0.949	0.967	0.939	0.898	0.844	0.934	4.77
20) 3+4-Methylphenols	1.357	1.323	1.372	1.400	1.366	1.296	1.166	1.326	5.90	
21) I Naphthalene-d8	-----ISTD-----									
22) Acetophenone	0.511	0.482	0.483	0.468	0.461	0.430	0.404	0.463	7.70	
23) S Nitrobenzene-d5	0.217	0.251	0.302	0.340	0.351	0.339	0.330	0.304	16.88	
24) Nitrobenzene	0.229	0.261	0.297	0.328	0.338	0.323	0.314	0.299	13.31	
25) Isophorone	0.660	0.630	0.642	0.660	0.654	0.626	0.601	0.639	3.36	
26) C 2-Nitrophenol	0.048	0.057	0.083	0.115	0.132	0.136	0.142	0.102	38.38#	
27) 2,4-Dimethylph...	0.315	0.319	0.333	0.337	0.333	0.315	0.302	0.322	3.96	
28) bis(2-Chloroet...	0.425	0.404	0.414	0.415	0.410	0.386	0.366	0.403	4.98	
29) C 2,4-Dichloroph...	0.246	0.253	0.274	0.291	0.286	0.272	0.263	0.269	6.12	
30) 1,2,4-Trichlor...	0.320	0.303	0.305	0.312	0.304	0.287	0.273	0.301	5.28	
31) Naphthalene	1.106	1.040	1.035	1.020	0.999	0.924	0.858	0.997	8.21	
32) Benzoic acid		0.048	0.087	0.130	0.146	0.159	0.179	0.125	39.24	
33) 4-Chloroaniline	0.446	0.414	0.429	0.429	0.423	0.396	0.371	0.415	5.98	
34) C Hexachlorobuta...	0.180	0.175	0.179	0.185	0.178	0.170	0.161	0.175	4.35	
35) Caprolactam	0.068	0.073	0.081	0.087	0.089	0.085	0.083	0.081	9.45	
36) C 4-Chloro-3-met...	0.262	0.265	0.279	0.295	0.293	0.276	0.266	0.277	4.84	
37) 2-Methylnaphth...	0.679	0.637	0.633	0.623	0.612	0.574	0.534	0.613	7.69	
38) 1-Methylnaphth...	0.700	0.655	0.667	0.652	0.641	0.590	0.542	0.635	8.29	

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF043025.M

39) I	Acenaphthene-d10	-----ISTD-----								
40)	1,2,4,5-Tetrac...	0.621	0.575	0.572	0.576	0.566	0.534	0.496	0.563	6.88
41) P	Hexachlorocycl...	0.244	0.265	0.306	0.352	0.368	0.358	0.343	0.319	15.23
42) S	2,4,6-Tribromo...	0.139	0.158	0.184	0.204	0.205	0.197	0.190	0.182	13.73
43) C	2,4,6-Trichlor...	0.275	0.329	0.353	0.370	0.375	0.381	0.357	0.349	10.50
44)	2,4,5-Trichlor...	0.322	0.330	0.362	0.404	0.411	0.371	0.365	0.367	9.17
45) S	2-Fluorobiphenyl	1.738	1.580	1.520	1.409	1.347	1.221	1.114	1.418	15.11
46)	1,1'-Biphenyl	1.773	1.621	1.612	1.581	1.547	1.435	1.316	1.555	9.38
47)	2-Chloronaphth...	1.289	1.176	1.191	1.185	1.157	1.085	1.014	1.157	7.51
48)	2-Nitroaniline	0.119	0.156	0.228	0.295	0.316	0.315	0.316	0.249	33.25
49)	Acenaphthylene	2.149	2.025	2.025	1.997	1.965	1.828	1.675	1.952	7.93
50)	Dimethylphthalate	1.362	1.284	1.301	1.321	1.309	1.227	1.152	1.280	5.42
51)	2,6-Dinitrotol...	0.098	0.137	0.193	0.241	0.259	0.254	0.249	0.204	31.45
52) C	Acenaphthene	1.271	1.175	1.156	1.172	1.138	1.072	0.995	1.140	7.63
53)	3-Nitroaniline	0.143	0.187	0.252	0.298	0.311	0.310	0.299	0.257	26.09
54) P	2,4-Dinitrophenol		0.025	0.036	0.057	0.069	0.076	0.085	0.058	40.36
55)	Dibenzofuran	1.930	1.775	1.751	1.743	1.713	1.581	1.466	1.709	8.67
56) P	4-Nitrophenol		0.131	0.177	0.216	0.233	0.230	0.220	0.201	19.82
57)	2,4-Dinitrotol...	0.104	0.145	0.213	0.285	0.311	0.309	0.306	0.239	35.96
58)	Fluorene	1.493	1.403	1.351	1.317	1.273	1.186	1.071	1.299	10.74
59)	2,3,4,6-Tetrac...	0.252	0.267	0.312	0.334	0.334	0.323	0.305	0.304	10.65
60)	Diethylphthalate	1.324	1.264	1.302	1.313	1.300	1.215	1.126	1.263	5.61
61)	4-Chlorophenyl...	0.702	0.642	0.632	0.632	0.614	0.570	0.526	0.617	9.06
62)	4-Nitroaniline	0.144	0.177	0.233	0.272	0.287	0.281	0.269	0.238	23.57
63)	Azobenzene	1.346	1.260	1.275	1.279	1.278	1.198	1.098	1.248	6.33
64) I	Phenanthrene-d10	-----ISTD-----								
65)	4,6-Dinitro-2-...		0.022	0.034	0.055	0.065	0.071	0.075	0.054	39.23
66) c	n-Nitrosodiphe...	0.727	0.686	0.693	0.712	0.689	0.653	0.623	0.683	5.11
67)	4-Bromophenyl-...	0.224	0.219	0.223	0.236	0.225	0.220	0.211	0.223	3.39
68)	Hexachlorobenzene	0.256	0.245	0.248	0.261	0.255	0.244	0.237	0.249	3.28
69)	Atrazine	0.166	0.172	0.186	0.190	0.190	0.182	0.174	0.180	5.29
70) C	Pentachlorophenol		0.093	0.121	0.144	0.145	0.144	0.144	0.132	16.25
71)	Phenanthrene	1.218	1.114	1.119	1.112	1.064	0.995	0.940	1.080	8.45
72)	Anthracene	1.209	1.161	1.151	1.128	1.099	1.025	0.959	1.105	7.77
73)	Carbazole	1.125	1.047	1.061	1.033	0.996	0.935	0.857	1.008	8.77
74)	Di-n-butylphth...	0.986	1.006	1.088	1.069	1.045	0.988	0.912	1.013	5.88
75) C	Fluoranthene	1.221	1.144	1.155	1.085	1.035	0.957	0.878	1.068	11.24
76) I	Chrysene-d12	-----ISTD-----								
77)	Benzidine	0.702	0.748	0.885	0.963	0.945	0.895	0.810	0.850	11.67
78)	Pyrene	1.833	1.801	1.849	2.026	1.937	1.833	1.635	1.845	6.54
79) S	Terphenyl-d14	1.472	1.413	1.422	1.481	1.392	1.290	1.155	1.375	8.41
80)	Butylbenzylphth...	0.218	0.293	0.405	0.505	0.525	0.524	0.509	0.425	29.32
81)	Benzo(a)anthra...	1.379	1.320	1.332	1.438	1.388	1.316	1.225	1.343	5.06
82)	3,3'-Dichlorob...	0.286	0.317	0.380	0.442	0.450	0.451	0.435	0.394	17.37
83)	Chrysene	1.302	1.210	1.249	1.232	1.227	1.197	1.138	1.222	4.12
84)	Bis(2-ethylhex...	0.384	0.448	0.549	0.695	0.726	0.722	0.707	0.604	23.69
85) c	Di-n-octyl pht...		0.668	0.875	1.262	1.320	1.347	1.329	1.133	25.55

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF043025.M

86) I	Perylene-d12	-----ISTD-----									
87)	Indeno(1,2,3-c...	1.323	1.384	1.466	1.613	1.556	1.490	1.422	1.465		6.78
88)	Benzo(b)fluora...	1.317	1.220	1.208	1.302	1.271	1.177	1.202	1.242		4.36
89)	Benzo(k)fluora...	1.211	1.159	1.222	1.145	1.146	1.110	0.962	1.136		7.59
90) C	Benzo(a)pyrene	1.121	1.098	1.146	1.203	1.180	1.137	1.076	1.137		3.91
91)	Dibenzo(a,h)an...	1.075	1.131	1.197	1.312	1.271	1.192	1.144	1.189		6.90
92)	Benzo(g,h,i)pe...	1.076	1.138	1.204	1.308	1.267	1.213	1.162	1.195		6.55

(#) = Out of Range

Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
 Method File : 8270-BM042825.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Apr 28 18:09:16 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BM050024.D 5 =BM050025.D 10 =BM050026.D 20 =BM050027.D 40 =BM050028.D 50 =BM050029.D 60 =BM050030.D 80 =BM050031.D

Compound		2.5	5	10	20	40	50	60	80	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----									
2)	1,4-Dioxane	0.506	0.464	0.478	0.511	0.504	0.491	0.473	0.490		3.75
3)	Pyridine	1.201	1.166	1.229	1.331	1.328	1.292	1.261	1.258		5.03
4)	n-Nitrosodimet...	0.487	0.459	0.475	0.522	0.518	0.504	0.490	0.493		4.62
5) S	2-Fluorophenol	1.074	1.079	1.107	1.208	1.211	1.173	1.142	1.142		5.04
6)	Aniline	1.657	1.689	1.768	1.942	1.922	1.875	1.835	1.813		6.16
7) S	Phenol-d6	1.290	1.318	1.398	1.543	1.537	1.494	1.469	1.436		7.13
8)	2-Chlorophenol	1.169	1.161	1.198	1.312	1.301	1.268	1.235	1.235		4.96
9)	Benzaldehyde	0.940	0.920	0.971	1.031	1.013	0.956	0.894	0.961		5.08
10) C	Phenol	1.349	1.344	1.420	1.539	1.537	1.501	1.483	1.453		5.73
11)	bis(2-Chloroet...	1.213	1.105	1.176	1.279	1.271	1.233	1.218	1.213		4.90
12)	1,3-Dichlorobe...	1.483	1.428	1.453	1.566	1.552	1.508	1.453	1.492		3.51
13) C	1,4-Dichlorobe...	1.518	1.418	1.463	1.561	1.557	1.516	1.458	1.499		3.60
14)	1,2-Dichlorobe...	1.420	1.393	1.423	1.522	1.506	1.464	1.405	1.448		3.50
15)	Benzyl Alcohol	0.866	0.881	0.966	1.080	1.074	1.050	1.050	0.996		9.16
16)	2,2'-oxybis(1-...	1.487	1.422	1.445	1.552	1.513	1.473	1.433	1.475		3.17
17)	2-Methylphenol	0.857	0.893	0.931	1.033	1.026	0.988	0.975	0.957		6.96
18)	Hexachloroethane	0.552	0.508	0.521	0.556	0.548	0.533	0.510	0.533		3.79
19) P	n-Nitroso-di-n...	0.802	0.889	0.885	0.921	1.003	0.988	0.949	0.928	0.921	6.95
20)	3+4-Methylphenols	1.137	1.170	1.267	1.393	1.392	1.349	1.341	1.293		8.09
21) I	Naphthalene-d8	-----ISTD-----									
22)	Acetophenone	0.483	0.464	0.485	0.531	0.525	0.506	0.487	0.497		4.94
23) S	Nitrobenzene-d5	0.386	0.374	0.397	0.434	0.433	0.417	0.400	0.406		5.68
24)	Nitrobenzene	0.337	0.325	0.345	0.374	0.372	0.358	0.346	0.351		5.16
25)	Isophorone	0.666	0.648	0.680	0.729	0.729	0.704	0.687	0.692		4.44
26) C	2-Nitrophenol	0.148	0.157	0.172	0.196	0.198	0.192	0.191	0.179		11.33
27)	2,4-Dimethylph...	0.263	0.266	0.289	0.324	0.322	0.312	0.305	0.297		8.42
28)	bis(2-Chloroet...	0.407	0.399	0.422	0.455	0.452	0.434	0.425	0.428		4.92
29) C	2,4-Dichloroph...	0.294	0.298	0.326	0.362	0.359	0.351	0.342	0.333		8.47
30)	1,2,4-Trichlor...	0.400	0.384	0.394	0.427	0.426	0.412	0.399	0.406		4.02
31)	Naphthalene	1.001	0.959	0.993	1.066	1.057	1.023	0.989	1.013		3.80
32)	Benzoic acid		0.138	0.172	0.197	0.206	0.212	0.215	0.190		15.67
33)	4-Chloroaniline	0.361	0.393	0.404	0.441	0.448	0.437	0.423	0.415		7.51
34) C	Hexachlorobuta...	0.253	0.240	0.249	0.270	0.270	0.263	0.258	0.258		4.34
35)	Caprolactam	0.090	0.085	0.095	0.104	0.104	0.102	0.100	0.097		7.73
36) C	4-Chloro-3-met...	0.289	0.283	0.302	0.325	0.329	0.317	0.309	0.308		5.72
37)	2-Methylnaphth...	0.656	0.639	0.662	0.713	0.714	0.692	0.678	0.679		4.22
38)	1-Methylnaphth...	0.703	0.675	0.708	0.753	0.751	0.731	0.710	0.719		3.92

Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
 Method File : 8270-BM042825.M

39) I	Acenaphthene-d10	-----ISTD-----									
40)	1,2,4,5-Tetrac...	0.624	0.622	0.655	0.742	0.745	0.725	0.733	0.692	8.17	
41) P	Hexachlorocycl...		0.303	0.362	0.455	0.468	0.467	0.483	0.423	17.30	
42) S	2,4,6-Tribromo...	0.253	0.249	0.281	0.318	0.329	0.318	0.323	0.296	11.61	
43) C	2,4,6-Trichlor...	0.372	0.381	0.423	0.476	0.485	0.467	0.469	0.439	10.76	
44)	2,4,5-Trichlor...	0.410	0.417	0.453	0.517	0.519	0.501	0.497	0.474	9.83	
45) S	2-Fluorobiphenyl	1.588	1.555	1.632	1.807	1.822	1.744	1.688	1.691	6.21	
46)	1,1'-Biphenyl	1.427	1.395	1.452	1.573	1.581	1.507	1.473	1.487	4.77	
47)	2-Chloronaphth...	1.126	1.112	1.155	1.250	1.248	1.192	1.158	1.177	4.69	
48)	2-Nitroaniline	0.238	0.242	0.269	0.302	0.304	0.294	0.286	0.276	9.96	
49)	Acenaphthylene	1.794	1.715	1.810	1.972	1.971	1.889	1.845	1.857	5.10	
50)	Dimethylphthalate	1.444	1.377	1.437	1.532	1.543	1.470	1.429	1.462	4.02	
51)	2,6-Dinitrotol...	0.266	0.271	0.299	0.327	0.331	0.316	0.309	0.303	8.50	
52) C	Acenaphthene	1.038	0.998	1.046	1.133	1.145	1.094	1.067	1.075	4.93	
53)	3-Nitroaniline	0.233	0.247	0.284	0.321	0.325	0.311	0.302	0.289	12.47	
54) P	2,4-Dinitrophenol		0.100	0.132	0.178	0.194	0.188	0.191	0.164	23.65	
55)	Dibenzofuran	1.749	1.681	1.749	1.876	1.888	1.810	1.763	1.788	4.17	
56) P	4-Nitrophenol		0.128	0.180	0.224	0.237	0.231	0.231	0.205	20.96	
57)	2,4-Dinitrotol...	0.338	0.366	0.412	0.456	0.468	0.447	0.440	0.418	11.72	
58)	Fluorene	1.387	1.343	1.431	1.558	1.573	1.500	1.453	1.463	5.84	
59)	2,3,4,6-Tetrac...	0.366	0.362	0.398	0.435	0.445	0.432	0.431	0.410	8.44	
60)	Diethylphthalate	1.372	1.301	1.380	1.433	1.451	1.376	1.328	1.377	3.86	
61)	4-Chlorophenyl...	0.741	0.716	0.770	0.857	0.871	0.836	0.836	0.804	7.58	
62)	4-Nitroaniline	0.208	0.234	0.277	0.313	0.321	0.301	0.295	0.279	15.17	
63)	Azobenzene	1.118	1.091	1.150	1.216	1.229	1.165	1.110	1.154	4.59	
64) I	Phenanthrene-d10	-----ISTD-----									
65)	4,6-Dinitro-2-...		0.088	0.110	0.135	0.142	0.138	0.137	0.125	17.16	
66) c	n-Nitrosodiphe...	0.580	0.572	0.596	0.655	0.657	0.629	0.610	0.614	5.56	
67)	4-Bromophenyl-...	0.223	0.216	0.229	0.258	0.258	0.256	0.253	0.242	7.60	
68)	Hexachlorobenzene	0.265	0.251	0.264	0.294	0.295	0.291	0.291	0.279	6.49	
69)	Atrazine	0.196	0.198	0.214	0.236	0.242	0.234	0.230	0.222	8.47	
70) C	Pentachlorophenol		0.117	0.141	0.167	0.172	0.170	0.174	0.157	14.52	
71)	Phenanthrene	1.075	1.035	1.083	1.187	1.212	1.173	1.132	1.128	5.84	
72)	Anthracene	1.062	1.038	1.096	1.214	1.236	1.191	1.154	1.142	6.78	
73)	Carbazole	0.909	0.900	0.960	1.054	1.075	1.038	0.999	0.991	7.05	
74)	Di-n-butylphth...	1.106	1.082	1.151	1.246	1.277	1.231	1.165	1.180	6.25	
75) C	Fluoranthene	1.178	1.142	1.250	1.410	1.455	1.425	1.411	1.324	9.86	
76) I	Chrysene-d12	-----ISTD-----									
77)	Benzidine		0.560	0.646	0.772	0.769	0.765	0.751	0.710	12.36	
78)	Pyrene	1.290	1.255	1.332	1.518	1.520	1.465	1.453	1.404	7.83	
79) S	Terphenyl-d14	1.176	1.189	1.332	1.524	1.541	1.458	1.239	1.352	11.58	
80)	Butylbenzylpht...	0.489	0.479	0.515	0.562	0.568	0.545	0.524	0.526	6.53	
81)	Benzo(a)anthra...	1.280	1.231	1.321	1.478	1.479	1.442	1.408	1.377	7.24	
82)	3,3'-Dichlorob...	0.434	0.432	0.483	0.580	0.597	0.583	0.581	0.527	14.14	
83)	Chrysene	1.208	1.170	1.214	1.342	1.369	1.325	1.296	1.275	6.03	
84)	Bis(2-ethylhex...	0.726	0.725	0.775	0.843	0.852	0.814	0.765	0.786	6.61	
85) c	Di-n-octyl pht...	1.221	1.197	1.267	1.366	1.404	1.341	1.275	1.296	5.92	

Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
Method File : 8270-BM042825.M

86) I	Perylene-d12	-----ISTD-----								
87)	Indeno(1,2,3-c...	1.301	1.293	1.404	1.602	1.646	1.591	1.595	1.490	10.27
88)	Benzo(b)fluora...	1.128	1.149	1.209	1.406	1.415	1.408	1.390	1.301	10.16
89)	Benzo(k)fluora...	1.219	1.150	1.251	1.387	1.449	1.383	1.372	1.316	8.29
90) C	Benzo(a)pyrene	1.094	1.066	1.147	1.295	1.330	1.298	1.297	1.218	9.15
91)	Dibenzo(a,h)an...	1.056	1.044	1.142	1.300	1.347	1.305	1.310	1.215	10.72
92)	Benzo(g,h,i)pe...	1.066	1.030	1.105	1.236	1.268	1.224	1.212	1.163	8.07

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG No.: Q1903

Instrument ID: BNA_F Calibration Date/Time: 05/01/2025 10:17

Lab File ID: BF142250.D Init. Calib. Date(s): 04/30/2025 04/30/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:24 14:43

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.252	1.250		-0.2	
Phenol-d6	1.512	1.575		4.2	
Nitrobenzene-d5	0.304	0.333		9.5	
Naphthalene	0.997	1.030		3.3	
2-Fluorobiphenyl	1.418	1.409		-0.6	
Fluorene	1.299	1.322		1.8	
2,4,6-Tribromophenol	0.182	0.196		7.7	
Phenanthrene	1.080	1.116		3.3	
Anthracene	1.105	1.142		3.3	
Pyrene	1.845	1.938		5.0	
Terphenyl-d14	1.375	1.422		3.4	
Benzo (a) anthracene	1.343	1.378		2.6	
Chrysene	1.222	1.297		6.1	
Benzo (b) fluoranthene	1.242	1.242		0.0	
Benzo (a) pyrene	1.137	1.193		4.9	20.0
Indeno (1,2,3-cd) pyrene	1.465	1.533		4.6	
Benzo (g,h,i) perylene	1.195	1.250		4.6	

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: POWE02

Lab Code: CHEM Case No.: Q1903 SAS No.: Q1903 SDG No.: Q1903

Instrument ID: BNA_M Calibration Date/Time: 05/01/2025 13:25

Lab File ID: BM050062.D Init. Calib. Date(s): 04/28/2025 04/28/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 12:30 17:04

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.142	1.339		17.3	
Phenol-d6	1.436	1.600		11.4	
Nitrobenzene-d5	0.406	0.470		15.8	
Naphthalene	1.013	1.086		7.2	
2-Fluorobiphenyl	1.691	1.996		18.0	
Fluorene	1.463	1.622		10.9	
2,4,6-Tribromophenol	0.296	0.313		5.7	
Phenanthrene	1.128	1.244		10.3	
Anthracene	1.142	1.274		11.6	
Pyrene	1.404	1.537		9.5	
Terphenyl-d14	1.352	1.606		18.8	
Benzo (a) anthracene	1.377	1.520		10.4	
Chrysene	1.275	1.371		7.5	
Benzo (b) fluoranthene	1.301	1.413		8.6	
Benzo (a) pyrene	1.218	1.323		8.6	20.0
Indeno (1,2,3-cd) pyrene	1.490	1.620		8.7	
Benzo (g,h,i) perylene	1.163	1.262		8.5	

All other compounds must meet a minimum RRF of 0.010.