

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM

SAS No.: Q1903

SDG NO.: Q1903

Lab File ID: BM050061.D

DFTPP Injection Date: 05/01/2025

Instrument ID: BNA\_M

DFTPP Injection Time: 12:24

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	22.1
68	Less than 2.0% of mass 69	0.4 ( 1.5 ) 1
69	Mass 69 relative abundance	26.4
70	Less than 2.0% of mass 69	0.1 ( 0.4 ) 1
127	10.0 - 80.0% of mass 198	34.3
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	27
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	12.6
442	Greater than 50% of mass 198	81.4
443	15.0 - 24.0% of mass 442	15.5 ( 19.1 ) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM050062.D	05/01/2025	13:25
PB167779BL	PB167779BL	BM050063.D	05/01/2025	14:04
PB167779BS	PB167779BS	BM050064.D	05/01/2025	14:43