



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**Walsh Construction Company II, LLC**  
**Project Name: Walsh CO-032 Sampling**  
**Project # N/A**  
**Chemtech Project # Q1907**  
**Test Name: Pesticide-TCL**

### **A. Number of Samples and Date of Receipt:**

2 Solid samples were received on 04/28/2025.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: Ammonia, COD, Corrosivity, Cyanide, EPH\_NF, Gasoline Range Organics, Herbicide, Hexavalent Chromium, Ignitability, Mercury, Metals ICP-TAL, METALS TAL+CN, Oil and Grease, Paint Filter, PCB, Pesticide-TCL, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SVOC-TCL BNA -20, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP METALS, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL, TPH GC, TS, TVS and VOC-TCLVOA-10. This data package contains results for Pesticide-TCL.

### **C. Analytical Techniques:**

The analysis was performed on instrument ECD\_D. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df.; Catalog # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 7HMG017- 11. The analysis of Pesticide-TCLs was based on method 8081B and extraction was done based on method 3541.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.  
The Surrogate recoveries met the acceptable criteria.  
The Retention Times were acceptable for all samples.  
The MS recoveries met the requirements for all compounds.  
The MSD recoveries met the acceptable requirements.  
The RPD met criteria .  
The Blank Spike met requirements for all samples .  
The Blank analysis did not indicate the presence of lab contamination.  
The Initial Calibration met the requirements .  
The Continuous Calibration met the requirements .

### **E. Additional Comments:**

The soil samples results are based on a dry weight basis.

### **F. Manual Integration Comments:**



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_