

Matrix Spike Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q1907
Project:	Walsh CO-032 Sampling	Sample ID:	Q1883-15
Client ID:	OU4-VSL-19-042325MS	Percent Solids for Spike Sample:	95.8

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/Kg	75-125	1.70		0.043	U	2	1	85		04/30/2025

Matrix Spike Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q1907
Project:	Walsh CO-032 Sampling	Sample ID:	Q1883-15
Client ID:	OU4-VSL-19-042325MSD	Percent Solids for Spike Sample:	95.8

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/Kg	75-125	1.70		0.043	U	2	1	85		04/30/2025

Matrix Spike Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q1907
Project:	Walsh CO-032 Sampling	Sample ID:	Q1889-01
Client ID:	COMP-1MS	Percent Solids for Spike Sample:	82.5

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/Kg	75-125	63.1		5.50	J	58.8	1	98		04/30/2025

Matrix Spike Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q1907
Project:	Walsh CO-032 Sampling	Sample ID:	Q1889-01
Client ID:	COMP-1MSD	Percent Solids for Spike Sample:	82.5

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/Kg	75-125	65.5		5.50	J	59.4	1	101		04/30/2025

Matrix Spike Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q1907
Project:	Walsh CO-032 Sampling	Sample ID:	Q1901-02
Client ID:	B-167-SB01MS	Percent Solids for Spike Sample:	82.6

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	75-125	1500		0.083	U	1550	40	97		04/29/2025

Matrix Spike Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q1907
Project:	Walsh CO-032 Sampling	Sample ID:	Q1901-02
Client ID:	B-167-SB01MS	Percent Solids for Spike Sample:	82.6

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	85-115	46.5		0.083	U	48.4	2	96		04/29/2025

Matrix Spike Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q1907
Project:	Walsh CO-032 Sampling	Sample ID:	Q1901-02
Client ID:	B-167-SB01MS	Percent Solids for Spike Sample:	82.6

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	75-125	40.8		0.083	U	48.4	2	84		04/29/2025

Matrix Spike Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q1907
Project:	Walsh CO-032 Sampling	Sample ID:	Q1907-01
Client ID:	CO-8R-WCMS	Percent Solids for Spike Sample:	84.6

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/Kg	75-125	2580		2480		118	1	83		05/06/2025
COD	mg/Kg	75-125	7930		5850		2760	1	75		05/06/2025

Matrix Spike Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q1907
Project:	Walsh CO-032 Sampling	Sample ID:	Q1907-01
Client ID:	CO-8R-WCMSD	Percent Solids for Spike Sample:	84.6

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/Kg	75-125	2620				118	1	114		05/06/2025
COD	mg/Kg	75-125	8100		5850		2840	1	79		05/06/2025