

Laboratory Control Sample Summary

Client: Walsh Construction Company II, LLC

SDG No.: Q1907

Project: Walsh CO-032 Sampling

Run No.: LB135676

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135676BS							
Oil and Grease	mg/Kg	100	95.0		95	1	80-120	05/06/2025

Laboratory Control Sample Summary

Client: Walsh Construction Company II, LLC

SDG No.: Q1907

Project: Walsh CO-032 Sampling

Run No.: LB135683

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135683BS							
COD	mg/Kg	2500	2550		102	1	90-110	05/06/2025

Laboratory Control Sample Summary

Client: Walsh Construction Company II, LLC

SDG No.: Q1907

Project: Walsh CO-032 Sampling

Run No.: LB135605

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167753BS							
Cyanide	mg/Kg	5	4.60		92	1	85-115	04/30/2025

Laboratory Control Sample Summary

Client: Walsh Construction Company II, LLC

SDG No.: Q1907

Project: Walsh CO-032 Sampling

Run No.: LB135596

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167785BS							
Hexavalent Chromium	mg/Kg	20	19.9		100	1	84-110	04/29/2025

Laboratory Control Sample Summary

Client: Walsh Construction Company II, LLC

SDG No.: Q1907

Project: Walsh CO-032 Sampling

Run No.: LB135612

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167793BS							
Ammonia as N	mg/Kg	50	47.5		95	1	90-110	04/30/2025