



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Walsh Construction Company II, LLC **SDG No.:** Q1907
Contract: WALS01 **Lab Code:** CHEM **Case No.:** Q1907 **SAS No.:** Q1907
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	106	40 - 160	CV	04/30/2025	13:45	LB135611
CRI01	Aluminum	97.5	100	98	40 - 160	P	04/30/2025	16:41	LB135614
	Antimony	54.6	50.0	109	40 - 160	P	04/30/2025	16:41	LB135614
	Arsenic	21.9	20.0	109	40 - 160	P	04/30/2025	16:41	LB135614
	Barium	90.6	100	91	40 - 160	P	04/30/2025	16:41	LB135614
	Beryllium	6.30	6.0	105	40 - 160	P	04/30/2025	16:41	LB135614
	Cadmium	6.34	6.0	106	40 - 160	P	04/30/2025	16:41	LB135614
	Calcium	2030	2000	101	40 - 160	P	04/30/2025	16:41	LB135614
	Chromium	9.88	10.0	99	40 - 160	P	04/30/2025	16:41	LB135614
	Cobalt	30.3	30.0	101	40 - 160	P	04/30/2025	16:41	LB135614
	Copper	23.2	20.0	116	40 - 160	P	04/30/2025	16:41	LB135614
	Iron	90.9	100	91	40 - 160	P	04/30/2025	16:41	LB135614
	Lead	13.5	12.0	112	40 - 160	P	04/30/2025	16:41	LB135614
	Magnesium	2140	2000	107	40 - 160	P	04/30/2025	16:41	LB135614
	Manganese	21.0	20.0	105	40 - 160	P	04/30/2025	16:41	LB135614
	Nickel	41.3	40.0	103	40 - 160	P	04/30/2025	16:41	LB135614
	Potassium	2030	2000	101	40 - 160	P	04/30/2025	16:41	LB135614
	Selenium	16.5	20.0	82	40 - 160	P	04/30/2025	16:41	LB135614
	Silver	10.8	10.0	108	40 - 160	P	04/30/2025	16:41	LB135614
	Sodium	1740	2000	87	40 - 160	P	04/30/2025	16:41	LB135614
	Thallium	44.9	40.0	112	40 - 160	P	04/30/2025	16:41	LB135614
	Vanadium	38.9	40.0	97	40 - 160	P	04/30/2025	16:41	LB135614
	Zinc	42.8	40.0	107	40 - 160	P	04/30/2025	16:41	LB135614
CRI01	Aluminum	102	100	102	40 - 160	P	05/06/2025	15:31	LB135689
	Antimony	50.4	50.0	101	40 - 160	P	05/06/2025	15:31	LB135689
	Arsenic	20.3	20.0	101	40 - 160	P	05/06/2025	15:31	LB135689
	Barium	89.4	100	89	40 - 160	P	05/06/2025	15:31	LB135689
	Beryllium	5.91	6.0	98	40 - 160	P	05/06/2025	15:31	LB135689
	Cadmium	5.91	6.0	98	40 - 160	P	05/06/2025	15:31	LB135689
	Calcium	1950	2000	97	40 - 160	P	05/06/2025	15:31	LB135689
	Chromium	9.54	10.0	95	40 - 160	P	05/06/2025	15:31	LB135689
	Cobalt	28.3	30.0	94	40 - 160	P	05/06/2025	15:31	LB135689
	Copper	21.6	20.0	108	40 - 160	P	05/06/2025	15:31	LB135689
	Iron	95.6	100	96	40 - 160	P	05/06/2025	15:31	LB135689
	Lead	10.0	12.0	83	40 - 160	P	05/06/2025	15:31	LB135689

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Walsh Construction Company II, LLC

SDG No.: Q1907

Contract: WALS01

Lab Code: CHEM

Case No.: Q1907

SAS No.: Q1907

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Magnesium	2050	2000	102	40 - 160	P	05/06/2025	15:31	LB135689
	Manganese	19.8	20.0	99	40 - 160	P	05/06/2025	15:31	LB135689
	Nickel	38.6	40.0	97	40 - 160	P	05/06/2025	15:31	LB135689
	Potassium	1730	2000	87	40 - 160	P	05/06/2025	15:31	LB135689
	Selenium	22.5	20.0	112	40 - 160	P	05/06/2025	15:31	LB135689
	Silver	9.66	10.0	97	40 - 160	P	05/06/2025	15:31	LB135689
	Sodium	1760	2000	88	40 - 160	P	05/06/2025	15:31	LB135689
	Thallium	37.5	40.0	94	40 - 160	P	05/06/2025	15:31	LB135689
	Vanadium	39.4	40.0	98	40 - 160	P	05/06/2025	15:31	LB135689
	Zinc	39.7	40.0	99	40 - 160	P	05/06/2025	15:31	LB135689