



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Walsh Construction Company II, LLC</u>		SDG No.: <u>Q1907</u>							
Contract:	<u>WALS01</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1907</u>	SAS No.:	<u>Q1907</u>		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB109	Mercury	0.20	+/-0.20	U	0.20	CV	04/30/2025	13:38	LB135611

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q1907

Contract: WALS01

Lab Code: CHEM

Case No.: Q1907

SAS No.: Q1907

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB52	Mercury	0.20	+/-0.20	U	0.20	CV	04/30/2025	13:42	LB135611
CCB53	Mercury	0.20	+/-0.20	U	0.20	CV	04/30/2025	14:13	LB135611
CCB54	Mercury	0.20	+/-0.20	U	0.20	CV	04/30/2025	14:40	LB135611
CCB55	Mercury	0.20	+/-0.20	U	0.20	CV	04/30/2025	14:58	LB135611

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC **SDG No.:** Q1907
Contract: WALS01 **Lab Code:** CHEM **Case No.:** Q1907 **SAS No.:** Q1907

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	04/30/2025	16:36	LB135614
	Antimony	50.0	+/-50.0	U	50.0	P	04/30/2025	16:36	LB135614
	Arsenic	20.0	+/-20.0	U	20.0	P	04/30/2025	16:36	LB135614
	Barium	100	+/-100	U	100	P	04/30/2025	16:36	LB135614
	Beryllium	6.00	+/-6.00	U	6.00	P	04/30/2025	16:36	LB135614
	Cadmium	6.00	+/-6.00	U	6.00	P	04/30/2025	16:36	LB135614
	Calcium	2000	+/-2000	U	2000	P	04/30/2025	16:36	LB135614
	Chromium	10.0	+/-10.0	U	10.0	P	04/30/2025	16:36	LB135614
	Cobalt	30.0	+/-30.0	U	30.0	P	04/30/2025	16:36	LB135614
	Copper	20.0	+/-20.0	U	20.0	P	04/30/2025	16:36	LB135614
	Iron	100	+/-100	U	100	P	04/30/2025	16:36	LB135614
	Lead	12.0	+/-12.0	U	12.0	P	04/30/2025	16:36	LB135614
	Magnesium	2000	+/-2000	U	2000	P	04/30/2025	16:36	LB135614
	Manganese	20.0	+/-20.0	U	20.0	P	04/30/2025	16:36	LB135614
	Nickel	40.0	+/-40.0	U	40.0	P	04/30/2025	16:36	LB135614
	Potassium	2000	+/-2000	U	2000	P	04/30/2025	16:36	LB135614
	Selenium	20.0	+/-20.0	U	20.0	P	04/30/2025	16:36	LB135614
	Silver	10.0	+/-10.0	U	10.0	P	04/30/2025	16:36	LB135614
	Sodium	2000	+/-2000	U	2000	P	04/30/2025	16:36	LB135614
	Thallium	40.0	+/-40.0	U	40.0	P	04/30/2025	16:36	LB135614
	Vanadium	40.0	+/-40.0	U	40.0	P	04/30/2025	16:36	LB135614
	Zinc	40.0	+/-40.0	U	40.0	P	04/30/2025	16:36	LB135614

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q1907

Contract: WALS01

Lab Code: CHEM

Case No.: Q1907

SAS No.: Q1907

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	04/30/2025	17:38	LB135614
	Antimony	50.0	+/-50.0	U	50.0	P	04/30/2025	17:38	LB135614
	Arsenic	20.0	+/-20.0	U	20.0	P	04/30/2025	17:38	LB135614
	Barium	100	+/-100	U	100	P	04/30/2025	17:38	LB135614
	Beryllium	6.00	+/-6.00	U	6.00	P	04/30/2025	17:38	LB135614
	Cadmium	6.00	+/-6.00	U	6.00	P	04/30/2025	17:38	LB135614
	Calcium	2000	+/-2000	U	2000	P	04/30/2025	17:38	LB135614
	Chromium	10.0	+/-10.0	U	10.0	P	04/30/2025	17:38	LB135614
	Cobalt	30.0	+/-30.0	U	30.0	P	04/30/2025	17:38	LB135614
	Copper	20.0	+/-20.0	U	20.0	P	04/30/2025	17:38	LB135614
	Iron	100	+/-100	U	100	P	04/30/2025	17:38	LB135614
	Lead	12.0	+/-12.0	U	12.0	P	04/30/2025	17:38	LB135614
	Magnesium	2000	+/-2000	U	2000	P	04/30/2025	17:38	LB135614
	Manganese	20.0	+/-20.0	U	20.0	P	04/30/2025	17:38	LB135614
	Nickel	40.0	+/-40.0	U	40.0	P	04/30/2025	17:38	LB135614
	Potassium	2000	+/-2000	U	2000	P	04/30/2025	17:38	LB135614
	Selenium	20.0	+/-20.0	U	20.0	P	04/30/2025	17:38	LB135614
	Silver	10.0	+/-10.0	U	10.0	P	04/30/2025	17:38	LB135614
	Sodium	2000	+/-2000	U	2000	P	04/30/2025	17:38	LB135614
	Thallium	40.0	+/-40.0	U	40.0	P	04/30/2025	17:38	LB135614
	Vanadium	40.0	+/-40.0	U	40.0	P	04/30/2025	17:38	LB135614
	Zinc	40.0	+/-40.0	U	40.0	P	04/30/2025	17:38	LB135614
CCB02	Aluminum	100	+/-100	U	100	P	04/30/2025	18:35	LB135614
	Antimony	50.0	+/-50.0	U	50.0	P	04/30/2025	18:35	LB135614
	Arsenic	20.0	+/-20.0	U	20.0	P	04/30/2025	18:35	LB135614
	Barium	100	+/-100	U	100	P	04/30/2025	18:35	LB135614
	Beryllium	6.00	+/-6.00	U	6.00	P	04/30/2025	18:35	LB135614
	Cadmium	6.00	+/-6.00	U	6.00	P	04/30/2025	18:35	LB135614
	Calcium	2000	+/-2000	U	2000	P	04/30/2025	18:35	LB135614
	Chromium	10.0	+/-10.0	U	10.0	P	04/30/2025	18:35	LB135614
	Cobalt	30.0	+/-30.0	U	30.0	P	04/30/2025	18:35	LB135614
	Copper	20.0	+/-20.0	U	20.0	P	04/30/2025	18:35	LB135614
	Iron	100	+/-100	U	100	P	04/30/2025	18:35	LB135614
	Lead	12.0	+/-12.0	U	12.0	P	04/30/2025	18:35	LB135614
	Magnesium	2000	+/-2000	U	2000	P	04/30/2025	18:35	LB135614
	Manganese	20.0	+/-20.0	U	20.0	P	04/30/2025	18:35	LB135614
	Nickel	40.0	+/-40.0	U	40.0	P	04/30/2025	18:35	LB135614
	Potassium	2000	+/-2000	U	2000	P	04/30/2025	18:35	LB135614
	Selenium	20.0	+/-20.0	U	20.0	P	04/30/2025	18:35	LB135614

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q1907

Contract: WALS01

Lab Code: CHEM

Case No.: Q1907

SAS No.: Q1907

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	04/30/2025	18:35	LB135614
	Sodium	2000	+/-2000	U	2000	P	04/30/2025	18:35	LB135614
	Thallium	40.0	+/-40.0	U	40.0	P	04/30/2025	18:35	LB135614
	Vanadium	40.0	+/-40.0	U	40.0	P	04/30/2025	18:35	LB135614
	Zinc	40.0	+/-40.0	U	40.0	P	04/30/2025	18:35	LB135614
CCB03	Aluminum	100	+/-100	U	100	P	04/30/2025	19:41	LB135614
	Antimony	50.0	+/-50.0	U	50.0	P	04/30/2025	19:41	LB135614
	Arsenic	20.0	+/-20.0	U	20.0	P	04/30/2025	19:41	LB135614
	Barium	100	+/-100	U	100	P	04/30/2025	19:41	LB135614
	Beryllium	6.00	+/-6.00	U	6.00	P	04/30/2025	19:41	LB135614
	Cadmium	1.42	+/-6.00	J	6.00	P	04/30/2025	19:41	LB135614
	Calcium	2000	+/-2000	U	2000	P	04/30/2025	19:41	LB135614
	Chromium	10.0	+/-10.0	U	10.0	P	04/30/2025	19:41	LB135614
	Cobalt	30.0	+/-30.0	U	30.0	P	04/30/2025	19:41	LB135614
	Copper	20.0	+/-20.0	U	20.0	P	04/30/2025	19:41	LB135614
	Iron	35.4	+/-100	J	100	P	04/30/2025	19:41	LB135614
	Lead	3.92	+/-12.0	J	12.0	P	04/30/2025	19:41	LB135614
	Magnesium	2000	+/-2000	U	2000	P	04/30/2025	19:41	LB135614
	Manganese	20.0	+/-20.0	U	20.0	P	04/30/2025	19:41	LB135614
	Nickel	40.0	+/-40.0	U	40.0	P	04/30/2025	19:41	LB135614
	Potassium	2000	+/-2000	U	2000	P	04/30/2025	19:41	LB135614
	Selenium	20.0	+/-20.0	U	20.0	P	04/30/2025	19:41	LB135614
	Silver	10.0	+/-10.0	U	10.0	P	04/30/2025	19:41	LB135614
	Sodium	2000	+/-2000	U	2000	P	04/30/2025	19:41	LB135614
	Thallium	40.0	+/-40.0	U	40.0	P	04/30/2025	19:41	LB135614
	Vanadium	40.0	+/-40.0	U	40.0	P	04/30/2025	19:41	LB135614
	Zinc	40.0	+/-40.0	U	40.0	P	04/30/2025	19:41	LB135614
CCB04	Aluminum	100	+/-100	U	100	P	04/30/2025	20:27	LB135614
	Antimony	50.0	+/-50.0	U	50.0	P	04/30/2025	20:27	LB135614
	Arsenic	20.0	+/-20.0	U	20.0	P	04/30/2025	20:27	LB135614
	Barium	100	+/-100	U	100	P	04/30/2025	20:27	LB135614
	Beryllium	6.00	+/-6.00	U	6.00	P	04/30/2025	20:27	LB135614
	Cadmium	0.57	+/-6.00	J	6.00	P	04/30/2025	20:27	LB135614
	Calcium	2000	+/-2000	U	2000	P	04/30/2025	20:27	LB135614
	Chromium	10.0	+/-10.0	U	10.0	P	04/30/2025	20:27	LB135614
	Cobalt	30.0	+/-30.0	U	30.0	P	04/30/2025	20:27	LB135614
	Copper	20.0	+/-20.0	U	20.0	P	04/30/2025	20:27	LB135614
	Iron	23.9	+/-100	J	100	P	04/30/2025	20:27	LB135614
	Lead	12.0	+/-12.0	U	12.0	P	04/30/2025	20:27	LB135614

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Walsh Construction Company II, LLC</u>		SDG No.: <u>Q1907</u>							
Contract: <u>WALS01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1907</u>	SAS No.: <u>Q1907</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	04/30/2025	20:27	LB135614
	Manganese	20.0	+/-20.0	U	20.0	P	04/30/2025	20:27	LB135614
	Nickel	40.0	+/-40.0	U	40.0	P	04/30/2025	20:27	LB135614
	Potassium	2000	+/-2000	U	2000	P	04/30/2025	20:27	LB135614
	Selenium	20.0	+/-20.0	U	20.0	P	04/30/2025	20:27	LB135614
	Silver	10.0	+/-10.0	U	10.0	P	04/30/2025	20:27	LB135614
	Sodium	2000	+/-2000	U	2000	P	04/30/2025	20:27	LB135614
	Thallium	40.0	+/-40.0	U	40.0	P	04/30/2025	20:27	LB135614
	Vanadium	40.0	+/-40.0	U	40.0	P	04/30/2025	20:27	LB135614
	Zinc	40.0	+/-40.0	U	40.0	P	04/30/2025	20:27	LB135614
CCB05	Aluminum	100	+/-100	U	100	P	04/30/2025	21:13	LB135614
	Antimony	50.0	+/-50.0	U	50.0	P	04/30/2025	21:13	LB135614
	Arsenic	20.0	+/-20.0	U	20.0	P	04/30/2025	21:13	LB135614
	Barium	100	+/-100	U	100	P	04/30/2025	21:13	LB135614
	Beryllium	6.00	+/-6.00	U	6.00	P	04/30/2025	21:13	LB135614
	Cadmium	0.94	+/-6.00	J	6.00	P	04/30/2025	21:13	LB135614
	Calcium	2000	+/-2000	U	2000	P	04/30/2025	21:13	LB135614
	Chromium	10.0	+/-10.0	U	10.0	P	04/30/2025	21:13	LB135614
	Cobalt	30.0	+/-30.0	U	30.0	P	04/30/2025	21:13	LB135614
	Copper	20.0	+/-20.0	U	20.0	P	04/30/2025	21:13	LB135614
	Iron	100	+/-100	U	100	P	04/30/2025	21:13	LB135614
	Lead	12.0	+/-12.0	U	12.0	P	04/30/2025	21:13	LB135614
	Magnesium	2000	+/-2000	U	2000	P	04/30/2025	21:13	LB135614
	Manganese	20.0	+/-20.0	U	20.0	P	04/30/2025	21:13	LB135614
	Nickel	40.0	+/-40.0	U	40.0	P	04/30/2025	21:13	LB135614
	Potassium	2000	+/-2000	U	2000	P	04/30/2025	21:13	LB135614
	Selenium	20.0	+/-20.0	U	20.0	P	04/30/2025	21:13	LB135614
	Silver	10.0	+/-10.0	U	10.0	P	04/30/2025	21:13	LB135614
	Sodium	2000	+/-2000	U	2000	P	04/30/2025	21:13	LB135614
	Thallium	40.0	+/-40.0	U	40.0	P	04/30/2025	21:13	LB135614
	Vanadium	40.0	+/-40.0	U	40.0	P	04/30/2025	21:13	LB135614
	Zinc	40.0	+/-40.0	U	40.0	P	04/30/2025	21:13	LB135614
CCB06	Aluminum	19.2	+/-100	J	100	P	04/30/2025	21:58	LB135614
	Antimony	50.0	+/-50.0	U	50.0	P	04/30/2025	21:58	LB135614
	Arsenic	20.0	+/-20.0	U	20.0	P	04/30/2025	21:58	LB135614
	Barium	100	+/-100	U	100	P	04/30/2025	21:58	LB135614
	Beryllium	6.00	+/-6.00	U	6.00	P	04/30/2025	21:58	LB135614
	Cadmium	6.00	+/-6.00	U	6.00	P	04/30/2025	21:58	LB135614
	Calcium	2000	+/-2000	U	2000	P	04/30/2025	21:58	LB135614

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Walsh Construction Company II, LLC</u>		SDG No.: <u>Q1907</u>							
Contract: <u>WALS01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1907</u>	SAS No.: <u>Q1907</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	04/30/2025	21:58	LB135614
	Cobalt	30.0	+/-30.0	U	30.0	P	04/30/2025	21:58	LB135614
	Copper	20.0	+/-20.0	U	20.0	P	04/30/2025	21:58	LB135614
	Iron	100	+/-100	U	100	P	04/30/2025	21:58	LB135614
	Lead	2.39	+/-12.0	J	12.0	P	04/30/2025	21:58	LB135614
	Magnesium	2000	+/-2000	U	2000	P	04/30/2025	21:58	LB135614
	Manganese	20.0	+/-20.0	U	20.0	P	04/30/2025	21:58	LB135614
	Nickel	40.0	+/-40.0	U	40.0	P	04/30/2025	21:58	LB135614
	Potassium	2000	+/-2000	U	2000	P	04/30/2025	21:58	LB135614
	Selenium	20.0	+/-20.0	U	20.0	P	04/30/2025	21:58	LB135614
	Silver	10.0	+/-10.0	U	10.0	P	04/30/2025	21:58	LB135614
	Sodium	2000	+/-2000	U	2000	P	04/30/2025	21:58	LB135614
	Thallium	40.0	+/-40.0	U	40.0	P	04/30/2025	21:58	LB135614
	Vanadium	40.0	+/-40.0	U	40.0	P	04/30/2025	21:58	LB135614
	Zinc	40.0	+/-40.0	U	40.0	P	04/30/2025	21:58	LB135614
CCB07	Aluminum	100	+/-100	U	100	P	04/30/2025	22:44	LB135614
	Antimony	50.0	+/-50.0	U	50.0	P	04/30/2025	22:44	LB135614
	Arsenic	20.0	+/-20.0	U	20.0	P	04/30/2025	22:44	LB135614
	Barium	100	+/-100	U	100	P	04/30/2025	22:44	LB135614
	Beryllium	6.00	+/-6.00	U	6.00	P	04/30/2025	22:44	LB135614
	Cadmium	0.88	+/-6.00	J	6.00	P	04/30/2025	22:44	LB135614
	Calcium	2000	+/-2000	U	2000	P	04/30/2025	22:44	LB135614
	Chromium	10.0	+/-10.0	U	10.0	P	04/30/2025	22:44	LB135614
	Cobalt	30.0	+/-30.0	U	30.0	P	04/30/2025	22:44	LB135614
	Copper	20.0	+/-20.0	U	20.0	P	04/30/2025	22:44	LB135614
	Iron	100	+/-100	U	100	P	04/30/2025	22:44	LB135614
	Lead	12.0	+/-12.0	U	12.0	P	04/30/2025	22:44	LB135614
	Magnesium	2000	+/-2000	U	2000	P	04/30/2025	22:44	LB135614
	Manganese	20.0	+/-20.0	U	20.0	P	04/30/2025	22:44	LB135614
	Nickel	40.0	+/-40.0	U	40.0	P	04/30/2025	22:44	LB135614
	Potassium	2000	+/-2000	U	2000	P	04/30/2025	22:44	LB135614
	Selenium	20.0	+/-20.0	U	20.0	P	04/30/2025	22:44	LB135614
	Silver	10.0	+/-10.0	U	10.0	P	04/30/2025	22:44	LB135614
	Sodium	2000	+/-2000	U	2000	P	04/30/2025	22:44	LB135614
	Thallium	40.0	+/-40.0	U	40.0	P	04/30/2025	22:44	LB135614
	Vanadium	40.0	+/-40.0	U	40.0	P	04/30/2025	22:44	LB135614
	Zinc	40.0	+/-40.0	U	40.0	P	04/30/2025	22:44	LB135614
CCB08	Aluminum	100	+/-100	U	100	P	04/30/2025	23:09	LB135614
	Antimony	50.0	+/-50.0	U	50.0	P	04/30/2025	23:09	LB135614

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC **SDG No.:** Q1907
Contract: WALS01 **Lab Code:** CHEM **Case No.:** Q1907 **SAS No.:** Q1907

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	04/30/2025	23:09	LB135614
	Barium	100	+/-100	U	100	P	04/30/2025	23:09	LB135614
	Beryllium	6.00	+/-6.00	U	6.00	P	04/30/2025	23:09	LB135614
	Cadmium	6.00	+/-6.00	U	6.00	P	04/30/2025	23:09	LB135614
	Calcium	2000	+/-2000	U	2000	P	04/30/2025	23:09	LB135614
	Chromium	10.0	+/-10.0	U	10.0	P	04/30/2025	23:09	LB135614
	Cobalt	30.0	+/-30.0	U	30.0	P	04/30/2025	23:09	LB135614
	Copper	20.0	+/-20.0	U	20.0	P	04/30/2025	23:09	LB135614
	Iron	100	+/-100	U	100	P	04/30/2025	23:09	LB135614
	Lead	12.0	+/-12.0	U	12.0	P	04/30/2025	23:09	LB135614
	Magnesium	2000	+/-2000	U	2000	P	04/30/2025	23:09	LB135614
	Manganese	20.0	+/-20.0	U	20.0	P	04/30/2025	23:09	LB135614
	Nickel	40.0	+/-40.0	U	40.0	P	04/30/2025	23:09	LB135614
	Potassium	2000	+/-2000	U	2000	P	04/30/2025	23:09	LB135614
	Selenium	20.0	+/-20.0	U	20.0	P	04/30/2025	23:09	LB135614
	Silver	10.0	+/-10.0	U	10.0	P	04/30/2025	23:09	LB135614
	Sodium	2000	+/-2000	U	2000	P	04/30/2025	23:09	LB135614
	Thallium	40.0	+/-40.0	U	40.0	P	04/30/2025	23:09	LB135614
	Vanadium	40.0	+/-40.0	U	40.0	P	04/30/2025	23:09	LB135614
	Zinc	40.0	+/-40.0	U	40.0	P	04/30/2025	23:09	LB135614

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q1907

Contract: WALS01

Lab Code: CHEM

Case No.: Q1907

SAS No.: Q1907

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	05/06/2025	15:27	LB135689
	Antimony	50.0	+/-50.0	U	50.0	P	05/06/2025	15:27	LB135689
	Arsenic	20.0	+/-20.0	U	20.0	P	05/06/2025	15:27	LB135689
	Barium	100	+/-100	U	100	P	05/06/2025	15:27	LB135689
	Beryllium	6.00	+/-6.00	U	6.00	P	05/06/2025	15:27	LB135689
	Cadmium	6.00	+/-6.00	U	6.00	P	05/06/2025	15:27	LB135689
	Calcium	2000	+/-2000	U	2000	P	05/06/2025	15:27	LB135689
	Chromium	10.0	+/-10.0	U	10.0	P	05/06/2025	15:27	LB135689
	Cobalt	30.0	+/-30.0	U	30.0	P	05/06/2025	15:27	LB135689
	Copper	20.0	+/-20.0	U	20.0	P	05/06/2025	15:27	LB135689
	Iron	100	+/-100	U	100	P	05/06/2025	15:27	LB135689
	Lead	12.0	+/-12.0	U	12.0	P	05/06/2025	15:27	LB135689
	Magnesium	2000	+/-2000	U	2000	P	05/06/2025	15:27	LB135689
	Manganese	20.0	+/-20.0	U	20.0	P	05/06/2025	15:27	LB135689
	Nickel	40.0	+/-40.0	U	40.0	P	05/06/2025	15:27	LB135689
	Potassium	2000	+/-2000	U	2000	P	05/06/2025	15:27	LB135689
	Selenium	20.0	+/-20.0	U	20.0	P	05/06/2025	15:27	LB135689
	Silver	10.0	+/-10.0	U	10.0	P	05/06/2025	15:27	LB135689
	Sodium	2000	+/-2000	U	2000	P	05/06/2025	15:27	LB135689
	Thallium	40.0	+/-40.0	U	40.0	P	05/06/2025	15:27	LB135689
	Vanadium	40.0	+/-40.0	U	40.0	P	05/06/2025	15:27	LB135689
	Zinc	40.0	+/-40.0	U	40.0	P	05/06/2025	15:27	LB135689

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q1907

Contract: WALS01

Lab Code: CHEM

Case No.: Q1907

SAS No.: Q1907

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	12.8	+/-100	J	100	P	05/06/2025	16:34	LB135689
	Antimony	50.0	+/-50.0	U	50.0	P	05/06/2025	16:34	LB135689
	Arsenic	20.0	+/-20.0	U	20.0	P	05/06/2025	16:34	LB135689
	Barium	100	+/-100	U	100	P	05/06/2025	16:34	LB135689
	Beryllium	6.00	+/-6.00	U	6.00	P	05/06/2025	16:34	LB135689
	Cadmium	6.00	+/-6.00	U	6.00	P	05/06/2025	16:34	LB135689
	Calcium	2000	+/-2000	U	2000	P	05/06/2025	16:34	LB135689
	Chromium	10.0	+/-10.0	U	10.0	P	05/06/2025	16:34	LB135689
	Cobalt	30.0	+/-30.0	U	30.0	P	05/06/2025	16:34	LB135689
	Copper	20.0	+/-20.0	U	20.0	P	05/06/2025	16:34	LB135689
	Iron	100	+/-100	U	100	P	05/06/2025	16:34	LB135689
	Lead	12.0	+/-12.0	U	12.0	P	05/06/2025	16:34	LB135689
	Magnesium	2000	+/-2000	U	2000	P	05/06/2025	16:34	LB135689
	Manganese	20.0	+/-20.0	U	20.0	P	05/06/2025	16:34	LB135689
	Nickel	40.0	+/-40.0	U	40.0	P	05/06/2025	16:34	LB135689
	Potassium	2000	+/-2000	U	2000	P	05/06/2025	16:34	LB135689
	Selenium	20.0	+/-20.0	U	20.0	P	05/06/2025	16:34	LB135689
	Silver	10.0	+/-10.0	U	10.0	P	05/06/2025	16:34	LB135689
	Sodium	2000	+/-2000	U	2000	P	05/06/2025	16:34	LB135689
	Thallium	40.0	+/-40.0	U	40.0	P	05/06/2025	16:34	LB135689
	Vanadium	40.0	+/-40.0	U	40.0	P	05/06/2025	16:34	LB135689
	Zinc	40.0	+/-40.0	U	40.0	P	05/06/2025	16:34	LB135689
CCB02	Aluminum	14.3	+/-100	J	100	P	05/06/2025	17:20	LB135689
	Antimony	50.0	+/-50.0	U	50.0	P	05/06/2025	17:20	LB135689
	Arsenic	20.0	+/-20.0	U	20.0	P	05/06/2025	17:20	LB135689
	Barium	100	+/-100	U	100	P	05/06/2025	17:20	LB135689
	Beryllium	6.00	+/-6.00	U	6.00	P	05/06/2025	17:20	LB135689
	Cadmium	0.60	+/-6.00	J	6.00	P	05/06/2025	17:20	LB135689
	Calcium	2000	+/-2000	U	2000	P	05/06/2025	17:20	LB135689
	Chromium	10.0	+/-10.0	U	10.0	P	05/06/2025	17:20	LB135689
	Cobalt	30.0	+/-30.0	U	30.0	P	05/06/2025	17:20	LB135689
	Copper	20.0	+/-20.0	U	20.0	P	05/06/2025	17:20	LB135689
	Iron	100	+/-100	U	100	P	05/06/2025	17:20	LB135689
	Lead	12.0	+/-12.0	U	12.0	P	05/06/2025	17:20	LB135689
	Magnesium	2000	+/-2000	U	2000	P	05/06/2025	17:20	LB135689
	Manganese	20.0	+/-20.0	U	20.0	P	05/06/2025	17:20	LB135689
	Nickel	40.0	+/-40.0	U	40.0	P	05/06/2025	17:20	LB135689
	Potassium	2000	+/-2000	U	2000	P	05/06/2025	17:20	LB135689
	Selenium	20.0	+/-20.0	U	20.0	P	05/06/2025	17:20	LB135689

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q1907

Contract: WALS01

Lab Code: CHEM

Case No.: Q1907

SAS No.: Q1907

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	05/06/2025	17:20	LB135689
	Sodium	2000	+/-2000	U	2000	P	05/06/2025	17:20	LB135689
	Thallium	40.0	+/-40.0	U	40.0	P	05/06/2025	17:20	LB135689
	Vanadium	40.0	+/-40.0	U	40.0	P	05/06/2025	17:20	LB135689
	Zinc	40.0	+/-40.0	U	40.0	P	05/06/2025	17:20	LB135689
CCB03	Aluminum	15.8	+/-100	J	100	P	05/06/2025	18:13	LB135689
	Antimony	50.0	+/-50.0	U	50.0	P	05/06/2025	18:13	LB135689
	Arsenic	20.0	+/-20.0	U	20.0	P	05/06/2025	18:13	LB135689
	Barium	100	+/-100	U	100	P	05/06/2025	18:13	LB135689
	Beryllium	6.00	+/-6.00	U	6.00	P	05/06/2025	18:13	LB135689
	Cadmium	0.59	+/-6.00	J	6.00	P	05/06/2025	18:13	LB135689
	Calcium	2000	+/-2000	U	2000	P	05/06/2025	18:13	LB135689
	Chromium	10.0	+/-10.0	U	10.0	P	05/06/2025	18:13	LB135689
	Cobalt	30.0	+/-30.0	U	30.0	P	05/06/2025	18:13	LB135689
	Copper	20.0	+/-20.0	U	20.0	P	05/06/2025	18:13	LB135689
	Iron	23.9	+/-100	J	100	P	05/06/2025	18:13	LB135689
	Lead	12.0	+/-12.0	U	12.0	P	05/06/2025	18:13	LB135689
	Magnesium	2000	+/-2000	U	2000	P	05/06/2025	18:13	LB135689
	Manganese	20.0	+/-20.0	U	20.0	P	05/06/2025	18:13	LB135689
	Nickel	40.0	+/-40.0	U	40.0	P	05/06/2025	18:13	LB135689
	Potassium	2000	+/-2000	U	2000	P	05/06/2025	18:13	LB135689
	Selenium	20.0	+/-20.0	U	20.0	P	05/06/2025	18:13	LB135689
	Silver	10.0	+/-10.0	U	10.0	P	05/06/2025	18:13	LB135689
	Sodium	2000	+/-2000	U	2000	P	05/06/2025	18:13	LB135689
	Thallium	40.0	+/-40.0	U	40.0	P	05/06/2025	18:13	LB135689
	Vanadium	40.0	+/-40.0	U	40.0	P	05/06/2025	18:13	LB135689
	Zinc	40.0	+/-40.0	U	40.0	P	05/06/2025	18:13	LB135689
CCB04	Aluminum	16.7	+/-100	J	100	P	05/06/2025	19:20	LB135689
	Antimony	50.0	+/-50.0	U	50.0	P	05/06/2025	19:20	LB135689
	Arsenic	20.0	+/-20.0	U	20.0	P	05/06/2025	19:20	LB135689
	Barium	100	+/-100	U	100	P	05/06/2025	19:20	LB135689
	Beryllium	6.00	+/-6.00	U	6.00	P	05/06/2025	19:20	LB135689
	Cadmium	6.00	+/-6.00	U	6.00	P	05/06/2025	19:20	LB135689
	Calcium	2000	+/-2000	U	2000	P	05/06/2025	19:20	LB135689
	Chromium	10.0	+/-10.0	U	10.0	P	05/06/2025	19:20	LB135689
	Cobalt	30.0	+/-30.0	U	30.0	P	05/06/2025	19:20	LB135689
	Copper	20.0	+/-20.0	U	20.0	P	05/06/2025	19:20	LB135689
	Iron	100	+/-100	U	100	P	05/06/2025	19:20	LB135689
	Lead	12.0	+/-12.0	U	12.0	P	05/06/2025	19:20	LB135689

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Walsh Construction Company II, LLC</u>		SDG No.: <u>Q1907</u>							
Contract: <u>WALS01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1907</u>	SAS No.: <u>Q1907</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	05/06/2025	19:20	LB135689
	Manganese	20.0	+/-20.0	U	20.0	P	05/06/2025	19:20	LB135689
	Nickel	40.0	+/-40.0	U	40.0	P	05/06/2025	19:20	LB135689
	Potassium	2000	+/-2000	U	2000	P	05/06/2025	19:20	LB135689
	Selenium	20.0	+/-20.0	U	20.0	P	05/06/2025	19:20	LB135689
	Silver	10.0	+/-10.0	U	10.0	P	05/06/2025	19:20	LB135689
	Sodium	2000	+/-2000	U	2000	P	05/06/2025	19:20	LB135689
	Thallium	40.0	+/-40.0	U	40.0	P	05/06/2025	19:20	LB135689
	Vanadium	40.0	+/-40.0	U	40.0	P	05/06/2025	19:20	LB135689
	Zinc	40.0	+/-40.0	U	40.0	P	05/06/2025	19:20	LB135689
CCB05	Aluminum	100	+/-100	U	100	P	05/06/2025	20:06	LB135689
	Antimony	50.0	+/-50.0	U	50.0	P	05/06/2025	20:06	LB135689
	Arsenic	20.0	+/-20.0	U	20.0	P	05/06/2025	20:06	LB135689
	Barium	100	+/-100	U	100	P	05/06/2025	20:06	LB135689
	Beryllium	6.00	+/-6.00	U	6.00	P	05/06/2025	20:06	LB135689
	Cadmium	6.00	+/-6.00	U	6.00	P	05/06/2025	20:06	LB135689
	Calcium	2000	+/-2000	U	2000	P	05/06/2025	20:06	LB135689
	Chromium	10.0	+/-10.0	U	10.0	P	05/06/2025	20:06	LB135689
	Cobalt	30.0	+/-30.0	U	30.0	P	05/06/2025	20:06	LB135689
	Copper	20.0	+/-20.0	U	20.0	P	05/06/2025	20:06	LB135689
	Iron	100	+/-100	U	100	P	05/06/2025	20:06	LB135689
	Lead	12.0	+/-12.0	U	12.0	P	05/06/2025	20:06	LB135689
	Magnesium	2000	+/-2000	U	2000	P	05/06/2025	20:06	LB135689
	Manganese	20.0	+/-20.0	U	20.0	P	05/06/2025	20:06	LB135689
	Nickel	40.0	+/-40.0	U	40.0	P	05/06/2025	20:06	LB135689
	Potassium	2000	+/-2000	U	2000	P	05/06/2025	20:06	LB135689
	Selenium	20.0	+/-20.0	U	20.0	P	05/06/2025	20:06	LB135689
	Silver	10.0	+/-10.0	U	10.0	P	05/06/2025	20:06	LB135689
	Sodium	2000	+/-2000	U	2000	P	05/06/2025	20:06	LB135689
	Thallium	40.0	+/-40.0	U	40.0	P	05/06/2025	20:06	LB135689
	Vanadium	40.0	+/-40.0	U	40.0	P	05/06/2025	20:06	LB135689
	Zinc	40.0	+/-40.0	U	40.0	P	05/06/2025	20:06	LB135689
CCB06	Aluminum	100	+/-100	U	100	P	05/06/2025	20:35	LB135689
	Antimony	50.0	+/-50.0	U	50.0	P	05/06/2025	20:35	LB135689
	Arsenic	20.0	+/-20.0	U	20.0	P	05/06/2025	20:35	LB135689
	Barium	100	+/-100	U	100	P	05/06/2025	20:35	LB135689
	Beryllium	6.00	+/-6.00	U	6.00	P	05/06/2025	20:35	LB135689
	Cadmium	6.00	+/-6.00	U	6.00	P	05/06/2025	20:35	LB135689
	Calcium	2000	+/-2000	U	2000	P	05/06/2025	20:35	LB135689

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC **SDG No.:** Q1907
Contract: WALS01 **Lab Code:** CHEM **Case No.:** Q1907 **SAS No.:** Q1907

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	05/06/2025	20:35	LB135689
	Cobalt	30.0	+/-30.0	U	30.0	P	05/06/2025	20:35	LB135689
	Copper	20.0	+/-20.0	U	20.0	P	05/06/2025	20:35	LB135689
	Iron	100	+/-100	U	100	P	05/06/2025	20:35	LB135689
	Lead	12.0	+/-12.0	U	12.0	P	05/06/2025	20:35	LB135689
	Magnesium	2000	+/-2000	U	2000	P	05/06/2025	20:35	LB135689
	Manganese	20.0	+/-20.0	U	20.0	P	05/06/2025	20:35	LB135689
	Nickel	40.0	+/-40.0	U	40.0	P	05/06/2025	20:35	LB135689
	Potassium	2000	+/-2000	U	2000	P	05/06/2025	20:35	LB135689
	Selenium	20.0	+/-20.0	U	20.0	P	05/06/2025	20:35	LB135689
	Silver	10.0	+/-10.0	U	10.0	P	05/06/2025	20:35	LB135689
	Sodium	2000	+/-2000	U	2000	P	05/06/2025	20:35	LB135689
	Thallium	40.0	+/-40.0	U	40.0	P	05/06/2025	20:35	LB135689
	Vanadium	40.0	+/-40.0	U	40.0	P	05/06/2025	20:35	LB135689
	Zinc	40.0	+/-40.0	U	40.0	P	05/06/2025	20:35	LB135689



- 3a -

Client: Walsh Construction Company II, LLC		SDG No.: Q1907	
Contract: WAL501	Lab Code: CHEM	Case No.: Q1907	SAS No.: Q1907

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
1	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
2	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
3	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
4	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
5	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
6	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
7	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
8	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
9	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
10	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
11	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
12	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
13	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
14	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
15	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
16	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
17	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
18	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
19	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
20	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
21	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
22	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
23	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
24	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
25	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
26	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
27	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
28	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
29	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
30	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
31	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
32	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
33	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
34	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
35	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
36	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
37	As	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00
38	As	0.00	0.05	0.00	0.00	0.00	0.00		