



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Walsh Construction Company II, LLC

SDG No.: Q1907

Contract: WALS01

Lab Code: CHEM

Case No.: Q1907

SAS No.: Q1907

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.17	0.2	86	40 - 160	CV	05/01/2025	12:05	LB135623
CRI01	Arsenic	23.3	20.0	116	40 - 160	P	05/01/2025	14:16	LB135636
	Barium	93.6	100	94	40 - 160	P	05/01/2025	14:16	LB135636
	Cadmium	6.36	6.0	106	40 - 160	P	05/01/2025	14:16	LB135636
	Chromium	10.6	10.0	106	40 - 160	P	05/01/2025	14:16	LB135636
	Lead	13.3	12.0	111	40 - 160	P	05/01/2025	14:16	LB135636
	Selenium	23.1	20.0	115	40 - 160	P	05/01/2025	14:16	LB135636
	Silver	10.6	10.0	106	40 - 160	P	05/01/2025	14:16	LB135636
CRI01	Arsenic	21.5	20.0	108	40 - 160	P	05/02/2025	15:12	LB135654
	Barium	80.2	100	80	40 - 160	P	05/02/2025	15:12	LB135654
	Cadmium	5.66	6.0	94	40 - 160	P	05/02/2025	15:12	LB135654
	Chromium	10.6	10.0	106	40 - 160	P	05/02/2025	15:12	LB135654
	Lead	12.1	12.0	101	40 - 160	P	05/02/2025	15:12	LB135654
	Selenium	23.5	20.0	117	40 - 160	P	05/02/2025	15:12	LB135654
	Silver	10.7	10.0	107	40 - 160	P	05/02/2025	15:12	LB135654