



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Walsh Construction Company II, LLC				SDG No.:	Q1907				
Contract:	WALS01	Lab Code:	CHEM		Case No.:	Q1907	SAS No.:	Q1907		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB112	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	11:49	LB135623	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q1907

Contract: WALS01

Lab Code: CHEM

Case No.: Q1907

SAS No.: Q1907

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB62	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	12:01	LB135623
CCB63	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	12:51	LB135623
CCB64	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	13:23	LB135623
CCB65	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	14:02	LB135623
CCB66	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	14:40	LB135623
CCB67	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	15:25	LB135623

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Walsh Construction Company II, LLC</u>		SDG No.: <u>Q1907</u>							
Contract: <u>WALS01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1907</u>	SAS No.: <u>Q1907</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	14:11	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	14:11	LB135636
	Cadmium	6.00	+/-6.00	U	6.00	P	05/01/2025	14:11	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	14:11	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	14:11	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	14:11	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	14:11	LB135636

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Walsh Construction Company II, LLC</u>		SDG No.: <u>Q1907</u>							
Contract: <u>WALS01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1907</u>	SAS No.: <u>Q1907</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	15:12	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	15:12	LB135636
	Cadmium	6.00	+/-6.00	U	6.00	P	05/01/2025	15:12	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	15:12	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	15:12	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	15:12	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	15:12	LB135636
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	16:04	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	16:04	LB135636
	Cadmium	0.60	+/-6.00	J	6.00	P	05/01/2025	16:04	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	16:04	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	16:04	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	16:04	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	16:04	LB135636
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	17:07	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	17:07	LB135636
	Cadmium	0.88	+/-6.00	J	6.00	P	05/01/2025	17:07	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	17:07	LB135636
	Lead	3.14	+/-12.0	J	12.0	P	05/01/2025	17:07	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	17:07	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	17:07	LB135636
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	17:53	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	17:53	LB135636
	Cadmium	0.92	+/-6.00	J	6.00	P	05/01/2025	17:53	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	17:53	LB135636
	Lead	2.63	+/-12.0	J	12.0	P	05/01/2025	17:53	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	17:53	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	17:53	LB135636
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	18:40	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	18:40	LB135636
	Cadmium	0.89	+/-6.00	J	6.00	P	05/01/2025	18:40	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	18:40	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	18:40	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	18:40	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	18:40	LB135636
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	19:27	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	19:27	LB135636
	Cadmium	6.00	+/-6.00	U	6.00	P	05/01/2025	19:27	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	19:27	LB135636

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC **SDG No.:** Q1907
Contract: WALS01 **Lab Code:** CHEM **Case No.:** Q1907 **SAS No.:** Q1907

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	19:27	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	19:27	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	19:27	LB135636
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	20:15	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	20:15	LB135636
	Cadmium	6.00	+/-6.00	U	6.00	P	05/01/2025	20:15	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	20:15	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	20:15	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	20:15	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	20:15	LB135636
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	20:28	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	20:28	LB135636
	Cadmium	6.00	+/-6.00	U	6.00	P	05/01/2025	20:28	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	20:28	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	20:28	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	20:28	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	20:28	LB135636

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q1907

Contract: WALS01

Lab Code: CHEM

Case No.: Q1907

SAS No.: Q1907

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/02/2025	15:08	LB135654
	Barium	100	+/-100	U	100	P	05/02/2025	15:08	LB135654
	Cadmium	6.00	+/-6.00	U	6.00	P	05/02/2025	15:08	LB135654
	Chromium	10.0	+/-10.0	U	10.0	P	05/02/2025	15:08	LB135654
	Lead	12.0	+/-12.0	U	12.0	P	05/02/2025	15:08	LB135654
	Selenium	20.0	+/-20.0	U	20.0	P	05/02/2025	15:08	LB135654
	Silver	10.0	+/-10.0	U	10.0	P	05/02/2025	15:08	LB135654

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Walsh Construction Company II, LLC</u>		SDG No.: <u>Q1907</u>							
Contract: <u>WALS01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1907</u>	SAS No.: <u>Q1907</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/02/2025	16:05	LB135654
	Barium	100	+/-100	U	100	P	05/02/2025	16:05	LB135654
	Cadmium	6.00	+/-6.00	U	6.00	P	05/02/2025	16:05	LB135654
	Chromium	10.0	+/-10.0	U	10.0	P	05/02/2025	16:05	LB135654
	Lead	12.0	+/-12.0	U	12.0	P	05/02/2025	16:05	LB135654
	Selenium	20.0	+/-20.0	U	20.0	P	05/02/2025	16:05	LB135654
	Silver	10.0	+/-10.0	U	10.0	P	05/02/2025	16:05	LB135654
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	05/02/2025	16:38	LB135654
	Barium	100	+/-100	U	100	P	05/02/2025	16:38	LB135654
	Cadmium	6.00	+/-6.00	U	6.00	P	05/02/2025	16:38	LB135654
	Chromium	10.0	+/-10.0	U	10.0	P	05/02/2025	16:38	LB135654
	Lead	12.0	+/-12.0	U	12.0	P	05/02/2025	16:38	LB135654
	Selenium	20.0	+/-20.0	U	20.0	P	05/02/2025	16:38	LB135654
	Silver	10.0	+/-10.0	U	10.0	P	05/02/2025	16:38	LB135654



- 3a -

Client:	Walsh Construction Company II, LLC			SDG No.:	Q1907
Contract:	WALS01	Lab Code:	CHEM	Case No.:	Q1907
				SAS No.:	Q1907

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------