

Continuing Calibration Report for SequenceID : FC050125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068729.D

Aliphatic C9-C12	5823575.000	60.000	3.343	6.644	97059.583	105282.003	7.810
Aliphatic C12-C16	3683316.000	40.000	6.645	10.046	92082.900	102322.575	10.007
Aliphatic C16-C21	5315964.000	60.000	10.047	13.414	88599.400	98816.539	10.340
Aliphatic C21-C28	6927625.000	80.000	13.415	17.079	86595.313	94185.337	8.059
Aliphatic C28-C40	11139642.000	120.000	17.080	22.084	92830.350	89882.835	-3.279
Aliphatic EPH	32890122.000	360.000	3.343	22.084	91361.450	96276.618	5.105