

Continuing Calibration Report for SequenceID : FC050125AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FC068739.D

|                   |              |         |        |        |           |            |        |
|-------------------|--------------|---------|--------|--------|-----------|------------|--------|
| Aliphatic C9-C12  | 5980952.000  | 60.000  | 3.343  | 6.644  | 99682.533 | 105282.003 | 5.319  |
| Aliphatic C12-C16 | 3725983.000  | 40.000  | 6.645  | 10.046 | 93149.575 | 102322.575 | 8.965  |
| Aliphatic C16-C21 | 5307738.000  | 60.000  | 10.047 | 13.414 | 88462.300 | 98816.539  | 10.478 |
| Aliphatic C21-C28 | 6952505.000  | 80.000  | 13.415 | 17.079 | 86906.313 | 94185.337  | 7.728  |
| Aliphatic C28-C40 | 11408908.000 | 120.000 | 17.080 | 22.084 | 95074.233 | 89882.835  | -5.776 |
| Aliphatic EPH     | 33376086.000 | 360.000 | 3.343  | 22.084 | 92711.350 | 96276.618  | 3.703  |