

Continuing Calibration Report for SequenceID : FE050125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE053602.D

Aliphatic C21-C28	10587831.000	80.000	13.578	17.245	132347.888	142174.626	6.912
Aliphatic C28-C40	15782705.000	120.000	17.246	22.133	131522.542	129049.185	-1.917
Aliphatic EPH	48513381.000	360.000	3.114	22.133	134759.392	137688.007	2.127
Aliphatic C9-C12	8249949.000	60.000	3.114	6.754	137499.150	138562.354	0.767
Aliphatic C12-C16	5616534.000	40.000	6.755	10.203	140413.350	141806.081	0.982
Aliphatic C16-C21	8276362.000	60.000	10.204	13.577	137939.367	145363.765	5.107