

SOLID EPH_NF LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Walsh Construction Company II, LLC
Lab Code: CHEM **Cas No:** Q1907 **SAS No :** Q1907 **SDG No:** Q1907
Sample No : PB167819BS **Datafile:** FC068731.D
Client ID : PB167819BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aliphatic C28-C40	30.0	30.3	101		(40-140)
Aliphatic C9-C28	99.9	85.9	86		(40-140)