

SOLID EPH_NF LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Walsh Construction Company II, LLC
Lab Code: CHEM **Cas No:** Q1907 **SAS No :** Q1907 **SDG No:** Q1907
Sample No : PB167819BSD **Datafile:** FC068732.D
Client ID : PB167819BSD

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C28-C40	30.0	31.1	103		2.7 (40-140)	50
Aliphatic C9-C28	99.9	88.6	89		3.2 (40-140)	50